

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3740
 Lab File ID: VY023829.D Date Analyzed: 12/01/2025
 Instrument ID: MSVOA_Y Time Analyzed: 07:55
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	375940	13.352				
UPPER LIMIT	751880	13.852				
LOWER LIMIT	187970	12.852				
EPA SAMPLE NO.						
WC1	445014	13.35				
VY1201SBL01	399677	13.35				
VY1201SBS01	360194	13.35				
VY1201SBSD01	353062	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.