

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, Inc

SDG No.: Q3741

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV141	Mercury	4.06	4.0	101	90 - 110	CV	12/05/2025	14:12	LB138128

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, Inc

SDG No.: Q3741

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV71	Mercury	4.93	5.0	99	90 - 110	CV	12/05/2025	14:16	LB138128
CCV72	Mercury	4.73	5.0	95	90 - 110	CV	12/05/2025	14:54	LB138128
CCV73	Mercury	4.85	5.0	97	90 - 110	CV	12/05/2025	15:23	LB138128

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, Inc

SDG No.: Q3741

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV147	Mercury	3.88	4.0	97	90 - 110	CV	12/09/2025	11:28	LB138165

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, Inc

SDG No.: Q3741

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV84	Mercury	5.20	5.0	104	90 - 110	CV	12/09/2025	11:33	LB138165
CCV85	Mercury	5.31	5.0	106	90 - 110	CV	12/09/2025	12:03	LB138165
CCV86	Mercury	5.28	5.0	106	90 - 110	CV	12/09/2025	12:40	LB138165
CCV87	Mercury	5.12	5.0	102	90 - 110	CV	12/09/2025	13:15	LB138165

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, Inc

SDG No.: Q3741

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7850	8000	98	90 - 110	P	12/02/2025	11:29	LB138085
	Antimony	3960	4000	99	90 - 110	P	12/02/2025	11:29	LB138085
	Arsenic	3880	4000	97	90 - 110	P	12/02/2025	11:29	LB138085
	Barium	8280	8000	104	90 - 110	P	12/02/2025	11:29	LB138085
	Beryllium	205	200	102	90 - 110	P	12/02/2025	11:29	LB138085
	Cadmium	1920	2000	96	90 - 110	P	12/02/2025	11:29	LB138085
	Calcium	19900	20000	100	90 - 110	P	12/02/2025	11:29	LB138085
	Chromium	781	800	98	90 - 110	P	12/02/2025	11:29	LB138085
	Cobalt	1930	2000	97	90 - 110	P	12/02/2025	11:29	LB138085
	Copper	999	1000	100	90 - 110	P	12/02/2025	11:29	LB138085
	Iron	3940	4000	99	90 - 110	P	12/02/2025	11:29	LB138085
	Lead	3790	4000	95	90 - 110	P	12/02/2025	11:29	LB138085
	Magnesium	19600	20000	98	90 - 110	P	12/02/2025	11:29	LB138085
	Manganese	2050	2000	102	90 - 110	P	12/02/2025	11:29	LB138085
	Nickel	1940	2000	97	90 - 110	P	12/02/2025	11:29	LB138085
	Potassium	19600	20000	98	90 - 110	P	12/02/2025	11:29	LB138085
	Selenium	3970	4000	99	90 - 110	P	12/02/2025	11:29	LB138085
	Silver	1080	1000	108	90 - 110	P	12/02/2025	11:29	LB138085
	Sodium	20300	20000	101	90 - 110	P	12/02/2025	11:29	LB138085
	Thallium	4060	4000	102	90 - 110	P	12/02/2025	11:29	LB138085
	Vanadium	1970	2000	99	90 - 110	P	12/02/2025	11:29	LB138085
	Zinc	2020	2000	101	90 - 110	P	12/02/2025	11:29	LB138085

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, Inc

SDG No.: Q3741

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	112	100	112	80 - 120	P	12/02/2025	11:34	LB138085
	Antimony	50.1	50.0	100	80 - 120	P	12/02/2025	11:34	LB138085
	Arsenic	22.6	20.0	113	80 - 120	P	12/02/2025	11:34	LB138085
	Barium	104	100	104	80 - 120	P	12/02/2025	11:34	LB138085
	Beryllium	6.61	6.0	110	80 - 120	P	12/02/2025	11:34	LB138085
	Cadmium	6.22	6.0	104	80 - 120	P	12/02/2025	11:34	LB138085
	Calcium	2050	2000	102	80 - 120	P	12/02/2025	11:34	LB138085
	Chromium	9.91	10.0	99	80 - 120	P	12/02/2025	11:34	LB138085
	Cobalt	29.9	30.0	100	80 - 120	P	12/02/2025	11:34	LB138085
	Copper	21.9	20.0	110	80 - 120	P	12/02/2025	11:34	LB138085
	Iron	103	100	103	80 - 120	P	12/02/2025	11:34	LB138085
	Lead	11.3	12.0	94	80 - 120	P	12/02/2025	11:34	LB138085
	Magnesium	2170	2000	109	80 - 120	P	12/02/2025	11:34	LB138085
	Manganese	21.9	20.0	110	80 - 120	P	12/02/2025	11:34	LB138085
	Nickel	40.7	40.0	102	80 - 120	P	12/02/2025	11:34	LB138085
	Potassium	1900	2000	95	80 - 120	P	12/02/2025	11:34	LB138085
	Selenium	18.8	20.0	94	80 - 120	P	12/02/2025	11:34	LB138085
	Silver	10.3	10.0	103	80 - 120	P	12/02/2025	11:34	LB138085
	Sodium	2010	2000	100	80 - 120	P	12/02/2025	11:34	LB138085
	Thallium	44.2	40.0	111	80 - 120	P	12/02/2025	11:34	LB138085
	Vanadium	40.2	40.0	100	80 - 120	P	12/02/2025	11:34	LB138085
	Zinc	40.6	40.0	102	80 - 120	P	12/02/2025	11:34	LB138085

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, Inc

SDG No.: Q3741

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	10100	10000	101	90 - 110	P	12/02/2025	12:10	LB138085
	Antimony	4970	5000	99	90 - 110	P	12/02/2025	12:10	LB138085
	Arsenic	4950	5000	99	90 - 110	P	12/02/2025	12:10	LB138085
	Barium	10200	10000	102	90 - 110	P	12/02/2025	12:10	LB138085
	Beryllium	254	250	102	90 - 110	P	12/02/2025	12:10	LB138085
	Cadmium	2470	2500	99	90 - 110	P	12/02/2025	12:10	LB138085
	Calcium	25300	25000	101	90 - 110	P	12/02/2025	12:10	LB138085
	Chromium	1000	1000	100	90 - 110	P	12/02/2025	12:10	LB138085
	Cobalt	2470	2500	99	90 - 110	P	12/02/2025	12:10	LB138085
	Copper	1250	1250	100	90 - 110	P	12/02/2025	12:10	LB138085
	Iron	4950	5000	99	90 - 110	P	12/02/2025	12:10	LB138085
	Lead	4950	5000	99	90 - 110	P	12/02/2025	12:10	LB138085
	Magnesium	25100	25000	101	90 - 110	P	12/02/2025	12:10	LB138085
	Manganese	2540	2500	101	90 - 110	P	12/02/2025	12:10	LB138085
	Nickel	2480	2500	99	90 - 110	P	12/02/2025	12:10	LB138085
	Potassium	24800	25000	99	90 - 110	P	12/02/2025	12:10	LB138085
	Selenium	4980	5000	100	90 - 110	P	12/02/2025	12:10	LB138085
	Silver	1250	1250	100	90 - 110	P	12/02/2025	12:10	LB138085
	Sodium	25400	25000	101	90 - 110	P	12/02/2025	12:10	LB138085
	Thallium	4980	5000	100	90 - 110	P	12/02/2025	12:10	LB138085
	Vanadium	2520	2500	101	90 - 110	P	12/02/2025	12:10	LB138085
	Zinc	2520	2500	101	90 - 110	P	12/02/2025	12:10	LB138085
CCV02	Aluminum	10100	10000	100	90 - 110	P	12/02/2025	13:13	LB138085
	Antimony	5120	5000	102	90 - 110	P	12/02/2025	13:13	LB138085
	Arsenic	5050	5000	101	90 - 110	P	12/02/2025	13:13	LB138085
	Barium	9890	10000	99	90 - 110	P	12/02/2025	13:13	LB138085
	Beryllium	249	250	100	90 - 110	P	12/02/2025	13:13	LB138085
	Cadmium	2500	2500	100	90 - 110	P	12/02/2025	13:13	LB138085
	Calcium	25000	25000	100	90 - 110	P	12/02/2025	13:13	LB138085
	Chromium	1020	1000	102	90 - 110	P	12/02/2025	13:13	LB138085
	Cobalt	2500	2500	100	90 - 110	P	12/02/2025	13:13	LB138085
	Copper	1280	1250	103	90 - 110	P	12/02/2025	13:13	LB138085
	Iron	5090	5000	102	90 - 110	P	12/02/2025	13:13	LB138085
	Lead	5020	5000	100	90 - 110	P	12/02/2025	13:13	LB138085

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, Inc

SDG No.: Q3741

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24800	25000	99	90 - 110	P	12/02/2025	13:13	LB138085
	Manganese	2480	2500	99	90 - 110	P	12/02/2025	13:13	LB138085
	Nickel	2520	2500	101	90 - 110	P	12/02/2025	13:13	LB138085
	Potassium	25900	25000	104	90 - 110	P	12/02/2025	13:13	LB138085
	Selenium	5180	5000	104	90 - 110	P	12/02/2025	13:13	LB138085
	Silver	1280	1250	103	90 - 110	P	12/02/2025	13:13	LB138085
	Sodium	26300	25000	105	90 - 110	P	12/02/2025	13:13	LB138085
	Thallium	4690	5000	94	90 - 110	P	12/02/2025	13:13	LB138085
	Vanadium	2490	2500	100	90 - 110	P	12/02/2025	13:13	LB138085
	Zinc	2590	2500	104	90 - 110	P	12/02/2025	13:13	LB138085
CCV03	Aluminum	10100	10000	101	90 - 110	P	12/02/2025	14:11	LB138085
	Antimony	5100	5000	102	90 - 110	P	12/02/2025	14:11	LB138085
	Arsenic	5040	5000	101	90 - 110	P	12/02/2025	14:11	LB138085
	Barium	10100	10000	102	90 - 110	P	12/02/2025	14:11	LB138085
	Beryllium	247	250	99	90 - 110	P	12/02/2025	14:11	LB138085
	Cadmium	2470	2500	99	90 - 110	P	12/02/2025	14:11	LB138085
	Calcium	25000	25000	100	90 - 110	P	12/02/2025	14:11	LB138085
	Chromium	987	1000	99	90 - 110	P	12/02/2025	14:11	LB138085
	Cobalt	2470	2500	99	90 - 110	P	12/02/2025	14:11	LB138085
	Copper	1280	1250	102	90 - 110	P	12/02/2025	14:11	LB138085
	Iron	5060	5000	101	90 - 110	P	12/02/2025	14:11	LB138085
	Lead	4970	5000	100	90 - 110	P	12/02/2025	14:11	LB138085
	Magnesium	24800	25000	99	90 - 110	P	12/02/2025	14:11	LB138085
	Manganese	2500	2500	100	90 - 110	P	12/02/2025	14:11	LB138085
	Nickel	2500	2500	100	90 - 110	P	12/02/2025	14:11	LB138085
	Potassium	25200	25000	101	90 - 110	P	12/02/2025	14:11	LB138085
	Selenium	5230	5000	105	90 - 110	P	12/02/2025	14:11	LB138085
	Silver	1260	1250	100	90 - 110	P	12/02/2025	14:11	LB138085
	Sodium	25900	25000	104	90 - 110	P	12/02/2025	14:11	LB138085
	Thallium	4900	5000	98	90 - 110	P	12/02/2025	14:11	LB138085
	Vanadium	2500	2500	100	90 - 110	P	12/02/2025	14:11	LB138085
	Zinc	2530	2500	101	90 - 110	P	12/02/2025	14:11	LB138085
CCV04	Aluminum	10200	10000	102	90 - 110	P	12/02/2025	15:15	LB138085
	Antimony	5200	5000	104	90 - 110	P	12/02/2025	15:15	LB138085

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, In

SDG No.: Q3741

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5140	5000	103	90 - 110	P	12/02/2025	15:15	LB138085
	Barium	10500	10000	105	90 - 110	P	12/02/2025	15:15	LB138085
	Beryllium	250	250	100	90 - 110	P	12/02/2025	15:15	LB138085
	Cadmium	2500	2500	100	90 - 110	P	12/02/2025	15:15	LB138085
	Calcium	25500	25000	102	90 - 110	P	12/02/2025	15:15	LB138085
	Chromium	990	1000	99	90 - 110	P	12/02/2025	15:15	LB138085
	Cobalt	2520	2500	101	90 - 110	P	12/02/2025	15:15	LB138085
	Copper	1300	1250	104	90 - 110	P	12/02/2025	15:15	LB138085
	Iron	5230	5000	105	90 - 110	P	12/02/2025	15:15	LB138085
	Lead	5050	5000	101	90 - 110	P	12/02/2025	15:15	LB138085
	Magnesium	24900	25000	100	90 - 110	P	12/02/2025	15:15	LB138085
	Manganese	2590	2500	104	90 - 110	P	12/02/2025	15:15	LB138085
	Nickel	2550	2500	102	90 - 110	P	12/02/2025	15:15	LB138085
	Potassium	26000	25000	104	90 - 110	P	12/02/2025	15:15	LB138085
	Selenium	5330	5000	107	90 - 110	P	12/02/2025	15:15	LB138085
	Silver	1280	1250	102	90 - 110	P	12/02/2025	15:15	LB138085
	Sodium	26900	25000	108	90 - 110	P	12/02/2025	15:15	LB138085
	Thallium	5220	5000	104	90 - 110	P	12/02/2025	15:15	LB138085
	Vanadium	2520	2500	101	90 - 110	P	12/02/2025	15:15	LB138085
	Zinc	2570	2500	103	90 - 110	P	12/02/2025	15:15	LB138085
CCV05	Aluminum	10200	10000	102	90 - 110	P	12/02/2025	16:05	LB138085
	Antimony	5110	5000	102	90 - 110	P	12/02/2025	16:05	LB138085
	Arsenic	5040	5000	101	90 - 110	P	12/02/2025	16:05	LB138085
	Barium	10600	10000	106	90 - 110	P	12/02/2025	16:05	LB138085
	Beryllium	249	250	100	90 - 110	P	12/02/2025	16:05	LB138085
	Cadmium	2460	2500	98	90 - 110	P	12/02/2025	16:05	LB138085
	Calcium	25600	25000	102	90 - 110	P	12/02/2025	16:05	LB138085
	Chromium	968	1000	97	90 - 110	P	12/02/2025	16:05	LB138085
	Cobalt	2480	2500	99	90 - 110	P	12/02/2025	16:05	LB138085
	Copper	1280	1250	102	90 - 110	P	12/02/2025	16:05	LB138085
	Iron	5170	5000	103	90 - 110	P	12/02/2025	16:05	LB138085
	Lead	4950	5000	99	90 - 110	P	12/02/2025	16:05	LB138085
	Magnesium	24900	25000	100	90 - 110	P	12/02/2025	16:05	LB138085
	Manganese	2640	2500	106	90 - 110	P	12/02/2025	16:05	LB138085

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, Inc

SDG No.: Q3741

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2520	2500	101	90 - 110	P	12/02/2025	16:05	LB138085
	Potassium	25600	25000	102	90 - 110	P	12/02/2025	16:05	LB138085
	Selenium	5220	5000	104	90 - 110	P	12/02/2025	16:05	LB138085
	Silver	1250	1250	100	90 - 110	P	12/02/2025	16:05	LB138085
	Sodium	27100	25000	108	90 - 110	P	12/02/2025	16:05	LB138085
	Thallium	5160	5000	103	90 - 110	P	12/02/2025	16:05	LB138085
	Vanadium	2530	2500	101	90 - 110	P	12/02/2025	16:05	LB138085
	Zinc	2510	2500	100	90 - 110	P	12/02/2025	16:05	LB138085
CCV06	Aluminum	10100	10000	101	90 - 110	P	12/02/2025	16:55	LB138085
	Antimony	5130	5000	102	90 - 110	P	12/02/2025	16:55	LB138085
	Arsenic	5050	5000	101	90 - 110	P	12/02/2025	16:55	LB138085
	Barium	10400	10000	104	90 - 110	P	12/02/2025	16:55	LB138085
	Beryllium	251	250	100	90 - 110	P	12/02/2025	16:55	LB138085
	Cadmium	2470	2500	99	90 - 110	P	12/02/2025	16:55	LB138085
	Calcium	25500	25000	102	90 - 110	P	12/02/2025	16:55	LB138085
	Chromium	983	1000	98	90 - 110	P	12/02/2025	16:55	LB138085
	Cobalt	2500	2500	100	90 - 110	P	12/02/2025	16:55	LB138085
	Copper	1290	1250	103	90 - 110	P	12/02/2025	16:55	LB138085
	Iron	5170	5000	103	90 - 110	P	12/02/2025	16:55	LB138085
	Lead	5000	5000	100	90 - 110	P	12/02/2025	16:55	LB138085
	Magnesium	24800	25000	99	90 - 110	P	12/02/2025	16:55	LB138085
	Manganese	2610	2500	104	90 - 110	P	12/02/2025	16:55	LB138085
	Nickel	2540	2500	102	90 - 110	P	12/02/2025	16:55	LB138085
	Potassium	25800	25000	103	90 - 110	P	12/02/2025	16:55	LB138085
	Selenium	5220	5000	104	90 - 110	P	12/02/2025	16:55	LB138085
	Silver	1260	1250	101	90 - 110	P	12/02/2025	16:55	LB138085
	Sodium	26700	25000	107	90 - 110	P	12/02/2025	16:55	LB138085
	Thallium	4600	5000	92	90 - 110	P	12/02/2025	16:55	LB138085
	Vanadium	2510	2500	101	90 - 110	P	12/02/2025	16:55	LB138085
	Zinc	2550	2500	102	90 - 110	P	12/02/2025	16:55	LB138085