

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Alliance

Contract: CORE02

ProjectID: Building 50 Probe Investigation

Lab Code: ACE

SDG No.: Q3741

Calibration Sequence : FG120125		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
1000	138381193	138381	FG017031.D
500	57796211	115592	FG017032.D
200	23582132	117911	FG017033.D
100	11130849	111308	FG017034.D
50	5220615	104412	FG017035.D
AVG RF : 117521		% RSD : 10.843	AVG RT : 15.1388