

DIESEL RANGE ORGANICS CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Alliane Contract: CORE02
ProjectID: Building 50 Probe Investigation
Lab Code: ACE SDG No.: Q3741
DataFile: FG017092.D Analyst Name: YP\AJ Analyst Date: 12-05-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
500	68622687	137245	117521	16.783