284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02

Lab Code: ACE **SDG NO.:** Q3741

Continuing Calib Date: 12/04/2025 **Initial Calibration Date(s):** 11/21/2025 11/21/2025

Continuing Calib Time: 03:10 **Initial Calibration Time(s):** 10:44 19:39

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1248-1	(1)	5.77	5.78	5.68	5.88	0.01
Aroclor-1248-2	(2)	6.04	6.05	5.95	6.15	0.01
Aroclor-1248-3	(3)	6.24	6.25	6.15	6.35	0.01
Aroclor-1248-4	(4)	6.64	6.65	6.55	6.75	0.01
Aroclor-1248-5	(5)	6.68	6.69	6.59	6.79	0.01
Tetrachloro-m-xylene		4.62	4.63	4.53	4.73	0.01
Decachlorobiphenyl		10.37	10.38	10.28	10.48	0.01

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02

Lab Code: ACE **SDG NO.:** Q3741

Continuing Calib Date: 12/04/2025 **Initial Calibration Date(s):** 11/21/2025 11/21/2025

Continuing Calib Time: 03:10 **Initial Calibration Time(s):** 10:44 19:39

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1248-1	(1)	4.85	4.86	4.76	4.96	0.01
Aroclor-1248-2	(2)	5.08	5.09	4.99	5.19	0.01
Aroclor-1248-3	(3)	5.13	5.13	5.03	5.23	0.00
Aroclor-1248-4	(4)	5.30	5.30	5.20	5.40	0.00
Aroclor-1248-5	(5)	5.69	5.69	5.59	5.79	0.00
Tetrachloro-m-xylene		3.77	3.78	3.68	3.88	0.01
Decachlorobiphenyl		8.76	8.77	8.67	8.87	0.01