

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q3741
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 12/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076744.D **Time Analyzed:** 10:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.770	5.667	5.867	510.750	500.000	2.2
Aroclor-1016-2	5.791	5.687	5.887	528.080	500.000	5.6
Aroclor-1016-3	5.854	5.750	5.950	523.240	500.000	4.6
Aroclor-1016-4	5.951	5.847	6.047	524.670	500.000	4.9
Aroclor-1016-5	6.243	6.140	6.340	522.370	500.000	4.5
Aroclor-1260-1	7.360	7.256	7.456	518.740	500.000	3.7
Aroclor-1260-2	7.613	7.509	7.709	506.520	500.000	1.3
Aroclor-1260-3	7.971	7.867	8.067	509.450	500.000	1.9
Aroclor-1260-4	8.197	8.093	8.293	504.520	500.000	0.9
Aroclor-1260-5	8.521	8.416	8.616	503.010	500.000	0.6
Decachlorobiphenyl	10.370	10.265	10.465	51.220	50.000	2.4
Tetrachloro-m-xylene	4.618	4.514	4.714	53.610	50.000	7.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q3741
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 12/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076744.D **Time Analyzed:** 10:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.851	4.750	4.950	567.100	500.000	13.4
Aroclor-1016-2	4.869	4.767	4.967	589.400	500.000	17.9
Aroclor-1016-3	5.044	4.942	5.142	581.550	500.000	16.3
Aroclor-1016-4	5.086	4.984	5.184	541.860	500.000	8.4
Aroclor-1016-5	5.298	5.197	5.397	629.900	500.000	26.0
Aroclor-1260-1	6.326	6.224	6.424	582.440	500.000	16.5
Aroclor-1260-2	6.513	6.412	6.612	587.660	500.000	17.5
Aroclor-1260-3	6.665	6.564	6.764	565.210	500.000	13.0
Aroclor-1260-4	7.134	7.033	7.233	570.270	500.000	14.1
Aroclor-1260-5	7.374	7.273	7.473	563.180	500.000	12.6
Decachlorobiphenyl	8.766	8.666	8.866	54.980	50.000	10.0
Tetrachloro-m-xylene	3.773	3.671	3.871	58.990	50.000	18.0