

## LAB CHRONICLE

|                                                                  |                                                 |
|------------------------------------------------------------------|-------------------------------------------------|
| <b>OrderID:</b> Q3741                                            | <b>OrderDate:</b> 11/26/2025 4:57:55 PM         |
| <b>Client:</b> Core Environmental Consultants and Services, Inc. | <b>Project:</b> Building 50 Probe Investigation |
| <b>Contact:</b> Roland Scardino                                  | <b>Location:</b> --Select--,VOA Lab             |

| LabID    | ClientID        | Matrix | Test | Method | Sample Date | Prep Date | Anal Date | Received |
|----------|-----------------|--------|------|--------|-------------|-----------|-----------|----------|
| Q3741-01 | B50-SP3-111925  | SOIL   | PCB  | 8082A  | 11/19/25    | 12/03/25  | 12/03/25  | 11/26/25 |
| Q3741-02 | B50-SP13-112025 | SOIL   | PCB  | 8082A  | 11/20/25    | 12/03/25  | 12/03/25  | 11/26/25 |
| Q3741-03 | B50-SP14-112025 | SOIL   | PCB  | 8082A  | 11/20/25    | 12/03/25  | 12/03/25  | 11/26/25 |
| Q3741-04 | B50-SP9-112425  | SOIL   | PCB  | 8082A  | 11/24/25    | 12/03/25  | 12/03/25  | 11/26/25 |
| Q3741-05 | B50-SP11-112525 | SOIL   | PCB  | 8082A  | 11/25/25    | 12/03/25  | 12/03/25  | 11/26/25 |
| Q3741-06 | B50-SP4-112525  | SOIL   | PCB  | 8082A  | 11/25/25    | 12/03/25  | 12/03/25  | 11/26/25 |

**Hit Summary Sheet**  
**SW-846**

**SDG No.:** Q3741

**Order ID:** Q3741

**Client:** Core Environmental Consultants and Services, In

**Project ID:** Building 50 Probe Investigation

| Sample ID          | Client ID              | Matrix | Parameter                   | Concentration | C    | MDL | RDL  | Units |
|--------------------|------------------------|--------|-----------------------------|---------------|------|-----|------|-------|
| <b>Client ID :</b> | <b>B50-SP11-112525</b> |        |                             |               |      |     |      |       |
| Q3741-05           | B50-SP11-112525        | SOIL   | Aroclor-1254                | 20.5          | 3.80 |     | 20.1 | ug/kg |
|                    |                        |        | <b>Total Concentration:</b> | <b>20.500</b> |      |     |      |       |



# SAMPLE DATA

## Report of Analysis

|                    |                                                   |                 |          |
|--------------------|---------------------------------------------------|-----------------|----------|
| Client:            | Core Environmental Consultants and Services, Inc. | Date Collected: | 11/19/25 |
| Project:           | Building 50 Probe Investigation                   | Date Received:  | 11/26/25 |
| Client Sample ID:  | B50-SP3-111925                                    | SDG No.:        | Q3741    |
| Lab Sample ID:     | Q3741-01                                          | Matrix:         | SOIL     |
| Analytical Method: | 8082A                                             | % Solid:        | 91.6     |
| Sample Wt/Vol:     | 30.06 g                                           | Final Vol:      | 10000 uL |
| Prep Method:       | SW3541B                                           | Test:           | PCB      |
|                    | Prep Date:                                        |                 | 12/03/25 |

| CAS Number        | Parameter            | Conc. | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana.      | BatchID  |
|-------------------|----------------------|-------|------|----|----------|------------|---------|----------------|----------|
| <b>TARGETS</b>    |                      |       |      |    |          |            |         |                |          |
| 12674-11-2        | Aroclor-1016         | 4.30  | U    | 1  | 4.30     | 18.5       | ug/kg   | 12/03/25 12:50 | PB170795 |
| 11104-28-2        | Aroclor-1221         | 4.40  | U    | 1  | 4.40     | 18.5       | ug/kg   | 12/03/25 12:50 | PB170795 |
| 11141-16-5        | Aroclor-1232         | 4.10  | U    | 1  | 4.10     | 18.5       | ug/kg   | 12/03/25 12:50 | PB170795 |
| 53469-21-9        | Aroclor-1242         | 4.40  | U    | 1  | 4.40     | 18.5       | ug/kg   | 12/03/25 12:50 | PB170795 |
| 12672-29-6        | Aroclor-1248         | 6.40  | U    | 1  | 6.40     | 18.5       | ug/kg   | 12/03/25 12:50 | PB170795 |
| 11097-69-1        | Aroclor-1254         | 3.50  | U    | 1  | 3.50     | 18.5       | ug/kg   | 12/03/25 12:50 | PB170795 |
| 37324-23-5        | Aroclor-1262         | 5.50  | U    | 1  | 5.50     | 18.5       | ug/kg   | 12/03/25 12:50 | PB170795 |
| 11100-14-4        | Aroclor-1268         | 3.90  | U    | 1  | 3.90     | 18.5       | ug/kg   | 12/03/25 12:50 | PB170795 |
| 11096-82-5        | Aroclor-1260         | 3.50  | U    | 1  | 3.50     | 18.5       | ug/kg   | 12/03/25 12:50 | PB170795 |
| <b>SURROGATES</b> |                      |       |      |    |          |            |         |                |          |
| 877-09-8          | Tetrachloro-m-xylene | 20.7  |      |    | 21 - 165 | 103%       | SPK: 20 |                |          |
| 2051-24-3         | Decachlorobiphenyl   | 22.6  |      |    | 10 - 170 | 113%       | SPK: 20 |                |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                                                   |                 |          |
|--------------------|---------------------------------------------------|-----------------|----------|
| Client:            | Core Environmental Consultants and Services, Inc. | Date Collected: | 11/20/25 |
| Project:           | Building 50 Probe Investigation                   | Date Received:  | 11/26/25 |
| Client Sample ID:  | B50-SP13-112025                                   | SDG No.:        | Q3741    |
| Lab Sample ID:     | Q3741-02                                          | Matrix:         | SOIL     |
| Analytical Method: | 8082A                                             | % Solid:        | 86.4     |
| Sample Wt/Vol:     | 30.04 g                                           | Final Vol:      | 10000 uL |
| Prep Method:       | SW3541B                                           | Test:           | PCB      |
|                    | Prep Date:                                        |                 | 12/03/25 |

| CAS Number        | Parameter            | Conc. | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana.      | BatchID  |
|-------------------|----------------------|-------|------|----|----------|------------|---------|----------------|----------|
| <b>TARGETS</b>    |                      |       |      |    |          |            |         |                |          |
| 12674-11-2        | Aroclor-1016         | 4.60  | U    | 1  | 4.60     | 19.6       | ug/kg   | 12/03/25 13:08 | PB170795 |
| 11104-28-2        | Aroclor-1221         | 4.70  | U    | 1  | 4.70     | 19.6       | ug/kg   | 12/03/25 13:08 | PB170795 |
| 11141-16-5        | Aroclor-1232         | 4.30  | U    | 1  | 4.30     | 19.6       | ug/kg   | 12/03/25 13:08 | PB170795 |
| 53469-21-9        | Aroclor-1242         | 4.60  | U    | 1  | 4.60     | 19.6       | ug/kg   | 12/03/25 13:08 | PB170795 |
| 12672-29-6        | Aroclor-1248         | 6.80  | U    | 1  | 6.80     | 19.6       | ug/kg   | 12/03/25 13:08 | PB170795 |
| 11097-69-1        | Aroclor-1254         | 3.70  | U    | 1  | 3.70     | 19.6       | ug/kg   | 12/03/25 13:08 | PB170795 |
| 37324-23-5        | Aroclor-1262         | 5.80  | U    | 1  | 5.80     | 19.6       | ug/kg   | 12/03/25 13:08 | PB170795 |
| 11100-14-4        | Aroclor-1268         | 4.20  | U    | 1  | 4.20     | 19.6       | ug/kg   | 12/03/25 13:08 | PB170795 |
| 11096-82-5        | Aroclor-1260         | 3.70  | U    | 1  | 3.70     | 19.6       | ug/kg   | 12/03/25 13:08 | PB170795 |
| <b>SURROGATES</b> |                      |       |      |    |          |            |         |                |          |
| 877-09-8          | Tetrachloro-m-xylene | 18.1  |      |    | 21 - 165 | 91%        | SPK: 20 |                |          |
| 2051-24-3         | Decachlorobiphenyl   | 17.3  |      |    | 10 - 170 | 86%        | SPK: 20 |                |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                                                   |                 |          |
|--------------------|---------------------------------------------------|-----------------|----------|
| Client:            | Core Environmental Consultants and Services, Inc. | Date Collected: | 11/20/25 |
| Project:           | Building 50 Probe Investigation                   | Date Received:  | 11/26/25 |
| Client Sample ID:  | B50-SP14-112025                                   | SDG No.:        | Q3741    |
| Lab Sample ID:     | Q3741-03                                          | Matrix:         | SOIL     |
| Analytical Method: | 8082A                                             | % Solid:        | 87.3     |
| Sample Wt/Vol:     | 30.08 g                                           | Final Vol:      | 10000 uL |
| Prep Method:       | SW3541B                                           | Test:           | PCB      |
|                    | Prep Date:                                        |                 | 12/03/25 |

| CAS Number        | Parameter            | Conc. | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana.      | BatchID  |
|-------------------|----------------------|-------|------|----|----------|------------|---------|----------------|----------|
| <b>TARGETS</b>    |                      |       |      |    |          |            |         |                |          |
| 12674-11-2        | Aroclor-1016         | 4.50  | U    | 1  | 4.50     | 19.4       | ug/kg   | 12/03/25 13:27 | PB170795 |
| 11104-28-2        | Aroclor-1221         | 4.60  | U    | 1  | 4.60     | 19.4       | ug/kg   | 12/03/25 13:27 | PB170795 |
| 11141-16-5        | Aroclor-1232         | 4.20  | U    | 1  | 4.20     | 19.4       | ug/kg   | 12/03/25 13:27 | PB170795 |
| 53469-21-9        | Aroclor-1242         | 4.60  | U    | 1  | 4.60     | 19.4       | ug/kg   | 12/03/25 13:27 | PB170795 |
| 12672-29-6        | Aroclor-1248         | 6.80  | U    | 1  | 6.80     | 19.4       | ug/kg   | 12/03/25 13:27 | PB170795 |
| 11097-69-1        | Aroclor-1254         | 3.70  | U    | 1  | 3.70     | 19.4       | ug/kg   | 12/03/25 13:27 | PB170795 |
| 37324-23-5        | Aroclor-1262         | 5.70  | U    | 1  | 5.70     | 19.4       | ug/kg   | 12/03/25 13:27 | PB170795 |
| 11100-14-4        | Aroclor-1268         | 4.10  | U    | 1  | 4.10     | 19.4       | ug/kg   | 12/03/25 13:27 | PB170795 |
| 11096-82-5        | Aroclor-1260         | 3.70  | U    | 1  | 3.70     | 19.4       | ug/kg   | 12/03/25 13:27 | PB170795 |
| <b>SURROGATES</b> |                      |       |      |    |          |            |         |                |          |
| 877-09-8          | Tetrachloro-m-xylene | 20.2  |      |    | 21 - 165 | 101%       | SPK: 20 |                |          |
| 2051-24-3         | Decachlorobiphenyl   | 18.4  |      |    | 10 - 170 | 92%        | SPK: 20 |                |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                                                   |                 |          |
|--------------------|---------------------------------------------------|-----------------|----------|
| Client:            | Core Environmental Consultants and Services, Inc. | Date Collected: | 11/24/25 |
| Project:           | Building 50 Probe Investigation                   | Date Received:  | 11/26/25 |
| Client Sample ID:  | B50-SP9-112425                                    | SDG No.:        | Q3741    |
| Lab Sample ID:     | Q3741-04                                          | Matrix:         | SOIL     |
| Analytical Method: | 8082A                                             | % Solid:        | 85.2     |
| Sample Wt/Vol:     | 30.03 g                                           | Final Vol:      | 10000 uL |
| Prep Method:       | SW3541B                                           | Test:           | PCB      |
|                    | Prep Date:                                        |                 | 12/03/25 |

| CAS Number        | Parameter            | Conc. | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana.      | BatchID  |
|-------------------|----------------------|-------|------|----|----------|------------|---------|----------------|----------|
| <b>TARGETS</b>    |                      |       |      |    |          |            |         |                |          |
| 12674-11-2        | Aroclor-1016         | 4.60  | U    | 1  | 4.60     | 19.9       | ug/kg   | 12/03/25 13:45 | PB170795 |
| 11104-28-2        | Aroclor-1221         | 4.70  | U    | 1  | 4.70     | 19.9       | ug/kg   | 12/03/25 13:45 | PB170795 |
| 11141-16-5        | Aroclor-1232         | 4.40  | U    | 1  | 4.40     | 19.9       | ug/kg   | 12/03/25 13:45 | PB170795 |
| 53469-21-9        | Aroclor-1242         | 4.70  | U    | 1  | 4.70     | 19.9       | ug/kg   | 12/03/25 13:45 | PB170795 |
| 12672-29-6        | Aroclor-1248         | 6.90  | U    | 1  | 6.90     | 19.9       | ug/kg   | 12/03/25 13:45 | PB170795 |
| 11097-69-1        | Aroclor-1254         | 3.80  | U    | 1  | 3.80     | 19.9       | ug/kg   | 12/03/25 13:45 | PB170795 |
| 37324-23-5        | Aroclor-1262         | 5.90  | U    | 1  | 5.90     | 19.9       | ug/kg   | 12/03/25 13:45 | PB170795 |
| 11100-14-4        | Aroclor-1268         | 4.20  | U    | 1  | 4.20     | 19.9       | ug/kg   | 12/03/25 13:45 | PB170795 |
| 11096-82-5        | Aroclor-1260         | 3.80  | U    | 1  | 3.80     | 19.9       | ug/kg   | 12/03/25 13:45 | PB170795 |
| <b>SURROGATES</b> |                      |       |      |    |          |            |         |                |          |
| 877-09-8          | Tetrachloro-m-xylene | 21.4  |      |    | 21 - 165 | 107%       | SPK: 20 |                |          |
| 2051-24-3         | Decachlorobiphenyl   | 18.6  |      |    | 10 - 170 | 93%        | SPK: 20 |                |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                                                   |                 |          |
|--------------------|---------------------------------------------------|-----------------|----------|
| Client:            | Core Environmental Consultants and Services, Inc. | Date Collected: | 11/25/25 |
| Project:           | Building 50 Probe Investigation                   | Date Received:  | 11/26/25 |
| Client Sample ID:  | B50-SP11-112525                                   | SDG No.:        | Q3741    |
| Lab Sample ID:     | Q3741-05                                          | Matrix:         | SOIL     |
| Analytical Method: | 8082A                                             | % Solid:        | 84.6     |
| Sample Wt/Vol:     | 30.02 g                                           | Final Vol:      | 10000 uL |
| Prep Method:       | SW3541B                                           | Test:           | PCB      |
|                    | Prep Date:                                        |                 | 12/03/25 |

| CAS Number        | Parameter            | Conc. | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana.      | BatchID  |
|-------------------|----------------------|-------|------|----|----------|------------|---------|----------------|----------|
| <b>TARGETS</b>    |                      |       |      |    |          |            |         |                |          |
| 12674-11-2        | Aroclor-1016         | 4.70  | U    | 1  | 4.70     | 20.1       | ug/kg   | 12/03/25 14:03 | PB170795 |
| 11104-28-2        | Aroclor-1221         | 4.80  | U    | 1  | 4.80     | 20.1       | ug/kg   | 12/03/25 14:03 | PB170795 |
| 11141-16-5        | Aroclor-1232         | 4.40  | U    | 1  | 4.40     | 20.1       | ug/kg   | 12/03/25 14:03 | PB170795 |
| 53469-21-9        | Aroclor-1242         | 4.70  | U    | 1  | 4.70     | 20.1       | ug/kg   | 12/03/25 14:03 | PB170795 |
| 12672-29-6        | Aroclor-1248         | 7.00  | U    | 1  | 7.00     | 20.1       | ug/kg   | 12/03/25 14:03 | PB170795 |
| 11097-69-1        | Aroclor-1254         | 20.5  |      | 1  | 3.80     | 20.1       | ug/kg   | 12/03/25 14:03 | PB170795 |
| 37324-23-5        | Aroclor-1262         | 5.90  | U    | 1  | 5.90     | 20.1       | ug/kg   | 12/03/25 14:03 | PB170795 |
| 11100-14-4        | Aroclor-1268         | 4.30  | U    | 1  | 4.30     | 20.1       | ug/kg   | 12/03/25 14:03 | PB170795 |
| 11096-82-5        | Aroclor-1260         | 3.80  | U    | 1  | 3.80     | 20.1       | ug/kg   | 12/03/25 14:03 | PB170795 |
| <b>SURROGATES</b> |                      |       |      |    |          |            |         |                |          |
| 877-09-8          | Tetrachloro-m-xylene | 18.8  |      |    | 21 - 165 | 94%        | SPK: 20 |                |          |
| 2051-24-3         | Decachlorobiphenyl   | 15.9  |      |    | 10 - 170 | 80%        | SPK: 20 |                |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                                                   |                 |          |
|--------------------|---------------------------------------------------|-----------------|----------|
| Client:            | Core Environmental Consultants and Services, Inc. | Date Collected: | 11/25/25 |
| Project:           | Building 50 Probe Investigation                   | Date Received:  | 11/26/25 |
| Client Sample ID:  | B50-SP4-112525                                    | SDG No.:        | Q3741    |
| Lab Sample ID:     | Q3741-06                                          | Matrix:         | SOIL     |
| Analytical Method: | 8082A                                             | % Solid:        | 85.9     |
| Sample Wt/Vol:     | 30.06 g                                           | Final Vol:      | 10000 uL |
| Prep Method:       | SW3541B                                           | Test:           | PCB      |
|                    | Prep Date:                                        |                 | 12/03/25 |

| CAS Number        | Parameter            | Conc. | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana.      | BatchID  |
|-------------------|----------------------|-------|------|----|----------|------------|---------|----------------|----------|
| <b>TARGETS</b>    |                      |       |      |    |          |            |         |                |          |
| 12674-11-2        | Aroclor-1016         | 4.60  | U    | 1  | 4.60     | 19.8       | ug/kg   | 12/03/25 14:53 | PB170795 |
| 11104-28-2        | Aroclor-1221         | 4.70  | U    | 1  | 4.70     | 19.8       | ug/kg   | 12/03/25 14:53 | PB170795 |
| 11141-16-5        | Aroclor-1232         | 4.30  | U    | 1  | 4.30     | 19.8       | ug/kg   | 12/03/25 14:53 | PB170795 |
| 53469-21-9        | Aroclor-1242         | 4.70  | U    | 1  | 4.70     | 19.8       | ug/kg   | 12/03/25 14:53 | PB170795 |
| 12672-29-6        | Aroclor-1248         | 6.90  | U    | 1  | 6.90     | 19.8       | ug/kg   | 12/03/25 14:53 | PB170795 |
| 11097-69-1        | Aroclor-1254         | 3.70  | U    | 1  | 3.70     | 19.8       | ug/kg   | 12/03/25 14:53 | PB170795 |
| 37324-23-5        | Aroclor-1262         | 5.80  | U    | 1  | 5.80     | 19.8       | ug/kg   | 12/03/25 14:53 | PB170795 |
| 11100-14-4        | Aroclor-1268         | 4.20  | U    | 1  | 4.20     | 19.8       | ug/kg   | 12/03/25 14:53 | PB170795 |
| 11096-82-5        | Aroclor-1260         | 3.80  | U    | 1  | 3.80     | 19.8       | ug/kg   | 12/03/25 14:53 | PB170795 |
| <b>SURROGATES</b> |                      |       |      |    |          |            |         |                |          |
| 877-09-8          | Tetrachloro-m-xylene | 20.9  |      |    | 21 - 165 | 104%       | SPK: 20 |                |          |
| 2051-24-3         | Decachlorobiphenyl   | 17.6  |      |    | 10 - 170 | 88%        | SPK: 20 |                |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



# QC SUMMARY

### Surrogate Summary

SDG No.: Q3741

Client: Core Environmental Consultants and Ser

Analytical Method: 8082A

| Lab Sample ID    | Client ID         | Parameter         | Column | Spike | Result | Recovery(%) | Qual | Limits(%) |      |
|------------------|-------------------|-------------------|--------|-------|--------|-------------|------|-----------|------|
|                  |                   |                   |        |       |        |             |      | Low       | High |
| I.BLK-PO114870.D | PIBLK-PO114870.D  | Tetrachloro-m-xyl | 1      | 20    | 15.5   | 77          |      | 60        | 140  |
|                  |                   | Decachlorobiphen  | 1      | 20    | 16.8   | 84          |      | 60        | 140  |
|                  |                   | Tetrachloro-m-xyl | 2      | 20    | 16.2   | 81          |      | 60        | 140  |
|                  |                   | Decachlorobiphen  | 2      | 20    | 17.3   | 86          |      | 60        | 140  |
| I.BLK-PO115556.D | PIBLK-PO115556.D  | Tetrachloro-m-xyl | 1      | 20    | 24.9   | 125         |      | 60        | 140  |
|                  |                   | Decachlorobiphen  | 1      | 20    | 26.0   | 130         |      | 60        | 140  |
|                  |                   | Tetrachloro-m-xyl | 2      | 20    | 23.6   | 118         |      | 60        | 140  |
|                  |                   | Decachlorobiphen  | 2      | 20    | 23.9   | 120         |      | 60        | 140  |
| Q3741-01         | B50-SP3-111925    | Tetrachloro-m-xyl | 1      | 20    | 20.7   | 103         |      | 21        | 165  |
|                  |                   | Decachlorobiphen  | 1      | 20    | 22.6   | 113         |      | 10        | 170  |
|                  |                   | Tetrachloro-m-xyl | 2      | 20    | 20.4   | 102         |      | 21        | 165  |
|                  |                   | Decachlorobiphen  | 2      | 20    | 19.6   | 98          |      | 10        | 170  |
| Q3741-02         | B50-SP13-112025   | Tetrachloro-m-xyl | 1      | 20    | 17.3   | 86          |      | 21        | 165  |
|                  |                   | Decachlorobiphen  | 1      | 20    | 17.3   | 86          |      | 10        | 170  |
|                  |                   | Tetrachloro-m-xyl | 2      | 20    | 18.1   | 91          |      | 21        | 165  |
|                  |                   | Decachlorobiphen  | 2      | 20    | 13.4   | 67          |      | 10        | 170  |
| Q3741-03         | B50-SP14-112025   | Tetrachloro-m-xyl | 1      | 20    | 20.1   | 101         |      | 21        | 165  |
|                  |                   | Decachlorobiphen  | 1      | 20    | 18.4   | 92          |      | 10        | 170  |
|                  |                   | Tetrachloro-m-xyl | 2      | 20    | 20.2   | 101         |      | 21        | 165  |
|                  |                   | Decachlorobiphen  | 2      | 20    | 15.8   | 79          |      | 10        | 170  |
| Q3741-04         | B50-SP9-112425    | Tetrachloro-m-xyl | 1      | 20    | 21.4   | 107         |      | 21        | 165  |
|                  |                   | Decachlorobiphen  | 1      | 20    | 18.6   | 93          |      | 10        | 170  |
|                  |                   | Tetrachloro-m-xyl | 2      | 20    | 21.3   | 107         |      | 21        | 165  |
|                  |                   | Decachlorobiphen  | 2      | 20    | 17.1   | 86          |      | 10        | 170  |
| I.BLK-PO115569.D | PIBLK-PO115569.D  | Tetrachloro-m-xyl | 1      | 20    | 25.5   | 127         |      | 60        | 140  |
|                  |                   | Decachlorobiphen  | 1      | 20    | 25.7   | 129         |      | 60        | 140  |
|                  |                   | Tetrachloro-m-xyl | 2      | 20    | 24.1   | 120         |      | 60        | 140  |
|                  |                   | Decachlorobiphen  | 2      | 20    | 23.8   | 119         |      | 60        | 140  |
| I.BLK-PP076598.D | PIBLK-PP076598.D  | Tetrachloro-m-xyl | 1      | 20    | 22.0   | 110         |      | 60        | 140  |
|                  |                   | Decachlorobiphen  | 1      | 20    | 23.3   | 117         |      | 60        | 140  |
|                  |                   | Tetrachloro-m-xyl | 2      | 20    | 21.3   | 106         |      | 60        | 140  |
|                  |                   | Decachlorobiphen  | 2      | 20    | 22.1   | 110         |      | 60        | 140  |
| I.BLK-PP076748.D | PIBLK-PP076748.D  | Tetrachloro-m-xyl | 1      | 20    | 25.5   | 127         |      | 60        | 140  |
|                  |                   | Decachlorobiphen  | 1      | 20    | 26.9   | 134         |      | 60        | 140  |
|                  |                   | Tetrachloro-m-xyl | 2      | 20    | 26.7   | 133         |      | 60        | 140  |
|                  |                   | Decachlorobiphen  | 2      | 20    | 26.8   | 134         |      | 60        | 140  |
| Q3741-05         | B50-SP11-112525   | Tetrachloro-m-xyl | 1      | 20    | 18.8   | 94          |      | 21        | 165  |
|                  |                   | Decachlorobiphen  | 1      | 20    | 15.9   | 80          |      | 10        | 170  |
|                  |                   | Tetrachloro-m-xyl | 2      | 20    | 18.2   | 91          |      | 21        | 165  |
|                  |                   | Decachlorobiphen  | 2      | 20    | 15.5   | 78          |      | 10        | 170  |
| Q3741-05MS       | B50-SP11-112525MS | Tetrachloro-m-xyl | 1      | 20    | 23.1   | 116         |      | 21        | 165  |

### Surrogate Summary

SDG No.: Q3741

Client: Core Environmental Consultants and Ser

Analytical Method: 8082A

| Lab Sample ID    | Client ID          | Parameter         | Column | Spike | Result | Recovery(%) | Qual | Limits(%) |      |
|------------------|--------------------|-------------------|--------|-------|--------|-------------|------|-----------|------|
|                  |                    |                   |        |       |        |             |      | Low       | High |
| Q3741-05MS       | B50-SP11-112525MS  | Decachlorobiphen  | 1      | 20    | 20.0   | 100         |      | 10        | 170  |
|                  |                    | Tetrachloro-m-xyl | 2      | 20    | 21.9   | 109         |      | 21        | 165  |
| Q3741-05MSD      | B50-SP11-112525MSD | Decachlorobiphen  | 2      | 20    | 19.6   | 98          |      | 10        | 170  |
|                  |                    | Tetrachloro-m-xyl | 1      | 20    | 23.3   | 116         |      | 21        | 165  |
|                  |                    | Decachlorobiphen  | 1      | 20    | 20.5   | 103         |      | 10        | 170  |
|                  |                    | Tetrachloro-m-xyl | 2      | 20    | 22.3   | 112         |      | 21        | 165  |
| Q3741-06         | B50-SP4-112525     | Decachlorobiphen  | 2      | 20    | 20.3   | 101         |      | 10        | 170  |
|                  |                    | Tetrachloro-m-xyl | 1      | 20    | 20.5   | 102         |      | 21        | 165  |
|                  |                    | Decachlorobiphen  | 1      | 20    | 17.6   | 88          |      | 10        | 170  |
|                  |                    | Tetrachloro-m-xyl | 2      | 20    | 20.9   | 104         |      | 21        | 165  |
| I.BLK-PP076762.D | PIBLK-PP076762.D   | Decachlorobiphen  | 2      | 20    | 17.1   | 86          |      | 10        | 170  |
|                  |                    | Tetrachloro-m-xyl | 1      | 20    | 24.9   | 124         |      | 60        | 140  |
|                  |                    | Decachlorobiphen  | 1      | 20    | 25.6   | 128         |      | 60        | 140  |
|                  |                    | Tetrachloro-m-xyl | 2      | 20    | 23.8   | 119         |      | 60        | 140  |
| PB170795BL       | PB170795BL         | Decachlorobiphen  | 2      | 20    | 24.0   | 120         |      | 60        | 140  |
|                  |                    | Tetrachloro-m-xyl | 1      | 20    | 22.1   | 111         |      | 21        | 165  |
|                  |                    | Decachlorobiphen  | 1      | 20    | 23.0   | 115         |      | 10        | 170  |
|                  |                    | Tetrachloro-m-xyl | 2      | 20    | 21.3   | 106         |      | 21        | 165  |
| PB170795BS       | PB170795BS         | Decachlorobiphen  | 2      | 20    | 21.6   | 108         |      | 10        | 170  |
|                  |                    | Tetrachloro-m-xyl | 1      | 20    | 22.8   | 114         |      | 21        | 165  |
|                  |                    | Decachlorobiphen  | 1      | 20    | 23.4   | 117         |      | 10        | 170  |
|                  |                    | Tetrachloro-m-xyl | 2      | 20    | 20.8   | 104         |      | 21        | 165  |
| I.BLK-PP076777.D | PIBLK-PP076777.D   | Decachlorobiphen  | 2      | 20    | 21.9   | 109         |      | 10        | 170  |
|                  |                    | Tetrachloro-m-xyl | 1      | 20    | 24.2   | 121         |      | 60        | 140  |
|                  |                    | Decachlorobiphen  | 1      | 20    | 25.2   | 126         |      | 60        | 140  |
|                  |                    | Tetrachloro-m-xyl | 2      | 20    | 23.4   | 117         |      | 60        | 140  |
|                  |                    | Decachlorobiphen  | 2      | 20    | 23.7   | 118         |      | 60        | 140  |

### Matrix Spike/Matrix Spike Duplicate Summary

SW-846

|                 |                                                |                           |                   |
|-----------------|------------------------------------------------|---------------------------|-------------------|
| <b>SDG No.:</b> | <u>Q3741</u>                                   | <b>Analytical Method:</b> | <u>8082A</u>      |
| <b>Client:</b>  | <u>Core Environmental Consultants and Serv</u> | <b>DataFile :</b>         | <u>PP076752.D</u> |

|                | Parameter                | Spike             | Sample |                   | Units | Rec | Rec  |     | RPD  |     | Limits |     | RPD |
|----------------|--------------------------|-------------------|--------|-------------------|-------|-----|------|-----|------|-----|--------|-----|-----|
|                |                          |                   | Result | Result            |       |     | Qual | RPD | Qual | Low | High   |     |     |
| Lab Sample ID: | Q3741-05MS<br>(Column 1) | Client Sample ID: |        | B50-SP11-112525MS |       |     |      |     |      |     |        |     |     |
|                | AR1016                   | 196.5             | 0      | 160               | ug/kg | 81  |      |     |      | 39  |        | 159 |     |
|                | AR1260                   | 196.5             | 0      | 147               | ug/kg | 75  |      |     |      | 19  |        | 160 |     |
| Lab Sample ID: | Q3741-05MS<br>(Column 2) | Client Sample ID: |        | B50-SP11-112525MS |       |     |      |     |      |     |        |     |     |
|                | AR1016                   | 196.5             | 0      | 156               | ug/kg | 79  |      |     |      | 39  |        | 159 |     |
|                | AR1260                   | 196.5             | 0      | 150               | ug/kg | 76  |      |     |      | 19  |        | 160 |     |

### Matrix Spike/Matrix Spike Duplicate Summary

SW-846

|                 |                                                |                           |                   |
|-----------------|------------------------------------------------|---------------------------|-------------------|
| <b>SDG No.:</b> | <u>Q3741</u>                                   | <b>Analytical Method:</b> | <u>8082A</u>      |
| <b>Client:</b>  | <u>Core Environmental Consultants and Serv</u> | <b>DataFile :</b>         | <u>PP076753.D</u> |

|                | Parameter                 | Spike             | Sample             |        | Units | Rec | Rec  |     | RPD  |     | Limits |     |
|----------------|---------------------------|-------------------|--------------------|--------|-------|-----|------|-----|------|-----|--------|-----|
|                |                           |                   | Result             | Result |       |     | Qual | RPD | Qual | Low | High   | RPD |
| Lab Sample ID: | Q3741-05MSD<br>(Column 1) | Client Sample ID: | B50-SP11-112525MSD |        |       |     |      |     |      |     |        |     |
|                | AR1016                    | 196.7             | 0                  | 163    | ug/kg | 83  |      | 2   |      | 39  | 159    | 21  |
|                | AR1260                    | 196.7             | 0                  | 149    | ug/kg | 76  |      | 1   |      | 19  | 160    | 15  |
| Lab Sample ID: | Q3741-05MSD<br>(Column 2) | Client Sample ID: | B50-SP11-112525MSD |        |       |     |      |     |      |     |        |     |
|                | AR1016                    | 196.7             | 0                  | 160    | ug/kg | 81  |      | 3   |      | 39  | 159    | 21  |
|                | AR1260                    | 196.7             | 0                  | 152    | ug/kg | 77  |      | 1   |      | 19  | 160    | 15  |

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

|                 |                                                |                           |                   |
|-----------------|------------------------------------------------|---------------------------|-------------------|
| <b>SDG No.:</b> | <u>Q3741</u>                                   | <b>Analytical Method:</b> | <u>8082A</u>      |
| <b>Client:</b>  | <u>Core Environmental Consultants and Serv</u> | <b>Datafile :</b>         | <u>PP076764.D</u> |

| Lab Sample ID            | Parameter | Spike | Result | Units | Rec | RPD | Qual | RPD  | Limits |      | RPD |
|--------------------------|-----------|-------|--------|-------|-----|-----|------|------|--------|------|-----|
|                          |           |       |        |       |     |     |      | Qual | Low    | High |     |
| PB170795BS<br>(Column 1) | AR1016    | 166.6 | 139    | ug/kg | 83  |     |      |      | 71     | 120  |     |
|                          | AR1260    | 166.6 | 134    | ug/kg | 80  |     |      |      | 65     | 130  |     |
| PB170795BS<br>(Column 2) | AR1016    | 166.6 | 136    | ug/kg | 82  |     |      |      | 71     | 120  |     |
|                          | AR1260    | 166.6 | 135    | ug/kg | 81  |     |      |      | 65     | 130  |     |

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID

PB170795BL

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG NO.: Q3741

Lab Sample ID: PB170795BL

Lab File ID: PP076763.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 12/03/2025

Date Analyzed (1): 12/03/2025

Date Analyzed (2): 12/03/2025

Time Analyzed (1): 18:40

Time Analyzed (2): 18:40

Instrument ID (1): ECD\_P

Instrument ID (2): ECD\_P

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

| EPA<br>SAMPLE NO.  | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED 1 | DATE<br>ANALYZED 2 |
|--------------------|------------------|----------------|--------------------|--------------------|
| B50-SP3-111925     | Q3741-01         | PO115561.D     | 12/03/2025         | 12/03/2025         |
| B50-SP13-112025    | Q3741-02         | PO115562.D     | 12/03/2025         | 12/03/2025         |
| B50-SP14-112025    | Q3741-03         | PO115563.D     | 12/03/2025         | 12/03/2025         |
| B50-SP9-112425     | Q3741-04         | PO115564.D     | 12/03/2025         | 12/03/2025         |
| B50-SP11-112525    | Q3741-05         | PP076751.D     | 12/03/2025         | 12/03/2025         |
| B50-SP11-112525MS  | Q3741-05MS       | PP076752.D     | 12/03/2025         | 12/03/2025         |
| B50-SP11-112525MSD | Q3741-05MSD      | PP076753.D     | 12/03/2025         | 12/03/2025         |
| B50-SP4-112525     | Q3741-06         | PP076754.D     | 12/03/2025         | 12/03/2025         |
| PB170795BS         | PB170795BS       | PP076764.D     | 12/03/2025         | 12/03/2025         |

COMMENTS: \_\_\_\_\_



# CALIBRATION SUMMARY

A  
B  
C  
D  
E  
F  
G

**Calibration Times:**                      09:39                      18:31

LAB FILE ID: RT 1000 = PO114871.D RT 750 = PO114872.D  
RT 500 = PO114873.D RT 250 = PO114874.D RT 050 = PO114875.D

[illegible]

**RETENTION TIMES OF INITIAL CALIBRATION**

|                      |      |      |      |      |      |      |      |       |
|----------------------|------|------|------|------|------|------|------|-------|
| Decachlorobiphenyl   | 9.93 | 9.93 | 9.93 | 9.93 | 9.93 | 9.93 | 9.83 | 10.03 |
| Tetrachloro-m-xylene | 4.35 | 4.35 | 4.35 | 4.35 | 4.35 | 4.35 | 4.25 | 4.45  |

A  
B  
C  
D  
E  
F  
G

A  
B  
C  
D  
E  
F  
G

**Contract:** CORE02

SDG NO.: Q3741

**Calibration Date(s):**           **11/04/2025**           **11/04/2025**

**Calibration Times:**                      09:39                      18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

[illegible]

**RETENTION TIMES OF INITIAL CALIBRATION**

|                      |      |      |      |      |      |      |      |      |
|----------------------|------|------|------|------|------|------|------|------|
| Decachlorobiphenyl   | 8.64 | 8.64 | 8.64 | 8.64 | 8.64 | 8.64 | 8.54 | 8.74 |
| Tetrachloro-m-xylene | 3.65 | 3.65 | 3.65 | 3.65 | 3.65 | 3.65 | 3.55 | 3.75 |

A  
B  
C  
D  
E  
F  
G

### CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance  
Lab Code: ACE  
Instrument ID: ECD\_O

Contract: CORE02  
SDG NO.: Q3741

Calibration Date(s): 11/04/2025 11/04/2025  
Calibration Times: 09:39 18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

| LAB FILE ID:               |     | CF 1000 = <u>PO114871.D</u> |            | CF 750 = <u>PO114872.D</u> |            |            |            |       |
|----------------------------|-----|-----------------------------|------------|----------------------------|------------|------------|------------|-------|
| CF 500 = <u>PO114873.D</u> |     | CF 250 = <u>PO114874.D</u>  |            | CF 050 = <u>PO114875.D</u> |            |            |            |       |
| COMPOUND                   |     | CF 1000                     | CF 750     | CF 500                     | CF 250     | CF 050     | CF         | % RSD |
| Aroclor-1016-1             | (1) | 235183609                   | 242405119  | 252300612                  | 252330724  | 247880040  | 246020021  | 3     |
| Aroclor-1016-2             | (2) | 380671765                   | 391174169  | 403435360                  | 408228592  | 401844700  | 397070917  | 3     |
| Aroclor-1016-3             | (3) | 214892716                   | 220811199  | 230510330                  | 231508016  | 217282560  | 223000964  | 3     |
| Aroclor-1016-4             | (4) | 180735079                   | 184273587  | 191842228                  | 192742164  | 194782360  | 188875084  | 3     |
| Aroclor-1016-5             | (5) | 166479276                   | 173610481  | 179587174                  | 180772736  | 190536000  | 178197133  | 5     |
| Aroclor-1260-1             | (1) | 291080181                   | 301659089  | 311914366                  | 329562032  | 293951540  | 305633442  | 5     |
| Aroclor-1260-2             | (2) | 393658847                   | 402595176  | 414352186                  | 418536840  | 350242960  | 395877202  | 7     |
| Aroclor-1260-3             | (3) | 331096941                   | 340266576  | 347857958                  | 348147780  | 291720640  | 331817979  | 7     |
| Aroclor-1260-4             | (4) | 303823322                   | 310642864  | 321126472                  | 321507392  | 268000240  | 305020058  | 7     |
| Aroclor-1260-5             | (5) | 714393776                   | 724690244  | 737226162                  | 719415648  | 670367440  | 713218654  | 3     |
| Decachlorobiphenyl         |     | 4451734610                  | 4516548787 | 4612561720                 | 4524869360 | 3995527600 | 4420248415 | 6     |
| Tetrachloro-m-xylene       |     | 7020403280                  | 7162783533 | 7307727940                 | 7156075200 | 6543618600 | 7038121711 | 4     |
| Aroclor-1242-1             | (1) | 180395803                   | 184480117  | 189796832                  | 192180524  | 184350480  | 186240751  | 2     |
| Aroclor-1242-2             | (2) | 290607708                   | 295647931  | 305830174                  | 302003732  | 299791360  | 298776181  | 2     |
| Aroclor-1242-3             | (3) | 164449629                   | 167684461  | 173929682                  | 174410736  | 176369820  | 171368866  | 3     |
| Aroclor-1242-4             | (4) | 136844790                   | 140633719  | 146402846                  | 139660276  | 152376540  | 143183634  | 4     |
| Aroclor-1242-5             | (5) | 138362947                   | 141839337  | 145002188                  | 138385972  | 141196180  | 140957325  | 2     |
| Decachlorobiphenyl         |     | 4729850460                  | 4804993427 | 4858375920                 | 4746069120 | 4160138600 | 4659885505 | 6     |
| Tetrachloro-m-xylene       |     | 7348004600                  | 7445204373 | 7532777260                 | 7262019280 | 6694579200 | 7256516943 | 5     |
| Aroclor-1248-1             | (1) | 130126160                   | 135244476  | 137909000                  | 141742196  | 154206600  | 139845686  | 6     |
| Aroclor-1248-2             | (2) | 182893435                   | 187680161  | 194064692                  | 196392780  | 211353600  | 194476934  | 5     |
| Aroclor-1248-3             | (3) | 203368877                   | 206306241  | 213983954                  | 212927108  | 217587500  | 210834736  | 3     |
| Aroclor-1248-4             | (4) | 240181487                   | 246210361  | 255848644                  | 255957264  | 264773260  | 252594203  | 4     |
| Aroclor-1248-5             | (5) | 225312398                   | 230738423  | 239469078                  | 240115800  | 243213880  | 235769916  | 3     |
| Decachlorobiphenyl         |     | 4590720290                  | 4647354867 | 4744288880                 | 4661202760 | 4211361200 | 4570985599 | 5     |
| Tetrachloro-m-xylene       |     | 7316541360                  | 7442934853 | 7528054520                 | 7402608560 | 7218400400 | 7381707939 | 2     |
| Aroclor-1254-1             | (1) | 240591852                   | 247416493  | 258765478                  | 257075152  | 263299240  | 253429643  | 3     |
| Aroclor-1254-2             | (2) | 346021204                   | 353529297  | 371589652                  | 368897016  | 412184820  | 370444398  | 7     |
| Aroclor-1254-3             | (3) | 372893964                   | 380446979  | 397009938                  | 384293176  | 389678940  | 384864599  | 2     |
| Aroclor-1254-4             | (4) | 315539209                   | 321424513  | 336840998                  | 326904952  | 342871860  | 328716306  | 3     |
| Aroclor-1254-5             | (5) | 295373361                   | 300690525  | 312155076                  | 300474784  | 300573260  | 301853401  | 2     |
| Decachlorobiphenyl         |     | 4791769470                  | 4896862187 | 5061968640                 | 4831244600 | 4661831200 | 4848735219 | 3     |
| Tetrachloro-m-xylene       |     | 7410585440                  | 7526809333 | 7802184660                 | 7481754040 | 7408041400 | 7525874975 | 2     |
| Aroclor-1268-1             | (1) | 876279188                   | 885717404  | 893041022                  | 887529864  | 813508780  | 871215252  | 4     |

**CALIBRATION FACTOR OF INITIAL CALIBRATION**

|                      |     |            |            |            |            |            |            |   |
|----------------------|-----|------------|------------|------------|------------|------------|------------|---|
| Aroclor-1268-2       | (2) | 748334854  | 756337347  | 765525536  | 759915712  | 702804360  | 746583562  | 3 |
| Aroclor-1268-3       | (3) | 663299706  | 672085827  | 679866912  | 671661004  | 617886840  | 660960058  | 4 |
| Aroclor-1268-4       | (4) | 274155738  | 277531193  | 282310020  | 282161716  | 286808980  | 280593529  | 2 |
| Aroclor-1268-5       | (5) | 1966581606 | 1979440064 | 1985729368 | 1956798376 | 1807416340 | 1939193151 | 4 |
| Decachlorobiphenyl   |     | 8288385350 | 8396434587 | 8504946880 | 8419357920 | 7722111200 | 8266247187 | 4 |
| Tetrachloro-m-xylene |     | 7475792880 | 7636683133 | 7655291620 | 7597134240 | 7121440400 | 7497268455 | 3 |

**CALIBRATION FACTOR OF INITIAL CALIBRATION**

**Lab Name:** Alliance  
**Lab Code:** ACE  
**Instrument ID:** ECD\_O

**Contract:** CORE02  
**SDG NO.:** Q3741

**Calibration Date(s):** 11/04/2025 11/04/2025  
**Calibration Times:** 09:39 18:31

**GC Column:** ZB-MR2 **ID:** 0.32 (mm)

| <b>LAB FILE ID:</b> CF 1000 = <u>PO114871.D</u> CF 750 = <u>PO114872.D</u><br>CF 500 = <u>PO114873.D</u> CF 250 = <u>PO114874.D</u> CF 050 = <u>PO114875.D</u> |     |            |            |            |            |            |       |
|----------------------------------------------------------------------------------------------------------------------------------------------------------------|-----|------------|------------|------------|------------|------------|-------|
| COMPOUND                                                                                                                                                       |     | CF 1000    | CF 750     | CF 500     | CF 250     | CF 050     | % RSD |
| Aroclor-1016-1                                                                                                                                                 | (1) | 178795127  | 185535456  | 193106802  | 200932132  | 192907960  | 4     |
| Aroclor-1016-2                                                                                                                                                 | (2) | 260787154  | 267435697  | 277593168  | 286495324  | 262010880  | 4     |
| Aroclor-1016-3                                                                                                                                                 | (3) | 140310175  | 144533207  | 150201526  | 154962192  | 149858780  | 4     |
| Aroclor-1016-4                                                                                                                                                 | (4) | 108607533  | 112021955  | 117721682  | 122827936  | 125909940  | 6     |
| Aroclor-1016-5                                                                                                                                                 | (5) | 142235561  | 147493437  | 153611756  | 157990124  | 162236540  | 5     |
| Aroclor-1260-1                                                                                                                                                 | (1) | 272865489  | 283696256  | 302549666  | 306994580  | 314431380  | 5     |
| Aroclor-1260-2                                                                                                                                                 | (2) | 410849028  | 420744668  | 440173148  | 441650472  | 436738740  | 3     |
| Aroclor-1260-3                                                                                                                                                 | (3) | 328919579  | 344471297  | 362206406  | 366201912  | 353957800  | 4     |
| Aroclor-1260-4                                                                                                                                                 | (4) | 295043629  | 302500340  | 313130718  | 316509244  | 287959420  | 4     |
| Aroclor-1260-5                                                                                                                                                 | (5) | 819932115  | 830411719  | 851082708  | 846532168  | 760181660  | 4     |
| Decachlorobiphenyl                                                                                                                                             |     | 3768213910 | 3862452533 | 4051635320 | 3997632280 | 3839185200 | 3     |
| Tetrachloro-m-xylene                                                                                                                                           |     | 4444199150 | 4532189360 | 4610968540 | 4573862320 | 4172678600 | 4     |
| Aroclor-1242-1                                                                                                                                                 | (1) | 138347633  | 144438088  | 149108388  | 148877560  | 150336280  | 3     |
| Aroclor-1242-2                                                                                                                                                 | (2) | 199908386  | 204961789  | 213926082  | 211880332  | 205062860  | 3     |
| Aroclor-1242-3                                                                                                                                                 | (3) | 108216818  | 112229521  | 116418372  | 113485296  | 109537600  | 3     |
| Aroclor-1242-4                                                                                                                                                 | (4) | 101919263  | 105301932  | 110004430  | 109236748  | 109193360  | 3     |
| Aroclor-1242-5                                                                                                                                                 | (5) | 135980897  | 140103160  | 145653410  | 150301700  | 142549620  | 4     |
| Decachlorobiphenyl                                                                                                                                             |     | 4093300380 | 4301738307 | 4403735460 | 4338775160 | 4380898600 | 3     |
| Tetrachloro-m-xylene                                                                                                                                           |     | 4800115840 | 4683347560 | 4961464680 | 4527345840 | 4437059000 | 4     |
| Aroclor-1248-1                                                                                                                                                 | (1) | 101027399  | 105625799  | 111262386  | 113093188  | 115539800  | 5     |
| Aroclor-1248-2                                                                                                                                                 | (2) | 136540550  | 142594189  | 148530518  | 152692968  | 164144180  | 7     |
| Aroclor-1248-3                                                                                                                                                 | (3) | 143887712  | 149887199  | 155474250  | 158349236  | 169440400  | 6     |
| Aroclor-1248-4                                                                                                                                                 | (4) | 171266913  | 177666692  | 184171754  | 187513908  | 195642380  | 5     |
| Aroclor-1248-5                                                                                                                                                 | (5) | 174649655  | 180220384  | 187038042  | 189642400  | 190092660  | 3     |
| Decachlorobiphenyl                                                                                                                                             |     | 4076042330 | 4166088587 | 4290124020 | 4403270520 | 4144788800 | 3     |
| Tetrachloro-m-xylene                                                                                                                                           |     | 4758509080 | 4829545453 | 4892334280 | 4811209560 | 4727034000 | 1     |
| Aroclor-1254-1                                                                                                                                                 | (1) | 275439068  | 284238351  | 300243386  | 299923620  | 341055960  | 8     |
| Aroclor-1254-2                                                                                                                                                 | (2) | 245333366  | 253279232  | 267984098  | 268318040  | 313122940  | 9     |
| Aroclor-1254-3                                                                                                                                                 | (3) | 398238109  | 410614839  | 430545692  | 426530308  | 445294740  | 4     |
| Aroclor-1254-4                                                                                                                                                 | (4) | 284785979  | 294895399  | 305027804  | 301955520  | 300306220  | 3     |
| Aroclor-1254-5                                                                                                                                                 | (5) | 385708423  | 392326712  | 409498118  | 398089984  | 402006780  | 2     |
| Decachlorobiphenyl                                                                                                                                             |     | 4273373360 | 4367226933 | 4604811080 | 4533103920 | 4272134600 | 3     |
| Tetrachloro-m-xylene                                                                                                                                           |     | 4791085320 | 4887004773 | 5044010700 | 4876428000 | 4843036000 | 2     |
| Aroclor-1268-1                                                                                                                                                 | (1) | 948830555  | 965398547  | 981732904  | 993582148  | 991082540  | 2     |

**CALIBRATION FACTOR OF INITIAL CALIBRATION**

|                      |     |            |            |            |            |            |            |   |
|----------------------|-----|------------|------------|------------|------------|------------|------------|---|
| Aroclor-1268-2       | (2) | 784312077  | 796694211  | 811227528  | 817006596  | 768929160  | 795633914  | 2 |
| Aroclor-1268-3       | (3) | 672635507  | 681509284  | 695723330  | 696225136  | 632763280  | 675771307  | 4 |
| Aroclor-1268-4       | (4) | 270862616  | 275402580  | 280086838  | 282372280  | 267609400  | 275266743  | 2 |
| Aroclor-1268-5       | (5) | 1768544208 | 1786014605 | 1818457210 | 1831952004 | 1717683400 | 1784530285 | 2 |
| Decachlorobiphenyl   |     | 7349980390 | 7437822467 | 7706201340 | 7861559120 | 8321979800 | 7735508623 | 5 |
| Tetrachloro-m-xylene |     | 4791521570 | 4902547960 | 4915920660 | 4883273320 | 4724156000 | 4843483902 | 2 |

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: CORE02

Lab Code: ACE SDG NO.: Q3741

Instrument ID: ECD\_O Date(s) Analyzed: 11/04/2025 11/04/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND     | AMOUNT<br>(ng) | PEAK | RT   | RT WINDOW |      | CALIBRATION<br>FACTOR |
|--------------|----------------|------|------|-----------|------|-----------------------|
|              |                |      |      | FROM      | TO   |                       |
| Aroclor-1221 | 500            | 1    | 4.55 | 4.45      | 4.65 | 84124000              |
|              |                | 2    | 4.63 | 4.53      | 4.73 | 60183600              |
|              |                | 3    | 4.71 | 4.61      | 4.81 | 193335000             |
|              |                | 4    | 0.00 |           |      | 0                     |
|              |                | 5    | 0.00 |           |      | 0                     |
| Aroclor-1232 | 500            | 1    | 4.71 | 4.61      | 4.81 | 150015000             |
|              |                | 2    | 5.22 | 5.12      | 5.32 | 72567400              |
|              |                | 3    | 5.51 | 5.41      | 5.61 | 158675000             |
|              |                | 4    | 5.67 | 5.57      | 5.77 | 73795600              |
|              |                | 5    | 5.76 | 5.66      | 5.86 | 51580600              |
| Aroclor-1262 | 500            | 1    | 7.70 | 7.60      | 7.80 | 453404000             |
|              |                | 2    | 8.23 | 8.13      | 8.33 | 707198000             |
|              |                | 3    | 8.53 | 8.43      | 8.63 | 451420000             |
|              |                | 4    | 8.62 | 8.52      | 8.72 | 336186000             |
|              |                | 5    | 9.24 | 9.14      | 9.34 | 241816000             |

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: CORE02

Lab Code: ACE SDG NO.: Q3741

Instrument ID: ECD\_O Date(s) Analyzed: 11/04/2025 11/04/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND     | AMOUNT<br>(ng) | PEAK | RT   | RT WINDOW |      | CALIBRATION<br>FACTOR |
|--------------|----------------|------|------|-----------|------|-----------------------|
|              |                |      |      | FROM      | TO   |                       |
| Aroclor-1221 | 500            | 1    | 3.86 | 3.76      | 3.96 | 61799200              |
|              |                | 2    | 3.95 | 3.85      | 4.05 | 45637000              |
|              |                | 3    | 4.02 | 3.92      | 4.12 | 140257000             |
|              |                | 4    | 0.00 |           |      | 0                     |
|              |                | 5    | 0.00 |           |      | 0                     |
| Aroclor-1232 | 500            | 1    | 4.02 | 3.92      | 4.12 | 111493000             |
|              |                | 2    | 4.75 | 4.65      | 4.85 | 112095000             |
|              |                | 3    | 4.92 | 4.82      | 5.02 | 59654200              |
|              |                | 4    | 5.01 | 4.91      | 5.11 | 50889000              |
|              |                | 5    | 5.18 | 5.08      | 5.28 | 56587600              |
| Aroclor-1262 | 500            | 1    | 6.76 | 6.66      | 6.86 | 494206000             |
|              |                | 2    | 7.26 | 7.16      | 7.36 | 795090000             |
|              |                | 3    | 7.54 | 7.44      | 7.64 | 329028000             |
|              |                | 4    | 7.61 | 7.51      | 7.71 | 538192000             |
|              |                | 5    | 8.10 | 8.00      | 8.20 | 238020000             |

A  
B  
C  
D  
E  
F  
G

**Contract:** CORE02

SDG NO.: Q3741

**Calibration Date(s):**           **11/21/2025**           **11/21/2025**

**Calibration Times:**                      10:44                      19:39

GC Column: ZB-MR1 ID: 0.32 (mm)

|              |            |           |                   |          |                   |
|--------------|------------|-----------|-------------------|----------|-------------------|
| LAB FILE ID: |            | RT 1000 = | <u>PP076599.D</u> | RT 750 = | <u>PP076600.D</u> |
| RT 500 =     | PP076601.D | RT 250 =  | PP076602.D        | RT 050 = | PP076603.D        |

[illegible]

**RETENTION TIMES OF INITIAL CALIBRATION**

|                      |       |       |       |       |       |       |       |       |
|----------------------|-------|-------|-------|-------|-------|-------|-------|-------|
| Decachlorobiphenyl   | 10.37 | 10.37 | 10.37 | 10.37 | 10.37 | 10.37 | 10.27 | 10.47 |
| Tetrachloro-m-xylene | 4.62  | 4.62  | 4.62  | 4.62  | 4.62  | 4.62  | 4.52  | 4.72  |

A  
B  
C  
D  
E  
F  
G

A  
B  
C  
D  
E  
F  
G

**Contract:** CORE02

SDG NO.: Q3741

**Calibration Date(s):**            **11/21/2025**            **11/21/2025**

**Calibration Times:**                    10:44                    19:39

GC Column: ZB-MR2 ID: 0.32 (mm)

|                     |           |                   |          |                   |
|---------------------|-----------|-------------------|----------|-------------------|
| LAB FILE ID:        | RT 1000 = | <u>PP076599.D</u> | RT 750 = | <u>PP076600.D</u> |
| RT 500 = PP076601.D | RT 250 =  | PP076602.D        | RT 050 = | PP076603.D        |

[illegible]

**RETENTION TIMES OF INITIAL CALIBRATION**

|                      |      |      |      |      |      |      |      |      |
|----------------------|------|------|------|------|------|------|------|------|
| Decachlorobiphenyl   | 8.77 | 8.77 | 8.77 | 8.77 | 8.77 | 8.77 | 8.67 | 8.87 |
| Tetrachloro-m-xylene | 3.78 | 3.77 | 3.78 | 3.77 | 3.77 | 3.77 | 3.67 | 3.87 |

A  
B  
C  
D  
E  
F  
G

**CALIBRATION FACTOR OF INITIAL CALIBRATION**

**Lab Name:** Alliance  
**Lab Code:** ACE  
**Instrument ID:** ECD\_P

**Contract:** CORE02  
**SDG NO.:** Q3741

**Calibration Date(s):** 11/21/2025 11/21/2025  
**Calibration Times:** 10:44 19:39

**GC Column:** ZB-MR1 **ID:** 0.32 (mm)

| LAB FILE ID:               |     | CF 1000 = <u>PP076599.D</u> |            | CF 750 = <u>PP076600.D</u> |            |            |            |
|----------------------------|-----|-----------------------------|------------|----------------------------|------------|------------|------------|
| CF 500 = <u>PP076601.D</u> |     | CF 250 = <u>PP076602.D</u>  |            | CF 050 = <u>PP076603.D</u> |            |            |            |
| COMPOUND                   |     | CF 1000                     | CF 750     | CF 500                     | CF 250     | CF 050     | % RSD      |
| Aroclor-1016-1             | (1) | 57708593                    | 60284011   | 62720674                   | 70146056   | 86715180   | 67514903   |
| Aroclor-1016-2             | (2) | 87370224                    | 91866497   | 94863778                   | 106525840  | 123912700  | 100907808  |
| Aroclor-1016-3             | (3) | 55014048                    | 58179041   | 60501046                   | 68146480   | 83131840   | 64994491   |
| Aroclor-1016-4             | (4) | 45866425                    | 48590819   | 50075960                   | 56156272   | 67027420   | 53543379   |
| Aroclor-1016-5             | (5) | 44467962                    | 47019184   | 48926274                   | 54576040   | 62856580   | 51569208   |
| Aroclor-1260-1             | (1) | 73805780                    | 77143584   | 80863582                   | 91288048   | 111181860  | 86856571   |
| Aroclor-1260-2             | (2) | 87101421                    | 91325823   | 95770194                   | 107029868  | 135552320  | 103355925  |
| Aroclor-1260-3             | (3) | 68153614                    | 71731653   | 74556476                   | 83896084   | 103586280  | 80384821   |
| Aroclor-1260-4             | (4) | 80664425                    | 84387037   | 87985074                   | 97767224   | 117554160  | 93671584   |
| Aroclor-1260-5             | (5) | 148193731                   | 153678800  | 158669232                  | 175972884  | 213566900  | 170016309  |
| Decachlorobiphenyl         |     | 1308548090                  | 1363697893 | 1416143240                 | 1583148160 | 1821861400 | 1498679757 |
| Tetrachloro-m-xylene       |     | 1632622270                  | 1686884480 | 1742740480                 | 1876774200 | 2238526800 | 1835509646 |
| Aroclor-1242-1             | (1) | 46820088                    | 49495555   | 51408518                   | 51802020   | 60136760   | 51932588   |
| Aroclor-1242-2             | (2) | 69927193                    | 74300271   | 78434750                   | 76478716   | 89342220   | 77696630   |
| Aroclor-1242-3             | (3) | 44603953                    | 47284257   | 49953000                   | 49585884   | 60152920   | 50316003   |
| Aroclor-1242-4             | (4) | 36970364                    | 39006845   | 41279374                   | 41298156   | 49350320   | 41581012   |
| Aroclor-1242-5             | (5) | 40380815                    | 43273208   | 45507256                   | 47172540   | 54801840   | 46227132   |
| Decachlorobiphenyl         |     | 1250955220                  | 1331366640 | 1381877340                 | 1379254120 | 1600423400 | 1388775344 |
| Tetrachloro-m-xylene       |     | 1571806280                  | 1652818307 | 1692613960                 | 1639542200 | 1904693000 | 1692294749 |
| Aroclor-1248-1             | (1) | 36073269                    | 37934815   | 40711200                   | 44451532   | 43127200   | 40459603   |
| Aroclor-1248-2             | (2) | 50000146                    | 52867352   | 55993870                   | 61932632   | 61150200   | 56388840   |
| Aroclor-1248-3             | (3) | 55566335                    | 58034408   | 62017348                   | 68318008   | 67641780   | 62315576   |
| Aroclor-1248-4             | (4) | 64403994                    | 67793604   | 72096762                   | 81581836   | 84852220   | 74145683   |
| Aroclor-1248-5             | (5) | 63599425                    | 66649801   | 71757734                   | 83225952   | 78337200   | 72714022   |
| Decachlorobiphenyl         |     | 1249447780                  | 1292070840 | 1384857840                 | 1528460280 | 1423697000 | 1375706748 |
| Tetrachloro-m-xylene       |     | 1574270000                  | 1618772787 | 1673222680                 | 1811616160 | 1785283000 | 1692632925 |
| Aroclor-1254-1             | (1) | 71656435                    | 73094801   | 76839598                   | 83916248   | 95588240   | 80219064   |
| Aroclor-1254-2             | (2) | 95196281                    | 100474461  | 108389070                  | 119616740  | 139960620  | 112727434  |
| Aroclor-1254-3             | (3) | 102399354                   | 107941216  | 115149900                  | 126570796  | 148664480  | 120145149  |
| Aroclor-1254-4             | (4) | 78180335                    | 82150911   | 87972230                   | 96257768   | 106029800  | 90118209   |
| Aroclor-1254-5             | (5) | 84891224                    | 89223744   | 94355464                   | 99145184   | 116356540  | 96794431   |
| Decachlorobiphenyl         |     | 1271379670                  | 1353876560 | 1432914580                 | 1578042720 | 1693770400 | 1465996786 |
| Tetrachloro-m-xylene       |     | 1587463770                  | 1656217453 | 1731172560                 | 1831394240 | 1978188800 | 1756887365 |
| Aroclor-1268-1             | (1) | 185284719                   | 196773012  | 206651882                  | 228623276  | 257246740  | 214915926  |

**CALIBRATION FACTOR OF INITIAL CALIBRATION**

|                      |     |            |            |            |            |            |            |    |
|----------------------|-----|------------|------------|------------|------------|------------|------------|----|
| Aroclor-1268-2       | (2) | 174556125  | 185120223  | 194982316  | 214398960  | 237120540  | 201235633  | 12 |
| Aroclor-1268-3       | (3) | 144029935  | 152659056  | 161257230  | 178389696  | 194435320  | 166154247  | 12 |
| Aroclor-1268-4       | (4) | 72720190   | 77125884   | 81820048   | 89236340   | 95590360   | 83298564   | 11 |
| Aroclor-1268-5       | (5) | 445375962  | 467824043  | 486648804  | 530281216  | 591465460  | 504319097  | 11 |
| Decachlorobiphenyl   |     | 2103145540 | 2229946733 | 2356050020 | 2605971520 | 2869837800 | 2432990323 | 13 |
| Tetrachloro-m-xylene |     | 1592145550 | 1682577760 | 1746782760 | 1859352840 | 2045449600 | 1785261702 | 10 |

**CALIBRATION FACTOR OF INITIAL CALIBRATION**

**Lab Name:** Alliance  
**Lab Code:** ACE  
**Instrument ID:** ECD\_P

**Contract:** CORE02  
**SDG NO.:** Q3741

**Calibration Date(s):** 11/21/2025 11/21/2025  
**Calibration Times:** 10:44 19:39

**GC Column:** ZB-MR2 **ID:** 0.32 (mm)

| LAB FILE ID:         |     |            |            |            |            |            |            |       |
|----------------------|-----|------------|------------|------------|------------|------------|------------|-------|
| CF 1000 =            |     | PP076599.D |            | CF 750 =   |            | PP076600.D |            |       |
| CF 500 =             |     | PP076601.D |            | CF 250 =   |            | PP076602.D |            |       |
|                      |     |            |            | CF 050 =   |            | PP076603.D |            |       |
| COMPOUND             |     | CF 1000    | CF 750     | CF 500     | CF 250     | CF 050     | CF         | % RSD |
| Aroclor-1016-1       | (1) | 69284330   | 70804845   | 71449300   | 76694208   | 87621040   | 75170745   | 10    |
| Aroclor-1016-2       | (2) | 104122902  | 106337112  | 106728386  | 113977468  | 124797320  | 111192638  | 8     |
| Aroclor-1016-3       | (3) | 53613162   | 55061171   | 55762260   | 59522960   | 65859300   | 57963771   | 8     |
| Aroclor-1016-4       | (4) | 41823980   | 43401780   | 44668820   | 48687932   | 55170840   | 46750670   | 11    |
| Aroclor-1016-5       | (5) | 55742749   | 57230741   | 58964166   | 64122844   | 73641320   | 61940364   | 12    |
| Aroclor-1260-1       | (1) | 99306847   | 101128783  | 103836720  | 111650784  | 129586700  | 109101967  | 11    |
| Aroclor-1260-2       | (2) | 121008321  | 123078656  | 126124030  | 136041204  | 158654880  | 132981418  | 12    |
| Aroclor-1260-3       | (3) | 113883356  | 115970629  | 117487594  | 125714324  | 146413640  | 123893909  | 11    |
| Aroclor-1260-4       | (4) | 96232422   | 98264441   | 99716326   | 106139636  | 119039280  | 103878421  | 9     |
| Aroclor-1260-5       | (5) | 235694647  | 238382773  | 237905118  | 249801344  | 273831240  | 247123024  | 6     |
| Decachlorobiphenyl   |     | 1631798890 | 1658457733 | 1674372880 | 1797333600 | 2060619400 | 1764516501 | 10    |
| Tetrachloro-m-xylene |     | 2029594040 | 2049794933 | 2020131300 | 2090549040 | 2224542800 | 2082922423 | 4     |
| Aroclor-1242-1       | (1) | 55846456   | 57386199   | 58152954   | 53306796   | 62776280   | 57493737   | 6     |
| Aroclor-1242-2       | (2) | 82292663   | 85381189   | 86354650   | 80143412   | 87242420   | 84282867   | 4     |
| Aroclor-1242-3       | (3) | 42952122   | 44235299   | 45123908   | 42052092   | 45839360   | 44040556   | 4     |
| Aroclor-1242-4       | (4) | 41020216   | 42556281   | 43998486   | 41892464   | 46781360   | 43249761   | 5     |
| Aroclor-1242-5       | (5) | 54892740   | 55577543   | 56552748   | 54423788   | 70165220   | 58322408   | 11    |
| Decachlorobiphenyl   |     | 1561601020 | 1626952053 | 1628254580 | 1539884520 | 1715565200 | 1614451475 | 4     |
| Tetrachloro-m-xylene |     | 1982565500 | 2020394053 | 2002276380 | 1786858880 | 1908073400 | 1940033643 | 5     |
| Aroclor-1248-1       | (1) | 41569992   | 42480253   | 44301150   | 47253764   | 44774880   | 44076008   | 5     |
| Aroclor-1248-2       | (2) | 54299284   | 55737199   | 58630046   | 62694512   | 59777260   | 58227660   | 6     |
| Aroclor-1248-3       | (3) | 57411768   | 58802539   | 61381284   | 66231484   | 63093400   | 61384095   | 6     |
| Aroclor-1248-4       | (4) | 68531642   | 69993235   | 72796730   | 78273768   | 76330020   | 73185079   | 6     |
| Aroclor-1248-5       | (5) | 72351458   | 74116725   | 87512792   | 87540176   | 99280300   | 84160290   | 13    |
| Decachlorobiphenyl   |     | 1502488940 | 1528042440 | 1588582860 | 1696630800 | 1607749600 | 1584698928 | 5     |
| Tetrachloro-m-xylene |     | 1965268790 | 1970101347 | 1992281480 | 2022502760 | 1786610000 | 1947352875 | 5     |
| Aroclor-1254-1       | (1) | 108376052  | 112150224  | 116858404  | 122914228  | 130906240  | 118241030  | 8     |
| Aroclor-1254-2       | (2) | 93680624   | 97777644   | 103342076  | 111983640  | 133837560  | 108124309  | 15    |
| Aroclor-1254-3       | (3) | 152849964  | 157484645  | 163154906  | 168476080  | 178969780  | 164187075  | 6     |
| Aroclor-1254-4       | (4) | 96670996   | 100414949  | 103848536  | 107894352  | 110135520  | 103792871  | 5     |
| Aroclor-1254-5       | (5) | 131464389  | 136042208  | 140591856  | 145925060  | 149848540  | 140774411  | 5     |
| Decachlorobiphenyl   |     | 1541731170 | 1604172760 | 1652126120 | 1739157040 | 1816837400 | 1670804898 | 7     |
| Tetrachloro-m-xylene |     | 1985357600 | 2026009013 | 2048265880 | 2040842960 | 1966667000 | 2013428491 | 2     |
| Aroclor-1268-1       | (1) | 264556427  | 274894625  | 279316682  | 289369552  | 298553780  | 281338213  | 5     |

**CALIBRATION FACTOR OF INITIAL CALIBRATION**

|                      |     |            |            |            |            |            |            |   |
|----------------------|-----|------------|------------|------------|------------|------------|------------|---|
| Aroclor-1268-2       | (2) | 240850862  | 249890508  | 253434678  | 261757772  | 266698340  | 254526432  | 4 |
| Aroclor-1268-3       | (3) | 194377273  | 201896009  | 205304164  | 211895436  | 218632500  | 206421076  | 5 |
| Aroclor-1268-4       | (4) | 87883512   | 91300353   | 93471410   | 98130748   | 98806800   | 93918565   | 5 |
| Aroclor-1268-5       | (5) | 616587385  | 634409216  | 638557228  | 650798496  | 642345620  | 636539589  | 2 |
| Decachlorobiphenyl   |     | 2670092970 | 2765280333 | 2823888140 | 2956250080 | 3086571200 | 2860416545 | 6 |
| Tetrachloro-m-xylene |     | 1979033710 | 2058080493 | 2057299240 | 2063660600 | 2008890000 | 2033392809 | 2 |

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: CORE02

Lab Code: ACE SDG NO.: Q3741

Instrument ID: ECD\_P Date(s) Analyzed: 11/21/2025 11/21/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND     | AMOUNT<br>(ng) | PEAK | RT   | RT WINDOW |      | CALIBRATION<br>FACTOR |
|--------------|----------------|------|------|-----------|------|-----------------------|
|              |                |      |      | FROM      | TO   |                       |
| Aroclor-1221 | 500            | 1    | 4.82 | 4.72      | 4.92 | 25833800              |
|              |                | 2    | 4.90 | 4.80      | 5.00 | 19927200              |
|              |                | 3    | 4.98 | 4.88      | 5.08 | 55758000              |
|              |                | 4    | 0.00 |           |      | 0                     |
|              |                | 5    | 0.00 |           |      | 0                     |
| Aroclor-1232 | 500            | 1    | 4.98 | 4.88      | 5.08 | 43472600              |
|              |                | 2    | 5.51 | 5.41      | 5.61 | 22364600              |
|              |                | 3    | 5.80 | 5.70      | 5.90 | 43924600              |
|              |                | 4    | 5.96 | 5.86      | 6.06 | 22708400              |
|              |                | 5    | 6.04 | 5.94      | 6.14 | 16825100              |
| Aroclor-1262 | 500            | 1    | 8.20 | 8.10      | 8.30 | 98956600              |
|              |                | 2    | 8.52 | 8.42      | 8.62 | 172451000             |
|              |                | 3    | 8.84 | 8.74      | 8.94 | 125724000             |
|              |                | 4    | 8.93 | 8.83      | 9.03 | 99218000              |
|              |                | 5    | 9.59 | 9.49      | 9.69 | 71040600              |

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: CORE02

Lab Code: ACE SDG NO.: Q3741

Instrument ID: ECD\_P Date(s) Analyzed: 11/21/2025 11/21/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND     | AMOUNT<br>(ng) | PEAK | RT   | RT WINDOW |      | CALIBRATION<br>FACTOR |
|--------------|----------------|------|------|-----------|------|-----------------------|
|              |                |      |      | FROM      | TO   |                       |
| Aroclor-1221 | 500            | 1    | 3.98 | 3.88      | 4.08 | 26717800              |
|              |                | 2    | 4.07 | 3.97      | 4.17 | 19991600              |
|              |                | 3    | 4.15 | 4.05      | 4.25 | 60271800              |
|              |                | 4    | 0.00 |           |      | 0                     |
|              |                | 5    | 0.00 |           |      | 0                     |
| Aroclor-1232 | 500            | 1    | 4.15 | 4.05      | 4.25 | 47081000              |
|              |                | 2    | 4.87 | 4.77      | 4.97 | 46115400              |
|              |                | 3    | 5.05 | 4.95      | 5.15 | 24241400              |
|              |                | 4    | 5.13 | 5.03      | 5.23 | 21532600              |
|              |                | 5    | 5.30 | 5.20      | 5.40 | 23536200              |
| Aroclor-1262 | 500            | 1    | 6.88 | 6.78      | 6.98 | 133315000             |
|              |                | 2    | 7.14 | 7.04      | 7.24 | 119918000             |
|              |                | 3    | 7.66 | 7.56      | 7.76 | 98515200              |
|              |                | 4    | 7.72 | 7.62      | 7.82 | 174485000             |
|              |                | 5    | 8.22 | 8.12      | 8.32 | 80357800              |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG NO.: Q3741

Continuing Calib Date: 12/03/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 09:34

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.47       | 5.49      | 5.39      | 5.59  | 0.02       |
| Aroclor-1016-2       | (2) | 5.50       | 5.51      | 5.41      | 5.61  | 0.01       |
| Aroclor-1016-3       | (3) | 5.56       | 5.57      | 5.47      | 5.67  | 0.01       |
| Aroclor-1016-4       | (4) | 5.65       | 5.67      | 5.57      | 5.77  | 0.02       |
| Aroclor-1016-5       | (5) | 5.95       | 5.96      | 5.86      | 6.06  | 0.01       |
| Aroclor-1260-1       | (1) | 7.06       | 7.07      | 6.97      | 7.17  | 0.01       |
| Aroclor-1260-2       | (2) | 7.31       | 7.33      | 7.23      | 7.43  | 0.02       |
| Aroclor-1260-3       | (3) | 7.67       | 7.68      | 7.58      | 7.78  | 0.01       |
| Aroclor-1260-4       | (4) | 7.89       | 7.90      | 7.80      | 8.00  | 0.01       |
| Aroclor-1260-5       | (5) | 8.20       | 8.21      | 8.11      | 8.31  | 0.01       |
| Tetrachloro-m-xylene |     | 4.34       | 4.35      | 4.25      | 4.45  | 0.01       |
| Decachlorobiphenyl   |     | 9.91       | 9.93      | 9.83      | 10.03 | 0.02       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG NO.: Q3741

Continuing Calib Date: 12/03/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 09:34

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.71       | 4.73      | 4.63      | 4.83 | 0.02       |
| Aroclor-1016-2       | (2) | 4.73       | 4.75      | 4.65      | 4.85 | 0.02       |
| Aroclor-1016-3       | (3) | 4.91       | 4.92      | 4.82      | 5.02 | 0.01       |
| Aroclor-1016-4       | (4) | 4.95       | 4.96      | 4.86      | 5.06 | 0.01       |
| Aroclor-1016-5       | (5) | 5.16       | 5.18      | 5.08      | 5.28 | 0.02       |
| Aroclor-1260-1       | (1) | 6.19       | 6.21      | 6.11      | 6.31 | 0.02       |
| Aroclor-1260-2       | (2) | 6.38       | 6.39      | 6.29      | 6.49 | 0.01       |
| Aroclor-1260-3       | (3) | 6.53       | 6.55      | 6.45      | 6.65 | 0.02       |
| Aroclor-1260-4       | (4) | 7.00       | 7.02      | 6.92      | 7.12 | 0.02       |
| Aroclor-1260-5       | (5) | 7.24       | 7.26      | 7.16      | 7.36 | 0.02       |
| Tetrachloro-m-xylene |     | 3.64       | 3.65      | 3.55      | 3.75 | 0.01       |
| Decachlorobiphenyl   |     | 8.62       | 8.64      | 8.54      | 8.74 | 0.02       |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CORE02  
**Lab Code:** ACE **SDG NO.:** Q3741  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

**Client Sample No.:** CCAL01 **Date Analyzed:** 12/03/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO115552.D **Time Analyzed:** 09:34

| COMPOUND             | RT    | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|--------|--------------------|-------------------|------|
|                      |       | FROM      | TO     |                    |                   |      |
| Aroclor-1016-1       | 5.474 | 5.387     | 5.587  | 470.340            | 500.000           | -5.9 |
| Aroclor-1016-2       | 5.497 | 5.408     | 5.608  | 505.670            | 500.000           | 1.1  |
| Aroclor-1016-3       | 5.559 | 5.470     | 5.670  | 559.330            | 500.000           | 11.9 |
| Aroclor-1016-4       | 5.654 | 5.566     | 5.766  | 541.740            | 500.000           | 8.3  |
| Aroclor-1016-5       | 5.946 | 5.858     | 6.058  | 522.030            | 500.000           | 4.4  |
| Aroclor-1260-1       | 7.060 | 6.971     | 7.171  | 518.800            | 500.000           | 3.8  |
| Aroclor-1260-2       | 7.313 | 7.225     | 7.425  | 539.800            | 500.000           | 8.0  |
| Aroclor-1260-3       | 7.671 | 7.582     | 7.782  | 522.570            | 500.000           | 4.5  |
| Aroclor-1260-4       | 7.893 | 7.803     | 8.003  | 501.910            | 500.000           | 0.4  |
| Aroclor-1260-5       | 8.199 | 8.110     | 8.310  | 491.180            | 500.000           | -1.8 |
| Decachlorobiphenyl   | 9.911 | 9.830     | 10.030 | 58.250             | 50.000            | 16.5 |
| Tetrachloro-m-xylene | 4.336 | 4.248     | 4.448  | 56.580             | 50.000            | 13.2 |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CORE02  
**Lab Code:** ACE **SDG NO.:** Q3741  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

**Client Sample No.:** CCAL01 **Date Analyzed:** 12/03/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO115552.D **Time Analyzed:** 09:34

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.714 | 4.628     | 4.828 | 483.410            | 500.000           | -3.3  |
| Aroclor-1016-2       | 4.734 | 4.647     | 4.847 | 463.460            | 500.000           | -7.3  |
| Aroclor-1016-3       | 4.908 | 4.821     | 5.021 | 430.720            | 500.000           | -13.9 |
| Aroclor-1016-4       | 4.950 | 4.864     | 5.064 | 406.600            | 500.000           | -18.7 |
| Aroclor-1016-5       | 5.162 | 5.076     | 5.276 | 538.120            | 500.000           | 7.6   |
| Aroclor-1260-1       | 6.191 | 6.106     | 6.306 | 441.910            | 500.000           | -11.6 |
| Aroclor-1260-2       | 6.379 | 6.294     | 6.494 | 409.120            | 500.000           | -18.2 |
| Aroclor-1260-3       | 6.531 | 6.446     | 6.646 | 416.000            | 500.000           | -16.8 |
| Aroclor-1260-4       | 7.002 | 6.917     | 7.117 | 410.560            | 500.000           | -17.9 |
| Aroclor-1260-5       | 7.243 | 7.158     | 7.358 | 412.150            | 500.000           | -17.6 |
| Decachlorobiphenyl   | 8.623 | 8.541     | 8.741 | 53.610             | 50.000            | 7.2   |
| Tetrachloro-m-xylene | 3.640 | 3.548     | 3.748 | 52.170             | 50.000            | 4.3   |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG NO.: Q3741

Continuing Calib Date: 12/03/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 14:22

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.48       | 5.49      | 5.39      | 5.59  | 0.02       |
| Aroclor-1016-2       | (2) | 5.50       | 5.51      | 5.41      | 5.61  | 0.01       |
| Aroclor-1016-3       | (3) | 5.56       | 5.57      | 5.47      | 5.67  | 0.01       |
| Aroclor-1016-4       | (4) | 5.65       | 5.67      | 5.57      | 5.77  | 0.02       |
| Aroclor-1016-5       | (5) | 5.95       | 5.96      | 5.86      | 6.06  | 0.01       |
| Aroclor-1260-1       | (1) | 7.06       | 7.07      | 6.97      | 7.17  | 0.01       |
| Aroclor-1260-2       | (2) | 7.31       | 7.33      | 7.23      | 7.43  | 0.02       |
| Aroclor-1260-3       | (3) | 7.67       | 7.68      | 7.58      | 7.78  | 0.01       |
| Aroclor-1260-4       | (4) | 7.89       | 7.90      | 7.80      | 8.00  | 0.01       |
| Aroclor-1260-5       | (5) | 8.20       | 8.21      | 8.11      | 8.31  | 0.01       |
| Tetrachloro-m-xylene |     | 4.34       | 4.35      | 4.25      | 4.45  | 0.01       |
| Decachlorobiphenyl   |     | 9.91       | 9.93      | 9.83      | 10.03 | 0.02       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG NO.: Q3741

Continuing Calib Date: 12/03/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 14:22

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.72       | 4.73      | 4.63      | 4.83 | 0.01       |
| Aroclor-1016-2       | (2) | 4.73       | 4.75      | 4.65      | 4.85 | 0.02       |
| Aroclor-1016-3       | (3) | 4.91       | 4.92      | 4.82      | 5.02 | 0.01       |
| Aroclor-1016-4       | (4) | 4.95       | 4.96      | 4.86      | 5.06 | 0.01       |
| Aroclor-1016-5       | (5) | 5.16       | 5.18      | 5.08      | 5.28 | 0.02       |
| Aroclor-1260-1       | (1) | 6.19       | 6.21      | 6.11      | 6.31 | 0.02       |
| Aroclor-1260-2       | (2) | 6.38       | 6.39      | 6.29      | 6.49 | 0.01       |
| Aroclor-1260-3       | (3) | 6.53       | 6.55      | 6.45      | 6.65 | 0.02       |
| Aroclor-1260-4       | (4) | 7.00       | 7.02      | 6.92      | 7.12 | 0.02       |
| Aroclor-1260-5       | (5) | 7.24       | 7.26      | 7.16      | 7.36 | 0.02       |
| Tetrachloro-m-xylene |     | 3.64       | 3.65      | 3.55      | 3.75 | 0.01       |
| Decachlorobiphenyl   |     | 8.62       | 8.64      | 8.54      | 8.74 | 0.02       |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CORE02  
**Lab Code:** ACE **SDG NO.:** Q3741  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

**Client Sample No.:** CCAL02 **Date Analyzed:** 12/03/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO115565.D **Time Analyzed:** 14:22

| COMPOUND             | RT    | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|--------|--------------------|-------------------|------|
|                      |       | FROM      | TO     |                    |                   |      |
| Aroclor-1016-1       | 5.475 | 5.387     | 5.587  | 486.940            | 500.000           | -2.6 |
| Aroclor-1016-2       | 5.496 | 5.408     | 5.608  | 462.390            | 500.000           | -7.5 |
| Aroclor-1016-3       | 5.557 | 5.470     | 5.670  | 489.160            | 500.000           | -2.2 |
| Aroclor-1016-4       | 5.654 | 5.566     | 5.766  | 479.290            | 500.000           | -4.1 |
| Aroclor-1016-5       | 5.945 | 5.858     | 6.058  | 473.190            | 500.000           | -5.4 |
| Aroclor-1260-1       | 7.059 | 6.971     | 7.171  | 504.760            | 500.000           | 1.0  |
| Aroclor-1260-2       | 7.313 | 7.225     | 7.425  | 514.490            | 500.000           | 2.9  |
| Aroclor-1260-3       | 7.670 | 7.582     | 7.782  | 508.230            | 500.000           | 1.6  |
| Aroclor-1260-4       | 7.893 | 7.803     | 8.003  | 477.910            | 500.000           | -4.4 |
| Aroclor-1260-5       | 8.199 | 8.110     | 8.310  | 468.510            | 500.000           | -6.3 |
| Decachlorobiphenyl   | 9.914 | 9.830     | 10.030 | 57.970             | 50.000            | 15.9 |
| Tetrachloro-m-xylene | 4.336 | 4.248     | 4.448  | 57.090             | 50.000            | 14.2 |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CORE02  
**Lab Code:** ACE **SDG NO.:** Q3741  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

**Client Sample No.:** CCAL02 **Date Analyzed:** 12/03/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO115565.D **Time Analyzed:** 14:22

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.715 | 4.628     | 4.828 | 488.550            | 500.000           | -2.3  |
| Aroclor-1016-2       | 4.734 | 4.647     | 4.847 | 467.850            | 500.000           | -6.4  |
| Aroclor-1016-3       | 4.908 | 4.821     | 5.021 | 469.280            | 500.000           | -6.1  |
| Aroclor-1016-4       | 4.949 | 4.864     | 5.064 | 415.460            | 500.000           | -16.9 |
| Aroclor-1016-5       | 5.162 | 5.076     | 5.276 | 547.590            | 500.000           | 9.5   |
| Aroclor-1260-1       | 6.191 | 6.106     | 6.306 | 431.240            | 500.000           | -13.8 |
| Aroclor-1260-2       | 6.379 | 6.294     | 6.494 | 402.390            | 500.000           | -19.5 |
| Aroclor-1260-3       | 6.531 | 6.446     | 6.646 | 394.340            | 500.000           | -21.1 |
| Aroclor-1260-4       | 7.002 | 6.917     | 7.117 | 392.140            | 500.000           | -21.6 |
| Aroclor-1260-5       | 7.244 | 7.158     | 7.358 | 391.850            | 500.000           | -21.6 |
| Decachlorobiphenyl   | 8.624 | 8.541     | 8.741 | 50.710             | 50.000            | 1.4   |
| Tetrachloro-m-xylene | 3.640 | 3.548     | 3.748 | 53.110             | 50.000            | 6.2   |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG NO.: Q3741

Continuing Calib Date: 12/03/2025

Initial Calibration Date(s): 11/21/2025

11/21/2025

Continuing Calib Time: 10:05

Initial Calibration Time(s): 10:44

19:39

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.77       | 5.77      | 5.67      | 5.87  | 0.00       |
| Aroclor-1016-2       | (2) | 5.79       | 5.79      | 5.69      | 5.89  | 0.00       |
| Aroclor-1016-3       | (3) | 5.85       | 5.85      | 5.75      | 5.95  | 0.00       |
| Aroclor-1016-4       | (4) | 5.95       | 5.95      | 5.85      | 6.05  | 0.00       |
| Aroclor-1016-5       | (5) | 6.24       | 6.24      | 6.14      | 6.34  | 0.00       |
| Aroclor-1260-1       | (1) | 7.36       | 7.36      | 7.26      | 7.46  | 0.00       |
| Aroclor-1260-2       | (2) | 7.61       | 7.61      | 7.51      | 7.71  | 0.00       |
| Aroclor-1260-3       | (3) | 7.97       | 7.97      | 7.87      | 8.07  | 0.00       |
| Aroclor-1260-4       | (4) | 8.20       | 8.19      | 8.09      | 8.29  | -0.01      |
| Aroclor-1260-5       | (5) | 8.52       | 8.52      | 8.42      | 8.62  | 0.00       |
| Tetrachloro-m-xylene |     | 4.62       | 4.61      | 4.51      | 4.71  | -0.01      |
| Decachlorobiphenyl   |     | 10.37      | 10.37     | 10.27     | 10.47 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG NO.: Q3741

Continuing Calib Date: 12/03/2025

Initial Calibration Date(s): 11/21/2025

11/21/2025

Continuing Calib Time: 10:05

Initial Calibration Time(s): 10:44

19:39

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.85       | 4.85      | 4.75      | 4.95 | 0.00       |
| Aroclor-1016-2       | (2) | 4.87       | 4.87      | 4.77      | 4.97 | 0.00       |
| Aroclor-1016-3       | (3) | 5.04       | 5.04      | 4.94      | 5.14 | 0.00       |
| Aroclor-1016-4       | (4) | 5.09       | 5.08      | 4.98      | 5.18 | -0.01      |
| Aroclor-1016-5       | (5) | 5.30       | 5.30      | 5.20      | 5.40 | 0.00       |
| Aroclor-1260-1       | (1) | 6.33       | 6.32      | 6.22      | 6.42 | -0.01      |
| Aroclor-1260-2       | (2) | 6.51       | 6.51      | 6.41      | 6.61 | 0.00       |
| Aroclor-1260-3       | (3) | 6.67       | 6.66      | 6.56      | 6.76 | 0.00       |
| Aroclor-1260-4       | (4) | 7.13       | 7.13      | 7.03      | 7.23 | 0.00       |
| Aroclor-1260-5       | (5) | 7.37       | 7.37      | 7.27      | 7.47 | 0.00       |
| Tetrachloro-m-xylene |     | 3.77       | 3.77      | 3.67      | 3.87 | 0.00       |
| Decachlorobiphenyl   |     | 8.77       | 8.77      | 8.67      | 8.87 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CORE02  
**Lab Code:** ACE **SDG NO.:** Q3741  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

**Client Sample No.:** CCAL03 **Date Analyzed:** 12/03/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP076744.D **Time Analyzed:** 10:05

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D  |
|----------------------|--------|-----------|--------|--------------------|-------------------|-----|
|                      |        | FROM      | TO     |                    |                   |     |
| Aroclor-1016-1       | 5.770  | 5.667     | 5.867  | 510.750            | 500.000           | 2.2 |
| Aroclor-1016-2       | 5.791  | 5.687     | 5.887  | 528.080            | 500.000           | 5.6 |
| Aroclor-1016-3       | 5.854  | 5.750     | 5.950  | 523.240            | 500.000           | 4.6 |
| Aroclor-1016-4       | 5.951  | 5.847     | 6.047  | 524.670            | 500.000           | 4.9 |
| Aroclor-1016-5       | 6.243  | 6.140     | 6.340  | 522.370            | 500.000           | 4.5 |
| Aroclor-1260-1       | 7.360  | 7.256     | 7.456  | 518.740            | 500.000           | 3.7 |
| Aroclor-1260-2       | 7.613  | 7.509     | 7.709  | 506.520            | 500.000           | 1.3 |
| Aroclor-1260-3       | 7.971  | 7.867     | 8.067  | 509.450            | 500.000           | 1.9 |
| Aroclor-1260-4       | 8.197  | 8.093     | 8.293  | 504.520            | 500.000           | 0.9 |
| Aroclor-1260-5       | 8.521  | 8.416     | 8.616  | 503.010            | 500.000           | 0.6 |
| Decachlorobiphenyl   | 10.370 | 10.265    | 10.465 | 51.220             | 50.000            | 2.4 |
| Tetrachloro-m-xylene | 4.618  | 4.514     | 4.714  | 53.610             | 50.000            | 7.2 |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CORE02  
**Lab Code:** ACE **SDG NO.:** Q3741  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

**Client Sample No.:** CCAL03 **Date Analyzed:** 12/03/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP076744.D **Time Analyzed:** 10:05

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.851 | 4.750     | 4.950 | 567.100            | 500.000           | 13.4 |
| Aroclor-1016-2       | 4.869 | 4.767     | 4.967 | 589.400            | 500.000           | 17.9 |
| Aroclor-1016-3       | 5.044 | 4.942     | 5.142 | 581.550            | 500.000           | 16.3 |
| Aroclor-1016-4       | 5.086 | 4.984     | 5.184 | 541.860            | 500.000           | 8.4  |
| Aroclor-1016-5       | 5.298 | 5.197     | 5.397 | 629.900            | 500.000           | 26.0 |
| Aroclor-1260-1       | 6.326 | 6.224     | 6.424 | 582.440            | 500.000           | 16.5 |
| Aroclor-1260-2       | 6.513 | 6.412     | 6.612 | 587.660            | 500.000           | 17.5 |
| Aroclor-1260-3       | 6.665 | 6.564     | 6.764 | 565.210            | 500.000           | 13.0 |
| Aroclor-1260-4       | 7.134 | 7.033     | 7.233 | 570.270            | 500.000           | 14.1 |
| Aroclor-1260-5       | 7.374 | 7.273     | 7.473 | 563.180            | 500.000           | 12.6 |
| Decachlorobiphenyl   | 8.766 | 8.666     | 8.866 | 54.980             | 50.000            | 10.0 |
| Tetrachloro-m-xylene | 3.773 | 3.671     | 3.871 | 58.990             | 50.000            | 18.0 |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG NO.: Q3741

Continuing Calib Date: 12/03/2025

Initial Calibration Date(s): 11/21/2025

11/21/2025

Continuing Calib Time: 17:01

Initial Calibration Time(s): 10:44

19:39

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.77       | 5.77      | 5.67      | 5.87  | 0.00       |
| Aroclor-1016-2       | (2) | 5.79       | 5.79      | 5.69      | 5.89  | 0.00       |
| Aroclor-1016-3       | (3) | 5.85       | 5.85      | 5.75      | 5.95  | 0.00       |
| Aroclor-1016-4       | (4) | 5.95       | 5.95      | 5.85      | 6.05  | 0.00       |
| Aroclor-1016-5       | (5) | 6.24       | 6.24      | 6.14      | 6.34  | 0.00       |
| Aroclor-1260-1       | (1) | 7.36       | 7.36      | 7.26      | 7.46  | 0.00       |
| Aroclor-1260-2       | (2) | 7.61       | 7.61      | 7.51      | 7.71  | 0.00       |
| Aroclor-1260-3       | (3) | 7.97       | 7.97      | 7.87      | 8.07  | 0.00       |
| Aroclor-1260-4       | (4) | 8.20       | 8.19      | 8.09      | 8.29  | -0.01      |
| Aroclor-1260-5       | (5) | 8.52       | 8.52      | 8.42      | 8.62  | 0.00       |
| Tetrachloro-m-xylene |     | 4.62       | 4.61      | 4.51      | 4.71  | -0.01      |
| Decachlorobiphenyl   |     | 10.37      | 10.37     | 10.27     | 10.47 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG NO.: Q3741

Continuing Calib Date: 12/03/2025

Initial Calibration Date(s): 11/21/2025

11/21/2025

Continuing Calib Time: 17:01

Initial Calibration Time(s): 10:44

19:39

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.85       | 4.85      | 4.75      | 4.95 | 0.00       |
| Aroclor-1016-2       | (2) | 4.87       | 4.87      | 4.77      | 4.97 | 0.00       |
| Aroclor-1016-3       | (3) | 5.04       | 5.04      | 4.94      | 5.14 | 0.00       |
| Aroclor-1016-4       | (4) | 5.09       | 5.08      | 4.98      | 5.18 | -0.01      |
| Aroclor-1016-5       | (5) | 5.30       | 5.30      | 5.20      | 5.40 | 0.00       |
| Aroclor-1260-1       | (1) | 6.33       | 6.32      | 6.22      | 6.42 | -0.01      |
| Aroclor-1260-2       | (2) | 6.51       | 6.51      | 6.41      | 6.61 | 0.00       |
| Aroclor-1260-3       | (3) | 6.67       | 6.66      | 6.56      | 6.76 | 0.00       |
| Aroclor-1260-4       | (4) | 7.13       | 7.13      | 7.03      | 7.23 | 0.00       |
| Aroclor-1260-5       | (5) | 7.38       | 7.37      | 7.27      | 7.47 | -0.01      |
| Tetrachloro-m-xylene |     | 3.77       | 3.77      | 3.67      | 3.87 | 0.00       |
| Decachlorobiphenyl   |     | 8.77       | 8.77      | 8.67      | 8.87 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CORE02  
**Lab Code:** ACE **SDG NO.:** Q3741  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

**Client Sample No.:** CCAL04 **Date Analyzed:** 12/03/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP076758.D **Time Analyzed:** 17:01

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|--------|-----------|--------|--------------------|-------------------|------|
|                      |        | FROM      | TO     |                    |                   |      |
| Aroclor-1016-1       | 5.771  | 5.667     | 5.867  | 498.120            | 500.000           | -0.4 |
| Aroclor-1016-2       | 5.792  | 5.687     | 5.887  | 504.450            | 500.000           | 0.9  |
| Aroclor-1016-3       | 5.854  | 5.750     | 5.950  | 500.530            | 500.000           | 0.1  |
| Aroclor-1016-4       | 5.951  | 5.847     | 6.047  | 501.070            | 500.000           | 0.2  |
| Aroclor-1016-5       | 6.244  | 6.140     | 6.340  | 510.060            | 500.000           | 2.0  |
| Aroclor-1260-1       | 7.361  | 7.256     | 7.456  | 500.150            | 500.000           | 0.0  |
| Aroclor-1260-2       | 7.613  | 7.509     | 7.709  | 489.260            | 500.000           | -2.1 |
| Aroclor-1260-3       | 7.972  | 7.867     | 8.067  | 493.290            | 500.000           | -1.3 |
| Aroclor-1260-4       | 8.198  | 8.093     | 8.293  | 488.420            | 500.000           | -2.3 |
| Aroclor-1260-5       | 8.521  | 8.416     | 8.616  | 488.930            | 500.000           | -2.2 |
| Decachlorobiphenyl   | 10.369 | 10.265    | 10.465 | 49.550             | 50.000            | -0.9 |
| Tetrachloro-m-xylene | 4.618  | 4.514     | 4.714  | 50.000             | 50.000            | 0.0  |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CORE02  
**Lab Code:** ACE **SDG NO.:** Q3741  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

**Client Sample No.:** CCAL04 **Date Analyzed:** 12/03/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP076758.D **Time Analyzed:** 17:01

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.852 | 4.750     | 4.950 | 508.130            | 500.000           | 1.6  |
| Aroclor-1016-2       | 4.869 | 4.767     | 4.967 | 516.310            | 500.000           | 3.3  |
| Aroclor-1016-3       | 5.044 | 4.942     | 5.142 | 515.230            | 500.000           | 3.0  |
| Aroclor-1016-4       | 5.087 | 4.984     | 5.184 | 484.990            | 500.000           | -3.0 |
| Aroclor-1016-5       | 5.299 | 5.197     | 5.397 | 552.140            | 500.000           | 10.4 |
| Aroclor-1260-1       | 6.326 | 6.224     | 6.424 | 512.690            | 500.000           | 2.5  |
| Aroclor-1260-2       | 6.513 | 6.412     | 6.612 | 519.620            | 500.000           | 3.9  |
| Aroclor-1260-3       | 6.665 | 6.564     | 6.764 | 497.970            | 500.000           | -0.4 |
| Aroclor-1260-4       | 7.134 | 7.033     | 7.233 | 503.740            | 500.000           | 0.7  |
| Aroclor-1260-5       | 7.375 | 7.273     | 7.473 | 500.140            | 500.000           | 0.0  |
| Decachlorobiphenyl   | 8.767 | 8.666     | 8.866 | 49.010             | 50.000            | -2.0 |
| Tetrachloro-m-xylene | 3.774 | 3.671     | 3.871 | 51.820             | 50.000            | 3.6  |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG NO.: Q3741

Continuing Calib Date: 12/03/2025

Initial Calibration Date(s): 11/21/2025

11/21/2025

Continuing Calib Time: 22:14

Initial Calibration Time(s): 10:44

19:39

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.77       | 5.77      | 5.67      | 5.87  | 0.00       |
| Aroclor-1016-2       | (2) | 5.79       | 5.79      | 5.69      | 5.89  | 0.00       |
| Aroclor-1016-3       | (3) | 5.85       | 5.85      | 5.75      | 5.95  | 0.00       |
| Aroclor-1016-4       | (4) | 5.95       | 5.95      | 5.85      | 6.05  | 0.00       |
| Aroclor-1016-5       | (5) | 6.24       | 6.24      | 6.14      | 6.34  | 0.00       |
| Aroclor-1260-1       | (1) | 7.36       | 7.36      | 7.26      | 7.46  | 0.00       |
| Aroclor-1260-2       | (2) | 7.61       | 7.61      | 7.51      | 7.71  | 0.00       |
| Aroclor-1260-3       | (3) | 7.97       | 7.97      | 7.87      | 8.07  | 0.00       |
| Aroclor-1260-4       | (4) | 8.19       | 8.19      | 8.09      | 8.29  | 0.00       |
| Aroclor-1260-5       | (5) | 8.52       | 8.52      | 8.42      | 8.62  | 0.00       |
| Tetrachloro-m-xylene |     | 4.61       | 4.61      | 4.51      | 4.71  | 0.00       |
| Decachlorobiphenyl   |     | 10.36      | 10.37     | 10.27     | 10.47 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG NO.: Q3741

Continuing Calib Date: 12/03/2025

Initial Calibration Date(s): 11/21/2025

11/21/2025

Continuing Calib Time: 22:14

Initial Calibration Time(s): 10:44

19:39

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.85       | 4.85      | 4.75      | 4.95 | 0.00       |
| Aroclor-1016-2       | (2) | 4.87       | 4.87      | 4.77      | 4.97 | 0.01       |
| Aroclor-1016-3       | (3) | 5.04       | 5.04      | 4.94      | 5.14 | 0.00       |
| Aroclor-1016-4       | (4) | 5.08       | 5.08      | 4.98      | 5.18 | 0.00       |
| Aroclor-1016-5       | (5) | 5.29       | 5.30      | 5.20      | 5.40 | 0.01       |
| Aroclor-1260-1       | (1) | 6.32       | 6.32      | 6.22      | 6.42 | 0.00       |
| Aroclor-1260-2       | (2) | 6.51       | 6.51      | 6.41      | 6.61 | 0.00       |
| Aroclor-1260-3       | (3) | 6.66       | 6.66      | 6.56      | 6.76 | 0.00       |
| Aroclor-1260-4       | (4) | 7.13       | 7.13      | 7.03      | 7.23 | 0.00       |
| Aroclor-1260-5       | (5) | 7.37       | 7.37      | 7.27      | 7.47 | 0.00       |
| Tetrachloro-m-xylene |     | 3.77       | 3.77      | 3.67      | 3.87 | 0.00       |
| Decachlorobiphenyl   |     | 8.76       | 8.77      | 8.67      | 8.87 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CORE02  
**Lab Code:** ACE **SDG NO.:** Q3741  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

**Client Sample No.:** CCAL05 **Date Analyzed:** 12/03/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP076773.D **Time Analyzed:** 22:14

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|--------|-----------|--------|--------------------|-------------------|------|
|                      |        | FROM      | TO     |                    |                   |      |
| Aroclor-1016-1       | 5.766  | 5.667     | 5.867  | 465.250            | 500.000           | -7.0 |
| Aroclor-1016-2       | 5.788  | 5.687     | 5.887  | 470.570            | 500.000           | -5.9 |
| Aroclor-1016-3       | 5.850  | 5.750     | 5.950  | 468.280            | 500.000           | -6.3 |
| Aroclor-1016-4       | 5.947  | 5.847     | 6.047  | 470.850            | 500.000           | -5.8 |
| Aroclor-1016-5       | 6.240  | 6.140     | 6.340  | 471.200            | 500.000           | -5.8 |
| Aroclor-1260-1       | 7.357  | 7.256     | 7.456  | 472.690            | 500.000           | -5.5 |
| Aroclor-1260-2       | 7.610  | 7.509     | 7.709  | 467.230            | 500.000           | -6.6 |
| Aroclor-1260-3       | 7.967  | 7.867     | 8.067  | 460.200            | 500.000           | -8.0 |
| Aroclor-1260-4       | 8.193  | 8.093     | 8.293  | 462.150            | 500.000           | -7.6 |
| Aroclor-1260-5       | 8.517  | 8.416     | 8.616  | 459.120            | 500.000           | -8.2 |
| Decachlorobiphenyl   | 10.364 | 10.265    | 10.465 | 46.590             | 50.000            | -6.8 |
| Tetrachloro-m-xylene | 4.614  | 4.514     | 4.714  | 46.540             | 50.000            | -6.9 |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CORE02  
**Lab Code:** ACE **SDG NO.:** Q3741  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

**Client Sample No.:** CCAL05 **Date Analyzed:** 12/03/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP076773.D **Time Analyzed:** 22:14

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.848 | 4.750     | 4.950 | 466.600            | 500.000           | -6.7  |
| Aroclor-1016-2       | 4.865 | 4.767     | 4.967 | 484.680            | 500.000           | -3.1  |
| Aroclor-1016-3       | 5.040 | 4.942     | 5.142 | 480.910            | 500.000           | -3.8  |
| Aroclor-1016-4       | 5.083 | 4.984     | 5.184 | 444.350            | 500.000           | -11.1 |
| Aroclor-1016-5       | 5.294 | 5.197     | 5.397 | 532.260            | 500.000           | 6.5   |
| Aroclor-1260-1       | 6.322 | 6.224     | 6.424 | 475.390            | 500.000           | -4.9  |
| Aroclor-1260-2       | 6.510 | 6.412     | 6.612 | 486.710            | 500.000           | -2.7  |
| Aroclor-1260-3       | 6.661 | 6.564     | 6.764 | 464.200            | 500.000           | -7.2  |
| Aroclor-1260-4       | 7.131 | 7.033     | 7.233 | 468.820            | 500.000           | -6.2  |
| Aroclor-1260-5       | 7.371 | 7.273     | 7.473 | 466.280            | 500.000           | -6.7  |
| Decachlorobiphenyl   | 8.763 | 8.666     | 8.866 | 46.330             | 50.000            | -7.3  |
| Tetrachloro-m-xylene | 3.770 | 3.671     | 3.871 | 48.400             | 50.000            | -3.2  |

## Analytical Sequence

|                                                  |                                                           |
|--------------------------------------------------|-----------------------------------------------------------|
| Client: Core Environmental Consultants and Servi | SDG No.: Q3741                                            |
| Project: Building 50 Probe Investigation         | Instrument ID: ECD_O                                      |
| GC Column: ZB-MR1                                | ID: 0.32 (mm) Inst. Calib. Date(s): 11/04/2025 11/04/2025 |

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,  
AND STANDARDS IS GIVEN BELOW:

| CLIENT ID       | LAB<br>SAMPLE ID | DATE<br>ANALYZED | TIME<br>ANALYZED | DATAFILE   | DCB<br>RT # | TCX<br>RT # |
|-----------------|------------------|------------------|------------------|------------|-------------|-------------|
| IBLK            | IBLK             | 11/04/2025       | 09:21            | PO114870.D | 9.93        | 4.35        |
| AR1660ICC1000   | AR1660ICC1000    | 11/04/2025       | 09:39            | PO114871.D | 9.93        | 4.35        |
| AR1660ICC750    | AR1660ICC750     | 11/04/2025       | 09:58            | PO114872.D | 9.93        | 4.35        |
| AR1660ICC500    | AR1660ICC500     | 11/04/2025       | 10:16            | PO114873.D | 9.93        | 4.35        |
| AR1660ICC250    | AR1660ICC250     | 11/04/2025       | 10:34            | PO114874.D | 9.93        | 4.35        |
| AR1660ICC050    | AR1660ICC050     | 11/04/2025       | 10:53            | PO114875.D | 9.93        | 4.35        |
| AR1221ICC500    | AR1221ICC500     | 11/04/2025       | 11:11            | PO114876.D | 9.93        | 4.35        |
| AR1232ICC500    | AR1232ICC500     | 11/04/2025       | 11:30            | PO114877.D | 9.93        | 4.35        |
| AR1242ICC1000   | AR1242ICC1000    | 11/04/2025       | 11:48            | PO114878.D | 9.93        | 4.35        |
| AR1242ICC750    | AR1242ICC750     | 11/04/2025       | 12:06            | PO114879.D | 9.93        | 4.35        |
| AR1242ICC500    | AR1242ICC500     | 11/04/2025       | 12:25            | PO114880.D | 9.93        | 4.35        |
| AR1242ICC250    | AR1242ICC250     | 11/04/2025       | 12:42            | PO114881.D | 9.93        | 4.35        |
| AR1242ICC050    | AR1242ICC050     | 11/04/2025       | 13:00            | PO114882.D | 9.93        | 4.35        |
| AR1248ICC1000   | AR1248ICC1000    | 11/04/2025       | 13:19            | PO114883.D | 9.93        | 4.35        |
| AR1248ICC750    | AR1248ICC750     | 11/04/2025       | 13:37            | PO114884.D | 9.93        | 4.35        |
| AR1248ICC500    | AR1248ICC500     | 11/04/2025       | 13:56            | PO114885.D | 9.93        | 4.35        |
| AR1248ICC250    | AR1248ICC250     | 11/04/2025       | 14:14            | PO114886.D | 9.93        | 4.35        |
| AR1248ICC050    | AR1248ICC050     | 11/04/2025       | 14:33            | PO114887.D | 9.93        | 4.35        |
| AR1254ICC1000   | AR1254ICC1000    | 11/04/2025       | 14:51            | PO114888.D | 9.93        | 4.35        |
| AR1254ICC750    | AR1254ICC750     | 11/04/2025       | 15:09            | PO114889.D | 9.93        | 4.35        |
| AR1254ICC500    | AR1254ICC500     | 11/04/2025       | 15:28            | PO114890.D | 9.93        | 4.35        |
| AR1254ICC250    | AR1254ICC250     | 11/04/2025       | 16:04            | PO114891.D | 9.93        | 4.35        |
| AR1254ICC050    | AR1254ICC050     | 11/04/2025       | 16:23            | PO114892.D | 9.93        | 4.35        |
| AR1262ICC500    | AR1262ICC500     | 11/04/2025       | 17:00            | PO114893.D | 9.95        | 4.37        |
| AR1268ICC1000   | AR1268ICC1000    | 11/04/2025       | 17:18            | PO114894.D | 9.93        | 4.35        |
| AR1268ICC750    | AR1268ICC750     | 11/04/2025       | 17:36            | PO114895.D | 9.93        | 4.35        |
| AR1268ICC500    | AR1268ICC500     | 11/04/2025       | 17:55            | PO114896.D | 9.93        | 4.35        |
| AR1268ICC250    | AR1268ICC250     | 11/04/2025       | 18:13            | PO114897.D | 9.93        | 4.35        |
| AR1268ICC050    | AR1268ICC050     | 11/04/2025       | 18:31            | PO114898.D | 9.93        | 4.35        |
| AR1660CCC500    | AR1660CCC500     | 12/03/2025       | 09:34            | PO115552.D | 9.91        | 4.34        |
| IBLK            | IBLK             | 12/03/2025       | 10:48            | PO115556.D | 9.91        | 4.34        |
| B50-SP3-111925  | Q3741-01         | 12/03/2025       | 12:50            | PO115561.D | 9.91        | 4.34        |
| B50-SP13-112025 | Q3741-02         | 12/03/2025       | 13:08            | PO115562.D | 9.92        | 4.34        |
| B50-SP14-112025 | Q3741-03         | 12/03/2025       | 13:27            | PO115563.D | 9.91        | 4.34        |
| B50-SP9-112425  | Q3741-04         | 12/03/2025       | 13:45            | PO115564.D | 9.91        | 4.34        |
| AR1660CCC500    | AR1660CCC500     | 12/03/2025       | 14:22            | PO115565.D | 9.91        | 4.34        |
| IBLK            | IBLK             | 12/03/2025       | 15:35            | PO115569.D | 9.91        | 4.34        |
| IBLK            | IBLK             | 11/21/2025       | 10:27            | PP076598.D | 10.37       | 4.62        |
| AR1660ICC1000   | AR1660ICC1000    | 11/21/2025       | 10:44            | PP076599.D | 10.37       | 4.62        |
| AR1660ICC750    | AR1660ICC750     | 11/21/2025       | 11:00            | PP076600.D | 10.37       | 4.62        |
| AR1660ICC500    | AR1660ICC500     | 11/21/2025       | 11:17            | PP076601.D | 10.37       | 4.61        |
| AR1660ICC250    | AR1660ICC250     | 11/21/2025       | 11:33            | PP076602.D | 10.37       | 4.62        |

### Analytical Sequence

|                    |               |            |       |            |       |      |
|--------------------|---------------|------------|-------|------------|-------|------|
| AR1660ICC050       | AR1660ICC050  | 11/21/2025 | 11:49 | PP076603.D | 10.36 | 4.62 |
| AR1221ICC500       | AR1221ICC500  | 11/21/2025 | 12:38 | PP076604.D | 10.37 | 4.62 |
| AR1232ICC500       | AR1232ICC500  | 11/21/2025 | 12:56 | PP076605.D | 10.38 | 4.62 |
| AR1242ICC1000      | AR1242ICC1000 | 11/21/2025 | 13:15 | PP076606.D | 10.38 | 4.62 |
| AR1242ICC750       | AR1242ICC750  | 11/21/2025 | 13:32 | PP076607.D | 10.37 | 4.62 |
| AR1242ICC500       | AR1242ICC500  | 11/21/2025 | 13:48 | PP076608.D | 10.37 | 4.62 |
| AR1242ICC250       | AR1242ICC250  | 11/21/2025 | 14:05 | PP076609.D | 10.37 | 4.62 |
| AR1242ICC050       | AR1242ICC050  | 11/21/2025 | 14:21 | PP076610.D | 10.37 | 4.62 |
| AR1248ICC1000      | AR1248ICC1000 | 11/21/2025 | 14:54 | PP076611.D | 10.37 | 4.62 |
| AR1248ICC750       | AR1248ICC750  | 11/21/2025 | 15:10 | PP076612.D | 10.37 | 4.62 |
| AR1248ICC500       | AR1248ICC500  | 11/21/2025 | 15:32 | PP076613.D | 10.38 | 4.63 |
| AR1248ICC250       | AR1248ICC250  | 11/21/2025 | 15:49 | PP076614.D | 10.37 | 4.62 |
| AR1248ICC050       | AR1248ICC050  | 11/21/2025 | 16:05 | PP076615.D | 10.37 | 4.62 |
| AR1254ICC1000      | AR1254ICC1000 | 11/21/2025 | 16:38 | PP076616.D | 10.37 | 4.62 |
| AR1254ICC750       | AR1254ICC750  | 11/21/2025 | 16:55 | PP076617.D | 10.37 | 4.62 |
| AR1254ICC500       | AR1254ICC500  | 11/21/2025 | 17:11 | PP076618.D | 10.37 | 4.62 |
| AR1254ICC250       | AR1254ICC250  | 11/21/2025 | 17:28 | PP076619.D | 10.36 | 4.61 |
| AR1254ICC050       | AR1254ICC050  | 11/21/2025 | 17:44 | PP076620.D | 10.37 | 4.62 |
| AR1262ICC500       | AR1262ICC500  | 11/21/2025 | 18:17 | PP076621.D | 10.37 | 4.62 |
| AR1268ICC1000      | AR1268ICC1000 | 11/21/2025 | 18:33 | PP076622.D | 10.37 | 4.62 |
| AR1268ICC750       | AR1268ICC750  | 11/21/2025 | 18:50 | PP076623.D | 10.37 | 4.62 |
| AR1268ICC500       | AR1268ICC500  | 11/21/2025 | 19:06 | PP076624.D | 10.37 | 4.62 |
| AR1268ICC250       | AR1268ICC250  | 11/21/2025 | 19:23 | PP076625.D | 10.37 | 4.62 |
| AR1268ICC050       | AR1268ICC050  | 11/21/2025 | 19:39 | PP076626.D | 10.37 | 4.62 |
| AR1660CCC500       | AR1660CCC500  | 12/03/2025 | 10:05 | PP076744.D | 10.37 | 4.62 |
| I.BLK              | I.BLK         | 12/03/2025 | 13:11 | PP076748.D | 10.37 | 4.62 |
| B50-SP11-112525    | Q3741-05      | 12/03/2025 | 14:03 | PP076751.D | 10.36 | 4.61 |
| B50-SP11-112525MS  | Q3741-05MS    | 12/03/2025 | 14:20 | PP076752.D | 10.37 | 4.62 |
| B50-SP11-112525MSD | Q3741-05MSD   | 12/03/2025 | 14:36 | PP076753.D | 10.37 | 4.62 |
| B50-SP4-112525     | Q3741-06      | 12/03/2025 | 14:53 | PP076754.D | 10.36 | 4.61 |
| AR1660CCC500       | AR1660CCC500  | 12/03/2025 | 17:01 | PP076758.D | 10.37 | 4.62 |
| I.BLK              | I.BLK         | 12/03/2025 | 18:23 | PP076762.D | 10.37 | 4.62 |
| PB170795BL         | PB170795BL    | 12/03/2025 | 18:40 | PP076763.D | 10.36 | 4.62 |
| PB170795BS         | PB170795BS    | 12/03/2025 | 18:56 | PP076764.D | 10.36 | 4.62 |
| AR1660CCC500       | AR1660CCC500  | 12/03/2025 | 22:14 | PP076773.D | 10.36 | 4.61 |
| I.BLK              | I.BLK         | 12/03/2025 | 23:36 | PP076777.D | 10.37 | 4.62 |

## Analytical Sequence

|                                                  |                                                                 |
|--------------------------------------------------|-----------------------------------------------------------------|
| Client: Core Environmental Consultants and Servi | SDG No.: Q3741                                                  |
| Project: Building 50 Probe Investigation         | Instrument ID: ECD_O                                            |
| GC Column: ZB-MR2                                | ID: 0.32 (mm)    Inst. Calib. Date(s): 11/04/2025    11/04/2025 |

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,  
AND STANDARDS IS GIVEN BELOW:

| CLIENT ID       | LAB<br>SAMPLE ID | DATE<br>ANALYZED | TIME<br>ANALYZED | DATAFILE   | DCB<br>RT # | TCX<br>RT # |
|-----------------|------------------|------------------|------------------|------------|-------------|-------------|
| IBLK            | IBLK             | 11/04/2025       | 09:21            | PO114870.D | 8.64        | 3.65        |
| AR1660ICC1000   | AR1660ICC1000    | 11/04/2025       | 09:39            | PO114871.D | 8.64        | 3.65        |
| AR1660ICC750    | AR1660ICC750     | 11/04/2025       | 09:58            | PO114872.D | 8.64        | 3.65        |
| AR1660ICC500    | AR1660ICC500     | 11/04/2025       | 10:16            | PO114873.D | 8.64        | 3.65        |
| AR1660ICC250    | AR1660ICC250     | 11/04/2025       | 10:34            | PO114874.D | 8.64        | 3.65        |
| AR1660ICC050    | AR1660ICC050     | 11/04/2025       | 10:53            | PO114875.D | 8.64        | 3.65        |
| AR1221ICC500    | AR1221ICC500     | 11/04/2025       | 11:11            | PO114876.D | 8.64        | 3.65        |
| AR1232ICC500    | AR1232ICC500     | 11/04/2025       | 11:30            | PO114877.D | 8.64        | 3.65        |
| AR1242ICC1000   | AR1242ICC1000    | 11/04/2025       | 11:48            | PO114878.D | 8.64        | 3.65        |
| AR1242ICC750    | AR1242ICC750     | 11/04/2025       | 12:06            | PO114879.D | 8.64        | 3.65        |
| AR1242ICC500    | AR1242ICC500     | 11/04/2025       | 12:25            | PO114880.D | 8.64        | 3.65        |
| AR1242ICC250    | AR1242ICC250     | 11/04/2025       | 12:42            | PO114881.D | 8.64        | 3.65        |
| AR1242ICC050    | AR1242ICC050     | 11/04/2025       | 13:00            | PO114882.D | 8.64        | 3.65        |
| AR1248ICC1000   | AR1248ICC1000    | 11/04/2025       | 13:19            | PO114883.D | 8.64        | 3.65        |
| AR1248ICC750    | AR1248ICC750     | 11/04/2025       | 13:37            | PO114884.D | 8.64        | 3.65        |
| AR1248ICC500    | AR1248ICC500     | 11/04/2025       | 13:56            | PO114885.D | 8.64        | 3.65        |
| AR1248ICC250    | AR1248ICC250     | 11/04/2025       | 14:14            | PO114886.D | 8.64        | 3.65        |
| AR1248ICC050    | AR1248ICC050     | 11/04/2025       | 14:33            | PO114887.D | 8.64        | 3.65        |
| AR1254ICC1000   | AR1254ICC1000    | 11/04/2025       | 14:51            | PO114888.D | 8.64        | 3.65        |
| AR1254ICC750    | AR1254ICC750     | 11/04/2025       | 15:09            | PO114889.D | 8.64        | 3.65        |
| AR1254ICC500    | AR1254ICC500     | 11/04/2025       | 15:28            | PO114890.D | 8.64        | 3.65        |
| AR1254ICC250    | AR1254ICC250     | 11/04/2025       | 16:04            | PO114891.D | 8.64        | 3.65        |
| AR1254ICC050    | AR1254ICC050     | 11/04/2025       | 16:23            | PO114892.D | 8.64        | 3.65        |
| AR1262ICC500    | AR1262ICC500     | 11/04/2025       | 17:00            | PO114893.D | 8.64        | 3.65        |
| AR1268ICC1000   | AR1268ICC1000    | 11/04/2025       | 17:18            | PO114894.D | 8.64        | 3.65        |
| AR1268ICC750    | AR1268ICC750     | 11/04/2025       | 17:36            | PO114895.D | 8.64        | 3.65        |
| AR1268ICC500    | AR1268ICC500     | 11/04/2025       | 17:55            | PO114896.D | 8.64        | 3.65        |
| AR1268ICC250    | AR1268ICC250     | 11/04/2025       | 18:13            | PO114897.D | 8.64        | 3.65        |
| AR1268ICC050    | AR1268ICC050     | 11/04/2025       | 18:31            | PO114898.D | 8.64        | 3.65        |
| AR1660CCC500    | AR1660CCC500     | 12/03/2025       | 09:34            | PO115552.D | 8.62        | 3.64        |
| IBLK            | IBLK             | 12/03/2025       | 10:48            | PO115556.D | 8.62        | 3.64        |
| B50-SP3-111925  | Q3741-01         | 12/03/2025       | 12:50            | PO115561.D | 8.62        | 3.64        |
| B50-SP13-112025 | Q3741-02         | 12/03/2025       | 13:08            | PO115562.D | 8.62        | 3.64        |
| B50-SP14-112025 | Q3741-03         | 12/03/2025       | 13:27            | PO115563.D | 8.62        | 3.64        |
| B50-SP9-112425  | Q3741-04         | 12/03/2025       | 13:45            | PO115564.D | 8.62        | 3.64        |
| AR1660CCC500    | AR1660CCC500     | 12/03/2025       | 14:22            | PO115565.D | 8.62        | 3.64        |
| IBLK            | IBLK             | 12/03/2025       | 15:35            | PO115569.D | 8.62        | 3.64        |
| IBLK            | IBLK             | 11/21/2025       | 10:27            | PP076598.D | 8.77        | 3.78        |
| AR1660ICC1000   | AR1660ICC1000    | 11/21/2025       | 10:44            | PP076599.D | 8.77        | 3.78        |
| AR1660ICC750    | AR1660ICC750     | 11/21/2025       | 11:00            | PP076600.D | 8.77        | 3.77        |
| AR1660ICC500    | AR1660ICC500     | 11/21/2025       | 11:17            | PP076601.D | 8.77        | 3.77        |
| AR1660ICC250    | AR1660ICC250     | 11/21/2025       | 11:33            | PP076602.D | 8.77        | 3.78        |

### Analytical Sequence

|                    |               |            |       |            |      |      |
|--------------------|---------------|------------|-------|------------|------|------|
| AR1660ICC050       | AR1660ICC050  | 11/21/2025 | 11:49 | PP076603.D | 8.77 | 3.77 |
| AR1221ICC500       | AR1221ICC500  | 11/21/2025 | 12:38 | PP076604.D | 8.77 | 3.77 |
| AR1232ICC500       | AR1232ICC500  | 11/21/2025 | 12:56 | PP076605.D | 8.77 | 3.77 |
| AR1242ICC1000      | AR1242ICC1000 | 11/21/2025 | 13:15 | PP076606.D | 8.77 | 3.77 |
| AR1242ICC750       | AR1242ICC750  | 11/21/2025 | 13:32 | PP076607.D | 8.77 | 3.77 |
| AR1242ICC500       | AR1242ICC500  | 11/21/2025 | 13:48 | PP076608.D | 8.77 | 3.77 |
| AR1242ICC250       | AR1242ICC250  | 11/21/2025 | 14:05 | PP076609.D | 8.77 | 3.77 |
| AR1242ICC050       | AR1242ICC050  | 11/21/2025 | 14:21 | PP076610.D | 8.77 | 3.77 |
| AR1248ICC1000      | AR1248ICC1000 | 11/21/2025 | 14:54 | PP076611.D | 8.77 | 3.77 |
| AR1248ICC750       | AR1248ICC750  | 11/21/2025 | 15:10 | PP076612.D | 8.77 | 3.77 |
| AR1248ICC500       | AR1248ICC500  | 11/21/2025 | 15:32 | PP076613.D | 8.77 | 3.78 |
| AR1248ICC250       | AR1248ICC250  | 11/21/2025 | 15:49 | PP076614.D | 8.77 | 3.77 |
| AR1248ICC050       | AR1248ICC050  | 11/21/2025 | 16:05 | PP076615.D | 8.77 | 3.77 |
| AR1254ICC1000      | AR1254ICC1000 | 11/21/2025 | 16:38 | PP076616.D | 8.77 | 3.77 |
| AR1254ICC750       | AR1254ICC750  | 11/21/2025 | 16:55 | PP076617.D | 8.77 | 3.78 |
| AR1254ICC500       | AR1254ICC500  | 11/21/2025 | 17:11 | PP076618.D | 8.77 | 3.77 |
| AR1254ICC250       | AR1254ICC250  | 11/21/2025 | 17:28 | PP076619.D | 8.77 | 3.77 |
| AR1254ICC050       | AR1254ICC050  | 11/21/2025 | 17:44 | PP076620.D | 8.77 | 3.78 |
| AR1262ICC500       | AR1262ICC500  | 11/21/2025 | 18:17 | PP076621.D | 8.77 | 3.77 |
| AR1268ICC1000      | AR1268ICC1000 | 11/21/2025 | 18:33 | PP076622.D | 8.77 | 3.78 |
| AR1268ICC750       | AR1268ICC750  | 11/21/2025 | 18:50 | PP076623.D | 8.77 | 3.77 |
| AR1268ICC500       | AR1268ICC500  | 11/21/2025 | 19:06 | PP076624.D | 8.77 | 3.78 |
| AR1268ICC250       | AR1268ICC250  | 11/21/2025 | 19:23 | PP076625.D | 8.77 | 3.77 |
| AR1268ICC050       | AR1268ICC050  | 11/21/2025 | 19:39 | PP076626.D | 8.77 | 3.77 |
| AR1660CCC500       | AR1660CCC500  | 12/03/2025 | 10:05 | PP076744.D | 8.77 | 3.77 |
| I.BLK              | I.BLK         | 12/03/2025 | 13:11 | PP076748.D | 8.77 | 3.77 |
| B50-SP11-112525    | Q3741-05      | 12/03/2025 | 14:03 | PP076751.D | 8.76 | 3.77 |
| B50-SP11-112525MS  | Q3741-05MS    | 12/03/2025 | 14:20 | PP076752.D | 8.77 | 3.77 |
| B50-SP11-112525MSD | Q3741-05MSD   | 12/03/2025 | 14:36 | PP076753.D | 8.77 | 3.77 |
| B50-SP4-112525     | Q3741-06      | 12/03/2025 | 14:53 | PP076754.D | 8.76 | 3.77 |
| AR1660CCC500       | AR1660CCC500  | 12/03/2025 | 17:01 | PP076758.D | 8.77 | 3.77 |
| I.BLK              | I.BLK         | 12/03/2025 | 18:23 | PP076762.D | 8.77 | 3.77 |
| PB170795BL         | PB170795BL    | 12/03/2025 | 18:40 | PP076763.D | 8.76 | 3.77 |
| PB170795BS         | PB170795BS    | 12/03/2025 | 18:56 | PP076764.D | 8.76 | 3.77 |
| AR1660CCC500       | AR1660CCC500  | 12/03/2025 | 22:14 | PP076773.D | 8.76 | 3.77 |
| I.BLK              | I.BLK         | 12/03/2025 | 23:36 | PP076777.D | 8.77 | 3.77 |



# QC SAMPLE DATA

## Report of Analysis

|                    |                                                   |                     |                 |       |
|--------------------|---------------------------------------------------|---------------------|-----------------|-------|
| Client:            | Core Environmental Consultants and Services, Inc. |                     | Date Collected: |       |
| Project:           | Building 50 Probe Investigation                   |                     | Date Received:  |       |
| Client Sample ID:  | PB170795BL                                        |                     | SDG No.:        | Q3741 |
| Lab Sample ID:     | PB170795BL                                        |                     | Matrix:         | SOIL  |
| Analytical Method: | 8082A                                             |                     | % Solid:        | 100   |
| Sample Wt/Vol:     | 30.03 g                                           | Final Vol: 10000 uL | Test:           | PCB   |
| Prep Method:       | SW3541B                                           | Prep Date: 12/03/25 |                 |       |

| CAS Number        | Parameter            | Conc. | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana.      | BatchID  |
|-------------------|----------------------|-------|------|----|----------|------------|---------|----------------|----------|
| <b>TARGETS</b>    |                      |       |      |    |          |            |         |                |          |
| 12674-11-2        | Aroclor-1016         | 3.90  | U    | 1  | 3.90     | 17.0       | ug/kg   | 12/03/25 18:40 | PB170795 |
| 11104-28-2        | Aroclor-1221         | 4.00  | U    | 1  | 4.00     | 17.0       | ug/kg   | 12/03/25 18:40 | PB170795 |
| 11141-16-5        | Aroclor-1232         | 3.70  | U    | 1  | 3.70     | 17.0       | ug/kg   | 12/03/25 18:40 | PB170795 |
| 53469-21-9        | Aroclor-1242         | 4.00  | U    | 1  | 4.00     | 17.0       | ug/kg   | 12/03/25 18:40 | PB170795 |
| 12672-29-6        | Aroclor-1248         | 5.90  | U    | 1  | 5.90     | 17.0       | ug/kg   | 12/03/25 18:40 | PB170795 |
| 11097-69-1        | Aroclor-1254         | 3.20  | U    | 1  | 3.20     | 17.0       | ug/kg   | 12/03/25 18:40 | PB170795 |
| 37324-23-5        | Aroclor-1262         | 5.00  | U    | 1  | 5.00     | 17.0       | ug/kg   | 12/03/25 18:40 | PB170795 |
| 11100-14-4        | Aroclor-1268         | 3.60  | U    | 1  | 3.60     | 17.0       | ug/kg   | 12/03/25 18:40 | PB170795 |
| 11096-82-5        | Aroclor-1260         | 3.20  | U    | 1  | 3.20     | 17.0       | ug/kg   | 12/03/25 18:40 | PB170795 |
| <b>SURROGATES</b> |                      |       |      |    |          |            |         |                |          |
| 877-09-8          | Tetrachloro-m-xylene | 22.1  |      |    | 21 - 165 | 111%       | SPK: 20 |                |          |
| 2051-24-3         | Decachlorobiphenyl   | 23.0  |      |    | 10 - 170 | 115%       | SPK: 20 |                |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                                                   |                     |                 |          |
|--------------------|---------------------------------------------------|---------------------|-----------------|----------|
| Client:            | Core Environmental Consultants and Services, Inc. |                     | Date Collected: | 11/04/25 |
| Project:           | Building 50 Probe Investigation                   |                     | Date Received:  | 11/04/25 |
| Client Sample ID:  | PIBLK-PO114870.D                                  |                     | SDG No.:        | Q3741    |
| Lab Sample ID:     | I.BLK-PO114870.D                                  |                     | Matrix:         | WATER    |
| Analytical Method: | 8082A                                             |                     | % Solid:        | 0        |
| Sample Wt/Vol:     | 1000 mL                                           | Final Vol: 10000 uL | Test:           | PCB      |
| Prep Method:       | 5030                                              | Prep Date:          |                 |          |

| CAS Number        | Parameter            | Conc. | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana. | BatchID  |
|-------------------|----------------------|-------|------|----|----------|------------|---------|-----------|----------|
| <b>TARGETS</b>    |                      |       |      |    |          |            |         |           |          |
| 12674-11-2        | Aroclor-1016         | 0.097 | U    | 1  | 0.097    | 0.50       | ug/L    | 11/04/25  | po110425 |
| 11104-28-2        | Aroclor-1221         | 0.13  | U    | 1  | 0.13     | 0.50       | ug/L    | 11/04/25  | po110425 |
| 11141-16-5        | Aroclor-1232         | 0.096 | U    | 1  | 0.096    | 0.50       | ug/L    | 11/04/25  | po110425 |
| 53469-21-9        | Aroclor-1242         | 0.12  | U    | 1  | 0.12     | 0.50       | ug/L    | 11/04/25  | po110425 |
| 12672-29-6        | Aroclor-1248         | 0.071 | U    | 1  | 0.071    | 0.50       | ug/L    | 11/04/25  | po110425 |
| 11097-69-1        | Aroclor-1254         | 0.094 | U    | 1  | 0.094    | 0.50       | ug/L    | 11/04/25  | po110425 |
| 11096-82-5        | Aroclor-1260         | 0.081 | U    | 1  | 0.081    | 0.50       | ug/L    | 11/04/25  | po110425 |
| 37324-23-5        | Aroclor-1262         | 0.14  | U    | 1  | 0.14     | 0.50       | ug/L    | 11/04/25  | po110425 |
| 11100-14-4        | Aroclor-1268         | 0.11  | U    | 1  | 0.11     | 0.50       | ug/L    | 11/04/25  | po110425 |
| <b>SURROGATES</b> |                      |       |      |    |          |            |         |           |          |
| 877-09-8          | Tetrachloro-m-xylene | 15.5  |      |    | 60 - 140 | 77%        | SPK: 20 |           |          |
| 2051-24-3         | Decachlorobiphenyl   | 16.8  |      |    | 60 - 140 | 84%        | SPK: 20 |           |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                                                   |                 |          |
|--------------------|---------------------------------------------------|-----------------|----------|
| Client:            | Core Environmental Consultants and Services, Inc. | Date Collected: | 12/03/25 |
| Project:           | Building 50 Probe Investigation                   | Date Received:  | 12/03/25 |
| Client Sample ID:  | PIBLK-PO115556.D                                  | SDG No.:        | Q3741    |
| Lab Sample ID:     | I.BLK-PO115556.D                                  | Matrix:         | WATER    |
| Analytical Method: | 8082A                                             | % Solid:        | 0        |
| Sample Wt/Vol:     | 1000 mL                                           | Final Vol:      | 10000 uL |
| Prep Method:       | 5030                                              | Test:           | PCB      |
|                    | Prep Date:                                        |                 |          |

| CAS Number        | Parameter            | Conc. | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana. | BatchID  |
|-------------------|----------------------|-------|------|----|----------|------------|---------|-----------|----------|
| <b>TARGETS</b>    |                      |       |      |    |          |            |         |           |          |
| 12674-11-2        | Aroclor-1016         | 0.097 | U    | 1  | 0.097    | 0.50       | ug/L    | 12/03/25  | PO120325 |
| 11104-28-2        | Aroclor-1221         | 0.13  | U    | 1  | 0.13     | 0.50       | ug/L    | 12/03/25  | PO120325 |
| 11141-16-5        | Aroclor-1232         | 0.096 | U    | 1  | 0.096    | 0.50       | ug/L    | 12/03/25  | PO120325 |
| 53469-21-9        | Aroclor-1242         | 0.12  | U    | 1  | 0.12     | 0.50       | ug/L    | 12/03/25  | PO120325 |
| 12672-29-6        | Aroclor-1248         | 0.071 | U    | 1  | 0.071    | 0.50       | ug/L    | 12/03/25  | PO120325 |
| 11097-69-1        | Aroclor-1254         | 0.094 | U    | 1  | 0.094    | 0.50       | ug/L    | 12/03/25  | PO120325 |
| 11096-82-5        | Aroclor-1260         | 0.081 | U    | 1  | 0.081    | 0.50       | ug/L    | 12/03/25  | PO120325 |
| 37324-23-5        | Aroclor-1262         | 0.14  | U    | 1  | 0.14     | 0.50       | ug/L    | 12/03/25  | PO120325 |
| 11100-14-4        | Aroclor-1268         | 0.11  | U    | 1  | 0.11     | 0.50       | ug/L    | 12/03/25  | PO120325 |
| <b>SURROGATES</b> |                      |       |      |    |          |            |         |           |          |
| 877-09-8          | Tetrachloro-m-xylene | 23.6  |      |    | 60 - 140 | 118%       | SPK: 20 |           |          |
| 2051-24-3         | Decachlorobiphenyl   | 23.9  |      |    | 60 - 140 | 120%       | SPK: 20 |           |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                                                   |                     |                 |          |
|--------------------|---------------------------------------------------|---------------------|-----------------|----------|
| Client:            | Core Environmental Consultants and Services, Inc. |                     | Date Collected: | 12/03/25 |
| Project:           | Building 50 Probe Investigation                   |                     | Date Received:  | 12/03/25 |
| Client Sample ID:  | PIBLK-PO115569.D                                  |                     | SDG No.:        | Q3741    |
| Lab Sample ID:     | I.BLK-PO115569.D                                  |                     | Matrix:         | WATER    |
| Analytical Method: | 8082A                                             |                     | % Solid:        | 0        |
| Sample Wt/Vol:     | 1000 mL                                           | Final Vol: 10000 uL | Test:           | PCB      |
| Prep Method:       | 5030                                              | Prep Date:          |                 |          |

| CAS Number        | Parameter            | Conc. | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana. | BatchID  |
|-------------------|----------------------|-------|------|----|----------|------------|---------|-----------|----------|
| <b>TARGETS</b>    |                      |       |      |    |          |            |         |           |          |
| 12674-11-2        | Aroclor-1016         | 0.097 | U    | 1  | 0.097    | 0.50       | ug/L    | 12/03/25  | PO120325 |
| 11104-28-2        | Aroclor-1221         | 0.13  | U    | 1  | 0.13     | 0.50       | ug/L    | 12/03/25  | PO120325 |
| 11141-16-5        | Aroclor-1232         | 0.096 | U    | 1  | 0.096    | 0.50       | ug/L    | 12/03/25  | PO120325 |
| 53469-21-9        | Aroclor-1242         | 0.12  | U    | 1  | 0.12     | 0.50       | ug/L    | 12/03/25  | PO120325 |
| 12672-29-6        | Aroclor-1248         | 0.071 | U    | 1  | 0.071    | 0.50       | ug/L    | 12/03/25  | PO120325 |
| 11097-69-1        | Aroclor-1254         | 0.094 | U    | 1  | 0.094    | 0.50       | ug/L    | 12/03/25  | PO120325 |
| 11096-82-5        | Aroclor-1260         | 0.081 | U    | 1  | 0.081    | 0.50       | ug/L    | 12/03/25  | PO120325 |
| 37324-23-5        | Aroclor-1262         | 0.14  | U    | 1  | 0.14     | 0.50       | ug/L    | 12/03/25  | PO120325 |
| 11100-14-4        | Aroclor-1268         | 0.11  | U    | 1  | 0.11     | 0.50       | ug/L    | 12/03/25  | PO120325 |
| <b>SURROGATES</b> |                      |       |      |    |          |            |         |           |          |
| 877-09-8          | Tetrachloro-m-xylene | 24.1  |      |    | 60 - 140 | 120%       | SPK: 20 |           |          |
| 2051-24-3         | Decachlorobiphenyl   | 23.8  |      |    | 60 - 140 | 119%       | SPK: 20 |           |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                                                   |                     |                 |          |
|--------------------|---------------------------------------------------|---------------------|-----------------|----------|
| Client:            | Core Environmental Consultants and Services, Inc. |                     | Date Collected: | 11/21/25 |
| Project:           | Building 50 Probe Investigation                   |                     | Date Received:  | 11/21/25 |
| Client Sample ID:  | PIBLK-PP076598.D                                  |                     | SDG No.:        | Q3741    |
| Lab Sample ID:     | I.BLK-PP076598.D                                  |                     | Matrix:         | WATER    |
| Analytical Method: | 8082A                                             |                     | % Solid:        | 0        |
| Sample Wt/Vol:     | 1000 mL                                           | Final Vol: 10000 uL | Test:           | PCB      |
| Prep Method:       | 5030                                              | Prep Date:          |                 |          |

| CAS Number        | Parameter            | Conc. | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana. | BatchID  |
|-------------------|----------------------|-------|------|----|----------|------------|---------|-----------|----------|
| <b>TARGETS</b>    |                      |       |      |    |          |            |         |           |          |
| 12674-11-2        | Aroclor-1016         | 0.097 | U    | 1  | 0.097    | 0.50       | ug/L    | 11/21/25  | PP112125 |
| 11104-28-2        | Aroclor-1221         | 0.13  | U    | 1  | 0.13     | 0.50       | ug/L    | 11/21/25  | PP112125 |
| 11141-16-5        | Aroclor-1232         | 0.096 | U    | 1  | 0.096    | 0.50       | ug/L    | 11/21/25  | PP112125 |
| 53469-21-9        | Aroclor-1242         | 0.12  | U    | 1  | 0.12     | 0.50       | ug/L    | 11/21/25  | PP112125 |
| 12672-29-6        | Aroclor-1248         | 0.071 | U    | 1  | 0.071    | 0.50       | ug/L    | 11/21/25  | PP112125 |
| 11097-69-1        | Aroclor-1254         | 0.094 | U    | 1  | 0.094    | 0.50       | ug/L    | 11/21/25  | PP112125 |
| 11096-82-5        | Aroclor-1260         | 0.081 | U    | 1  | 0.081    | 0.50       | ug/L    | 11/21/25  | PP112125 |
| 37324-23-5        | Aroclor-1262         | 0.14  | U    | 1  | 0.14     | 0.50       | ug/L    | 11/21/25  | PP112125 |
| 11100-14-4        | Aroclor-1268         | 0.11  | U    | 1  | 0.11     | 0.50       | ug/L    | 11/21/25  | PP112125 |
| <b>SURROGATES</b> |                      |       |      |    |          |            |         |           |          |
| 877-09-8          | Tetrachloro-m-xylene | 21.3  |      |    | 60 - 140 | 106%       | SPK: 20 |           |          |
| 2051-24-3         | Decachlorobiphenyl   | 22.1  |      |    | 60 - 140 | 110%       | SPK: 20 |           |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                                                   |                     |                 |          |
|--------------------|---------------------------------------------------|---------------------|-----------------|----------|
| Client:            | Core Environmental Consultants and Services, Inc. |                     | Date Collected: | 12/03/25 |
| Project:           | Building 50 Probe Investigation                   |                     | Date Received:  | 12/03/25 |
| Client Sample ID:  | PIBLK-PP076748.D                                  |                     | SDG No.:        | Q3741    |
| Lab Sample ID:     | I.BLK-PP076748.D                                  |                     | Matrix:         | WATER    |
| Analytical Method: | 8082A                                             |                     | % Solid:        | 0        |
| Sample Wt/Vol:     | 1000 mL                                           | Final Vol: 10000 uL | Test:           | PCB      |
| Prep Method:       | 5030                                              | Prep Date:          |                 |          |

| CAS Number        | Parameter            | Conc. | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana. | BatchID  |
|-------------------|----------------------|-------|------|----|----------|------------|---------|-----------|----------|
| <b>TARGETS</b>    |                      |       |      |    |          |            |         |           |          |
| 12674-11-2        | Aroclor-1016         | 0.097 | U    | 1  | 0.097    | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 11104-28-2        | Aroclor-1221         | 0.13  | U    | 1  | 0.13     | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 11141-16-5        | Aroclor-1232         | 0.096 | U    | 1  | 0.096    | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 53469-21-9        | Aroclor-1242         | 0.12  | U    | 1  | 0.12     | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 12672-29-6        | Aroclor-1248         | 0.071 | U    | 1  | 0.071    | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 11097-69-1        | Aroclor-1254         | 0.094 | U    | 1  | 0.094    | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 11096-82-5        | Aroclor-1260         | 0.081 | U    | 1  | 0.081    | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 37324-23-5        | Aroclor-1262         | 0.14  | U    | 1  | 0.14     | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 11100-14-4        | Aroclor-1268         | 0.11  | U    | 1  | 0.11     | 0.50       | ug/L    | 12/03/25  | PP120325 |
| <b>SURROGATES</b> |                      |       |      |    |          |            |         |           |          |
| 877-09-8          | Tetrachloro-m-xylene | 25.5  |      |    | 60 - 140 | 127%       | SPK: 20 |           |          |
| 2051-24-3         | Decachlorobiphenyl   | 26.8  |      |    | 60 - 140 | 134%       | SPK: 20 |           |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                                                   |                     |                 |          |
|--------------------|---------------------------------------------------|---------------------|-----------------|----------|
| Client:            | Core Environmental Consultants and Services, Inc. |                     | Date Collected: | 12/03/25 |
| Project:           | Building 50 Probe Investigation                   |                     | Date Received:  | 12/03/25 |
| Client Sample ID:  | PIBLK-PP076762.D                                  |                     | SDG No.:        | Q3741    |
| Lab Sample ID:     | I.BLK-PP076762.D                                  |                     | Matrix:         | WATER    |
| Analytical Method: | 8082A                                             |                     | % Solid:        | 0        |
| Sample Wt/Vol:     | 1000 mL                                           | Final Vol: 10000 uL | Test:           | PCB      |
| Prep Method:       | 5030                                              | Prep Date:          |                 |          |

| CAS Number        | Parameter            | Conc. | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana. | BatchID  |
|-------------------|----------------------|-------|------|----|----------|------------|---------|-----------|----------|
| <b>TARGETS</b>    |                      |       |      |    |          |            |         |           |          |
| 12674-11-2        | Aroclor-1016         | 0.097 | U    | 1  | 0.097    | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 11104-28-2        | Aroclor-1221         | 0.13  | U    | 1  | 0.13     | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 11141-16-5        | Aroclor-1232         | 0.096 | U    | 1  | 0.096    | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 53469-21-9        | Aroclor-1242         | 0.12  | U    | 1  | 0.12     | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 12672-29-6        | Aroclor-1248         | 0.071 | U    | 1  | 0.071    | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 11097-69-1        | Aroclor-1254         | 0.094 | U    | 1  | 0.094    | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 11096-82-5        | Aroclor-1260         | 0.081 | U    | 1  | 0.081    | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 37324-23-5        | Aroclor-1262         | 0.14  | U    | 1  | 0.14     | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 11100-14-4        | Aroclor-1268         | 0.11  | U    | 1  | 0.11     | 0.50       | ug/L    | 12/03/25  | PP120325 |
| <b>SURROGATES</b> |                      |       |      |    |          |            |         |           |          |
| 877-09-8          | Tetrachloro-m-xylene | 23.8  |      |    | 60 - 140 | 119%       | SPK: 20 |           |          |
| 2051-24-3         | Decachlorobiphenyl   | 24.0  |      |    | 60 - 140 | 120%       | SPK: 20 |           |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                                                   |                     |                 |          |
|--------------------|---------------------------------------------------|---------------------|-----------------|----------|
| Client:            | Core Environmental Consultants and Services, Inc. |                     | Date Collected: | 12/03/25 |
| Project:           | Building 50 Probe Investigation                   |                     | Date Received:  | 12/03/25 |
| Client Sample ID:  | PIBLK-PP076777.D                                  |                     | SDG No.:        | Q3741    |
| Lab Sample ID:     | I.BLK-PP076777.D                                  |                     | Matrix:         | WATER    |
| Analytical Method: | 8082A                                             |                     | % Solid:        | 0        |
| Sample Wt/Vol:     | 1000 mL                                           | Final Vol: 10000 uL | Test:           | PCB      |
| Prep Method:       | 5030                                              | Prep Date:          |                 |          |

| CAS Number        | Parameter            | Conc. | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana. | BatchID  |
|-------------------|----------------------|-------|------|----|----------|------------|---------|-----------|----------|
| <b>TARGETS</b>    |                      |       |      |    |          |            |         |           |          |
| 12674-11-2        | Aroclor-1016         | 0.097 | U    | 1  | 0.097    | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 11104-28-2        | Aroclor-1221         | 0.13  | U    | 1  | 0.13     | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 11141-16-5        | Aroclor-1232         | 0.096 | U    | 1  | 0.096    | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 53469-21-9        | Aroclor-1242         | 0.12  | U    | 1  | 0.12     | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 12672-29-6        | Aroclor-1248         | 0.071 | U    | 1  | 0.071    | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 11097-69-1        | Aroclor-1254         | 0.094 | U    | 1  | 0.094    | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 11096-82-5        | Aroclor-1260         | 0.081 | U    | 1  | 0.081    | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 37324-23-5        | Aroclor-1262         | 0.14  | U    | 1  | 0.14     | 0.50       | ug/L    | 12/03/25  | PP120325 |
| 11100-14-4        | Aroclor-1268         | 0.11  | U    | 1  | 0.11     | 0.50       | ug/L    | 12/03/25  | PP120325 |
| <b>SURROGATES</b> |                      |       |      |    |          |            |         |           |          |
| 877-09-8          | Tetrachloro-m-xylene | 23.4  |      |    | 60 - 140 | 117%       | SPK: 20 |           |          |
| 2051-24-3         | Decachlorobiphenyl   | 23.7  |      |    | 60 - 140 | 118%       | SPK: 20 |           |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                                                   |                     |                 |       |
|--------------------|---------------------------------------------------|---------------------|-----------------|-------|
| Client:            | Core Environmental Consultants and Services, Inc. |                     | Date Collected: |       |
| Project:           | Building 50 Probe Investigation                   |                     | Date Received:  |       |
| Client Sample ID:  | PB170795BS                                        |                     | SDG No.:        | Q3741 |
| Lab Sample ID:     | PB170795BS                                        |                     | Matrix:         | SOIL  |
| Analytical Method: | 8082A                                             |                     | % Solid:        | 100   |
| Sample Wt/Vol:     | 30.01 g                                           | Final Vol: 10000 uL | Test:           | PCB   |
| Prep Method:       | SW3541B                                           | Prep Date: 12/03/25 |                 |       |

| CAS Number        | Parameter            | Conc. | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana.      | BatchID  |
|-------------------|----------------------|-------|------|----|----------|------------|---------|----------------|----------|
| <b>TARGETS</b>    |                      |       |      |    |          |            |         |                |          |
| 12674-11-2        | Aroclor-1016         | 139   |      | 1  | 3.90     | 17.0       | ug/kg   | 12/03/25 18:56 | PB170795 |
| 11104-28-2        | Aroclor-1221         | 4.00  | U    | 1  | 4.00     | 17.0       | ug/kg   | 12/03/25 18:56 | PB170795 |
| 11141-16-5        | Aroclor-1232         | 3.70  | U    | 1  | 3.70     | 17.0       | ug/kg   | 12/03/25 18:56 | PB170795 |
| 53469-21-9        | Aroclor-1242         | 4.00  | U    | 1  | 4.00     | 17.0       | ug/kg   | 12/03/25 18:56 | PB170795 |
| 12672-29-6        | Aroclor-1248         | 5.90  | U    | 1  | 5.90     | 17.0       | ug/kg   | 12/03/25 18:56 | PB170795 |
| 11097-69-1        | Aroclor-1254         | 3.20  | U    | 1  | 3.20     | 17.0       | ug/kg   | 12/03/25 18:56 | PB170795 |
| 37324-23-5        | Aroclor-1262         | 5.00  | U    | 1  | 5.00     | 17.0       | ug/kg   | 12/03/25 18:56 | PB170795 |
| 11100-14-4        | Aroclor-1268         | 3.60  | U    | 1  | 3.60     | 17.0       | ug/kg   | 12/03/25 18:56 | PB170795 |
| 11096-82-5        | Aroclor-1260         | 135   |      | 1  | 3.20     | 17.0       | ug/kg   | 12/03/25 18:56 | PB170795 |
| <b>SURROGATES</b> |                      |       |      |    |          |            |         |                |          |
| 877-09-8          | Tetrachloro-m-xylene | 22.8  |      |    | 21 - 165 | 114%       | SPK: 20 |                |          |
| 2051-24-3         | Decachlorobiphenyl   | 23.4  |      |    | 10 - 170 | 117%       | SPK: 20 |                |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                                                   |                 |          |
|--------------------|---------------------------------------------------|-----------------|----------|
| Client:            | Core Environmental Consultants and Services, Inc. | Date Collected: | 11/25/25 |
| Project:           | Building 50 Probe Investigation                   | Date Received:  | 11/26/25 |
| Client Sample ID:  | B50-SP11-112525MS                                 | SDG No.:        | Q3741    |
| Lab Sample ID:     | Q3741-05MS                                        | Matrix:         | SOIL     |
| Analytical Method: | 8082A                                             | % Solid:        | 84.6     |
| Sample Wt/Vol:     | 30.07 g                                           | Final Vol:      | 10000 uL |
| Prep Method:       | SW3541B                                           | Test:           | PCB      |
|                    | Prep Date:                                        | 12/03/25        |          |

| CAS Number        | Parameter            | Conc. | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana.      | BatchID  |
|-------------------|----------------------|-------|------|----|----------|------------|---------|----------------|----------|
| <b>TARGETS</b>    |                      |       |      |    |          |            |         |                |          |
| 12674-11-2        | Aroclor-1016         | 160   |      | 1  | 4.70     | 20.0       | ug/kg   | 12/03/25 14:20 | PB170795 |
| 11104-28-2        | Aroclor-1221         | 4.80  | U    | 1  | 4.80     | 20.0       | ug/kg   | 12/03/25 14:20 | PB170795 |
| 11141-16-5        | Aroclor-1232         | 4.40  | U    | 1  | 4.40     | 20.0       | ug/kg   | 12/03/25 14:20 | PB170795 |
| 53469-21-9        | Aroclor-1242         | 4.70  | U    | 1  | 4.70     | 20.0       | ug/kg   | 12/03/25 14:20 | PB170795 |
| 12672-29-6        | Aroclor-1248         | 7.00  | U    | 1  | 7.00     | 20.0       | ug/kg   | 12/03/25 14:20 | PB170795 |
| 11097-69-1        | Aroclor-1254         | 107   |      | 1  | 3.80     | 20.0       | ug/kg   | 12/03/25 14:20 | PB170795 |
| 37324-23-5        | Aroclor-1262         | 5.90  | U    | 1  | 5.90     | 20.0       | ug/kg   | 12/03/25 14:20 | PB170795 |
| 11100-14-4        | Aroclor-1268         | 4.20  | U    | 1  | 4.20     | 20.0       | ug/kg   | 12/03/25 14:20 | PB170795 |
| 11096-82-5        | Aroclor-1260         | 150   |      | 1  | 3.80     | 20.0       | ug/kg   | 12/03/25 14:20 | PB170795 |
| <b>SURROGATES</b> |                      |       |      |    |          |            |         |                |          |
| 877-09-8          | Tetrachloro-m-xylene | 23.1  |      |    | 21 - 165 | 116%       | SPK: 20 |                |          |
| 2051-24-3         | Decachlorobiphenyl   | 20.0  |      |    | 10 - 170 | 100%       | SPK: 20 |                |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                                                   |                 |          |
|--------------------|---------------------------------------------------|-----------------|----------|
| Client:            | Core Environmental Consultants and Services, Inc. | Date Collected: | 11/25/25 |
| Project:           | Building 50 Probe Investigation                   | Date Received:  | 11/26/25 |
| Client Sample ID:  | B50-SP11-112525MSD                                | SDG No.:        | Q3741    |
| Lab Sample ID:     | Q3741-05MSD                                       | Matrix:         | SOIL     |
| Analytical Method: | 8082A                                             | % Solid:        | 84.6     |
| Sample Wt/Vol:     | 30.05 g                                           | Final Vol:      | 10000 uL |
| Prep Method:       | SW3541B                                           | Test:           | PCB      |
|                    | Prep Date:                                        | 12/03/25        |          |

| CAS Number        | Parameter            | Conc. | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana.      | BatchID  |
|-------------------|----------------------|-------|------|----|----------|------------|---------|----------------|----------|
| <b>TARGETS</b>    |                      |       |      |    |          |            |         |                |          |
| 12674-11-2        | Aroclor-1016         | 163   |      | 1  | 4.70     | 20.1       | ug/kg   | 12/03/25 14:36 | PB170795 |
| 11104-28-2        | Aroclor-1221         | 4.80  | U    | 1  | 4.80     | 20.1       | ug/kg   | 12/03/25 14:36 | PB170795 |
| 11141-16-5        | Aroclor-1232         | 4.40  | U    | 1  | 4.40     | 20.1       | ug/kg   | 12/03/25 14:36 | PB170795 |
| 53469-21-9        | Aroclor-1242         | 4.70  | U    | 1  | 4.70     | 20.1       | ug/kg   | 12/03/25 14:36 | PB170795 |
| 12672-29-6        | Aroclor-1248         | 7.00  | U    | 1  | 7.00     | 20.1       | ug/kg   | 12/03/25 14:36 | PB170795 |
| 11097-69-1        | Aroclor-1254         | 108   |      | 1  | 3.80     | 20.1       | ug/kg   | 12/03/25 14:36 | PB170795 |
| 37324-23-5        | Aroclor-1262         | 5.90  | U    | 1  | 5.90     | 20.1       | ug/kg   | 12/03/25 14:36 | PB170795 |
| 11100-14-4        | Aroclor-1268         | 4.20  | U    | 1  | 4.20     | 20.1       | ug/kg   | 12/03/25 14:36 | PB170795 |
| 11096-82-5        | Aroclor-1260         | 152   |      | 1  | 3.80     | 20.1       | ug/kg   | 12/03/25 14:36 | PB170795 |
| <b>SURROGATES</b> |                      |       |      |    |          |            |         |                |          |
| 877-09-8          | Tetrachloro-m-xylene | 23.3  |      |    | 21 - 165 | 116%       | SPK: 20 |                |          |
| 2051-24-3         | Decachlorobiphenyl   | 20.5  |      |    | 10 - 170 | 103%       | SPK: 20 |                |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit