



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3741</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Core Environmental Consultants and Serv</u>	Datafile :	<u>PP076764.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB170795BS (Column 1)	AR1016	166.6	139	ug/kg	83				71	120	
	AR1260	166.6	134	ug/kg	80				65	130	
PB170795BS (Column 2)	AR1016	166.6	136	ug/kg	82				71	120	
	AR1260	166.6	135	ug/kg	81				65	130	