

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111825\
 Data File : BP026143.D
 Acq On : 18 Nov 2025 14:31
 Operator : RC/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC2.5

Quant Time: Nov 19 00:37:46 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 00:37:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.655	152	296435	20.000	ng	-0.02
21) Naphthalene-d8	10.431	136	1184362	20.000	ng	-0.02
39) Acenaphthene-d10	14.313	164	759057	20.000	ng	-0.01
64) Phenanthrene-d10	17.125	188	1539519	20.000	ng	0.00
76) Chrysene-d12	21.583	240	1732194	20.000	ng	0.01
86) Perylene-d12	24.918	264	1973418	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						
19) n-Nitroso-di-n-propyla...	8.449	70	34712	2.457	ng	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111825\
Data File : BP026143.D
Acq On : 18 Nov 2025 14:31
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Quant Time: Nov 19 00:37:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 19 00:37:08 2025
Response via : Initial Calibration

