

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: BF120325

SequenceID : BF120325

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	<u>✓</u>	<u> </u>	<u> </u>
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)	<u>✓</u>	<u> </u>	<u> </u>
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series	<u>✓</u>	<u> </u>	<u> </u>
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series	<u>✓</u>	<u> </u>	<u> </u>
5. GC/MS Calibration Met:	<u>✓</u>	<u> </u>	<u> </u>
a. Initial calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	<u>✓</u>	<u> </u>	<u> </u>
b. Continuous Calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. Compounds#9,54,77 were biased high in the CCC and not present in the samples.	<u>✓</u>	<u> </u>	<u> </u>
6. Blank Contamination - If yes, list compounds and concentrations in each blank:	<u>✓</u>	<u> </u>	<u> </u>
a. B/N Fraction			
d. Acid Fraction			

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

a. B/N Fraction

d. Acid Fraction

✓

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

a. B/N Fraction

d. Acid Fraction

✓

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

✓

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

✓

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

✓

ADDITIONAL COMMENTS:

Internal standard failed in the sample Q3753-04 & Needed 2X Dilution.

Jagrut
Analyst

12/03/2025
Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: BF120325

SequenceID : BF120325 NA NO YES