

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3741

Client ID : SSTDCCC040

Date Analyzed: 12/03/2025

Lab File ID: BF144415.D

Time Analyzed: 12:25

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	58261	6.863	220069	8.15	120401	9.90
UPPER LIMIT	116522	7.363	440138	8.645	240802	10.404
LOWER LIMIT	29130.5	6.363	110035	7.645	60200.5	9.404
EPA SAMPLE NO.						
01 B50-SP14-112025	52469	6.86	193698	8.14	103555	9.90
02 B50-SP3-111925	81142	6.86	286107	8.15	142110	9.90
03 B50-SP13-112025	88360	6.86	321604	8.15	169125	9.90
04 B50-SP9-112425	45091	6.86	171081	8.15	84880	9.90

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.