

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>CORE02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3741</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>12/02/2025 09:14</u>
Lab File ID:	<u>VY023849.D</u>	Init. Calib. Date(s):	<u>11/26/2025 11/26/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>08:04 10:02</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.377	0.394		4.51	20
Chloromethane	0.576	0.577	0.1	0.17	20
Vinyl Chloride	0.624	0.668		7.05	20
Bromomethane	0.440	0.438		-0.46	20
Chloroethane	0.406	0.428		5.42	20
Trichlorofluoromethane	0.843	0.887		5.22	20
1,1,2-Trichlorotrifluoroethane	0.492	0.518		5.28	20
1,1-Dichloroethene	0.484	0.518		7.03	20
Acetone	0.140	0.184		31.43	20
Carbon Disulfide	1.562	1.710		9.48	20
Methyl tert-butyl Ether	1.237	1.274		2.99	20
Methyl Acetate	0.271	0.262		-3.32	20
Methylene Chloride	0.593	0.551		-7.08	20
trans-1,2-Dichloroethene	0.536	0.581		8.4	20
1,1-Dichloroethane	0.976	1.037	0.1	6.25	20
Cyclohexane	0.936	0.999		6.73	20
2-Butanone	0.171	0.201		17.54	20
Carbon Tetrachloride	0.498	0.544		9.24	20
cis-1,2-Dichloroethene	0.622	0.654		5.14	20
Bromochloromethane	0.406	0.406		0	20
Chloroform	0.981	1.035		5.51	20
1,1,1-Trichloroethane	0.871	0.928		6.54	20
Methylcyclohexane	0.600	0.659		9.83	20
Benzene	1.398	1.520		8.73	20
1,2-Dichloroethane	0.360	0.381		5.83	20
Trichloroethene	0.359	0.385		7.24	20
1,2-Dichloropropane	0.330	0.357		8.18	20
Bromodichloromethane	0.467	0.493		5.57	20
4-Methyl-2-Pentanone	0.210	0.226		7.62	20
Toluene	0.866	0.960		10.85	20
t-1,3-Dichloropropene	0.428	0.458		7.01	20
cis-1,3-Dichloropropene	0.512	0.547		6.84	20
1,1,2-Trichloroethane	0.240	0.251		4.58	20
2-Hexanone	0.155	0.185		19.35	20
Dibromochloromethane	0.321	0.336		4.67	20
1,2-Dibromoethane	0.222	0.234		5.41	20
Tetrachloroethene	0.474	0.519		9.49	20
Chlorobenzene	1.098	1.184	0.3	7.83	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>CORE02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3741</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>12/02/2025</u> <u>09:14</u>
Lab File ID:	<u>VY023849.D</u>	Init. Calib. Date(s):	<u>11/26/2025</u> <u>11/26/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>08:04</u> <u>10:02</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.875	2.111		12.59	20
m/p-Xylenes	0.722	0.812		12.47	20
o-Xylene	0.677	0.758		11.97	20
Styrene	1.116	1.255		12.45	20
Bromoform	0.217	0.229	0.1	5.53	20
Isopropylbenzene	3.549	3.929		10.71	20
1,1,2,2-Tetrachloroethane	0.584	0.601	0.3	2.91	20
1,3-Dichlorobenzene	1.673	1.755		4.9	20
1,4-Dichlorobenzene	1.670	1.719		2.93	20
1,2-Dichlorobenzene	1.463	1.541		5.33	20
1,2-Dibromo-3-Chloropropane	0.098	0.096		-2.04	20
1,2,4-Trichlorobenzene	0.872	0.895		2.64	20
1,2,3-Trichlorobenzene	0.743	0.758		2.02	20
1,2-Dichloroethane-d4	0.487	0.485		-0.41	20
Dibromofluoromethane	0.296	0.307		3.72	20
Toluene-d8	1.176	1.213		3.15	20
4-Bromofluorobenzene	0.387	0.392		1.29	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.