

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN112425

Review By	John Carlone	Review On	11/25/2025 9:36:43 AM
Supervise By	Maresh Dadoda	Supervise On	11/26/2025 1:52:19 PM
SubDirectory	VN112425	HP Acquire Method	HP Processing Method 82N112425W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP136378		
Initial Calibration Stds	VP136444,VP136445,VP136446,VP136447,VP136448,VP136449		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP136450		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN088343.D	24 Nov 2025 08:17	JC\MD	Ok
2	VSTDIC001	VN088344.D	24 Nov 2025 12:38	JC\MD	Ok,M
3	VSTDIC005	VN088345.D	24 Nov 2025 13:11	JC\MD	Ok,M
4	VSTDIC020	VN088346.D	24 Nov 2025 13:32	JC\MD	Ok,M
5	VSTDIC050	VN088347.D	24 Nov 2025 13:53	JC\MD	Ok,M
6	VSTDIC100	VN088348.D	24 Nov 2025 14:14	JC\MD	Ok,M
7	VSTDIC150	VN088349.D	24 Nov 2025 14:35	JC\MD	Ok,M
8	IBLK	VN088350.D	24 Nov 2025 14:56	JC\MD	Ok
9	VSTDICV050	VN088351.D	24 Nov 2025 15:18	JC\MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN120125

Review By	Review On
Supervise By	Supervise On
SubDirectory VN120125	HP Acquire Method HP Processing Method 82N112425W.M
STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136451
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136452,VP136453

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN088369.D	01 Dec 2025 08:49	JC\MD	Ok
2	VSTDCCC050	VN088370.D	01 Dec 2025 09:58	JC\MD	Ok,M
3	VN1201MBL01	VN088371.D	01 Dec 2025 10:33	JC\MD	Ok
4	VN1201WBL01	VN088372.D	01 Dec 2025 10:53	JC\MD	Ok
5	VN1201WBS01	VN088373.D	01 Dec 2025 11:27	JC\MD	Ok,M
6	VN1201WBSD01	VN088374.D	01 Dec 2025 12:10	JC\MD	Ok,M
7	Q3715-02	VN088375.D	01 Dec 2025 12:30	JC\MD	Ok,M
8	Q3741-07	VN088376.D	01 Dec 2025 12:51	JC\MD	Ok
9	Q3716-01	VN088377.D	01 Dec 2025 13:11	JC\MD	Dilution
10	Q3716-03	VN088378.D	01 Dec 2025 13:32	JC\MD	Not Ok
11	Q3716-04	VN088379.D	01 Dec 2025 13:52	JC\MD	Dilution
12	Q3715-01	VN088380.D	01 Dec 2025 14:13	JC\MD	ReRun
13	Q3716-02	VN088381.D	01 Dec 2025 14:33	JC\MD	Ok
14	Q3738-01	VN088382.D	01 Dec 2025 14:54	JC\MD	Not Ok
15	Q3738-03	VN088383.D	01 Dec 2025 15:15	JC\MD	Ok
16	Q3738-02	VN088384.D	01 Dec 2025 15:35	JC\MD	Ok,M
17	Q3745-01	VN088385.D	01 Dec 2025 15:56	JC\MD	ReRun
18	Q3745-02	VN088386.D	01 Dec 2025 16:16	JC\MD	Ok
19	Q3745-03	VN088387.D	01 Dec 2025 16:37	JC\MD	Ok
20	Q3745-04	VN088388.D	01 Dec 2025 16:58	JC\MD	Ok
21	Q3743-03	VN088389.D	01 Dec 2025 17:18	JC\MD	Ok

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QCBatch ID # VN120125

Review By	Review On
Supervise By	Supervise On
SubDirectory VN120125	HP Acquire Method HP Processing Method 82N112425W.M
STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136451
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136452,VP136453

22	Q3743-04	VN088390.D	01 Dec 2025 17:39	JC\MD	Ok
23	Q3743-05	VN088391.D	01 Dec 2025 18:00	JC\MD	Not Ok
24	Q3743-01	VN088392.D	01 Dec 2025 18:20	JC\MD	Ok
25	Q3743-02	VN088393.D	01 Dec 2025 18:41	JC\MD	Ok
26	VSTDCCC050	VN088394.D	01 Dec 2025 19:01	JC\MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QC Batch ID # VY112625

Review By	Semsettin Yesilyurt	Review On	12/1/2025 9:50:11 AM
Supervise By	John Carlone	Supervise On	12/1/2025 2:45:01 PM
SubDirectory	VY112625	HP Acquire Method	MSVOA_Y
		HP Processing Method	82y112625s.m
STD. NAME	STD REF.#		
Tune/Reschk	VP136432		
Initial Calibration Stds	VP136433,VP136434,VP136435,VP136436,VP136437,VP136438		
CCC	VP136440		
Internal Standard/PEM	VP136146		
ICV/I.BLK	VP136439		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY023808.D	26 Nov 2025 07:11	SY/MD	Ok
2	VSTDICCC005	VY023809.D	26 Nov 2025 08:04	SY/MD	Ok,M
3	VSTDICCC010	VY023810.D	26 Nov 2025 08:27	SY/MD	Ok,M
4	VSTDICCC020	VY023811.D	26 Nov 2025 08:49	SY/MD	Ok,M
5	VSTDICCC050	VY023812.D	26 Nov 2025 09:17	SY/MD	Ok,M
6	VSTDICCC100	VY023813.D	26 Nov 2025 09:39	SY/MD	Ok,M
7	VSTDICCC150	VY023814.D	26 Nov 2025 10:02	SY/MD	Ok,M
8	VIBLK	VY023815.D	26 Nov 2025 10:27	SY/MD	Ok
9	VSTDICV050	VY023816.D	26 Nov 2025 10:50	SY/MD	Ok,M
10	VY1126SBL01	VY023817.D	26 Nov 2025 11:13	SY/MD	Ok
11	VY1126SBS01	VY023818.D	26 Nov 2025 11:47	SY/MD	Ok,M
12	VY1126SBSD01	VY023819.D	26 Nov 2025 12:09	SY/MD	Ok,M
13	Q3604-04	VY023820.D	26 Nov 2025 12:34	SY/MD	Not Ok
14	Q3604-05	VY023821.D	26 Nov 2025 12:57	SY/MD	Not Ok
15	VIBLK	VY023822.D	26 Nov 2025 13:21	SY/MD	Ok
16	Q3721-01	VY023823.D	26 Nov 2025 13:44	SY/MD	Ok
17	Q3723-01	VY023824.D	26 Nov 2025 14:08	SY/MD	Ok
18	Q3725-04	VY023825.D	26 Nov 2025 14:31	SY/MD	Ok
19	VIBLK	VY023826.D	26 Nov 2025 14:54	SY/MD	Ok
20	VSTDCCC050	VY023827.D	26 Nov 2025 15:17	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QCBatch ID # VY120125

Review By	John Carlone	Review On	12/2/2025 9:00:18 AM		
Supervise By	Maresh Dadoda	Supervise On	12/2/2025 10:16:40 AM		
SubDirectory	VY120125	HP Acquire Method	MSVOA_Y	HP Processing Method	82y112625s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP136454			
CCC		VP136455,VP136456			
Internal Standard/PEM		VP136146			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY023828.D	01 Dec 2025 07:24	SY/MD	Ok
2	VSTDCCC050	VY023829.D	01 Dec 2025 07:55	SY/MD	Ok,M
3	VY1201SBL01	VY023830.D	01 Dec 2025 09:13	SY/MD	Ok
4	VY1201SBS01	VY023831.D	01 Dec 2025 09:37	SY/MD	Ok,M
5	VY1201SBSD01	VY023832.D	01 Dec 2025 10:00	SY/MD	Ok,M
6	Q3719-09RE	VY023833.D	01 Dec 2025 10:38	SY/MD	Confirms
7	Q3719-13	VY023834.D	01 Dec 2025 11:25	SY/MD	Ok
8	Q3735-01	VY023835.D	01 Dec 2025 12:02	SY/MD	ReRun
9	Q3740-01	VY023836.D	01 Dec 2025 12:26	SY/MD	Ok
10	Q3741-02	VY023837.D	01 Dec 2025 12:49	SY/MD	Ok
11	Q3741-03	VY023838.D	01 Dec 2025 13:12	SY/MD	Ok
12	Q3741-04	VY023839.D	01 Dec 2025 13:36	SY/MD	ReRun
13	Q3741-05	VY023840.D	01 Dec 2025 13:59	SY/MD	ReRun
14	Q3741-01	VY023841.D	01 Dec 2025 14:23	SY/MD	Ok
15	Q3741-06	VY023842.D	01 Dec 2025 14:46	SY/MD	ReRun
16	Q3742-01	VY023843.D	01 Dec 2025 15:10	SY/MD	Ok
17	Q3742-02	VY023844.D	01 Dec 2025 15:33	SY/MD	Ok
18	Q3742-03	VY023845.D	01 Dec 2025 15:57	SY/MD	Ok
19	Q3742-04	VY023846.D	01 Dec 2025 16:20	SY/MD	Ok
20	VSTDCCC050	VY023847.D	01 Dec 2025 17:47	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QC Batch ID # VY120225

Review By	Semsettin Yesilyurt	Review On	12/3/2025 12:40:53 PM		
Supervise By	Mahesh Dadoda	Supervise On	12/3/2025 2:29:19 PM		
SubDirectory	VY120225	HP Acquire Method	MSVOA_Y	HP Processing Method	82y112625s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		AP136460			
CCC		VP136461,VP136462			
Internal Standard/PEM		VP136146			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY023848.D	02 Dec 2025 06:50	SY/MD	Ok
2	VSTDCCC050	VY023849.D	02 Dec 2025 09:14	SY/MD	Ok,M
3	VY1202SBL01	VY023850.D	02 Dec 2025 10:24	SY/MD	Ok
4	VY1202SBS01	VY023851.D	02 Dec 2025 10:53	SY/MD	Ok,M
5	VY1202SBS01	VY023852.D	02 Dec 2025 11:16	SY/MD	Ok,M
6	Q3742-05	VY023853.D	02 Dec 2025 12:08	SY/MD	Ok
7	Q3735-01	VY023854.D	02 Dec 2025 12:31	SY/MD	Ok
8	Q3741-04RE	VY023855.D	02 Dec 2025 12:55	SY/MD	Confirms
9	Q3741-05RE	VY023856.D	02 Dec 2025 13:18	SY/MD	Confirms
10	Q3741-06RE	VY023857.D	02 Dec 2025 13:42	SY/MD	Confirms
11	Q3746-01	VY023858.D	02 Dec 2025 14:05	SY/MD	Ok
12	Q3742-06	VY023859.D	02 Dec 2025 14:29	SY/MD	Ok
13	VIBLK	VY023860.D	02 Dec 2025 15:33	SY/MD	Ok
14	Q3742-08	VY023861.D	02 Dec 2025 15:57	SY/MD	Ok
15	Q3742-09	VY023862.D	02 Dec 2025 16:20	SY/MD	ReRun
16	VSTDCCC050	VY023863.D	02 Dec 2025 17:20	SY/MD	Ok,M

M : Manual Integration