

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN112425

Review By	John Carlone	Review On	11/25/2025 9:36:43 AM
Supervise By	Maresh Dadoda	Supervise On	11/26/2025 1:52:19 PM
SubDirectory	VN112425	HP Acquire Method	HP Processing Method 82N112425W.M

STD. NAME	STD REF.#
Tune/Reschk	VP136378
Initial Calibration Stds	VP136444,VP136445,VP136446,VP136447,VP136448,VP136449
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP136450
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN088343.D	24 Nov 2025 08:17		JC\MD	Ok
2	VSTDICC001	VSTDICC001	VN088344.D	24 Nov 2025 12:38		JC\MD	Ok,M
3	VSTDICC005	VSTDICC005	VN088345.D	24 Nov 2025 13:11	COM.#74 Peak not detected in 1 and 5 PPB	JC\MD	Ok,M
4	VSTDICC020	VSTDICC020	VN088346.D	24 Nov 2025 13:32		JC\MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VN088347.D	24 Nov 2025 13:53	Method failed for Comp. #74	JC\MD	Ok,M
6	VSTDICC100	VSTDICC100	VN088348.D	24 Nov 2025 14:14		JC\MD	Ok,M
7	VSTDICC150	VSTDICC150	VN088349.D	24 Nov 2025 14:35		JC\MD	Ok,M
8	IBLK	IBLK	VN088350.D	24 Nov 2025 14:56		JC\MD	Ok
9	VSTDICV050	ICVVN112425	VN088351.D	24 Nov 2025 15:18		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN120125

Review By	Review On
Supervise By	Supervise On
SubDirectory	VN120125
HP Acquire Method	HP Processing Method
	82N112425W.M
STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136451
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136452,VP136453

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN088369.D	01 Dec 2025 08:49		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN088370.D	01 Dec 2025 09:58	pH#Lot#V12668	JC\MD	Ok,M
3	VN1201MBL01	VN1201MBL01	VN088371.D	01 Dec 2025 10:33		JC\MD	Ok
4	VN1201WBL01	VN1201WBL01	VN088372.D	01 Dec 2025 10:53		JC\MD	Ok
5	VN1201WBS01	VN1201WBS01	VN088373.D	01 Dec 2025 11:27		JC\MD	Ok,M
6	VN1201WBSD01	VN1201WBSD01	VN088374.D	01 Dec 2025 12:10		JC\MD	Ok,M
7	Q3715-02	MW6	VN088375.D	01 Dec 2025 12:30	vial A pH<2	JC\MD	Ok,M
8	Q3741-07	TB-1	VN088376.D	01 Dec 2025 12:51	vial A pH<2 TB	JC\MD	Ok
9	Q3716-01	MW5	VN088377.D	01 Dec 2025 13:11	vial A pH<2 Need 20X	JC\MD	Dilution
10	Q3716-03	MW4	VN088378.D	01 Dec 2025 13:32	Need Straight Run	JC\MD	Not Ok
11	Q3716-04	MW7	VN088379.D	01 Dec 2025 13:52	Need 200X	JC\MD	Dilution
12	Q3715-01	MW2	VN088380.D	01 Dec 2025 14:13	E flag of previous sample	JC\MD	ReRun
13	Q3716-02	MW3	VN088381.D	01 Dec 2025 14:33	vial A pH<2	JC\MD	Ok
14	Q3738-01	MW4	VN088382.D	01 Dec 2025 14:54	Run @ Lower Dilution	JC\MD	Not Ok
15	Q3738-03	Field	VN088383.D	01 Dec 2025 15:15	vial A pH<2 FB	JC\MD	Ok
16	Q3738-02	MW5	VN088384.D	01 Dec 2025 15:35	vial A pH<2	JC\MD	Ok,M
17	Q3745-01	STORAGE-BLANK-SO	VN088385.D	01 Dec 2025 15:56	vial A pH<2 Tics Contamination	JC\MD	ReRun
18	Q3745-02	STORAGE-BLANK-WA	VN088386.D	01 Dec 2025 16:16	vial A pH<2	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN120125

Review By	Review On
Supervise By	Supervise On
SubDirectory VN120125	HP Acquire Method HP Processing Method 82N112425W.M
STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136451
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136452,VP136453

19	Q3745-03	STORAGE-BLANK-WA	VN088387.D	01 Dec 2025 16:37	vial A pH<2	JC\MD	Ok
20	Q3745-04	STORAGE-BLANK-SAM	VN088388.D	01 Dec 2025 16:58	vial A pH<2	JC\MD	Ok
21	Q3743-03	TB	VN088389.D	01 Dec 2025 17:18	vial A pH<2 TB	JC\MD	Ok
22	Q3743-04	TB	VN088390.D	01 Dec 2025 17:39	vial A pH<2 TB	JC\MD	Ok
23	Q3743-05	TB	VN088391.D	01 Dec 2025 18:00	SIM Method on login	JC\MD	Not Ok
24	Q3743-01	TW-1125-1	VN088392.D	01 Dec 2025 18:20	vial A pH<2	JC\MD	Ok
25	Q3743-02	TW-1125-2	VN088393.D	01 Dec 2025 18:41	vial A pH<2	JC\MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VN088394.D	01 Dec 2025 19:01		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY112625

Review By	Semsettin Yesilyurt	Review On	12/1/2025 9:50:11 AM		
Supervise By	John Carlone	Supervise On	12/1/2025 2:45:01 PM		
SubDirectory	VY112625	HP Acquire Method	MSVOA_Y	HP Processing Method	82y112625s.m

STD. NAME	STD REF.#
Tune/Reschk	VP136432
Initial Calibration Stds	VP136433,VP136434,VP136435,VP136436,VP136437,VP136438
CCC	VP136440
Internal Standard/PEM	VP136146
ICV/I.BLK	VP136439
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023808.D	26 Nov 2025 07:11		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY023809.D	26 Nov 2025 08:04		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY023810.D	26 Nov 2025 08:27		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY023811.D	26 Nov 2025 08:49		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY023812.D	26 Nov 2025 09:17		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY023813.D	26 Nov 2025 09:39		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY023814.D	26 Nov 2025 10:02		SY/MD	Ok,M
8	VIBLK	VIBLK	VY023815.D	26 Nov 2025 10:27		SY/MD	Ok
9	VSTDICV050	ICVVY112625	VY023816.D	26 Nov 2025 10:50		SY/MD	Ok,M
10	VY1126SBL01	VY1126SBL01	VY023817.D	26 Nov 2025 11:13		SY/MD	Ok
11	VY1126SBS01	VY1126SBS01	VY023818.D	26 Nov 2025 11:47		SY/MD	Ok,M
12	VY1126SBSD01	VY1126SBSD01	VY023819.D	26 Nov 2025 12:09		SY/MD	Ok,M
13	Q3604-04	S-1	VY023820.D	26 Nov 2025 12:34	not reported,vial A	SY/MD	Not Ok
14	Q3604-05	S-2	VY023821.D	26 Nov 2025 12:57	not reported,vial A	SY/MD	Not Ok
15	VIBLK	VIBLK	VY023822.D	26 Nov 2025 13:21		SY/MD	Ok
16	Q3721-01	EO-2-112525	VY023823.D	26 Nov 2025 13:44	vial A	SY/MD	Ok
17	Q3723-01	CC-1-112525	VY023824.D	26 Nov 2025 14:08	vial A	SY/MD	Ok
18	Q3725-04	STOCKPILE-SAMPLES	VY023825.D	26 Nov 2025 14:31	vial A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY112625

Review By	Semsettin Yesilyurt	Review On	12/1/2025 9:50:11 AM		
Supervise By	John Carlone	Supervise On	12/1/2025 2:45:01 PM		
SubDirectory	VY112625	HP Acquire Method	MSVOA_Y	HP Processing Method	82y112625s.m

STD. NAME	STD REF.#
Tune/Reschk	VP136432
Initial Calibration Stds	VP136433,VP136434,VP136435,VP136436,VP136437,VP136438
CCC	VP136440
Internal Standard/PEM	VP136146
ICV/I.BLK	VP136439
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	VIBLK	VIBLK	VY023826.D	26 Nov 2025 14:54		SY/MD	Ok
20	VSTDCCC050	VSTDCCC050EC	VY023827.D	26 Nov 2025 15:17		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY120125

Review By	John Carlone	Review On	12/2/2025 9:00:18 AM
Supervise By	Maresh Dadoda	Supervise On	12/2/2025 10:16:40 AM
SubDirectory	VY120125	HP Acquire Method	MSVOA_Y HP Processing Method 82y112625s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136454
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136455,VP136456 VP136146

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023828.D	01 Dec 2025 07:24		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY023829.D	01 Dec 2025 07:55		SY/MD	Ok,M
3	VY1201SBL01	VY1201SBL01	VY023830.D	01 Dec 2025 09:13		SY/MD	Ok
4	VY1201SBS01	VY1201SBS01	VY023831.D	01 Dec 2025 09:37		SY/MD	Ok,M
5	VY1201SBSD01	VY1201SBSD01	VY023832.D	01 Dec 2025 10:00		SY/MD	Ok,M
6	Q3719-09RE	SB-1125-15ARE	VY023833.D	01 Dec 2025 10:38	Internal standard fail;Surrogate fail,vial B	SY/MD	Confirms
7	Q3719-13	SB-1125-17A	VY023834.D	01 Dec 2025 11:25	vial B	SY/MD	Ok
8	Q3735-01	BU-3-112625	VY023835.D	01 Dec 2025 12:02	Internal Standard Fail,vial A	SY/MD	ReRun
9	Q3740-01	WC1	VY023836.D	01 Dec 2025 12:26	vial A	SY/MD	Ok
10	Q3741-02	B50-SP13-112025	VY023837.D	01 Dec 2025 12:49	vial A	SY/MD	Ok
11	Q3741-03	B50-SP14-112025	VY023838.D	01 Dec 2025 13:12	vial A	SY/MD	Ok
12	Q3741-04	B50-SP9-112425	VY023839.D	01 Dec 2025 13:36	Surrogate fail,vial A	SY/MD	ReRun
13	Q3741-05	B50-SP11-112525	VY023840.D	01 Dec 2025 13:59	Internal standard fail;Surrogate fail,vial A	SY/MD	ReRun
14	Q3741-01	B50-SP3-111925	VY023841.D	01 Dec 2025 14:23	vial A	SY/MD	Ok
15	Q3741-06	B50-SP4-112525	VY023842.D	01 Dec 2025 14:46	Internal standard fail;Surrogate fail,vial A	SY/MD	ReRun
16	Q3742-01	SB-1125-23	VY023843.D	01 Dec 2025 15:10	vial A	SY/MD	Ok
17	Q3742-02	SB-1125-19	VY023844.D	01 Dec 2025 15:33	vial A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY120125

Review By	John Carlone	Review On	12/2/2025 9:00:18 AM
Supervise By	Maresh Dadoda	Supervise On	12/2/2025 10:16:40 AM
SubDirectory	VY120125	HP Acquire Method	MSVOA_Y HP Processing Method 82y112625s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136454
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136455,VP136456 VP136146

18	Q3742-03	SB-1125-8	VY023845.D	01 Dec 2025 15:57	vial A	SY/MD	Ok
19	Q3742-04	SB-1125-9	VY023846.D	01 Dec 2025 16:20	vial A	SY/MD	Ok
20	VSTDCCC050	VSTDCCC050EC	VY023847.D	01 Dec 2025 17:47		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY120225

Review By	Semsettin Yesilyurt	Review On	12/3/2025 12:40:53 PM
Supervise By	Maresh Dadoda	Supervise On	12/3/2025 2:29:19 PM
SubDirectory	VY120225	HP Acquire Method	MSVOA_Y HP Processing Method 82y112625s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	AP136460
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136461,VP136462 VP136146

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023848.D	02 Dec 2025 06:50		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY023849.D	02 Dec 2025 09:14	CCAL failed high for com.#16	SY/MD	Ok,M
3	VY1202SBL01	VY1202SBL01	VY023850.D	02 Dec 2025 10:24		SY/MD	Ok
4	VY1202SBS01	VY1202SBS01	VY023851.D	02 Dec 2025 10:53		SY/MD	Ok,M
5	VY1202SBSD01	VY1202SBSD01	VY023852.D	02 Dec 2025 11:16		SY/MD	Ok,M
6	Q3742-05	SB-1125-21	VY023853.D	02 Dec 2025 12:08	vial B	SY/MD	Ok
7	Q3735-01	BU-3-112625	VY023854.D	02 Dec 2025 12:31	vial B	SY/MD	Ok
8	Q3741-04RE	B50-SP9-112425RE	VY023855.D	02 Dec 2025 12:55	Internal standard fail;Surrogate fail,vial B	SY/MD	Confirms
9	Q3741-05RE	B50-SP11-112525RE	VY023856.D	02 Dec 2025 13:18	Surrogate fail,vial B	SY/MD	Confirms
10	Q3741-06RE	B50-SP4-112525RE	VY023857.D	02 Dec 2025 13:42	Internal standard fail;Not match with first run,vial B	SY/MD	Confirms
11	Q3746-01	ROW	VY023858.D	02 Dec 2025 14:05	vial A	SY/MD	Ok
12	Q3742-06	SB-1125-13	VY023859.D	02 Dec 2025 14:29	vial A	SY/MD	Ok
13	VIBLK	VIBLK	VY023860.D	02 Dec 2025 15:33		SY/MD	Ok
14	Q3742-08	SB-1125-20	VY023861.D	02 Dec 2025 15:57	vial A	SY/MD	Ok
15	Q3742-09	SB-1125-22	VY023862.D	02 Dec 2025 16:20	Internal Standard Fail; CCC Failed High,vial A	SY/MD	ReRun
16	VSTDCCC050	VSTDCCC050EC	VY023863.D	02 Dec 2025 17:20		SY/MD	Ok,M

M : Manual Integration