

Initial and Continuing Calibration Verification

Client: Remington & Vernick

SDG No.: Q3742

Project: Saddler Property

RunNo.: LB138097

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.093	0.099	94	90-110	12/03/2025
Sample ID: CCV1 Cyanide	mg/L	0.23	0.25	92	90-110	12/03/2025
Sample ID: CCV2 Cyanide	mg/L	0.24	0.25	96	90-110	12/03/2025
Sample ID: CCV3 Cyanide	mg/L	0.23	0.25	92	90-110	12/03/2025
Sample ID: CCV4 Cyanide	mg/L	0.24	0.25	96	90-110	12/03/2025
Sample ID: CCV5 Cyanide	mg/L	0.25	0.25	100	90-110	12/03/2025