

DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

J	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
U	Indicates the analyte was analyzed for, but not detected.
ND	Indicates the analyte was analyzed for, but not detected
E	Indicates the reported value is estimated because of the presence of interference
M	Indicates Duplicate injection precision not met.
N	Indicates the spiked sample recovery is not within control limits.
S	Indicates the reported value was determined by the Method of Standard Addition (MSA).
*	Indicates that the duplicate analysis is not within control limits.
+	Indicates the correlation coefficient for the MSA is less than 0.995.
D	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
M	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
OR	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
Q	Indicates the LCS did not meet the control limits requirements
H	Sample Analysis Out Of Hold Time

LAB CHRONICLE

OrderID:	Q3742	OrderDate:	12/1/2025 9:06:00 AM
Client:	Remington & Vernick	Project:	Saddler Property
Contact:	Justin Zarzecki	Location:	--Select--,A21

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3742-01	SB-1125-23	SOIL			11/25/25			11/26/25
			Mercury	7471B		12/02/25	12/03/25	
			Metals ICP-TAL	6010D		12/02/25	12/02/25	
Q3742-02	SB-1125-19	SOIL			11/25/25			11/26/25
			Mercury	7471B		12/02/25	12/03/25	
			Metals ICP-TAL	6010D		12/02/25	12/02/25	
Q3742-03	SB-1125-8	SOIL			11/25/25			11/26/25
			Mercury	7471B		12/02/25	12/03/25	
			Metals ICP-TAL	6010D		12/02/25	12/02/25	
Q3742-04	SB-1125-9	SOIL			11/25/25			11/26/25
			Mercury	7471B		12/02/25	12/03/25	
			Metals ICP-TAL	6010D		12/02/25	12/02/25	
Q3742-05	SB-1125-21	SOIL			11/25/25			11/26/25
			Mercury	7471B		12/02/25	12/03/25	
			Metals ICP-TAL	6010D		12/02/25	12/02/25	
Q3742-06	SB-1125-13	SOIL			11/25/25			11/26/25
			Mercury	7471B		12/02/25	12/03/25	
			Metals ICP-TAL	6010D		12/02/25	12/02/25	
Q3742-07	SB-1125-18	SOIL			11/25/25			11/26/25
			Mercury	7471B		12/02/25	12/03/25	
			Metals ICP-TAL	6010D		12/02/25	12/02/25	
Q3742-08	SB-1125-20	SOIL			11/25/25			11/26/25
			Mercury	7471B		12/02/25	12/03/25	
			Metals ICP-TAL	6010D		12/02/25	12/02/25	
Q3742-09	SB-1125-22	SOIL			11/25/25			11/26/25

LAB CHRONICLE

Q3742-10	SB-1125-2	SOIL	Mercury	7471B	12/02/25	12/03/25	11/25/25	11/26/25
			Metals ICP-TAL	6010D	12/02/25	12/02/25		
Q3742-16	SB-1125-1	SOIL	Mercury	7471B	12/02/25	12/03/25	11/25/25	11/26/25
			Metals ICP-TAL	6010D	12/02/25	12/02/25		
Q3742-17	SB-1125-25	SOIL	Mercury	7471B	12/02/25	12/03/25	11/25/25	11/26/25
			Metals ICP-TAL	6010D	12/02/25	12/02/25		
Q3742-18	SB-1125-24	SOIL	Mercury	7471B	12/02/25	12/03/25	11/25/25	11/26/25
			Metals ICP-TAL	6010D	12/02/25	12/02/25		

Hit Summary Sheet SW-846

SDG No.: Q3742

Order ID: Q3742

Client: Remington & Vernick

Project ID: Saddler Property

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : SB-1125-23								
Q3742-01	SB-1125-23	SOIL	Aluminum	7710		1.19	7.10	mg/Kg
Q3742-01	SB-1125-23	SOIL	Arsenic	12.6		0.27	1.42	mg/Kg
Q3742-01	SB-1125-23	SOIL	Barium	133		1.04	7.10	mg/Kg
Q3742-01	SB-1125-23	SOIL	Beryllium	0.87		0.035	0.43	mg/Kg
Q3742-01	SB-1125-23	SOIL	Cadmium	2.32		0.034	0.43	mg/Kg
Q3742-01	SB-1125-23	SOIL	Calcium	3270		15.8	142	mg/Kg
Q3742-01	SB-1125-23	SOIL	Chromium	8.16		0.067	0.71	mg/Kg
Q3742-01	SB-1125-23	SOIL	Cobalt	10.0		0.14	2.13	mg/Kg
Q3742-01	SB-1125-23	SOIL	Copper	51.8		0.31	1.42	mg/Kg
Q3742-01	SB-1125-23	SOIL	Iron	16000		5.66	7.10	mg/Kg
Q3742-01	SB-1125-23	SOIL	Lead	55.5		0.19	0.85	mg/Kg
Q3742-01	SB-1125-23	SOIL	Magnesium	338		17.0	142	mg/Kg
Q3742-01	SB-1125-23	SOIL	Manganese	150		0.20	1.42	mg/Kg
Q3742-01	SB-1125-23	SOIL	Mercury	0.031		0.011	0.020	mg/Kg
Q3742-01	SB-1125-23	SOIL	Nickel	18.5		0.19	2.84	mg/Kg
Q3742-01	SB-1125-23	SOIL	Potassium	594		39.3	142	mg/Kg
Q3742-01	SB-1125-23	SOIL	Silver	0.80		0.17	0.71	mg/Kg
Q3742-01	SB-1125-23	SOIL	Sodium	222		25.3	142	mg/Kg
Q3742-01	SB-1125-23	SOIL	Vanadium	54.9		0.36	2.84	mg/Kg
Q3742-01	SB-1125-23	SOIL	Zinc	341		0.33	2.84	mg/Kg
Client ID : SB-1125-19								
Q3742-02	SB-1125-19	SOIL	Aluminum	8170		1.80	10.7	mg/Kg
Q3742-02	SB-1125-19	SOIL	Antimony	0.73	J	0.47	5.36	mg/Kg
Q3742-02	SB-1125-19	SOIL	Arsenic	94.9		0.41	2.14	mg/Kg
Q3742-02	SB-1125-19	SOIL	Barium	169		1.57	10.7	mg/Kg
Q3742-02	SB-1125-19	SOIL	Beryllium	1.22		0.054	0.64	mg/Kg
Q3742-02	SB-1125-19	SOIL	Cadmium	4.47		0.051	0.64	mg/Kg
Q3742-02	SB-1125-19	SOIL	Calcium	3340		23.8	214	mg/Kg
Q3742-02	SB-1125-19	SOIL	Chromium	331		0.10	1.07	mg/Kg
Q3742-02	SB-1125-19	SOIL	Cobalt	32.5		0.21	3.22	mg/Kg
Q3742-02	SB-1125-19	SOIL	Copper	232		0.47	2.14	mg/Kg
Q3742-02	SB-1125-19	SOIL	Iron	23400		8.56	10.7	mg/Kg
Q3742-02	SB-1125-19	SOIL	Lead	493		0.28	1.29	mg/Kg
Q3742-02	SB-1125-19	SOIL	Magnesium	2490		25.7	214	mg/Kg
Q3742-02	SB-1125-19	SOIL	Manganese	904		0.30	2.14	mg/Kg
Q3742-02	SB-1125-19	SOIL	Mercury	1.40		0.018	0.032	mg/Kg
Q3742-02	SB-1125-19	SOIL	Nickel	44.0		0.28	4.29	mg/Kg

Hit Summary Sheet
SW-846

SDG No.: Q3742

Order ID: Q3742

Client: Remington & Vernick

Project ID: Saddler Property

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q3742-02	SB-1125-19	SOIL	Potassium	802		59.4	214	mg/Kg
Q3742-02	SB-1125-19	SOIL	Silver	1.59		0.26	1.07	mg/Kg
Q3742-02	SB-1125-19	SOIL	Sodium	50.9	J	38.2	214	mg/Kg
Q3742-02	SB-1125-19	SOIL	Thallium	0.57	J	0.49	4.29	mg/Kg
Q3742-02	SB-1125-19	SOIL	Vanadium	18.6		0.54	4.29	mg/Kg
Q3742-02	SB-1125-19	SOIL	Zinc	1180		0.49	4.29	mg/Kg
Client ID : SB-1125-8								
Q3742-03	SB-1125-8	SOIL	Aluminum	1830		0.88	5.26	mg/Kg
Q3742-03	SB-1125-8	SOIL	Arsenic	20.0		0.20	1.05	mg/Kg
Q3742-03	SB-1125-8	SOIL	Barium	66.8		0.77	5.26	mg/Kg
Q3742-03	SB-1125-8	SOIL	Beryllium	0.81		0.026	0.32	mg/Kg
Q3742-03	SB-1125-8	SOIL	Cadmium	0.65		0.025	0.32	mg/Kg
Q3742-03	SB-1125-8	SOIL	Calcium	532		11.7	105	mg/Kg
Q3742-03	SB-1125-8	SOIL	Chromium	3.95		0.049	0.53	mg/Kg
Q3742-03	SB-1125-8	SOIL	Cobalt	4.27		0.11	1.58	mg/Kg
Q3742-03	SB-1125-8	SOIL	Copper	47.2		0.23	1.05	mg/Kg
Q3742-03	SB-1125-8	SOIL	Iron	25700		4.20	5.26	mg/Kg
Q3742-03	SB-1125-8	SOIL	Lead	39.1		0.14	0.63	mg/Kg
Q3742-03	SB-1125-8	SOIL	Magnesium	147		12.6	105	mg/Kg
Q3742-03	SB-1125-8	SOIL	Manganese	29.3		0.15	1.05	mg/Kg
Q3742-03	SB-1125-8	SOIL	Mercury	0.080		0.0080	0.015	mg/Kg
Q3742-03	SB-1125-8	SOIL	Nickel	10.7		0.14	2.11	mg/Kg
Q3742-03	SB-1125-8	SOIL	Potassium	313		29.2	105	mg/Kg
Q3742-03	SB-1125-8	SOIL	Silver	1.47		0.13	0.53	mg/Kg
Q3742-03	SB-1125-8	SOIL	Sodium	39.7	J	18.7	105	mg/Kg
Q3742-03	SB-1125-8	SOIL	Thallium	0.78	J	0.24	2.11	mg/Kg
Q3742-03	SB-1125-8	SOIL	Vanadium	24.5		0.26	2.11	mg/Kg
Q3742-03	SB-1125-8	SOIL	Zinc	24.5		0.24	2.11	mg/Kg
Client ID : SB-1125-9								
Q3742-04	SB-1125-9	SOIL	Aluminum	2710		0.96	5.73	mg/Kg
Q3742-04	SB-1125-9	SOIL	Antimony	3.59		0.25	2.87	mg/Kg
Q3742-04	SB-1125-9	SOIL	Arsenic	57.1		0.22	1.15	mg/Kg
Q3742-04	SB-1125-9	SOIL	Barium	103		0.84	5.73	mg/Kg
Q3742-04	SB-1125-9	SOIL	Beryllium	0.92		0.029	0.34	mg/Kg
Q3742-04	SB-1125-9	SOIL	Cadmium	1.10		0.028	0.34	mg/Kg
Q3742-04	SB-1125-9	SOIL	Calcium	1160		12.7	115	mg/Kg
Q3742-04	SB-1125-9	SOIL	Chromium	10.7		0.054	0.57	mg/Kg
Q3742-04	SB-1125-9	SOIL	Cobalt	5.12		0.12	1.72	mg/Kg
Q3742-04	SB-1125-9	SOIL	Copper	83.3		0.25	1.15	mg/Kg

Hit Summary Sheet
SW-846

SDG No.: Q3742

Order ID: Q3742

Client: Remington & Vernick

Project ID: Saddler Property

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q3742-04	SB-1125-9	SOIL	Iron	26000		4.57	5.73	mg/Kg
Q3742-04	SB-1125-9	SOIL	Lead	206		0.15	0.69	mg/Kg
Q3742-04	SB-1125-9	SOIL	Magnesium	370		13.8	115	mg/Kg
Q3742-04	SB-1125-9	SOIL	Manganese	46.0		0.16	1.15	mg/Kg
Q3742-04	SB-1125-9	SOIL	Mercury	0.16		0.0090	0.017	mg/Kg
Q3742-04	SB-1125-9	SOIL	Nickel	16.1		0.15	2.29	mg/Kg
Q3742-04	SB-1125-9	SOIL	Potassium	488		31.7	115	mg/Kg
Q3742-04	SB-1125-9	SOIL	Selenium	1.05	J	0.30	1.15	mg/Kg
Q3742-04	SB-1125-9	SOIL	Silver	1.50		0.14	0.57	mg/Kg
Q3742-04	SB-1125-9	SOIL	Sodium	60.1	J	20.4	115	mg/Kg
Q3742-04	SB-1125-9	SOIL	Vanadium	68.8		0.29	2.29	mg/Kg
Q3742-04	SB-1125-9	SOIL	Zinc	85.6		0.26	2.29	mg/Kg
Client ID : SB-1125-21								
Q3742-05	SB-1125-21	SOIL	Aluminum	2690		0.90	5.38	mg/Kg
Q3742-05	SB-1125-21	SOIL	Antimony	0.70	J	0.24	2.69	mg/Kg
Q3742-05	SB-1125-21	SOIL	Arsenic	17.6		0.20	1.08	mg/Kg
Q3742-05	SB-1125-21	SOIL	Barium	57.9		0.79	5.38	mg/Kg
Q3742-05	SB-1125-21	SOIL	Beryllium	0.84		0.027	0.32	mg/Kg
Q3742-05	SB-1125-21	SOIL	Cadmium	0.86		0.026	0.32	mg/Kg
Q3742-05	SB-1125-21	SOIL	Calcium	1650		11.9	108	mg/Kg
Q3742-05	SB-1125-21	SOIL	Chromium	7.28		0.051	0.54	mg/Kg
Q3742-05	SB-1125-21	SOIL	Cobalt	5.74		0.11	1.61	mg/Kg
Q3742-05	SB-1125-21	SOIL	Copper	73.7		0.24	1.08	mg/Kg
Q3742-05	SB-1125-21	SOIL	Iron	27200		4.29	5.38	mg/Kg
Q3742-05	SB-1125-21	SOIL	Lead	177		0.14	0.65	mg/Kg
Q3742-05	SB-1125-21	SOIL	Magnesium	339		12.9	108	mg/Kg
Q3742-05	SB-1125-21	SOIL	Manganese	41.4		0.15	1.08	mg/Kg
Q3742-05	SB-1125-21	SOIL	Mercury	0.16		0.010	0.017	mg/Kg
Q3742-05	SB-1125-21	SOIL	Nickel	16.6		0.14	2.15	mg/Kg
Q3742-05	SB-1125-21	SOIL	Potassium	443		29.8	108	mg/Kg
Q3742-05	SB-1125-21	SOIL	Silver	1.50		0.13	0.54	mg/Kg
Q3742-05	SB-1125-21	SOIL	Sodium	58.6	J	19.1	108	mg/Kg
Q3742-05	SB-1125-21	SOIL	Vanadium	49.0		0.27	2.15	mg/Kg
Q3742-05	SB-1125-21	SOIL	Zinc	51.2		0.25	2.15	mg/Kg
Client ID : SB-1125-13								
Q3742-06	SB-1125-13	SOIL	Aluminum	2370		0.89	5.32	mg/Kg
Q3742-06	SB-1125-13	SOIL	Arsenic	13.4		0.20	1.06	mg/Kg
Q3742-06	SB-1125-13	SOIL	Barium	64.9		0.78	5.32	mg/Kg
Q3742-06	SB-1125-13	SOIL	Beryllium	0.88		0.027	0.32	mg/Kg

Hit Summary Sheet
SW-846

SDG No.: Q3742

Order ID: Q3742

Client: Remington & Vernick

Project ID: Saddler Property

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q3742-06	SB-1125-13	SOIL	Cadmium	0.90		0.026	0.32	mg/Kg
Q3742-06	SB-1125-13	SOIL	Calcium	1770		11.8	106	mg/Kg
Q3742-06	SB-1125-13	SOIL	Chromium	12.4		0.050	0.53	mg/Kg
Q3742-06	SB-1125-13	SOIL	Cobalt	5.45		0.11	1.59	mg/Kg
Q3742-06	SB-1125-13	SOIL	Copper	58.0		0.23	1.06	mg/Kg
Q3742-06	SB-1125-13	SOIL	Iron	30100		4.24	5.32	mg/Kg
Q3742-06	SB-1125-13	SOIL	Lead	89.5		0.14	0.64	mg/Kg
Q3742-06	SB-1125-13	SOIL	Magnesium	229		12.8	106	mg/Kg
Q3742-06	SB-1125-13	SOIL	Manganese	78.4		0.15	1.06	mg/Kg
Q3742-06	SB-1125-13	SOIL	Mercury	0.068		0.0080	0.014	mg/Kg
Q3742-06	SB-1125-13	SOIL	Nickel	16.8		0.14	2.13	mg/Kg
Q3742-06	SB-1125-13	SOIL	Potassium	509		29.5	106	mg/Kg
Q3742-06	SB-1125-13	SOIL	Silver	1.73		0.13	0.53	mg/Kg
Q3742-06	SB-1125-13	SOIL	Sodium	117		18.9	106	mg/Kg
Q3742-06	SB-1125-13	SOIL	Vanadium	27.5		0.27	2.13	mg/Kg
Q3742-06	SB-1125-13	SOIL	Zinc	54.4		0.25	2.13	mg/Kg
Client ID : SB-1125-18								
Q3742-07	SB-1125-18	SOIL	Aluminum	10900		1.18	7.01	mg/Kg
Q3742-07	SB-1125-18	SOIL	Arsenic	12.8		0.27	1.40	mg/Kg
Q3742-07	SB-1125-18	SOIL	Barium	123		1.02	7.01	mg/Kg
Q3742-07	SB-1125-18	SOIL	Beryllium	0.92		0.035	0.42	mg/Kg
Q3742-07	SB-1125-18	SOIL	Cadmium	2.04		0.034	0.42	mg/Kg
Q3742-07	SB-1125-18	SOIL	Calcium	8220		15.6	140	mg/Kg
Q3742-07	SB-1125-18	SOIL	Chromium	63.8		0.066	0.70	mg/Kg
Q3742-07	SB-1125-18	SOIL	Cobalt	10.2		0.14	2.10	mg/Kg
Q3742-07	SB-1125-18	SOIL	Copper	79.9		0.31	1.40	mg/Kg
Q3742-07	SB-1125-18	SOIL	Iron	19800		5.59	7.01	mg/Kg
Q3742-07	SB-1125-18	SOIL	Lead	85.6		0.18	0.84	mg/Kg
Q3742-07	SB-1125-18	SOIL	Magnesium	5510		16.8	140	mg/Kg
Q3742-07	SB-1125-18	SOIL	Manganese	457		0.20	1.40	mg/Kg
Q3742-07	SB-1125-18	SOIL	Mercury	0.24		0.010	0.019	mg/Kg
Q3742-07	SB-1125-18	SOIL	Nickel	23.0		0.18	2.80	mg/Kg
Q3742-07	SB-1125-18	SOIL	Potassium	1480		38.8	140	mg/Kg
Q3742-07	SB-1125-18	SOIL	Silver	1.23		0.17	0.70	mg/Kg
Q3742-07	SB-1125-18	SOIL	Sodium	129	J	24.9	140	mg/Kg
Q3742-07	SB-1125-18	SOIL	Thallium	0.36	J	0.32	2.80	mg/Kg
Q3742-07	SB-1125-18	SOIL	Vanadium	25.7		0.35	2.80	mg/Kg
Q3742-07	SB-1125-18	SOIL	Zinc	218		0.32	2.80	mg/Kg

Client ID : SB-1125-20

Hit Summary Sheet
SW-846

SDG No.: Q3742

Order ID: Q3742

Client: Remington & Vernick

Project ID: Saddler Property

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q3742-08	SB-1125-20	SOIL	Aluminum	8170		0.89	5.32	mg/Kg
Q3742-08	SB-1125-20	SOIL	Arsenic	29.1		0.20	1.06	mg/Kg
Q3742-08	SB-1125-20	SOIL	Barium	97.1		0.78	5.32	mg/Kg
Q3742-08	SB-1125-20	SOIL	Beryllium	0.94		0.027	0.32	mg/Kg
Q3742-08	SB-1125-20	SOIL	Cadmium	1.78		0.026	0.32	mg/Kg
Q3742-08	SB-1125-20	SOIL	Calcium	2290		11.8	106	mg/Kg
Q3742-08	SB-1125-20	SOIL	Chromium	72.1		0.050	0.53	mg/Kg
Q3742-08	SB-1125-20	SOIL	Cobalt	13.4		0.11	1.59	mg/Kg
Q3742-08	SB-1125-20	SOIL	Copper	170		0.23	1.06	mg/Kg
Q3742-08	SB-1125-20	SOIL	Iron	23200		4.24	5.32	mg/Kg
Q3742-08	SB-1125-20	SOIL	Lead	292		0.14	0.64	mg/Kg
Q3742-08	SB-1125-20	SOIL	Magnesium	3200		12.8	106	mg/Kg
Q3742-08	SB-1125-20	SOIL	Manganese	323		0.15	1.06	mg/Kg
Q3742-08	SB-1125-20	SOIL	Mercury	0.62		0.0090	0.015	mg/Kg
Q3742-08	SB-1125-20	SOIL	Nickel	22.9		0.14	2.13	mg/Kg
Q3742-08	SB-1125-20	SOIL	Potassium	2070		29.4	106	mg/Kg
Q3742-08	SB-1125-20	SOIL	Silver	1.95		0.13	0.53	mg/Kg
Q3742-08	SB-1125-20	SOIL	Sodium	116		18.9	106	mg/Kg
Q3742-08	SB-1125-20	SOIL	Vanadium	22.4		0.27	2.13	mg/Kg
Q3742-08	SB-1125-20	SOIL	Zinc	354		0.25	2.13	mg/Kg
Client ID : SB-1125-22								
Q3742-09	SB-1125-22	SOIL	Aluminum	7130		1.00	5.97	mg/Kg
Q3742-09	SB-1125-22	SOIL	Arsenic	7.51		0.23	1.19	mg/Kg
Q3742-09	SB-1125-22	SOIL	Barium	120		0.87	5.97	mg/Kg
Q3742-09	SB-1125-22	SOIL	Beryllium	0.67		0.030	0.36	mg/Kg
Q3742-09	SB-1125-22	SOIL	Cadmium	0.84		0.029	0.36	mg/Kg
Q3742-09	SB-1125-22	SOIL	Calcium	2420		13.3	119	mg/Kg
Q3742-09	SB-1125-22	SOIL	Chromium	24.2		0.056	0.60	mg/Kg
Q3742-09	SB-1125-22	SOIL	Cobalt	7.75		0.12	1.79	mg/Kg
Q3742-09	SB-1125-22	SOIL	Copper	21.4		0.26	1.19	mg/Kg
Q3742-09	SB-1125-22	SOIL	Iron	17300		4.77	5.97	mg/Kg
Q3742-09	SB-1125-22	SOIL	Lead	408		0.16	0.72	mg/Kg
Q3742-09	SB-1125-22	SOIL	Magnesium	2560		14.3	119	mg/Kg
Q3742-09	SB-1125-22	SOIL	Manganese	702		0.17	1.19	mg/Kg
Q3742-09	SB-1125-22	SOIL	Mercury	1.03	D	0.019	0.035	mg/Kg
Q3742-09	SB-1125-22	SOIL	Nickel	17.1		0.16	2.39	mg/Kg
Q3742-09	SB-1125-22	SOIL	Potassium	938		33.1	119	mg/Kg
Q3742-09	SB-1125-22	SOIL	Silver	1.01		0.14	0.60	mg/Kg
Q3742-09	SB-1125-22	SOIL	Sodium	105	J	21.3	119	mg/Kg
Q3742-09	SB-1125-22	SOIL	Vanadium	16.3		0.30	2.39	mg/Kg

Hit Summary Sheet
SW-846

SDG No.: Q3742
Client: Remington & Vernick

Order ID: Q3742
Project ID: Saddler Property

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q3742-09	SB-1125-22	SOIL	Zinc	117		0.28	2.39	mg/Kg
Client ID : SB-1125-2								
Q3742-10	SB-1125-2	SOIL	Aluminum	7130		0.97	5.77	mg/Kg
Q3742-10	SB-1125-2	SOIL	Arsenic	28.3		0.22	1.15	mg/Kg
Q3742-10	SB-1125-2	SOIL	Barium	91.7		0.84	5.77	mg/Kg
Q3742-10	SB-1125-2	SOIL	Beryllium	0.77		0.029	0.35	mg/Kg
Q3742-10	SB-1125-2	SOIL	Cadmium	1.28		0.028	0.35	mg/Kg
Q3742-10	SB-1125-2	SOIL	Calcium	1650		12.8	115	mg/Kg
Q3742-10	SB-1125-2	SOIL	Chromium	93.7		0.054	0.58	mg/Kg
Q3742-10	SB-1125-2	SOIL	Cobalt	11.9		0.12	1.73	mg/Kg
Q3742-10	SB-1125-2	SOIL	Copper	38.7		0.25	1.15	mg/Kg
Q3742-10	SB-1125-2	SOIL	Iron	17000		4.60	5.77	mg/Kg
Q3742-10	SB-1125-2	SOIL	Lead	99.5		0.15	0.69	mg/Kg
Q3742-10	SB-1125-2	SOIL	Magnesium	2270		13.8	115	mg/Kg
Q3742-10	SB-1125-2	SOIL	Manganese	299		0.16	1.15	mg/Kg
Q3742-10	SB-1125-2	SOIL	Mercury	0.44		0.0090	0.017	mg/Kg
Q3742-10	SB-1125-2	SOIL	Nickel	22.2		0.15	2.31	mg/Kg
Q3742-10	SB-1125-2	SOIL	Potassium	845		31.9	115	mg/Kg
Q3742-10	SB-1125-2	SOIL	Silver	1.07		0.14	0.58	mg/Kg
Q3742-10	SB-1125-2	SOIL	Sodium	74.9	J	20.5	115	mg/Kg
Q3742-10	SB-1125-2	SOIL	Vanadium	18.1		0.29	2.31	mg/Kg
Q3742-10	SB-1125-2	SOIL	Zinc	282		0.27	2.31	mg/Kg
Client ID : SB-1125-1								
Q3742-16	SB-1125-1	SOIL	Aluminum	8020		1.10	6.54	mg/Kg
Q3742-16	SB-1125-1	SOIL	Arsenic	10.6		0.25	1.31	mg/Kg
Q3742-16	SB-1125-1	SOIL	Barium	71.1		0.96	6.54	mg/Kg
Q3742-16	SB-1125-1	SOIL	Beryllium	0.72		0.033	0.39	mg/Kg
Q3742-16	SB-1125-1	SOIL	Cadmium	0.81		0.031	0.39	mg/Kg
Q3742-16	SB-1125-1	SOIL	Calcium	2230		14.5	131	mg/Kg
Q3742-16	SB-1125-1	SOIL	Chromium	25.9		0.061	0.65	mg/Kg
Q3742-16	SB-1125-1	SOIL	Cobalt	8.69		0.13	1.96	mg/Kg
Q3742-16	SB-1125-1	SOIL	Copper	24.7		0.29	1.31	mg/Kg
Q3742-16	SB-1125-1	SOIL	Iron	18200		5.22	6.54	mg/Kg
Q3742-16	SB-1125-1	SOIL	Lead	129		0.17	0.79	mg/Kg
Q3742-16	SB-1125-1	SOIL	Magnesium	3750		15.7	131	mg/Kg
Q3742-16	SB-1125-1	SOIL	Manganese	461		0.18	1.31	mg/Kg
Q3742-16	SB-1125-1	SOIL	Mercury	0.22		0.011	0.019	mg/Kg
Q3742-16	SB-1125-1	SOIL	Nickel	20.4		0.17	2.62	mg/Kg
Q3742-16	SB-1125-1	SOIL	Potassium	2540		36.2	131	mg/Kg

Hit Summary Sheet
SW-846

SDG No.: Q3742
Client: Remington & Vernick

Order ID: Q3742
Project ID: Saddler Property

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q3742-16	SB-1125-1	SOIL	Silver	1.06		0.16	0.65	mg/Kg
Q3742-16	SB-1125-1	SOIL	Sodium	145		23.3	131	mg/Kg
Q3742-16	SB-1125-1	SOIL	Vanadium	19.7		0.33	2.62	mg/Kg
Q3742-16	SB-1125-1	SOIL	Zinc	138		0.30	2.62	mg/Kg
Client ID : SB-1125-25								
Q3742-17	SB-1125-25	SOIL	Aluminum	7600		1.18	7.05	mg/Kg
Q3742-17	SB-1125-25	SOIL	Arsenic	38.1		0.27	1.41	mg/Kg
Q3742-17	SB-1125-25	SOIL	Barium	80.2		1.03	7.05	mg/Kg
Q3742-17	SB-1125-25	SOIL	Beryllium	0.86		0.035	0.42	mg/Kg
Q3742-17	SB-1125-25	SOIL	Cadmium	1.49		0.034	0.42	mg/Kg
Q3742-17	SB-1125-25	SOIL	Calcium	2110		15.7	141	mg/Kg
Q3742-17	SB-1125-25	SOIL	Chromium	128		0.066	0.71	mg/Kg
Q3742-17	SB-1125-25	SOIL	Cobalt	15.0		0.14	2.12	mg/Kg
Q3742-17	SB-1125-25	SOIL	Copper	64.0		0.31	1.41	mg/Kg
Q3742-17	SB-1125-25	SOIL	Iron	16500		5.63	7.05	mg/Kg
Q3742-17	SB-1125-25	SOIL	Lead	123		0.18	0.85	mg/Kg
Q3742-17	SB-1125-25	SOIL	Magnesium	2360		16.9	141	mg/Kg
Q3742-17	SB-1125-25	SOIL	Manganese	394		0.20	1.41	mg/Kg
Q3742-17	SB-1125-25	SOIL	Mercury	0.89		0.011	0.020	mg/Kg
Q3742-17	SB-1125-25	SOIL	Nickel	26.6		0.18	2.82	mg/Kg
Q3742-17	SB-1125-25	SOIL	Potassium	910		39.1	141	mg/Kg
Q3742-17	SB-1125-25	SOIL	Silver	1.12		0.17	0.71	mg/Kg
Q3742-17	SB-1125-25	SOIL	Sodium	76.7	J	25.1	141	mg/Kg
Q3742-17	SB-1125-25	SOIL	Vanadium	17.4		0.35	2.82	mg/Kg
Q3742-17	SB-1125-25	SOIL	Zinc	305		0.32	2.82	mg/Kg
Client ID : SB-1125-24								
Q3742-18	SB-1125-24	SOIL	Aluminum	8950		1.11	6.58	mg/Kg
Q3742-18	SB-1125-24	SOIL	Arsenic	41.4		0.25	1.32	mg/Kg
Q3742-18	SB-1125-24	SOIL	Barium	115		0.96	6.58	mg/Kg
Q3742-18	SB-1125-24	SOIL	Beryllium	1.16		0.033	0.40	mg/Kg
Q3742-18	SB-1125-24	SOIL	Cadmium	2.46		0.032	0.40	mg/Kg
Q3742-18	SB-1125-24	SOIL	Calcium	5210		14.6	132	mg/Kg
Q3742-18	SB-1125-24	SOIL	Chromium	235		0.062	0.66	mg/Kg
Q3742-18	SB-1125-24	SOIL	Cobalt	16.9		0.13	1.97	mg/Kg
Q3742-18	SB-1125-24	SOIL	Copper	103		0.29	1.32	mg/Kg
Q3742-18	SB-1125-24	SOIL	Iron	22100		5.25	6.58	mg/Kg
Q3742-18	SB-1125-24	SOIL	Lead	226		0.17	0.79	mg/Kg
Q3742-18	SB-1125-24	SOIL	Magnesium	2710		15.8	132	mg/Kg
Q3742-18	SB-1125-24	SOIL	Manganese	546		0.18	1.32	mg/Kg

Hit Summary Sheet
SW-846

SDG No.: Q3742

Order ID: Q3742

Client: Remington & Vernick

Project ID: Saddler Property

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q3742-18	SB-1125-24	SOIL	Mercury	0.75		0.010	0.019	mg/Kg
Q3742-18	SB-1125-24	SOIL	Nickel	30.1		0.17	2.63	mg/Kg
Q3742-18	SB-1125-24	SOIL	Potassium	1290		36.5	132	mg/Kg
Q3742-18	SB-1125-24	SOIL	Silver	1.69		0.16	0.66	mg/Kg
Q3742-18	SB-1125-24	SOIL	Sodium	132		23.4	132	mg/Kg
Q3742-18	SB-1125-24	SOIL	Thallium	0.44	J	0.30	2.63	mg/Kg
Q3742-18	SB-1125-24	SOIL	Vanadium	23.2		0.33	2.63	mg/Kg
Q3742-18	SB-1125-24	SOIL	Zinc	434		0.30	2.63	mg/Kg



SAMPLE DATA

Report of Analysis

Client: Remington & Vernick
Project: Saddler Property
Client Sample ID: SB-1125-23
Lab Sample ID: Q3742-01
Level (low/med): low

Date Collected: 11/25/25
Date Received: 11/26/25
SDG No.: Q3742
Matrix: SOIL
% Solid: 70.1

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	7710		1	1.19	7.10	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7440-36-0	Antimony	0.31	UN*	1	0.31	3.55	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7440-38-2	Arsenic	12.6	N	1	0.27	1.42	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7440-39-3	Barium	133		1	1.04	7.10	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7440-41-7	Beryllium	0.87		1	0.035	0.43	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7440-43-9	Cadmium	2.32		1	0.034	0.43	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7440-70-2	Calcium	3270	*	1	15.8	142	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7440-47-3	Chromium	8.16		1	0.067	0.71	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7440-48-4	Cobalt	10.0		1	0.14	2.13	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7440-50-8	Copper	51.8		1	0.31	1.42	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7439-89-6	Iron	16000		1	5.66	7.10	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7439-92-1	Lead	55.5	N*	1	0.19	0.85	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7439-95-4	Magnesium	338		1	17.0	142	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7439-96-5	Manganese	150		1	0.20	1.42	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7439-97-6	Mercury	0.031		1	0.011	0.020	mg/Kg	12/02/25 13:35	12/03/25 09:06	7471B	M7471B
7440-02-0	Nickel	18.5		1	0.19	2.84	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7440-09-7	Potassium	594	N	1	39.3	142	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7782-49-2	Selenium	0.37	U	1	0.37	1.42	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7440-22-4	Silver	0.80		1	0.17	0.71	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7440-23-5	Sodium	222	N	1	25.3	142	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7440-28-0	Thallium	0.33	U	1	0.33	2.84	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7440-62-2	Vanadium	54.9	N	1	0.36	2.84	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050
7440-66-6	Zinc	341		1	0.33	2.84	mg/Kg	12/02/25 10:30	12/02/25 15:24	6010D	SW3050

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: Remington & Vernick
Project: Saddler Property
Client Sample ID: SB-1125-19
Lab Sample ID: Q3742-02
Level (low/med): low

Date Collected: 11/25/25
Date Received: 11/26/25
SDG No.: Q3742
Matrix: SOIL
% Solid: 40.9

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	8170		1	1.80	10.7	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7440-36-0	Antimony	0.73	JN*	1	0.47	5.36	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7440-38-2	Arsenic	94.9	N	1	0.41	2.14	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7440-39-3	Barium	169		1	1.57	10.7	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7440-41-7	Beryllium	1.22		1	0.054	0.64	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7440-43-9	Cadmium	4.47		1	0.051	0.64	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7440-70-2	Calcium	3340	*	1	23.8	214	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7440-47-3	Chromium	331		1	0.10	1.07	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7440-48-4	Cobalt	32.5		1	0.21	3.22	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7440-50-8	Copper	232		1	0.47	2.14	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7439-89-6	Iron	23400		1	8.56	10.7	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7439-92-1	Lead	493	N*	1	0.28	1.29	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7439-95-4	Magnesium	2490		1	25.7	214	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7439-96-5	Manganese	904		1	0.30	2.14	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7439-97-6	Mercury	1.40		1	0.018	0.032	mg/Kg	12/02/25 13:35	12/03/25 09:09	7471B	M7471B
7440-02-0	Nickel	44.0		1	0.28	4.29	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7440-09-7	Potassium	802	N	1	59.4	214	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7782-49-2	Selenium	0.56	U	1	0.56	2.14	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7440-22-4	Silver	1.59		1	0.26	1.07	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7440-23-5	Sodium	50.9	JN	1	38.2	214	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7440-28-0	Thallium	0.57	J	1	0.49	4.29	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7440-62-2	Vanadium	18.6	N	1	0.54	4.29	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050
7440-66-6	Zinc	1180		1	0.49	4.29	mg/Kg	12/02/25 10:30	12/02/25 15:28	6010D	SW3050

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: Remington & Vernick
Project: Saddler Property
Client Sample ID: SB-1125-8
Lab Sample ID: Q3742-03
Level (low/med): low

Date Collected: 11/25/25
Date Received: 11/26/25
SDG No.: Q3742
Matrix: SOIL
% Solid: 85.2

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	1830		1	0.88	5.26	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7440-36-0	Antimony	0.23	UN*	1	0.23	2.63	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7440-38-2	Arsenic	20.0	N	1	0.20	1.05	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7440-39-3	Barium	66.8		1	0.77	5.26	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7440-41-7	Beryllium	0.81		1	0.026	0.32	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7440-43-9	Cadmium	0.65		1	0.025	0.32	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7440-70-2	Calcium	532	*	1	11.7	105	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7440-47-3	Chromium	3.95		1	0.049	0.53	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7440-48-4	Cobalt	4.27		1	0.11	1.58	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7440-50-8	Copper	47.2		1	0.23	1.05	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7439-89-6	Iron	25700		1	4.20	5.26	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7439-92-1	Lead	39.1	N*	1	0.14	0.63	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7439-95-4	Magnesium	147		1	12.6	105	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7439-96-5	Manganese	29.3		1	0.15	1.05	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7439-97-6	Mercury	0.080		1	0.0080	0.015	mg/Kg	12/02/25 13:35	12/03/25 09:11	7471B	M7471B
7440-02-0	Nickel	10.7		1	0.14	2.11	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7440-09-7	Potassium	313	N	1	29.2	105	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7782-49-2	Selenium	0.27	U	1	0.27	1.05	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7440-22-4	Silver	1.47		1	0.13	0.53	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7440-23-5	Sodium	39.7	JN	1	18.7	105	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7440-28-0	Thallium	0.78	J	1	0.24	2.11	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7440-62-2	Vanadium	24.5	N	1	0.26	2.11	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050
7440-66-6	Zinc	24.5		1	0.24	2.11	mg/Kg	12/02/25 10:30	12/02/25 15:33	6010D	SW3050

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: Remington & Vernick
Project: Saddler Property
Client Sample ID: SB-1125-9
Lab Sample ID: Q3742-04
Level (low/med): low

Date Collected: 11/25/25
Date Received: 11/26/25
SDG No.: Q3742
Matrix: SOIL
% Solid: 78.6

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	2710		1	0.96	5.73	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7440-36-0	Antimony	3.59	N*	1	0.25	2.87	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7440-38-2	Arsenic	57.1	N	1	0.22	1.15	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7440-39-3	Barium	103		1	0.84	5.73	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7440-41-7	Beryllium	0.92		1	0.029	0.34	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7440-43-9	Cadmium	1.10		1	0.028	0.34	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7440-70-2	Calcium	1160	*	1	12.7	115	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7440-47-3	Chromium	10.7		1	0.054	0.57	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7440-48-4	Cobalt	5.12		1	0.12	1.72	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7440-50-8	Copper	83.3		1	0.25	1.15	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7439-89-6	Iron	26000		1	4.57	5.73	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7439-92-1	Lead	206	N*	1	0.15	0.69	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7439-95-4	Magnesium	370		1	13.8	115	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7439-96-5	Manganese	46.0		1	0.16	1.15	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7439-97-6	Mercury	0.16		1	0.0090	0.017	mg/Kg	12/02/25 13:35	12/03/25 09:13	7471B	M7471B
7440-02-0	Nickel	16.1		1	0.15	2.29	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7440-09-7	Potassium	488	N	1	31.7	115	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7782-49-2	Selenium	1.05	J	1	0.30	1.15	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7440-22-4	Silver	1.50		1	0.14	0.57	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7440-23-5	Sodium	60.1	JN	1	20.4	115	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7440-28-0	Thallium	0.26	U	1	0.26	2.29	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7440-62-2	Vanadium	68.8	N	1	0.29	2.29	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050
7440-66-6	Zinc	85.6		1	0.26	2.29	mg/Kg	12/02/25 10:30	12/02/25 15:37	6010D	SW3050

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: Remington & Vernick
Project: Saddler Property
Client Sample ID: SB-1125-21
Lab Sample ID: Q3742-05
Level (low/med): low

Date Collected: 11/25/25
Date Received: 11/26/25
SDG No.: Q3742
Matrix: SOIL
% Solid: 79.5

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	2690		1	0.90	5.38	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7440-36-0	Antimony	0.70	JN*	1	0.24	2.69	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7440-38-2	Arsenic	17.6	N	1	0.20	1.08	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7440-39-3	Barium	57.9		1	0.79	5.38	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7440-41-7	Beryllium	0.84		1	0.027	0.32	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7440-43-9	Cadmium	0.86		1	0.026	0.32	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7440-70-2	Calcium	1650	*	1	11.9	108	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7440-47-3	Chromium	7.28		1	0.051	0.54	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7440-48-4	Cobalt	5.74		1	0.11	1.61	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7440-50-8	Copper	73.7		1	0.24	1.08	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7439-89-6	Iron	27200		1	4.29	5.38	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7439-92-1	Lead	177	N*	1	0.14	0.65	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7439-95-4	Magnesium	339		1	12.9	108	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7439-96-5	Manganese	41.4		1	0.15	1.08	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7439-97-6	Mercury	0.16		1	0.010	0.017	mg/Kg	12/02/25 13:35	12/03/25 09:16	7471B	M7471B
7440-02-0	Nickel	16.6		1	0.14	2.15	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7440-09-7	Potassium	443	N	1	29.8	108	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7782-49-2	Selenium	0.28	U	1	0.28	1.08	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7440-22-4	Silver	1.50		1	0.13	0.54	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7440-23-5	Sodium	58.6	JN	1	19.1	108	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7440-28-0	Thallium	0.25	U	1	0.25	2.15	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7440-62-2	Vanadium	49.0	N	1	0.27	2.15	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050
7440-66-6	Zinc	51.2		1	0.25	2.15	mg/Kg	12/02/25 10:30	12/02/25 15:41	6010D	SW3050

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: Remington & Vernick
Project: Saddler Property
Client Sample ID: SB-1125-13
Lab Sample ID: Q3742-06
Level (low/med): low

Date Collected: 11/25/25
Date Received: 11/26/25
SDG No.: Q3742
Matrix: SOIL
% Solid: 85.5

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	2370		1	0.89	5.32	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7440-36-0	Antimony	0.23	UN*	1	0.23	2.66	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7440-38-2	Arsenic	13.4	N	1	0.20	1.06	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7440-39-3	Barium	64.9		1	0.78	5.32	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7440-41-7	Beryllium	0.88		1	0.027	0.32	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7440-43-9	Cadmium	0.90		1	0.026	0.32	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7440-70-2	Calcium	1770	*	1	11.8	106	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7440-47-3	Chromium	12.4		1	0.050	0.53	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7440-48-4	Cobalt	5.45		1	0.11	1.59	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7440-50-8	Copper	58.0		1	0.23	1.06	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7439-89-6	Iron	30100		1	4.24	5.32	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7439-92-1	Lead	89.5	N*	1	0.14	0.64	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7439-95-4	Magnesium	229		1	12.8	106	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7439-96-5	Manganese	78.4		1	0.15	1.06	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7439-97-6	Mercury	0.068		1	0.0080	0.014	mg/Kg	12/02/25 13:35	12/03/25 09:22	7471B	M7471B
7440-02-0	Nickel	16.8		1	0.14	2.13	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7440-09-7	Potassium	509	N	1	29.5	106	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7782-49-2	Selenium	0.28	U	1	0.28	1.06	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7440-22-4	Silver	1.73		1	0.13	0.53	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7440-23-5	Sodium	117	N	1	18.9	106	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7440-28-0	Thallium	0.25	U	1	0.25	2.13	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7440-62-2	Vanadium	27.5	N	1	0.27	2.13	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050
7440-66-6	Zinc	54.4		1	0.25	2.13	mg/Kg	12/02/25 10:30	12/02/25 15:45	6010D	SW3050

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: Remington & Vernick
Project: Saddler Property
Client Sample ID: SB-1125-18
Lab Sample ID: Q3742-07
Level (low/med): low

Date Collected: 11/25/25
Date Received: 11/26/25
SDG No.: Q3742
Matrix: SOIL
% Solid: 68.3

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	10900		1	1.18	7.01	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7440-36-0	Antimony	0.31	UN*	1	0.31	3.50	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7440-38-2	Arsenic	12.8	N	1	0.27	1.40	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7440-39-3	Barium	123		1	1.02	7.01	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7440-41-7	Beryllium	0.92		1	0.035	0.42	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7440-43-9	Cadmium	2.04		1	0.034	0.42	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7440-70-2	Calcium	8220	*	1	15.6	140	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7440-47-3	Chromium	63.8		1	0.066	0.70	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7440-48-4	Cobalt	10.2		1	0.14	2.10	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7440-50-8	Copper	79.9		1	0.31	1.40	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7439-89-6	Iron	19800		1	5.59	7.01	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7439-92-1	Lead	85.6	N*	1	0.18	0.84	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7439-95-4	Magnesium	5510		1	16.8	140	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7439-96-5	Manganese	457		1	0.20	1.40	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7439-97-6	Mercury	0.24		1	0.010	0.019	mg/Kg	12/02/25 13:35	12/03/25 09:25	7471B	M7471B
7440-02-0	Nickel	23.0		1	0.18	2.80	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7440-09-7	Potassium	1480	N	1	38.8	140	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7782-49-2	Selenium	0.36	U	1	0.36	1.40	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7440-22-4	Silver	1.23		1	0.17	0.70	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7440-23-5	Sodium	129	JN	1	24.9	140	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7440-28-0	Thallium	0.36	J	1	0.32	2.80	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7440-62-2	Vanadium	25.7	N	1	0.35	2.80	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050
7440-66-6	Zinc	218		1	0.32	2.80	mg/Kg	12/02/25 10:30	12/02/25 15:49	6010D	SW3050

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: Remington & Vernick
Project: Saddler Property
Client Sample ID: SB-1125-20
Lab Sample ID: Q3742-08
Level (low/med): low

Date Collected: 11/25/25
Date Received: 11/26/25
SDG No.: Q3742
Matrix: SOIL
% Solid: 80.4

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	8170		1	0.89	5.32	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7440-36-0	Antimony	0.23	UN*	1	0.23	2.66	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7440-38-2	Arsenic	29.1	N	1	0.20	1.06	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7440-39-3	Barium	97.1		1	0.78	5.32	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7440-41-7	Beryllium	0.94		1	0.027	0.32	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7440-43-9	Cadmium	1.78		1	0.026	0.32	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7440-70-2	Calcium	2290	*	1	11.8	106	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7440-47-3	Chromium	72.1		1	0.050	0.53	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7440-48-4	Cobalt	13.4		1	0.11	1.59	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7440-50-8	Copper	170		1	0.23	1.06	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7439-89-6	Iron	23200		1	4.24	5.32	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7439-92-1	Lead	292	N*	1	0.14	0.64	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7439-95-4	Magnesium	3200		1	12.8	106	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7439-96-5	Manganese	323		1	0.15	1.06	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7439-97-6	Mercury	0.62		1	0.0090	0.015	mg/Kg	12/02/25 13:35	12/03/25 09:27	7471B	M7471B
7440-02-0	Nickel	22.9		1	0.14	2.13	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7440-09-7	Potassium	2070	N	1	29.4	106	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7782-49-2	Selenium	0.28	U	1	0.28	1.06	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7440-22-4	Silver	1.95		1	0.13	0.53	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7440-23-5	Sodium	116	N	1	18.9	106	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7440-28-0	Thallium	0.25	U	1	0.25	2.13	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7440-62-2	Vanadium	22.4	N	1	0.27	2.13	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050
7440-66-6	Zinc	354		1	0.25	2.13	mg/Kg	12/02/25 10:30	12/02/25 15:53	6010D	SW3050

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: Remington & Vernick
Project: Saddler Property
Client Sample ID: SB-1125-22
Lab Sample ID: Q3742-09
Level (low/med): low

Date Collected: 11/25/25
Date Received: 11/26/25
SDG No.: Q3742
Matrix: SOIL
% Solid: 72.8

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	7130		1	1.00	5.97	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7440-36-0	Antimony	0.26	UN*	1	0.26	2.99	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7440-38-2	Arsenic	7.51	N	1	0.23	1.19	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7440-39-3	Barium	120		1	0.87	5.97	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7440-41-7	Beryllium	0.67		1	0.030	0.36	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7440-43-9	Cadmium	0.84		1	0.029	0.36	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7440-70-2	Calcium	2420	*	1	13.3	119	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7440-47-3	Chromium	24.2		1	0.056	0.60	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7440-48-4	Cobalt	7.75		1	0.12	1.79	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7440-50-8	Copper	21.4		1	0.26	1.19	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7439-89-6	Iron	17300		1	4.77	5.97	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7439-92-1	Lead	408	N*	1	0.16	0.72	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7439-95-4	Magnesium	2560		1	14.3	119	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7439-96-5	Manganese	702		1	0.17	1.19	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7439-97-6	Mercury	1.03	D	2	0.019	0.035	mg/Kg	12/02/25 13:35	12/03/25 10:11	7471B	M7471B
7440-02-0	Nickel	17.1		1	0.16	2.39	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7440-09-7	Potassium	938	N	1	33.1	119	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7782-49-2	Selenium	0.31	U	1	0.31	1.19	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7440-22-4	Silver	1.01		1	0.14	0.60	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7440-23-5	Sodium	105	JN	1	21.3	119	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7440-28-0	Thallium	0.28	U	1	0.28	2.39	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7440-62-2	Vanadium	16.3	N	1	0.30	2.39	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050
7440-66-6	Zinc	117		1	0.28	2.39	mg/Kg	12/02/25 10:30	12/02/25 15:57	6010D	SW3050

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: Remington & Vernick
Project: Saddler Property
Client Sample ID: SB-1125-2
Lab Sample ID: Q3742-10
Level (low/med): low

Date Collected: 11/25/25
Date Received: 11/26/25
SDG No.: Q3742
Matrix: SOIL
% Solid: 76.4

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	7130		1	0.97	5.77	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7440-36-0	Antimony	0.25	UN*	1	0.25	2.88	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7440-38-2	Arsenic	28.3	N	1	0.22	1.15	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7440-39-3	Barium	91.7		1	0.84	5.77	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7440-41-7	Beryllium	0.77		1	0.029	0.35	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7440-43-9	Cadmium	1.28		1	0.028	0.35	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7440-70-2	Calcium	1650	*	1	12.8	115	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7440-47-3	Chromium	93.7		1	0.054	0.58	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7440-48-4	Cobalt	11.9		1	0.12	1.73	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7440-50-8	Copper	38.7		1	0.25	1.15	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7439-89-6	Iron	17000		1	4.60	5.77	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7439-92-1	Lead	99.5	N*	1	0.15	0.69	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7439-95-4	Magnesium	2270		1	13.8	115	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7439-96-5	Manganese	299		1	0.16	1.15	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7439-97-6	Mercury	0.44		1	0.0090	0.017	mg/Kg	12/02/25 13:35	12/03/25 09:31	7471B	M7471B
7440-02-0	Nickel	22.2		1	0.15	2.31	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7440-09-7	Potassium	845	N	1	31.9	115	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7782-49-2	Selenium	0.30	U	1	0.30	1.15	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7440-22-4	Silver	1.07		1	0.14	0.58	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7440-23-5	Sodium	74.9	JN	1	20.5	115	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7440-28-0	Thallium	0.27	U	1	0.27	2.31	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7440-62-2	Vanadium	18.1	N	1	0.29	2.31	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050
7440-66-6	Zinc	282		1	0.27	2.31	mg/Kg	12/02/25 10:30	12/02/25 16:01	6010D	SW3050

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: Remington & Vernick
Project: Saddler Property
Client Sample ID: SB-1125-1
Lab Sample ID: Q3742-16
Level (low/med): low

Date Collected: 11/25/25
Date Received: 11/26/25
SDG No.: Q3742
Matrix: SOIL
% Solid: 73.5

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	8020		1	1.10	6.54	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7440-36-0	Antimony	0.29	UN*	1	0.29	3.27	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7440-38-2	Arsenic	10.6	N	1	0.25	1.31	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7440-39-3	Barium	71.1		1	0.96	6.54	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7440-41-7	Beryllium	0.72		1	0.033	0.39	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7440-43-9	Cadmium	0.81		1	0.031	0.39	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7440-70-2	Calcium	2230	*	1	14.5	131	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7440-47-3	Chromium	25.9		1	0.061	0.65	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7440-48-4	Cobalt	8.69		1	0.13	1.96	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7440-50-8	Copper	24.7		1	0.29	1.31	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7439-89-6	Iron	18200		1	5.22	6.54	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7439-92-1	Lead	129	N*	1	0.17	0.79	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7439-95-4	Magnesium	3750		1	15.7	131	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7439-96-5	Manganese	461		1	0.18	1.31	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7439-97-6	Mercury	0.22		1	0.011	0.019	mg/Kg	12/02/25 13:35	12/03/25 09:34	7471B	M7471B
7440-02-0	Nickel	20.4		1	0.17	2.62	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7440-09-7	Potassium	2540	N	1	36.2	131	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7782-49-2	Selenium	0.34	U	1	0.34	1.31	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7440-22-4	Silver	1.06		1	0.16	0.65	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7440-23-5	Sodium	145	N	1	23.3	131	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7440-28-0	Thallium	0.30	U	1	0.30	2.62	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7440-62-2	Vanadium	19.7	N	1	0.33	2.62	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050
7440-66-6	Zinc	138		1	0.30	2.62	mg/Kg	12/02/25 10:30	12/02/25 16:14	6010D	SW3050

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: Remington & Vernick
Project: Saddler Property
Client Sample ID: SB-1125-25
Lab Sample ID: Q3742-17
Level (low/med): low

Date Collected: 11/25/25
Date Received: 11/26/25
SDG No.: Q3742
Matrix: SOIL
% Solid: 63.3

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	7600		1	1.18	7.05	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7440-36-0	Antimony	0.31	UN*	1	0.31	3.53	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7440-38-2	Arsenic	38.1	N	1	0.27	1.41	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7440-39-3	Barium	80.2		1	1.03	7.05	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7440-41-7	Beryllium	0.86		1	0.035	0.42	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7440-43-9	Cadmium	1.49		1	0.034	0.42	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7440-70-2	Calcium	2110	*	1	15.7	141	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7440-47-3	Chromium	128		1	0.066	0.71	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7440-48-4	Cobalt	15.0		1	0.14	2.12	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7440-50-8	Copper	64.0		1	0.31	1.41	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7439-89-6	Iron	16500		1	5.63	7.05	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7439-92-1	Lead	123	N*	1	0.18	0.85	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7439-95-4	Magnesium	2360		1	16.9	141	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7439-96-5	Manganese	394		1	0.20	1.41	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7439-97-6	Mercury	0.89		1	0.011	0.020	mg/Kg	12/02/25 13:35	12/03/25 09:36	7471B	M7471B
7440-02-0	Nickel	26.6		1	0.18	2.82	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7440-09-7	Potassium	910	N	1	39.1	141	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7782-49-2	Selenium	0.37	U	1	0.37	1.41	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7440-22-4	Silver	1.12		1	0.17	0.71	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7440-23-5	Sodium	76.7	JN	1	25.1	141	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7440-28-0	Thallium	0.32	U	1	0.32	2.82	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7440-62-2	Vanadium	17.4	N	1	0.35	2.82	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050
7440-66-6	Zinc	305		1	0.32	2.82	mg/Kg	12/02/25 10:30	12/02/25 16:18	6010D	SW3050

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: Remington & Vernick
Project: Saddler Property
Client Sample ID: SB-1125-24
Lab Sample ID: Q3742-18
Level (low/med): low

Date Collected: 11/25/25
Date Received: 11/26/25
SDG No.: Q3742
Matrix: SOIL
% Solid: 62.8

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	8950		1	1.11	6.58	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7440-36-0	Antimony	0.29	UN*	1	0.29	3.29	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7440-38-2	Arsenic	41.4	N	1	0.25	1.32	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7440-39-3	Barium	115		1	0.96	6.58	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7440-41-7	Beryllium	1.16		1	0.033	0.40	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7440-43-9	Cadmium	2.46		1	0.032	0.40	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7440-70-2	Calcium	5210	*	1	14.6	132	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7440-47-3	Chromium	235		1	0.062	0.66	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7440-48-4	Cobalt	16.9		1	0.13	1.97	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7440-50-8	Copper	103		1	0.29	1.32	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7439-89-6	Iron	22100		1	5.25	6.58	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7439-92-1	Lead	226	N*	1	0.17	0.79	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7439-95-4	Magnesium	2710		1	15.8	132	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7439-96-5	Manganese	546		1	0.18	1.32	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7439-97-6	Mercury	0.75		1	0.010	0.019	mg/Kg	12/02/25 13:35	12/03/25 09:38	7471B	M7471B
7440-02-0	Nickel	30.1		1	0.17	2.63	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7440-09-7	Potassium	1290	N	1	36.5	132	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7782-49-2	Selenium	0.34	U	1	0.34	1.32	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7440-22-4	Silver	1.69		1	0.16	0.66	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7440-23-5	Sodium	132	N	1	23.4	132	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7440-28-0	Thallium	0.44	J	1	0.30	2.63	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7440-62-2	Vanadium	23.2	N	1	0.33	2.63	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050
7440-66-6	Zinc	434		1	0.30	2.63	mg/Kg	12/02/25 10:30	12/02/25 16:22	6010D	SW3050

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV135	Mercury	4.04	4.0	101	90 - 110	CV	12/03/2025	08:46	LB138086

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV53	Mercury	4.88	5.0	98	90 - 110	CV	12/03/2025	08:50	LB138086
CCV54	Mercury	5.08	5.0	102	90 - 110	CV	12/03/2025	09:18	LB138086
CCV55	Mercury	5.15	5.0	103	90 - 110	CV	12/03/2025	09:45	LB138086
CCV56	Mercury	4.99	5.0	100	90 - 110	CV	12/03/2025	11:09	LB138086

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7850	8000	98	90 - 110	P	12/02/2025	11:29	LB138085
	Antimony	3960	4000	99	90 - 110	P	12/02/2025	11:29	LB138085
	Arsenic	3880	4000	97	90 - 110	P	12/02/2025	11:29	LB138085
	Barium	8280	8000	104	90 - 110	P	12/02/2025	11:29	LB138085
	Beryllium	205	200	102	90 - 110	P	12/02/2025	11:29	LB138085
	Cadmium	1920	2000	96	90 - 110	P	12/02/2025	11:29	LB138085
	Calcium	19900	20000	100	90 - 110	P	12/02/2025	11:29	LB138085
	Chromium	781	800	98	90 - 110	P	12/02/2025	11:29	LB138085
	Cobalt	1930	2000	97	90 - 110	P	12/02/2025	11:29	LB138085
	Copper	999	1000	100	90 - 110	P	12/02/2025	11:29	LB138085
	Iron	3940	4000	99	90 - 110	P	12/02/2025	11:29	LB138085
	Lead	3790	4000	95	90 - 110	P	12/02/2025	11:29	LB138085
	Magnesium	19600	20000	98	90 - 110	P	12/02/2025	11:29	LB138085
	Manganese	2050	2000	102	90 - 110	P	12/02/2025	11:29	LB138085
	Nickel	1940	2000	97	90 - 110	P	12/02/2025	11:29	LB138085
	Potassium	19600	20000	98	90 - 110	P	12/02/2025	11:29	LB138085
	Selenium	3970	4000	99	90 - 110	P	12/02/2025	11:29	LB138085
	Silver	1080	1000	108	90 - 110	P	12/02/2025	11:29	LB138085
	Sodium	20300	20000	101	90 - 110	P	12/02/2025	11:29	LB138085
	Thallium	4060	4000	102	90 - 110	P	12/02/2025	11:29	LB138085
	Vanadium	1970	2000	99	90 - 110	P	12/02/2025	11:29	LB138085
	Zinc	2020	2000	101	90 - 110	P	12/02/2025	11:29	LB138085

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	112	100	112	80 - 120	P	12/02/2025	11:34	LB138085
	Antimony	50.1	50.0	100	80 - 120	P	12/02/2025	11:34	LB138085
	Arsenic	22.6	20.0	113	80 - 120	P	12/02/2025	11:34	LB138085
	Barium	104	100	104	80 - 120	P	12/02/2025	11:34	LB138085
	Beryllium	6.61	6.0	110	80 - 120	P	12/02/2025	11:34	LB138085
	Cadmium	6.22	6.0	104	80 - 120	P	12/02/2025	11:34	LB138085
	Calcium	2050	2000	102	80 - 120	P	12/02/2025	11:34	LB138085
	Chromium	9.91	10.0	99	80 - 120	P	12/02/2025	11:34	LB138085
	Cobalt	29.9	30.0	100	80 - 120	P	12/02/2025	11:34	LB138085
	Copper	21.9	20.0	110	80 - 120	P	12/02/2025	11:34	LB138085
	Iron	103	100	103	80 - 120	P	12/02/2025	11:34	LB138085
	Lead	11.3	12.0	94	80 - 120	P	12/02/2025	11:34	LB138085
	Magnesium	2170	2000	109	80 - 120	P	12/02/2025	11:34	LB138085
	Manganese	21.9	20.0	110	80 - 120	P	12/02/2025	11:34	LB138085
	Nickel	40.7	40.0	102	80 - 120	P	12/02/2025	11:34	LB138085
	Potassium	1900	2000	95	80 - 120	P	12/02/2025	11:34	LB138085
	Selenium	18.8	20.0	94	80 - 120	P	12/02/2025	11:34	LB138085
	Silver	10.3	10.0	103	80 - 120	P	12/02/2025	11:34	LB138085
	Sodium	2010	2000	100	80 - 120	P	12/02/2025	11:34	LB138085
	Thallium	44.2	40.0	111	80 - 120	P	12/02/2025	11:34	LB138085
	Vanadium	40.2	40.0	100	80 - 120	P	12/02/2025	11:34	LB138085
	Zinc	40.6	40.0	102	80 - 120	P	12/02/2025	11:34	LB138085

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	10100	10000	101	90 - 110	P	12/02/2025	12:10	LB138085
	Antimony	4970	5000	99	90 - 110	P	12/02/2025	12:10	LB138085
	Arsenic	4950	5000	99	90 - 110	P	12/02/2025	12:10	LB138085
	Barium	10200	10000	102	90 - 110	P	12/02/2025	12:10	LB138085
	Beryllium	254	250	102	90 - 110	P	12/02/2025	12:10	LB138085
	Cadmium	2470	2500	99	90 - 110	P	12/02/2025	12:10	LB138085
	Calcium	25300	25000	101	90 - 110	P	12/02/2025	12:10	LB138085
	Chromium	1000	1000	100	90 - 110	P	12/02/2025	12:10	LB138085
	Cobalt	2470	2500	99	90 - 110	P	12/02/2025	12:10	LB138085
	Copper	1250	1250	100	90 - 110	P	12/02/2025	12:10	LB138085
	Iron	4950	5000	99	90 - 110	P	12/02/2025	12:10	LB138085
	Lead	4950	5000	99	90 - 110	P	12/02/2025	12:10	LB138085
	Magnesium	25100	25000	101	90 - 110	P	12/02/2025	12:10	LB138085
	Manganese	2540	2500	101	90 - 110	P	12/02/2025	12:10	LB138085
	Nickel	2480	2500	99	90 - 110	P	12/02/2025	12:10	LB138085
	Potassium	24800	25000	99	90 - 110	P	12/02/2025	12:10	LB138085
	Selenium	4980	5000	100	90 - 110	P	12/02/2025	12:10	LB138085
	Silver	1250	1250	100	90 - 110	P	12/02/2025	12:10	LB138085
	Sodium	25400	25000	101	90 - 110	P	12/02/2025	12:10	LB138085
	Thallium	4980	5000	100	90 - 110	P	12/02/2025	12:10	LB138085
	Vanadium	2520	2500	101	90 - 110	P	12/02/2025	12:10	LB138085
	Zinc	2520	2500	101	90 - 110	P	12/02/2025	12:10	LB138085
CCV02	Aluminum	10100	10000	100	90 - 110	P	12/02/2025	13:13	LB138085
	Antimony	5120	5000	102	90 - 110	P	12/02/2025	13:13	LB138085
	Arsenic	5050	5000	101	90 - 110	P	12/02/2025	13:13	LB138085
	Barium	9890	10000	99	90 - 110	P	12/02/2025	13:13	LB138085
	Beryllium	249	250	100	90 - 110	P	12/02/2025	13:13	LB138085
	Cadmium	2500	2500	100	90 - 110	P	12/02/2025	13:13	LB138085
	Calcium	25000	25000	100	90 - 110	P	12/02/2025	13:13	LB138085
	Chromium	1020	1000	102	90 - 110	P	12/02/2025	13:13	LB138085
	Cobalt	2500	2500	100	90 - 110	P	12/02/2025	13:13	LB138085
	Copper	1280	1250	103	90 - 110	P	12/02/2025	13:13	LB138085
	Iron	5090	5000	102	90 - 110	P	12/02/2025	13:13	LB138085
	Lead	5020	5000	100	90 - 110	P	12/02/2025	13:13	LB138085

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24800	25000	99	90 - 110	P	12/02/2025	13:13	LB138085
	Manganese	2480	2500	99	90 - 110	P	12/02/2025	13:13	LB138085
	Nickel	2520	2500	101	90 - 110	P	12/02/2025	13:13	LB138085
	Potassium	25900	25000	104	90 - 110	P	12/02/2025	13:13	LB138085
	Selenium	5180	5000	104	90 - 110	P	12/02/2025	13:13	LB138085
	Silver	1280	1250	103	90 - 110	P	12/02/2025	13:13	LB138085
	Sodium	26300	25000	105	90 - 110	P	12/02/2025	13:13	LB138085
	Thallium	4690	5000	94	90 - 110	P	12/02/2025	13:13	LB138085
	Vanadium	2490	2500	100	90 - 110	P	12/02/2025	13:13	LB138085
	Zinc	2590	2500	104	90 - 110	P	12/02/2025	13:13	LB138085
CCV03	Aluminum	10100	10000	101	90 - 110	P	12/02/2025	14:11	LB138085
	Antimony	5100	5000	102	90 - 110	P	12/02/2025	14:11	LB138085
	Arsenic	5040	5000	101	90 - 110	P	12/02/2025	14:11	LB138085
	Barium	10100	10000	102	90 - 110	P	12/02/2025	14:11	LB138085
	Beryllium	247	250	99	90 - 110	P	12/02/2025	14:11	LB138085
	Cadmium	2470	2500	99	90 - 110	P	12/02/2025	14:11	LB138085
	Calcium	25000	25000	100	90 - 110	P	12/02/2025	14:11	LB138085
	Chromium	987	1000	99	90 - 110	P	12/02/2025	14:11	LB138085
	Cobalt	2470	2500	99	90 - 110	P	12/02/2025	14:11	LB138085
	Copper	1280	1250	102	90 - 110	P	12/02/2025	14:11	LB138085
	Iron	5060	5000	101	90 - 110	P	12/02/2025	14:11	LB138085
	Lead	4970	5000	100	90 - 110	P	12/02/2025	14:11	LB138085
	Magnesium	24800	25000	99	90 - 110	P	12/02/2025	14:11	LB138085
	Manganese	2500	2500	100	90 - 110	P	12/02/2025	14:11	LB138085
	Nickel	2500	2500	100	90 - 110	P	12/02/2025	14:11	LB138085
	Potassium	25200	25000	101	90 - 110	P	12/02/2025	14:11	LB138085
	Selenium	5230	5000	105	90 - 110	P	12/02/2025	14:11	LB138085
	Silver	1260	1250	100	90 - 110	P	12/02/2025	14:11	LB138085
	Sodium	25900	25000	104	90 - 110	P	12/02/2025	14:11	LB138085
	Thallium	4900	5000	98	90 - 110	P	12/02/2025	14:11	LB138085
	Vanadium	2500	2500	100	90 - 110	P	12/02/2025	14:11	LB138085
	Zinc	2530	2500	101	90 - 110	P	12/02/2025	14:11	LB138085
CCV04	Aluminum	10200	10000	102	90 - 110	P	12/02/2025	15:15	LB138085
	Antimony	5200	5000	104	90 - 110	P	12/02/2025	15:15	LB138085

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5140	5000	103	90 - 110	P	12/02/2025	15:15	LB138085
	Barium	10500	10000	105	90 - 110	P	12/02/2025	15:15	LB138085
	Beryllium	250	250	100	90 - 110	P	12/02/2025	15:15	LB138085
	Cadmium	2500	2500	100	90 - 110	P	12/02/2025	15:15	LB138085
	Calcium	25500	25000	102	90 - 110	P	12/02/2025	15:15	LB138085
	Chromium	990	1000	99	90 - 110	P	12/02/2025	15:15	LB138085
	Cobalt	2520	2500	101	90 - 110	P	12/02/2025	15:15	LB138085
	Copper	1300	1250	104	90 - 110	P	12/02/2025	15:15	LB138085
	Iron	5230	5000	105	90 - 110	P	12/02/2025	15:15	LB138085
	Lead	5050	5000	101	90 - 110	P	12/02/2025	15:15	LB138085
	Magnesium	24900	25000	100	90 - 110	P	12/02/2025	15:15	LB138085
	Manganese	2590	2500	104	90 - 110	P	12/02/2025	15:15	LB138085
	Nickel	2550	2500	102	90 - 110	P	12/02/2025	15:15	LB138085
	Potassium	26000	25000	104	90 - 110	P	12/02/2025	15:15	LB138085
	Selenium	5330	5000	107	90 - 110	P	12/02/2025	15:15	LB138085
	Silver	1280	1250	102	90 - 110	P	12/02/2025	15:15	LB138085
	Sodium	26900	25000	108	90 - 110	P	12/02/2025	15:15	LB138085
	Thallium	5220	5000	104	90 - 110	P	12/02/2025	15:15	LB138085
	Vanadium	2520	2500	101	90 - 110	P	12/02/2025	15:15	LB138085
	Zinc	2570	2500	103	90 - 110	P	12/02/2025	15:15	LB138085
CCV05	Aluminum	10200	10000	102	90 - 110	P	12/02/2025	16:05	LB138085
	Antimony	5110	5000	102	90 - 110	P	12/02/2025	16:05	LB138085
	Arsenic	5040	5000	101	90 - 110	P	12/02/2025	16:05	LB138085
	Barium	10600	10000	106	90 - 110	P	12/02/2025	16:05	LB138085
	Beryllium	249	250	100	90 - 110	P	12/02/2025	16:05	LB138085
	Cadmium	2460	2500	98	90 - 110	P	12/02/2025	16:05	LB138085
	Calcium	25600	25000	102	90 - 110	P	12/02/2025	16:05	LB138085
	Chromium	968	1000	97	90 - 110	P	12/02/2025	16:05	LB138085
	Cobalt	2480	2500	99	90 - 110	P	12/02/2025	16:05	LB138085
	Copper	1280	1250	102	90 - 110	P	12/02/2025	16:05	LB138085
	Iron	5170	5000	103	90 - 110	P	12/02/2025	16:05	LB138085
	Lead	4950	5000	99	90 - 110	P	12/02/2025	16:05	LB138085
	Magnesium	24900	25000	100	90 - 110	P	12/02/2025	16:05	LB138085
	Manganese	2640	2500	106	90 - 110	P	12/02/2025	16:05	LB138085

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2520	2500	101	90 - 110	P	12/02/2025	16:05	LB138085
	Potassium	25600	25000	102	90 - 110	P	12/02/2025	16:05	LB138085
	Selenium	5220	5000	104	90 - 110	P	12/02/2025	16:05	LB138085
	Silver	1250	1250	100	90 - 110	P	12/02/2025	16:05	LB138085
	Sodium	27100	25000	108	90 - 110	P	12/02/2025	16:05	LB138085
	Thallium	5160	5000	103	90 - 110	P	12/02/2025	16:05	LB138085
	Vanadium	2530	2500	101	90 - 110	P	12/02/2025	16:05	LB138085
	Zinc	2510	2500	100	90 - 110	P	12/02/2025	16:05	LB138085
CCV06	Aluminum	10100	10000	101	90 - 110	P	12/02/2025	16:55	LB138085
	Antimony	5130	5000	102	90 - 110	P	12/02/2025	16:55	LB138085
	Arsenic	5050	5000	101	90 - 110	P	12/02/2025	16:55	LB138085
	Barium	10400	10000	104	90 - 110	P	12/02/2025	16:55	LB138085
	Beryllium	251	250	100	90 - 110	P	12/02/2025	16:55	LB138085
	Cadmium	2470	2500	99	90 - 110	P	12/02/2025	16:55	LB138085
	Calcium	25500	25000	102	90 - 110	P	12/02/2025	16:55	LB138085
	Chromium	983	1000	98	90 - 110	P	12/02/2025	16:55	LB138085
	Cobalt	2500	2500	100	90 - 110	P	12/02/2025	16:55	LB138085
	Copper	1290	1250	103	90 - 110	P	12/02/2025	16:55	LB138085
	Iron	5170	5000	103	90 - 110	P	12/02/2025	16:55	LB138085
	Lead	5000	5000	100	90 - 110	P	12/02/2025	16:55	LB138085
	Magnesium	24800	25000	99	90 - 110	P	12/02/2025	16:55	LB138085
	Manganese	2610	2500	104	90 - 110	P	12/02/2025	16:55	LB138085
	Nickel	2540	2500	102	90 - 110	P	12/02/2025	16:55	LB138085
	Potassium	25800	25000	103	90 - 110	P	12/02/2025	16:55	LB138085
	Selenium	5220	5000	104	90 - 110	P	12/02/2025	16:55	LB138085
	Silver	1260	1250	101	90 - 110	P	12/02/2025	16:55	LB138085
	Sodium	26700	25000	107	90 - 110	P	12/02/2025	16:55	LB138085
	Thallium	4600	5000	92	90 - 110	P	12/02/2025	16:55	LB138085
	Vanadium	2510	2500	101	90 - 110	P	12/02/2025	16:55	LB138085
	Zinc	2550	2500	102	90 - 110	P	12/02/2025	16:55	LB138085

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Aluminum	108	100	108	65 - 135	P	12/02/2025	11:42	LB138085
	Antimony	52.1	50.0	104	65 - 135	P	12/02/2025	11:42	LB138085
	Arsenic	20.2	20.0	101	65 - 135	P	12/02/2025	11:42	LB138085
	Barium	102	100	102	65 - 135	P	12/02/2025	11:42	LB138085
	Beryllium	6.67	6.0	111	65 - 135	P	12/02/2025	11:42	LB138085
	Cadmium	6.34	6.0	106	65 - 135	P	12/02/2025	11:42	LB138085
	Calcium	2050	2000	102	65 - 135	P	12/02/2025	11:42	LB138085
	Chromium	10.6	10.0	106	65 - 135	P	12/02/2025	11:42	LB138085
	Cobalt	30.0	30.0	100	65 - 135	P	12/02/2025	11:42	LB138085
	Copper	21.5	20.0	108	65 - 135	P	12/02/2025	11:42	LB138085
	Iron	110	100	110	65 - 135	P	12/02/2025	11:42	LB138085
	Lead	10.2	12.0	85	65 - 135	P	12/02/2025	11:42	LB138085
	Magnesium	2160	2000	108	65 - 135	P	12/02/2025	11:42	LB138085
	Manganese	21.6	20.0	108	65 - 135	P	12/02/2025	11:42	LB138085
	Nickel	41.1	40.0	103	65 - 135	P	12/02/2025	11:42	LB138085
	Potassium	1980	2000	99	65 - 135	P	12/02/2025	11:42	LB138085
	Selenium	19.3	20.0	96	65 - 135	P	12/02/2025	11:42	LB138085
	Silver	10.1	10.0	101	65 - 135	P	12/02/2025	11:42	LB138085
	Sodium	1990	2000	100	65 - 135	P	12/02/2025	11:42	LB138085
	Thallium	49.9	40.0	125	65 - 135	P	12/02/2025	11:42	LB138085
	Vanadium	42.0	40.0	105	65 - 135	P	12/02/2025	11:42	LB138085
	Zinc	41.5	40.0	104	65 - 135	P	12/02/2025	11:42	LB138085
CRA	Mercury	0.18	0.2	92	70 - 130	CV	12/03/2025	08:55	LB138086



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB135	Mercury	0.076	+/-0.2	U	0.20	CV	12/03/2025	08:48	LB138086

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB53	Mercury	0.076	+/-0.2	U	0.20	CV	12/03/2025	08:53	LB138086
CCB54	Mercury	0.076	+/-0.2	U	0.20	CV	12/03/2025	09:20	LB138086
CCB55	Mercury	0.076	+/-0.2	U	0.20	CV	12/03/2025	09:48	LB138086
CCB56	Mercury	0.076	+/-0.2	U	0.20	CV	12/03/2025	11:11	LB138086

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	12/02/2025	11:38	LB138085
	Antimony	6.76	+/-25	U	50.0	P	12/02/2025	11:38	LB138085
	Arsenic	5.12	+/-10	U	20.0	P	12/02/2025	11:38	LB138085
	Barium	14.6	+/-50	U	100	P	12/02/2025	11:38	LB138085
	Beryllium	0.56	+/-3	U	6.00	P	12/02/2025	11:38	LB138085
	Cadmium	0.50	+/-3	U	6.00	P	12/02/2025	11:38	LB138085
	Calcium	234	+/-1000	U	2000	P	12/02/2025	11:38	LB138085
	Chromium	2.12	+/-5	U	10.0	P	12/02/2025	11:38	LB138085
	Cobalt	2.26	+/-15	U	30.0	P	12/02/2025	11:38	LB138085
	Copper	4.60	+/-10	U	20.0	P	12/02/2025	11:38	LB138085
	Iron	23.4	+/-50	U	100	P	12/02/2025	11:38	LB138085
	Lead	2.30	+/-6	U	12.0	P	12/02/2025	11:38	LB138085
	Magnesium	244	+/-1000	U	2000	P	12/02/2025	11:38	LB138085
	Manganese	5.94	+/-10	U	20.0	P	12/02/2025	11:38	LB138085
	Nickel	3.06	+/-20	U	40.0	P	12/02/2025	11:38	LB138085
	Potassium	918	+/-1000	U	2000	P	12/02/2025	11:38	LB138085
	Selenium	9.64	+/-10	U	20.0	P	12/02/2025	11:38	LB138085
	Silver	1.62	+/-5	U	10.0	P	12/02/2025	11:38	LB138085
	Sodium	868	+/-1000	U	2000	P	12/02/2025	11:38	LB138085
	Thallium	5.65	+/-20	J	40.0	P	12/02/2025	11:38	LB138085
	Vanadium	6.26	+/-20	U	40.0	P	12/02/2025	11:38	LB138085
	Zinc	16.7	+/-20	U	40.0	P	12/02/2025	11:38	LB138085

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	12/02/2025	12:14	LB138085
	Antimony	6.76	+/-25	U	50.0	P	12/02/2025	12:14	LB138085
	Arsenic	5.12	+/-10	U	20.0	P	12/02/2025	12:14	LB138085
	Barium	14.6	+/-50	U	100	P	12/02/2025	12:14	LB138085
	Beryllium	0.56	+/-3	U	6.00	P	12/02/2025	12:14	LB138085
	Cadmium	0.50	+/-3	U	6.00	P	12/02/2025	12:14	LB138085
	Calcium	234	+/-1000	U	2000	P	12/02/2025	12:14	LB138085
	Chromium	2.12	+/-5	U	10.0	P	12/02/2025	12:14	LB138085
	Cobalt	2.26	+/-15	U	30.0	P	12/02/2025	12:14	LB138085
	Copper	4.60	+/-10	U	20.0	P	12/02/2025	12:14	LB138085
	Iron	23.4	+/-50	U	100	P	12/02/2025	12:14	LB138085
	Lead	2.30	+/-6	U	12.0	P	12/02/2025	12:14	LB138085
	Magnesium	244	+/-1000	U	2000	P	12/02/2025	12:14	LB138085
	Manganese	5.94	+/-10	U	20.0	P	12/02/2025	12:14	LB138085
	Nickel	3.06	+/-20	U	40.0	P	12/02/2025	12:14	LB138085
	Potassium	918	+/-1000	U	2000	P	12/02/2025	12:14	LB138085
	Selenium	9.64	+/-10	U	20.0	P	12/02/2025	12:14	LB138085
	Silver	1.62	+/-5	U	10.0	P	12/02/2025	12:14	LB138085
	Sodium	868	+/-1000	U	2000	P	12/02/2025	12:14	LB138085
	Thallium	4.38	+/-20	U	40.0	P	12/02/2025	12:14	LB138085
	Vanadium	6.26	+/-20	U	40.0	P	12/02/2025	12:14	LB138085
	Zinc	16.7	+/-20	U	40.0	P	12/02/2025	12:14	LB138085
CCB02	Aluminum	11.3	+/-50	U	100	P	12/02/2025	13:17	LB138085
	Antimony	6.76	+/-25	U	50.0	P	12/02/2025	13:17	LB138085
	Arsenic	5.12	+/-10	U	20.0	P	12/02/2025	13:17	LB138085
	Barium	14.6	+/-50	U	100	P	12/02/2025	13:17	LB138085
	Beryllium	0.56	+/-3	U	6.00	P	12/02/2025	13:17	LB138085
	Cadmium	0.50	+/-3	U	6.00	P	12/02/2025	13:17	LB138085
	Calcium	234	+/-1000	U	2000	P	12/02/2025	13:17	LB138085
	Chromium	2.12	+/-5	U	10.0	P	12/02/2025	13:17	LB138085
	Cobalt	2.26	+/-15	U	30.0	P	12/02/2025	13:17	LB138085
	Copper	4.60	+/-10	U	20.0	P	12/02/2025	13:17	LB138085
	Iron	23.4	+/-50	U	100	P	12/02/2025	13:17	LB138085
	Lead	2.30	+/-6	U	12.0	P	12/02/2025	13:17	LB138085
	Magnesium	244	+/-1000	U	2000	P	12/02/2025	13:17	LB138085
	Manganese	5.94	+/-10	U	20.0	P	12/02/2025	13:17	LB138085
	Nickel	3.06	+/-20	U	40.0	P	12/02/2025	13:17	LB138085
	Potassium	918	+/-1000	U	2000	P	12/02/2025	13:17	LB138085
	Selenium	9.64	+/-10	U	20.0	P	12/02/2025	13:17	LB138085

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	12/02/2025	13:17	LB138085
	Sodium	868	+/-1000	U	2000	P	12/02/2025	13:17	LB138085
	Thallium	7.81	+/-20	J	40.0	P	12/02/2025	13:17	LB138085
	Vanadium	6.26	+/-20	U	40.0	P	12/02/2025	13:17	LB138085
	Zinc	16.7	+/-20	U	40.0	P	12/02/2025	13:17	LB138085
CCB03	Aluminum	13.3	+/-50	J	100	P	12/02/2025	14:15	LB138085
	Antimony	6.76	+/-25	U	50.0	P	12/02/2025	14:15	LB138085
	Arsenic	5.12	+/-10	U	20.0	P	12/02/2025	14:15	LB138085
	Barium	14.6	+/-50	U	100	P	12/02/2025	14:15	LB138085
	Beryllium	0.56	+/-3	U	6.00	P	12/02/2025	14:15	LB138085
	Cadmium	0.50	+/-3	U	6.00	P	12/02/2025	14:15	LB138085
	Calcium	234	+/-1000	U	2000	P	12/02/2025	14:15	LB138085
	Chromium	2.12	+/-5	U	10.0	P	12/02/2025	14:15	LB138085
	Cobalt	2.26	+/-15	U	30.0	P	12/02/2025	14:15	LB138085
	Copper	4.60	+/-10	U	20.0	P	12/02/2025	14:15	LB138085
	Iron	25.1	+/-50	J	100	P	12/02/2025	14:15	LB138085
	Lead	2.30	+/-6	U	12.0	P	12/02/2025	14:15	LB138085
	Magnesium	244	+/-1000	U	2000	P	12/02/2025	14:15	LB138085
	Manganese	5.94	+/-10	U	20.0	P	12/02/2025	14:15	LB138085
	Nickel	3.06	+/-20	U	40.0	P	12/02/2025	14:15	LB138085
	Potassium	918	+/-1000	U	2000	P	12/02/2025	14:15	LB138085
	Selenium	9.64	+/-10	U	20.0	P	12/02/2025	14:15	LB138085
	Silver	1.62	+/-5	U	10.0	P	12/02/2025	14:15	LB138085
	Sodium	868	+/-1000	U	2000	P	12/02/2025	14:15	LB138085
	Thallium	4.38	+/-20	U	40.0	P	12/02/2025	14:15	LB138085
	Vanadium	6.26	+/-20	U	40.0	P	12/02/2025	14:15	LB138085
	Zinc	16.7	+/-20	U	40.0	P	12/02/2025	14:15	LB138085
CCB04	Aluminum	22.4	+/-50	J	100	P	12/02/2025	15:20	LB138085
	Antimony	6.76	+/-25	U	50.0	P	12/02/2025	15:20	LB138085
	Arsenic	5.12	+/-10	U	20.0	P	12/02/2025	15:20	LB138085
	Barium	14.6	+/-50	U	100	P	12/02/2025	15:20	LB138085
	Beryllium	0.56	+/-3	U	6.00	P	12/02/2025	15:20	LB138085
	Cadmium	0.50	+/-3	U	6.00	P	12/02/2025	15:20	LB138085
	Calcium	234	+/-1000	U	2000	P	12/02/2025	15:20	LB138085
	Chromium	2.12	+/-5	U	10.0	P	12/02/2025	15:20	LB138085
	Cobalt	2.26	+/-15	U	30.0	P	12/02/2025	15:20	LB138085
	Copper	4.60	+/-10	U	20.0	P	12/02/2025	15:20	LB138085
	Iron	23.4	+/-50	U	100	P	12/02/2025	15:20	LB138085
	Lead	2.30	+/-6	U	12.0	P	12/02/2025	15:20	LB138085

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	12/02/2025	15:20	LB138085
	Manganese	5.94	+/-10	U	20.0	P	12/02/2025	15:20	LB138085
	Nickel	3.06	+/-20	U	40.0	P	12/02/2025	15:20	LB138085
	Potassium	918	+/-1000	U	2000	P	12/02/2025	15:20	LB138085
	Selenium	9.64	+/-10	U	20.0	P	12/02/2025	15:20	LB138085
	Silver	1.62	+/-5	U	10.0	P	12/02/2025	15:20	LB138085
	Sodium	868	+/-1000	U	2000	P	12/02/2025	15:20	LB138085
	Thallium	5.38	+/-20	J	40.0	P	12/02/2025	15:20	LB138085
	Vanadium	6.26	+/-20	U	40.0	P	12/02/2025	15:20	LB138085
	Zinc	16.7	+/-20	U	40.0	P	12/02/2025	15:20	LB138085
CCB05	Aluminum	11.3	+/-50	U	100	P	12/02/2025	16:09	LB138085
	Antimony	6.76	+/-25	U	50.0	P	12/02/2025	16:09	LB138085
	Arsenic	5.12	+/-10	U	20.0	P	12/02/2025	16:09	LB138085
	Barium	14.6	+/-50	U	100	P	12/02/2025	16:09	LB138085
	Beryllium	0.56	+/-3	U	6.00	P	12/02/2025	16:09	LB138085
	Cadmium	0.50	+/-3	U	6.00	P	12/02/2025	16:09	LB138085
	Calcium	234	+/-1000	U	2000	P	12/02/2025	16:09	LB138085
	Chromium	2.12	+/-5	U	10.0	P	12/02/2025	16:09	LB138085
	Cobalt	2.26	+/-15	U	30.0	P	12/02/2025	16:09	LB138085
	Copper	4.60	+/-10	U	20.0	P	12/02/2025	16:09	LB138085
	Iron	23.4	+/-50	U	100	P	12/02/2025	16:09	LB138085
	Lead	2.30	+/-6	U	12.0	P	12/02/2025	16:09	LB138085
	Magnesium	244	+/-1000	U	2000	P	12/02/2025	16:09	LB138085
	Manganese	5.94	+/-10	U	20.0	P	12/02/2025	16:09	LB138085
	Nickel	3.06	+/-20	U	40.0	P	12/02/2025	16:09	LB138085
	Potassium	918	+/-1000	U	2000	P	12/02/2025	16:09	LB138085
	Selenium	9.64	+/-10	U	20.0	P	12/02/2025	16:09	LB138085
	Silver	1.62	+/-5	U	10.0	P	12/02/2025	16:09	LB138085
	Sodium	868	+/-1000	U	2000	P	12/02/2025	16:09	LB138085
	Thallium	5.53	+/-20	J	40.0	P	12/02/2025	16:09	LB138085
	Vanadium	6.26	+/-20	U	40.0	P	12/02/2025	16:09	LB138085
	Zinc	16.7	+/-20	U	40.0	P	12/02/2025	16:09	LB138085
CCB06	Aluminum	11.3	+/-50	U	100	P	12/02/2025	16:59	LB138085
	Antimony	6.76	+/-25	U	50.0	P	12/02/2025	16:59	LB138085
	Arsenic	5.12	+/-10	U	20.0	P	12/02/2025	16:59	LB138085
	Barium	14.6	+/-50	U	100	P	12/02/2025	16:59	LB138085
	Beryllium	0.56	+/-3	U	6.00	P	12/02/2025	16:59	LB138085
	Cadmium	0.50	+/-3	U	6.00	P	12/02/2025	16:59	LB138085
	Calcium	234	+/-1000	U	2000	P	12/02/2025	16:59	LB138085

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	12/02/2025	16:59	LB138085
	Cobalt	2.26	+/-15	U	30.0	P	12/02/2025	16:59	LB138085
	Copper	4.60	+/-10	U	20.0	P	12/02/2025	16:59	LB138085
	Iron	23.4	+/-50	U	100	P	12/02/2025	16:59	LB138085
	Lead	2.30	+/-6	U	12.0	P	12/02/2025	16:59	LB138085
	Magnesium	244	+/-1000	U	2000	P	12/02/2025	16:59	LB138085
	Manganese	5.94	+/-10	U	20.0	P	12/02/2025	16:59	LB138085
	Nickel	3.06	+/-20	U	40.0	P	12/02/2025	16:59	LB138085
	Potassium	918	+/-1000	U	2000	P	12/02/2025	16:59	LB138085
	Selenium	9.64	+/-10	U	20.0	P	12/02/2025	16:59	LB138085
	Silver	1.62	+/-5	U	10.0	P	12/02/2025	16:59	LB138085
	Sodium	868	+/-1000	U	2000	P	12/02/2025	16:59	LB138085
	Thallium	4.85	+/-20	J	40.0	P	12/02/2025	16:59	LB138085
	Vanadium	6.26	+/-20	U	40.0	P	12/02/2025	16:59	LB138085
	Zinc	16.7	+/-20	U	40.0	P	12/02/2025	16:59	LB138085

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3742

Instrument: CV1

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB170790BL		SOLID		Batch Number:	PB170790		Prep Date:	12/02/2025	
	Mercury	0.0080	<0.014	U	0.014	CV	12/03/2025	09:02	LB138086

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3742

Instrument: P4

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB170778BL	SOLID			Batch Number:	PB170778		Prep Date:	12/02/2025	
	Aluminum	1.07	<2.5	J	5.00	P	12/02/2025	13:34	LB138085
	Antimony	0.22	<1.25	U	2.50	P	12/02/2025	13:34	LB138085
	Arsenic	0.19	<0.5	U	1.00	P	12/02/2025	13:34	LB138085
	Barium	0.73	<2.5	U	5.00	P	12/02/2025	13:34	LB138085
	Beryllium	0.025	<0.15	U	0.30	P	12/02/2025	13:34	LB138085
	Cadmium	0.024	<0.15	U	0.30	P	12/02/2025	13:34	LB138085
	Calcium	11.1	<50	U	100	P	12/02/2025	13:34	LB138085
	Chromium	0.047	<0.25	U	0.50	P	12/02/2025	13:34	LB138085
	Cobalt	0.10	<0.75	U	1.50	P	12/02/2025	13:34	LB138085
	Copper	0.26	<0.5	J	1.00	P	12/02/2025	13:34	LB138085
	Iron	3.99	<2.5	U	5.00	P	12/02/2025	13:34	LB138085
	Lead	0.13	<0.3	U	0.60	P	12/02/2025	13:34	LB138085
	Magnesium	12.0	<50	U	100	P	12/02/2025	13:34	LB138085
	Manganese	0.14	<0.5	U	1.00	P	12/02/2025	13:34	LB138085
	Nickel	0.13	<1	U	2.00	P	12/02/2025	13:34	LB138085
	Potassium	27.7	<50	U	100	P	12/02/2025	13:34	LB138085
	Selenium	0.26	<0.5	U	1.00	P	12/02/2025	13:34	LB138085
	Silver	0.12	<0.25	U	0.50	P	12/02/2025	13:34	LB138085
	Sodium	17.8	<50	U	100	P	12/02/2025	13:34	LB138085
	Thallium	0.39	<1	J	2.00	P	12/02/2025	13:34	LB138085
	Vanadium	0.25	<1	U	2.00	P	12/02/2025	13:34	LB138085
	Zinc	0.27	<1	J	2.00	P	12/02/2025	13:34	LB138085

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Remington & Vernick</u>	SDG No.: <u>Q3742</u>
Contract: <u>REMI01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	268000	255000	105	216000	294000	12/02/2025	11:52	LB138085
	Antimony	-6.65			-50	50	12/02/2025	11:52	LB138085
	Arsenic	2.54			-20	20	12/02/2025	11:52	LB138085
	Barium	4.18	6.0	70	-94	106	12/02/2025	11:52	LB138085
	Beryllium	1.60			-6	6	12/02/2025	11:52	LB138085
	Cadmium	4.20	1.0	420	-5	7	12/02/2025	11:52	LB138085
	Calcium	256000	245000	104	208000	282000	12/02/2025	11:52	LB138085
	Chromium	46.6	52.0	90	42	62	12/02/2025	11:52	LB138085
	Cobalt	2.36			-30	30	12/02/2025	11:52	LB138085
	Copper	0.50	2.0	25	-18	22	12/02/2025	11:52	LB138085
	Iron	103000	101000	102	85600	116500	12/02/2025	11:52	LB138085
	Lead	5.64			-12	12	12/02/2025	11:52	LB138085
	Magnesium	274000	255000	108	216000	294000	12/02/2025	11:52	LB138085
	Manganese	9.91	7.0	142	-13	27	12/02/2025	11:52	LB138085
	Nickel	6.08	2.0	304	-38	42	12/02/2025	11:52	LB138085
	Potassium	41.4			0	0	12/02/2025	11:52	LB138085
	Selenium	-2.16			-20	20	12/02/2025	11:52	LB138085
	Silver	8.41			-10	10	12/02/2025	11:52	LB138085
	Sodium	133			0	0	12/02/2025	11:52	LB138085
	Thallium	17.4			-40	40	12/02/2025	11:52	LB138085
	Vanadium	5.74			-40	40	12/02/2025	11:52	LB138085
	Zinc	3.36			-40	40	12/02/2025	11:52	LB138085
ICSAB01	Aluminum	251000	247000	102	209000	285000	12/02/2025	11:57	LB138085
	Antimony	600	618	97	525	711	12/02/2025	11:57	LB138085
	Arsenic	108	104	104	88.4	120	12/02/2025	11:57	LB138085
	Barium	500	537	93	437	637	12/02/2025	11:57	LB138085
	Beryllium	505	495	102	420	570	12/02/2025	11:57	LB138085
	Cadmium	1000	972	103	826	1120	12/02/2025	11:57	LB138085
	Calcium	240000	235000	102	199000	271000	12/02/2025	11:57	LB138085
	Chromium	538	542	99	460	624	12/02/2025	11:57	LB138085
	Cobalt	499	476	105	404	548	12/02/2025	11:57	LB138085
	Copper	480	511	94	434	588	12/02/2025	11:57	LB138085
	Iron	92700	99300	93	84400	114500	12/02/2025	11:57	LB138085
	Lead	52.6	49.0	107	37	61	12/02/2025	11:57	LB138085
	Magnesium	257000	248000	104	210000	286000	12/02/2025	11:57	LB138085
	Manganese	499	507	98	430	584	12/02/2025	11:57	LB138085
	Nickel	994	954	104	810	1100	12/02/2025	11:57	LB138085
	Potassium	19.2			0	0	12/02/2025	11:57	LB138085
	Selenium	41.4	46.0	90	26	66	12/02/2025	11:57	LB138085
	Silver	185	201	92	170	232	12/02/2025	11:57	LB138085
	Sodium	101			0	0	12/02/2025	11:57	LB138085
	Thallium	128	108	118	68	148	12/02/2025	11:57	LB138085

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	486	491	99	417	565	12/02/2025	11:57	LB138085
	Zinc	1030	952	108	809	1095	12/02/2025	11:57	LB138085



METAL QC DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Remington & Vernick **level:** low **sdg no.:** Q3742
contract: REMI01 **lab code:** ACE
matrix: Solid **sample id:** Q3742-18 **client id:** SB-1125-24MS

Percent Solids for Sample: 62.8 **Spiked ID:** Q3742-18MS **Percent Solids for Spike Sample:** 62.8

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	mg/Kg	75 - 125	9630		8950		130	520		P
Antimony	mg/Kg	75 - 125	19.7		3.29	U	53.1	37	N	P
Arsenic	mg/Kg	75 - 125	77.9		41.4		53.1	69	N	P
Barium	mg/Kg	75 - 125	116		115		13.3	5		P
Beryllium	mg/Kg	75 - 125	11.4		1.16		13.3	77		P
Cadmium	mg/Kg	75 - 125	14.9		2.46		13.3	94		P
Calcium	mg/Kg	75 - 125	5320		5210		66.3	168		P
Chromium	mg/Kg	75 - 125	207		235		26.5	-106		P
Cobalt	mg/Kg	75 - 125	29.0		16.9		13.3	92		P
Copper	mg/Kg	75 - 125	101		103		19.9	-7		P
Iron	mg/Kg	75 - 125	21800		22100		200	-110		P
Lead	mg/Kg	75 - 125	257		226		66.3	47	N	P
Magnesium	mg/Kg	75 - 125	2830		2710		130	96		P
Manganese	mg/Kg	75 - 125	558		546		13.3	92		P
Nickel	mg/Kg	75 - 125	61.6		30.1		33.2	95		P
Potassium	mg/Kg	75 - 125	1930		1290		660	96		P
Selenium	mg/Kg	75 - 125	107		1.32	U	130	82		P
Silver	mg/Kg	75 - 125	5.49		1.69		5.0	76		P
Sodium	mg/Kg	75 - 125	309		132		200	88		P
Thallium	mg/Kg	75 - 125	114		0.44	J	130	88		P
Vanadium	mg/Kg	75 - 125	39.2		23.2		19.9	81		P
Zinc	mg/Kg	75 - 125	386		434		13.3	-360		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Remington & Vernick</u>	level: <u>low</u>	sdg no.: <u>Q3742</u>
contract: <u>REMI01</u>		lab code: <u>ACE</u>
matrix: <u>Solid</u>	sample id: <u>Q3742-18</u>	client id: <u>SB-1125-24MSD</u>
Percent Solids for Sample: 62.8	Spiked ID: Q3742-18MSD	Percent Solids for Spike Sample: 62.8

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	mg/Kg	75 - 125	8890		8950		140	-44		P
Antimony	mg/Kg	75 - 125	29.5		3.29	U	56.6	52	N	P
Arsenic	mg/Kg	75 - 125	81.1		41.4		56.6	70	N	P
Barium	mg/Kg	75 - 125	104		115		14.2	-78		P
Beryllium	mg/Kg	75 - 125	12.4		1.16		14.2	79		P
Cadmium	mg/Kg	75 - 125	15.6		2.46		14.2	93		P
Calcium	mg/Kg	75 - 125	3510		5210		70.8	-2392		P
Chromium	mg/Kg	75 - 125	209		235		28.3	-93		P
Cobalt	mg/Kg	75 - 125	28.6		16.9		14.2	83		P
Copper	mg/Kg	75 - 125	98.6		103		21.2	-20		P
Iron	mg/Kg	75 - 125	19700		22100		210	-1138		P
Lead	mg/Kg	75 - 125	381		226		70.8	218	N	P
Magnesium	mg/Kg	75 - 125	2600		2710		140	-78		P
Manganese	mg/Kg	75 - 125	499		546		14.2	-328		P
Nickel	mg/Kg	75 - 125	61.3		30.1		35.4	88		P
Potassium	mg/Kg	75 - 125	1690		1290		710	57	N	P
Selenium	mg/Kg	75 - 125	117		1.32	U	140	84		P
Silver	mg/Kg	75 - 125	5.70		1.69		5.3	76		P
Sodium	mg/Kg	75 - 125	289		132		210	74	N	P
Thallium	mg/Kg	75 - 125	119		0.44	J	140	85		P
Vanadium	mg/Kg	75 - 125	37.7		23.2		21.2	68	N	P
Zinc	mg/Kg	75 - 125	363		434		14.2	-501		P

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Remington & Vernick **level:** low **sdg no.:** Q3742
contract: REMI01 **lab code:** ACE
matrix: Solid **sample id:** Q3750-01 **client id:** FENCE WC-1MS
Percent Solids for Sample: 86.8 **Spiked ID:** Q3750-01MS **Percent Solids for Spike Sample:** 86.8

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	mg/Kg	80 - 120	0.52		0.24		0.28	101		CV

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client:	<u>Remington & Vernick</u>	level:	<u>low</u>	sdg no.:	<u>Q3742</u>
contract:	<u>REMI01</u>			lab code:	<u>ACE</u>
matrix:	<u>Solid</u>	sample id:	<u>Q3750-01</u>	client id:	<u>FENCE WC-1MSD</u>
Percent Solids for Sample:	86.8	Spiked ID:	Q3750-01MSD	Percent Solids for Spike Sample:	86.8

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	mg/Kg	80 - 120	0.59		0.24		0.3	117		CV

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Matrix: Solid

Level: LOW

Client ID: SB-1125-24A

Sample ID: Q3742-18

Spiked ID: Q3742-18A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Antimony	mg/Kg	75 - 125	4.02		3.29	U	52.6	8	N	P
Arsenic	mg/Kg	75 - 125	59.0		41.4		52.6	33	N	P
Lead	mg/Kg	75 - 125	238		226		65.8	18	N	P
Potassium	mg/Kg	75 - 125	1270		1290		660	-3	N	P
Sodium	mg/Kg	75 - 125	132		132		200	0	N	P
Vanadium	mg/Kg	75 - 125	22.6		23.2		19.7	-3	N	P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Remington & Vernick</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3742</u>
Contract:	<u>REMI01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Solid</u>	Sample ID:	<u>Q3742-18</u>	Client ID:	<u>SB-1125-24DUP</u>
Percent Solids for Sample:	62.8	Duplicate ID	Q3742-18DUP	Percent Solids for Spike Sample:	62.8

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	mg/Kg	20	8950		8200		9		P
Antimony	mg/Kg	20	3.29	U	3.37	U			P
Arsenic	mg/Kg	20	41.4		37.2		11		P
Barium	mg/Kg	20	115		101		13		P
Beryllium	mg/Kg	20	1.16		1.06		9		P
Cadmium	mg/Kg	20	2.46		2.12		15		P
Calcium	mg/Kg	20	5210		3760		32	*	P
Chromium	mg/Kg	20	235		210		11		P
Cobalt	mg/Kg	20	16.9		15.6		8		P
Copper	mg/Kg	20	103		91.9		11		P
Iron	mg/Kg	20	22100		21000		5		P
Lead	mg/Kg	20	226		196		14		P
Magnesium	mg/Kg	20	2710		2530		7		P
Manganese	mg/Kg	20	546		508		7		P
Nickel	mg/Kg	20	30.1		27.7		8		P
Potassium	mg/Kg	20	1290		1200		7		P
Selenium	mg/Kg	20	1.32	U	1.35	U			P
Silver	mg/Kg	20	1.69		1.65		2		P
Sodium	mg/Kg	20	132		120	J	10		P
Thallium	mg/Kg	20	0.44	J	2.70	U	200.0		P
Vanadium	mg/Kg	20	23.2		21.8		6		P
Zinc	mg/Kg	20	434		406		7		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Remington & Vernick</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3742</u>
Contract:	<u>REMI01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Solid</u>	Sample ID:	<u>Q3742-18MS</u>	Client ID:	<u>SB-1125-24MSD</u>
Percent Solids for Sample:	62.8	Duplicate ID	Q3742-18MSD	Percent Solids for Spike Sample:	62.8

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	mg/Kg	20	9630		8890		8		P
Antimony	mg/Kg	20	19.7		29.5		40	*	P
Arsenic	mg/Kg	20	77.9		81.1		4		P
Barium	mg/Kg	20	116		104		11		P
Beryllium	mg/Kg	20	11.4		12.4		8		P
Cadmium	mg/Kg	20	14.9		15.6		5		P
Calcium	mg/Kg	20	5320		3510		41	*	P
Chromium	mg/Kg	20	207		209		1		P
Cobalt	mg/Kg	20	29.0		28.6		1		P
Copper	mg/Kg	20	101		98.6		2		P
Iron	mg/Kg	20	21800		19700		10		P
Lead	mg/Kg	20	257		381		39	*	P
Magnesium	mg/Kg	20	2830		2600		8		P
Manganese	mg/Kg	20	558		499		11		P
Nickel	mg/Kg	20	61.6		61.3		0		P
Potassium	mg/Kg	20	1930		1690		13		P
Selenium	mg/Kg	20	107		117		9		P
Silver	mg/Kg	20	5.49		5.70		4		P
Sodium	mg/Kg	20	309		289		7		P
Thallium	mg/Kg	20	114		119		4		P
Vanadium	mg/Kg	20	39.2		37.7		4		P
Zinc	mg/Kg	20	386		363		6		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Remington & Vernick</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3742</u>
Contract:	<u>REMI01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Solid</u>	Sample ID:	<u>Q3750-01</u>	Client ID:	<u>FENCE WC-1DUP</u>
Percent Solids for Sample:	86.8	Duplicate ID	Q3750-01DUP	Percent Solids for Spike Sample:	86.8

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	mg/Kg	20	0.24		0.23		3		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Remington & Vernick</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3742</u>
Contract:	<u>REMI01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Solid</u>	Sample ID:	<u>Q3750-01MS</u>	Client ID:	<u>FENCE WC-1MSD</u>
Percent Solids for Sample:	86.8	Duplicate ID	Q3750-01MSD	Percent Solids for Spike Sample:	86.8

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	mg/Kg	20	0.52		0.59		12		CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170778BS							
Aluminum	mg/Kg	100	99.1		99	80 - 120	P
Antimony	mg/Kg	40.0	40.9		102	80 - 120	P
Arsenic	mg/Kg	40.0	39.8		100	80 - 120	P
Barium	mg/Kg	10.0	10.0		100	80 - 120	P
Beryllium	mg/Kg	10.0	10.2		102	80 - 120	P
Cadmium	mg/Kg	10.0	9.77		98	80 - 120	P
Calcium	mg/Kg	50.0	47.7	J	95	80 - 120	P
Chromium	mg/Kg	20.0	19.8		99	80 - 120	P
Cobalt	mg/Kg	10.0	9.73		97	80 - 120	P
Copper	mg/Kg	15.0	15.6		104	80 - 120	P
Iron	mg/Kg	150	149		99	80 - 120	P
Lead	mg/Kg	50.0	47.9		96	80 - 120	P
Magnesium	mg/Kg	100	98.3	J	98	80 - 120	P
Manganese	mg/Kg	10.0	10.3		103	80 - 120	P
Nickel	mg/Kg	25.0	24.8		99	80 - 120	P
Potassium	mg/Kg	500	489		98	80 - 120	P
Selenium	mg/Kg	100	106		106	80 - 120	P
Silver	mg/Kg	3.8	3.64		96	80 - 120	P
Sodium	mg/Kg	150	157		105	80 - 120	P
Thallium	mg/Kg	100	96.2		96	80 - 120	P
Vanadium	mg/Kg	15.0	15.0		100	80 - 120	P
Zinc	mg/Kg	10.0	10.7		107	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170790BS Mercury	mg/Kg	0.28	0.25		90	80 - 120	CV

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

SB-1125-24L

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **Lb No.:** lb138085 **Lab Sample ID :** Q3742-18L **SDG No.:** Q3742
Matrix (soil/water): Solid **Level (low/med):** LOW
Concentration Units: mg/Kg

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Differ- ence	Q	M
	C		C				
Aluminum	8950		11100		24		P
Antimony	3.29	U	16.5	U			P
Arsenic	41.4		46.7		13		P
Barium	115		143		24		P
Beryllium	1.16		1.52	J	31		P
Cadmium	2.46		1.97		20		P
Calcium	5210		6660		28		P
Chromium	235		297		26		P
Cobalt	16.9		17.1		2		P
Copper	103		130		26		P
Iron	22100		27700		25		P
Lead	226		238		5		P
Magnesium	2710		3430		27		P
Manganese	546		701		28		P
Nickel	30.1		31.0		3		P
Potassium	1290		1530		19		P
Selenium	1.32	U	6.58	U			P
Silver	1.69		2.12	J	26		P
Sodium	132		169	J	27		P
Thallium	0.44	J	1.61	J	264		P
Vanadium	23.2		29.5		27		P
Zinc	434		567		31		P

Metals

-9 -

ICP SERIAL DILUTIONS

SAMPLE NO.

FENCE WC-1L

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

Lb No.: lb138086

Lab Sample ID : Q3750-01L

SDG No.: Q3742

Matrix (soil/water): Solid

Level (low/med): LOW

Concentration Units: mg/Kg

Analyte	Initial Sample Result (I) C	Serial Dilution Result (S) C	% Differ- ence	Q	M
Mercury	0.24	0.25	5		CV



METAL PREPARATION & INSTRUMENT DATA

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Al	Ca	Fe	Mg	Ag
Aluminum	396.100	0.0000000	-0.0002060	0.0000000	0.0000000	0.0000000
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	-0.0000440	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000930	0.0000000	0.0000000
Calcium	373.690	0.0000000	0.0000000	-0.0075970	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0007850	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	-0.0000920	0.0000000	0.0000380	0.0000000	0.0000000
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	-0.0001440	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	-0.0001490	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0001050	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		As	Ba	Be	Cd	Co
Aluminum	396.100	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0002870
Calcium	373.690	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0000000	0.0009530
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	-0.0039600
Lead	220.353	0.0000000	0.0003170	0.0000000	0.0000000	0.0000000
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	-0.0003570
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0054900
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Cr	Cu	K	Mn	Mo
Aluminum	396.100	0.0000000	0.0000000	0.0000590	0.0000000	0.0396900
Antimony	206.833	0.0122000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	-0.0029000	0.0000000	0.0000000	0.0000000	0.0004900
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	-0.0000710	-0.0003400
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Calcium	373.690	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000070	0.0002200	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	-0.0007860
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0006510	0.0020500
Iron	240.488	0.0000000	0.0000000	0.0000730	0.0000000	-0.0015250
Lead	220.353	0.0000000	0.0000000	0.0000000	0.0001400	-0.0008600
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0007460	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	-0.0000120
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0017400	-0.0100400
Vanadium	292.402	-0.0025100	0.0000000	0.0000000	0.0000000	-0.0072000
Zinc	213.800	0.0000000	0.0009010	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Na	Ni	Pb	Sb	Se
Aluminum	396.100	0.0000000	0.0000000	0.0012800	0.0000000	0.0000000
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Calcium	373.690	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	-0.0047000	0.0036100	0.0000000	0.0000000
Iron	240.488	0.0000000	-0.0017000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	0.0006580	0.0000000	0.0000000	0.0001290
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0003330	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0067600	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMEN CORRECTION FACTORS

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Instrument ID:

Date:

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Sn	Ti	Tl	V	Zn
Aluminum	396.100	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Antimony	206.833	-0.0035600	-0.0007970	0.0000000	-0.0018900	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000630	0.0001280	0.0000000	0.0000000
Calcium	373.690	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0001110	0.0000000
Cobalt	228.616	0.0000000	0.0018800	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0003840	0.0000000	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	-0.0003610	0.0000000	0.0000000	0.0000000
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Silver	328.068	0.0000000	-0.0007420	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	-0.0039700	0.0000000	-0.0115600	0.0000000
Vanadium	292.402	0.0000000	0.0005320	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000



METAL PREPARATION & ANALYICAL SUMMARY

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Method:

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(g)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB170778							
PB170778BL	PB170778BL	MB	SOLID	12/02/2025	2.00	100.0	100.00
PB170778BS	PB170778BS	LCS	SOLID	12/02/2025	2.00	100.0	100.00
Q3742-01	SB-1125-23	SAM	SOLID	12/02/2025	2.01	100.0	70.10
Q3742-02	SB-1125-19	SAM	SOLID	12/02/2025	2.28	100.0	40.90
Q3742-03	SB-1125-8	SAM	SOLID	12/02/2025	2.23	100.0	85.20
Q3742-04	SB-1125-9	SAM	SOLID	12/02/2025	2.22	100.0	78.60
Q3742-05	SB-1125-21	SAM	SOLID	12/02/2025	2.34	100.0	79.50
Q3742-06	SB-1125-13	SAM	SOLID	12/02/2025	2.20	100.0	85.50
Q3742-07	SB-1125-18	SAM	SOLID	12/02/2025	2.09	100.0	68.30
Q3742-08	SB-1125-20	SAM	SOLID	12/02/2025	2.34	100.0	80.40
Q3742-09	SB-1125-22	SAM	SOLID	12/02/2025	2.30	100.0	72.80
Q3742-10	SB-1125-2	SAM	SOLID	12/02/2025	2.27	100.0	76.40
Q3742-16	SB-1125-1	SAM	SOLID	12/02/2025	2.08	100.0	73.50
Q3742-17	SB-1125-25	SAM	SOLID	12/02/2025	2.24	100.0	63.30
Q3742-18	SB-1125-24	SAM	SOLID	12/02/2025	2.42	100.0	62.80
Q3742-18DUP	SB-1125-24DUP	DUP	SOLID	12/02/2025	2.36	100.0	62.80
Q3742-18MS	SB-1125-24MS	MS	SOLID	12/02/2025	2.40	100.0	62.80
Q3742-18MSD	SB-1125-24MSD	MSD	SOLID	12/02/2025	2.25	100.0	62.80

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client: Remington & Vernick

SDG No.: Q3742

Contract: REMI01

Lab Code: ACE

Method:

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(g)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB170790							
PB170790BL	PB170790BL	MB	SOLID	12/02/2025	0.50	35.0	100.00
PB170790BS	PB170790BS	LCS	SOLID	12/02/2025	0.50	35.0	100.00
Q3742-01	SB-1125-23	SAM	SOLID	12/02/2025	0.51	35.0	70.10
Q3742-02	SB-1125-19	SAM	SOLID	12/02/2025	0.53	35.0	40.90
Q3742-03	SB-1125-8	SAM	SOLID	12/02/2025	0.55	35.0	85.20
Q3742-04	SB-1125-9	SAM	SOLID	12/02/2025	0.53	35.0	78.60
Q3742-05	SB-1125-21	SAM	SOLID	12/02/2025	0.51	35.0	79.50
Q3742-06	SB-1125-13	SAM	SOLID	12/02/2025	0.57	35.0	85.50
Q3742-07	SB-1125-18	SAM	SOLID	12/02/2025	0.55	35.0	68.30
Q3742-08	SB-1125-20	SAM	SOLID	12/02/2025	0.57	35.0	80.40
Q3742-09	SB-1125-22	SAM	SOLID	12/02/2025	0.55	35.0	72.80
Q3742-10	SB-1125-2	SAM	SOLID	12/02/2025	0.54	35.0	76.40
Q3742-16	SB-1125-1	SAM	SOLID	12/02/2025	0.50	35.0	73.50
Q3742-17	SB-1125-25	SAM	SOLID	12/02/2025	0.56	35.0	63.30
Q3742-18	SB-1125-24	SAM	SOLID	12/02/2025	0.60	35.0	62.80
Q3750-01DUP	FENCE WC-1DUP	DUP	SOLID	12/02/2025	0.54	35.0	86.80
Q3750-01MS	FENCE WC-1MS	MS	SOLID	12/02/2025	0.58	35.0	86.80
Q3750-01MSD	FENCE WC-1MSD	MSD	SOLID	12/02/2025	0.53	35.0	86.80

metals
- 14 -
ANALYSIS RUN LOG

Client: Remington & Vernick

Contract: REMI01

Lab code: ACE

Sdg no.: Q3742

Instrument id number: _____

Method: _____

Run number: LB138085

Start date: 12/02/2025

End date: 12/02/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1104	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S1	S1	1	1108	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S2	S2	1	1112	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S3	S3	1	1116	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S4	S4	1	1121	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S5	S5	1	1125	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICV01	ICV01	1	1129	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
LLICV01	LLICV01	1	1134	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICB01	ICB01	1	1138	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CRI01	CRI01	1	1142	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSA01	ICSA01	1	1152	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSAB01	ICSAB01	1	1157	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV01	CCV01	1	1210	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB01	CCB01	1	1214	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV02	CCV02	1	1313	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB02	CCB02	1	1317	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
PB170778BL	PB170778BL	1	1334	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
PB170778BS	PB170778BS	1	1339	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV03	CCV03	1	1411	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB03	CCB03	1	1415	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV04	CCV04	1	1515	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB04	CCB04	1	1520	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3742-01	SB-1125-23	1	1524	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3742-02	SB-1125-19	1	1528	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3742-03	SB-1125-8	1	1533	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3742-04	SB-1125-9	1	1537	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3742-05	SB-1125-21	1	1541	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3742-06	SB-1125-13	1	1545	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3742-07	SB-1125-18	1	1549	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3742-08	SB-1125-20	1	1553	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3742-09	SB-1125-22	1	1557	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3742-10	SB-1125-2	1	1601	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV05	CCV05	1	1605	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB05	CCB05	1	1609	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3742-16	SB-1125-1	1	1614	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3742-17	SB-1125-25	1	1618	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3742-18	SB-1125-24	1	1622	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3742-18DUP	SB-1125-24DUP	1	1626	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3742-18L	SB-1125-24L	5	1630	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3742-18MS	SB-1125-24MS	1	1634	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3742-18MSD	SB-1125-24MSD	1	1638	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn

metals
- 14 -
ANALYSIS RUN LOG

Client: Remington & Vernick

Contract: REMI01

Lab code: ACE

Sdg no.: Q3742

Instrument id number: _____

Method: _____

Run number: LB138085

Start date: 12/02/2025

End date: 12/02/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
Q3742-18A	SB-1125-24A	1	1642	As,K,Na,Pb,Sb,V
CCV06	CCV06	1	1655	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB06	CCB06	1	1659	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn

metals
- 14 -
ANALYSIS RUN LOG

Client: Remington & Vernick

Contract: REMI01

Lab code: ACE

Sdg no.: Q3742

Instrument id number: _____

Method: _____

Run number: LB138086


Start date: 12/03/2025

End date: 12/03/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	0831	HG
S0.2	S0.2	1	0833	HG
S2.5	S2.5	1	0835	HG
S5	S5	1	0838	HG
S7.5	S7.5	1	0840	HG
S10	S10	1	0842	HG
ICV135	ICV135	1	0846	HG
ICB135	ICB135	1	0848	HG
CCV53	CCV53	1	0850	HG
CCB53	CCB53	1	0853	HG
CRA	CRA	1	0855	HG
PB170790BL	PB170790BL	1	0902	HG
PB170790BS	PB170790BS	1	0904	HG
Q3742-01	SB-1125-23	1	0906	HG
Q3742-02	SB-1125-19	1	0909	HG
Q3742-03	SB-1125-8	1	0911	HG
Q3742-04	SB-1125-9	1	0913	HG
Q3742-05	SB-1125-21	1	0916	HG
CCV54	CCV54	1	0918	HG
CCB54	CCB54	1	0920	HG
Q3742-06	SB-1125-13	1	0922	HG
Q3742-07	SB-1125-18	1	0925	HG
Q3742-08	SB-1125-20	1	0927	HG
Q3742-10	SB-1125-2	1	0931	HG
Q3742-16	SB-1125-1	1	0934	HG
Q3742-17	SB-1125-25	1	0936	HG
Q3742-18	SB-1125-24	1	0938	HG
CCV55	CCV55	1	0945	HG
CCB55	CCB55	1	0948	HG
Q3750-01DUP	FENCE WC-1DUP	1	0957	HG
Q3750-01MS	FENCE WC-1MS	1	0959	HG
Q3750-01MSD	FENCE WC-1MSD	1	1002	HG
Q3750-01L	FENCE WC-1L	5	1006	HG
Q3742-09	SB-1125-22	2	1011	HG
CCV56	CCV56	1	1109	HG
CCB56	CCB56	1	1111	HG



METAL RAW DATA





As 189.042 {479}

Date of Fit: 12/2/2025 14:24:34 Type of Fit: Linear Weighting: 1/Conc

A0 (Offset): -0.000086 Re-Slope: 1.000000
A1 (Gain): 0.036179 Y-int: 0.000000
A2 (Curvature): 0.000000
n (Exponent): 1.000000
Correlation: 0.999972 Status: OK.
Std Error of Est: 0.000006
Predicted MDL: 0.004118
Predicted MQL: 0.013727

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	-.00009	.000	1
S1	.02000	.02290	.003	14.5	.00074	.000	1
S3	2.5000	2.4920	-.008	-.320	.08998	.001	1
S4	5.0000	4.9809	-.019	-.383	.17994	.001	1
S5	10.000	10.004	.004	.040	.36150	.002	1
S2	.80000	.82019	.020	2.52	.02956	.000	1







Pb 220.353 {453}

Date of Fit: 12/2/2025 14:24:34 Type of Fit: Linear Weighting: 1/Conc

A0 (Offset): -0.000053 Re-Slope: 1.000000
A1 (Gain): 0.139020 Y-int: 0.000000
A2 (Curvature): 0.000000
n (Exponent): 1.000000
Correlation: 0.999979 Status: OK.
Std Error of Est: 0.000016
Predicted MDL: 0.002080
Predicted MQL: 0.006934

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	-.00005	.000	1
S1	.01200	.01154	-.000	-3.83	.00152	.000	1
S3	2.5000	2.4854	-.015	-.583	.34508	.001	1
S4	5.0000	4.9628	-.037	-.745	.68910	.005	1
S5	10.000	10.038	.038	.380	1.3939	.006	1
S2	.80000	.81430	.014	1.79	.11303	.000	1







Sb 206.833 {463}


Date of Fit: 12/2/2025 14:24:34 Type of Fit: Linear Weighting: 1/Conc


A0 (Offset): 0.000759 Re-Slope: 1.000000
A1 (Gain): 0.080947 Y-int: 0.000000
A2 (Curvature): 0.000000
n (Exponent): 1.000000
Correlation: 0.999974 Status: OK.
Std Error of Est: 0.000022
Predicted MDL: 0.002735
Predicted MQL: 0.009115


Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	.00076	.000	1
S1	.05000	.05036	.000	.714	.00484	.000	1
S3	2.5000	2.4944	-.006	-.225	.20322	.001	1
S4	5.0000	4.9640	-.036	-.720	.40368	.002	1
S5	10.000	10.019	.019	.186	.81393	.004	1
S2	.80000	.82270	.023	2.84	.06753	.000	1






















Cu 224.700 {450}


Date of Fit: 12/2/2025 14:24:34 Type of Fit: Linear Weighting: 1/Conc


A0 (Offset): 0.000561 Re-Slope: 1.000000
A1 (Gain): 0.609590 Y-int: 0.000000
A2 (Curvature): 0.000000
n (Exponent): 1.000000
Correlation: 0.999860 Status: OK.
Std Error of Est: 0.000121
Predicted MDL: 0.000551
Predicted MQL: 0.001836

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	.00056	.000	1
S1	.02000	.02178	.002	8.91	.01398	.000	1
S3	.62500	.63655	.012	1.85	.39186	.001	1
S4	1.2500	1.2459	-.004	-.332	.76655	.005	1
S5	2.5000	2.4786	-.021	-.854	1.5246	.006	1
S2	.20000	.21217	.012	6.09	.13094	.001	1











Ni 231.604 {446}


Date of Fit: 12/2/2025 14:24:34 Type of Fit: Linear Weighting: 1/Conc


A0 (Offset): -0.000714 Re-Slope: 1.000000
A1 (Gain): 0.553457 Y-int: 0.000000
A2 (Curvature): 0.000000
n (Exponent): 1.000000
Correlation: 0.999957 Status: OK.
Std Error of Est: 0.000120
Predicted MDL: 0.000441
Predicted MQL: 0.001470


Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	-.00071	.000	1
S1	.04000	.04115	.001	2.86	.02206	.000	1
S3	1.2500	1.2458	-.004	-.333	.68880	.002	1
S4	2.5000	2.4696	-.030	-1.22	1.3661	.008	1
S5	5.0000	5.0232	.023	.465	2.7794	.016	1
S2	.40000	.41019	.010	2.55	.22631	.001	1













Zn 213.856 {158}


Date of Fit: 12/2/2025 14:24:34 Type of Fit: Linear Weighting: 1/Conc


A0 (Offset): 0.011998 Re-Slope: 1.000000
A1 (Gain): 2.137339 Y-int: 0.000000
A2 (Curvature): 0.000000
n (Exponent): 1.000000
Correlation: 0.999860 Status: OK.
Std Error of Est: 0.000846
Predicted MDL: 0.000359
Predicted MQL: 0.001195


Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	.01199	.001	1
S1	.04000	.04161	.002	4.02	.10153	.000	1
S3	1.2500	1.2700	.020	1.60	2.7450	.020	1
S4	2.5000	2.5077	.008	.307	5.4089	.012	1
S5	5.0000	4.9454	-.055	-1.09	10.656	.054	1
S2	.40000	.42535	.025	6.34	.92707	.006	1
























Date of Fit: 12/2/2025 14:24:34 Type of Fit: Linear Weighting: 1/Conc

A0 (Offset): 0.007241 Re-Slope: 1.000000
A1 (Gain): 0.734416 Y-int: 0.000000
A2 (Curvature): 0.000000
n (Exponent): 1.000000
Correlation: 0.999884 Status: OK.
Std Error of Est: 0.000263
Predicted MDL: 0.001841
Predicted MQL: 0.006138




Sr 407.771 { 83}

Date of Fit: 12/2/2025 14:24:34 Type of Fit: Linear Weighting: 1/Conc

A0 (Offset):	-0.000149	Re-Slope:	1.000000
A1 (Gain):	8.402244	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999906	Status:	OK.
Std Error of Est:	0.002700		
Predicted MDL:	0.000089		
Predicted MQL:	0.000298		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	-.00017	.000	1
S1	.02000	.02135	.001	6.75	.18004	.002	1
S3	2.5000	2.5607	.061	2.43	21.536	.030	1
S4	5.0000	4.9709	-.029	-.581	41.807	.120	1
S5	10.000	9.9349	-.065	-.651	83.556	1.26	1
S2	.80000	.83207	.032	4.01	6.9976	.013	1



Date of Fit: 12/2/2025 14:24:34 Type of Fit: Linear Weighting: 1/Conc


A0 (Offset):	0.000000	Re-Slope:	1.000000
A1 (Gain):	0.000000	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.000000	Status:	Warning Zero Gain
Std Error of Est:	0.000000		
Predicted MDL:	n/a		
Predicted MQL:	n/a		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
-----------	--------------	-------------	------------	---------	-------	---------	----------

A scatter plot with a grid. The x-axis is labeled from -2e35 to 1.3e36 in increments of 1e35. The y-axis is labeled from 0 to 1 in increments of 0.1. A single red data point is plotted at the coordinates (0, 1). The plot title is "Y 360.073 { 94}*".

Date of Fit: <not fit> Type of Fit: Linear Weighting: 1/Conc


A0 (Offset):	0.000000	Re-Slope:	1.000000
A1 (Gain):	0.000000	Y-int:	0.000000



A2 (Curvature):		0.000000		Status:	Warning	Zero Gain	
n (Exponent):		1.000000					
Correlation:		0.000000					
Std Error of Est:		0.000000					
Predicted MDL:		n/a					
Predicted MQL:		n/a					
Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis

LB138086

7471B INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.006	-0.006	21	0.000	21				
0.2	0.200	0.193	-0.007	1042	0.0 %	1042				
2.5	2.500	2.572	0.072	13228	0.0 %	13228				
5.0	5.000	4.857	-0.143	24935	0.0 %	24935				
7.5	7.500	7.639	0.139	39186	0.0 %	39186				
10.0	10.000	9.944	-0.056	50992	0.0 %	50992				

LB138086 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Dec 2025 07:56:05

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Method	Std Cond	Date
S	0.0 - 1	50	-	21	PPB	7471B	0.0000	03 Dec 2025 08:31:15
S	0.2 - 1	51.2	-	1042	PPB	7471B	0.2000	03 Dec 2025 08:33:31
S	2.5 - 1	52.5	-	13228	PPB	7471B	2.5000	03 Dec 2025 08:35:48
S	5.0 - 1	55	-	24935	PPB	7471B	5.0000	03 Dec 2025 08:38:05
S	7.5 - 1	53.5	-	39186	PPB	7471B	7.5000	03 Dec 2025 08:40:23
S	10.0 - 1	510	-	50992	PPB	7471B	10.0000	03 Dec 2025 08:42:40
U	ICV135 - 1	ICV135	4.0389	20742	PPB	7471B		03 Dec 2025 08:46:23
U	ICB135 - 1	ICB135	-0.0291	-97	PPB	7471B		03 Dec 2025 08:48:38
U	CCV53 - 1	CCV53	4.8847	25075	PPB	7471B		03 Dec 2025 08:50:54
U	CCB53 - 1	CCB53	-0.0381	-143	PPB	7471B		03 Dec 2025 08:55:26
U	CRA - 1	CRA	0.1832	991	PPB	7471B		03 Dec 2025 08:57:41
U	HighStd - 1	HighStd	10.0352	51459	PPB	7471B		03 Dec 2025 08:59:57
U	ChkStd - 1	ChkStd	7.3105	37501	PPB	7471B		03 Dec 2025 09:02:15
U	PB170790BL - 1	PBS	-0.0520	-214	PPB	7471B		03 Dec 2025 09:04:34
U	PB170790BS - 1	LCSS	3.6202	18597	PPB	7471B		03 Dec 2025 09:06:50
U	Q3742-01 - 1	SB-1125-23	0.3213	1698	PPB	7471B		03 Dec 2025 09:09:08
U	Q3742-02 - 1	SB-1125-19	8.6945	44591	PPB	7471B		03 Dec 2025 09:11:26
U	Q3742-03 - 1	SB-1125-8	1.0761	5565	PPB	7471B		03 Dec 2025 09:13:45
U	Q3742-04 - 1	SB-1125-9	1.9402	9991	PPB	7471B		03 Dec 2025 09:16:01
U	Q3742-05 - 1	SB-1125-21	1.8779	9672	PPB	7471B		03 Dec 2025 09:18:16
U	CCV54 - 1	CCV54	5.0798	26074	PPB	7471B		03 Dec 2025 09:20:32
U	CCB54 - 1	CCB54	-0.0233	-67	PPB	7471B		03 Dec 2025 09:22:54
U	Q3742-06 - 1	SB-1125-13	0.9405	4870	PPB	7471B		03 Dec 2025 09:25:10
U	Q3742-07 - 1	SB-1125-18	2.5170	12946	PPB	7471B		03 Dec 2025 09:27:25
U	Q3742-08 - 1	SB-1125-20	8.1059	41576	PPB	7471B		03 Dec 2025 09:29:41
U	Q3742-09 - 1	SB-1125-22	11.3035	57956	PPB	7471B		03 Dec 2025 09:31:59
U	Q3742-10 - 1	SB-1125-2	5.1528	26448	PPB	7471B		03 Dec 2025 09:34:18
U	Q3742-16 - 1	SB-1125-1	2.2742	11702	PPB	7471B		03 Dec 2025 09:36:36
U	Q3742-17 - 1	SB-1125-25	9.0570	46448	PPB	7471B		03 Dec 2025 09:38:53
U	Q3742-18 - 1	SB-1125-24	8.0829	41458	PPB	7471B		03 Dec 2025 09:41:12
U	Q3746-01 - 1	ROW	0.1177	655	PPB	7471B		03 Dec 2025 09:43:30
U	Q3748-01 - 1	AU-05-12-1-2025	0.5637	2940	PPB	7471B		03 Dec 2025 09:45:46
U	CCV55 - 1	CCV55	5.1518	26443	PPB	7471B		03 Dec 2025 09:48:01
U	CCB55 - 1	CCB55	-0.0233	-67	PPB	7471B		03 Dec 2025 09:55:27
U	Q3750-01 - 1	FENCE WC-1	3.1649	16265	PPB	7471B		03 Dec 2025 09:57:42
U	Q3750-01DUP - 1	FENCE WC-1DUP	3.1235	16053	PPB	7471B		03 Dec 2025 09:59:59
U	Q3750-01MS - 1	FENCE WC-1MS	7.5228	38589	PPB	7471B		03 Dec 2025 10:02:14
U	Q3750-01MSD - 1	FENCE WC-1MSD	7.7762	39887	PPB	7471B		03 Dec 2025 10:04:32
U	Q3750-04 - 1	FENCE WC-2	13.6351	69900	PPB	7471B		03 Dec 2025 10:06:51
U	Q3750-01LX5 - 1		0.6631	3449	PPB	7471B		03 Dec 2025 10:09:10
U	Q3750-01A - 1		8.4476	43326	PPB	7471B		03 Dec 2025 10:11:27
U	Q3742-09DLX2 - 1		5.8909	30229	PPB	7471B		03 Dec 2025 11:01:13
U	Q3750-04DLX5 - 1		2.5918	13329	PPB	7471B		03 Dec 2025 11:09:09
U	CCV56 - 1	CCV56	4.9919	25624	PPB	7471B		03 Dec 2025 11:11:24
U	CCB56 - 1	CCB56	-0.0510	-209	PPB	7471B		

SOP ID : M3050B-Digestion-20

SDG No : N/A

Matrix : SOIL

Pipette ID: ICP A

Balance ID : M SC-2

Filter paper ID : N/A

pH Strip ID : N/A

Hood ID : #3

Block ID: 1. HOT BLOCK #2 2. N/A

Start Digest Date: 12/02/2025 Time : 10:30 Temp : 96 °C

End Digest Date: 12/02/2025 Time : 11:32 Temp : 96 °C

Digestion tube ID: M6054

Block thermometer ID: MET-DIG. #2

Dig Technician Signature: *SPS*

Supervisor Signature: *[Signature]*

Temp : 1. 96°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
LFS-1	1.00	M6209
LFS-2	1.00	M6201
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Conc. HNO3	5.00	M6187
1:1 HNO3	10.00	MP88081
30% H2O2	3.00	M6170
Conc. HCL	10.00	M6151
PTFE Boiling Stones	N/A	M5581
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

HOT BLOCK#2 CELL#35 96 C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
12/02/25 12:32	<i>SPS met dtd</i>	<i>SA [Signature]</i>
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	pH	Initial Weight (g)	Final Vol (ml)	Color Before	Color After	Texture	Artifact	Comment	Prep Pos
PB170778BL	PBS778	N/A	2.00	100	Colorless	Colorless	Fine	No	N/A	17
PB170778BS	LCS778	N/A	2.00	100	Colorless	Colorless	Fine	No	M6209,M6201	18
Q3742-01	SB-1125-23	N/A	2.01	100	Brown	Yellow	Medium	No	N/A	19
Q3742-02	SB-1125-19	N/A	2.28	100	Black	Yellow	Medium	No	N/A	20
Q3742-03	SB-1125-8	N/A	2.23	100	Black	Yellow	Medium	No	N/A	21
Q3742-04	SB-1125-9	N/A	2.22	100	Black	Yellow	Medium	No	N/A	22
Q3742-05	SB-1125-21	N/A	2.34	100	Black	Yellow	Medium	No	N/A	23
Q3742-06	SB-1125-13	N/A	2.20	100	Black	Yellow	Medium	No	N/A	24
Q3742-07	SB-1125-18	N/A	2.09	100	Black	Yellow	Medium	No	N/A	25
Q3742-08	SB-1125-20	N/A	2.34	100	Black	Yellow	Medium	No	N/A	26
Q3742-09	SB-1125-22	N/A	2.30	100	Black	Yellow	Medium	No	N/A	27
Q3742-10	SB-1125-2	N/A	2.27	100	Black	Yellow	Medium	No	N/A	28
Q3742-16	SB-1125-1	N/A	2.08	100	Black	Yellow	Medium	No	N/A	29
Q3742-17	SB-1125-25	N/A	2.24	100	Black	Yellow	Medium	No	N/A	30
Q3742-18	SB-1125-24	N/A	2.42	100	Black	Yellow	Medium	No	N/A	31
Q3742-18MS	SB-1125-24MS	N/A	2.40	100	Black	Yellow	Medium	No	M6209,M6201	33
Q3742-18MSD	SB-1125-24MSD	N/A	2.25	100	Black	Yellow	Medium	No	M6209,M6201	34
Q3742-18DUP	SB-1125-24DUP	N/A	2.36	100	Black	Yellow	Medium	No	N/A	32

WORKLIST(Hardcopy Internal Chain)

Worklist Name : PB170778

Worklist ID : 193415

Department : Digestion

Date : 12-02-2025 08:35:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3742-01	SB-1125-23	Solid	Metals ICP-TAL	Cool 4 deg C	REMIO1	A21	11/25/6025	6010D
Q3742-02	SB-1125-19	Solid	Metals ICP-TAL	Cool 4 deg C	REMIO1	A21	11/25/2025	6010D
Q3742-03	SB-1125-8	Solid	Metals ICP-TAL	Cool 4 deg C	REMIO1	A21	11/25/2025	6010D
Q3742-04	SB-1125-9	Solid	Metals ICP-TAL	Cool 4 deg C	REMIO1	A21	11/25/2025	6010D
Q3742-05	SB-1125-21	Solid	Metals ICP-TAL	Cool 4 deg C	REMIO1	A21	11/25/2025	6010D
Q3742-06	SB-1125-13	Solid	Metals ICP-TAL	Cool 4 deg C	REMIO1	A21	11/25/2025	6010D
Q3742-07	SB-1125-18	Solid	Metals ICP-TAL	Cool 4 deg C	REMIO1	A21	11/25/2025	6010D
Q3742-08	SB-1125-20	Solid	Metals ICP-TAL	Cool 4 deg C	REMIO1	A21	11/25/2025	6010D
Q3742-09	SB-1125-22	Solid	Metals ICP-TAL	Cool 4 deg C	REMIO1	A21	11/25/2025	6010D
Q3742-10	SB-1125-2	Solid	Metals ICP-TAL	Cool 4 deg C	REMIO1	A21	11/25/2025	6010D
Q3742-16	SB-1125-1	Solid	Metals ICP-TAL	Cool 4 deg C	REMIO1	A21	11/25/2025	6010D
Q3742-17	SB-1125-25	Solid	Metals ICP-TAL	Cool 4 deg C	REMIO1	A21	11/25/2025	6010D
Q3742-18	SB-1125-24	Solid	Metals ICP-TAL	Cool 4 deg C	REMIO1	A21	11/25/2025	6010D

Date/Time 12/02/2025 09:03

Raw Sample Received by: SPS met dig

Raw Sample Relinquished by: SHU SH

Date/Time 12/02/25 10:03

Raw Sample Received by: SPS met dig

Raw Sample Relinquished by: ASH

SOP ID : M7471B-Mercury-19

SDG No : N/A

Matrix : SOIL

Pipette ID: HG A

Balance ID : M SC-3

Filter paper ID : N/A

pH Strip ID : N/A

Hood ID : #1

Block ID: 1. HOT BLOCK #3 2. N/A

Start Digest Date: 12/02/2025 Time : 13:35 Temp : 96 °C

End Digest Date: 12/02/2025 Time : 14:05 Temp : 96 °C

Digestion tube ID: M5595

Block thermometer ID: MET-DIG. #3

Dig Technician Signature: *[Signature]* MB

Supervisor Signature: *[Signature]* MB

Temp : 1. 96°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
ICV	30mL	MP88185
CCV	30mL	MP88187
CRA	30mL	MP88189
Blank Spike	0.48mL	MP88177
Matrix Spike	0.48mL	MP88177

Chemical Used	ML/SAMPLE USED	Lot Number
AQUA REGIA	1.5MI	MP88191
KMnO4 (5%)	4.5mL	MP87634
Hydroxylamine HCL (12%)	2.0MI	MP87635
PTFE Boiling Stones	-----	M5582
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

LAB SAMPLE ID	CLIENT SAMPLE ID	Wt(g)/Vol(ml)	Comment
0.0 ppb	S0	30mL	MP88178
0.05 ppb	S0.05	N/A	N/A
0.2 ppb	S0.2	30mL	MP88180
2.5 ppb	S2.5	30mL	MP88181
5.0 ppb	S5.0	30mL	MP88182
7.5 ppb	S7.5	30mL	MP88183
10.0 ppb	S10.0	30mL	MP88184
ICV	ICV	30mL	MP88185
ICB	ICB	30mL	MP88186
CCV	CCV	30mL	MP88187
CCB	CCB	30mL	MP88188
CRI	CRI	30mL	MP88189
CHK STD	CHK STD	30mL	MP88190

Extraction Conformance/Non-Conformance Comments:

N/A		
Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
12/02/25 15:05	<i>[Signature]</i> MB Met/143	<i>[Signature]</i> MB Met/143
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Weight (g)	Final Vol (ml)	pH	Comment	Prep Pos
PB170790BL	PBS790	0.50	35	N/A	N/A	3-1
PB170790BS	LCS790	0.50	35	N/A	MP88177	2
Q3742-01	SB-1125-23	0.51	35	N/A	N/A	3
Q3742-02	SB-1125-19	0.53	35	N/A	N/A	4
Q3742-03	SB-1125-8	0.55	35	N/A	N/A	5
Q3742-04	SB-1125-9	0.53	35	N/A	N/A	6
Q3742-05	SB-1125-21	0.51	35	N/A	N/A	7
Q3742-06	SB-1125-13	0.57	35	N/A	N/A	8
Q3742-07	SB-1125-18	0.55	35	N/A	N/A	9
Q3742-08	SB-1125-20	0.57	35	N/A	N/A	10
Q3742-09	SB-1125-22	0.55	35	N/A	N/A	11
Q3742-10	SB-1125-2	0.54	35	N/A	N/A	12
Q3742-16	SB-1125-1	0.50	35	N/A	N/A	13
Q3742-17	SB-1125-25	0.56	35	N/A	N/A	14
Q3742-18	SB-1125-24	0.60	35	N/A	N/A	15
Q3746-01	ROW	0.57	35	N/A	N/A	16
Q3748-01	AU-05-12-1-2025	0.57	35	N/A	N/A	17
Q3750-01	FENCE WC-1	0.53	35	N/A	N/A	18
Q3750-01MS	FENCE WC-1MS	0.58	35	N/A	MP88177	20
Q3750-01MSD	FENCE WC-1MSD	0.53	35	N/A	MP88177	21
Q3750-01DUP	FENCE WC-1DUP	0.54	35	N/A	N/A	19
Q3750-04	FENCE WC-2	0.60	35	N/A	N/A	22

WORKLIST(Hardcopy Internal Chain)

Worklist Name : 120225_7471

Worklist ID : 193423

Department : Digestion

Date : 12-02-2025 13:01:07

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3742-01	SB-1125-23	Solid	Mercury	Cool 4 deg C	REM101	A21	11/25/6025	7471B
Q3742-02	SB-1125-19	Solid	Mercury	Cool 4 deg C	REM101	A21	11/25/2025	7471B
Q3742-03	SB-1125-8	Solid	Mercury	Cool 4 deg C	REM101	A21	11/25/2025	7471B
Q3742-04	SB-1125-9	Solid	Mercury	Cool 4 deg C	REM101	A21	11/25/2025	7471B
Q3742-05	SB-1125-21	Solid	Mercury	Cool 4 deg C	REM101	A21	11/25/2025	7471B
Q3742-06	SB-1125-13	Solid	Mercury	Cool 4 deg C	REM101	A21	11/25/2025	7471B
Q3742-07	SB-1125-18	Solid	Mercury	Cool 4 deg C	REM101	A21	11/25/2025	7471B
Q3742-08	SB-1125-20	Solid	Mercury	Cool 4 deg C	REM101	A21	11/25/2025	7471B
Q3742-09	SB-1125-22	Solid	Mercury	Cool 4 deg C	REM101	A21	11/25/2025	7471B
Q3742-10	SB-1125-2	Solid	Mercury	Cool 4 deg C	REM101	A21	11/25/2025	7471B
Q3742-16	SB-1125-1	Solid	Mercury	Cool 4 deg C	REM101	A21	11/25/2025	7471B
Q3742-17	SB-1125-25	Solid	Mercury	Cool 4 deg C	REM101	A21	11/25/2025	7471B
Q3742-18	SB-1125-24	Solid	Mercury	Cool 4 deg C	REM101	A21	11/25/2025	7471B
Q3746-01	ROW	Solid	Mercury	Cool 4 deg C	REM101	A21	11/25/2025	7471B
Q3748-01	AU-05-12-1-2025	Solid	Mercury	Cool 4 deg C	PSEG05	A11	12/02/2025	7471B
Q3750-01	FENCE WC-1	Solid	Mercury	Cool 4 deg C	PSEG05	A12	12/01/2025	7471B
Q3750-04	FENCE WC-2	Solid	Mercury	Cool 4 deg C	PSEG05	A11	12/01/2025	7471B

Date/Time 12/02/25 13:26

Raw Sample Received by: Sismetdig

Raw Sample Relinquished by: CWSM)

Date/Time 12/02/2025 14:20

Raw Sample Received by: CWSM)

Raw Sample Relinquished by: Sismetdig

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/2/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 12/01/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:25
Out Date: 12/02/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB138066

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3741-01	B50-SP3-111925	1	1.15	10.38	11.53	10.66	91.6	
Q3741-02	B50-SP13-112025	2	1.18	10.33	11.51	10.1	86.4	
Q3741-03	B50-SP14-112025	3	1.19	10.25	11.44	10.14	87.3	
Q3741-04	B50-SP9-112425	4	1.16	10.61	11.77	10.2	85.2	
Q3741-05	B50-SP11-112525	5	1.16	10.47	11.63	10.02	84.6	
Q3741-06	B50-SP4-112525	6	1.18	10.50	11.68	10.2	85.9	
Q3742-01	SB-1125-23	7	1.11	10.61	11.72	8.55	70.1	
Q3742-02	SB-1125-19	8	1.19	11.07	12.26	5.72	40.9	
Q3742-03	SB-1125-8	9	1.17	10.55	11.72	10.16	85.2	
Q3742-04	SB-1125-9	10	1.19	10.57	11.76	9.5	78.6	
Q3742-05	SB-1125-21	11	1.16	10.61	11.77	9.6	79.5	
Q3742-06	SB-1125-13	12	1.18	10.18	11.36	9.88	85.5	
Q3742-07	SB-1125-18	13	1.11	10.19	11.3	8.07	68.3	
Q3742-08	SB-1125-20	14	1.15	10.84	11.99	9.86	80.4	
Q3742-09	SB-1125-22	15	1.12	10.90	12.02	9.05	72.8	
Q3742-10	SB-1125-2	16	1.16	10.76	11.92	9.38	76.4	
Q3742-11	SB-1125-3A	17	1.19	10.58	11.77	8.93	73.2	
Q3742-12	SB-1125-3B	18	1.17	10.50	11.67	10.55	89.3	
Q3742-13	SB-1125-3C	19	1.16	10.50	11.66	9.31	77.6	
Q3742-14	SB-1125-3D	20	1.13	10.45	11.58	8.5	70.5	
Q3742-14D	SB-1125-3DDUP	21	1.15	10.45	11.6	8.53	70.6	
Q3742-15	SB-1125-3E	22	1.13	10.39	11.52	8.05	66.6	
Q3742-16	SB-1125-1	23	1.19	10.33	11.52	8.78	73.5	
Q3742-17	SB-1125-25	24	1.18	11.18	12.36	8.26	63.3	
Q3742-18	SB-1125-24	25	1.15	10.90	12.05	7.99	62.8	
Q3745-05	SVOC-GPC-BLANK	31	1.00	1.00	2.00	2.00	100.0	
Q3745-06	PEST-GPC-BLANK	32	1.00	1.00	2.00	2.00	100.0	
Q3745-07	PEST-GPC-BLANK-SPIKE	33	1.00	1.00	2.00	2.00	100.0	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/2/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 12/01/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:25
Out Date: 12/02/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB138066

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3745-08	SVOC-GPC2-BLANK	34	1.00	1.00	2.00	2.00	100.0	
Q3745-09	PEST-GPC2-BLANK	35	1.00	1.00	2.00	2.00	100.0	
Q3745-10	PEST -GPC2-BLANK-SPIKE	36	1.00	1.00	2.00	2.00	100.0	
Q3746-01	ROW	26	1.15	10.84	11.99	8.59	68.6	
Q3746-02	ROW-E2	27	1.14	10.58	11.72	5.2	38.4	
Q3747-01	1	28	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3747-02	2	29	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3747-03	3	30	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3748-01	AU-05-12-1-2025	37	1.18	10.75	11.93	11.03	91.6	
Q3748-02	AU-05-12-1-2025	38	1.18	10.55	11.73	10.88	91.9	
Q3750-01	FENCE WC-1	39	1.15	10.77	11.92	10.5	86.8	
Q3750-02	FENCE WC-1-EPH	40	1.19	10.31	11.5	10.15	86.9	
Q3750-03	FENCE WC-1-VOC	41	1.16	10.58	11.74	9.72	80.9	
Q3750-04	FENCE WC-2	42	1.11	10.51	11.62	10.2	86.5	
Q3750-05	FENCE WC-2-EPH	43	1.13	10.21	11.34	9.97	86.6	
Q3750-06	FENCE WC-2-VOC	44	1.19	10.38	11.57	10.52	89.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

128066

WorkList Name : %1-120125

WorkList ID : 193391

Department : Wet-Chemistry

Date : 12-01-2025 08:55:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3741-01	B50-SP3-111925	Solid	Percent Solids	Cool 4 deg C	CORE02	--Sele	11/19/2025	Chemtech -SO
Q3741-02	B50-SP13-112025	Solid	Percent Solids	Cool 4 deg C	CORE02	--Sele	11/20/2025	Chemtech -SO
Q3741-03	B50-SP14-112025	Solid	Percent Solids	Cool 4 deg C	CORE02	--Sele	11/20/2025	Chemtech -SO
Q3741-04	B50-SP9-112425	Solid	Percent Solids	Cool 4 deg C	CORE02	--Sele	11/24/2025	Chemtech -SO
Q3741-05	B50-SP11-112525	Solid	Percent Solids	Cool 4 deg C	CORE02	--Sele	11/25/2025	Chemtech -SO
Q3741-06	B50-SP4-112525	Solid	Percent Solids	Cool 4 deg C	CORE02	--Sele	11/25/2025	Chemtech -SO
Q3742-01	SB-1125-23	Solid	Percent Solids	Cool 4 deg C	REMI01	A21	11/25/6025	Chemtech -SO
Q3742-02	SB-1125-19	Solid	Percent Solids	Cool 4 deg C	REMI01	A21	11/25/2025	Chemtech -SO
Q3742-03	SB-1125-8	Solid	Percent Solids	Cool 4 deg C	REMI01	A21	11/25/2025	Chemtech -SO
Q3742-04	SB-1125-9	Solid	Percent Solids	Cool 4 deg C	REMI01	A21	11/25/2025	Chemtech -SO
Q3742-05	SB-1125-21	Solid	Percent Solids	Cool 4 deg C	REMI01	A21	11/25/2025	Chemtech -SO
Q3742-06	SB-1125-13	Solid	Percent Solids	Cool 4 deg C	REMI01	A21	11/25/2025	Chemtech -SO
Q3742-07	SB-1125-18	Solid	Percent Solids	Cool 4 deg C	REMI01	A21	11/25/2025	Chemtech -SO
Q3742-08	SB-1125-20	Solid	Percent Solids	Cool 4 deg C	REMI01	A21	11/25/2025	Chemtech -SO
Q3742-09	SB-1125-22	Solid	Percent Solids	Cool 4 deg C	REMI01	A21	11/25/2025	Chemtech -SO
Q3742-10	SB-1125-2	Solid	Percent Solids	Cool 4 deg C	REMI01	A21	11/25/2025	Chemtech -SO
Q3742-11	SB-1125-3A	Solid	Percent Solids	Cool 4 deg C	REMI01	A21	11/25/2025	Chemtech -SO
Q3742-12	SB-1125-3B	Solid	Percent Solids	Cool 4 deg C	REMI01	A21	11/25/2025	Chemtech -SO
Q3742-13	SB-1125-3C	Solid	Percent Solids	Cool 4 deg C	REMI01	A21	11/25/2025	Chemtech -SO
Q3742-14	SB-1125-3D	Solid	Percent Solids	Cool 4 deg C	REMI01	A21	11/25/2025	Chemtech -SO
Q3742-15	SB-1125-3E	Solid	Percent Solids	Cool 4 deg C	REMI01	A21	11/25/2025	Chemtech -SO

Date/Time 12/01/25 15:13
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 12/01/25 17:40
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

g 172066

WorkList Name : %1-120125

WorkList ID : 193391

Department : Wet-Chemistry

Date : 12-01-2025 08:55:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3742-16	SB-1125-1	Solid	Percent Solids	Cool 4 deg C	REMI01	A21	11/25/2025	Chemtech -SO
Q3742-17	SB-1125-25	Solid	Percent Solids	Cool 4 deg C	REMI01	A21	11/25/2025	Chemtech -SO
Q3742-18	SB-1125-24	Solid	Percent Solids	Cool 4 deg C	REMI01	A21	11/25/2025	Chemtech -SO
Q3745-05	SVOC-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	A12	11/21/2025	Chemtech -SO
Q3745-06	PEST-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	A12	11/21/2025	Chemtech -SO
Q3745-07	PEST-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	A12	11/21/2025	Chemtech -SO
Q3745-08	SVOC-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	A12	11/21/2025	Chemtech -SO
Q3745-09	PEST-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	A12	11/21/2025	Chemtech -SO
Q3745-10	PEST -GPC2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	A12	11/21/2025	Chemtech -SO
Q3746-01	ROW	Solid	Percent Solids	Cool 4 deg C	PSEG05	A11	12/02/2025	Chemtech -SO
Q3746-02	ROW-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	A11	12/02/2025	Chemtech -SO
Q3747-01	1	Solid	Percent Solids	Cool 4 deg C	PSEG03	A11	12/01/2025	Chemtech -SO
Q3747-02	2	Solid	Percent Solids	Cool 4 deg C	PSEG03	A11	12/01/2025	Chemtech -SO
Q3747-03	3	Solid	Percent Solids	Cool 4 deg C	PSEG03	A11	12/01/2025	Chemtech -SO
Q3748-01	AU-05-12-1-2025	Solid	Percent Solids	Cool 4 deg C	PSEG03	A11	12/01/2025	Chemtech -SO
Q3748-02	AU-05-12-1-2025	Solid	Percent Solids	Cool 4 deg C	PSEG05	A12	12/01/2025	Chemtech -SO
Q3750-01	FENCE WC-1	Solid	Percent Solids	Cool 4 deg C	PSEG05	A11	12/01/2025	Chemtech -SO
Q3750-02	FENCE WC-1-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG05	A11	12/01/2025	Chemtech -SO
Q3750-03	FENCE WC-1-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG05	A11	12/01/2025	Chemtech -SO
Q3750-04	FENCE WC-2	Solid	Percent Solids	Cool 4 deg C	PSEG05	A11	12/01/2025	Chemtech -SO
Q3750-05	FENCE WC-2-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG05	A11	12/01/2025	Chemtech -SO

Date/Time 12/01/25 15:15
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 12/01/25 17:40
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

138066

WorkList Name : %1-120125

WorkList ID : 193391

Department : Wet-Chemistry

Date : 12-01-2025 08:55:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3750-06	FENCE WC-2-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG05	A11	12/01/2025	Chemtech -SO

Date/Time 12/01/25 13:15
 Raw Sample Received by: SB WPC
 Raw Sample Relinquished by: CR 8

Date/Time 12/01/25 17:40
 Raw Sample Received by: CR 8
 Raw Sample Relinquished by: SB WPC

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB138085

Review By	Janvi	Review On	12/3/2025 10:09:18 AM
Supervise By	jaswal	Supervise On	12/5/2025 9:46:28 PM

STD. NAME	STD REF.#
ICAL Standard	MP88101,MP88123,MP88121,MP88120,MP88119,MP88118
ICV Standard	MP88124,MP88123
CCV Standard	MP88127
ICSA Standard	MP88125,MP88126
CRI Standard	MP88120
LCS Standard	
Chk Standard	MP88128,MP88129,MP88130,MP88131,MP88132

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	12/02/25 11:04		Jaswal	OK
2	S1	S1	CAL2	12/02/25 11:08		Jaswal	OK
3	S2	S2	CAL3	12/02/25 11:12		Jaswal	OK
4	S3	S3	CAL4	12/02/25 11:16		Jaswal	OK
5	S4	S4	CAL5	12/02/25 11:21		Jaswal	OK
6	S5	S5	CAL6	12/02/25 11:25		Jaswal	OK
7	ICV01	ICV01	ICV	12/02/25 11:29		Jaswal	OK
8	LLICV01	LLICV01	LLICV	12/02/25 11:34		Jaswal	OK
9	ICB01	ICB01	ICB	12/02/25 11:38		Jaswal	OK
10	CRI01	CRI01	CRDL	12/02/25 11:42		Jaswal	OK
11	ICSA01	ICSA01	ICSA	12/02/25 11:52		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	12/02/25 11:57		Jaswal	OK
13	ICSADL	ICSADL	ICSA	12/02/25 12:01		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	12/02/25 12:05		Jaswal	OK
15	CCV01	CCV01	CCV	12/02/25 12:10		Jaswal	OK
16	CCB01	CCB01	CCB	12/02/25 12:14		Jaswal	OK
17	Q3739-01DL2	GSB1ADL2	SAM	12/02/25 12:19	NOT USE	Jaswal	Not Ok
18	Q3739-02DL2	GSB1BDL2	SAM	12/02/25 12:23	25x for Cr	Jaswal	Confirms

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB138085

Review By	Janvi	Review On	12/3/2025 10:09:18 AM
Supervise By	jaswal	Supervise On	12/5/2025 9:46:28 PM
STD. NAME	STD REF.#		
ICAL Standard	MP88101,MP88123,MP88121,MP88120,MP88119,MP88118		
ICV Standard	MP88124,MP88123		
CCV Standard	MP88127		
ICSA Standard	MP88125,MP88126		
CRI Standard	MP88120		
LCS Standard			
Chk Standard	MP88128,MP88129,MP88130,MP88131,MP88132		

19	Q3740-01DL2	WC1DL2	SAM	12/02/25 12:27	25x for Cr	Jaswal	Confirms
20	Q3740-01DUPDL2	WC1DUPDL2	DUP	12/02/25 12:31	25x for Cr	Jaswal	Confirms
21	Q3740-01LDL2	WC1LDL2	SD	12/02/25 12:36	Not Required	Jaswal	Not Ok
22	Q3740-01MSDL2	WC1MSDL2	MS	12/02/25 12:40	25x for Cr	Jaswal	Confirms
23	Q3740-01MSDDL2	WC1MSDDL2	MSD	12/02/25 12:44	25x for Cr	Jaswal	Confirms
24	Q3740-01ADL2	WC1ADL2	PS	12/02/25 12:48	Not Required (0.1 ML OF EACH M6008 AND M6017 IN 10ML SAMPLE BEFORE DILUTION)	Jaswal	Not Ok
25	LR-2	LR-2	HIGH STD	12/02/25 12:57		Jaswal	OK
26	LR-1	LR-1	HIGH STD	12/02/25 13:08		Jaswal	OK
27	CCV02	CCV02	CCV	12/02/25 13:13		Jaswal	OK
28	CCB02	CCB02	CCB	12/02/25 13:17		Jaswal	OK
29	LR-3	LR-3	HIGH STD	12/02/25 13:21		Jaswal	OK
30	PB170777BL	PB170777BL	MB	12/02/25 13:26		Jaswal	OK
31	PB170777BS	PB170777BS	LCS	12/02/25 13:30		Jaswal	OK
32	PB170778BL	PB170778BL	MB	12/02/25 13:34		Jaswal	OK
33	PB170778BS	PB170778BS	LCS	12/02/25 13:39		Jaswal	OK
34	Q3746-01	ROW	SAM	12/02/25 13:51		Jaswal	OK
35	Q3748-01	AU-05-12-1-2025	SAM	12/02/25 13:55		Jaswal	OK
36	Q3750-04	FENCE WC-2	SAM	12/02/25 13:59		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB138085

Review By	Janvi	Review On	12/3/2025 10:09:18 AM
Supervise By	jaswal	Supervise On	12/5/2025 9:46:28 PM
STD. NAME	STD REF.#		
ICAL Standard	MP88101,MP88123,MP88121,MP88120,MP88119,MP88118		
ICV Standard	MP88124,MP88123		
CCV Standard	MP88127		
ICSA Standard	MP88125,MP88126		
CRI Standard	MP88120		
LCS Standard			
Chk Standard	MP88128,MP88129,MP88130,MP88131,MP88132		

37	Q3741-01	B50-SP3-111925	SAM	12/02/25 14:03		Jaswal	OK
38	Q3741-02	B50-SP13-112025	SAM	12/02/25 14:07		Jaswal	OK
39	CCV03	CCV03	CCV	12/02/25 14:11		Jaswal	OK
40	CCB03	CCB03	CCB	12/02/25 14:15		Jaswal	OK
41	Q3741-03	B50-SP14-112025	SAM	12/02/25 14:24		Jaswal	OK
42	Q3741-04	B50-SP9-112425	SAM	12/02/25 14:28		Jaswal	OK
43	Q3741-05	B50-SP11-112525	SAM	12/02/25 14:33		Jaswal	OK
44	Q3741-06	B50-SP4-112525	SAM	12/02/25 14:47		Jaswal	OK
45	Q3750-01	FENCE WC-1	SAM	12/02/25 14:51		Jaswal	OK
46	Q3750-01DUP	FENCE WC-1DUP	DUP	12/02/25 14:55		Jaswal	OK
47	Q3750-01L	FENCE WC-1L	SD	12/02/25 14:59		Jaswal	OK
48	Q3750-01MS	FENCE WC-1MS	MS	12/02/25 15:03		Jaswal	OK
49	Q3750-01MSD	FENCE WC-1MSD	MSD	12/02/25 15:07		Jaswal	OK
50	Q3750-01A	FENCE WC-1A	PS	12/02/25 15:11	0.1 ML OF EACH M6008 AND M6017 IN 10ML SAMPLE.	Jaswal	OK
51	CCV04	CCV04	CCV	12/02/25 15:15		Jaswal	OK
52	CCB04	CCB04	CCB	12/02/25 15:20		Jaswal	OK
53	Q3742-01	SB-1125-23	SAM	12/02/25 15:24		Jaswal	OK
54	Q3742-02	SB-1125-19	SAM	12/02/25 15:28		Jaswal	OK
55	Q3742-03	SB-1125-8	SAM	12/02/25 15:33		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB138085

Review By	Janvi	Review On	12/3/2025 10:09:18 AM
Supervise By	jaswal	Supervise On	12/5/2025 9:46:28 PM
STD. NAME	STD REF.#		
ICAL Standard	MP88101,MP88123,MP88121,MP88120,MP88119,MP88118		
ICV Standard	MP88124,MP88123		
CCV Standard	MP88127		
ICSA Standard	MP88125,MP88126		
CRI Standard	MP88120		
LCS Standard			
Chk Standard	MP88128,MP88129,MP88130,MP88131,MP88132		

56	Q3742-04	SB-1125-9	SAM	12/02/25 15:37		Jaswal	OK
57	Q3742-05	SB-1125-21	SAM	12/02/25 15:41		Jaswal	OK
58	Q3742-06	SB-1125-13	SAM	12/02/25 15:45		Jaswal	OK
59	Q3742-07	SB-1125-18	SAM	12/02/25 15:49		Jaswal	OK
60	Q3742-08	SB-1125-20	SAM	12/02/25 15:53		Jaswal	OK
61	Q3742-09	SB-1125-22	SAM	12/02/25 15:57		Jaswal	OK
62	Q3742-10	SB-1125-2	SAM	12/02/25 16:01		Jaswal	OK
63	CCV05	CCV05	CCV	12/02/25 16:05		Jaswal	OK
64	CCB05	CCB05	CCB	12/02/25 16:09		Jaswal	OK
65	Q3742-16	SB-1125-1	SAM	12/02/25 16:14		Jaswal	OK
66	Q3742-17	SB-1125-25	SAM	12/02/25 16:18		Jaswal	OK
67	Q3742-18	SB-1125-24	SAM	12/02/25 16:22		Jaswal	OK
68	Q3742-18DUP	SB-1125-24DUP	DUP	12/02/25 16:26		Jaswal	OK
69	Q3742-18L	SB-1125-24L	SD	12/02/25 16:30		Jaswal	OK
70	Q3742-18MS	SB-1125-24MS	MS	12/02/25 16:34		Jaswal	OK
71	Q3742-18MSD	SB-1125-24MSD	MSD	12/02/25 16:38		Jaswal	OK
72	Q3742-18A	SB-1125-24A	PS	12/02/25 16:42	0.1 ML OF EACH M6008 AND M6017 IN 10ML SAMPLE.	Jaswal	OK
73	PB170754TB	PB170754TB	MB	12/02/25 16:46		Jaswal	OK
74	PB170755TB	PB170755TB	MB	12/02/25 16:51		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB138085

Review By	Janvi	Review On	12/3/2025 10:09:18 AM
Supervise By	jaswal	Supervise On	12/5/2025 9:46:28 PM

STD. NAME	STD REF.#
ICAL Standard	MP88101,MP88123,MP88121,MP88120,MP88119,MP88118
ICV Standard	MP88124,MP88123
CCV Standard	MP88127
ICSA Standard	MP88125,MP88126
CRI Standard	MP88120
LCS Standard	
Chk Standard	MP88128,MP88129,MP88130,MP88131,MP88132

75	CCV06	CCV06	CCV	12/02/25 16:55		Jaswal	OK
76	CCB06	CCB06	CCB	12/02/25 16:59		Jaswal	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB138086

Review By	jaswal	Review On	12/5/2025 8:54:33 PM
Supervise By	mohan	Supervise On	12/5/2025 8:59:54 PM

STD. NAME	STD REF.#
ICAL Standard	MP88178,MP88180,MP88181,MP88182,MP88183,MP88184
ICV Standard	MP88185
CCV Standard	MP88187
ICSA Standard	
CRI Standard	MP88189
LCS Standard	
Chk Standard	MP88186,MP88188,MP88190,MP88194

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	12/03/25 08:31		mohan	OK
2	S0.2	S0.2	CAL2	12/03/25 08:33		mohan	OK
3	S2.5	S2.5	CAL3	12/03/25 08:35		mohan	OK
4	S5	S5	CAL4	12/03/25 08:38		mohan	OK
5	S7.5	S7.5	CAL5	12/03/25 08:40		mohan	OK
6	S10	S10	CAL6	12/03/25 08:42		mohan	OK
7	ICV135	ICV135	ICV	12/03/25 08:46		mohan	OK
8	ICB135	ICB135	ICB	12/03/25 08:48		mohan	OK
9	CCV53	CCV53	CCV	12/03/25 08:50		mohan	OK
10	CCB53	CCB53	CCB	12/03/25 08:53		mohan	OK
11	CRA	CRA	CRDL	12/03/25 08:55		mohan	OK
12	HighStd	HighStd	HIGH STD	12/03/25 08:57		mohan	OK
13	ChkStd	ChkStd	SAM	12/03/25 08:59		mohan	OK
14	PB170790BL	PB170790BL	MB	12/03/25 09:02		mohan	OK
15	PB170790BS	PB170790BS	LCS	12/03/25 09:04		mohan	OK
16	Q3742-01	SB-1125-23	SAM	12/03/25 09:06		mohan	OK
17	Q3742-02	SB-1125-19	SAM	12/03/25 09:09		mohan	OK
18	Q3742-03	SB-1125-8	SAM	12/03/25 09:11		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB138086

Review By	jaswal	Review On	12/5/2025 8:54:33 PM
Supervise By	mohan	Supervise On	12/5/2025 8:59:54 PM
STD. NAME	STD REF.#		
ICAL Standard	MP88178,MP88180,MP88181,MP88182,MP88183,MP88184		
ICV Standard	MP88185		
CCV Standard	MP88187		
ICSA Standard			
CRI Standard	MP88189		
LCS Standard			
Chk Standard	MP88186,MP88188,MP88190,MP88194		

19	Q3742-04	SB-1125-9	SAM	12/03/25 09:13		mohan	OK
20	Q3742-05	SB-1125-21	SAM	12/03/25 09:16		mohan	OK
21	CCV54	CCV54	CCV	12/03/25 09:18		mohan	OK
22	CCB54	CCB54	CCB	12/03/25 09:20		mohan	OK
23	Q3742-06	SB-1125-13	SAM	12/03/25 09:22		mohan	OK
24	Q3742-07	SB-1125-18	SAM	12/03/25 09:25		mohan	OK
25	Q3742-08	SB-1125-20	SAM	12/03/25 09:27		mohan	OK
26	Q3742-09	SB-1125-22	SAM	12/03/25 09:29	Hg is high	mohan	Dilution
27	Q3742-10	SB-1125-2	SAM	12/03/25 09:31		mohan	OK
28	Q3742-16	SB-1125-1	SAM	12/03/25 09:34		mohan	OK
29	Q3742-17	SB-1125-25	SAM	12/03/25 09:36		mohan	OK
30	Q3742-18	SB-1125-24	SAM	12/03/25 09:38		mohan	OK
31	Q3746-01	ROW	SAM	12/03/25 09:41		mohan	OK
32	Q3748-01	AU-05-12-1-2025	SAM	12/03/25 09:43		mohan	OK
33	CCV55	CCV55	CCV	12/03/25 09:45		mohan	OK
34	CCB55	CCB55	CCB	12/03/25 09:48		mohan	OK
35	Q3750-01	FENCE WC-1	SAM	12/03/25 09:55		mohan	OK
36	Q3750-01DUP	FENCE WC-1DUP	DUP	12/03/25 09:57		mohan	OK
37	Q3750-01MS	FENCE WC-1MS	MS	12/03/25 09:59		mohan	OK
38	Q3750-01MSD	FENCE WC-1MSD	MSD	12/03/25 10:02		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB138086

Review By	jaswal	Review On	12/5/2025 8:54:33 PM
Supervise By	mohan	Supervise On	12/5/2025 8:59:54 PM
STD. NAME	STD REF.#		
ICAL Standard	MP88178,MP88180,MP88181,MP88182,MP88183,MP88184		
ICV Standard	MP88185		
CCV Standard	MP88187		
ICSA Standard			
CRI Standard	MP88189		
LCS Standard			
Chk Standard	MP88186,MP88188,MP88190,MP88194		

39	Q3750-04	FENCE WC-2	SAM	12/03/25 10:04	Hg is high	mohan	Dilution
40	Q3750-01L	FENCE WC-1L	SD	12/03/25 10:06		mohan	OK
41	Q3750-01A	FENCE WC-1A	PS	12/03/25 10:09		mohan	OK
42	Q3742-09DL	SB-1125-22DL	SAM	12/03/25 10:11	2X For Hg	mohan	Confirms
43	Q3750-04DL	FENCE WC-2DL	SAM	12/03/25 11:01	5X For Hg	mohan	Confirms
44	CCV56	CCV56	CCV	12/03/25 11:09		mohan	OK
45	CCB56	CCB56	CCB	12/03/25 11:11		mohan	OK



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:
COMPANY: Remington Vernick Engineering
ADDRESS: 2059 Sprydale Road
CITY: Cherry Hill STATE: NJ ZIP: 08003
ATTENTION: Kyle Carlson @ RVE.com
PHONE: 609-682-3049 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Saddler Property
PROJECT NO.: 04002003 LOCATION: Camden, NJ
PROJECT MANAGER: Kyle Carlson
e-mail: Kyle.Carlson@RVE.com
PHONE: 609-682-3049 FAX:

CLIENT BILLING INFORMATION

BILL TO: AP Invoices @ RVE.com PO#: 04002003
ADDRESS: 2059 Sprydale Road
CITY: Cherry Hill STATE: NJ ZIP: 08003
ATTENTION: Kyle Carlson PHONE: 609-682-3049

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): Standard DAYS*
EDD: _____ DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☒ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other
☒ EDD FORMAT NJ Haz Site

TAL/TCL/3049
Cyanide
EPH Cat 2 (non-fuel)
VOCs
BNA + 15
PCBs + TAL Method
TAL Method (MC)
EPH Cat 1

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS
			COMP	GRAB	DATE	TIME		E	E	F	E	E	E	E	E	E	
1.	SB-1125-23	Soil		X	11/26/25	850	7	3		1	3						F-Encore Samples
2.	SB-1125-19	Soil		X	11/25/25	855	7	3		1	3						
3.	SB-1125-8	Soil		X	11/25/25	930	6			3	1	1		1			See Below Comments *
4.	SB-1125-9	Soil		X	11/25/25	935	6			3	1	1		1			See Below Comments *
5.	SB-1125-8 (MC)	Soil		X	11/25/25	930	6			3							
6.	SB-1125-21	Soil		X	11/25/25	1105	7	3		1	3						
7.	SB-1125-13	Soil		X	11/25/25	1115	7	3		1	3						
8.	SB-1125-18	Soil		X	11/25/25	1125	7	3		1	3						
9.	SB-1125-20	Soil		X	11/25/25	1205	7	3		1	3						
10.	SB-1125-22	Soil		X	11/25/25	1245	7	3		1	3						

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP
1. <u>Mario Carilli @ RVE</u>	<u>11-26-25 11:00</u>	<u>[Signature]</u> <u>11-26-25</u>	Comments: <u>* For Samples SB-1125-8 and SB-1125-9 Run ALL Analysis, but only Report EPH First then if EPH is detected RVE will choose to report Remaining Parameters.</u>
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	
2.		2.	
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	
3. <u>[Signature]</u>	<u>11-26-25</u>	3.	

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: Remington + Vernick Engineers
ADDRESS: 2059 Springdale Road
CITY: Cherry Hill STATE: NJ ZIP: 08003
ATTENTION: Kyle Carlson & RVE, Inc.
PHONE: 609-682-3049 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Saddle Property
PROJECT NO. 04002003 LOCATION: Camden, NJ
PROJECT MANAGER: Kyle Carlson
e-mail: Kyle.Carlson@RVE.com
PHONE: 609-682-3049 FAX:

CLIENT BILLING INFORMATION

BILL TO: AP Invoices @ RVE, Inc. PO#: 04002003
ADDRESS: 2059 Springdale Road
CITY: Cherry Hill STATE: NJ ZIP: 08003
ATTENTION: Kyle Carlson PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): Standard _____ DAYS*
EDD: _____ DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☒ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other _____
EDD FORMAT: NJ Hazmat

TAL/TEL 200 2000
EPH Car 2 Nbr 1000
VOCs
EPH Car 1

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS
			COMP	GRAB	DATE	TIME		E	E	F	E						
1.	SB-1125-2	Soil		X	11/25/25	1305	7	3	1	3	1						F = Encores
2.	SB-1125-3A	Soil		X	11/25/25	1330					1						
3.	SB-1125-3B	Soil		X	11/25/25	1335					1						
4.	SB-1125-3C	Soil		X	11/25/25	1340					1						
5.	SB-1125-3D	Soil		X	11/25/25	1355					1						
6.	SB-1125-3E	Soil		X	11/25/25	1355					1						
7.	SB-1125-1	Soil		X	11/25/25	1430	7	3	1	3							
8.	SB-1125-25	Soil		X	11/25/25	1430	7	3	1	3							
9.	SB-1125-24	Soil		X	11/25/25	1505	7	3	1	3							
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	Conditions of bottles or coolers at receipt:
1. Marco Carillo @ RVE	11/26/25 12PM	1. [Signature] 11-26-25	☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 2.5 °C
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	Comments:
2.		2.	
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	
3. [Signature]	11-26-25	3.	

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q3742 REMI01

Order Date : 12/1/2025 9:06:00 AM

Project Mgr :

Client Name : Remington & Vernick

Project Name : Sadler Property

Report Type : Level 1

Client Contact : Justin Zarzecki

Receive DateTime : 11/26/2025 6:20:00 PM

EDD Type : EXCEL NJCLEANUP

Invoice Name : Remington & Vernick

Purchase Order :

Hard Copy Date :

Invoice Contact : Justin Zarzecki

Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q3742-01	SB-1125-23	Solid	11/25/6025	08:50	VOC-TCLVOA-10	TCL+30/TAL	8260C	10 Bus. Days	
Q3742-02	SB-1125-19	Solid	11/25/2025	08:55	VOC-TCLVOA-10	TCL+30/TAL	8260C	10 Bus. Days	
Q3742-03	SB-1125-8	Solid	11/25/2025	09:30	VOC-TCLVOA-10	TCL+30/TAL	8260C	10 Bus. Days	
Q3742-04	SB-1125-9	Solid	11/25/2025	09:35	VOC-TCLVOA-10	TCL+30/TAL	8260C	10 Bus. Days	
Q3742-05	SB-1125-21	Solid	11/25/2025	11:05	VOC-TCLVOA-10	TCL+30/TAL	8260C	10 Bus. Days	
Q3742-06	SB-1125-13	Solid	11/25/2025	11:15	VOC-TCLVOA-10	TCL+30/TAL	8260C	10 Bus. Days	
Q3742-07	SB-1125-18	Solid	11/25/2025	11:25	VOC-TCLVOA-10	TCL+30/TAL	8260C	10 Bus. Days	
Q3742-08	SB-1125-20	Solid	11/25/2025	12:05	VOC-TCLVOA-10	TCL+30/TAL	8260C	10 Bus. Days	

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q3742	REMI01	Order Date : 12/1/2025 9:06:00 AM	Project Mgr :
Client Name : Remington & Vernick		Project Name : Sadler Property	Report Type : Level 1
Client Contact : Justin Zarzecki		Receive DateTime : 11/26/2025 6:20:00 PM	EDD Type : EXCEL NJCLEANUP
Invoice Name : Remington & Vernick		Purchase Order :	Hard Copy Date :
Invoice Contact : Justin Zarzecki			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q3742-09	SB-1125-22	Solid	11/25/2025	12:45	VOC-TCLVOA-10	TCL+30/TAL	8260C	10 Bus. Days	
Q3742-10	SB-1125-2	Solid	11/25/2025	13:05	VOC-TCLVOA-10	TCL+30/TAL	8260C	10 Bus. Days	
Q3742-16	SB-1125-1	Solid	11/25/2025	14:30	VOC-TCLVOA-10	TCL+30/TAL	8260C	10 Bus. Days	
Q3742-17	SB-1125-25	Solid	11/25/2025	14:30	VOC-TCLVOA-10	TCL+30/TAL	8260C	10 Bus. Days	
Q3742-18	SB-1125-24	Solid	11/25/2025	15:05	VOC-TCLVOA-10	TCL+30/TAL	8260C	10 Bus. Days	

Relinquished By :

Date / Time :

CP
12/1/25 11:05

Received By :

Date / Time :

Sam
12/01/25 11:05 *NGH 6*
PJZ

Storage Area : VOA Refridgerator Room