

Continuing Calibration Report for SequenceID : FE120425AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE057171.D

Aliphatic C9-C12	12890824.000	60.000	3.194	6.819	214847.067	200655.681	-7.073
Aliphatic C12-C16	8655794.000	40.000	6.820	10.265	216394.850	203177.346	-6.505
Aliphatic C16-C21	12639459.000	60.000	10.266	13.639	210657.650	198315.768	-6.223
Aliphatic C21-C28	15912676.000	80.000	13.640	17.309	198908.450	198747.045	-0.081
Aliphatic C28-C40	22627036.000	120.000	17.310	22.235	188558.633	199899.890	5.673
Aliphatic EPH	72725789.000	360.000	3.194	22.235	202016.081	199869.809	-1.074