

Continuing Calibration Report for SequenceID : FE120425AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FE057187.D

|                   |              |         |        |        |            |            |         |
|-------------------|--------------|---------|--------|--------|------------|------------|---------|
| Aliphatic C9-C12  | 14323176.000 | 60.000  | 3.194  | 6.819  | 238719.600 | 200655.681 | -18.970 |
| Aliphatic C12-C16 | 9622943.000  | 40.000  | 6.820  | 10.265 | 240573.575 | 203177.346 | -18.406 |
| Aliphatic C16-C21 | 14001578.000 | 60.000  | 10.266 | 13.639 | 233359.633 | 198315.768 | -17.671 |
| Aliphatic C21-C28 | 17502830.000 | 80.000  | 13.640 | 17.309 | 218785.375 | 198747.045 | -10.082 |
| Aliphatic C28-C40 | 23665435.000 | 120.000 | 17.310 | 22.235 | 197211.958 | 199899.890 | 1.345   |
| Aliphatic EPH     | 79115962.000 | 360.000 | 3.194  | 22.235 | 219766.561 | 199869.809 | -9.955  |