

Continuing Calibration Report for SequenceID : FE120925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE057250.D

Aliphatic C9-C12	12792144.000	60.000	3.193	6.818	213202.400	200655.681	-6.253
Aliphatic C12-C16	8687540.000	40.000	6.819	10.263	217188.500	203177.346	-6.896
Aliphatic C16-C21	12552618.000	60.000	10.264	13.636	209210.300	198315.768	-5.494
Aliphatic C21-C28	15717983.000	80.000	13.637	17.304	196474.788	198747.045	1.143
Aliphatic C28-C40	22176911.000	120.000	17.305	22.230	184807.592	199899.890	7.550
Aliphatic EPH	71927196.000	360.000	3.193	22.230	199797.767	199869.809	0.036