

Continuing Calibration Report for SequenceID : FE120225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE057118.D

Aliphatic C9-C12	13883102.000	60.000	3.199	6.825	231385.033	200655.681	-15.314
Aliphatic C12-C16	9286544.000	40.000	6.826	10.272	232163.600	203177.346	-14.266
Aliphatic C21-C28	16747772.000	80.000	13.647	17.317	209347.150	198747.045	-5.333
Aliphatic C28-C40	23093544.000	120.000	17.318	22.251	192446.200	199899.890	3.729
Aliphatic EPH	76425229.000	360.000	3.199	22.251	212292.303	199869.809	-6.215
Aliphatic C16-C21	13414267.000	60.000	10.273	13.646	223571.117	198315.768	-12.735