

Continuing Calibration Report for SequenceID : FE120225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE057143.D

Aliphatic C9-C12	12214582.000	60.000	3.199	6.825	203576.367	200655.681	-1.456
Aliphatic C12-C16	8179345.000	40.000	6.826	10.272	204483.625	203177.346	-0.643
Aliphatic C21-C28	14776603.000	80.000	13.647	17.317	184707.538	198747.045	7.064
Aliphatic C28-C40	20281555.000	120.000	17.318	22.251	169012.958	199899.890	15.451
Aliphatic EPH	67280886.000	360.000	3.199	22.251	186891.350	199869.809	6.493
Aliphatic C16-C21	11828801.000	60.000	10.273	13.646	197146.683	198315.768	0.590