

Continuing Calibration Report for SequenceID : FE120225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE057149.D

Aliphatic C9-C12	12109422.000	60.000	3.199	6.825	201823.700	200655.681	-0.582
Aliphatic C12-C16	8134734.000	40.000	6.826	10.272	203368.350	203177.346	-0.094
Aliphatic C21-C28	14775055.000	80.000	13.647	17.317	184688.188	198747.045	7.074
Aliphatic C28-C40	20241882.000	120.000	17.318	22.251	168682.350	199899.890	15.617
Aliphatic EPH	67068270.000	360.000	3.199	22.251	186300.750	199869.809	6.789
Aliphatic C16-C21	11807177.000	60.000	10.273	13.646	196786.283	198315.768	0.771