

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Continuing Calib Date: 12/02/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 10:01

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.48       | 5.49      | 5.39      | 5.59  | 0.01       |
| Aroclor-1016-2       | (2) | 5.51       | 5.51      | 5.41      | 5.61  | 0.00       |
| Aroclor-1016-3       | (3) | 5.57       | 5.57      | 5.47      | 5.67  | 0.00       |
| Aroclor-1016-4       | (4) | 5.66       | 5.67      | 5.57      | 5.77  | 0.01       |
| Aroclor-1016-5       | (5) | 5.95       | 5.96      | 5.86      | 6.06  | 0.01       |
| Aroclor-1260-1       | (1) | 7.07       | 7.07      | 6.97      | 7.17  | 0.00       |
| Aroclor-1260-2       | (2) | 7.32       | 7.33      | 7.23      | 7.43  | 0.01       |
| Aroclor-1260-3       | (3) | 7.68       | 7.68      | 7.58      | 7.78  | 0.00       |
| Aroclor-1260-4       | (4) | 7.90       | 7.90      | 7.80      | 8.00  | 0.00       |
| Aroclor-1260-5       | (5) | 8.21       | 8.21      | 8.11      | 8.31  | 0.00       |
| Tetrachloro-m-xylene |     | 4.34       | 4.35      | 4.25      | 4.45  | 0.01       |
| Decachlorobiphenyl   |     | 9.93       | 9.93      | 9.83      | 10.03 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Continuing Calib Date: 12/02/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 10:01

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.72       | 4.73      | 4.63      | 4.83 | 0.01       |
| Aroclor-1016-2       | (2) | 4.74       | 4.75      | 4.65      | 4.85 | 0.01       |
| Aroclor-1016-3       | (3) | 4.91       | 4.92      | 4.82      | 5.02 | 0.01       |
| Aroclor-1016-4       | (4) | 4.95       | 4.96      | 4.86      | 5.06 | 0.01       |
| Aroclor-1016-5       | (5) | 5.17       | 5.18      | 5.08      | 5.28 | 0.01       |
| Aroclor-1260-1       | (1) | 6.20       | 6.21      | 6.11      | 6.31 | 0.01       |
| Aroclor-1260-2       | (2) | 6.38       | 6.39      | 6.29      | 6.49 | 0.01       |
| Aroclor-1260-3       | (3) | 6.54       | 6.55      | 6.45      | 6.65 | 0.02       |
| Aroclor-1260-4       | (4) | 7.01       | 7.02      | 6.92      | 7.12 | 0.01       |
| Aroclor-1260-5       | (5) | 7.25       | 7.26      | 7.16      | 7.36 | 0.01       |
| Tetrachloro-m-xylene |     | 3.64       | 3.65      | 3.55      | 3.75 | 0.01       |
| Decachlorobiphenyl   |     | 8.63       | 8.64      | 8.54      | 8.74 | 0.01       |