

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Continuing Calib Date: 12/03/2025

Initial Calibration Date(s): 11/04/2025 11/04/2025

Continuing Calib Time: 05:29

Initial Calibration Time(s): 09:39 18:31

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Tetrachloro-m-xylene	4.34	4.35	4.25	4.45	0.01
Decachlorobiphenyl	9.92	9.93	9.83	10.03	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Continuing Calib Date: 12/03/2025

Initial Calibration Date(s): 11/04/2025 11/04/2025

Continuing Calib Time: 05:29

Initial Calibration Time(s): 09:39 18:31

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Tetrachloro-m-xylene	3.64	3.65	3.55	3.75	0.01
Decachlorobiphenyl	8.63	8.64	8.54	8.74	0.01