

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 12/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115506.D **Time Analyzed:** 10:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.483	5.387	5.587	487.950	500.000	-2.4
Aroclor-1016-2	5.505	5.408	5.608	479.350	500.000	-4.1
Aroclor-1016-3	5.567	5.470	5.670	508.910	500.000	1.8
Aroclor-1016-4	5.663	5.566	5.766	495.360	500.000	-0.9
Aroclor-1016-5	5.954	5.858	6.058	480.870	500.000	-3.8
Aroclor-1260-1	7.068	6.971	7.171	473.490	500.000	-5.3
Aroclor-1260-2	7.322	7.225	7.425	513.790	500.000	2.8
Aroclor-1260-3	7.680	7.582	7.782	510.170	500.000	2.0
Aroclor-1260-4	7.902	7.803	8.003	485.740	500.000	-2.9
Aroclor-1260-5	8.208	8.110	8.310	475.810	500.000	-4.8
Decachlorobiphenyl	9.926	9.830	10.030	57.010	50.000	14.0
Tetrachloro-m-xylene	4.344	4.248	4.448	54.810	50.000	9.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 12/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115506.D **Time Analyzed:** 10:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.718	4.628	4.828	439.360	500.000	-12.1
Aroclor-1016-2	4.737	4.647	4.847	447.790	500.000	-10.4
Aroclor-1016-3	4.912	4.821	5.021	440.520	500.000	-11.9
Aroclor-1016-4	4.954	4.864	5.064	435.480	500.000	-12.9
Aroclor-1016-5	5.166	5.076	5.276	465.510	500.000	-6.9
Aroclor-1260-1	6.195	6.106	6.306	409.460	500.000	-18.1
Aroclor-1260-2	6.383	6.294	6.494	382.210	500.000	-23.6
Aroclor-1260-3	6.535	6.446	6.646	401.450	500.000	-19.7
Aroclor-1260-4	7.006	6.917	7.117	405.760	500.000	-18.8
Aroclor-1260-5	7.247	7.158	7.358	408.290	500.000	-18.3
Decachlorobiphenyl	8.628	8.541	8.741	53.010	50.000	6.0
Tetrachloro-m-xylene	3.642	3.548	3.748	50.430	50.000	0.9