

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 12/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115521.D **Time Analyzed:** 18:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.479	5.387	5.587	492.890	500.000	-1.4
Aroclor-1016-2	5.500	5.408	5.608	482.580	500.000	-3.5
Aroclor-1016-3	5.562	5.470	5.670	517.590	500.000	3.5
Aroclor-1016-4	5.658	5.566	5.766	499.780	500.000	0.0
Aroclor-1016-5	5.949	5.858	6.058	486.840	500.000	-2.6
Aroclor-1260-1	7.063	6.971	7.171	472.040	500.000	-5.6
Aroclor-1260-2	7.317	7.225	7.425	509.620	500.000	1.9
Aroclor-1260-3	7.674	7.582	7.782	504.970	500.000	1.0
Aroclor-1260-4	7.896	7.803	8.003	483.140	500.000	-3.4
Aroclor-1260-5	8.203	8.110	8.310	478.530	500.000	-4.3
Decachlorobiphenyl	9.921	9.830	10.030	57.170	50.000	14.3
Tetrachloro-m-xylene	4.340	4.248	4.448	55.570	50.000	11.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 12/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115521.D **Time Analyzed:** 18:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.718	4.628	4.828	457.190	500.000	-8.6
Aroclor-1016-2	4.737	4.647	4.847	457.230	500.000	-8.6
Aroclor-1016-3	4.911	4.821	5.021	449.550	500.000	-10.1
Aroclor-1016-4	4.953	4.864	5.064	444.230	500.000	-11.2
Aroclor-1016-5	5.165	5.076	5.276	477.670	500.000	-4.5
Aroclor-1260-1	6.195	6.106	6.306	410.750	500.000	-17.9
Aroclor-1260-2	6.383	6.294	6.494	385.760	500.000	-22.8
Aroclor-1260-3	6.535	6.446	6.646	401.860	500.000	-19.6
Aroclor-1260-4	7.006	6.917	7.117	402.580	500.000	-19.5
Aroclor-1260-5	7.247	7.158	7.358	404.800	500.000	-19.0
Decachlorobiphenyl	8.627	8.541	8.741	53.010	50.000	6.0
Tetrachloro-m-xylene	3.642	3.548	3.748	51.510	50.000	3.0