

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL04 **Date Analyzed:** 12/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115546.D **Time Analyzed:** 04:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.477	5.387	5.587	494.540	500.000	-1.1
Aroclor-1016-2	5.499	5.408	5.608	479.710	500.000	-4.1
Aroclor-1016-3	5.560	5.470	5.670	523.180	500.000	4.6
Aroclor-1016-4	5.656	5.566	5.766	508.100	500.000	1.6
Aroclor-1016-5	5.948	5.858	6.058	494.930	500.000	-1.0
Aroclor-1260-1	7.061	6.971	7.171	492.050	500.000	-1.6
Aroclor-1260-2	7.315	7.225	7.425	518.550	500.000	3.7
Aroclor-1260-3	7.672	7.582	7.782	503.470	500.000	0.7
Aroclor-1260-4	7.894	7.803	8.003	481.970	500.000	-3.6
Aroclor-1260-5	8.201	8.110	8.310	475.100	500.000	-5.0
Decachlorobiphenyl	9.914	9.830	10.030	56.920	50.000	13.8
Tetrachloro-m-xylene	4.338	4.248	4.448	54.480	50.000	9.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL04 **Date Analyzed:** 12/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115546.D **Time Analyzed:** 04:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.717	4.628	4.828	469.120	500.000	-6.2
Aroclor-1016-2	4.736	4.647	4.847	454.360	500.000	-9.1
Aroclor-1016-3	4.910	4.821	5.021	440.320	500.000	-11.9
Aroclor-1016-4	4.952	4.864	5.064	421.770	500.000	-15.6
Aroclor-1016-5	5.164	5.076	5.276	514.510	500.000	2.9
Aroclor-1260-1	6.194	6.106	6.306	426.390	500.000	-14.7
Aroclor-1260-2	6.381	6.294	6.494	402.930	500.000	-19.4
Aroclor-1260-3	6.534	6.446	6.646	402.420	500.000	-19.5
Aroclor-1260-4	7.005	6.917	7.117	400.910	500.000	-19.8
Aroclor-1260-5	7.245	7.158	7.358	400.030	500.000	-20.0
Decachlorobiphenyl	8.626	8.541	8.741	51.620	50.000	3.2
Tetrachloro-m-xylene	3.642	3.548	3.748	50.880	50.000	1.8