

284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

Report of Analysis

Client: Remington & Vernick
Project: Saddler Property
Client Sample ID: SB-1125-25
Lab Sample ID: Q3742-17

Analytical Method: 8082A Sample Wt/Vol: 30.06 g

Prep Method: SW3541B

Date Collected: 11/25/25
Date Received: 11/26/25
SDG No.: Q3742
Matrix: SOIL
% Solid: 63.3
Test: PCB

Final Vol: 10000 uL Prep Date: 12/02/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	6.20	U	1	6.20	26.8	ug/kg	12/02/25 20:38	PB170775
11104-28-2	Aroclor-1221	6.40	U	1	6.40	26.8	ug/kg	12/02/25 20:38	PB170775
11141-16-5	Aroclor-1232	5.90	U	1	5.90	26.8	ug/kg	12/02/25 20:38	PB170775
53469-21-9	Aroclor-1242	6.30	U	1	6.30	26.8	ug/kg	12/02/25 20:38	PB170775
12672-29-6	Aroclor-1248	9.30	U	1	9.30	26.8	ug/kg	12/02/25 20:38	PB170775
11097-69-1	Aroclor-1254	5.10	U	1	5.10	26.8	ug/kg	12/02/25 20:38	PB170775
37324-23-5	Aroclor-1262	7.90	U	1	7.90	26.8	ug/kg	12/02/25 20:38	PB170775
11100-14-4	Aroclor-1268	5.70	U	1	5.70	26.8	ug/kg	12/02/25 20:38	PB170775
11096-82-5	Aroclor-1260	5.10	U	1	5.10	26.8	ug/kg	12/02/25 20:38	PB170775
SURROGATES									
877-09-8	Tetrachloro-m-xylene	18.5			21 - 165	92%	SPK: 20		
2051-24-3	Decachlorobiphenyl	17.6			10 - 170	88%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit