

Surrogate Summary

SDG No.: **Q3742**

Client: **Remington & Vernick**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO114870.D	PIBLK-PO114870.D	Tetrachloro-m-xyl	1	20	15.5	77		60	140
		Decachlorobiphen	1	20	16.8	84		60	140
		Tetrachloro-m-xyl	2	20	16.2	81		60	140
		Decachlorobiphen	2	20	17.3	86		60	140
I.BLK-PO115510.D	PIBLK-PO115510.D	Tetrachloro-m-xyl	1	20	26.3	131		60	140
		Decachlorobiphen	1	20	28.1	140		60	140
		Tetrachloro-m-xyl	2	20	24.6	123		60	140
		Decachlorobiphen	2	20	26.5	133		60	140
Q3742-01	SB-1125-23	Tetrachloro-m-xyl	1	20	22.5	112		21	165
		Decachlorobiphen	1	20	20.4	102		10	170
		Tetrachloro-m-xyl	2	20	21.0	105		21	165
		Decachlorobiphen	2	20	19.6	98		10	170
Q3742-02	SB-1125-19	Tetrachloro-m-xyl	1	20	22.7	113		21	165
		Decachlorobiphen	1	20	21.0	105		10	170
		Tetrachloro-m-xyl	2	20	22.0	110		21	165
		Decachlorobiphen	2	20	19.8	99		10	170
Q3742-03	SB-1125-8	Tetrachloro-m-xyl	1	20	21.8	109		21	165
		Decachlorobiphen	1	20	20.5	102		10	170
		Tetrachloro-m-xyl	2	20	21.1	105		21	165
		Decachlorobiphen	2	20	19.7	98		10	170
Q3742-04	SB-1125-9	Tetrachloro-m-xyl	1	20	24.4	122		21	165
		Decachlorobiphen	1	20	24.9	124		10	170
		Tetrachloro-m-xyl	2	20	23.7	119		21	165
		Decachlorobiphen	2	20	23.7	118		10	170
Q3742-05	SB-1125-21	Tetrachloro-m-xyl	1	20	25.6	128		21	165
		Decachlorobiphen	1	20	21.7	109		10	170
		Tetrachloro-m-xyl	2	20	24.9	125		21	165
		Decachlorobiphen	2	20	20.7	104		10	170
Q3742-06	SB-1125-13	Tetrachloro-m-xyl	1	20	20.7	104		21	165
		Decachlorobiphen	1	20	21.4	107		10	170
		Tetrachloro-m-xyl	2	20	20.2	101		21	165
		Decachlorobiphen	2	20	20.0	100		10	170
Q3742-07	SB-1125-18	Tetrachloro-m-xyl	1	20	20.4	102		21	165
		Decachlorobiphen	1	20	23.2	116		10	170
		Tetrachloro-m-xyl	2	20	19.9	100		21	165
		Decachlorobiphen	2	20	21.2	106		10	170
Q3742-07MS	SB-1125-18MS	Tetrachloro-m-xyl	1	20	19.9	100		21	165
		Decachlorobiphen	1	20	20.5	102		10	170
		Tetrachloro-m-xyl	2	20	18.6	93		21	165
		Decachlorobiphen	2	20	18.7	94		10	170
Q3742-07MSD	SB-1125-18MSD	Tetrachloro-m-xyl	1	20	18.8	94		21	165

Surrogate Summary

SDG No.: Q3742

Client: Remington & Vernick

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q3742-07MSD	SB-1125-18MSD	Decachlorobiphen	1	20	19.9	99		10	170
		Tetrachloro-m-xyl	2	20	18.6	93		21	165
		Decachlorobiphen	2	20	18.4	92		10	170
Q3742-08	SB-1125-20	Tetrachloro-m-xyl	1	20	22.1	111		21	165
		Decachlorobiphen	1	20	22.1	110		10	170
		Tetrachloro-m-xyl	2	20	21.6	108		21	165
		Decachlorobiphen	2	20	20.9	104		10	170
I.BLK-PO115525.D	PIBLK-PO115525.D	Tetrachloro-m-xyl	1	20	25.0	125		60	140
		Decachlorobiphen	1	20	26.0	130		60	140
		Tetrachloro-m-xyl	2	20	23.7	118		60	140
		Decachlorobiphen	2	20	24.6	123		60	140
Q3742-09	SB-1125-22	Tetrachloro-m-xyl	1	20	22.3	111		21	165
		Decachlorobiphen	1	20	21.8	109		10	170
		Tetrachloro-m-xyl	2	20	21.7	108		21	165
		Decachlorobiphen	2	20	20.7	104		10	170
Q3742-10	SB-1125-2	Tetrachloro-m-xyl	1	20	23.1	115		21	165
		Decachlorobiphen	1	20	20.8	104		10	170
		Tetrachloro-m-xyl	2	20	22.5	113		21	165
		Decachlorobiphen	2	20	19.7	98		10	170
Q3742-16	SB-1125-1	Tetrachloro-m-xyl	1	20	24.4	122		21	165
		Decachlorobiphen	1	20	24.8	124		10	170
		Tetrachloro-m-xyl	2	20	23.7	119		21	165
		Decachlorobiphen	2	20	23.3	116		10	170
Q3742-17	SB-1125-25	Tetrachloro-m-xyl	1	20	18.5	92		21	165
		Decachlorobiphen	1	20	17.6	88		10	170
		Tetrachloro-m-xyl	2	20	18.3	92		21	165
		Decachlorobiphen	2	20	17.0	85		10	170
Q3742-18	SB-1125-24	Tetrachloro-m-xyl	1	20	20.8	104		21	165
		Decachlorobiphen	1	20	18.8	94		10	170
		Tetrachloro-m-xyl	2	20	20.6	103		21	165
		Decachlorobiphen	2	20	18.1	90		10	170
I.BLK-PO115540.D	PIBLK-PO115540.D	Tetrachloro-m-xyl	1	20	25.1	126		60	140
		Decachlorobiphen	1	20	25.5	127		60	140
		Tetrachloro-m-xyl	2	20	23.1	116		60	140
		Decachlorobiphen	2	20	24.0	120		60	140
PB170775BL	PB170775BL	Tetrachloro-m-xyl	1	20	26.7	134		21	165
		Decachlorobiphen	1	20	27.2	136		10	170
		Tetrachloro-m-xyl	2	20	25.0	125		21	165
		Decachlorobiphen	2	20	25.5	127		10	170
PB170775BS	PB170775BS	Tetrachloro-m-xyl	1	20	26.4	132		21	165
		Decachlorobiphen	1	20	26.8	134		10	170

Surrogate Summary

SDG No.: Q3742

Client: Remington & Vernick

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
PB170775BS	PB170775BS	Tetrachloro-m-xyl	2	20	23.7	118		21	165
		Decachlorobiphen	2	20	25.3	127		10	170
I.BLK-PO115550.D	PIBLK-PO115550.D	Tetrachloro-m-xyl	1	20	24.7	123		60	140
		Decachlorobiphen	1	20	25.6	128		60	140
		Tetrachloro-m-xyl	2	20	23.6	118		60	140
		Decachlorobiphen	2	20	23.7	119		60	140