

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL02 **Date Analyzed:** 12/02/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091426.D **Time Analyzed:** 16:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.698	6.599	6.799	44.440	50.000	-11.1
4,4'-DDE	6.187	6.089	6.289	45.000	50.000	-10.0
4,4'-DDT	7.014	6.914	7.114	45.890	50.000	-8.2
Aldrin	5.261	5.163	5.363	46.020	50.000	-8.0
alpha-BHC	3.991	3.893	4.093	47.590	50.000	-4.8
alpha-Chlordane	6.017	5.919	6.119	43.170	50.000	-13.7
beta-BHC	4.509	4.411	4.611	45.680	50.000	-8.6
Decachlorobiphenyl	9.064	8.963	9.163	40.930	50.000	-18.1
delta-BHC	4.758	4.660	4.860	47.180	50.000	-5.6
Dieldrin	6.338	6.239	6.439	45.510	50.000	-9.0
Endosulfan I	6.065	5.966	6.166	44.980	50.000	-10.0
Endosulfan II	6.778	6.679	6.879	44.780	50.000	-10.4
Endosulfan sulfate	7.142	7.043	7.243	43.570	50.000	-12.9
Endrin	6.565	6.466	6.666	47.890	50.000	-4.2
Endrin aldehyde	6.907	6.808	7.008	42.870	50.000	-14.3
Endrin ketone	7.623	7.524	7.724	43.740	50.000	-12.5
gamma-BHC (Lindane)	4.321	4.224	4.424	47.230	50.000	-5.5
gamma-Chlordane	5.937	5.838	6.038	46.000	50.000	-8.0
Heptachlor	4.919	4.821	5.021	45.680	50.000	-8.6
Heptachlor epoxide	5.681	5.582	5.782	45.200	50.000	-9.6
Methoxychlor	7.487	7.387	7.587	43.190	50.000	-13.6
Tetrachloro-m-xylene	3.541	3.444	3.644	46.040	50.000	-7.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL02 **Date Analyzed:** 12/02/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091426.D **Time Analyzed:** 16:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.916	5.817	6.017	50.280	50.000	0.6
4,4'-DDE	5.360	5.263	5.463	52.360	50.000	4.7
4,4'-DDT	6.170	6.071	6.271	53.700	50.000	7.4
Aldrin	4.356	4.258	4.458	53.520	50.000	7.0
alpha-BHC	3.383	3.286	3.486	53.430	50.000	6.9
alpha-Chlordane	5.178	5.079	5.279	52.980	50.000	6.0
beta-BHC	4.016	3.918	4.118	52.290	50.000	4.6
Decachlorobiphenyl	8.058	7.958	8.158	44.240	50.000	-11.5
delta-BHC	4.252	4.153	4.353	53.010	50.000	6.0
Dieldrin	5.500	5.401	5.601	52.390	50.000	4.8
Endosulfan I	5.234	5.135	5.335	51.440	50.000	2.9
Endosulfan II	6.068	5.969	6.169	51.310	50.000	2.6
Endosulfan sulfate	6.470	6.371	6.571	50.390	50.000	0.8
Endrin	5.776	5.678	5.878	55.300	50.000	10.6
Endrin aldehyde	6.246	6.148	6.348	49.180	50.000	-1.6
Endrin ketone	6.979	6.880	7.080	48.270	50.000	-3.5
gamma-BHC (Lindane)	3.719	3.621	3.821	53.250	50.000	6.5
gamma-Chlordane	5.113	5.014	5.214	53.280	50.000	6.6
Heptachlor	4.071	3.973	4.173	52.430	50.000	4.9
Heptachlor epoxide	4.860	4.762	4.962	52.480	50.000	5.0
Methoxychlor	6.741	6.641	6.841	50.390	50.000	0.8
Tetrachloro-m-xylene	2.872	2.774	2.974	52.510	50.000	5.0