

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL03 **Date Analyzed:** 12/02/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091434.D **Time Analyzed:** 19:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.698	6.599	6.799	49.770	50.000	-0.5
4,4'-DDE	6.188	6.089	6.289	50.340	50.000	0.7
4,4'-DDT	7.014	6.914	7.114	51.070	50.000	2.1
Aldrin	5.261	5.163	5.363	50.920	50.000	1.8
alpha-BHC	3.991	3.893	4.093	51.990	50.000	4.0
alpha-Chlordane	6.018	5.919	6.119	50.030	50.000	0.1
beta-BHC	4.510	4.411	4.611	49.750	50.000	-0.5
Decachlorobiphenyl	9.063	8.963	9.163	45.580	50.000	-8.8
delta-BHC	4.758	4.660	4.860	52.450	50.000	4.9
Dieldrin	6.338	6.239	6.439	50.690	50.000	1.4
Endosulfan I	6.065	5.966	6.166	50.060	50.000	0.1
Endosulfan II	6.779	6.679	6.879	48.490	50.000	-3.0
Endosulfan sulfate	7.142	7.043	7.243	48.880	50.000	-2.2
Endrin	6.566	6.466	6.666	53.130	50.000	6.3
Endrin aldehyde	6.908	6.808	7.008	47.960	50.000	-4.1
Endrin ketone	7.623	7.524	7.724	49.330	50.000	-1.3
gamma-BHC (Lindane)	4.322	4.224	4.424	51.710	50.000	3.4
gamma-Chlordane	5.937	5.838	6.038	50.740	50.000	1.5
Heptachlor	4.919	4.821	5.021	50.210	50.000	0.4
Heptachlor epoxide	5.682	5.582	5.782	51.340	50.000	2.7
Methoxychlor	7.487	7.387	7.587	48.330	50.000	-3.3
Tetrachloro-m-xylene	3.542	3.444	3.644	49.450	50.000	-1.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL03 **Date Analyzed:** 12/02/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091434.D **Time Analyzed:** 19:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.916	5.817	6.017	53.790	50.000	7.6
4,4'-DDE	5.362	5.263	5.463	56.070	50.000	12.1
4,4'-DDT	6.168	6.071	6.271	56.860	50.000	13.7
Aldrin	4.356	4.258	4.458	56.540	50.000	13.1
alpha-BHC	3.383	3.286	3.486	56.260	50.000	12.5
alpha-Chlordane	5.177	5.079	5.279	56.340	50.000	12.7
beta-BHC	4.016	3.918	4.118	55.090	50.000	10.2
Decachlorobiphenyl	8.058	7.958	8.158	51.040	50.000	2.1
delta-BHC	4.252	4.153	4.353	55.990	50.000	12.0
Dieldrin	5.499	5.401	5.601	55.940	50.000	11.9
Endosulfan I	5.234	5.135	5.335	54.820	50.000	9.6
Endosulfan II	6.068	5.969	6.169	56.730	50.000	13.5
Endosulfan sulfate	6.469	6.371	6.571	50.500	50.000	1.0
Endrin	5.775	5.678	5.878	59.630	50.000	19.3
Endrin aldehyde	6.246	6.148	6.348	54.610	50.000	9.2
Endrin ketone	6.979	6.880	7.080	53.580	50.000	7.2
gamma-BHC (Lindane)	3.719	3.621	3.821	56.180	50.000	12.4
gamma-Chlordane	5.112	5.014	5.214	57.000	50.000	14.0
Heptachlor	4.071	3.973	4.173	54.500	50.000	9.0
Heptachlor epoxide	4.860	4.762	4.962	55.850	50.000	11.7
Methoxychlor	6.740	6.641	6.841	54.300	50.000	8.6
Tetrachloro-m-xylene	2.873	2.774	2.974	55.230	50.000	10.5