

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL04 **Date Analyzed:** 12/02/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091446.D **Time Analyzed:** 21:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.698	6.599	6.799	46.020	50.000	-8.0
4,4'-DDE	6.188	6.089	6.289	46.980	50.000	-6.0
4,4'-DDT	7.015	6.914	7.114	46.950	50.000	-6.1
Aldrin	5.262	5.163	5.363	47.590	50.000	-4.8
alpha-BHC	3.992	3.893	4.093	48.690	50.000	-2.6
alpha-Chlordane	6.018	5.919	6.119	45.560	50.000	-8.9
beta-BHC	4.510	4.411	4.611	46.800	50.000	-6.4
Decachlorobiphenyl	9.064	8.963	9.163	42.580	50.000	-14.8
delta-BHC	4.758	4.660	4.860	48.830	50.000	-2.3
Dieldrin	6.339	6.239	6.439	47.410	50.000	-5.2
Endosulfan I	6.066	5.966	6.166	46.810	50.000	-6.4
Endosulfan II	6.779	6.679	6.879	46.530	50.000	-6.9
Endosulfan sulfate	7.144	7.043	7.243	45.500	50.000	-9.0
Endrin	6.566	6.466	6.666	49.270	50.000	-1.5
Endrin aldehyde	6.909	6.808	7.008	45.000	50.000	-10.0
Endrin ketone	7.622	7.524	7.724	44.870	50.000	-10.3
gamma-BHC (Lindane)	4.323	4.224	4.424	48.530	50.000	-2.9
gamma-Chlordane	5.936	5.838	6.038	47.120	50.000	-5.8
Heptachlor	4.920	4.821	5.021	46.270	50.000	-7.5
Heptachlor epoxide	5.682	5.582	5.782	47.110	50.000	-5.8
Methoxychlor	7.488	7.387	7.587	43.900	50.000	-12.2
Tetrachloro-m-xylene	3.543	3.444	3.644	46.630	50.000	-6.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL04 **Date Analyzed:** 12/02/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091446.D **Time Analyzed:** 21:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.917	5.817	6.017	49.780	50.000	-0.4
4,4'-DDE	5.361	5.263	5.463	51.790	50.000	3.6
4,4'-DDT	6.171	6.071	6.271	53.360	50.000	6.7
Aldrin	4.357	4.258	4.458	52.900	50.000	5.8
alpha-BHC	3.384	3.286	3.486	52.690	50.000	5.4
alpha-Chlordane	5.178	5.079	5.279	52.490	50.000	5.0
beta-BHC	4.016	3.918	4.118	51.790	50.000	3.6
Decachlorobiphenyl	8.058	7.958	8.158	46.070	50.000	-7.9
delta-BHC	4.252	4.153	4.353	52.440	50.000	4.9
Dieldrin	5.500	5.401	5.601	52.230	50.000	4.5
Endosulfan I	5.235	5.135	5.335	51.090	50.000	2.2
Endosulfan II	6.069	5.969	6.169	51.690	50.000	3.4
Endosulfan sulfate	6.470	6.371	6.571	49.120	50.000	-1.8
Endrin	5.778	5.678	5.878	55.300	50.000	10.6
Endrin aldehyde	6.247	6.148	6.348	50.150	50.000	0.3
Endrin ketone	6.980	6.880	7.080	49.530	50.000	-0.9
gamma-BHC (Lindane)	3.720	3.621	3.821	52.570	50.000	5.1
gamma-Chlordane	5.113	5.014	5.214	53.070	50.000	6.1
Heptachlor	4.072	3.973	4.173	50.410	50.000	0.8
Heptachlor epoxide	4.861	4.762	4.962	52.120	50.000	4.2
Methoxychlor	6.741	6.641	6.841	49.760	50.000	-0.5
Tetrachloro-m-xylene	2.873	2.774	2.974	51.650	50.000	3.3