

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/08/2025 12/08/2025

Client Sample No.: CCAL05 **Date Analyzed:** 12/09/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091525.D **Time Analyzed:** 14:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.700	6.595	6.795	51.400	50.000	2.8
4,4'-DDE	6.190	6.085	6.285	47.060	50.000	-5.9
4,4'-DDT	7.015	6.911	7.111	46.830	50.000	-6.3
Aldrin	5.263	5.158	5.358	49.330	50.000	-1.3
alpha-BHC	3.994	3.889	4.089	47.500	50.000	-5.0
alpha-Chlordane	6.020	5.915	6.115	46.360	50.000	-7.3
beta-BHC	4.511	4.408	4.608	46.230	50.000	-7.5
Decachlorobiphenyl	9.064	8.957	9.157	45.660	50.000	-8.7
delta-BHC	4.760	4.656	4.856	56.240	50.000	12.5
Dieldrin	6.340	6.235	6.435	47.720	50.000	-4.6
Endosulfan I	6.067	5.963	6.163	47.490	50.000	-5.0
Endosulfan II	6.779	6.676	6.876	42.290	50.000	-15.4
Endosulfan sulfate	7.144	7.040	7.240	46.370	50.000	-7.3
Endrin	6.567	6.463	6.663	48.250	50.000	-3.5
Endrin aldehyde	6.910	6.805	7.005	48.440	50.000	-3.1
Endrin ketone	7.625	7.521	7.721	46.690	50.000	-6.6
gamma-BHC (Lindane)	4.324	4.220	4.420	47.610	50.000	-4.8
gamma-Chlordane	5.940	5.834	6.034	48.440	50.000	-3.1
Heptachlor	4.922	4.817	5.017	48.830	50.000	-2.3
Heptachlor epoxide	5.684	5.579	5.779	48.240	50.000	-3.5
Methoxychlor	7.489	7.384	7.584	45.370	50.000	-9.3
Tetrachloro-m-xylene	3.544	3.440	3.640	45.490	50.000	-9.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/08/2025 12/08/2025

Client Sample No.: CCAL05 **Date Analyzed:** 12/09/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091525.D **Time Analyzed:** 14:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.916	5.814	6.014	52.190	50.000	4.4
4,4'-DDE	5.360	5.260	5.460	51.410	50.000	2.8
4,4'-DDT	6.170	6.068	6.268	52.560	50.000	5.1
Aldrin	4.353	4.254	4.454	51.550	50.000	3.1
alpha-BHC	3.381	3.281	3.481	49.490	50.000	-1.0
alpha-Chlordane	5.176	5.076	5.276	52.250	50.000	4.5
beta-BHC	4.015	3.914	4.114	50.940	50.000	1.9
Decachlorobiphenyl	8.056	7.955	8.155	51.800	50.000	3.6
delta-BHC	4.250	4.150	4.350	51.690	50.000	3.4
Dieldrin	5.499	5.398	5.598	51.820	50.000	3.6
Endosulfan I	5.233	5.132	5.332	43.540	50.000	-12.9
Endosulfan II	6.067	5.966	6.166	51.720	50.000	3.4
Endosulfan sulfate	6.469	6.367	6.567	51.500	50.000	3.0
Endrin	5.776	5.674	5.874	51.890	50.000	3.8
Endrin aldehyde	6.246	6.144	6.344	51.210	50.000	2.4
Endrin ketone	6.978	6.876	7.076	51.600	50.000	3.2
gamma-BHC (Lindane)	3.718	3.617	3.817	54.430	50.000	8.9
gamma-Chlordane	5.112	5.011	5.211	52.990	50.000	6.0
Heptachlor	4.070	3.969	4.169	51.840	50.000	3.7
Heptachlor epoxide	4.859	4.758	4.958	52.390	50.000	4.8
Methoxychlor	6.740	6.638	6.838	52.010	50.000	4.0
Tetrachloro-m-xylene	2.870	2.770	2.970	48.630	50.000	-2.7