

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/08/2025 12/08/2025

Client Sample No.: CCAL06 **Date Analyzed:** 12/09/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091529.D **Time Analyzed:** 16:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.696	6.595	6.795	51.230	50.000	2.5
4,4'-DDE	6.186	6.085	6.285	47.770	50.000	-4.5
4,4'-DDT	7.012	6.911	7.111	47.240	50.000	-5.5
Aldrin	5.259	5.158	5.358	50.220	50.000	0.4
alpha-BHC	3.989	3.889	4.089	49.100	50.000	-1.8
alpha-Chlordane	6.016	5.915	6.115	47.430	50.000	-5.1
beta-BHC	4.508	4.408	4.608	48.170	50.000	-3.7
Decachlorobiphenyl	9.061	8.957	9.157	45.860	50.000	-8.3
delta-BHC	4.756	4.656	4.856	60.750	50.000	21.5
Dieldrin	6.337	6.235	6.435	48.850	50.000	-2.3
Endosulfan I	6.063	5.963	6.163	48.600	50.000	-2.8
Endosulfan II	6.777	6.676	6.876	44.880	50.000	-10.2
Endosulfan sulfate	7.141	7.040	7.240	47.300	50.000	-5.4
Endrin	6.562	6.463	6.663	48.120	50.000	-3.8
Endrin aldehyde	6.906	6.805	7.005	48.680	50.000	-2.6
Endrin ketone	7.620	7.521	7.721	47.350	50.000	-5.3
gamma-BHC (Lindane)	4.320	4.220	4.420	49.110	50.000	-1.8
gamma-Chlordane	5.935	5.834	6.034	49.640	50.000	-0.7
Heptachlor	4.918	4.817	5.017	50.410	50.000	0.8
Heptachlor epoxide	5.680	5.579	5.779	51.720	50.000	3.4
Methoxychlor	7.485	7.384	7.584	46.860	50.000	-6.3
Tetrachloro-m-xylene	3.540	3.440	3.640	47.650	50.000	-4.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/08/2025 12/08/2025

Client Sample No.: CCAL06 **Date Analyzed:** 12/09/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091529.D **Time Analyzed:** 16:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.914	5.814	6.014	54.610	50.000	9.2
4,4'-DDE	5.358	5.260	5.460	54.280	50.000	8.6
4,4'-DDT	6.168	6.068	6.268	54.610	50.000	9.2
Aldrin	4.352	4.254	4.454	54.070	50.000	8.1
alpha-BHC	3.380	3.281	3.481	51.860	50.000	3.7
alpha-Chlordane	5.175	5.076	5.276	54.850	50.000	9.7
beta-BHC	4.013	3.914	4.114	53.330	50.000	6.7
Decachlorobiphenyl	8.055	7.955	8.155	54.480	50.000	9.0
delta-BHC	4.249	4.150	4.350	54.130	50.000	8.3
Dieldrin	5.498	5.398	5.598	54.280	50.000	8.6
Endosulfan I	5.231	5.132	5.332	45.620	50.000	-8.8
Endosulfan II	6.066	5.966	6.166	54.310	50.000	8.6
Endosulfan sulfate	6.468	6.367	6.567	53.890	50.000	7.8
Endrin	5.773	5.674	5.874	54.340	50.000	8.7
Endrin aldehyde	6.244	6.144	6.344	53.430	50.000	6.9
Endrin ketone	6.976	6.876	7.076	54.470	50.000	8.9
gamma-BHC (Lindane)	3.716	3.617	3.817	57.430	50.000	14.9
gamma-Chlordane	5.110	5.011	5.211	55.610	50.000	11.2
Heptachlor	4.069	3.969	4.169	54.030	50.000	8.1
Heptachlor epoxide	4.857	4.758	4.958	54.840	50.000	9.7
Methoxychlor	6.739	6.638	6.838	53.850	50.000	7.7
Tetrachloro-m-xylene	2.869	2.770	2.970	50.750	50.000	1.5