

Hit Summary Sheet
SW-846

SDG No.: Q3742

Order ID: Q3742

Client: Remington & Vernick

Project ID: Saddler Property

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : SB-1125-21								
Q3742-05	SB-1125-21	SOIL	Dieldrin	0.93	JP	0.18	2.10	ug/kg
Q3742-05	SB-1125-21	SOIL	4,4-DDE	1.80	JP	0.18	2.10	ug/kg
Q3742-05	SB-1125-21	SOIL	Endrin	0.84	JP	0.18	2.10	ug/kg
Q3742-05	SB-1125-21	SOIL	4,4-DDT	4.10		0.18	2.10	ug/kg
Total Concentration:				7.670				
Client ID : SB-1125-22								
Q3742-09	SB-1125-22	SOIL	4,4-DDE	0.48	JP	0.19	2.30	ug/kg
Total Concentration:				0.480				
Client ID : SB-1125-1								
Q3742-16	SB-1125-1	SOIL	gamma-BHC (Lindane)	22.0		0.19	2.30	ug/kg
Q3742-16	SB-1125-1	SOIL	Heptachlor	0.94	J	0.16	2.30	ug/kg
Q3742-16	SB-1125-1	SOIL	Heptachlor epoxide	9.40		0.26	2.30	ug/kg
Q3742-16	SB-1125-1	SOIL	Dieldrin	0.46	J	0.19	2.30	ug/kg
Q3742-16	SB-1125-1	SOIL	alpha-Chlordane	23.6		0.16	2.30	ug/kg
Q3742-16	SB-1125-1	SOIL	gamma-Chlordane	30.9		0.20	2.30	ug/kg
Total Concentration:				87.300				



SAMPLE DATA

Report of Analysis

Client: Remington & Vernick	Date Collected: 11/26/25	
Project: Saddler Property	Date Received: 11/26/25	
Client Sample ID: SB-1125-23	SDG No.: Q3742	
Lab Sample ID: Q3742-01	Matrix: SOIL	
Analytical Method: 8081B	% Solid: 70.1	
Sample Wt/Vol: 30.06 g	Test: Pesticide-TCL	Final Vol: 10000 uL
Prep Method: SW3541B		Prep Date: 12/02/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.19	U	1	0.19	2.40	ug/kg	12/02/25 17:23	PB170776
319-85-7	beta-BHC	0.26	U	1	0.26	2.40	ug/kg	12/02/25 17:23	PB170776
319-86-8	delta-BHC	0.56	U	1	0.56	2.40	ug/kg	12/02/25 17:23	PB170776
58-89-9	gamma-BHC (Lindane)	0.20	U	1	0.20	2.40	ug/kg	12/02/25 17:23	PB170776
76-44-8	Heptachlor	0.17	U	1	0.17	2.40	ug/kg	12/02/25 17:23	PB170776
309-00-2	Aldrin	0.17	U	1	0.17	2.40	ug/kg	12/02/25 17:23	PB170776
1024-57-3	Heptachlor epoxide	0.27	U	1	0.27	2.40	ug/kg	12/02/25 17:23	PB170776
959-98-8	Endosulfan I	0.20	U	1	0.20	2.40	ug/kg	12/02/25 17:23	PB170776
60-57-1	Dieldrin	0.20	U	1	0.20	2.40	ug/kg	12/02/25 17:23	PB170776
72-55-9	4,4-DDE	0.20	U	1	0.20	2.40	ug/kg	12/02/25 17:23	PB170776
72-20-8	Endrin	0.20	U	1	0.20	2.40	ug/kg	12/02/25 17:23	PB170776
33213-65-9	Endosulfan II	0.41	U	1	0.41	2.40	ug/kg	12/02/25 17:23	PB170776
72-54-8	4,4-DDD	0.21	U	1	0.21	2.40	ug/kg	12/02/25 17:23	PB170776
1031-07-8	Endosulfan Sulfate	0.19	U	1	0.19	2.40	ug/kg	12/02/25 17:23	PB170776
50-29-3	4,4-DDT	0.20	U	1	0.20	2.40	ug/kg	12/02/25 17:23	PB170776
72-43-5	Methoxychlor	0.53	U	1	0.53	2.40	ug/kg	12/02/25 17:23	PB170776
53494-70-5	Endrin ketone	0.27	U	1	0.27	2.40	ug/kg	12/02/25 17:23	PB170776
7421-93-4	Endrin aldehyde	0.53	U	1	0.53	2.40	ug/kg	12/02/25 17:23	PB170776
5103-71-9	alpha-Chlordane	0.17	U	1	0.17	2.40	ug/kg	12/02/25 17:23	PB170776
5103-74-2	gamma-Chlordane	0.21	U	1	0.21	2.40	ug/kg	12/02/25 17:23	PB170776
8001-35-2	Toxaphene	7.70	U	1	7.70	47.0	ug/kg	12/02/25 17:23	PB170776
SURROGATES									
2051-24-3	Decachlorobiphenyl	16.5			10 - 143	82%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	21.1			10 - 131	105%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick		Date Collected:	11/25/25
Project:	Saddler Property		Date Received:	11/26/25
Client Sample ID:	SB-1125-19		SDG No.:	Q3742
Lab Sample ID:	Q3742-02		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	40.9
Sample Wt/Vol:	30.02 g	Final Vol: 10000 uL	Test:	Pesticide-TCL
Prep Method:	SW3541B	Prep Date: 12/02/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.32	U	1	0.32	4.20	ug/kg	12/02/25 17:37	PB170776
319-85-7	beta-BHC	0.44	U	1	0.44	4.20	ug/kg	12/02/25 17:37	PB170776
319-86-8	delta-BHC	0.95	U	1	0.95	4.20	ug/kg	12/02/25 17:37	PB170776
58-89-9	gamma-BHC (Lindane)	0.34	U	1	0.34	4.20	ug/kg	12/02/25 17:37	PB170776
76-44-8	Heptachlor	0.29	U	1	0.29	4.20	ug/kg	12/02/25 17:37	PB170776
309-00-2	Aldrin	0.29	U	1	0.29	4.20	ug/kg	12/02/25 17:37	PB170776
1024-57-3	Heptachlor epoxide	0.46	U	1	0.46	4.20	ug/kg	12/02/25 17:37	PB170776
959-98-8	Endosulfan I	0.34	U	1	0.34	4.20	ug/kg	12/02/25 17:37	PB170776
60-57-1	Dieldrin	0.34	U	1	0.34	4.20	ug/kg	12/02/25 17:37	PB170776
72-55-9	4,4-DDE	0.34	U	1	0.34	4.20	ug/kg	12/02/25 17:37	PB170776
72-20-8	Endrin	0.34	U	1	0.34	4.20	ug/kg	12/02/25 17:37	PB170776
33213-65-9	Endosulfan II	0.71	U	1	0.71	4.20	ug/kg	12/02/25 17:37	PB170776
72-54-8	4,4-DDD	0.37	U	1	0.37	4.20	ug/kg	12/02/25 17:37	PB170776
1031-07-8	Endosulfan Sulfate	0.32	U	1	0.32	4.20	ug/kg	12/02/25 17:37	PB170776
50-29-3	4,4-DDT	0.34	U	1	0.34	4.20	ug/kg	12/02/25 17:37	PB170776
72-43-5	Methoxychlor	0.90	U	1	0.90	4.20	ug/kg	12/02/25 17:37	PB170776
53494-70-5	Endrin ketone	0.46	U	1	0.46	4.20	ug/kg	12/02/25 17:37	PB170776
7421-93-4	Endrin aldehyde	0.90	U	1	0.90	4.20	ug/kg	12/02/25 17:37	PB170776
5103-71-9	alpha-Chlordane	0.29	U	1	0.29	4.20	ug/kg	12/02/25 17:37	PB170776
5103-74-2	gamma-Chlordane	0.37	U	1	0.37	4.20	ug/kg	12/02/25 17:37	PB170776
8001-35-2	Toxaphene	13.2	U	1	13.2	80.6	ug/kg	12/02/25 17:37	PB170776
SURROGATES									
2051-24-3	Decachlorobiphenyl	16.7			10 - 143	83%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	21.9			10 - 131	109%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick	Date Collected:	11/25/25
Project:	Saddler Property	Date Received:	11/26/25
Client Sample ID:	SB-1125-21	SDG No.:	Q3742
Lab Sample ID:	Q3742-05	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	79.5
Sample Wt/Vol:	30.08 g	Final Vol:	10000 uL
Prep Method:	SW3541B	Test:	Pesticide-TCL
	Prep Date:	12/02/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.16	U	1	0.16	2.10	ug/kg	12/02/25 18:17	PB170776
319-85-7	beta-BHC	0.23	U	1	0.23	2.10	ug/kg	12/02/25 18:17	PB170776
319-86-8	delta-BHC	0.49	U	1	0.49	2.10	ug/kg	12/02/25 18:17	PB170776
58-89-9	gamma-BHC (Lindane)	0.18	U	1	0.18	2.10	ug/kg	12/02/25 18:17	PB170776
76-44-8	Heptachlor	0.15	U	1	0.15	2.10	ug/kg	12/02/25 18:17	PB170776
309-00-2	Aldrin	0.15	U	1	0.15	2.10	ug/kg	12/02/25 18:17	PB170776
1024-57-3	Heptachlor epoxide	0.24	U	1	0.24	2.10	ug/kg	12/02/25 18:17	PB170776
959-98-8	Endosulfan I	0.18	U	1	0.18	2.10	ug/kg	12/02/25 18:17	PB170776
60-57-1	Dieldrin	0.93	JP	1	0.18	2.10	ug/kg	12/02/25 18:17	PB170776
72-55-9	4,4-DDE	1.80	JP	1	0.18	2.10	ug/kg	12/02/25 18:17	PB170776
72-20-8	Endrin	0.84	JP	1	0.18	2.10	ug/kg	12/02/25 18:17	PB170776
33213-65-9	Endosulfan II	0.36	U	1	0.36	2.10	ug/kg	12/02/25 18:17	PB170776
72-54-8	4,4-DDD	0.19	U	1	0.19	2.10	ug/kg	12/02/25 18:17	PB170776
1031-07-8	Endosulfan Sulfate	0.16	U	1	0.16	2.10	ug/kg	12/02/25 18:17	PB170776
50-29-3	4,4-DDT	4.10		1	0.18	2.10	ug/kg	12/02/25 18:17	PB170776
72-43-5	Methoxychlor	0.46	U	1	0.46	2.10	ug/kg	12/02/25 18:17	PB170776
53494-70-5	Endrin ketone	0.24	U	1	0.24	2.10	ug/kg	12/02/25 18:17	PB170776
7421-93-4	Endrin aldehyde	0.46	U	1	0.46	2.10	ug/kg	12/02/25 18:17	PB170776
5103-71-9	alpha-Chlordane	0.15	U	1	0.15	2.10	ug/kg	12/02/25 18:17	PB170776
5103-74-2	gamma-Chlordane	0.19	U	1	0.19	2.10	ug/kg	12/02/25 18:17	PB170776
8001-35-2	Toxaphene	6.80	U	1	6.80	41.4	ug/kg	12/02/25 18:17	PB170776
SURROGATES									
2051-24-3	Decachlorobiphenyl	17.8			10 - 143	89%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	23.0			10 - 131	115%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick	Date Collected:	11/25/25
Project:	Saddler Property	Date Received:	11/26/25
Client Sample ID:	SB-1125-13	SDG No.:	Q3742
Lab Sample ID:	Q3742-06	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	85.5
Sample Wt/Vol:	30.05 g	Final Vol:	10000 uL
Prep Method:	SW3541B	Test:	Pesticide-TCL
	Prep Date:	12/02/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.15	U	1	0.15	2.00	ug/kg	12/02/25 19:26	PB170776
319-85-7	beta-BHC	0.21	U	1	0.21	2.00	ug/kg	12/02/25 19:26	PB170776
319-86-8	delta-BHC	0.46	U	1	0.46	2.00	ug/kg	12/02/25 19:26	PB170776
58-89-9	gamma-BHC (Lindane)	0.16	U	1	0.16	2.00	ug/kg	12/02/25 19:26	PB170776
76-44-8	Heptachlor	0.14	U	1	0.14	2.00	ug/kg	12/02/25 19:26	PB170776
309-00-2	Aldrin	0.14	U	1	0.14	2.00	ug/kg	12/02/25 19:26	PB170776
1024-57-3	Heptachlor epoxide	0.22	U	1	0.22	2.00	ug/kg	12/02/25 19:26	PB170776
959-98-8	Endosulfan I	0.16	U	1	0.16	2.00	ug/kg	12/02/25 19:26	PB170776
60-57-1	Dieldrin	0.16	U	1	0.16	2.00	ug/kg	12/02/25 19:26	PB170776
72-55-9	4,4-DDE	0.16	U	1	0.16	2.00	ug/kg	12/02/25 19:26	PB170776
72-20-8	Endrin	0.16	U	1	0.16	2.00	ug/kg	12/02/25 19:26	PB170776
33213-65-9	Endosulfan II	0.34	U	1	0.34	2.00	ug/kg	12/02/25 19:26	PB170776
72-54-8	4,4-DDD	0.18	U	1	0.18	2.00	ug/kg	12/02/25 19:26	PB170776
1031-07-8	Endosulfan Sulfate	0.15	U	1	0.15	2.00	ug/kg	12/02/25 19:26	PB170776
50-29-3	4,4-DDT	0.16	U	1	0.16	2.00	ug/kg	12/02/25 19:26	PB170776
72-43-5	Methoxychlor	0.43	U	1	0.43	2.00	ug/kg	12/02/25 19:26	PB170776
53494-70-5	Endrin ketone	0.22	U	1	0.22	2.00	ug/kg	12/02/25 19:26	PB170776
7421-93-4	Endrin aldehyde	0.43	U	1	0.43	2.00	ug/kg	12/02/25 19:26	PB170776
5103-71-9	alpha-Chlordane	0.14	U	1	0.14	2.00	ug/kg	12/02/25 19:26	PB170776
5103-74-2	gamma-Chlordane	0.18	U	1	0.18	2.00	ug/kg	12/02/25 19:26	PB170776
8001-35-2	Toxaphene	6.30	U	1	6.30	38.5	ug/kg	12/02/25 19:26	PB170776
SURROGATES									
2051-24-3	Decachlorobiphenyl	16.0			10 - 143	80%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	15.4			10 - 131	77%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick	Date Collected:	11/25/25
Project:	Saddler Property	Date Received:	11/26/25
Client Sample ID:	SB-1125-18	SDG No.:	Q3742
Lab Sample ID:	Q3742-07	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	68.3
Sample Wt/Vol:	30.03 g	Final Vol:	10000 uL
Prep Method:	SW3541B	Test:	Pesticide-TCL
	Prep Date:	12/02/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.19	U	1	0.19	2.50	ug/kg	12/02/25 19:39	PB170776
319-85-7	beta-BHC	0.26	U	1	0.26	2.50	ug/kg	12/02/25 19:39	PB170776
319-86-8	delta-BHC	0.57	U	1	0.57	2.50	ug/kg	12/02/25 19:39	PB170776
58-89-9	gamma-BHC (Lindane)	0.20	U	1	0.20	2.50	ug/kg	12/02/25 19:39	PB170776
76-44-8	Heptachlor	0.18	U	1	0.18	2.50	ug/kg	12/02/25 19:39	PB170776
309-00-2	Aldrin	0.18	U	1	0.18	2.50	ug/kg	12/02/25 19:39	PB170776
1024-57-3	Heptachlor epoxide	0.28	U	1	0.28	2.50	ug/kg	12/02/25 19:39	PB170776
959-98-8	Endosulfan I	0.20	U	1	0.20	2.50	ug/kg	12/02/25 19:39	PB170776
60-57-1	Dieldrin	0.20	U	1	0.20	2.50	ug/kg	12/02/25 19:39	PB170776
72-55-9	4,4-DDE	0.20	U	1	0.20	2.50	ug/kg	12/02/25 19:39	PB170776
72-20-8	Endrin	0.20	U	1	0.20	2.50	ug/kg	12/02/25 19:39	PB170776
33213-65-9	Endosulfan II	0.42	U	1	0.42	2.50	ug/kg	12/02/25 19:39	PB170776
72-54-8	4,4-DDD	0.22	U	1	0.22	2.50	ug/kg	12/02/25 19:39	PB170776
1031-07-8	Endosulfan Sulfate	0.19	U	1	0.19	2.50	ug/kg	12/02/25 19:39	PB170776
50-29-3	4,4-DDT	0.20	U	1	0.20	2.50	ug/kg	12/02/25 19:39	PB170776
72-43-5	Methoxychlor	0.54	U	1	0.54	2.50	ug/kg	12/02/25 19:39	PB170776
53494-70-5	Endrin ketone	0.28	U	1	0.28	2.50	ug/kg	12/02/25 19:39	PB170776
7421-93-4	Endrin aldehyde	0.54	U	1	0.54	2.50	ug/kg	12/02/25 19:39	PB170776
5103-71-9	alpha-Chlordane	0.18	U	1	0.18	2.50	ug/kg	12/02/25 19:39	PB170776
5103-74-2	gamma-Chlordane	0.22	U	1	0.22	2.50	ug/kg	12/02/25 19:39	PB170776
8001-35-2	Toxaphene	7.90	U	1	7.90	48.3	ug/kg	12/02/25 19:39	PB170776
SURROGATES									
2051-24-3	Decachlorobiphenyl	18.0			10 - 143	90%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	18.8			10 - 131	94%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick	Date Collected:	11/25/25
Project:	Saddler Property	Date Received:	11/26/25
Client Sample ID:	SB-1125-20	SDG No.:	Q3742
Lab Sample ID:	Q3742-08	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	80.4
Sample Wt/Vol:	30.05 g	Final Vol:	10000 uL
Prep Method:	SW3541B	Test:	Pesticide-TCL
	Prep Date:	12/02/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.16	U	1	0.16	2.10	ug/kg	12/02/25 20:20	PB170776
319-85-7	beta-BHC	0.22	U	1	0.22	2.10	ug/kg	12/02/25 20:20	PB170776
319-86-8	delta-BHC	0.48	U	1	0.48	2.10	ug/kg	12/02/25 20:20	PB170776
58-89-9	gamma-BHC (Lindane)	0.17	U	1	0.17	2.10	ug/kg	12/02/25 20:20	PB170776
76-44-8	Heptachlor	0.15	U	1	0.15	2.10	ug/kg	12/02/25 20:20	PB170776
309-00-2	Aldrin	0.15	U	1	0.15	2.10	ug/kg	12/02/25 20:20	PB170776
1024-57-3	Heptachlor epoxide	0.24	U	1	0.24	2.10	ug/kg	12/02/25 20:20	PB170776
959-98-8	Endosulfan I	0.17	U	1	0.17	2.10	ug/kg	12/02/25 20:20	PB170776
60-57-1	Dieldrin	0.17	U	1	0.17	2.10	ug/kg	12/02/25 20:20	PB170776
72-55-9	4,4-DDE	0.17	U	1	0.17	2.10	ug/kg	12/02/25 20:20	PB170776
72-20-8	Endrin	0.17	U	1	0.17	2.10	ug/kg	12/02/25 20:20	PB170776
33213-65-9	Endosulfan II	0.36	U	1	0.36	2.10	ug/kg	12/02/25 20:20	PB170776
72-54-8	4,4-DDD	0.19	U	1	0.19	2.10	ug/kg	12/02/25 20:20	PB170776
1031-07-8	Endosulfan Sulfate	0.16	U	1	0.16	2.10	ug/kg	12/02/25 20:20	PB170776
50-29-3	4,4-DDT	0.17	U	1	0.17	2.10	ug/kg	12/02/25 20:20	PB170776
72-43-5	Methoxychlor	0.46	U	1	0.46	2.10	ug/kg	12/02/25 20:20	PB170776
53494-70-5	Endrin ketone	0.24	U	1	0.24	2.10	ug/kg	12/02/25 20:20	PB170776
7421-93-4	Endrin aldehyde	0.46	U	1	0.46	2.10	ug/kg	12/02/25 20:20	PB170776
5103-71-9	alpha-Chlordane	0.15	U	1	0.15	2.10	ug/kg	12/02/25 20:20	PB170776
5103-74-2	gamma-Chlordane	0.19	U	1	0.19	2.10	ug/kg	12/02/25 20:20	PB170776
8001-35-2	Toxaphene	6.70	U	1	6.70	41.0	ug/kg	12/02/25 20:20	PB170776
SURROGATES									
2051-24-3	Decachlorobiphenyl	18.1			10 - 143	90%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	21.2			10 - 131	106%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick		Date Collected:	11/25/25
Project:	Saddler Property		Date Received:	11/26/25
Client Sample ID:	SB-1125-22		SDG No.:	Q3742
Lab Sample ID:	Q3742-09		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	72.8
Sample Wt/Vol:	30.03 g	Final Vol: 10000 uL	Test:	Pesticide-TCL
Prep Method:	SW3541B	Prep Date: 12/02/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.18	U	1	0.18	2.30	ug/kg	12/02/25 20:34	PB170776
319-85-7	beta-BHC	0.25	U	1	0.25	2.30	ug/kg	12/02/25 20:34	PB170776
319-86-8	delta-BHC	0.54	U	1	0.54	2.30	ug/kg	12/02/25 20:34	PB170776
58-89-9	gamma-BHC (Lindane)	0.19	U	1	0.19	2.30	ug/kg	12/02/25 20:34	PB170776
76-44-8	Heptachlor	0.16	U	1	0.16	2.30	ug/kg	12/02/25 20:34	PB170776
309-00-2	Aldrin	0.16	U	1	0.16	2.30	ug/kg	12/02/25 20:34	PB170776
1024-57-3	Heptachlor epoxide	0.26	U	1	0.26	2.30	ug/kg	12/02/25 20:34	PB170776
959-98-8	Endosulfan I	0.19	U	1	0.19	2.30	ug/kg	12/02/25 20:34	PB170776
60-57-1	Dieldrin	0.19	U	1	0.19	2.30	ug/kg	12/02/25 20:34	PB170776
72-55-9	4,4-DDE	0.48	JP	1	0.19	2.30	ug/kg	12/02/25 20:34	PB170776
72-20-8	Endrin	0.19	U	1	0.19	2.30	ug/kg	12/02/25 20:34	PB170776
33213-65-9	Endosulfan II	0.40	U	1	0.40	2.30	ug/kg	12/02/25 20:34	PB170776
72-54-8	4,4-DDD	0.21	U	1	0.21	2.30	ug/kg	12/02/25 20:34	PB170776
1031-07-8	Endosulfan Sulfate	0.18	U	1	0.18	2.30	ug/kg	12/02/25 20:34	PB170776
50-29-3	4,4-DDT	0.19	U	1	0.19	2.30	ug/kg	12/02/25 20:34	PB170776
72-43-5	Methoxychlor	0.51	U	1	0.51	2.30	ug/kg	12/02/25 20:34	PB170776
53494-70-5	Endrin ketone	0.26	U	1	0.26	2.30	ug/kg	12/02/25 20:34	PB170776
7421-93-4	Endrin aldehyde	0.51	U	1	0.51	2.30	ug/kg	12/02/25 20:34	PB170776
5103-71-9	alpha-Chlordane	0.16	U	1	0.16	2.30	ug/kg	12/02/25 20:34	PB170776
5103-74-2	gamma-Chlordane	0.21	U	1	0.21	2.30	ug/kg	12/02/25 20:34	PB170776
8001-35-2	Toxaphene	7.40	U	1	7.40	45.3	ug/kg	12/02/25 20:34	PB170776
SURROGATES									
2051-24-3	Decachlorobiphenyl	16.2			10 - 143	81%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.6			10 - 131	103%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick	Date Collected:	11/25/25
Project:	Saddler Property	Date Received:	11/26/25
Client Sample ID:	SB-1125-2	SDG No.:	Q3742
Lab Sample ID:	Q3742-10	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	76.4
Sample Wt/Vol:	30.06 g	Final Vol:	10000 uL
Prep Method:	SW3541B	Test:	Pesticide-TCL
	Prep Date:	12/02/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.17	U	1	0.17	2.20	ug/kg	12/02/25 20:47	PB170776
319-85-7	beta-BHC	0.24	U	1	0.24	2.20	ug/kg	12/02/25 20:47	PB170776
319-86-8	delta-BHC	0.51	U	1	0.51	2.20	ug/kg	12/02/25 20:47	PB170776
58-89-9	gamma-BHC (Lindane)	0.18	U	1	0.18	2.20	ug/kg	12/02/25 20:47	PB170776
76-44-8	Heptachlor	0.16	U	1	0.16	2.20	ug/kg	12/02/25 20:47	PB170776
309-00-2	Aldrin	0.16	U	1	0.16	2.20	ug/kg	12/02/25 20:47	PB170776
1024-57-3	Heptachlor epoxide	0.25	U	1	0.25	2.20	ug/kg	12/02/25 20:47	PB170776
959-98-8	Endosulfan I	0.18	U	1	0.18	2.20	ug/kg	12/02/25 20:47	PB170776
60-57-1	Dieldrin	0.18	U	1	0.18	2.20	ug/kg	12/02/25 20:47	PB170776
72-55-9	4,4-DDE	0.18	U	1	0.18	2.20	ug/kg	12/02/25 20:47	PB170776
72-20-8	Endrin	0.18	U	1	0.18	2.20	ug/kg	12/02/25 20:47	PB170776
33213-65-9	Endosulfan II	0.38	U	1	0.38	2.20	ug/kg	12/02/25 20:47	PB170776
72-54-8	4,4-DDD	0.20	U	1	0.20	2.20	ug/kg	12/02/25 20:47	PB170776
1031-07-8	Endosulfan Sulfate	0.17	U	1	0.17	2.20	ug/kg	12/02/25 20:47	PB170776
50-29-3	4,4-DDT	0.18	U	1	0.18	2.20	ug/kg	12/02/25 20:47	PB170776
72-43-5	Methoxychlor	0.48	U	1	0.48	2.20	ug/kg	12/02/25 20:47	PB170776
53494-70-5	Endrin ketone	0.25	U	1	0.25	2.20	ug/kg	12/02/25 20:47	PB170776
7421-93-4	Endrin aldehyde	0.48	U	1	0.48	2.20	ug/kg	12/02/25 20:47	PB170776
5103-71-9	alpha-Chlordane	0.16	U	1	0.16	2.20	ug/kg	12/02/25 20:47	PB170776
5103-74-2	gamma-Chlordane	0.20	U	1	0.20	2.20	ug/kg	12/02/25 20:47	PB170776
8001-35-2	Toxaphene	7.10	U	1	7.10	43.1	ug/kg	12/02/25 20:47	PB170776
SURROGATES									
2051-24-3	Decachlorobiphenyl	16.0			10 - 143	80%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	21.0			10 - 131	105%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client: Remington & Vernick	Date Collected: 11/25/25	
Project: Saddler Property	Date Received: 11/26/25	
Client Sample ID: SB-1125-1	SDG No.: Q3742	
Lab Sample ID: Q3742-16	Matrix: SOIL	
Analytical Method: 8081B	% Solid: 73.5	
Sample Wt/Vol: 30.05 g	Test: Pesticide-TCL	Final Vol: 10000 uL
Prep Method: SW3541B		Prep Date: 12/02/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.18	U	1	0.18	2.30	ug/kg	12/02/25 21:01	PB170776
319-85-7	beta-BHC	0.24	U	1	0.24	2.30	ug/kg	12/02/25 21:01	PB170776
319-86-8	delta-BHC	0.53	U	1	0.53	2.30	ug/kg	12/02/25 21:01	PB170776
58-89-9	gamma-BHC (Lindane)	22.0		1	0.19	2.30	ug/kg	12/02/25 21:01	PB170776
76-44-8	Heptachlor	0.94	J	1	0.16	2.30	ug/kg	12/02/25 21:01	PB170776
309-00-2	Aldrin	0.16	U	1	0.16	2.30	ug/kg	12/02/25 21:01	PB170776
1024-57-3	Heptachlor epoxide	9.40		1	0.26	2.30	ug/kg	12/02/25 21:01	PB170776
959-98-8	Endosulfan I	0.19	U	1	0.19	2.30	ug/kg	12/02/25 21:01	PB170776
60-57-1	Dieldrin	0.46	J	1	0.19	2.30	ug/kg	12/02/25 21:01	PB170776
72-55-9	4,4-DDE	0.19	U	1	0.19	2.30	ug/kg	12/02/25 21:01	PB170776
72-20-8	Endrin	0.19	U	1	0.19	2.30	ug/kg	12/02/25 21:01	PB170776
33213-65-9	Endosulfan II	0.39	U	1	0.39	2.30	ug/kg	12/02/25 21:01	PB170776
72-54-8	4,4-DDD	0.20	U	1	0.20	2.30	ug/kg	12/02/25 21:01	PB170776
1031-07-8	Endosulfan Sulfate	0.18	U	1	0.18	2.30	ug/kg	12/02/25 21:01	PB170776
50-29-3	4,4-DDT	0.19	U	1	0.19	2.30	ug/kg	12/02/25 21:01	PB170776
72-43-5	Methoxychlor	0.50	U	1	0.50	2.30	ug/kg	12/02/25 21:01	PB170776
53494-70-5	Endrin ketone	0.26	U	1	0.26	2.30	ug/kg	12/02/25 21:01	PB170776
7421-93-4	Endrin aldehyde	0.50	U	1	0.50	2.30	ug/kg	12/02/25 21:01	PB170776
5103-71-9	alpha-Chlordane	23.6		1	0.16	2.30	ug/kg	12/02/25 21:01	PB170776
5103-74-2	gamma-Chlordane	30.9		1	0.20	2.30	ug/kg	12/02/25 21:01	PB170776
8001-35-2	Toxaphene	7.30	U	1	7.30	44.8	ug/kg	12/02/25 21:01	PB170776
SURROGATES									
2051-24-3	Decachlorobiphenyl	20.5			10 - 143	103%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	22.1			10 - 131	111%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick	Date Collected:	11/25/25
Project:	Saddler Property	Date Received:	11/26/25
Client Sample ID:	SB-1125-25	SDG No.:	Q3742
Lab Sample ID:	Q3742-17	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	63.3
Sample Wt/Vol:	30.06 g	Final Vol:	10000 uL
Prep Method:	SW3541B	Test:	Pesticide-TCL
	Prep Date:	12/02/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.21	U	1	0.21	2.70	ug/kg	12/02/25 21:15	PB170776
319-85-7	beta-BHC	0.28	U	1	0.28	2.70	ug/kg	12/02/25 21:15	PB170776
319-86-8	delta-BHC	0.61	U	1	0.61	2.70	ug/kg	12/02/25 21:15	PB170776
58-89-9	gamma-BHC (Lindane)	0.22	U	1	0.22	2.70	ug/kg	12/02/25 21:15	PB170776
76-44-8	Heptachlor	0.19	U	1	0.19	2.70	ug/kg	12/02/25 21:15	PB170776
309-00-2	Aldrin	0.19	U	1	0.19	2.70	ug/kg	12/02/25 21:15	PB170776
1024-57-3	Heptachlor epoxide	0.30	U	1	0.30	2.70	ug/kg	12/02/25 21:15	PB170776
959-98-8	Endosulfan I	0.22	U	1	0.22	2.70	ug/kg	12/02/25 21:15	PB170776
60-57-1	Dieldrin	0.22	U	1	0.22	2.70	ug/kg	12/02/25 21:15	PB170776
72-55-9	4,4-DDE	0.22	U	1	0.22	2.70	ug/kg	12/02/25 21:15	PB170776
72-20-8	Endrin	0.22	U	1	0.22	2.70	ug/kg	12/02/25 21:15	PB170776
33213-65-9	Endosulfan II	0.46	U	1	0.46	2.70	ug/kg	12/02/25 21:15	PB170776
72-54-8	4,4-DDD	0.24	U	1	0.24	2.70	ug/kg	12/02/25 21:15	PB170776
1031-07-8	Endosulfan Sulfate	0.21	U	1	0.21	2.70	ug/kg	12/02/25 21:15	PB170776
50-29-3	4,4-DDT	0.22	U	1	0.22	2.70	ug/kg	12/02/25 21:15	PB170776
72-43-5	Methoxychlor	0.58	U	1	0.58	2.70	ug/kg	12/02/25 21:15	PB170776
53494-70-5	Endrin ketone	0.30	U	1	0.30	2.70	ug/kg	12/02/25 21:15	PB170776
7421-93-4	Endrin aldehyde	0.58	U	1	0.58	2.70	ug/kg	12/02/25 21:15	PB170776
5103-71-9	alpha-Chlordane	0.19	U	1	0.19	2.70	ug/kg	12/02/25 21:15	PB170776
5103-74-2	gamma-Chlordane	0.24	U	1	0.24	2.70	ug/kg	12/02/25 21:15	PB170776
8001-35-2	Toxaphene	8.50	U	1	8.50	52.0	ug/kg	12/02/25 21:15	PB170776
SURROGATES									
2051-24-3	Decachlorobiphenyl	14.6			10 - 143	73%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	15.9			10 - 131	80%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client: Remington & Vernick	Date Collected: 11/25/25	
Project: Saddler Property	Date Received: 11/26/25	
Client Sample ID: SB-1125-24	SDG No.: Q3742	
Lab Sample ID: Q3742-18	Matrix: SOIL	
Analytical Method: 8081B	% Solid: 62.8	
Sample Wt/Vol: 30.04 g	Test: Pesticide-TCL	Final Vol: 10000 uL
Prep Method: SW3541B		Prep Date: 12/02/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.21	U	1	0.21	2.70	ug/kg	12/02/25 21:28	PB170776
319-85-7	beta-BHC	0.29	U	1	0.29	2.70	ug/kg	12/02/25 21:28	PB170776
319-86-8	delta-BHC	0.62	U	1	0.62	2.70	ug/kg	12/02/25 21:28	PB170776
58-89-9	gamma-BHC (Lindane)	0.22	U	1	0.22	2.70	ug/kg	12/02/25 21:28	PB170776
76-44-8	Heptachlor	0.19	U	1	0.19	2.70	ug/kg	12/02/25 21:28	PB170776
309-00-2	Aldrin	0.19	U	1	0.19	2.70	ug/kg	12/02/25 21:28	PB170776
1024-57-3	Heptachlor epoxide	0.30	U	1	0.30	2.70	ug/kg	12/02/25 21:28	PB170776
959-98-8	Endosulfan I	0.22	U	1	0.22	2.70	ug/kg	12/02/25 21:28	PB170776
60-57-1	Dieldrin	0.22	U	1	0.22	2.70	ug/kg	12/02/25 21:28	PB170776
72-55-9	4,4-DDE	0.22	U	1	0.22	2.70	ug/kg	12/02/25 21:28	PB170776
72-20-8	Endrin	0.22	U	1	0.22	2.70	ug/kg	12/02/25 21:28	PB170776
33213-65-9	Endosulfan II	0.46	U	1	0.46	2.70	ug/kg	12/02/25 21:28	PB170776
72-54-8	4,4-DDD	0.24	U	1	0.24	2.70	ug/kg	12/02/25 21:28	PB170776
1031-07-8	Endosulfan Sulfate	0.21	U	1	0.21	2.70	ug/kg	12/02/25 21:28	PB170776
50-29-3	4,4-DDT	0.22	U	1	0.22	2.70	ug/kg	12/02/25 21:28	PB170776
72-43-5	Methoxychlor	0.59	U	1	0.59	2.70	ug/kg	12/02/25 21:28	PB170776
53494-70-5	Endrin ketone	0.30	U	1	0.30	2.70	ug/kg	12/02/25 21:28	PB170776
7421-93-4	Endrin aldehyde	0.59	U	1	0.59	2.70	ug/kg	12/02/25 21:28	PB170776
5103-71-9	alpha-Chlordane	0.19	U	1	0.19	2.70	ug/kg	12/02/25 21:28	PB170776
5103-74-2	gamma-Chlordane	0.24	U	1	0.24	2.70	ug/kg	12/02/25 21:28	PB170776
8001-35-2	Toxaphene	8.60	U	1	8.60	52.5	ug/kg	12/02/25 21:28	PB170776
SURROGATES									
2051-24-3	Decachlorobiphenyl	15.0			10 - 143	75%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	19.3			10 - 131	97%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



QC SUMMARY

Surrogate Summary

SDG No.: **Q3742**

Client: **Remington & Vernick**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD091193.D	PIBLK-PD091193.D	Decachlorobiphen	1	20	20.8	104		31	149
		Tetrachloro-m-xyl	1	20	19.5	98		41	134
		Decachlorobiphen	2	20	20.1	100		31	149
		Tetrachloro-m-xyl	2	20	20.0	100		41	134
I.BLK-PD091416.D	PIBLK-PD091416.D	Decachlorobiphen	1	20	21.0	105		31	149
		Tetrachloro-m-xyl	1	20	21.3	107		41	134
		Decachlorobiphen	2	20	23.4	117		31	149
		Tetrachloro-m-xyl	2	20	25.8	129		41	134
PB170776BL	PB170776BL	Decachlorobiphen	1	20	22.3	112		10	143
		Tetrachloro-m-xyl	1	20	21.8	109		10	131
		Decachlorobiphen	2	20	25.1	125		10	143
		Tetrachloro-m-xyl	2	20	25.5	128		10	131
I.BLK-PD091425.D	PIBLK-PD091425.D	Decachlorobiphen	1	20	22.3	112		31	149
		Tetrachloro-m-xyl	1	20	22.1	110		41	134
		Decachlorobiphen	2	20	21.3	106		31	149
		Tetrachloro-m-xyl	2	20	25.6	128		41	134
Q3742-01	SB-1125-23	Decachlorobiphen	1	20	16.0	80		10	143
		Tetrachloro-m-xyl	1	20	18.0	90		10	131
		Decachlorobiphen	2	20	16.5	82		10	143
		Tetrachloro-m-xyl	2	20	21.1	105		10	131
Q3742-02	SB-1125-19	Decachlorobiphen	1	20	16.7	83		10	143
		Tetrachloro-m-xyl	1	20	18.8	94		10	131
		Decachlorobiphen	2	20	16.1	81		10	143
		Tetrachloro-m-xyl	2	20	21.9	109		10	131
Q3742-05	SB-1125-21	Decachlorobiphen	1	20	17.8	89		10	143
		Tetrachloro-m-xyl	1	20	22.6	113		10	131
		Decachlorobiphen	2	20	17.2	86		10	143
		Tetrachloro-m-xyl	2	20	23.0	115		10	131
I.BLK-PD091432.D	PIBLK-PD091432.D	Decachlorobiphen	1	20	20.2	101		31	149
		Tetrachloro-m-xyl	1	20	20.8	104		41	134
		Decachlorobiphen	2	20	21.9	109		31	149
		Tetrachloro-m-xyl	2	20	24.0	120		41	134
Q3742-06	SB-1125-13	Decachlorobiphen	1	20	15.7	78		10	143
		Tetrachloro-m-xyl	1	20	14.6	73		10	131
		Decachlorobiphen	2	20	16.0	80		10	143
		Tetrachloro-m-xyl	2	20	15.4	77		10	131
Q3742-07	SB-1125-18	Decachlorobiphen	1	20	18.0	90		10	143
		Tetrachloro-m-xyl	1	20	16.6	83		10	131
		Decachlorobiphen	2	20	14.3	71		10	143
		Tetrachloro-m-xyl	2	20	18.8	94		10	131
Q3742-07MS	SB-1125-18MS	Decachlorobiphen	1	20	16.3	81		10	143

Surrogate Summary

SDG No.: Q3742

Client: Remington & Vernick

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q3742-07MS	SB-1125-18MS	Tetrachloro-m-xyl	1	20	14.6	73		10	131
		Decachlorobiphen	2	20	12.6	63		10	143
Q3742-07MSD	SB-1125-18MSD	Tetrachloro-m-xyl	2	20	16.6	83		10	131
		Decachlorobiphen	1	20	15.8	79		10	143
		Tetrachloro-m-xyl	1	20	14.5	72		10	131
		Decachlorobiphen	2	20	12.4	62		10	143
		Tetrachloro-m-xyl	2	20	16.7	83		10	131
Q3742-08	SB-1125-20	Decachlorobiphen	1	20	18.1	90		10	143
		Tetrachloro-m-xyl	1	20	18.3	91		10	131
		Decachlorobiphen	2	20	15.8	79		10	143
		Tetrachloro-m-xyl	2	20	21.2	106		10	131
Q3742-09	SB-1125-22	Decachlorobiphen	1	20	16.2	81		10	143
		Tetrachloro-m-xyl	1	20	17.6	88		10	131
		Decachlorobiphen	2	20	16.0	80		10	143
		Tetrachloro-m-xyl	2	20	20.6	103		10	131
Q3742-10	SB-1125-2	Decachlorobiphen	1	20	15.3	77		10	143
		Tetrachloro-m-xyl	1	20	17.7	89		10	131
		Decachlorobiphen	2	20	16.0	80		10	143
		Tetrachloro-m-xyl	2	20	21.0	105		10	131
Q3742-16	SB-1125-1	Decachlorobiphen	1	20	19.2	96		10	143
		Tetrachloro-m-xyl	1	20	18.8	94		10	131
		Decachlorobiphen	2	20	20.5	103		10	143
		Tetrachloro-m-xyl	2	20	22.1	111		10	131
Q3742-17	SB-1125-25	Decachlorobiphen	1	20	14.6	73		10	143
		Tetrachloro-m-xyl	1	20	14.2	71		10	131
		Decachlorobiphen	2	20	12.6	63		10	143
		Tetrachloro-m-xyl	2	20	15.9	80		10	131
Q3742-18	SB-1125-24	Decachlorobiphen	1	20	15.0	75		10	143
		Tetrachloro-m-xyl	1	20	16.6	83		10	131
		Decachlorobiphen	2	20	14.6	73		10	143
		Tetrachloro-m-xyl	2	20	19.3	97		10	131
I.BLK-PD091445.D	PIBLK-PD091445.D	Decachlorobiphen	1	20	19.6	98		31	149
		Tetrachloro-m-xyl	1	20	20.3	101		41	134
		Decachlorobiphen	2	20	21.1	105		31	149
		Tetrachloro-m-xyl	2	20	22.9	115		41	134
I.BLK-PD091478.D	PIBLK-PD091478.D	Decachlorobiphen	1	20	20.6	103		31	149
		Tetrachloro-m-xyl	1	20	21.1	105		41	134
		Decachlorobiphen	2	20	19.6	98		31	149
		Tetrachloro-m-xyl	2	20	21.1	105		41	134
I.BLK-PD091523.D	PIBLK-PD091523.D	Decachlorobiphen	1	20	24.4	122		31	149
		Tetrachloro-m-xyl	1	20	22.3	111		41	134

Surrogate Summary

SDG No.: Q3742

Client: Remington & Vernick

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD091523.D	PIBLK-PD091523.D	Decachlorobiphen	2	20	27.1	135		31	149
		Tetrachloro-m-xyl	2	20	24.3	122		41	134
PB170776BS	PB170776BS	Decachlorobiphen	1	20	20.7	103		10	143
		Tetrachloro-m-xyl	1	20	18.3	91		10	131
		Decachlorobiphen	2	20	22.5	112		10	143
		Tetrachloro-m-xyl	2	20	20.5	103		10	131
I.BLK-PD091528.D	PIBLK-PD091528.D	Decachlorobiphen	1	20	23.1	115		31	149
		Tetrachloro-m-xyl	1	20	19.8	99		41	134
		Decachlorobiphen	2	20	24.2	121		31	149
		Tetrachloro-m-xyl	2	20	22.2	111		41	134

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3742</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Remington & Vernick</u>	DataFile :	<u>PD091437.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits		RPD
			Result								High		
Lab Sample ID:	Q3742-07MS (Column 1)	Client Sample ID:		SB-1125-18MS									
	alpha-BHC	24.35	0	23.2	ug/kg	95				60	144		
	beta-BHC	24.35	0	20.8	ug/kg	85				54	143		
	delta-BHC	24.35	0	23.1	ug/kg	95				47	129		
	gamma-BHC (Lindane)	24.35	0	22.6	ug/kg	93				39	136		
	Heptachlor	24.35	0	21.3	ug/kg	87				33	148		
	Aldrin	24.35	0	22.1	ug/kg	91				38	139		
	Heptachlor epoxide	24.35	0	21.4	ug/kg	88				23	155		
	Endosulfan I	24.35	0	21.3	ug/kg	87				30	142		
	Dieldrin	24.35	0	20.6	ug/kg	85				32	140		
	4,4'-DDE	24.35	0	20.3	ug/kg	83				20	156		
	Endrin	24.35	0	22.8	ug/kg	94				57	139		
	Endosulfan II	24.35	0	20.4	ug/kg	84				41	125		
	4,4'-DDD	24.35	0	20.3	ug/kg	83				47	163		
	Endosulfan sulfate	24.35	0	21.2	ug/kg	87				43	125		
	4,4'-DDT	24.35	0	23.5	ug/kg	97				25	149		
	Methoxychlor	24.35	0	23.0	ug/kg	94				32	132		
	Endrin ketone	24.35	0	21.6	ug/kg	89				27	142		
	Endrin aldehyde	24.35	0	21.7	ug/kg	89				19	148		
	alpha-Chlordane	24.35	0	19.8	ug/kg	81				26	147		
	gamma-Chlordane	24.35	0	20.8	ug/kg	85				30	142		
Lab Sample ID:	Q3742-07MS (Column 2)	Client Sample ID:		SB-1125-18MS									
	alpha-BHC	24.35	0	21.6	ug/kg	89				60	144		
	beta-BHC	24.35	0	20.8	ug/kg	85				54	143		
	delta-BHC	24.35	0	20.1	ug/kg	83				47	129		
	gamma-BHC (Lindane)	24.35	0	20.2	ug/kg	83				39	136		
	Heptachlor	24.35	0	19.3	ug/kg	79				33	148		
	Aldrin	24.35	0	20.7	ug/kg	85				38	139		
	Heptachlor epoxide	24.35	0	19.3	ug/kg	79				23	155		
	Endosulfan I	24.35	0	17.9	ug/kg	74				30	142		
	Dieldrin	24.35	0	18.7	ug/kg	77				32	140		
	4,4'-DDE	24.35	0	18.1	ug/kg	74				20	156		
	Endrin	24.35	0	19.6	ug/kg	80				57	139		
	Endosulfan II	24.35	0	18.1	ug/kg	74				41	125		
	4,4'-DDD	24.35	0	15.6	ug/kg	64				47	163		
	Endosulfan sulfate	24.35	0	18.1	ug/kg	74				43	125		

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3742</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Remington & Vernick</u>	DataFile :	<u>PD091437.D</u>

Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits		RPD
		Result	Result							High		
4,4'-DDT	24.35	0	19.7	ug/kg	81				25	149		
Methoxychlor	24.35	0	18.9	ug/kg	78				32	132		
Endrin ketone	24.35	0	17.0	ug/kg	70				27	142		
Endrin aldehyde	24.35	0	18.9	ug/kg	78				19	148		
alpha-Chlordane	24.35	0	18.5	ug/kg	76				26	147		
gamma-Chlordane	24.35	0	18.5	ug/kg	76				30	142		

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3742</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Remington & Vernick</u>	DataFile :	<u>PD091438.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD		Low	Limits	
			Result					Qual			High	RPD
Lab Sample ID:	Q3742-07MSD	Client Sample ID:		SB-1125-18MSD								
	(Column 1)											
	alpha-BHC	24.37	0	22.8	ug/kg	94		1		60	144	15
	beta-BHC	24.37	0	20.3	ug/kg	83		2		54	143	15
	delta-BHC	24.37	0	23.0	ug/kg	94		1		47	129	20
	gamma-BHC (Lindane)	24.37	0	22.2	ug/kg	91		2		39	136	30
	Heptachlor	24.37	0	21.4	ug/kg	88		1		33	148	20
	Aldrin	24.37	0	21.6	ug/kg	89		2		38	139	30
	Heptachlor epoxide	24.37	0	21.1	ug/kg	87		1		23	155	16
	Endosulfan I	24.37	0	20.9	ug/kg	86		1		30	142	15
	Dieldrin	24.37	0	20.3	ug/kg	83		2		32	140	20
	4,4'-DDE	24.37	0	20.0	ug/kg	82		1		20	156	16
	Endrin	24.37	0	23.0	ug/kg	94		0		57	139	16
	Endosulfan II	24.37	0	20.0	ug/kg	82		2		41	125	15
	4,4'-DDD	24.37	0	19.8	ug/kg	81		2		47	163	20
	Endosulfan sulfate	24.37	0	20.9	ug/kg	86		1		43	125	15
	4,4'-DDT	24.37	0	23.5	ug/kg	96		1		25	149	15
	Methoxychlor	24.37	0	23.2	ug/kg	95		1		32	132	20
	Endrin ketone	24.37	0	21.3	ug/kg	87		2		27	142	15
	Endrin aldehyde	24.37	0	21.4	ug/kg	88		1		19	148	15
	alpha-Chlordane	24.37	0	19.6	ug/kg	80		1		26	147	20
	gamma-Chlordane	24.37	0	20.5	ug/kg	84		1		30	142	15
Lab Sample ID:	Q3742-07MSD	Client Sample ID:		SB-1125-18MSD								
	(Column 2)											
	alpha-BHC	24.37	0	21.2	ug/kg	87		2		60	144	15
	beta-BHC	24.37	0	20.6	ug/kg	85		0		54	143	15
	delta-BHC	24.37	0	19.9	ug/kg	82		1		47	129	20
	gamma-BHC (Lindane)	24.37	0	19.8	ug/kg	81		2		39	136	30
	Heptachlor	24.37	0	19.4	ug/kg	80		1		33	148	20
	Aldrin	24.37	0	20.4	ug/kg	84		1		38	139	30
	Heptachlor epoxide	24.37	0	19.0	ug/kg	78		1		23	155	16
	Endosulfan I	24.37	0	17.7	ug/kg	73		1		30	142	15
	Dieldrin	24.37	0	18.5	ug/kg	76		1		32	140	20
	4,4'-DDE	24.37	0	17.9	ug/kg	73		1		20	156	16
	Endrin	24.37	0	19.6	ug/kg	80		0		57	139	16
	Endosulfan II	24.37	0	18.8	ug/kg	77		4		41	125	15
	4,4'-DDD	24.37	0	15.3	ug/kg	63		2		47	163	20
	Endosulfan sulfate	24.37	0	17.9	ug/kg	73		1		43	125	15

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3742</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Remington & Vernick</u>	DataFile :	<u>PD091438.D</u>

Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits		RPD
		Result	Result							High		
4,4'-DDT	24.37	0	20.0	ug/kg	82		1		25	149		15
Methoxychlor	24.37	0	19.2	ug/kg	79		1		32	132		20
Endrin ketone	24.37	0	16.8	ug/kg	69		1		27	142		15
Endrin aldehyde	24.37	0	18.7	ug/kg	77		1		19	148		15
alpha-Chlordane	24.37	0	18.2	ug/kg	75		1		26	147		20
gamma-Chlordane	24.37	0	18.3	ug/kg	75		1		30	142		15

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3742</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Remington & Vernick</u>	Datafile :	<u>PD091526.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB170776BS (Column 1)	alpha-BHC	16.66	15.7	ug/kg	94				84	123	
	beta-BHC	16.66	15.8	ug/kg	95				82	123	
	delta-BHC	16.66	19.2	ug/kg	115				83	126	
	gamma-BHC (Lindane)	16.66	15.9	ug/kg	95				83	125	
	Heptachlor	16.66	16.5	ug/kg	99				83	122	
	Aldrin	16.66	16.7	ug/kg	100				82	124	
	Heptachlor epoxide	16.66	16.5	ug/kg	99				83	120	
	Endosulfan I	16.66	16.3	ug/kg	98				81	124	
	Dieldrin	16.66	16.5	ug/kg	99				85	121	
	4,4'-DDE	16.66	16.4	ug/kg	98				81	123	
	Endrin	16.66	16.1	ug/kg	97				76	130	
	Endosulfan II	16.66	14.7	ug/kg	88				80	125	
	4,4'-DDD	16.66	16.8	ug/kg	101				80	131	
	Endosulfan sulfate	16.66	16.1	ug/kg	97				81	122	
	4,4'-DDT	16.66	15.8	ug/kg	95				70	129	
	Methoxychlor	16.66	15.5	ug/kg	93				60	119	
	Endrin ketone	16.66	16.3	ug/kg	98				77	132	
	Endrin aldehyde	16.66	16.2	ug/kg	97				79	124	
	alpha-Chlordane	16.66	15.8	ug/kg	95				84	120	
	gamma-Chlordane	16.66	16.5	ug/kg	99				83	122	
PB170776BS (Column 2)	alpha-BHC	16.66	16.8	ug/kg	101				84	123	
	beta-BHC	16.66	18.1	ug/kg	109				82	123	
	delta-BHC	16.66	17.5	ug/kg	105				83	126	
	gamma-BHC (Lindane)	16.66	18.7	ug/kg	112				83	125	
	Heptachlor	16.66	17.7	ug/kg	106				83	122	
	Aldrin	16.66	17.8	ug/kg	107				82	124	
	Heptachlor epoxide	16.66	18.3	ug/kg	110				83	120	
	Endosulfan I	16.66	17.9	ug/kg	107				81	124	
	Dieldrin	16.66	18.0	ug/kg	108				85	121	
	4,4'-DDE	16.66	18.0	ug/kg	108				81	123	
	Endrin	16.66	17.3	ug/kg	104				76	130	
	Endosulfan II	16.66	18.0	ug/kg	108				80	125	
	4,4'-DDD	16.66	18.3	ug/kg	110				80	131	
	Endosulfan sulfate	16.66	17.9	ug/kg	107				81	122	
	4,4'-DDT	16.66	18.0	ug/kg	108				70	129	
	Methoxychlor	16.66	17.7	ug/kg	106				60	119	
	Endrin ketone	16.66	18.2	ug/kg	109				77	132	
	Endrin aldehyde	16.66	17.7	ug/kg	106				79	124	
	alpha-Chlordane	16.66	18.2	ug/kg	109				84	120	
	gamma-Chlordane	16.66	18.0	ug/kg	108				83	122	

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID

PB170776BL

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Lab Sample ID: PB170776BL

Lab File ID: PD091419.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 12/02/2025

Date Analyzed (1): 12/02/2025

Date Analyzed (2): 12/02/2025

Time Analyzed (1): 13:37

Time Analyzed (2): 13:37

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
SB-1125-23	Q3742-01	PD091427.D	12/02/2025	12/02/2025
SB-1125-19	Q3742-02	PD091428.D	12/02/2025	12/02/2025
SB-1125-21	Q3742-05	PD091431.D	12/02/2025	12/02/2025
SB-1125-13	Q3742-06	PD091435.D	12/02/2025	12/02/2025
SB-1125-18	Q3742-07	PD091436.D	12/02/2025	12/02/2025
SB-1125-18MS	Q3742-07MS	PD091437.D	12/02/2025	12/02/2025
SB-1125-18MSD	Q3742-07MSD	PD091438.D	12/02/2025	12/02/2025
SB-1125-20	Q3742-08	PD091439.D	12/02/2025	12/02/2025
SB-1125-22	Q3742-09	PD091440.D	12/02/2025	12/02/2025
SB-1125-2	Q3742-10	PD091441.D	12/02/2025	12/02/2025
SB-1125-1	Q3742-16	PD091442.D	12/02/2025	12/02/2025
SB-1125-25	Q3742-17	PD091443.D	12/02/2025	12/02/2025
SB-1125-24	Q3742-18	PD091444.D	12/02/2025	12/02/2025
PB170776BS	PB170776BS	PD091526.D	12/09/2025	12/09/2025

COMMENTS: _____



QC SAMPLE DATA



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client: Remington & Vernick
Project: Saddler Property
Client Sample ID: PB170776BL
Lab Sample ID: PB170776BL
Analytical Method: 8081B
Sample Wt/Vol: 30.02 g Final Vol: 10000 uL
Prep Method: SW3541B Prep Date: 12/02/25

Date Collected:
Date Received:
SDG No.: Q3742
Matrix: SOIL
% Solid: 100
Test: Pesticide-TCL

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.13	U	1	0.13	1.70	ug/kg	12/02/25 13:37	PB170776
319-85-7	beta-BHC	0.18	U	1	0.18	1.70	ug/kg	12/02/25 13:37	PB170776
319-86-8	delta-BHC	0.39	U	1	0.39	1.70	ug/kg	12/02/25 13:37	PB170776
58-89-9	gamma-BHC (Lindane)	0.14	U	1	0.14	1.70	ug/kg	12/02/25 13:37	PB170776
76-44-8	Heptachlor	0.12	U	1	0.12	1.70	ug/kg	12/02/25 13:37	PB170776
309-00-2	Aldrin	0.12	U	1	0.12	1.70	ug/kg	12/02/25 13:37	PB170776
1024-57-3	Heptachlor epoxide	0.19	U	1	0.19	1.70	ug/kg	12/02/25 13:37	PB170776
959-98-8	Endosulfan I	0.14	U	1	0.14	1.70	ug/kg	12/02/25 13:37	PB170776
60-57-1	Dieldrin	0.14	U	1	0.14	1.70	ug/kg	12/02/25 13:37	PB170776
72-55-9	4,4-DDE	0.14	U	1	0.14	1.70	ug/kg	12/02/25 13:37	PB170776
72-20-8	Endrin	0.14	U	1	0.14	1.70	ug/kg	12/02/25 13:37	PB170776
33213-65-9	Endosulfan II	0.29	U	1	0.29	1.70	ug/kg	12/02/25 13:37	PB170776
72-54-8	4,4-DDD	0.15	U	1	0.15	1.70	ug/kg	12/02/25 13:37	PB170776
1031-07-8	Endosulfan Sulfate	0.13	U	1	0.13	1.70	ug/kg	12/02/25 13:37	PB170776
50-29-3	4,4-DDT	0.14	U	1	0.14	1.70	ug/kg	12/02/25 13:37	PB170776
72-43-5	Methoxychlor	0.37	U	1	0.37	1.70	ug/kg	12/02/25 13:37	PB170776
53494-70-5	Endrin ketone	0.19	U	1	0.19	1.70	ug/kg	12/02/25 13:37	PB170776
7421-93-4	Endrin aldehyde	0.37	U	1	0.37	1.70	ug/kg	12/02/25 13:37	PB170776
5103-71-9	alpha-Chlordane	0.12	U	1	0.12	1.70	ug/kg	12/02/25 13:37	PB170776
5103-74-2	gamma-Chlordane	0.15	U	1	0.15	1.70	ug/kg	12/02/25 13:37	PB170776
8001-35-2	Toxaphene	5.40	U	1	5.40	33.0	ug/kg	12/02/25 13:37	PB170776
SURROGATES									
2051-24-3	Decachlorobiphenyl	25.1			10 - 143	125%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	25.5			10 - 131	128%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Remington & Vernick		Date Collected:	11/14/25
Project:	Saddler Property		Date Received:	11/14/25
Client Sample ID:	PIBLK-PD091193.D		SDG No.:	Q3742
Lab Sample ID:	I.BLK-PD091193.D		Matrix:	WATER
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	Pesticide-TCL
Prep Method:	3510C	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.0039	U	1	0.0039	0.050	ug/L	11/14/25	PD111525
319-85-7	beta-BHC	0.0049	U	1	0.0049	0.050	ug/L	11/14/25	PD111525
319-86-8	delta-BHC	0.011	U	1	0.011	0.050	ug/L	11/14/25	PD111525
58-89-9	gamma-BHC (Lindane)	0.0037	U	1	0.0037	0.050	ug/L	11/14/25	PD111525
76-44-8	Heptachlor	0.0027	U	1	0.0027	0.050	ug/L	11/14/25	PD111525
309-00-2	Aldrin	0.0036	U	1	0.0036	0.050	ug/L	11/14/25	PD111525
1024-57-3	Heptachlor epoxide	0.0096	U	1	0.0096	0.050	ug/L	11/14/25	PD111525
959-98-8	Endosulfan I	0.0031	U	1	0.0031	0.050	ug/L	11/14/25	PD111525
60-57-1	Dieldrin	0.0036	U	1	0.0036	0.050	ug/L	11/14/25	PD111525
72-55-9	4,4-DDE	0.0037	U	1	0.0037	0.050	ug/L	11/14/25	PD111525
72-20-8	Endrin	0.0032	U	1	0.0032	0.050	ug/L	11/14/25	PD111525
33213-65-9	Endosulfan II	0.0079	U	1	0.0079	0.050	ug/L	11/14/25	PD111525
72-54-8	4,4-DDD	0.0071	U	1	0.0071	0.050	ug/L	11/14/25	PD111525
1031-07-8	Endosulfan Sulfate	0.0037	U	1	0.0037	0.050	ug/L	11/14/25	PD111525
50-29-3	4,4-DDT	0.0035	U	1	0.0035	0.050	ug/L	11/14/25	PD111525
72-43-5	Methoxychlor	0.011	U	1	0.011	0.050	ug/L	11/14/25	PD111525
53494-70-5	Endrin ketone	0.0093	U	1	0.0093	0.050	ug/L	11/14/25	PD111525
7421-93-4	Endrin aldehyde	0.011	U	1	0.011	0.050	ug/L	11/14/25	PD111525
5103-71-9	alpha-Chlordane	0.0035	U	1	0.0035	0.050	ug/L	11/14/25	PD111525
5103-74-2	gamma-Chlordane	0.0039	U	1	0.0039	0.050	ug/L	11/14/25	PD111525
8001-35-2	Toxaphene	0.17	U	1	0.17	1.00	ug/L	11/14/25	PD111525
SURROGATES									
2051-24-3	Decachlorobiphenyl	20.8			31 - 149	104%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.0			41 - 134	100%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Remington & Vernick		Date Collected:	12/02/25
Project:	Saddler Property		Date Received:	12/02/25
Client Sample ID:	PIBLK-PD091416.D		SDG No.:	Q3742
Lab Sample ID:	I.BLK-PD091416.D		Matrix:	WATER
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	Pesticide-TCL
Prep Method:	3510C	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.0039	U	1	0.0039	0.050	ug/L	12/02/25	PD120225
319-85-7	beta-BHC	0.0049	U	1	0.0049	0.050	ug/L	12/02/25	PD120225
319-86-8	delta-BHC	0.011	U	1	0.011	0.050	ug/L	12/02/25	PD120225
58-89-9	gamma-BHC (Lindane)	0.0037	U	1	0.0037	0.050	ug/L	12/02/25	PD120225
76-44-8	Heptachlor	0.0027	U	1	0.0027	0.050	ug/L	12/02/25	PD120225
309-00-2	Aldrin	0.0036	U	1	0.0036	0.050	ug/L	12/02/25	PD120225
1024-57-3	Heptachlor epoxide	0.0096	U	1	0.0096	0.050	ug/L	12/02/25	PD120225
959-98-8	Endosulfan I	0.0031	U	1	0.0031	0.050	ug/L	12/02/25	PD120225
60-57-1	Dieldrin	0.0036	U	1	0.0036	0.050	ug/L	12/02/25	PD120225
72-55-9	4,4-DDE	0.0037	U	1	0.0037	0.050	ug/L	12/02/25	PD120225
72-20-8	Endrin	0.0032	U	1	0.0032	0.050	ug/L	12/02/25	PD120225
33213-65-9	Endosulfan II	0.0079	U	1	0.0079	0.050	ug/L	12/02/25	PD120225
72-54-8	4,4-DDD	0.0071	U	1	0.0071	0.050	ug/L	12/02/25	PD120225
1031-07-8	Endosulfan Sulfate	0.0037	U	1	0.0037	0.050	ug/L	12/02/25	PD120225
50-29-3	4,4-DDT	0.0035	U	1	0.0035	0.050	ug/L	12/02/25	PD120225
72-43-5	Methoxychlor	0.011	U	1	0.011	0.050	ug/L	12/02/25	PD120225
53494-70-5	Endrin ketone	0.0093	U	1	0.0093	0.050	ug/L	12/02/25	PD120225
7421-93-4	Endrin aldehyde	0.011	U	1	0.011	0.050	ug/L	12/02/25	PD120225
5103-71-9	alpha-Chlordane	0.0035	U	1	0.0035	0.050	ug/L	12/02/25	PD120225
5103-74-2	gamma-Chlordane	0.0039	U	1	0.0039	0.050	ug/L	12/02/25	PD120225
8001-35-2	Toxaphene	0.17	U	1	0.17	1.00	ug/L	12/02/25	PD120225
SURROGATES									
2051-24-3	Decachlorobiphenyl	23.4			31 - 149	117%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	25.8			41 - 134	129%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Remington & Vernick		Date Collected:	12/02/25
Project:	Saddler Property		Date Received:	12/02/25
Client Sample ID:	PIBLK-PD091425.D		SDG No.:	Q3742
Lab Sample ID:	I.BLK-PD091425.D		Matrix:	WATER
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	Pesticide-TCL
Prep Method:	3510C	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.0039	U	1	0.0039	0.050	ug/L	12/02/25	PD120225
319-85-7	beta-BHC	0.0049	U	1	0.0049	0.050	ug/L	12/02/25	PD120225
319-86-8	delta-BHC	0.011	U	1	0.011	0.050	ug/L	12/02/25	PD120225
58-89-9	gamma-BHC (Lindane)	0.0037	U	1	0.0037	0.050	ug/L	12/02/25	PD120225
76-44-8	Heptachlor	0.0027	U	1	0.0027	0.050	ug/L	12/02/25	PD120225
309-00-2	Aldrin	0.0036	U	1	0.0036	0.050	ug/L	12/02/25	PD120225
1024-57-3	Heptachlor epoxide	0.0096	U	1	0.0096	0.050	ug/L	12/02/25	PD120225
959-98-8	Endosulfan I	0.0031	U	1	0.0031	0.050	ug/L	12/02/25	PD120225
60-57-1	Dieldrin	0.0036	U	1	0.0036	0.050	ug/L	12/02/25	PD120225
72-55-9	4,4-DDE	0.0037	U	1	0.0037	0.050	ug/L	12/02/25	PD120225
72-20-8	Endrin	0.0032	U	1	0.0032	0.050	ug/L	12/02/25	PD120225
33213-65-9	Endosulfan II	0.0079	U	1	0.0079	0.050	ug/L	12/02/25	PD120225
72-54-8	4,4-DDD	0.0071	U	1	0.0071	0.050	ug/L	12/02/25	PD120225
1031-07-8	Endosulfan Sulfate	0.0037	U	1	0.0037	0.050	ug/L	12/02/25	PD120225
50-29-3	4,4-DDT	0.0035	U	1	0.0035	0.050	ug/L	12/02/25	PD120225
72-43-5	Methoxychlor	0.011	U	1	0.011	0.050	ug/L	12/02/25	PD120225
53494-70-5	Endrin ketone	0.0093	U	1	0.0093	0.050	ug/L	12/02/25	PD120225
7421-93-4	Endrin aldehyde	0.011	U	1	0.011	0.050	ug/L	12/02/25	PD120225
5103-71-9	alpha-Chlordane	0.0035	U	1	0.0035	0.050	ug/L	12/02/25	PD120225
5103-74-2	gamma-Chlordane	0.0039	U	1	0.0039	0.050	ug/L	12/02/25	PD120225
8001-35-2	Toxaphene	0.17	U	1	0.17	1.00	ug/L	12/02/25	PD120225
SURROGATES									
2051-24-3	Decachlorobiphenyl	22.3			31 - 149	112%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	25.6			41 - 134	128%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Remington & Vernick		Date Collected:	12/02/25
Project:	Saddler Property		Date Received:	12/02/25
Client Sample ID:	PIBLK-PD091432.D		SDG No.:	Q3742
Lab Sample ID:	I.BLK-PD091432.D		Matrix:	WATER
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	Pesticide-TCL
Prep Method:	3510C	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.0039	U	1	0.0039	0.050	ug/L	12/02/25	PD120225
319-85-7	beta-BHC	0.0049	U	1	0.0049	0.050	ug/L	12/02/25	PD120225
319-86-8	delta-BHC	0.011	U	1	0.011	0.050	ug/L	12/02/25	PD120225
58-89-9	gamma-BHC (Lindane)	0.0037	U	1	0.0037	0.050	ug/L	12/02/25	PD120225
76-44-8	Heptachlor	0.0027	U	1	0.0027	0.050	ug/L	12/02/25	PD120225
309-00-2	Aldrin	0.0036	U	1	0.0036	0.050	ug/L	12/02/25	PD120225
1024-57-3	Heptachlor epoxide	0.0096	U	1	0.0096	0.050	ug/L	12/02/25	PD120225
959-98-8	Endosulfan I	0.0031	U	1	0.0031	0.050	ug/L	12/02/25	PD120225
60-57-1	Dieldrin	0.0036	U	1	0.0036	0.050	ug/L	12/02/25	PD120225
72-55-9	4,4-DDE	0.0037	U	1	0.0037	0.050	ug/L	12/02/25	PD120225
72-20-8	Endrin	0.0032	U	1	0.0032	0.050	ug/L	12/02/25	PD120225
33213-65-9	Endosulfan II	0.0079	U	1	0.0079	0.050	ug/L	12/02/25	PD120225
72-54-8	4,4-DDD	0.0071	U	1	0.0071	0.050	ug/L	12/02/25	PD120225
1031-07-8	Endosulfan Sulfate	0.0037	U	1	0.0037	0.050	ug/L	12/02/25	PD120225
50-29-3	4,4-DDT	0.0035	U	1	0.0035	0.050	ug/L	12/02/25	PD120225
72-43-5	Methoxychlor	0.011	U	1	0.011	0.050	ug/L	12/02/25	PD120225
53494-70-5	Endrin ketone	0.0093	U	1	0.0093	0.050	ug/L	12/02/25	PD120225
7421-93-4	Endrin aldehyde	0.011	U	1	0.011	0.050	ug/L	12/02/25	PD120225
5103-71-9	alpha-Chlordane	0.0035	U	1	0.0035	0.050	ug/L	12/02/25	PD120225
5103-74-2	gamma-Chlordane	0.0039	U	1	0.0039	0.050	ug/L	12/02/25	PD120225
8001-35-2	Toxaphene	0.17	U	1	0.17	1.00	ug/L	12/02/25	PD120225
SURROGATES									
2051-24-3	Decachlorobiphenyl	21.9			31 - 149	109%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	24.0			41 - 134	120%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Remington & Vernick		Date Collected:	12/02/25
Project:	Saddler Property		Date Received:	12/02/25
Client Sample ID:	PIBLK-PD091445.D		SDG No.:	Q3742
Lab Sample ID:	I.BLK-PD091445.D		Matrix:	WATER
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	Pesticide-TCL
Prep Method:	3510C	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.0039	U	1	0.0039	0.050	ug/L	12/02/25	PD120225
319-85-7	beta-BHC	0.0049	U	1	0.0049	0.050	ug/L	12/02/25	PD120225
319-86-8	delta-BHC	0.011	U	1	0.011	0.050	ug/L	12/02/25	PD120225
58-89-9	gamma-BHC (Lindane)	0.0037	U	1	0.0037	0.050	ug/L	12/02/25	PD120225
76-44-8	Heptachlor	0.0027	U	1	0.0027	0.050	ug/L	12/02/25	PD120225
309-00-2	Aldrin	0.0036	U	1	0.0036	0.050	ug/L	12/02/25	PD120225
1024-57-3	Heptachlor epoxide	0.0096	U	1	0.0096	0.050	ug/L	12/02/25	PD120225
959-98-8	Endosulfan I	0.0031	U	1	0.0031	0.050	ug/L	12/02/25	PD120225
60-57-1	Dieldrin	0.0036	U	1	0.0036	0.050	ug/L	12/02/25	PD120225
72-55-9	4,4-DDE	0.0037	U	1	0.0037	0.050	ug/L	12/02/25	PD120225
72-20-8	Endrin	0.0032	U	1	0.0032	0.050	ug/L	12/02/25	PD120225
33213-65-9	Endosulfan II	0.0079	U	1	0.0079	0.050	ug/L	12/02/25	PD120225
72-54-8	4,4-DDD	0.0071	U	1	0.0071	0.050	ug/L	12/02/25	PD120225
1031-07-8	Endosulfan Sulfate	0.0037	U	1	0.0037	0.050	ug/L	12/02/25	PD120225
50-29-3	4,4-DDT	0.0035	U	1	0.0035	0.050	ug/L	12/02/25	PD120225
72-43-5	Methoxychlor	0.011	U	1	0.011	0.050	ug/L	12/02/25	PD120225
53494-70-5	Endrin ketone	0.0093	U	1	0.0093	0.050	ug/L	12/02/25	PD120225
7421-93-4	Endrin aldehyde	0.011	U	1	0.011	0.050	ug/L	12/02/25	PD120225
5103-71-9	alpha-Chlordane	0.0035	U	1	0.0035	0.050	ug/L	12/02/25	PD120225
5103-74-2	gamma-Chlordane	0.0039	U	1	0.0039	0.050	ug/L	12/02/25	PD120225
8001-35-2	Toxaphene	0.17	U	1	0.17	1.00	ug/L	12/02/25	PD120225
SURROGATES									
2051-24-3	Decachlorobiphenyl	21.1			31 - 149	105%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	22.9			41 - 134	115%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Remington & Vernick		Date Collected:	12/08/25
Project:	Saddler Property		Date Received:	12/08/25
Client Sample ID:	PIBLK-PD091478.D		SDG No.:	Q3742
Lab Sample ID:	I.BLK-PD091478.D		Matrix:	WATER
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	Pesticide-TCL
Prep Method:	3510C	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.0039	U	1	0.0039	0.050	ug/L	12/08/25	PD120825
319-85-7	beta-BHC	0.0049	U	1	0.0049	0.050	ug/L	12/08/25	PD120825
319-86-8	delta-BHC	0.011	U	1	0.011	0.050	ug/L	12/08/25	PD120825
58-89-9	gamma-BHC (Lindane)	0.0037	U	1	0.0037	0.050	ug/L	12/08/25	PD120825
76-44-8	Heptachlor	0.0027	U	1	0.0027	0.050	ug/L	12/08/25	PD120825
309-00-2	Aldrin	0.0036	U	1	0.0036	0.050	ug/L	12/08/25	PD120825
1024-57-3	Heptachlor epoxide	0.0096	U	1	0.0096	0.050	ug/L	12/08/25	PD120825
959-98-8	Endosulfan I	0.0031	U	1	0.0031	0.050	ug/L	12/08/25	PD120825
60-57-1	Dieldrin	0.0036	U	1	0.0036	0.050	ug/L	12/08/25	PD120825
72-55-9	4,4-DDE	0.0037	U	1	0.0037	0.050	ug/L	12/08/25	PD120825
72-20-8	Endrin	0.0032	U	1	0.0032	0.050	ug/L	12/08/25	PD120825
33213-65-9	Endosulfan II	0.0079	U	1	0.0079	0.050	ug/L	12/08/25	PD120825
72-54-8	4,4-DDD	0.0071	U	1	0.0071	0.050	ug/L	12/08/25	PD120825
1031-07-8	Endosulfan Sulfate	0.0037	U	1	0.0037	0.050	ug/L	12/08/25	PD120825
50-29-3	4,4-DDT	0.0035	U	1	0.0035	0.050	ug/L	12/08/25	PD120825
72-43-5	Methoxychlor	0.011	U	1	0.011	0.050	ug/L	12/08/25	PD120825
53494-70-5	Endrin ketone	0.0093	U	1	0.0093	0.050	ug/L	12/08/25	PD120825
7421-93-4	Endrin aldehyde	0.011	U	1	0.011	0.050	ug/L	12/08/25	PD120825
5103-71-9	alpha-Chlordane	0.0035	U	1	0.0035	0.050	ug/L	12/08/25	PD120825
5103-74-2	gamma-Chlordane	0.0039	U	1	0.0039	0.050	ug/L	12/08/25	PD120825
8001-35-2	Toxaphene	0.17	U	1	0.17	1.00	ug/L	12/08/25	PD120825
SURROGATES									
2051-24-3	Decachlorobiphenyl	20.6			31 - 149	103%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	21.1			41 - 134	105%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Remington & Vernick	Date Collected:	12/09/25
Project:	Saddler Property	Date Received:	12/09/25
Client Sample ID:	PIBLK-PD091523.D	SDG No.:	Q3742
Lab Sample ID:	I.BLK-PD091523.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol:	10000 uL
Prep Method:	3510C	Test:	Pesticide-TCL
	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.0039	U	1	0.0039	0.050	ug/L	12/09/25	Pd120925
319-85-7	beta-BHC	0.0049	U	1	0.0049	0.050	ug/L	12/09/25	Pd120925
319-86-8	delta-BHC	0.011	U	1	0.011	0.050	ug/L	12/09/25	Pd120925
58-89-9	gamma-BHC (Lindane)	0.0037	U	1	0.0037	0.050	ug/L	12/09/25	Pd120925
76-44-8	Heptachlor	0.0027	U	1	0.0027	0.050	ug/L	12/09/25	Pd120925
309-00-2	Aldrin	0.0036	U	1	0.0036	0.050	ug/L	12/09/25	Pd120925
1024-57-3	Heptachlor epoxide	0.0096	U	1	0.0096	0.050	ug/L	12/09/25	Pd120925
959-98-8	Endosulfan I	0.0031	U	1	0.0031	0.050	ug/L	12/09/25	Pd120925
60-57-1	Dieldrin	0.0036	U	1	0.0036	0.050	ug/L	12/09/25	Pd120925
72-55-9	4,4-DDE	0.0037	U	1	0.0037	0.050	ug/L	12/09/25	Pd120925
72-20-8	Endrin	0.0032	U	1	0.0032	0.050	ug/L	12/09/25	Pd120925
33213-65-9	Endosulfan II	0.0079	U	1	0.0079	0.050	ug/L	12/09/25	Pd120925
72-54-8	4,4-DDD	0.0071	U	1	0.0071	0.050	ug/L	12/09/25	Pd120925
1031-07-8	Endosulfan Sulfate	0.0037	U	1	0.0037	0.050	ug/L	12/09/25	Pd120925
50-29-3	4,4-DDT	0.0035	U	1	0.0035	0.050	ug/L	12/09/25	Pd120925
72-43-5	Methoxychlor	0.011	U	1	0.011	0.050	ug/L	12/09/25	Pd120925
53494-70-5	Endrin ketone	0.0093	U	1	0.0093	0.050	ug/L	12/09/25	Pd120925
7421-93-4	Endrin aldehyde	0.011	U	1	0.011	0.050	ug/L	12/09/25	Pd120925
5103-71-9	alpha-Chlordane	0.0035	U	1	0.0035	0.050	ug/L	12/09/25	Pd120925
5103-74-2	gamma-Chlordane	0.0039	U	1	0.0039	0.050	ug/L	12/09/25	Pd120925
8001-35-2	Toxaphene	0.17	U	1	0.17	1.00	ug/L	12/09/25	Pd120925
SURROGATES									
2051-24-3	Decachlorobiphenyl	27.1			31 - 149	135%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	24.3			41 - 134	122%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Remington & Vernick		Date Collected:	12/09/25
Project:	Saddler Property		Date Received:	12/09/25
Client Sample ID:	PIBLK-PD091528.D		SDG No.:	Q3742
Lab Sample ID:	I.BLK-PD091528.D		Matrix:	WATER
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	Pesticide-TCL
Prep Method:	3510C	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.0039	U	1	0.0039	0.050	ug/L	12/09/25	PD120925
319-85-7	beta-BHC	0.0049	U	1	0.0049	0.050	ug/L	12/09/25	PD120925
319-86-8	delta-BHC	0.011	U	1	0.011	0.050	ug/L	12/09/25	PD120925
58-89-9	gamma-BHC (Lindane)	0.0037	U	1	0.0037	0.050	ug/L	12/09/25	PD120925
76-44-8	Heptachlor	0.0027	U	1	0.0027	0.050	ug/L	12/09/25	PD120925
309-00-2	Aldrin	0.0036	U	1	0.0036	0.050	ug/L	12/09/25	PD120925
1024-57-3	Heptachlor epoxide	0.0096	U	1	0.0096	0.050	ug/L	12/09/25	PD120925
959-98-8	Endosulfan I	0.0031	U	1	0.0031	0.050	ug/L	12/09/25	PD120925
60-57-1	Dieldrin	0.0036	U	1	0.0036	0.050	ug/L	12/09/25	PD120925
72-55-9	4,4-DDE	0.0037	U	1	0.0037	0.050	ug/L	12/09/25	PD120925
72-20-8	Endrin	0.0032	U	1	0.0032	0.050	ug/L	12/09/25	PD120925
33213-65-9	Endosulfan II	0.0079	U	1	0.0079	0.050	ug/L	12/09/25	PD120925
72-54-8	4,4-DDD	0.0071	U	1	0.0071	0.050	ug/L	12/09/25	PD120925
1031-07-8	Endosulfan Sulfate	0.0037	U	1	0.0037	0.050	ug/L	12/09/25	PD120925
50-29-3	4,4-DDT	0.0035	U	1	0.0035	0.050	ug/L	12/09/25	PD120925
72-43-5	Methoxychlor	0.011	U	1	0.011	0.050	ug/L	12/09/25	PD120925
53494-70-5	Endrin ketone	0.0093	U	1	0.0093	0.050	ug/L	12/09/25	PD120925
7421-93-4	Endrin aldehyde	0.011	U	1	0.011	0.050	ug/L	12/09/25	PD120925
5103-71-9	alpha-Chlordane	0.0035	U	1	0.0035	0.050	ug/L	12/09/25	PD120925
5103-74-2	gamma-Chlordane	0.0039	U	1	0.0039	0.050	ug/L	12/09/25	PD120925
8001-35-2	Toxaphene	0.17	U	1	0.17	1.00	ug/L	12/09/25	PD120925
SURROGATES									
2051-24-3	Decachlorobiphenyl	24.2			31 - 149	121%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	22.2			41 - 134	111%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick		Date Collected:	
Project:	Saddler Property		Date Received:	
Client Sample ID:	PB170776BS		SDG No.:	Q3742
Lab Sample ID:	PB170776BS		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	100
Sample Wt/Vol:	30.01 g	Final Vol: 10000 uL	Test:	Pesticide-TCL
Prep Method:	SW3541B	Prep Date: 12/02/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	16.8		1	0.13	1.70	ug/kg	12/09/25 14:22	PB170776
319-85-7	beta-BHC	18.1		1	0.18	1.70	ug/kg	12/09/25 14:22	PB170776
319-86-8	delta-BHC	19.2		1	0.39	1.70	ug/kg	12/09/25 14:22	PB170776
58-89-9	gamma-BHC (Lindane)	18.7		1	0.14	1.70	ug/kg	12/09/25 14:22	PB170776
76-44-8	Heptachlor	17.7		1	0.12	1.70	ug/kg	12/09/25 14:22	PB170776
309-00-2	Aldrin	17.8		1	0.12	1.70	ug/kg	12/09/25 14:22	PB170776
1024-57-3	Heptachlor epoxide	18.3		1	0.19	1.70	ug/kg	12/09/25 14:22	PB170776
959-98-8	Endosulfan I	17.9		1	0.14	1.70	ug/kg	12/09/25 14:22	PB170776
60-57-1	Dieldrin	18.0		1	0.14	1.70	ug/kg	12/09/25 14:22	PB170776
72-55-9	4,4-DDE	18.0		1	0.14	1.70	ug/kg	12/09/25 14:22	PB170776
72-20-8	Endrin	17.3		1	0.14	1.70	ug/kg	12/09/25 14:22	PB170776
33213-65-9	Endosulfan II	18.0		1	0.29	1.70	ug/kg	12/09/25 14:22	PB170776
72-54-8	4,4-DDD	18.3		1	0.15	1.70	ug/kg	12/09/25 14:22	PB170776
1031-07-8	Endosulfan Sulfate	17.9		1	0.13	1.70	ug/kg	12/09/25 14:22	PB170776
50-29-3	4,4-DDT	18.0		1	0.14	1.70	ug/kg	12/09/25 14:22	PB170776
72-43-5	Methoxychlor	17.7		1	0.37	1.70	ug/kg	12/09/25 14:22	PB170776
53494-70-5	Endrin ketone	18.2		1	0.19	1.70	ug/kg	12/09/25 14:22	PB170776
7421-93-4	Endrin aldehyde	17.7		1	0.37	1.70	ug/kg	12/09/25 14:22	PB170776
5103-71-9	alpha-Chlordane	18.2		1	0.12	1.70	ug/kg	12/09/25 14:22	PB170776
5103-74-2	gamma-Chlordane	18.0		1	0.15	1.70	ug/kg	12/09/25 14:22	PB170776
8001-35-2	Toxaphene	5.40	U	1	5.40	33.0	ug/kg	12/09/25 14:22	PB170776
SURROGATES									
2051-24-3	Decachlorobiphenyl	22.5			10 - 143	112%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.5			10 - 131	103%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick	Date Collected:	11/25/25
Project:	Saddler Property	Date Received:	11/26/25
Client Sample ID:	SB-1125-18MS	SDG No.:	Q3742
Lab Sample ID:	Q3742-07MS	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	68.3
Sample Wt/Vol:	30.07 g	Final Vol:	10000 uL
Prep Method:	SW3541B	Test:	Pesticide-TCL
	Prep Date:	12/02/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	23.2		1	0.19	2.50	ug/kg	12/02/25 19:53	PB170776
319-85-7	beta-BHC	20.8		1	0.26	2.50	ug/kg	12/02/25 19:53	PB170776
319-86-8	delta-BHC	23.1		1	0.57	2.50	ug/kg	12/02/25 19:53	PB170776
58-89-9	gamma-BHC (Lindane)	22.6		1	0.20	2.50	ug/kg	12/02/25 19:53	PB170776
76-44-8	Heptachlor	21.3		1	0.18	2.50	ug/kg	12/02/25 19:53	PB170776
309-00-2	Aldrin	22.1		1	0.18	2.50	ug/kg	12/02/25 19:53	PB170776
1024-57-3	Heptachlor epoxide	21.4		1	0.28	2.50	ug/kg	12/02/25 19:53	PB170776
959-98-8	Endosulfan I	21.3		1	0.20	2.50	ug/kg	12/02/25 19:53	PB170776
60-57-1	Dieldrin	20.6		1	0.20	2.50	ug/kg	12/02/25 19:53	PB170776
72-55-9	4,4-DDE	20.3		1	0.20	2.50	ug/kg	12/02/25 19:53	PB170776
72-20-8	Endrin	22.8		1	0.20	2.50	ug/kg	12/02/25 19:53	PB170776
33213-65-9	Endosulfan II	20.4		1	0.42	2.50	ug/kg	12/02/25 19:53	PB170776
72-54-8	4,4-DDD	20.3		1	0.22	2.50	ug/kg	12/02/25 19:53	PB170776
1031-07-8	Endosulfan Sulfate	21.2		1	0.19	2.50	ug/kg	12/02/25 19:53	PB170776
50-29-3	4,4-DDT	23.5		1	0.20	2.50	ug/kg	12/02/25 19:53	PB170776
72-43-5	Methoxychlor	23.0		1	0.54	2.50	ug/kg	12/02/25 19:53	PB170776
53494-70-5	Endrin ketone	21.6		1	0.28	2.50	ug/kg	12/02/25 19:53	PB170776
7421-93-4	Endrin aldehyde	21.7		1	0.54	2.50	ug/kg	12/02/25 19:53	PB170776
5103-71-9	alpha-Chlordane	19.8		1	0.18	2.50	ug/kg	12/02/25 19:53	PB170776
5103-74-2	gamma-Chlordane	20.8		1	0.22	2.50	ug/kg	12/02/25 19:53	PB170776
8001-35-2	Toxaphene	7.90	U	1	7.90	48.2	ug/kg	12/02/25 19:53	PB170776
SURROGATES									
2051-24-3	Decachlorobiphenyl	16.3			10 - 143	81%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	16.6			10 - 131	83%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick	Date Collected:	11/25/25
Project:	Saddler Property	Date Received:	11/26/25
Client Sample ID:	SB-1125-18MSD	SDG No.:	Q3742
Lab Sample ID:	Q3742-07MSD	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	68.3
Sample Wt/Vol:	30.04 g	Final Vol:	10000 uL
Prep Method:	SW3541B	Test:	Pesticide-TCL
	Prep Date:	12/02/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	22.8		1	0.19	2.50	ug/kg	12/02/25 20:06	PB170776
319-85-7	beta-BHC	20.6		1	0.26	2.50	ug/kg	12/02/25 20:06	PB170776
319-86-8	delta-BHC	23.0		1	0.57	2.50	ug/kg	12/02/25 20:06	PB170776
58-89-9	gamma-BHC (Lindane)	22.2		1	0.20	2.50	ug/kg	12/02/25 20:06	PB170776
76-44-8	Heptachlor	21.4		1	0.18	2.50	ug/kg	12/02/25 20:06	PB170776
309-00-2	Aldrin	21.6		1	0.18	2.50	ug/kg	12/02/25 20:06	PB170776
1024-57-3	Heptachlor epoxide	21.1		1	0.28	2.50	ug/kg	12/02/25 20:06	PB170776
959-98-8	Endosulfan I	20.9		1	0.20	2.50	ug/kg	12/02/25 20:06	PB170776
60-57-1	Dieldrin	20.3		1	0.20	2.50	ug/kg	12/02/25 20:06	PB170776
72-55-9	4,4-DDE	20.0		1	0.20	2.50	ug/kg	12/02/25 20:06	PB170776
72-20-8	Endrin	23.0		1	0.20	2.50	ug/kg	12/02/25 20:06	PB170776
33213-65-9	Endosulfan II	20.0		1	0.42	2.50	ug/kg	12/02/25 20:06	PB170776
72-54-8	4,4-DDD	19.8		1	0.22	2.50	ug/kg	12/02/25 20:06	PB170776
1031-07-8	Endosulfan Sulfate	20.9		1	0.19	2.50	ug/kg	12/02/25 20:06	PB170776
50-29-3	4,4-DDT	23.5		1	0.20	2.50	ug/kg	12/02/25 20:06	PB170776
72-43-5	Methoxychlor	23.2		1	0.54	2.50	ug/kg	12/02/25 20:06	PB170776
53494-70-5	Endrin ketone	21.3		1	0.28	2.50	ug/kg	12/02/25 20:06	PB170776
7421-93-4	Endrin aldehyde	21.4		1	0.54	2.50	ug/kg	12/02/25 20:06	PB170776
5103-71-9	alpha-Chlordane	19.6		1	0.18	2.50	ug/kg	12/02/25 20:06	PB170776
5103-74-2	gamma-Chlordane	20.5		1	0.22	2.50	ug/kg	12/02/25 20:06	PB170776
8001-35-2	Toxaphene	7.90	U	1	7.90	48.3	ug/kg	12/02/25 20:06	PB170776
SURROGATES									
2051-24-3	Decachlorobiphenyl	15.8			10 - 143	79%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	16.7			10 - 131	83%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



CALIBRATION SUMMARY

A
B
C
D
E
F
G
H
I
J
K
L

Contract: REMI01

SDG NO.: Q3742

Calibration Date(s): 11/14/2025 11/14/2025

Calibration Times: 20:56 21:51

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		RT 100 =	<u>PD091196.D</u>	RT 075 =	<u>PD091197.D</u>
RT 050 =	PD091198.D	RT 025 =	PD091199.D	RT 005 =	PD091200.D

[illegible]



Lab Name:	<u>Alliance</u>	Contract:	<u>REMI01</u>	
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3742</u>	
Instrument ID:	<u>ECD_D</u>	Calibration Date(s):	<u>11/14/2025</u>	<u>11/14/2025</u>
		Calibration Times:	<u>20:56</u>	<u>21:51</u>

LAB FILE ID:	RT 100 =	<u>PD091196.D</u>	RT 075 =	<u>PD091197.D</u>
RT 050 =	<u>PD091198.D</u>	RT 025 =	<u>PD091199.D</u>	RT 005 =
			<u>PD091200.D</u>	

[illegible]

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_D

Contract: REMI01
SDG NO.: Q3742

Calibration Date(s): 11/14/2025 11/14/2025
Calibration Times: 20:56 21:51

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD091196.D</u>	CF 075 =	<u>PD091197.D</u>		
CF 050 =		<u>PD091198.D</u>	CF 025 =	<u>PD091199.D</u>	CF 005 =	<u>PD091200.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	4592720000	4750410000	4706670000	4762630000	5745110000	4911510000	10
4,4'-DDE	5537410000	5724420000	5594750000	5596930000	6543500000	5799400000	7
4,4'-DDT	4033900000	4103640000	4023370000	3973800000	4542840000	4135510000	6
Aldrin	6913630000	7108360000	6994820000	7033330000	8231840000	7256400000	8
alpha-BHC	7991490000	8120730000	7870330000	7613270000	8381750000	7995510000	4
alpha-Chlordane	6168980000	6320470000	6262210000	6530520000	9131880000	6882810000	18
beta-BHC	2584470000	2701070000	2721620000	2850470000	3669860000	2905500000	15
Decachlorobiphenyl	4266550000	4520270000	4633000000	4951710000	6537810000	4981870000	18
delta-BHC	7343040000	7477970000	7279360000	7082560000	7977500000	7432080000	5
Dieldrin	6112350000	6323390000	6220240000	6264150000	7404460000	6464920000	8
Endosulfan I	5622190000	5847020000	5809210000	5991120000	7445430000	6142990000	12
Endosulfan II	5022110000	5221970000	5253890000	5419940000	6878740000	5559330000	14
Endosulfan sulfate	4635330000	4824080000	4867360000	5023730000	6373440000	5144790000	14
Endrin	4763010000	4861760000	4775160000	4762720000	5654790000	4963490000	8
Endrin aldehyde	3642590000	3820110000	3882150000	4087850000	5397480000	4166040000	17
Endrin ketone	5131250000	5317460000	5324710000	5507230000	6887120000	5633560000	13
gamma-BHC (Lindane)	7349570000	7492470000	7318000000	7162850000	8041460000	7472870000	5
gamma-Chlordane	6123460000	6343120000	6211230000	6313670000	7512390000	6500770000	9
Heptachlor	6873030000	7051630000	6922840000	6902860000	8058350000	7161740000	7
Heptachlor epoxide	6045880000	6198590000	6197710000	6327850000	7713630000	6496730000	11
Methoxychlor	1967810000	2033910000	2048060000	2143440000	2678460000	2174340000	13
Tetrachloro-m-xylene	3100550000	3224630000	3262160000	3454950000	4200860000	3448630000	13

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_D

Contract: REMI01
SDG NO.: Q3742

Calibration Date(s): 11/14/2025 11/14/2025
Calibration Times: 20:56 21:51

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD091196.D</u>	CF 075 =	<u>PD091197.D</u>		
CF 050 =		<u>PD091198.D</u>	CF 025 =	<u>PD091199.D</u>	CF 005 =	<u>PD091200.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	42169800000	44652200000	46098100000	48914200000	62508300000	48868500000	16
4,4'-DDE	46784900000	49153600000	50413800000	53041500000	67816900000	53442100000	16
4,4'-DDT	36565400000	37844700000	38225900000	38748200000	44278500000	39132500000	8
Aldrin	49756900000	52224800000	53642000000	56183400000	70485000000	56458400000	14
alpha-BHC	57081100000	59401300000	60005600000	62092800000	77433200000	63202800000	13
alpha-Chlordane	46162300000	48496400000	49694800000	52042000000	65261400000	52331400000	14
beta-BHC	20855800000	21963600000	22629700000	23880400000	30805800000	24027100000	16
Decachlorobiphenyl	35391500000	36530600000	37125400000	38755300000	50071600000	39574900000	15
delta-BHC	51951800000	54520500000	55333700000	57523500000	73056200000	58477100000	14
Dieldrin	47342900000	49974500000	51581700000	54390600000	68265900000	54311100000	15
Endosulfan I	41219300000	43810000000	45334200000	48114600000	60781000000	47851800000	16
Endosulfan II	40233600000	42630400000	44155200000	46739900000	59639900000	46679800000	16
Endosulfan sulfate	38807500000	40991800000	42555300000	44718700000	57101400000	44834900000	16
Endrin	39723100000	41747800000	43010100000	44994600000	56487600000	45192700000	15
Endrin aldehyde	29904600000	31738200000	33159100000	35293100000	46350200000	35289100000	18
Endrin ketone	43743400000	46381200000	48213200000	51142400000	65299300000	50955900000	17
gamma-BHC (Lindane)	52036700000	54159800000	54955200000	56993600000	70961400000	57821300000	13
gamma-Chlordane	48541600000	50822000000	51970500000	54222400000	67580500000	54627400000	14
Heptachlor	48452800000	51096600000	52341200000	55061200000	69095300000	55209400000	15
Heptachlor epoxide	43794700000	46344600000	48042100000	50883900000	65728100000	50958700000	17
Methoxychlor	17875100000	18763500000	19368600000	20213700000	24036200000	20051400000	12
Tetrachloro-m-xylene	32274400000	33875900000	34577100000	36418900000	46004800000	36630200000	15

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Instrument ID: ECD_D

Date(s) Analyzed: 11/14/2025 11/14/2025

GC Column: ZB-MR1
ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.23	6.13	6.33	42812100
		2	6.44	6.34	6.54	68979500
		3	7.14	7.04	7.24	112624000
		4	7.56	7.46	7.66	130712000
		5	7.92	7.82	8.02	79291700

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: REMI01

Lab Code: ACE SDG NO.: Q3742

Instrument ID: ECD_D Date(s) Analyzed: 11/14/2025 11/14/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.46	5.36	5.56	348471000
		2	5.64	5.54	5.74	231072000
		3	6.74	6.64	6.84	1091020000
		4	7.19	7.09	7.29	705058000
		5	7.32	7.22	7.42	442502000

A
B
C
D
E
F
G
H
I
J
K
L

Contract: REMI01

SDG NO.: Q3742

Calibration Date(s): 12/08/2025 12/08/2025

Calibration Times: **15:09** **16:03**

LAB FILE ID:	RT 100 =	<u>PD091481.D</u>	RT 075 =	<u>PD091482.D</u>
	RT 050 =	PD091483.D	RT 025 =	PD091484.D
			RT 005 =	PD091485.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Calibration Times: 15:09 16:03

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 100 =	<u>PD091481.D</u>	RT 075 =	<u>PD091482.D</u>
RT 050 = PD091483.D	RT 025 =	PD091484.D	RT 005 =	PD091485.D

[illegible]

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_D

Contract: REMI01
SDG NO.: Q3742

Calibration Date(s): 12/08/2025 12/08/2025
Calibration Times: 15:09 16:03

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD091481.D</u>	CF 075 =	<u>PD091482.D</u>		
CF 050 =		<u>PD091483.D</u>	CF 025 =	<u>PD091484.D</u>	CF 005 =	<u>PD091485.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	4579120000	4639720000	4523160000	4555750000	4716360000	4602820000	2
4,4'-DDE	5779280000	5892820000	5756320000	5753890000	5864650000	5809390000	1
4,4'-DDT	5121650000	5175250000	5048960000	5090770000	5345670000	5156460000	2
Aldrin	6819680000	7059090000	6886500000	6926640000	7097310000	6957840000	2
alpha-BHC	8353280000	8376480000	8051960000	7877410000	7595380000	8050900000	4
alpha-Chlordane	6287560000	6438540000	6369880000	6674420000	7810600000	6716200000	9
beta-BHC	2693260000	2773490000	2767330000	2932590000	3268690000	2887070000	8
Decachlorobiphenyl	4683640000	4854830000	4922070000	5372580000	6157000000	5198020000	11
delta-BHC	6218880000	6272710000	5977640000	5871060000	5571050000	5982270000	5
Dieldrin	6427120000	6542330000	6414320000	6454450000	6695670000	6506780000	2
Endosulfan I	5840790000	5990310000	5931890000	6120850000	6665110000	6109790000	5
Endosulfan II	5362960000	5505160000	5533320000	5965250000	8037190000	6080780000	18
Endosulfan sulfate	5050450000	5142130000	5111170000	5303530000	5788930000	5279240000	6
Endrin	5557100000	5635030000	5505600000	5563010000	5780280000	5608200000	2
Endrin aldehyde	3894900000	4000290000	4002290000	4219360000	4770070000	4177380000	8
Endrin ketone	5404130000	5495010000	5447080000	5659030000	6164150000	5633880000	6
gamma-BHC (Lindane)	7671910000	7735050000	7465630000	7375770000	7330240000	7515720000	2
gamma-Chlordane	6227370000	6372860000	6222700000	6304590000	6757580000	6377020000	3
Heptachlor	6952400000	7072060000	6917500000	6987030000	7565530000	7098900000	4
Heptachlor epoxide	5883560000	6080570000	5982150000	6343250000	6672070000	6192320000	5
Methoxychlor	2493520000	2555740000	2569740000	2728230000	3092680000	2687980000	9
Tetrachloro-m-xylene	3251060000	3343560000	3349890000	3549710000	3872320000	3473310000	7

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_D

Contract: REMI01
SDG NO.: Q3742

Calibration Date(s): 12/08/2025 12/08/2025
Calibration Times: 15:09 16:03

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD091481.D</u>	CF 075 =	<u>PD091482.D</u>		
CF 050 =		<u>PD091483.D</u>	CF 025 =	<u>PD091484.D</u>	CF 005 =	<u>PD091485.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	40563700000	42868900000	43750200000	47169200000	53085700000	45487500000	11
4,4'-DDE	48787600000	51617900000	52713500000	56864300000	64054400000	54807500000	11
4,4'-DDT	43900100000	46021200000	46721600000	48760000000	51590700000	47398700000	6
Aldrin	51907000000	54725200000	55832800000	60028000000	67438000000	57986200000	10
alpha-BHC	63667000000	65476500000	65662500000	69227800000	76834900000	68173700000	8
alpha-Chlordane	47953600000	50307400000	51451000000	55287900000	62754900000	53551000000	11
beta-BHC	21190200000	22242700000	22924300000	24571000000	28867000000	23959000000	13
Decachlorobiphenyl	38176600000	39143100000	39996700000	41508800000	48228700000	41410800000	10
delta-BHC	55002700000	57028900000	57836400000	61224100000	68342200000	59886900000	9
Dieldrin	49241300000	52176000000	53646800000	57602500000	64653700000	55464000000	11
Endosulfan I	42921600000	45232500000	47126000000	51141000000	58523800000	48989000000	13
Endosulfan II	41394600000	43750200000	45159600000	48122500000	57145300000	47114400000	13
Endosulfan sulfate	41155800000	43302700000	44767900000	46802000000	54439800000	46093700000	11
Endrin	45317500000	47843400000	49487500000	53828200000	64913700000	52278100000	15
Endrin aldehyde	30861500000	32780400000	33996700000	36384500000	44432300000	35691100000	15
Endrin ketone	43085500000	45422600000	47390000000	48876700000	58439800000	48642900000	12
gamma-BHC (Lindane)	51269700000	53765100000	54102400000	57888300000	64120200000	56229200000	9
gamma-Chlordane	50182800000	52763300000	53670800000	57546600000	64941100000	55820900000	10
Heptachlor	51929700000	54754900000	56222500000	60275100000	68237000000	58283800000	11
Heptachlor epoxide	44426000000	47110900000	48711500000	52646000000	60090500000	50597000000	12
Methoxychlor	20918300000	22241400000	23265200000	24000900000	27564500000	23598100000	11
Tetrachloro-m-xylene	35806100000	36994700000	37706700000	40254600000	46254000000	39403200000	11

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance
Contract: REMI01

Lab Code: ACE
SDG NO.: Q3742

Instrument ID: ECD_D
Date(s) Analyzed: 12/08/2025 12/08/2025

GC Column: ZB-MR1
ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.23	6.13	6.33	37811600
		2	6.43	6.33	6.53	55917700
		3	7.14	7.04	7.24	103526000
		4	7.55	7.45	7.65	133679000
		5	7.92	7.82	8.02	76866900

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance
Contract: REMI01

Lab Code: ACE
SDG NO.: Q3742

Instrument ID: ECD_D
Date(s) Analyzed: 12/08/2025 12/08/2025

GC Column: ZB-MR2
ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.46	5.36	5.56	318132000
		2	5.63	5.53	5.73	218900000
		3	6.74	6.64	6.84	1071720000
		4	7.18	7.08	7.28	690678000
		5	7.31	7.21	7.41	511709000

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Continuing Calib Date: 12/02/2025

Initial Calibration Date(s): 11/14/2025

11/14/2025

Continuing Calib Time: 12:44

Initial Calibration Time(s): 20:56

21:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.07	9.06	8.96	9.16	-0.01
Tetrachloro-m-xylene	3.55	3.54	3.44	3.64	-0.01
alpha-BHC	4.00	3.99	3.89	4.09	-0.01
beta-BHC	4.52	4.51	4.41	4.61	-0.01
delta-BHC	4.77	4.76	4.66	4.86	0.00
gamma-BHC (Lindane)	4.33	4.32	4.22	4.42	-0.01
Heptachlor	4.93	4.92	4.82	5.02	-0.01
Aldrin	5.27	5.26	5.16	5.36	-0.01
Heptachlor epoxide	5.69	5.68	5.58	5.78	-0.01
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.35	6.34	6.24	6.44	-0.01
4,4'-DDE	6.20	6.19	6.09	6.29	-0.01
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.79	6.78	6.68	6.88	-0.01
4,4'-DDD	6.71	6.70	6.60	6.80	0.00
Endosulfan sulfate	7.15	7.14	7.04	7.24	-0.01
4,4'-DDT	7.02	7.01	6.91	7.11	-0.01
Methoxychlor	7.49	7.49	7.39	7.59	0.00
Endrin ketone	7.63	7.62	7.52	7.72	-0.01
Endrin aldehyde	6.92	6.91	6.81	7.01	0.00
alpha-Chlordane	6.03	6.02	5.92	6.12	-0.01
gamma-Chlordane	5.95	5.94	5.84	6.04	-0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Continuing Calib Date: 12/02/2025

Initial Calibration Date(s): 11/14/2025

11/14/2025

Continuing Calib Time: 12:44

Initial Calibration Time(s): 20:56

21:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.87	2.77	2.97	0.00
alpha-BHC	3.38	3.39	3.29	3.49	0.01
beta-BHC	4.02	4.02	3.92	4.12	0.00
delta-BHC	4.25	4.25	4.15	4.35	0.00
gamma-BHC (Lindane)	3.72	3.72	3.62	3.82	0.00
Heptachlor	4.07	4.07	3.97	4.17	0.00
Aldrin	4.36	4.36	4.26	4.46	0.00
Heptachlor epoxide	4.86	4.86	4.76	4.96	0.00
Endosulfan I	5.23	5.24	5.14	5.34	0.01
Dieldrin	5.50	5.50	5.40	5.60	0.00
4,4'-DDE	5.36	5.36	5.26	5.46	0.00
Endrin	5.78	5.78	5.68	5.88	0.00
Endosulfan II	6.07	6.07	5.97	6.17	0.00
4,4'-DDD	5.92	5.92	5.82	6.02	0.00
Endosulfan sulfate	6.47	6.47	6.37	6.57	0.00
4,4'-DDT	6.17	6.17	6.07	6.27	0.00
Methoxychlor	6.74	6.74	6.64	6.84	0.00
Endrin ketone	6.98	6.98	6.88	7.08	0.00
Endrin aldehyde	6.25	6.25	6.15	6.35	0.00
alpha-Chlordane	5.18	5.18	5.08	5.28	0.00
gamma-Chlordane	5.11	5.11	5.01	5.21	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL01 **Date Analyzed:** 12/02/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091418.D **Time Analyzed:** 12:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.705	6.599	6.799	45.570	50.000	-8.9
4,4'-DDE	6.195	6.089	6.289	46.680	50.000	-6.6
4,4'-DDT	7.021	6.914	7.114	49.110	50.000	-1.8
Aldrin	5.268	5.163	5.363	46.850	50.000	-6.3
alpha-BHC	3.998	3.893	4.093	47.920	50.000	-4.2
alpha-Chlordane	6.025	5.919	6.119	44.400	50.000	-11.2
beta-BHC	4.517	4.411	4.611	45.910	50.000	-8.2
Decachlorobiphenyl	9.071	8.963	9.163	42.490	50.000	-15.0
delta-BHC	4.765	4.660	4.860	47.460	50.000	-5.1
Dieldrin	6.346	6.239	6.439	47.200	50.000	-5.6
Endosulfan I	6.073	5.966	6.166	46.560	50.000	-6.9
Endosulfan II	6.786	6.679	6.879	46.780	50.000	-6.4
Endosulfan sulfate	7.150	7.043	7.243	45.390	50.000	-9.2
Endrin	6.573	6.466	6.666	49.290	50.000	-1.4
Endrin aldehyde	6.915	6.808	7.008	45.090	50.000	-9.8
Endrin ketone	7.631	7.524	7.724	45.190	50.000	-9.6
gamma-BHC (Lindane)	4.330	4.224	4.424	47.510	50.000	-5.0
gamma-Chlordane	5.945	5.838	6.038	47.390	50.000	-5.2
Heptachlor	4.927	4.821	5.021	47.410	50.000	-5.2
Heptachlor epoxide	5.689	5.582	5.782	46.310	50.000	-7.4
Methoxychlor	7.494	7.387	7.587	46.480	50.000	-7.0
Tetrachloro-m-xylene	3.548	3.444	3.644	45.560	50.000	-8.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL01 **Date Analyzed:** 12/02/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091418.D **Time Analyzed:** 12:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.917	5.817	6.017	50.880	50.000	1.8
4,4'-DDE	5.362	5.263	5.463	53.810	50.000	7.6
4,4'-DDT	6.171	6.071	6.271	57.840	50.000	15.7
Aldrin	4.357	4.258	4.458	53.660	50.000	7.3
alpha-BHC	3.383	3.286	3.486	53.080	50.000	6.2
alpha-Chlordane	5.178	5.079	5.279	54.100	50.000	8.2
beta-BHC	4.016	3.918	4.118	52.030	50.000	4.1
Decachlorobiphenyl	8.059	7.958	8.158	48.120	50.000	-3.8
delta-BHC	4.252	4.153	4.353	52.840	50.000	5.7
Dieldrin	5.501	5.401	5.601	53.390	50.000	6.8
Endosulfan I	5.234	5.135	5.335	52.610	50.000	5.2
Endosulfan II	6.069	5.969	6.169	53.350	50.000	6.7
Endosulfan sulfate	6.471	6.371	6.571	53.030	50.000	6.1
Endrin	5.777	5.678	5.878	55.700	50.000	11.4
Endrin aldehyde	6.248	6.148	6.348	51.900	50.000	3.8
Endrin ketone	6.980	6.880	7.080	51.440	50.000	2.9
gamma-BHC (Lindane)	3.719	3.621	3.821	52.890	50.000	5.8
gamma-Chlordane	5.113	5.014	5.214	54.400	50.000	8.8
Heptachlor	4.071	3.973	4.173	52.420	50.000	4.8
Heptachlor epoxide	4.860	4.762	4.962	53.100	50.000	6.2
Methoxychlor	6.742	6.641	6.841	53.400	50.000	6.8
Tetrachloro-m-xylene	2.871	2.774	2.974	52.030	50.000	4.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Continuing Calib Date: 12/02/2025

Initial Calibration Date(s): 11/14/2025

11/14/2025

Continuing Calib Time: 16:39

Initial Calibration Time(s): 20:56

21:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
alpha-BHC	3.99	3.99	3.89	4.09	0.00
beta-BHC	4.51	4.51	4.41	4.61	0.00
delta-BHC	4.76	4.76	4.66	4.86	0.00
gamma-BHC (Lindane)	4.32	4.32	4.22	4.42	0.00
Heptachlor	4.92	4.92	4.82	5.02	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.78	6.78	6.68	6.88	0.00
4,4'-DDD	6.70	6.70	6.60	6.80	0.00
Endosulfan sulfate	7.14	7.14	7.04	7.24	0.00
4,4'-DDT	7.01	7.01	6.91	7.11	0.00
Methoxychlor	7.49	7.49	7.39	7.59	0.00
Endrin ketone	7.62	7.62	7.52	7.72	0.00
Endrin aldehyde	6.91	6.91	6.81	7.01	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Continuing Calib Date: 12/02/2025

Initial Calibration Date(s): 11/14/2025

11/14/2025

Continuing Calib Time: 16:39

Initial Calibration Time(s): 20:56

21:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.87	2.77	2.97	0.00
alpha-BHC	3.38	3.39	3.29	3.49	0.01
beta-BHC	4.02	4.02	3.92	4.12	0.00
delta-BHC	4.25	4.25	4.15	4.35	0.00
gamma-BHC (Lindane)	3.72	3.72	3.62	3.82	0.00
Heptachlor	4.07	4.07	3.97	4.17	0.00
Aldrin	4.36	4.36	4.26	4.46	0.00
Heptachlor epoxide	4.86	4.86	4.76	4.96	0.00
Endosulfan I	5.23	5.24	5.14	5.34	0.01
Dieldrin	5.50	5.50	5.40	5.60	0.00
4,4'-DDE	5.36	5.36	5.26	5.46	0.00
Endrin	5.78	5.78	5.68	5.88	0.00
Endosulfan II	6.07	6.07	5.97	6.17	0.00
4,4'-DDD	5.92	5.92	5.82	6.02	0.00
Endosulfan sulfate	6.47	6.47	6.37	6.57	0.00
4,4'-DDT	6.17	6.17	6.07	6.27	0.00
Methoxychlor	6.74	6.74	6.64	6.84	0.00
Endrin ketone	6.98	6.98	6.88	7.08	0.00
Endrin aldehyde	6.25	6.25	6.15	6.35	0.00
alpha-Chlordane	5.18	5.18	5.08	5.28	0.00
gamma-Chlordane	5.11	5.11	5.01	5.21	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL02 **Date Analyzed:** 12/02/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091426.D **Time Analyzed:** 16:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.698	6.599	6.799	44.440	50.000	-11.1
4,4'-DDE	6.187	6.089	6.289	45.000	50.000	-10.0
4,4'-DDT	7.014	6.914	7.114	45.890	50.000	-8.2
Aldrin	5.261	5.163	5.363	46.020	50.000	-8.0
alpha-BHC	3.991	3.893	4.093	47.590	50.000	-4.8
alpha-Chlordane	6.017	5.919	6.119	43.170	50.000	-13.7
beta-BHC	4.509	4.411	4.611	45.680	50.000	-8.6
Decachlorobiphenyl	9.064	8.963	9.163	40.930	50.000	-18.1
delta-BHC	4.758	4.660	4.860	47.180	50.000	-5.6
Dieldrin	6.338	6.239	6.439	45.510	50.000	-9.0
Endosulfan I	6.065	5.966	6.166	44.980	50.000	-10.0
Endosulfan II	6.778	6.679	6.879	44.780	50.000	-10.4
Endosulfan sulfate	7.142	7.043	7.243	43.570	50.000	-12.9
Endrin	6.565	6.466	6.666	47.890	50.000	-4.2
Endrin aldehyde	6.907	6.808	7.008	42.870	50.000	-14.3
Endrin ketone	7.623	7.524	7.724	43.740	50.000	-12.5
gamma-BHC (Lindane)	4.321	4.224	4.424	47.230	50.000	-5.5
gamma-Chlordane	5.937	5.838	6.038	46.000	50.000	-8.0
Heptachlor	4.919	4.821	5.021	45.680	50.000	-8.6
Heptachlor epoxide	5.681	5.582	5.782	45.200	50.000	-9.6
Methoxychlor	7.487	7.387	7.587	43.190	50.000	-13.6
Tetrachloro-m-xylene	3.541	3.444	3.644	46.040	50.000	-7.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL02 **Date Analyzed:** 12/02/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091426.D **Time Analyzed:** 16:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.916	5.817	6.017	50.280	50.000	0.6
4,4'-DDE	5.360	5.263	5.463	52.360	50.000	4.7
4,4'-DDT	6.170	6.071	6.271	53.700	50.000	7.4
Aldrin	4.356	4.258	4.458	53.520	50.000	7.0
alpha-BHC	3.383	3.286	3.486	53.430	50.000	6.9
alpha-Chlordane	5.178	5.079	5.279	52.980	50.000	6.0
beta-BHC	4.016	3.918	4.118	52.290	50.000	4.6
Decachlorobiphenyl	8.058	7.958	8.158	44.240	50.000	-11.5
delta-BHC	4.252	4.153	4.353	53.010	50.000	6.0
Dieldrin	5.500	5.401	5.601	52.390	50.000	4.8
Endosulfan I	5.234	5.135	5.335	51.440	50.000	2.9
Endosulfan II	6.068	5.969	6.169	51.310	50.000	2.6
Endosulfan sulfate	6.470	6.371	6.571	50.390	50.000	0.8
Endrin	5.776	5.678	5.878	55.300	50.000	10.6
Endrin aldehyde	6.246	6.148	6.348	49.180	50.000	-1.6
Endrin ketone	6.979	6.880	7.080	48.270	50.000	-3.5
gamma-BHC (Lindane)	3.719	3.621	3.821	53.250	50.000	6.5
gamma-Chlordane	5.113	5.014	5.214	53.280	50.000	6.6
Heptachlor	4.071	3.973	4.173	52.430	50.000	4.9
Heptachlor epoxide	4.860	4.762	4.962	52.480	50.000	5.0
Methoxychlor	6.741	6.641	6.841	50.390	50.000	0.8
Tetrachloro-m-xylene	2.872	2.774	2.974	52.510	50.000	5.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Continuing Calib Date: 12/02/2025

Initial Calibration Date(s): 11/14/2025

11/14/2025

Continuing Calib Time: 19:12

Initial Calibration Time(s): 20:56

21:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
alpha-BHC	3.99	3.99	3.89	4.09	0.00
beta-BHC	4.51	4.51	4.41	4.61	0.00
delta-BHC	4.76	4.76	4.66	4.86	0.00
gamma-BHC (Lindane)	4.32	4.32	4.22	4.42	0.00
Heptachlor	4.92	4.92	4.82	5.02	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.78	6.78	6.68	6.88	0.00
4,4'-DDD	6.70	6.70	6.60	6.80	0.00
Endosulfan sulfate	7.14	7.14	7.04	7.24	0.00
4,4'-DDT	7.01	7.01	6.91	7.11	0.00
Methoxychlor	7.49	7.49	7.39	7.59	0.00
Endrin ketone	7.62	7.62	7.52	7.72	0.00
Endrin aldehyde	6.91	6.91	6.81	7.01	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Continuing Calib Date: 12/02/2025

Initial Calibration Date(s): 11/14/2025

11/14/2025

Continuing Calib Time: 19:12

Initial Calibration Time(s): 20:56

21:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.87	2.77	2.97	0.00
alpha-BHC	3.38	3.39	3.29	3.49	0.01
beta-BHC	4.02	4.02	3.92	4.12	0.00
delta-BHC	4.25	4.25	4.15	4.35	0.00
gamma-BHC (Lindane)	3.72	3.72	3.62	3.82	0.00
Heptachlor	4.07	4.07	3.97	4.17	0.00
Aldrin	4.36	4.36	4.26	4.46	0.00
Heptachlor epoxide	4.86	4.86	4.76	4.96	0.00
Endosulfan I	5.23	5.24	5.14	5.34	0.01
Dieldrin	5.50	5.50	5.40	5.60	0.00
4,4'-DDE	5.36	5.36	5.26	5.46	0.00
Endrin	5.78	5.78	5.68	5.88	0.00
Endosulfan II	6.07	6.07	5.97	6.17	0.00
4,4'-DDD	5.92	5.92	5.82	6.02	0.00
Endosulfan sulfate	6.47	6.47	6.37	6.57	0.00
4,4'-DDT	6.17	6.17	6.07	6.27	0.00
Methoxychlor	6.74	6.74	6.64	6.84	0.00
Endrin ketone	6.98	6.98	6.88	7.08	0.00
Endrin aldehyde	6.25	6.25	6.15	6.35	0.00
alpha-Chlordane	5.18	5.18	5.08	5.28	0.00
gamma-Chlordane	5.11	5.11	5.01	5.21	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL03 **Date Analyzed:** 12/02/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091434.D **Time Analyzed:** 19:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.698	6.599	6.799	49.770	50.000	-0.5
4,4'-DDE	6.188	6.089	6.289	50.340	50.000	0.7
4,4'-DDT	7.014	6.914	7.114	51.070	50.000	2.1
Aldrin	5.261	5.163	5.363	50.920	50.000	1.8
alpha-BHC	3.991	3.893	4.093	51.990	50.000	4.0
alpha-Chlordane	6.018	5.919	6.119	50.030	50.000	0.1
beta-BHC	4.510	4.411	4.611	49.750	50.000	-0.5
Decachlorobiphenyl	9.063	8.963	9.163	45.580	50.000	-8.8
delta-BHC	4.758	4.660	4.860	52.450	50.000	4.9
Dieldrin	6.338	6.239	6.439	50.690	50.000	1.4
Endosulfan I	6.065	5.966	6.166	50.060	50.000	0.1
Endosulfan II	6.779	6.679	6.879	48.490	50.000	-3.0
Endosulfan sulfate	7.142	7.043	7.243	48.880	50.000	-2.2
Endrin	6.566	6.466	6.666	53.130	50.000	6.3
Endrin aldehyde	6.908	6.808	7.008	47.960	50.000	-4.1
Endrin ketone	7.623	7.524	7.724	49.330	50.000	-1.3
gamma-BHC (Lindane)	4.322	4.224	4.424	51.710	50.000	3.4
gamma-Chlordane	5.937	5.838	6.038	50.740	50.000	1.5
Heptachlor	4.919	4.821	5.021	50.210	50.000	0.4
Heptachlor epoxide	5.682	5.582	5.782	51.340	50.000	2.7
Methoxychlor	7.487	7.387	7.587	48.330	50.000	-3.3
Tetrachloro-m-xylene	3.542	3.444	3.644	49.450	50.000	-1.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL03 **Date Analyzed:** 12/02/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091434.D **Time Analyzed:** 19:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.916	5.817	6.017	53.790	50.000	7.6
4,4'-DDE	5.362	5.263	5.463	56.070	50.000	12.1
4,4'-DDT	6.168	6.071	6.271	56.860	50.000	13.7
Aldrin	4.356	4.258	4.458	56.540	50.000	13.1
alpha-BHC	3.383	3.286	3.486	56.260	50.000	12.5
alpha-Chlordane	5.177	5.079	5.279	56.340	50.000	12.7
beta-BHC	4.016	3.918	4.118	55.090	50.000	10.2
Decachlorobiphenyl	8.058	7.958	8.158	51.040	50.000	2.1
delta-BHC	4.252	4.153	4.353	55.990	50.000	12.0
Dieldrin	5.499	5.401	5.601	55.940	50.000	11.9
Endosulfan I	5.234	5.135	5.335	54.820	50.000	9.6
Endosulfan II	6.068	5.969	6.169	56.730	50.000	13.5
Endosulfan sulfate	6.469	6.371	6.571	50.500	50.000	1.0
Endrin	5.775	5.678	5.878	59.630	50.000	19.3
Endrin aldehyde	6.246	6.148	6.348	54.610	50.000	9.2
Endrin ketone	6.979	6.880	7.080	53.580	50.000	7.2
gamma-BHC (Lindane)	3.719	3.621	3.821	56.180	50.000	12.4
gamma-Chlordane	5.112	5.014	5.214	57.000	50.000	14.0
Heptachlor	4.071	3.973	4.173	54.500	50.000	9.0
Heptachlor epoxide	4.860	4.762	4.962	55.850	50.000	11.7
Methoxychlor	6.740	6.641	6.841	54.300	50.000	8.6
Tetrachloro-m-xylene	2.873	2.774	2.974	55.230	50.000	10.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Continuing Calib Date: 12/02/2025

Initial Calibration Date(s): 11/14/2025

11/14/2025

Continuing Calib Time: 21:55

Initial Calibration Time(s): 20:56

21:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
alpha-BHC	3.99	3.99	3.89	4.09	0.00
beta-BHC	4.51	4.51	4.41	4.61	0.00
delta-BHC	4.76	4.76	4.66	4.86	0.00
gamma-BHC (Lindane)	4.32	4.32	4.22	4.42	0.00
Heptachlor	4.92	4.92	4.82	5.02	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.78	6.78	6.68	6.88	0.00
4,4'-DDD	6.70	6.70	6.60	6.80	0.00
Endosulfan sulfate	7.14	7.14	7.04	7.24	0.00
4,4'-DDT	7.02	7.01	6.91	7.11	0.00
Methoxychlor	7.49	7.49	7.39	7.59	0.00
Endrin ketone	7.62	7.62	7.52	7.72	0.00
Endrin aldehyde	6.91	6.91	6.81	7.01	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Continuing Calib Date: 12/02/2025

Initial Calibration Date(s): 11/14/2025

11/14/2025

Continuing Calib Time: 21:55

Initial Calibration Time(s): 20:56

21:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.87	2.77	2.97	0.00
alpha-BHC	3.38	3.39	3.29	3.49	0.01
beta-BHC	4.02	4.02	3.92	4.12	0.00
delta-BHC	4.25	4.25	4.15	4.35	0.00
gamma-BHC (Lindane)	3.72	3.72	3.62	3.82	0.00
Heptachlor	4.07	4.07	3.97	4.17	0.00
Aldrin	4.36	4.36	4.26	4.46	0.00
Heptachlor epoxide	4.86	4.86	4.76	4.96	0.00
Endosulfan I	5.24	5.24	5.14	5.34	0.00
Dieldrin	5.50	5.50	5.40	5.60	0.00
4,4'-DDE	5.36	5.36	5.26	5.46	0.00
Endrin	5.78	5.78	5.68	5.88	0.00
Endosulfan II	6.07	6.07	5.97	6.17	0.00
4,4'-DDD	5.92	5.92	5.82	6.02	0.00
Endosulfan sulfate	6.47	6.47	6.37	6.57	0.00
4,4'-DDT	6.17	6.17	6.07	6.27	0.00
Methoxychlor	6.74	6.74	6.64	6.84	0.00
Endrin ketone	6.98	6.98	6.88	7.08	0.00
Endrin aldehyde	6.25	6.25	6.15	6.35	0.00
alpha-Chlordane	5.18	5.18	5.08	5.28	0.00
gamma-Chlordane	5.11	5.11	5.01	5.21	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL04 **Date Analyzed:** 12/02/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091446.D **Time Analyzed:** 21:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.698	6.599	6.799	46.020	50.000	-8.0
4,4'-DDE	6.188	6.089	6.289	46.980	50.000	-6.0
4,4'-DDT	7.015	6.914	7.114	46.950	50.000	-6.1
Aldrin	5.262	5.163	5.363	47.590	50.000	-4.8
alpha-BHC	3.992	3.893	4.093	48.690	50.000	-2.6
alpha-Chlordane	6.018	5.919	6.119	45.560	50.000	-8.9
beta-BHC	4.510	4.411	4.611	46.800	50.000	-6.4
Decachlorobiphenyl	9.064	8.963	9.163	42.580	50.000	-14.8
delta-BHC	4.758	4.660	4.860	48.830	50.000	-2.3
Dieldrin	6.339	6.239	6.439	47.410	50.000	-5.2
Endosulfan I	6.066	5.966	6.166	46.810	50.000	-6.4
Endosulfan II	6.779	6.679	6.879	46.530	50.000	-6.9
Endosulfan sulfate	7.144	7.043	7.243	45.500	50.000	-9.0
Endrin	6.566	6.466	6.666	49.270	50.000	-1.5
Endrin aldehyde	6.909	6.808	7.008	45.000	50.000	-10.0
Endrin ketone	7.622	7.524	7.724	44.870	50.000	-10.3
gamma-BHC (Lindane)	4.323	4.224	4.424	48.530	50.000	-2.9
gamma-Chlordane	5.936	5.838	6.038	47.120	50.000	-5.8
Heptachlor	4.920	4.821	5.021	46.270	50.000	-7.5
Heptachlor epoxide	5.682	5.582	5.782	47.110	50.000	-5.8
Methoxychlor	7.488	7.387	7.587	43.900	50.000	-12.2
Tetrachloro-m-xylene	3.543	3.444	3.644	46.630	50.000	-6.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL04 **Date Analyzed:** 12/02/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091446.D **Time Analyzed:** 21:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.917	5.817	6.017	49.780	50.000	-0.4
4,4'-DDE	5.361	5.263	5.463	51.790	50.000	3.6
4,4'-DDT	6.171	6.071	6.271	53.360	50.000	6.7
Aldrin	4.357	4.258	4.458	52.900	50.000	5.8
alpha-BHC	3.384	3.286	3.486	52.690	50.000	5.4
alpha-Chlordane	5.178	5.079	5.279	52.490	50.000	5.0
beta-BHC	4.016	3.918	4.118	51.790	50.000	3.6
Decachlorobiphenyl	8.058	7.958	8.158	46.070	50.000	-7.9
delta-BHC	4.252	4.153	4.353	52.440	50.000	4.9
Dieldrin	5.500	5.401	5.601	52.230	50.000	4.5
Endosulfan I	5.235	5.135	5.335	51.090	50.000	2.2
Endosulfan II	6.069	5.969	6.169	51.690	50.000	3.4
Endosulfan sulfate	6.470	6.371	6.571	49.120	50.000	-1.8
Endrin	5.778	5.678	5.878	55.300	50.000	10.6
Endrin aldehyde	6.247	6.148	6.348	50.150	50.000	0.3
Endrin ketone	6.980	6.880	7.080	49.530	50.000	-0.9
gamma-BHC (Lindane)	3.720	3.621	3.821	52.570	50.000	5.1
gamma-Chlordane	5.113	5.014	5.214	53.070	50.000	6.1
Heptachlor	4.072	3.973	4.173	50.410	50.000	0.8
Heptachlor epoxide	4.861	4.762	4.962	52.120	50.000	4.2
Methoxychlor	6.741	6.641	6.841	49.760	50.000	-0.5
Tetrachloro-m-xylene	2.873	2.774	2.974	51.650	50.000	3.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Continuing Calib Date: 12/09/2025

Initial Calibration Date(s): 12/08/2025

12/08/2025

Continuing Calib Time: 14:05

Initial Calibration Time(s): 15:09

16:03

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
alpha-BHC	3.99	3.99	3.89	4.09	0.00
beta-BHC	4.51	4.51	4.41	4.61	0.00
delta-BHC	4.76	4.76	4.66	4.86	0.00
gamma-BHC (Lindane)	4.32	4.32	4.22	4.42	0.00
Heptachlor	4.92	4.92	4.82	5.02	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.06	5.96	6.16	-0.01
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.56	6.46	6.66	-0.01
Endosulfan II	6.78	6.78	6.68	6.88	0.00
4,4'-DDD	6.70	6.70	6.60	6.80	0.00
Endosulfan sulfate	7.14	7.14	7.04	7.24	0.00
4,4'-DDT	7.02	7.01	6.91	7.11	0.00
Methoxychlor	7.49	7.48	7.38	7.58	-0.01
Endrin ketone	7.63	7.62	7.52	7.72	-0.01
Endrin aldehyde	6.91	6.91	6.81	7.01	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.93	5.83	6.03	-0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Continuing Calib Date: 12/09/2025

Initial Calibration Date(s): 12/08/2025

12/08/2025

Continuing Calib Time: 14:05

Initial Calibration Time(s): 15:09

16:03

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.87	2.77	2.97	0.00
alpha-BHC	3.38	3.38	3.28	3.48	0.00
beta-BHC	4.02	4.01	3.91	4.11	0.00
delta-BHC	4.25	4.25	4.15	4.35	0.00
gamma-BHC (Lindane)	3.72	3.72	3.62	3.82	0.00
Heptachlor	4.07	4.07	3.97	4.17	0.00
Aldrin	4.35	4.35	4.25	4.45	0.00
Heptachlor epoxide	4.86	4.86	4.76	4.96	0.00
Endosulfan I	5.23	5.23	5.13	5.33	0.00
Dieldrin	5.50	5.50	5.40	5.60	0.00
4,4'-DDE	5.36	5.36	5.26	5.46	0.00
Endrin	5.78	5.77	5.67	5.87	-0.01
Endosulfan II	6.07	6.07	5.97	6.17	0.00
4,4'-DDD	5.92	5.91	5.81	6.01	-0.01
Endosulfan sulfate	6.47	6.47	6.37	6.57	0.00
4,4'-DDT	6.17	6.17	6.07	6.27	0.00
Methoxychlor	6.74	6.74	6.64	6.84	0.00
Endrin ketone	6.98	6.98	6.88	7.08	0.00
Endrin aldehyde	6.25	6.24	6.14	6.34	-0.01
alpha-Chlordane	5.18	5.18	5.08	5.28	0.00
gamma-Chlordane	5.11	5.11	5.01	5.21	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/08/2025 12/08/2025

Client Sample No.: CCAL05 **Date Analyzed:** 12/09/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091525.D **Time Analyzed:** 14:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.700	6.595	6.795	51.400	50.000	2.8
4,4'-DDE	6.190	6.085	6.285	47.060	50.000	-5.9
4,4'-DDT	7.015	6.911	7.111	46.830	50.000	-6.3
Aldrin	5.263	5.158	5.358	49.330	50.000	-1.3
alpha-BHC	3.994	3.889	4.089	47.500	50.000	-5.0
alpha-Chlordane	6.020	5.915	6.115	46.360	50.000	-7.3
beta-BHC	4.511	4.408	4.608	46.230	50.000	-7.5
Decachlorobiphenyl	9.064	8.957	9.157	45.660	50.000	-8.7
delta-BHC	4.760	4.656	4.856	56.240	50.000	12.5
Dieldrin	6.340	6.235	6.435	47.720	50.000	-4.6
Endosulfan I	6.067	5.963	6.163	47.490	50.000	-5.0
Endosulfan II	6.779	6.676	6.876	42.290	50.000	-15.4
Endosulfan sulfate	7.144	7.040	7.240	46.370	50.000	-7.3
Endrin	6.567	6.463	6.663	48.250	50.000	-3.5
Endrin aldehyde	6.910	6.805	7.005	48.440	50.000	-3.1
Endrin ketone	7.625	7.521	7.721	46.690	50.000	-6.6
gamma-BHC (Lindane)	4.324	4.220	4.420	47.610	50.000	-4.8
gamma-Chlordane	5.940	5.834	6.034	48.440	50.000	-3.1
Heptachlor	4.922	4.817	5.017	48.830	50.000	-2.3
Heptachlor epoxide	5.684	5.579	5.779	48.240	50.000	-3.5
Methoxychlor	7.489	7.384	7.584	45.370	50.000	-9.3
Tetrachloro-m-xylene	3.544	3.440	3.640	45.490	50.000	-9.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/08/2025 12/08/2025

Client Sample No.: CCAL05 **Date Analyzed:** 12/09/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091525.D **Time Analyzed:** 14:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.916	5.814	6.014	52.190	50.000	4.4
4,4'-DDE	5.360	5.260	5.460	51.410	50.000	2.8
4,4'-DDT	6.170	6.068	6.268	52.560	50.000	5.1
Aldrin	4.353	4.254	4.454	51.550	50.000	3.1
alpha-BHC	3.381	3.281	3.481	49.490	50.000	-1.0
alpha-Chlordane	5.176	5.076	5.276	52.250	50.000	4.5
beta-BHC	4.015	3.914	4.114	50.940	50.000	1.9
Decachlorobiphenyl	8.056	7.955	8.155	51.800	50.000	3.6
delta-BHC	4.250	4.150	4.350	51.690	50.000	3.4
Dieldrin	5.499	5.398	5.598	51.820	50.000	3.6
Endosulfan I	5.233	5.132	5.332	43.540	50.000	-12.9
Endosulfan II	6.067	5.966	6.166	51.720	50.000	3.4
Endosulfan sulfate	6.469	6.367	6.567	51.500	50.000	3.0
Endrin	5.776	5.674	5.874	51.890	50.000	3.8
Endrin aldehyde	6.246	6.144	6.344	51.210	50.000	2.4
Endrin ketone	6.978	6.876	7.076	51.600	50.000	3.2
gamma-BHC (Lindane)	3.718	3.617	3.817	54.430	50.000	8.9
gamma-Chlordane	5.112	5.011	5.211	52.990	50.000	6.0
Heptachlor	4.070	3.969	4.169	51.840	50.000	3.7
Heptachlor epoxide	4.859	4.758	4.958	52.390	50.000	4.8
Methoxychlor	6.740	6.638	6.838	52.010	50.000	4.0
Tetrachloro-m-xylene	2.870	2.770	2.970	48.630	50.000	-2.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Continuing Calib Date: 12/09/2025

Initial Calibration Date(s): 12/08/2025

12/08/2025

Continuing Calib Time: 16:47

Initial Calibration Time(s): 15:09

16:03

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
alpha-BHC	3.99	3.99	3.89	4.09	0.00
beta-BHC	4.51	4.51	4.41	4.61	0.00
delta-BHC	4.76	4.76	4.66	4.86	0.00
gamma-BHC (Lindane)	4.32	4.32	4.22	4.42	0.00
Heptachlor	4.92	4.92	4.82	5.02	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endosulfan I	6.06	6.06	5.96	6.16	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.56	6.56	6.46	6.66	0.00
Endosulfan II	6.78	6.78	6.68	6.88	0.00
4,4'-DDD	6.70	6.70	6.60	6.80	0.00
Endosulfan sulfate	7.14	7.14	7.04	7.24	0.00
4,4'-DDT	7.01	7.01	6.91	7.11	0.00
Methoxychlor	7.49	7.48	7.38	7.58	-0.01
Endrin ketone	7.62	7.62	7.52	7.72	0.00
Endrin aldehyde	6.91	6.91	6.81	7.01	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.93	5.83	6.03	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Continuing Calib Date: 12/09/2025

Initial Calibration Date(s): 12/08/2025

12/08/2025

Continuing Calib Time: 16:47

Initial Calibration Time(s): 15:09

16:03

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.87	2.77	2.97	0.00
alpha-BHC	3.38	3.38	3.28	3.48	0.00
beta-BHC	4.01	4.01	3.91	4.11	0.00
delta-BHC	4.25	4.25	4.15	4.35	0.00
gamma-BHC (Lindane)	3.72	3.72	3.62	3.82	0.00
Heptachlor	4.07	4.07	3.97	4.17	0.00
Aldrin	4.35	4.35	4.25	4.45	0.00
Heptachlor epoxide	4.86	4.86	4.76	4.96	0.00
Endosulfan I	5.23	5.23	5.13	5.33	0.00
Dieldrin	5.50	5.50	5.40	5.60	0.00
4,4'-DDE	5.36	5.36	5.26	5.46	0.00
Endrin	5.77	5.77	5.67	5.87	0.00
Endosulfan II	6.07	6.07	5.97	6.17	0.00
4,4'-DDD	5.91	5.91	5.81	6.01	0.00
Endosulfan sulfate	6.47	6.47	6.37	6.57	0.00
4,4'-DDT	6.17	6.17	6.07	6.27	0.00
Methoxychlor	6.74	6.74	6.64	6.84	0.00
Endrin ketone	6.98	6.98	6.88	7.08	0.00
Endrin aldehyde	6.24	6.24	6.14	6.34	0.00
alpha-Chlordane	5.18	5.18	5.08	5.28	0.00
gamma-Chlordane	5.11	5.11	5.01	5.21	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/08/2025 12/08/2025

Client Sample No.: CCAL06 **Date Analyzed:** 12/09/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091529.D **Time Analyzed:** 16:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.696	6.595	6.795	51.230	50.000	2.5
4,4'-DDE	6.186	6.085	6.285	47.770	50.000	-4.5
4,4'-DDT	7.012	6.911	7.111	47.240	50.000	-5.5
Aldrin	5.259	5.158	5.358	50.220	50.000	0.4
alpha-BHC	3.989	3.889	4.089	49.100	50.000	-1.8
alpha-Chlordane	6.016	5.915	6.115	47.430	50.000	-5.1
beta-BHC	4.508	4.408	4.608	48.170	50.000	-3.7
Decachlorobiphenyl	9.061	8.957	9.157	45.860	50.000	-8.3
delta-BHC	4.756	4.656	4.856	60.750	50.000	21.5
Dieldrin	6.337	6.235	6.435	48.850	50.000	-2.3
Endosulfan I	6.063	5.963	6.163	48.600	50.000	-2.8
Endosulfan II	6.777	6.676	6.876	44.880	50.000	-10.2
Endosulfan sulfate	7.141	7.040	7.240	47.300	50.000	-5.4
Endrin	6.562	6.463	6.663	48.120	50.000	-3.8
Endrin aldehyde	6.906	6.805	7.005	48.680	50.000	-2.6
Endrin ketone	7.620	7.521	7.721	47.350	50.000	-5.3
gamma-BHC (Lindane)	4.320	4.220	4.420	49.110	50.000	-1.8
gamma-Chlordane	5.935	5.834	6.034	49.640	50.000	-0.7
Heptachlor	4.918	4.817	5.017	50.410	50.000	0.8
Heptachlor epoxide	5.680	5.579	5.779	51.720	50.000	3.4
Methoxychlor	7.485	7.384	7.584	46.860	50.000	-6.3
Tetrachloro-m-xylene	3.540	3.440	3.640	47.650	50.000	-4.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3742
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/08/2025 12/08/2025

Client Sample No.: CCAL06 **Date Analyzed:** 12/09/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091529.D **Time Analyzed:** 16:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.914	5.814	6.014	54.610	50.000	9.2
4,4'-DDE	5.358	5.260	5.460	54.280	50.000	8.6
4,4'-DDT	6.168	6.068	6.268	54.610	50.000	9.2
Aldrin	4.352	4.254	4.454	54.070	50.000	8.1
alpha-BHC	3.380	3.281	3.481	51.860	50.000	3.7
alpha-Chlordane	5.175	5.076	5.276	54.850	50.000	9.7
beta-BHC	4.013	3.914	4.114	53.330	50.000	6.7
Decachlorobiphenyl	8.055	7.955	8.155	54.480	50.000	9.0
delta-BHC	4.249	4.150	4.350	54.130	50.000	8.3
Dieldrin	5.498	5.398	5.598	54.280	50.000	8.6
Endosulfan I	5.231	5.132	5.332	45.620	50.000	-8.8
Endosulfan II	6.066	5.966	6.166	54.310	50.000	8.6
Endosulfan sulfate	6.468	6.367	6.567	53.890	50.000	7.8
Endrin	5.773	5.674	5.874	54.340	50.000	8.7
Endrin aldehyde	6.244	6.144	6.344	53.430	50.000	6.9
Endrin ketone	6.976	6.876	7.076	54.470	50.000	8.9
gamma-BHC (Lindane)	3.716	3.617	3.817	57.430	50.000	14.9
gamma-Chlordane	5.110	5.011	5.211	55.610	50.000	11.2
Heptachlor	4.069	3.969	4.169	54.030	50.000	8.1
Heptachlor epoxide	4.857	4.758	4.958	54.840	50.000	9.7
Methoxychlor	6.739	6.638	6.838	53.850	50.000	7.7
Tetrachloro-m-xylene	2.869	2.770	2.970	50.750	50.000	1.5

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No. (PEM): PEM - PD091194.D Date Analyzed: 11/14/2025

Lab Sample No.(PEM): PEM Time Analyzed: 20:15

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.064	8.960	9.160	20.170	20.000	0.9
Tetrachloro-m-xylene	3.544	3.490	3.590	19.500	20.000	-2.5
alpha-BHC	3.993	3.940	4.040	8.250	10.000	-17.5
beta-BHC	4.511	4.460	4.560	9.240	10.000	-7.6
gamma-BHC (Lindane)	4.324	4.270	4.370	8.450	10.000	-15.5
Endrin	6.567	6.500	6.640	44.390	50.000	-11.2
4,4'-DDT	7.015	6.940	7.090	88.290	100.000	-11.7
Methoxychlor	7.487	7.420	7.560	207.040	250.000	-17.2

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No. (PEM): PEM - PD091194.D Date Analyzed: 11/14/2025

Lab Sample No.(PEM): PEM Time Analyzed: 20:15

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.059	7.960	8.160	19.620	20.000	-1.9
Tetrachloro-m-xylene	2.874	2.820	2.920	19.370	20.000	-3.2
alpha-BHC	3.385	3.330	3.440	9.380	10.000	-6.2
beta-BHC	4.018	3.970	4.070	9.530	10.000	-4.7
gamma-BHC (Lindane)	3.721	3.670	3.770	9.460	10.000	-5.4
Endrin	5.779	5.710	5.850	45.450	50.000	-9.1
4,4'-DDT	6.172	6.100	6.240	89.340	100.000	-10.7
Methoxychlor	6.742	6.670	6.810	186.630	250.000	-25.3

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No. (PEM): PEM - PD091417.D Date Analyzed: 12/02/2025

Lab Sample No.(PEM): PEM Time Analyzed: 12:04

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.063	8.960	9.160	20.040	20.000	0.2
Tetrachloro-m-xylene	3.541	3.490	3.590	21.200	20.000	6.0
alpha-BHC	3.990	3.940	4.040	9.100	10.000	-9.0
beta-BHC	4.509	4.460	4.560	10.140	10.000	1.4
gamma-BHC (Lindane)	4.321	4.270	4.370	9.300	10.000	-7.0
Endrin	6.565	6.490	6.640	50.900	50.000	1.8
4,4'-DDT	7.014	6.940	7.080	97.180	100.000	-2.8
Methoxychlor	7.486	7.420	7.560	219.850	250.000	-12.1

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No. (PEM): PEM - PD091417.D Date Analyzed: 12/02/2025

Lab Sample No.(PEM): PEM Time Analyzed: 12:04

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.056	7.960	8.160	22.260	20.000	11.3
Tetrachloro-m-xylene	2.871	2.820	2.920	23.970	20.000	19.9
alpha-BHC	3.382	3.330	3.430	11.750	10.000	17.5
beta-BHC	4.015	3.960	4.070	11.670	10.000	16.7
gamma-BHC (Lindane)	3.718	3.670	3.770	11.820	10.000	18.2
Endrin	5.776	5.710	5.850	57.760	50.000	15.5
4,4'-DDT	6.169	6.100	6.240	114.940	100.000	14.9
Methoxychlor	6.740	6.670	6.810	216.570	250.000	-13.4

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No. (PEM): PEM - PD091433.D Date Analyzed: 12/02/2025

Lab Sample No.(PEM): PEM Time Analyzed: 18:45

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.064	8.960	9.160	19.770	20.000	-1.2
Tetrachloro-m-xylene	3.541	3.490	3.590	20.970	20.000	4.9
alpha-BHC	3.990	3.940	4.040	9.050	10.000	-9.5
beta-BHC	4.509	4.460	4.560	10.220	10.000	2.2
gamma-BHC (Lindane)	4.321	4.270	4.370	9.260	10.000	-7.4
Endrin	6.565	6.490	6.640	53.720	50.000	7.4
4,4'-DDT	7.013	6.940	7.080	104.670	100.000	4.7
Methoxychlor	7.486	7.420	7.560	237.210	250.000	-5.1

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No. (PEM): PEM - PD091433.D Date Analyzed: 12/02/2025

Lab Sample No.(PEM): PEM Time Analyzed: 18:45

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.058	7.960	8.160	22.240	20.000	11.2
Tetrachloro-m-xylene	2.872	2.820	2.920	23.470	20.000	17.4
alpha-BHC	3.383	3.330	3.430	11.610	10.000	16.1
beta-BHC	4.016	3.970	4.070	11.420	10.000	14.2
gamma-BHC (Lindane)	3.719	3.670	3.770	11.720	10.000	17.2
Endrin	5.777	5.710	5.850	62.670	50.000	25.3
4,4'-DDT	6.171	6.100	6.240	120.980	100.000	21.0
Methoxychlor	6.741	6.670	6.810	239.330	250.000	-4.3

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/08/2025 12/08/2025

Client Sample No. (PEM): PEM - PD091479.D Date Analyzed: 12/08/2025

Lab Sample No.(PEM): PEM Time Analyzed: 12:59

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.058	8.960	9.160	20.680	20.000	3.4
Tetrachloro-m-xylene	3.540	3.490	3.590	20.340	20.000	1.7
alpha-BHC	3.988	3.940	4.040	8.510	10.000	-14.9
beta-BHC	4.507	4.460	4.560	9.600	10.000	-4.0
gamma-BHC (Lindane)	4.319	4.270	4.370	8.650	10.000	-13.5
Endrin	6.563	6.490	6.630	45.750	50.000	-8.5
4,4'-DDT	7.010	6.940	7.080	93.730	100.000	-6.3
Methoxychlor	7.484	7.410	7.550	216.760	250.000	-13.3

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/08/2025 12/08/2025

Client Sample No. (PEM): PEM - PD091479.D Date Analyzed: 12/08/2025

Lab Sample No.(PEM): PEM Time Analyzed: 12:59

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.054	7.950	8.150	19.300	20.000	-3.5
Tetrachloro-m-xylene	2.870	2.820	2.920	20.110	20.000	0.6
alpha-BHC	3.382	3.330	3.430	9.600	10.000	-4.0
beta-BHC	4.014	3.960	4.060	9.320	10.000	-6.8
gamma-BHC (Lindane)	3.717	3.670	3.770	8.320	10.000	-16.8
Endrin	5.774	5.700	5.840	42.820	50.000	-14.4
4,4'-DDT	6.167	6.100	6.240	87.960	100.000	-12.0
Methoxychlor	6.738	6.670	6.810	187.020	250.000	-25.2

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/08/2025 12/08/2025

Client Sample No. (PEM): PEM - PD091524.D Date Analyzed: 12/09/2025

Lab Sample No.(PEM): PEM Time Analyzed: 13:20

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.057	8.960	9.160	22.960	20.000	14.8
Tetrachloro-m-xylene	3.539	3.490	3.590	22.010	20.000	10.1
alpha-BHC	3.989	3.940	4.040	9.430	10.000	-5.7
beta-BHC	4.507	4.460	4.560	10.620	10.000	6.2
gamma-BHC (Lindane)	4.319	4.270	4.370	9.780	10.000	-2.2
Endrin	6.563	6.490	6.630	51.200	50.000	2.4
4,4'-DDT	7.010	6.940	7.080	99.040	100.000	-1.0
Methoxychlor	7.484	7.410	7.550	227.530	250.000	-9.0

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/08/2025 12/08/2025

Client Sample No. (PEM): PEM - PD091524.D Date Analyzed: 12/09/2025

Lab Sample No.(PEM): PEM Time Analyzed: 13:20

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.054	7.950	8.150	25.200	20.000	26.0
Tetrachloro-m-xylene	2.870	2.820	2.920	23.340	20.000	16.7
alpha-BHC	3.381	3.330	3.430	11.360	10.000	13.6
beta-BHC	4.014	3.960	4.060	12.460	10.000	24.6
gamma-BHC (Lindane)	3.717	3.670	3.770	12.690	10.000	26.9
Endrin	5.774	5.700	5.840	54.760	50.000	9.5
4,4'-DDT	6.168	6.100	6.240	111.970	100.000	12.0
Methoxychlor	6.739	6.670	6.810	227.310	250.000	-9.1

Analytical Sequence

Client: Remington & Vernick

SDG No.: Q3742

Project: Saddler Property

Instrument ID: ECD_D

GC Column: ZB-MR1

ID: 0.32 (mm)

Inst. Calib. Date(s): 11/14/2025

11/14/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	11/14/2025	20:01	PD091193.D	9.06	3.54
PEM	PEM	11/14/2025	20:15	PD091194.D	9.06	3.54
RESCHK	RESCHK	11/14/2025	20:43	PD091195.D	9.06	3.54
PSTDICC100	PSTDICC100	11/14/2025	20:56	PD091196.D	9.06	3.54
PSTDICC075	PSTDICC075	11/14/2025	21:10	PD091197.D	9.06	3.54
PSTDICC050	PSTDICC050	11/14/2025	21:24	PD091198.D	9.06	3.54
PSTDICC025	PSTDICC025	11/14/2025	21:37	PD091199.D	9.07	3.54
PSTDICC005	PSTDICC005	11/14/2025	21:51	PD091200.D	9.06	3.54
PCHLORICC500	PCHLORICC500	11/14/2025	22:32	PD091203.D	9.06	3.54
PTOXICC500	PTOXICC500	11/14/2025	23:40	PD091208.D	9.06	3.54
IBLK	IBLK	12/02/2025	11:50	PD091416.D	9.06	3.54
PEM	PEM	12/02/2025	12:04	PD091417.D	9.06	3.54
PSTDCCC050	PSTDCCC050	12/02/2025	12:44	PD091418.D	9.07	3.55
PB170776BL	PB170776BL	12/02/2025	13:37	PD091419.D	9.07	3.55
IBLK	IBLK	12/02/2025	16:25	PD091425.D	9.06	3.54
PSTDCCC050	PSTDCCC050	12/02/2025	16:39	PD091426.D	9.06	3.54
SB-1125-23	Q3742-01	12/02/2025	17:23	PD091427.D	9.08	3.55
SB-1125-19	Q3742-02	12/02/2025	17:37	PD091428.D	9.07	3.54
SB-1125-21	Q3742-05	12/02/2025	18:17	PD091431.D	9.07	3.54
IBLK	IBLK	12/02/2025	18:31	PD091432.D	9.06	3.54
PEM	PEM	12/02/2025	18:45	PD091433.D	9.06	3.54
PSTDCCC050	PSTDCCC050	12/02/2025	19:12	PD091434.D	9.06	3.54
SB-1125-13	Q3742-06	12/02/2025	19:26	PD091435.D	9.06	3.54
SB-1125-18	Q3742-07	12/02/2025	19:39	PD091436.D	9.07	3.54
SB-1125-18MS	Q3742-07MS	12/02/2025	19:53	PD091437.D	9.06	3.54
SB-1125-18MSD	Q3742-07MSD	12/02/2025	20:06	PD091438.D	9.06	3.54
SB-1125-20	Q3742-08	12/02/2025	20:20	PD091439.D	9.07	3.54
SB-1125-22	Q3742-09	12/02/2025	20:34	PD091440.D	9.06	3.54
SB-1125-2	Q3742-10	12/02/2025	20:47	PD091441.D	9.06	3.54
SB-1125-1	Q3742-16	12/02/2025	21:01	PD091442.D	9.06	3.54
SB-1125-25	Q3742-17	12/02/2025	21:15	PD091443.D	9.06	3.54
SB-1125-24	Q3742-18	12/02/2025	21:28	PD091444.D	9.06	3.54
IBLK	IBLK	12/02/2025	21:42	PD091445.D	9.06	3.54
PSTDCCC050	PSTDCCC050	12/02/2025	21:55	PD091446.D	9.06	3.54
IBLK	IBLK	12/08/2025	12:45	PD091478.D	9.06	3.54
PEM	PEM	12/08/2025	12:59	PD091479.D	9.06	3.54
RESCHK	RESCHK	12/08/2025	14:55	PD091480.D	9.07	3.55
PSTDICC100	PSTDICC100	12/08/2025	15:09	PD091481.D	9.06	3.54
PSTDICC075	PSTDICC075	12/08/2025	15:22	PD091482.D	9.06	3.54
PSTDICC050	PSTDICC050	12/08/2025	15:36	PD091483.D	9.06	3.54
PSTDICC025	PSTDICC025	12/08/2025	15:50	PD091484.D	9.06	3.54
PSTDICC005	PSTDICC005	12/08/2025	16:03	PD091485.D	9.06	3.54

Analytical Sequence

PCHLORICC500	PCHLORICC500	12/08/2025	16:44	PD091488.D	9.06	3.54
PTOXICC500	PTOXICC500	12/08/2025	17:53	PD091493.D	9.06	3.54
LBLK	LBLK	12/09/2025	13:07	PD091523.D	9.06	3.54
PEM	PEM	12/09/2025	13:20	PD091524.D	9.06	3.54
PSTDCCC050	PSTDCCC050	12/09/2025	14:05	PD091525.D	9.06	3.54
PB170776BS	PB170776BS	12/09/2025	14:22	PD091526.D	9.06	3.54
LBLK	LBLK	12/09/2025	16:33	PD091528.D	9.07	3.55
PSTDCCC050	PSTDCCC050	12/09/2025	16:47	PD091529.D	9.06	3.54

Analytical Sequence

Client: Remington & Vernick

SDG No.: Q3742

Project: Saddler Property

Instrument ID: ECD_D

GC Column: ZB-MR2

ID: 0.32 (mm)

Inst. Calib. Date(s): 11/14/2025

11/14/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	11/14/2025	20:01	PD091193.D	8.06	2.88
PEM	PEM	11/14/2025	20:15	PD091194.D	8.06	2.87
RESCHK	RESCHK	11/14/2025	20:43	PD091195.D	8.06	2.87
PSTDICCC100	PSTDICCC100	11/14/2025	20:56	PD091196.D	8.06	2.87
PSTDICCC075	PSTDICCC075	11/14/2025	21:10	PD091197.D	8.06	2.87
PSTDICCC050	PSTDICCC050	11/14/2025	21:24	PD091198.D	8.06	2.87
PSTDICCC025	PSTDICCC025	11/14/2025	21:37	PD091199.D	8.06	2.87
PSTDICCC005	PSTDICCC005	11/14/2025	21:51	PD091200.D	8.06	2.87
PCHLORICC500	PCHLORICC500	11/14/2025	22:32	PD091203.D	8.06	2.87
PTOXICC500	PTOXICC500	11/14/2025	23:40	PD091208.D	8.06	2.87
IBLK	IBLK	12/02/2025	11:50	PD091416.D	8.06	2.87
PEM	PEM	12/02/2025	12:04	PD091417.D	8.06	2.87
PSTDCCC050	PSTDCCC050	12/02/2025	12:44	PD091418.D	8.06	2.87
PB170776BL	PB170776BL	12/02/2025	13:37	PD091419.D	8.06	2.87
IBLK	IBLK	12/02/2025	16:25	PD091425.D	8.06	2.87
PSTDCCC050	PSTDCCC050	12/02/2025	16:39	PD091426.D	8.06	2.87
SB-1125-23	Q3742-01	12/02/2025	17:23	PD091427.D	8.06	2.87
SB-1125-19	Q3742-02	12/02/2025	17:37	PD091428.D	8.06	2.87
SB-1125-21	Q3742-05	12/02/2025	18:17	PD091431.D	8.06	2.87
IBLK	IBLK	12/02/2025	18:31	PD091432.D	8.06	2.87
PEM	PEM	12/02/2025	18:45	PD091433.D	8.06	2.87
PSTDCCC050	PSTDCCC050	12/02/2025	19:12	PD091434.D	8.06	2.87
SB-1125-13	Q3742-06	12/02/2025	19:26	PD091435.D	8.06	2.87
SB-1125-18	Q3742-07	12/02/2025	19:39	PD091436.D	8.06	2.87
SB-1125-18MS	Q3742-07MS	12/02/2025	19:53	PD091437.D	8.06	2.87
SB-1125-18MSD	Q3742-07MSD	12/02/2025	20:06	PD091438.D	8.06	2.87
SB-1125-20	Q3742-08	12/02/2025	20:20	PD091439.D	8.06	2.87
SB-1125-22	Q3742-09	12/02/2025	20:34	PD091440.D	8.06	2.87
SB-1125-2	Q3742-10	12/02/2025	20:47	PD091441.D	8.06	2.87
SB-1125-1	Q3742-16	12/02/2025	21:01	PD091442.D	8.06	2.87
SB-1125-25	Q3742-17	12/02/2025	21:15	PD091443.D	8.06	2.87
SB-1125-24	Q3742-18	12/02/2025	21:28	PD091444.D	8.06	2.87
IBLK	IBLK	12/02/2025	21:42	PD091445.D	8.06	2.87
PSTDCCC050	PSTDCCC050	12/02/2025	21:55	PD091446.D	8.06	2.87
IBLK	IBLK	12/08/2025	12:45	PD091478.D	8.05	2.87
PEM	PEM	12/08/2025	12:59	PD091479.D	8.05	2.87
RESCHK	RESCHK	12/08/2025	14:55	PD091480.D	8.06	2.87
PSTDICCC100	PSTDICCC100	12/08/2025	15:09	PD091481.D	8.06	2.87
PSTDICCC075	PSTDICCC075	12/08/2025	15:22	PD091482.D	8.05	2.87
PSTDICCC050	PSTDICCC050	12/08/2025	15:36	PD091483.D	8.06	2.87
PSTDICCC025	PSTDICCC025	12/08/2025	15:50	PD091484.D	8.06	2.87
PSTDICCC005	PSTDICCC005	12/08/2025	16:03	PD091485.D	8.05	2.87

Analytical Sequence

PCHLORICC500	PCHLORICC500	12/08/2025	16:44	PD091488.D	8.05	2.87
PTOXICC500	PTOXICC500	12/08/2025	17:53	PD091493.D	8.05	2.87
LBLK	LBLK	12/09/2025	13:07	PD091523.D	8.06	2.87
PEM	PEM	12/09/2025	13:20	PD091524.D	8.05	2.87
PSTDCCC050	PSTDCCC050	12/09/2025	14:05	PD091525.D	8.06	2.87
PB170776BS	PB170776BS	12/09/2025	14:22	PD091526.D	8.06	2.87
LBLK	LBLK	12/09/2025	16:33	PD091528.D	8.06	2.87
PSTDCCC050	PSTDCCC050	12/09/2025	16:47	PD091529.D	8.06	2.87

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB170776BS

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Lab Sample ID: PB170776BS

Date(s) Analyzed: 12/09/2025 12/09/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDD	1	6.70	6.65	6.75	16.8	8.5
	2	5.92	5.87	5.97	18.3	
4,4'-DDE	1	6.19	6.14	6.24	16.4	9.3
	2	5.36	5.31	5.41	18.0	
4,4'-DDT	1	7.02	6.97	7.07	15.8	13
	2	6.17	6.12	6.22	18.0	
alpha-BHC	1	3.99	3.94	4.04	15.7	6.8
	2	3.38	3.33	3.43	16.8	
Aldrin	1	5.26	5.21	5.31	16.7	6.4
	2	4.36	4.31	4.41	17.8	
alpha-Chlordane	1	6.02	5.97	6.07	15.8	14.1
	2	5.18	5.13	5.23	18.2	
Endosulfan II	1	6.78	6.73	6.83	14.7	20.2
	2	6.07	6.02	6.12	18.0	
Endosulfan sulfate	1	7.14	7.09	7.19	16.1	10.6
	2	6.47	6.42	6.52	17.9	
beta-BHC	1	4.51	4.46	4.56	15.8	13.6
	2	4.01	3.96	4.06	18.1	
delta-BHC	1	4.76	4.71	4.81	19.2	9.3
	2	4.25	4.20	4.30	17.5	
Endosulfan I	1	6.07	6.02	6.12	16.3	9.4
	2	5.23	5.18	5.28	17.9	
Dieldrin	1	6.34	6.29	6.39	16.5	8.7
	2	5.50	5.45	5.55	18.0	
Endrin aldehyde	1	6.91	6.86	6.96	16.2	8.8
	2	6.25	6.20	6.30	17.7	
Methoxychlor	1	7.49	7.44	7.54	15.5	13.3
	2	6.74	6.69	6.79	17.7	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB170776BS

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Lab Sample ID: PB170776BS

Date(s) Analyzed: 12/09/2025 12/09/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endrin ketone	1	7.62	7.57	7.67	16.3	11
	2	6.98	6.93	7.03	18.2	
gamma-BHC (Lindane)	1	4.32	4.27	4.37	15.9	16.2
	2	3.72	3.67	3.77	18.7	
Heptachlor	1	4.92	4.87	4.97	16.5	7
	2	4.07	4.02	4.12	17.7	
Heptachlor epoxide	1	5.68	5.63	5.73	16.5	10.3
	2	4.86	4.81	4.91	18.3	
gamma-Chlordane	1	5.94	5.89	5.99	16.5	8.7
	2	5.11	5.06	5.16	18.0	
Endrin	1	6.57	6.52	6.62	16.1	7.2
	2	5.78	5.73	5.83	17.3	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

SB-1125-1

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Lab Sample ID: Q3742-16

Date(s) Analyzed: 12/02/2025 12/02/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
gamma-BHC (Lindane)	1	4.32	4.27	4.37	19.8	10.5
	2	3.72	3.67	3.77	22.0	
Heptachlor	1	4.92	4.87	4.97	0.66	34.6
	2	4.07	4.02	4.12	0.94	
Heptachlor epoxide	1	5.68	5.63	5.73	9.40	17.3
	2	4.86	4.81	4.91	7.90	
gamma-Chlordane	1	5.94	5.89	5.99	30.9	14.2
	2	5.11	5.06	5.16	26.8	
alpha-Chlordane	1	6.02	5.97	6.07	23.6	33.1
	2	5.18	5.13	5.23	16.9	
Dieldrin	1	6.34	6.29	6.39	0.38	19.4
	2	5.50	5.45	5.55	0.46	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

SB-1125-18MS

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Lab Sample ID: Q3742-07MS

Date(s) Analyzed: 12/02/2025 12/02/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDE	1	6.19	6.14	6.24	20.3	11.5
	2	5.36	5.31	5.41	18.1	
Dieldrin	1	6.34	6.29	6.39	20.6	9.7
	2	5.50	5.45	5.55	18.7	
Endrin	1	6.56	6.51	6.61	22.8	15.1
	2	5.78	5.73	5.83	19.6	
Endosulfan II	1	6.78	6.73	6.83	20.4	11.9
	2	6.07	6.02	6.12	18.1	
4,4'-DDD	1	6.70	6.65	6.75	20.3	26.2
	2	5.92	5.87	5.97	15.6	
4,4'-DDT	1	7.01	6.96	7.06	23.5	17.6
	2	6.17	6.12	6.22	19.7	
Endrin aldehyde	1	6.91	6.86	6.96	21.7	13.8
	2	6.25	6.20	6.30	18.9	
Endosulfan sulfate	1	7.14	7.09	7.19	21.2	15.8
	2	6.47	6.42	6.52	18.1	
Methoxychlor	1	7.49	7.44	7.54	23.0	19.6
	2	6.74	6.69	6.79	18.9	
Endrin ketone	1	7.62	7.57	7.67	21.6	23.8
	2	6.98	6.93	7.03	17.0	
alpha-BHC	1	3.99	3.94	4.04	23.2	7.1
	2	3.38	3.33	3.43	21.6	
gamma-BHC (Lindane)	1	4.32	4.27	4.37	22.6	11.2
	2	3.72	3.67	3.77	20.2	
Heptachlor	1	4.92	4.87	4.97	21.3	9.9
	2	4.07	4.02	4.12	19.3	
Aldrin	1	5.26	5.21	5.31	22.1	6.5
	2	4.36	4.31	4.41	20.7	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

SB-1125-18MS

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Lab Sample ID: Q3742-07MS

Date(s) Analyzed: 12/02/2025 12/02/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
beta-BHC	1	4.51	4.46	4.56	20.8	0
	2	4.02	3.97	4.07	20.8	
delta-BHC	1	4.76	4.71	4.81	23.1	13.9
	2	4.25	4.20	4.30	20.1	
Heptachlor epoxide	1	5.68	5.63	5.73	21.4	10.3
	2	4.86	4.81	4.91	19.3	
Endosulfan I	1	6.07	6.02	6.12	21.3	17.3
	2	5.23	5.18	5.28	17.9	
gamma-Chlordane	1	5.94	5.89	5.99	20.8	11.7
	2	5.11	5.06	5.16	18.5	
alpha-Chlordane	1	6.02	5.97	6.07	19.8	6.8
	2	5.18	5.13	5.23	18.5	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

SB-1125-18MSD

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Lab Sample ID: Q3742-07MSD

Date(s) Analyzed: 12/02/2025 12/02/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDD	1	6.70	6.65	6.75	19.8	25.6
	2	5.92	5.87	5.97	15.3	
4,4'-DDT	1	7.01	6.96	7.06	23.5	16.1
	2	6.17	6.12	6.22	20.0	
alpha-BHC	1	3.99	3.94	4.04	22.8	7.3
	2	3.38	3.33	3.43	21.2	
Aldrin	1	5.26	5.21	5.31	21.6	5.7
	2	4.36	4.31	4.41	20.4	
4,4'-DDE	1	6.19	6.14	6.24	20.0	11.1
	2	5.36	5.31	5.41	17.9	
Endosulfan II	1	6.78	6.73	6.83	20.0	6.2
	2	6.07	6.02	6.12	18.8	
Endrin aldehyde	1	6.91	6.86	6.96	21.4	13.5
	2	6.25	6.20	6.30	18.7	
Endosulfan sulfate	1	7.14	7.09	7.19	20.9	15.5
	2	6.47	6.42	6.52	17.9	
Methoxychlor	1	7.49	7.44	7.54	23.2	18.9
	2	6.74	6.69	6.79	19.2	
Endrin ketone	1	7.62	7.57	7.67	21.3	23.6
	2	6.98	6.93	7.03	16.8	
gamma-BHC (Lindane)	1	4.32	4.27	4.37	22.2	11.4
	2	3.72	3.67	3.77	19.8	
Heptachlor	1	4.92	4.87	4.97	21.4	9.8
	2	4.07	4.02	4.12	19.4	
beta-BHC	1	4.51	4.46	4.56	20.3	1.5
	2	4.02	3.97	4.07	20.6	
delta-BHC	1	4.76	4.71	4.81	23.0	14.5
	2	4.25	4.20	4.30	19.9	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

SB-1125-18MSD

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Lab Sample ID: Q3742-07MSD

Date(s) Analyzed: 12/02/2025 12/02/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Heptachlor epoxide	1	5.68	5.63	5.73	21.1	10.5
	2	4.86	4.81	4.91	19.0	
Endosulfan I	1	6.07	6.02	6.12	20.9	16.6
	2	5.24	5.19	5.29	17.7	
gamma-Chlordane	1	5.94	5.89	5.99	20.5	11.3
	2	5.11	5.06	5.16	18.3	
alpha-Chlordane	1	6.02	5.97	6.07	19.6	7.4
	2	5.18	5.13	5.23	18.2	
Dieldrin	1	6.34	6.29	6.39	20.3	9.3
	2	5.50	5.45	5.55	18.5	
Endrin	1	6.57	6.52	6.62	23.0	16
	2	5.78	5.73	5.83	19.6	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

SB-1125-21

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Lab Sample ID: Q3742-05

Date(s) Analyzed: 12/02/2025 12/02/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDT	1	7.01	6.96	7.06	3.80	7.6
	2	6.17	6.12	6.22	4.10	
4,4'-DDE	1	6.19	6.14	6.24	0.95	61.9
	2	5.36	5.31	5.41	1.80	
Dieldrin	1	6.34	6.29	6.39	0.44	71.1
	2	5.50	5.45	5.55	0.93	
Endrin	1	6.57	6.52	6.62	0.32	88.8
	2	5.78	5.73	5.83	0.84	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

SB-1125-22

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Lab Sample ID: Q3742-09

Date(s) Analyzed: 12/02/2025 12/02/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDE	1	6.19	6.14	6.24	0.22	71.9
	2	5.36	5.31	5.41	0.48	



SAMPLE RAW DATA

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
 Data File : PD091427.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Dec 2025 17:23
 Operator : AR\AJ
 Sample : Q3742-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 SB-1125-23

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 12/03/2025
 Supervised By : Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 01:09:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.548	2.871	61999307	772.5E6	17.978m	21.090
28) SA Decachlor...	9.075	8.061	79869136	651.3E6	16.032	16.459

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
Data File : PD091427.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2025 17:23
Operator : AR\AJ
Sample : Q3742-01
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
SB-1125-23

Manual Integrations
APPROVED


Reviewed By : Abdul Mirza 12/03/2025
Supervised By : Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 01:09:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PD091427.D\ECD1A.ch



#1 Tetrachloro-m-xylene


R.T.: 3.548 min
Delta R.T.: 0.004 min
Response: 61999307
Conc: 17.98 ng/ml

Instrument :
ECD_D
ClientSampleId :
SB-1125-23

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 12/03/2025
Supervised By :Sohil Jodhani 12/03/2025


Time Response_ Signal: PD091427.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.871 min
Delta R.T.: -0.003 min
Response: 772515638
Conc: 21.09 ng/ml


Time Response_ Signal: PD091427.D\ECD1A.ch



#28 Decachlorobiphenyl

R.T.: 9.075 min
Delta R.T.: 0.011 min
Response: 79869136
Conc: 16.03 ng/ml

Time Response_ Signal: PD091427.D\ECD2B.ch



#28 Decachlorobiphenyl

R.T.: 8.061 min
Delta R.T.: 0.002 min
Response: 651344172
Conc: 16.46 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
 Data File : PD091428.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Dec 2025 17:37
 Operator : AR\AJ
 Sample : Q3742-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 SB-1125-19

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 12/03/2025
 Supervised By : Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 01:09:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.540	2.870	64723812	802.0E6	18.768m	21.894m
28) SA Decachlor...	9.065	8.059	83107090	637.2E6	16.682m	16.100

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
Data File : PD091428.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2025 17:37
Operator : AR\AJ
Sample : Q3742-02
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
SB-1125-19


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 12/03/2025
Supervised By : Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 01:09:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.540 min
Delta R.T.: -0.004 min
Response: 64723812
Conc: 18.77 ng/ml

Instrument :
ECD_D
ClientSampleId :
SB-1125-19


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 12/03/2025
Supervised By :Sohil Jodhani 12/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.870 min
Delta R.T.: -0.004 min
Response: 801983123
Conc: 21.89 ng/ml m



#28 Decachlorobiphenyl

R.T.: 9.065 min
Delta R.T.: 0.001 min
Response: 83107090
Conc: 16.68 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.059 min
Delta R.T.: 0.000 min
Response: 637160943
Conc: 16.10 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
 Data File : PD091431.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Dec 2025 18:17
 Operator : AR\AJ
 Sample : Q3742-05
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 SB-1125-21

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 12/03/2025
 Supervised By : Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 01:10:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.540	2.871	78003668	842.8E6	22.619m	23.009
2) SA Decachlor...	9.065	8.058	88479728	679.6E6	17.760	17.173
Target Compounds						
12) B 4,4'-DDE	6.187	5.360	13191335	223.9E6	2.275m	4.190m#
13) MA Dieldrin	6.341	5.499	6881729	121.3E6	1.064m	2.233m#
14) MA Endrin	6.567	5.779	3828003	90600758	0.771m	2.005m#
17) MA 4,4'-DDT	7.014	6.169	37618259	386.2E6	9.096	9.870

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
Data File : PD091431.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2025 18:17
Operator : AR\AJ
Sample : Q3742-05
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

SB-1125-21

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 12/03/2025

Supervised By :Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 01:10:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.540 min
Delta R.T.: -0.004 min
Response: 78003668
Conc: 22.62 ng/ml

Instrument :
ECD_D
ClientSampleId :
SB-1125-21


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 12/03/2025
Supervised By :Sohil Jodhani 12/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.871 min
Delta R.T.: -0.003 min
Response: 842819102
Conc: 23.01 ng/ml



#12 4,4'-DDE

R.T.: 6.187 min
Delta R.T.: -0.001 min
Response: 13191335
Conc: 2.27 ng/ml m



#12 4,4'-DDE

R.T.: 5.360 min
Delta R.T.: -0.003 min
Response: 223914036
Conc: 4.19 ng/ml m

#13 Dieldrin

Instrument :
ECD_D
ClientSampleId :
SB-1125-21

Reviewed By :Abdul Mirza 12/03/2025
Supervised By :Sohil Jodhani 12/03/2025

#13 Dieldrin

#14 Endrin


R.T.: 6.567 min
Delta R.T.: 0.001 min
Response: 3828003
Conc: 0.77 ng/ml m

#14 Endrin

```

      R.T.:      5.779 min
Delta R.T.:      0.002 min
Response: 90600758
      Conc:      2.00 ng/ml m

```




#17 4,4'-DDT

R.T.: 7.014 min
Delta R.T.: 0.000 min
Response: 37618259
Conc: 9.10 ng/ml

Instrument :
ECD_D
ClientSampleId :
SB-1125-21


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 12/03/2025
Supervised By :Sohil Jodhani 12/03/2025




#17 4,4'-DDT

R.T.: 6.169 min
Delta R.T.: -0.002 min
Response: 386218619
Conc: 9.87 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.065 min
Delta R.T.: 0.001 min
Response: 88479728
Conc: 17.76 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.058 min
Delta R.T.: 0.000 min
Response: 679599890
Conc: 17.17 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
 Data File : PD091435.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Dec 2025 19:26
 Operator : AR\AJ
 Sample : Q3742-06
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 SB-1125-13

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 12/03/2025
 Supervised By : Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 01:11:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.541	2.871	50414787	565.9E6	14.619m	15.448m
28) SA Decachlor...	9.064	8.059	78110422	633.9E6	15.679	16.017

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
Data File : PD091435.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2025 19:26
Operator : AR\AJ
Sample : Q3742-06
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

SB-1125-13

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 12/03/2025

Supervised By :Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 01:11:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.541 min
Delta R.T.: -0.003 min
Response: 50414787
Conc: 14.62 ng/ml

Instrument :


ECD_D

ClientSampleId :

SB-1125-13


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 12/03/2025
Supervised By :Sohil Jodhani 12/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.871 min
Delta R.T.: -0.003 min
Response: 565851945
Conc: 15.45 ng/ml m



#28 Decachlorobiphenyl

R.T.: 9.064 min
Delta R.T.: 0.000 min
Response: 78110422
Conc: 15.68 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.059 min
Delta R.T.: 0.000 min
Response: 633869627
Conc: 16.02 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
 Data File : PD091436.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Dec 2025 19:39
 Operator : AR\AJ
 Sample : Q3742-07
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 SB-1125-18

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 12/03/2025
 Supervised By : Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 01:11:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.540	2.872	57372515	690.0E6	16.636m	18.837
28) SA Decachlor...	9.065	8.058	89567253	565.0E6	17.979	14.276

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
Data File : PD091436.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2025 19:39
Operator : AR\AJ
Sample : Q3742-07
Misc :
ALS Vial : 17 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
SB-1125-18


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 12/03/2025
Supervised By : Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 01:11:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.540 min
Delta R.T.: -0.004 min
Response: 57372515
Conc: 16.64 ng/ml

Instrument :
ECD_D
ClientSampleId :
SB-1125-18


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 12/03/2025
Supervised By :Sohil Jodhani 12/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.872 min
Delta R.T.: -0.003 min
Response: 690007903
Conc: 18.84 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.065 min
Delta R.T.: 0.002 min
Response: 89567253
Conc: 17.98 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.058 min
Delta R.T.: 0.000 min
Response: 564986089
Conc: 14.28 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
 Data File : PD091439.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Dec 2025 20:20
 Operator : AR\AJ
 Sample : Q3742-08
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 SB-1125-20

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 12/03/2025
 Supervised By : Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 01:11:40 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.540	2.872	62956443	776.5E6	18.256m	21.197
28) SA Decachlor...	9.065	8.058	89950811	627.3E6	18.056	15.851

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
Data File : PD091439.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2025 20:20
Operator : AR\AJ
Sample : Q3742-08
Misc :
ALS Vial : 20 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
SB-1125-20


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 12/03/2025
Supervised By : Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 01:11:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.540 min
Delta R.T.: -0.004 min
Response: 62956443
Conc: 18.26 ng/ml

Instrument :
ECD_D
ClientSampleId :
SB-1125-20


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 12/03/2025
Supervised By :Sohil Jodhani 12/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.872 min
Delta R.T.: -0.002 min
Response: 776455823
Conc: 21.20 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.065 min
Delta R.T.: 0.002 min
Response: 89950811
Conc: 18.06 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.058 min
Delta R.T.: 0.000 min
Response: 627303506
Conc: 15.85 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
 Data File : PD091440.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Dec 2025 20:34
 Operator : AR\AJ
 Sample : Q3742-09
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 SB-1125-22

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 12/03/2025
 Supervised By : Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 01:11:50 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.540	2.872	60552023	755.7E6	17.558m	20.632
28)	SA Decachlor...	9.064	8.058	80795907	634.6E6	16.218	16.036
Target Compounds							
12)	B 4,4'-DDE	6.187	5.361	2812944	55504496	0.485	1.039 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
Data File : PD091440.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2025 20:34
Operator : AR\AJ
Sample : Q3742-09
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
SB-1125-22


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 12/03/2025
Supervised By : Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 01:11:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.540 min
Delta R.T.: -0.004 min
Response: 60552023
Conc: 17.56 ng/ml

Instrument :
ECD_D
ClientSampleId :
SB-1125-22


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 12/03/2025
Supervised By :Sohil Jodhani 12/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.872 min
Delta R.T.: -0.002 min
Response: 755743711
Conc: 20.63 ng/ml




#12 4,4'-DDE

R.T.: 6.187 min
Delta R.T.: -0.002 min
Response: 2812944
Conc: 0.49 ng/ml



#12 4,4'-DDE

R.T.: 5.361 min
Delta R.T.: -0.002 min
Response: 55504496
Conc: 1.04 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.064 min
Delta R.T.: 0.000 min
Response: 80795907
Conc: 16.22 ng/ml

Instrument :
ECD_D
ClientSampleId :
SB-1125-22

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 12/03/2025
Supervised By :Sohil Jodhani 12/03/2025



#28 Decachlorobiphenyl

R.T.: 8.058 min
Delta R.T.: 0.000 min
Response: 634611906
Conc: 16.04 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
 Data File : PD091441.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Dec 2025 20:47
 Operator : AR\AJ
 Sample : Q3742-10
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 SB-1125-2

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 12/03/2025
 Supervised By : Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 01:11:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.540	2.872	61092074	769.5E6	17.715m	21.008
28) SA Decachlor...	9.063	8.057	76301241	632.0E6	15.316	15.970

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
Data File : PD091441.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2025 20:47
Operator : AR\AJ
Sample : Q3742-10
Misc :
ALS Vial : 22 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
SB-1125-2


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 12/03/2025
Supervised By : Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 01:11:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.540 min
Delta R.T.: -0.004 min
Response: 61092074
Conc: 17.71 ng/ml

Instrument :
ECD_D
ClientSampleId :
SB-1125-2


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 12/03/2025
Supervised By :Sohil Jodhani 12/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.872 min
Delta R.T.: -0.002 min
Response: 769539691
Conc: 21.01 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.063 min
Delta R.T.: 0.000 min
Response: 76301241
Conc: 15.32 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.057 min
Delta R.T.: -0.001 min
Response: 631996770
Conc: 15.97 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
 Data File : PD091442.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Dec 2025 21:01
 Operator : AR\AJ
 Sample : Q3742-16
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 SB-1125-1

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 12/03/2025
 Supervised By : Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 01:12:09 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.540	2.872	64900245	809.6E6	18.819m	22.102
28) SA Decachlor...	9.063	8.057	95600096	811.1E6	19.190	20.495
Target Compounds						
3) MA gamma-BHC...	4.321	3.719	327.3E6	2808.0E6	43.798	48.563
4) MA Heptachlor	4.919	4.071	10445966	114.2E6	1.459	2.069 #
8) B Heptachlo...	5.680	4.860	135.6E6	888.3E6	20.869	17.432
10) B gamma-Chl...	5.937	5.113	444.1E6	3233.8E6	68.321	59.197
11) B alpha-Chl...	6.022	5.177	358.2E6	1958.1E6	52.050	37.418 #
13) MA Dieldrin	6.337	5.498	5425719	55193887	0.839	1.016m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
Data File : PD091442.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2025 21:01
Operator : AR\AJ
Sample : Q3742-16
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :

ECD_D

Client Sample Id :

SB-1125-1

Manual Integrations



APPROVED


Reviewed By : Abdul Mirza 12/03/2025

Supervised By : Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 01:12:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.540 min
Delta R.T.: -0.004 min
Response: 64900245
Conc: 18.82 ng/ml

Instrument :

ECD_D

ClientSampleId :


SB-1125-1

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 12/03/2025

Supervised By :Sohil Jodhani 12/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.872 min
Delta R.T.: -0.002 min
Response: 809598643
Conc: 22.10 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.321 min
Delta R.T.: -0.002 min
Response: 327300065
Conc: 43.80 ng/ml



#3 gamma-BHC (Lindane)

R.T.: 3.719 min
Delta R.T.: -0.002 min
Response: 2807981294
Conc: 48.56 ng/ml




#4 Heptachlor

R.T.: 4.919 min
Delta R.T.: -0.002 min
Response: 10445966
Conc: 1.46 ng/ml

Instrument :
ECD_D
ClientSampleId :
SB-1125-1


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 12/03/2025
Supervised By :Sohil Jodhani 12/03/2025




#4 Heptachlor

R.T.: 4.071 min
Delta R.T.: -0.002 min
Response: 114218674
Conc: 2.07 ng/ml



#8 Heptachlor epoxide


R.T.: 5.680 min
Delta R.T.: -0.002 min
Response: 135583425
Conc: 20.87 ng/ml



#8 Heptachlor epoxide

R.T.: 4.860 min
Delta R.T.: -0.001 min
Response: 888308485
Conc: 17.43 ng/ml

Response_ Signal: PD091442.D\ECD1A.ch



#10 gamma-Chlordane

R.T.: 5.937 min
Delta R.T.: 0.000 min
Response: 444136459
Conc: 68.32 ng/ml

Instrument :

ECD_D

ClientSampleId :

SB-1125-1


Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 12/03/2025

Supervised By :Sohil Jodhani 12/03/2025


Time Response_ Signal: PD091442.D\ECD2B.ch



#10 gamma-Chlordane

R.T.: 5.113 min
Delta R.T.: -0.002 min
Response: 3233766844
Conc: 59.20 ng/ml


Time Response_ Signal: PD091442.D\ECD1A.ch



#11 alpha-Chlordane


R.T.: 6.022 min
Delta R.T.: 0.003 min
Response: 358249384
Conc: 52.05 ng/ml

Time Response_ Signal: PD091442.D\ECD2B.ch



#11 alpha-Chlordane

R.T.: 5.177 min
Delta R.T.: -0.002 min
Response: 1958111592
Conc: 37.42 ng/ml




#13 Dieldrin

R.T.: 6.337 min
Delta R.T.: -0.002 min
Response: 5425719
Conc: 0.84 ng/ml

Instrument :
ECD_D
ClientSampleId :
SB-1125-1


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 12/03/2025
Supervised By :Sohil Jodhani 12/03/2025




#13 Dieldrin

R.T.: 5.498 min
Delta R.T.: -0.003 min
Response: 55193887
Conc: 1.02 ng/ml m



#28 Decachlorobiphenyl

R.T.: 9.063 min
Delta R.T.: 0.000 min
Response: 95600096
Conc: 19.19 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.057 min
Delta R.T.: -0.002 min
Response: 811090399
Conc: 20.50 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
 Data File : PD091443.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Dec 2025 21:15
 Operator : AR\AJ
 Sample : Q3742-17
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 SB-1125-25

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 12/03/2025
 Supervised By :Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 01:12:18 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.540	2.870	48829556	583.8E6	14.159m	15.938m
28) SA Decachlor...	9.063	8.057	72839292	498.0E6	14.621	12.583

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
Data File : PD091443.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2025 21:15
Operator : AR\AJ
Sample : Q3742-17
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
SB-1125-25


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 12/03/2025
Supervised By : Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 01:12:18 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.540 min
Delta R.T.: -0.004 min
Response: 48829556
Conc: 14.16 ng/ml

Instrument :
ECD_D
ClientSampleId :
SB-1125-25


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 12/03/2025
Supervised By :Sohil Jodhani 12/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.870 min
Delta R.T.: -0.004 min
Response: 583805489
Conc: 15.94 ng/ml m



#28 Decachlorobiphenyl

R.T.: 9.063 min
Delta R.T.: 0.000 min
Response: 72839292
Conc: 14.62 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.057 min
Delta R.T.: -0.001 min
Response: 497953596
Conc: 12.58 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
 Data File : PD091444.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Dec 2025 21:28
 Operator : AR\AJ
 Sample : Q3742-18
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 SB-1125-24

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 12/03/2025
 Supervised By : Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 01:12:28 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.540	2.872	57178599	707.2E6	16.580m	19.306
28) SA Decachlor...	9.063	8.057	74798776	577.5E6	15.014m	14.592

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
Data File : PD091444.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2025 21:28
Operator : AR\AJ
Sample : Q3742-18
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

SB-1125-24

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 12/03/2025

Supervised By :Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 01:12:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.540 min
Delta R.T.: -0.004 min
Response: 57178599
Conc: 16.58 ng/ml

Instrument :
ECD_D
ClientSampleId :
SB-1125-24


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 12/03/2025
Supervised By :Sohil Jodhani 12/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.872 min
Delta R.T.: -0.002 min
Response: 707165635
Conc: 19.31 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.063 min
Delta R.T.: 0.000 min
Response: 74798776
Conc: 15.01 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.057 min
Delta R.T.: -0.001 min
Response: 577488146
Conc: 14.59 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
Data File : PD091419.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2025 13:37
Operator : AR\AJ
Sample : PB170776BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB170776BL

A
B
C
D
E
F
G
H
I
J
K
L

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 01:08:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.548	2.870	75177881	933.9E6	21.799	25.496
28)	SA Decachlor...	9.074	8.060	111.1E6	991.2E6	22.304	25.046

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
Data File : PD091419.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2025 13:37
Operator : AR\AJ
Sample : PB170776BL
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
PB170776BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 01:08:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PD091419.D\ECD1A.ch




#1 Tetrachloro-m-xylene

R.T.: 3.548 min
Delta R.T.: 0.004 min
Response: 75177881
Conc: 21.80 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB170776BL


Response_ Signal: PD091419.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.870 min
Delta R.T.: -0.005 min
Response: 933932038
Conc: 25.50 ng/ml


Response_ Signal: PD091419.D\ECD1A.ch



#28 Decachlorobiphenyl

R.T.: 9.074 min
Delta R.T.: 0.011 min
Response: 111113111
Conc: 22.30 ng/ml

Response_ Signal: PD091419.D\ECD2B.ch



#28 Decachlorobiphenyl

R.T.: 8.060 min
Delta R.T.: 0.001 min
Response: 991199531
Conc: 25.05 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120925\
 Data File : PD091526.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Dec 2025 14:22
 Operator : AR\AJ
 Sample : PB170776BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PB170776BS

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 12/10/2025

Supervised By :mohammad ahmed 12/10/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 10 00:37:49 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120825.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 09 04:29:17 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.870	63433818	807.9E6	18.263	20.503
28)	SA Decachlor...	9.063	8.057	107.4E6	930.6E6	20.663	22.472
Target Compounds							
2)	A alpha-BHC	3.994	3.382	380.2E6	3431.6E6	47.218	50.336
3)	MA gamma-BHC...	4.324	3.718	359.4E6	3154.2E6	47.817	56.095
4)	MA Heptachlor	4.922	4.070	351.7E6	3098.2E6	49.544	53.158
5)	MB Aldrin	5.263	4.355	348.5E6	3092.2E6	50.088	53.326
6)	B beta-BHC	4.511	4.014	136.9E6	1299.5E6	47.435m	54.237
7)	B delta-BHC	4.760	4.250	345.4E6	3148.6E6	57.736	52.577
8)	B Heptachlo...	5.683	4.858	306.3E6	2774.6E6	49.461	54.837
9)	A Endosulfan I	6.067	5.233	299.7E6	2632.1E6	49.044	53.728
10)	B gamma-Chl...	5.939	5.110	316.2E6	3022.7E6	49.582	54.149m
11)	B alpha-Chl...	6.019	5.176	319.2E6	2922.4E6	47.532	54.573
12)	B 4,4'-DDE	6.189	5.360	286.7E6	2964.1E6	49.357	54.083
13)	MA Dieldrin	6.340	5.498	322.3E6	2992.8E6	49.532	53.958
14)	MA Endrin	6.567	5.776	270.2E6	2715.0E6	48.182	51.933
15)	B Endosulfa...	6.780	6.067	269.0E6	2538.3E6	44.232	53.876
16)	A 4,4'-DDD	6.699	5.915	231.6E6	2496.8E6	50.307	54.889
17)	MA 4,4'-DDT	7.015	6.169	245.0E6	2555.1E6	47.506	53.906
18)	B Endrin al...	6.909	6.245	203.3E6	1900.4E6	48.666	53.245
19)	B Endosulfa...	7.144	6.469	255.0E6	2478.1E6	48.303	53.763
20)	A Methoxychlor	7.488	6.740	124.9E6	1251.1E6	46.454	53.015
21)	B Endrin ke...	7.624	6.978	275.9E6	2661.2E6	48.970	54.709
22)	Mirex	8.106	7.170	170.4E6	1714.8E6	45.853	52.321

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120925\
Data File : PD091526.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Dec 2025 14:22
Operator : AR\AJ
Sample : PB170776BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PB170776BS

Manual Integrations



APPROVED

Reviewed By :Yogesh Patel 12/10/2025

Supervised By :mohammad ahmed 12/10/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 10 00:37:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120825.M
Quant Title : GC Extractables
QLast Update : Tue Dec 09 04:29:17 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
 Data File : PD091437.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Dec 2025 19:53
 Operator : AR\AJ
 Sample : Q3742-07MS
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 SB-1125-18MS

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 12/03/2025
 Supervised By : Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 01:11:21 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.540	2.870	50394202	609.1E6	14.613m	16.628m
28)	SA Decachlor...	9.064	8.058	81143427	497.1E6	16.288	12.561
Target Compounds							
2)	A alpha-BHC	3.991	3.382	381.8E6	2806.1E6	47.754	44.399
3)	MA gamma-BHC...	4.322	3.719	347.3E6	2395.8E6	46.471	41.434
4)	MA Heptachlor	4.919	4.070	313.4E6	2192.1E6	43.760	39.705
5)	MB Aldrin	5.261	4.356	328.6E6	2404.7E6	45.290	42.592
6)	B beta-BHC	4.510	4.015	124.1E6	1027.1E6	42.718	42.746
7)	B delta-BHC	4.756	4.251	352.3E6	2418.8E6	47.403m	41.363
8)	B Heptachlo...	5.682	4.859	285.5E6	2023.4E6	43.943	39.706
9)	A Endosulfan I	6.065	5.234	269.0E6	1759.0E6	43.785	36.760
10)	B gamma-Chl...	5.936	5.112	277.4E6	2075.4E6	42.679m	37.991m
11)	B alpha-Chl...	6.018	5.177	280.1E6	1984.7E6	40.699	37.926
12)	B 4,4'-DDE	6.187	5.361	242.2E6	1987.0E6	41.762	37.181
13)	MA Dieldrin	6.338	5.500	274.2E6	2082.1E6	42.409	38.336
14)	MA Endrin	6.564	5.776	232.6E6	1823.0E6	46.871m	40.339
15)	B Endosulfa...	6.779	6.067	233.0E6	1736.4E6	41.906	37.199m
16)	A 4,4'-DDD	6.698	5.917	204.4E6	1560.7E6	41.625	31.937
17)	MA 4,4'-DDT	7.014	6.170	199.4E6	1584.8E6	48.224	40.498
18)	B Endrin al...	6.908	6.246	185.4E6	1372.9E6	44.502	38.904
19)	B Endosulfa...	7.143	6.470	224.3E6	1663.3E6	43.600	37.099
20)	A Methoxychlor	7.486	6.741	102.6E6	780.1E6	47.165	38.905
21)	B Endrin ke...	7.624	6.979	249.6E6	1778.2E6	44.303	34.897
22)	Mirex	8.104	7.172	130.6E6	1052.2E6	36.714m	31.750

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
Data File : PD091437.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2025 19:53
Operator : AR\AJ
Sample : Q3742-07MS
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

SB-1125-18MS

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 12/03/2025

Supervised By :Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 01:11:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
 Data File : PD091438.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Dec 2025 20:06
 Operator : AR\AJ
 Sample : Q3742-07MSD
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 SB-1125-18MSD

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 12/03/2025
 Supervised By : Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 01:11:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.540	2.872	49857228	610.5E6	14.457m	16.666
28)	SA Decachlor...	9.064	8.058	78868103	492.1E6	15.831	12.435
Target Compounds							
2)	A alpha-BHC	3.991	3.383	373.3E6	2753.1E6	46.683	43.560
3)	MA gamma-BHC...	4.322	3.719	341.2E6	2351.6E6	45.653	40.669
4)	MA Heptachlor	4.920	4.071	314.1E6	2199.7E6	43.852	39.843
5)	MB Aldrin	5.261	4.356	322.1E6	2360.5E6	44.393	41.810
6)	B beta-BHC	4.509	4.015	120.8E6	1013.6E6	41.591	42.185
7)	B delta-BHC	4.758	4.251	350.4E6	2390.2E6	47.142	40.873
8)	B Heptachlo...	5.681	4.860	280.9E6	1990.3E6	43.239	39.057
9)	A Endosulfan I	6.065	5.235	263.6E6	1734.9E6	42.903	36.256
10)	B gamma-Chl...	5.937	5.112	272.9E6	2046.5E6	41.986	37.463m
11)	B alpha-Chl...	6.018	5.178	276.9E6	1954.4E6	40.230	37.347
12)	B 4,4'-DDE	6.188	5.362	238.0E6	1963.8E6	41.041	36.745
13)	MA Dieldrin	6.338	5.500	269.0E6	2059.0E6	41.617	37.910
14)	MA Endrin	6.566	5.777	234.4E6	1814.1E6	47.218	40.142
15)	B Endosulfa...	6.778	6.069	228.0E6	1802.2E6	41.004	38.607
16)	A 4,4'-DDD	6.698	5.917	199.8E6	1533.8E6	40.688	31.386
17)	MA 4,4'-DDT	7.014	6.170	199.3E6	1604.8E6	48.197	41.010
18)	B Endrin al...	6.908	6.246	182.8E6	1357.2E6	43.874	38.460
19)	B Endosulfa...	7.143	6.470	220.1E6	1647.3E6	42.778	36.741
20)	A Methoxychlor	7.487	6.740	103.4E6	791.2E6	47.556	39.460
21)	B Endrin ke...	7.623	6.979	245.9E6	1756.2E6	43.654	34.465
22)	Mirex	8.104	7.172	127.4E6	1038.8E6	35.816m	31.347

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120225\
Data File : PD091438.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2025 20:06
Operator : AR\AJ
Sample : Q3742-07MSD
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

SB-1125-18MSD

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 12/03/2025

Supervised By :Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 01:11:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integration Report

Sequence:	PD111525	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PD091194.D	4,4"-DDE #2	Abdul	11/17/2025 8:53:25 AM	Sohil	11/17/2025 2:13:45	Peak Integrated by Software
PEM	PD091194.D	Endrin aldehyde	Abdul	11/17/2025 8:53:25 AM	Sohil	11/17/2025 2:13:45	Peak Integrated by Software
PEM	PD091194.D	Endrin aldehyde #2	Abdul	11/17/2025 8:53:25 AM	Sohil	11/17/2025 2:13:45	Peak Integrated by Software
PSTDICC005	PD091200.D	delta-BHC	Abdul	11/17/2025 8:54:07 AM	Sohil	11/17/2025 2:13:50	Peak Integrated by Software
PSTDICC005	PD091200.D	Heptachlor	Abdul	11/17/2025 8:54:07 AM	Sohil	11/17/2025 2:13:50	Peak Integrated by Software
PSTDICC005	PD091200.D	Heptachlor epoxide	Abdul	11/17/2025 8:54:07 AM	Sohil	11/17/2025 2:13:50	Peak Integrated by Software
PSTDICC005	PD091200.D	Tetrachloro-m-xylene	Abdul	11/17/2025 8:54:07 AM	Sohil	11/17/2025 2:13:50	Peak Integrated by Software
PSTDICC005	PD091200.D	Tetrachloro-m-xylene #2	Abdul	11/17/2025 8:54:07 AM	Sohil	11/17/2025 2:13:50	Peak Integrated by Software
PEM	PD091215.D	4,4"-DDE #2	Abdul	11/17/2025 8:53:18 AM	Sohil	11/17/2025 2:14:03	Peak Integrated by Software

Manual Integration Report

Sequence:	PD120225	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
I.BLK	PD091416.D	Tetrachloro-m-xylene	Abdul	12/3/2025 8:28:45 AM	Sohil	12/3/2025 3:24:17	Peak Integrated by Software
PEM	PD091417.D	4,4"-DDE #2	Abdul	12/3/2025 8:28:38 AM	Sohil	12/3/2025 3:24:23	Peak Integrated by Software
PEM	PD091417.D	Endrin ketone #2	Abdul	12/3/2025 8:28:38 AM	Sohil	12/3/2025 3:24:23	Peak Integrated by Software
PSTDCCC050	PD091426.D	4,4"-DDE #2	Abdul	12/3/2025 8:28:09 AM	Sohil	12/3/2025 3:24:40	Peak Integrated by Software
Q3742-01	PD091427.D	Tetrachloro-m-xylene	Abdul	12/3/2025 8:29:05 AM	Sohil	12/3/2025 3:24:44	Peak Integrated by Software
Q3742-02	PD091428.D	Decachlorobiphenyl	Abdul	12/3/2025 8:28:06 AM	Sohil	12/3/2025 3:24:46	Peak Integrated by Software
Q3742-02	PD091428.D	Tetrachloro-m-xylene	Abdul	12/3/2025 8:28:06 AM	Sohil	12/3/2025 3:24:46	Peak Integrated by Software
Q3742-02	PD091428.D	Tetrachloro-m-xylene #2	Abdul	12/3/2025 8:28:06 AM	Sohil	12/3/2025 3:24:46	Peak Integrated by Software
Q3742-05	PD091431.D	4,4"-DDE	Abdul	12/3/2025 10:40:54 AM	Sohil	12/3/2025 3:24:54	Peak Integrated by Software
Q3742-05	PD091431.D	4,4"-DDE #2	Abdul	12/3/2025 10:40:54 AM	Sohil	12/3/2025 3:24:54	Peak Integrated by Software
Q3742-05	PD091431.D	Dieldrin	Abdul	12/3/2025 10:40:54 AM	Sohil	12/3/2025 3:24:54	Peak Integrated by Software
Q3742-05	PD091431.D	Dieldrin #2	Abdul	12/3/2025 10:40:54 AM	Sohil	12/3/2025 3:24:54	Peak Integrated by Software
Q3742-05	PD091431.D	Endrin	Abdul	12/3/2025 10:40:54 AM	Sohil	12/3/2025 3:24:54	Peak Integrated by Software

Manual Integration Report

Sequence:	PD120225	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q3742-05	PD091431.D	Endrin #2	Abdul	12/3/2025 10:40:54 AM	Sohil	12/3/2025 3:24:54	Peak Integrated by Software
Q3742-05	PD091431.D	Tetrachloro-m-xylene	Abdul	12/3/2025 10:40:54 AM	Sohil	12/3/2025 3:24:54	Peak Integrated by Software
PEM	PD091433.D	4,4"-DDD	Abdul	12/3/2025 8:22:49 AM	Sohil	12/3/2025 3:24:57	Peak Integrated by Software
PEM	PD091433.D	Endrin aldehyde	Abdul	12/3/2025 8:22:49 AM	Sohil	12/3/2025 3:24:57	Peak Integrated by Software
PEM	PD091433.D	Endrin ketone #2	Abdul	12/3/2025 8:22:49 AM	Sohil	12/3/2025 3:24:57	Peak Integrated by Software
PSTDCCC050	PD091434.D	4,4"-DDT #2	Abdul	12/3/2025 10:40:48 AM	Sohil	12/3/2025 3:25:00	Peak Integrated by Software
PSTDCCC050	PD091434.D	Endrin #2	Abdul	12/3/2025 10:40:48 AM	Sohil	12/3/2025 3:25:00	Peak Integrated by Software
Q3742-06	PD091435.D	Tetrachloro-m-xylene	Abdul	12/3/2025 8:22:40 AM	Sohil	12/3/2025 3:25:03	Peak Integrated by Software
Q3742-06	PD091435.D	Tetrachloro-m-xylene #2	Abdul	12/3/2025 8:22:40 AM	Sohil	12/3/2025 3:25:03	Peak Integrated by Software
Q3742-07	PD091436.D	Tetrachloro-m-xylene	Abdul	12/3/2025 8:22:35 AM	Sohil	12/3/2025 3:25:05	Peak Integrated by Software
Q3742-07MS	PD091437.D	delta-BHC	Abdul	12/3/2025 8:22:31 AM	Sohil	12/3/2025 3:25:07	Peak Integrated by Software
Q3742-07MS	PD091437.D	Endosulfan II #2	Abdul	12/3/2025 8:22:31 AM	Sohil	12/3/2025 3:25:07	Peak Integrated by Software
Q3742-07MS	PD091437.D	Endrin	Abdul	12/3/2025 8:22:31 AM	Sohil	12/3/2025 3:25:07	Peak Integrated by Software

Manual Integration Report

Sequence:	PD120225	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q3742-07MS	PD091437.D	gamma-Chlordane	Abdul	12/3/2025 8:22:31 AM	Sohil	12/3/2025 3:25:07	Peak Integrated by Software
Q3742-07MS	PD091437.D	gamma-Chlordane #2	Abdul	12/3/2025 8:22:31 AM	Sohil	12/3/2025 3:25:07	Peak Integrated by Software
Q3742-07MS	PD091437.D	Mirex	Abdul	12/3/2025 8:22:31 AM	Sohil	12/3/2025 3:25:07	Peak Integrated by Software
Q3742-07MS	PD091437.D	Tetrachloro-m-xylene	Abdul	12/3/2025 8:22:31 AM	Sohil	12/3/2025 3:25:07	Peak Integrated by Software
Q3742-07MS	PD091437.D	Tetrachloro-m-xylene #2	Abdul	12/3/2025 8:22:31 AM	Sohil	12/3/2025 3:25:07	Peak Integrated by Software
Q3742-07MSD	PD091438.D	gamma-Chlordane #2	Abdul	12/3/2025 8:22:28 AM	Sohil	12/3/2025 3:25:09	Peak Integrated by Software
Q3742-07MSD	PD091438.D	Mirex	Abdul	12/3/2025 8:22:28 AM	Sohil	12/3/2025 3:25:09	Peak Integrated by Software
Q3742-07MSD	PD091438.D	Tetrachloro-m-xylene	Abdul	12/3/2025 8:22:28 AM	Sohil	12/3/2025 3:25:09	Peak Integrated by Software
Q3742-08	PD091439.D	Tetrachloro-m-xylene	Abdul	12/3/2025 8:22:23 AM	Sohil	12/3/2025 3:25:11	Peak Integrated by Software
Q3742-09	PD091440.D	Tetrachloro-m-xylene	Abdul	12/3/2025 8:27:37 AM	Sohil	12/3/2025 3:25:14	Peak Integrated by Software
Q3742-10	PD091441.D	Tetrachloro-m-xylene	Abdul	12/3/2025 8:22:19 AM	Sohil	12/3/2025 3:25:15	Peak Integrated by Software
Q3742-16	PD091442.D	Dieldrin #2	Abdul	12/3/2025 8:22:15 AM	Sohil	12/3/2025 3:25:18	Peak Integrated by Software
Q3742-16	PD091442.D	Tetrachloro-m-xylene	Abdul	12/3/2025 8:22:15 AM	Sohil	12/3/2025 3:25:18	Peak Integrated by Software

Manual Integration Report

Sequence:	PD120225	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q3742-17	PD091443.D	Tetrachloro-m-xylene	Abdul	12/3/2025 8:22:11 AM	Sohil	12/3/2025 3:25:27	Peak Integrated by Software
Q3742-17	PD091443.D	Tetrachloro-m-xylene #2	Abdul	12/3/2025 8:22:11 AM	Sohil	12/3/2025 3:25:27	Peak Integrated by Software
Q3742-18	PD091444.D	Decachlorobiphenyl	Abdul	12/3/2025 8:22:06 AM	Sohil	12/3/2025 3:25:30	Peak Integrated by Software
Q3742-18	PD091444.D	Tetrachloro-m-xylene	Abdul	12/3/2025 8:22:06 AM	Sohil	12/3/2025 3:25:30	Peak Integrated by Software
PSTDCCC050	PD091446.D	4,4"-DDE #2	Abdul	12/3/2025 8:21:56 AM	Sohil	12/3/2025 3:25:33	Peak Integrated by Software
PSTDCCC050	PD091446.D	Endrin ketone	Abdul	12/3/2025 8:21:56 AM	Sohil	12/3/2025 3:25:33	Peak Integrated by Software
PSTDCCC050	PD091446.D	gamma-Chlordane	Abdul	12/3/2025 8:21:56 AM	Sohil	12/3/2025 3:25:33	Peak Integrated by Software

Manual Integration Report

Sequence:	PD120825	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PD091479.D	Endrin aldehyde	Abdul	12/9/2025 8:50:37 AM	yogesh	12/9/2025 10:36:17	Peak Integrated by Software
RESCHK	PD091480.D	gamma-Chlordane #2	Abdul	12/9/2025 8:50:33 AM	yogesh	12/9/2025 10:36:18	Peak Integrated by Software
PSTDICC005	PD091485.D	delta-BHC	Abdul	12/9/2025 8:50:29 AM	yogesh	12/9/2025 10:36:20	Peak Integrated by Software
PSTDICC005	PD091485.D	Endosulfan II	Abdul	12/9/2025 8:50:29 AM	yogesh	12/9/2025 10:36:20	Peak Integrated by Software
PSTDICC005	PD091485.D	Endosulfan II #2	Abdul	12/9/2025 8:50:29 AM	yogesh	12/9/2025 10:36:20	Peak Integrated by Software
PSTDICC005	PD091485.D	Heptachlor epoxide	Abdul	12/9/2025 8:50:29 AM	yogesh	12/9/2025 10:36:20	Peak Integrated by Software
PSTDICC005	PD091485.D	Tetrachloro-m-xylene	Abdul	12/9/2025 8:50:29 AM	yogesh	12/9/2025 10:36:20	Peak Integrated by Software
PSTDICV050	PD091496.D	delta-BHC	Abdul	12/9/2025 8:50:20 AM	yogesh	12/9/2025 10:36:25	Peak Integrated by Software
PTOXICV500	PD091498.D	Toxaphene-5	Abdul	12/9/2025 8:50:16 AM	yogesh	12/9/2025 10:36:27	Peak Integrated by Software
PEM	PD091500.D	Endrin aldehyde	Abdul	12/9/2025 8:50:12 AM	yogesh	12/9/2025 10:36:30	Peak Integrated by Software
PSTDCCC050	PD091501.D	4,4"-DDE #2	Abdul	12/9/2025 8:50:09 AM	yogesh	12/9/2025 10:36:32	Peak Integrated by Software
PSTDCCC050	PD091513.D	4,4"-DDE #2	Abdul	12/9/2025 8:48:10 AM	yogesh	12/9/2025 10:36:51	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

PD120825

Instrument

ECD_d

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:	Pd120925	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PSTDCCC050	PD091525.D	4,4"-DDE #2	yogesh	12/10/2025 12:59:33 AM	mohammad	12/10/2025 1:01:23	Peak Integrated by Software
PSTDCCC050	PD091525.D	Aldrin #2	yogesh	12/10/2025 12:59:33 AM	mohammad	12/10/2025 1:01:23	Peak Integrated by Software
PSTDCCC050	PD091525.D	beta-BHC	yogesh	12/10/2025 12:59:33 AM	mohammad	12/10/2025 1:01:23	Peak Integrated by Software
PSTDCCC050	PD091525.D	Endosulfan II	yogesh	12/10/2025 12:59:33 AM	mohammad	12/10/2025 1:01:23	Peak Integrated by Software
PB170776BS	PD091526.D	beta-BHC	yogesh	12/10/2025 12:59:35 AM	mohammad	12/10/2025 1:01:23	Peak Integrated by Software
PB170776BS	PD091526.D	gamma-Chlordane #2	yogesh	12/10/2025 12:59:35 AM	mohammad	12/10/2025 1:01:23	Peak Integrated by Software
PSTDCCC050	PD091529.D	4,4"-DDE #2	yogesh	12/10/2025 12:59:38 AM	mohammad	12/10/2025 1:01:23	Peak Integrated by Software
PSTDCCC050	PD091529.D	Aldrin #2	yogesh	12/10/2025 12:59:38 AM	mohammad	12/10/2025 1:01:23	Peak Integrated by Software
PSTDCCC050	PD091529.D	Endrin	yogesh	12/10/2025 12:59:38 AM	mohammad	12/10/2025 1:01:23	Peak Integrated by Software
PSTDCCC050	PD091529.D	Endrin #2	yogesh	12/10/2025 12:59:38 AM	mohammad	12/10/2025 1:01:23	Peak Integrated by Software
PSTDCCC050	PD091529.D	Endrin ketone	yogesh	12/10/2025 12:59:38 AM	mohammad	12/10/2025 1:01:23	Peak Integrated by Software
PSTDCCC050	PD091529.D	Endrin ketone #2	yogesh	12/10/2025 12:59:38 AM	mohammad	12/10/2025 1:01:23	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	Pd120925	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
-----------	---------	-----------	-----------	-----------	---------------	---------------	--------

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111525

Review By	rahul	Review On	11/14/2025 3:12:58 PM
Supervise By	Sohil	Supervise On	11/17/2025 2:14:18 PM
SubDirectory	PD111525	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP25075		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	I.BLK	PD091193.D	14 Nov 2025 20:01	AR\AJ	Ok
2	PEM	PD091194.D	14 Nov 2025 20:15	AR\AJ	Ok,M
3	RESCHK	PD091195.D	14 Nov 2025 20:43	AR\AJ	Ok
4	PSTDICC100	PD091196.D	14 Nov 2025 20:56	AR\AJ	Ok
5	PSTDICC075	PD091197.D	14 Nov 2025 21:10	AR\AJ	Ok
6	PSTDICC050	PD091198.D	14 Nov 2025 21:24	AR\AJ	Ok
7	PSTDICC025	PD091199.D	14 Nov 2025 21:37	AR\AJ	Ok
8	PSTDICC005	PD091200.D	14 Nov 2025 21:51	AR\AJ	Ok,M
9	PCHLORICC1000	PD091201.D	14 Nov 2025 22:04	AR\AJ	Ok
10	PCHLORICC750	PD091202.D	14 Nov 2025 22:18	AR\AJ	Ok
11	PCHLORICC500	PD091203.D	14 Nov 2025 22:32	AR\AJ	Ok
12	PCHLORICC250	PD091204.D	14 Nov 2025 22:46	AR\AJ	Ok,M
13	PCHLORICC50	PD091205.D	14 Nov 2025 22:59	AR\AJ	Ok,M
14	PTOXICC1000	PD091206.D	14 Nov 2025 23:13	AR\AJ	Ok
15	PTOXICC750	PD091207.D	14 Nov 2025 23:27	AR\AJ	Ok
16	PTOXICC500	PD091208.D	14 Nov 2025 23:40	AR\AJ	Ok
17	PTOXICC250	PD091209.D	14 Nov 2025 23:54	AR\AJ	Ok
18	PTOXICC100	PD091210.D	15 Nov 2025 00:08	AR\AJ	Ok
19	PSTDICV050	PD091211.D	15 Nov 2025 00:21	AR\AJ	Ok
20	PCHLORICV050	PD091212.D	15 Nov 2025 00:49	AR\AJ	Ok
21	PTOXICV050	PD091213.D	15 Nov 2025 01:30	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111525

Review By	rahul	Review On	11/14/2025 3:12:58 PM
Supervise By	Sohil	Supervise On	11/17/2025 2:14:18 PM
SubDirectory	PD111525	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP25075		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	I.BLK	PD091214.D	15 Nov 2025 01:44	AR\AJ	Ok
23	PEM	PD091215.D	15 Nov 2025 01:57	AR\AJ	Ok,M
24	PSTDCCC050	PD091216.D	15 Nov 2025 02:11	AR\AJ	Ok

M : Manual Integration



Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD120225

Review By	Abdul	Review On	12/3/2025 8:29:21 AM
Supervise By	Sohil	Supervise On	12/3/2025 3:25:55 PM
SubDirectory	PD120225	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP25075		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD091415.D	02 Dec 2025 11:37	AR\AJ	Ok
2	I.BLK	PD091416.D	02 Dec 2025 11:50	AR\AJ	Ok,M
3	PEM	PD091417.D	02 Dec 2025 12:04	AR\AJ	Ok,M
4	PSTDCCC050	PD091418.D	02 Dec 2025 12:44	AR\AJ	Ok
5	PB170776BL	PD091419.D	02 Dec 2025 13:37	AR\AJ	Ok
6	PB170776BS	PD091420.D	02 Dec 2025 14:21	AR\AJ	Not Ok
7	Q3746-01	PD091421.D	02 Dec 2025 15:30	AR\AJ	Ok,M
8	Q3750-01	PD091422.D	02 Dec 2025 15:44	AR\AJ	Ok,M
9	Q3750-04	PD091423.D	02 Dec 2025 15:58	AR\AJ	Ok,M
10	Q3748-01	PD091424.D	02 Dec 2025 16:11	AR\AJ	Ok,M
11	I.BLK	PD091425.D	02 Dec 2025 16:25	AR\AJ	Ok
12	PSTDCCC050	PD091426.D	02 Dec 2025 16:39	AR\AJ	Ok,M
13	Q3742-01	PD091427.D	02 Dec 2025 17:23	AR\AJ	Ok,M
14	Q3742-02	PD091428.D	02 Dec 2025 17:37	AR\AJ	Ok,M
15	Q3742-03	PD091429.D	02 Dec 2025 17:50	AR\AJ	Not Ok
16	Q3742-04	PD091430.D	02 Dec 2025 18:04	AR\AJ	Not Ok
17	Q3742-05	PD091431.D	02 Dec 2025 18:17	AR\AJ	Ok,M
18	I.BLK	PD091432.D	02 Dec 2025 18:31	AR\AJ	Ok
19	PEM	PD091433.D	02 Dec 2025 18:45	AR\AJ	Ok,M
20	PSTDCCC050	PD091434.D	02 Dec 2025 19:12	AR\AJ	Ok,M
21	Q3742-06	PD091435.D	02 Dec 2025 19:26	AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD120225

Review By	Abdul	Review On	12/3/2025 8:29:21 AM
Supervise By	Sohil	Supervise On	12/3/2025 3:25:55 PM
SubDirectory	PD120225	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP25075		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q3742-07	PD091436.D	02 Dec 2025 19:39	AR\AJ	Ok,M
23	Q3742-07MS	PD091437.D	02 Dec 2025 19:53	AR\AJ	Ok,M
24	Q3742-07MSD	PD091438.D	02 Dec 2025 20:06	AR\AJ	Ok,M
25	Q3742-08	PD091439.D	02 Dec 2025 20:20	AR\AJ	Ok,M
26	Q3742-09	PD091440.D	02 Dec 2025 20:34	AR\AJ	Ok,M
27	Q3742-10	PD091441.D	02 Dec 2025 20:47	AR\AJ	Ok,M
28	Q3742-16	PD091442.D	02 Dec 2025 21:01	AR\AJ	Ok,M
29	Q3742-17	PD091443.D	02 Dec 2025 21:15	AR\AJ	Ok,M
30	Q3742-18	PD091444.D	02 Dec 2025 21:28	AR\AJ	Ok,M
31	I.BLK	PD091445.D	02 Dec 2025 21:42	AR\AJ	Ok
32	PSTDCCC050	PD091446.D	02 Dec 2025 21:55	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD120825

Review By	Abdul	Review On	12/9/2025 8:51:43 AM
Supervise By	yogesh	Supervise On	12/9/2025 10:37:14 AM
SubDirectory	PD120825	HP Acquire Method	HP Processing Method pd120825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP25089		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761PP24754,PP24759,PP24761PP24754,PP24759,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD091477.D	08 Dec 2025 12:32	AR\AJ	Ok
2	I.BLK	PD091478.D	08 Dec 2025 12:45	AR\AJ	Ok
3	PEM	PD091479.D	08 Dec 2025 12:59	AR\AJ	Ok,M
4	RESCHK	PD091480.D	08 Dec 2025 14:55	AR\AJ	Ok,M
5	PSTDICC100	PD091481.D	08 Dec 2025 15:09	AR\AJ	Ok
6	PSTDICC075	PD091482.D	08 Dec 2025 15:22	AR\AJ	Ok
7	PSTDICC050	PD091483.D	08 Dec 2025 15:36	AR\AJ	Ok
8	PSTDICC025	PD091484.D	08 Dec 2025 15:50	AR\AJ	Ok
9	PSTDICC005	PD091485.D	08 Dec 2025 16:03	AR\AJ	Ok,M
10	PCHLORICC1000	PD091486.D	08 Dec 2025 16:17	AR\AJ	Ok
11	PCHLORICC750	PD091487.D	08 Dec 2025 16:31	AR\AJ	Ok
12	PCHLORICC500	PD091488.D	08 Dec 2025 16:44	AR\AJ	Ok
13	PCHLORICC250	PD091489.D	08 Dec 2025 16:58	AR\AJ	Ok
14	PCHLORICC050	PD091490.D	08 Dec 2025 17:12	AR\AJ	Ok,M
15	PTOXICC1000	PD091491.D	08 Dec 2025 17:25	AR\AJ	Not Ok
16	PTOXICC750	PD091492.D	08 Dec 2025 17:39	AR\AJ	Not Ok
17	PTOXICC500	PD091493.D	08 Dec 2025 17:53	AR\AJ	Ok
18	PTOXICC250	PD091494.D	08 Dec 2025 18:06	AR\AJ	Not Ok
19	PTOXICC100	PD091495.D	08 Dec 2025 18:20	AR\AJ	Not Ok
20	PSTDICV050	PD091496.D	08 Dec 2025 18:34	AR\AJ	Ok,M
21	PCHLORICV500	PD091497.D	08 Dec 2025 18:47	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD120825

Review By	Abdul	Review On	12/9/2025 8:51:43 AM
Supervise By	yogesh	Supervise On	12/9/2025 10:37:14 AM
SubDirectory	PD120825	HP Acquire Method	HP Processing Method pd120825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP25089		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761PP24754,PP24759,PP24761PP24754,PP24759,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PTOXICV500	PD091498.D	09 Dec 2025 00:41	AR\AJ	Not Ok
23	I.BLK	PD091499.D	09 Dec 2025 00:54	AR\AJ	Ok
24	PEM	PD091500.D	09 Dec 2025 01:08	AR\AJ	Ok,M
25	PSTDCCC050	PD091501.D	09 Dec 2025 01:22	AR\AJ	Ok,M
26	PB170864BL	PD091502.D	09 Dec 2025 01:35	AR\AJ	Ok,M
27	PB170864BS	PD091503.D	09 Dec 2025 01:49	AR\AJ	Ok
28	Q3795-01	PD091504.D	09 Dec 2025 02:02	AR\AJ	Ok,M
29	Q3795-01MS	PD091505.D	09 Dec 2025 02:16	AR\AJ	Ok,M
30	Q3795-01MSD	PD091506.D	09 Dec 2025 02:30	AR\AJ	Ok,M
31	Q3795-05	PD091507.D	09 Dec 2025 02:43	AR\AJ	Not Ok
32	Q3801-01	PD091508.D	09 Dec 2025 02:57	AR\AJ	Ok,M
33	Q3801-06	PD091509.D	09 Dec 2025 03:11	AR\AJ	Ok,M
34	Q3806-01	PD091510.D	09 Dec 2025 03:24	AR\AJ	Ok,M
35	Q3806-04	PD091511.D	09 Dec 2025 03:38	AR\AJ	Ok,M
36	I.BLK	PD091512.D	09 Dec 2025 03:52	AR\AJ	Ok
37	PSTDCCC050	PD091513.D	09 Dec 2025 04:05	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD120925

Review By	yogesh	Review On	12/10/2025 12:59:58 AM
Supervise By	mohammad	Supervise On	12/10/2025 1:01:23 AM
SubDirectory	PD120925	HP Acquire Method	HP Processing Method pd120625 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP25089		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761PP24754,PP24759,PP24761PP24754,PP24759,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD091514.D	09 Dec 2025 09:39	AR\AJ	Ok
2	I.BLK	PD091515.D	09 Dec 2025 09:52	AR\AJ	Ok
3	PEM	PD091516.D	09 Dec 2025 10:06	AR\AJ	Not Ok
4	PTOXICC1000	PD091517.D	09 Dec 2025 10:20	AR\AJ	Ok
5	PTOXICC750	PD091518.D	09 Dec 2025 10:37	AR\AJ	Ok
6	PTOXICC500	PD091519.D	09 Dec 2025 10:50	AR\AJ	Ok
7	PTOXICC250	PD091520.D	09 Dec 2025 11:04	AR\AJ	Ok
8	PTOXICC100	PD091521.D	09 Dec 2025 12:26	AR\AJ	Ok
9	PTOXICV500	PD091522.D	09 Dec 2025 12:53	AR\AJ	Ok
10	I.BLK	PD091523.D	09 Dec 2025 13:07	AR\AJ	Ok
11	PEM	PD091524.D	09 Dec 2025 13:20	AR\AJ	Ok
12	PSTDCCC050	PD091525.D	09 Dec 2025 14:05	AR\AJ	Ok,M
13	PB170776BS	PD091526.D	09 Dec 2025 14:22	AR\AJ	Ok,M
14	Q3795-05	PD091527.D	09 Dec 2025 14:50	AR\AJ	Ok,M
15	I.BLK	PD091528.D	09 Dec 2025 16:33	AR\AJ	Ok
16	PSTDCCC050	PD091529.D	09 Dec 2025 16:47	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111525

Review By	rahul	Review On	11/14/2025 3:12:58 PM
Supervise By	Sohil	Supervise On	11/17/2025 2:14:18 PM
SubDirectory	PD111525	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP25075		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	I.BLK	I.BLK	PD091193.D	14 Nov 2025 20:01		AR\AJ	Ok
2	PEM	PEM	PD091194.D	14 Nov 2025 20:15		AR\AJ	Ok,M
3	RESCHK	RESCHK	PD091195.D	14 Nov 2025 20:43		AR\AJ	Ok
4	PSTDICC100	PSTDICC100	PD091196.D	14 Nov 2025 20:56		AR\AJ	Ok
5	PSTDICC075	PSTDICC075	PD091197.D	14 Nov 2025 21:10		AR\AJ	Ok
6	PSTDICC050	PSTDICC050	PD091198.D	14 Nov 2025 21:24		AR\AJ	Ok
7	PSTDICC025	PSTDICC025	PD091199.D	14 Nov 2025 21:37		AR\AJ	Ok
8	PSTDICC005	PSTDICC005	PD091200.D	14 Nov 2025 21:51		AR\AJ	Ok,M
9	PCHLORICC1000	PCHLORICC1000	PD091201.D	14 Nov 2025 22:04		AR\AJ	Ok
10	PCHLORICC750	PCHLORICC750	PD091202.D	14 Nov 2025 22:18		AR\AJ	Ok
11	PCHLORICC500	PCHLORICC500	PD091203.D	14 Nov 2025 22:32		AR\AJ	Ok
12	PCHLORICC250	PCHLORICC250	PD091204.D	14 Nov 2025 22:46		AR\AJ	Ok,M
13	PCHLORICC50	PCHLORICC50	PD091205.D	14 Nov 2025 22:59		AR\AJ	Ok,M
14	PTOXICC1000	PTOXICC1000	PD091206.D	14 Nov 2025 23:13		AR\AJ	Ok
15	PTOXICC750	PTOXICC750	PD091207.D	14 Nov 2025 23:27		AR\AJ	Ok
16	PTOXICC500	PTOXICC500	PD091208.D	14 Nov 2025 23:40		AR\AJ	Ok
17	PTOXICC250	PTOXICC250	PD091209.D	14 Nov 2025 23:54		AR\AJ	Ok
18	PTOXICC100	PTOXICC100	PD091210.D	15 Nov 2025 00:08		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111525

Review By	rahul	Review On	11/14/2025 3:12:58 PM
Supervise By	Sohil	Supervise On	11/17/2025 2:14:18 PM
SubDirectory	PD111525	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP25075		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PSTDICV050	ICVPD111525	PD091211.D	15 Nov 2025 00:21		AR\AJ	Ok
20	PCHLORICV050	ICVPD111525CHLOR	PD091212.D	15 Nov 2025 00:49		AR\AJ	Ok
21	PTOXICV050	ICVPD111525TOX	PD091213.D	15 Nov 2025 01:30		AR\AJ	Ok
22	I.BLK	I.BLK	PD091214.D	15 Nov 2025 01:44		AR\AJ	Ok
23	PEM	PEM	PD091215.D	15 Nov 2025 01:57		AR\AJ	Ok,M
24	PSTDCCC050	PSTDCCC050	PD091216.D	15 Nov 2025 02:11		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD120225

Review By	Abdul	Review On	12/3/2025 8:29:21 AM
Supervise By	Sohil	Supervise On	12/3/2025 3:25:55 PM
SubDirectory	PD120225	HP Acquire Method	HP Processing Method pd111525 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24942,PP25075
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD091415.D	02 Dec 2025 11:37		AR\AJ	Ok
2	I.BLK	I.BLK	PD091416.D	02 Dec 2025 11:50		AR\AJ	Ok,M
3	PEM	PEM	PD091417.D	02 Dec 2025 12:04		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PD091418.D	02 Dec 2025 12:44		AR\AJ	Ok
5	PB170776BL	PB170776BL	PD091419.D	02 Dec 2025 13:37		AR\AJ	Ok
6	PB170776BS	PB170776BS	PD091420.D	02 Dec 2025 14:21	F flag in 2nd column for most of compounds	AR\AJ	Not Ok
7	Q3746-01	ROW	PD091421.D	02 Dec 2025 15:30		AR\AJ	Ok,M
8	Q3750-01	FENCE WC-1	PD091422.D	02 Dec 2025 15:44		AR\AJ	Ok,M
9	Q3750-04	FENCE WC-2	PD091423.D	02 Dec 2025 15:58		AR\AJ	Ok,M
10	Q3748-01	AU-05-12-1-2025	PD091424.D	02 Dec 2025 16:11		AR\AJ	Ok,M
11	I.BLK	I.BLK	PD091425.D	02 Dec 2025 16:25		AR\AJ	Ok
12	PSTDCCC050	PSTDCCC050	PD091426.D	02 Dec 2025 16:39		AR\AJ	Ok,M
13	Q3742-01	SB-1125-23	PD091427.D	02 Dec 2025 17:23		AR\AJ	Ok,M
14	Q3742-02	SB-1125-19	PD091428.D	02 Dec 2025 17:37		AR\AJ	Ok,M
15	Q3742-03	SB-1125-8	PD091429.D	02 Dec 2025 17:50		AR\AJ	Not Ok
16	Q3742-04	SB-1125-9	PD091430.D	02 Dec 2025 18:04		AR\AJ	Not Ok
17	Q3742-05	SB-1125-21	PD091431.D	02 Dec 2025 18:17		AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD120225

Review By	Abdul	Review On	12/3/2025 8:29:21 AM
Supervise By	Sohil	Supervise On	12/3/2025 3:25:55 PM
SubDirectory	PD120225	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP25075		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	I.BLK	I.BLK	PD091432.D	02 Dec 2025 18:31		AR\AJ	Ok
19	PEM	PEM	PD091433.D	02 Dec 2025 18:45		AR\AJ	Ok,M
20	PSTDCCC050	PSTDCCC050	PD091434.D	02 Dec 2025 19:12		AR\AJ	Ok,M
21	Q3742-06	SB-1125-13	PD091435.D	02 Dec 2025 19:26		AR\AJ	Ok,M
22	Q3742-07	SB-1125-18	PD091436.D	02 Dec 2025 19:39		AR\AJ	Ok,M
23	Q3742-07MS	SB-1125-18MS	PD091437.D	02 Dec 2025 19:53		AR\AJ	Ok,M
24	Q3742-07MSD	SB-1125-18MSD	PD091438.D	02 Dec 2025 20:06		AR\AJ	Ok,M
25	Q3742-08	SB-1125-20	PD091439.D	02 Dec 2025 20:20		AR\AJ	Ok,M
26	Q3742-09	SB-1125-22	PD091440.D	02 Dec 2025 20:34		AR\AJ	Ok,M
27	Q3742-10	SB-1125-2	PD091441.D	02 Dec 2025 20:47		AR\AJ	Ok,M
28	Q3742-16	SB-1125-1	PD091442.D	02 Dec 2025 21:01		AR\AJ	Ok,M
29	Q3742-17	SB-1125-25	PD091443.D	02 Dec 2025 21:15		AR\AJ	Ok,M
30	Q3742-18	SB-1125-24	PD091444.D	02 Dec 2025 21:28		AR\AJ	Ok,M
31	I.BLK	I.BLK	PD091445.D	02 Dec 2025 21:42		AR\AJ	Ok
32	PSTDCCC050	PSTDCCC050	PD091446.D	02 Dec 2025 21:55		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD120825

Review By	Abdul	Review On	12/9/2025 8:51:43 AM
Supervise By	yogesh	Supervise On	12/9/2025 10:37:14 AM
SubDirectory	PD120825	HP Acquire Method	HP Processing Method pd120825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24942,PP25089
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764
CCC	PP24751,PP24756,PP24761PP24754,PP24759,PP24761PP24754,PP24759,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD091477.D	08 Dec 2025 12:32		AR\AJ	Ok
2	I.BLK	I.BLK	PD091478.D	08 Dec 2025 12:45		AR\AJ	Ok
3	PEM	PEM	PD091479.D	08 Dec 2025 12:59		AR\AJ	Ok,M
4	RESCHK	RESCHK	PD091480.D	08 Dec 2025 14:55		AR\AJ	Ok,M
5	PSTDICC100	PSTDICC100	PD091481.D	08 Dec 2025 15:09		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PD091482.D	08 Dec 2025 15:22		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PD091483.D	08 Dec 2025 15:36		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PD091484.D	08 Dec 2025 15:50		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PD091485.D	08 Dec 2025 16:03		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PD091486.D	08 Dec 2025 16:17		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PD091487.D	08 Dec 2025 16:31		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PD091488.D	08 Dec 2025 16:44		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PD091489.D	08 Dec 2025 16:58		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PD091490.D	08 Dec 2025 17:12		AR\AJ	Ok,M
15	PTOXICC1000	PTOXICC1000	PD091491.D	08 Dec 2025 17:25	Not used	AR\AJ	Not Ok
16	PTOXICC750	PTOXICC750	PD091492.D	08 Dec 2025 17:39	Not used	AR\AJ	Not Ok
17	PTOXICC500	PTOXICC500	PD091493.D	08 Dec 2025 17:53		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PD091494.D	08 Dec 2025 18:06	Not used	AR\AJ	Not Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD120825

Review By	Abdul	Review On	12/9/2025 8:51:43 AM
Supervise By	yogesh	Supervise On	12/9/2025 10:37:14 AM
SubDirectory	PD120825	HP Acquire Method	HP Processing Method pd120825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP25089		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761PP24754,PP24759,PP24761PP24754,PP24759,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PD091495.D	08 Dec 2025 18:20	Not used	AR\AJ	Not Ok
20	PSTDICV050	ICVPD120825	PD091496.D	08 Dec 2025 18:34		AR\AJ	Ok,M
21	PCHLORICV500	ICVPD120825CHLOR	PD091497.D	08 Dec 2025 18:47		AR\AJ	Ok
22	PTOXICV500	ICVPD120825TOX	PD091498.D	09 Dec 2025 00:41	failing low	AR\AJ	Not Ok
23	I.BLK	I.BLK	PD091499.D	09 Dec 2025 00:54		AR\AJ	Ok
24	PEM	PEM	PD091500.D	09 Dec 2025 01:08		AR\AJ	Ok,M
25	PSTDCCC050	PSTDCCC050	PD091501.D	09 Dec 2025 01:22		AR\AJ	Ok,M
26	PB170864BL	PB170864BL	PD091502.D	09 Dec 2025 01:35		AR\AJ	Ok,M
27	PB170864BS	PB170864BS	PD091503.D	09 Dec 2025 01:49		AR\AJ	Ok
28	Q3795-01	72-11966	PD091504.D	09 Dec 2025 02:02		AR\AJ	Ok,M
29	Q3795-01MS	72-11966MS	PD091505.D	09 Dec 2025 02:16	some compounds fail for recovery	AR\AJ	Ok,M
30	Q3795-01MSD	72-11966MSD	PD091506.D	09 Dec 2025 02:30	some compounds fail for recovery	AR\AJ	Ok,M
31	Q3795-05	RBR 200059	PD091507.D	09 Dec 2025 02:43	Need to run again	AR\AJ	Not Ok
32	Q3801-01	SOIL#1	PD091508.D	09 Dec 2025 02:57		AR\AJ	Ok,M
33	Q3801-06	SOIL#2	PD091509.D	09 Dec 2025 03:11		AR\AJ	Ok,M
34	Q3806-01	TR-01-12-5-2025	PD091510.D	09 Dec 2025 03:24		AR\AJ	Ok,M
35	Q3806-04	TR-05-12-5-2025	PD091511.D	09 Dec 2025 03:38		AR\AJ	Ok,M
36	I.BLK	I.BLK	PD091512.D	09 Dec 2025 03:52		AR\AJ	Ok
37	PSTDCCC050	PSTDCCC050	PD091513.D	09 Dec 2025 04:05		AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD120825

Review By	Abdul	Review On	12/9/2025 8:51:43 AM	
Supervise By	yogesh	Supervise On	12/9/2025 10:37:14 AM	
SubDirectory	PD120825	HP Acquire Method	HP Processing Method	pd120825 8081
STD. NAME	STD REF.#			
Tune/Reschk	PP24942,PP25089			
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764			
CCC	PP24751,PP24756,PP24761PP24754,PP24759,PP24761PP24754,PP24759,PP24761			
Internal Standard/PEM				
ICV/I.BLK	PP24754,PP24759,PP24761			
Surrogate Standard				
MS/MSD Standard				
LCS Standard				

M : Manual Integration

A
B
C
D
E
F
G
H
I
J
K
L

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD120925

Review By	yogesh	Review On	12/10/2025 12:59:58 AM
Supervise By	mohammad	Supervise On	12/10/2025 1:01:23 AM
SubDirectory	PD120925	HP Acquire Method	HP Processing Method pd120625 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP25089		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761PP24754,PP24759,PP24761PP24754,PP24759,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD091514.D	09 Dec 2025 09:39		AR\AJ	Ok
2	I.BLK	I.BLK	PD091515.D	09 Dec 2025 09:52		AR\AJ	Ok
3	PEM	PEM	PD091516.D	09 Dec 2025 10:06	Not use	AR\AJ	Not Ok
4	PTOXICC1000	PTOXICC1000	PD091517.D	09 Dec 2025 10:20		AR\AJ	Ok
5	PTOXICC750	PTOXICC750	PD091518.D	09 Dec 2025 10:37		AR\AJ	Ok
6	PTOXICC500	PTOXICC500	PD091519.D	09 Dec 2025 10:50		AR\AJ	Ok
7	PTOXICC250	PTOXICC250	PD091520.D	09 Dec 2025 11:04		AR\AJ	Ok
8	PTOXICC100	PTOXICC100	PD091521.D	09 Dec 2025 12:26		AR\AJ	Ok
9	PTOXICV500	ICVPD120825TOX	PD091522.D	09 Dec 2025 12:53		AR\AJ	Ok
10	I.BLK	I.BLK	PD091523.D	09 Dec 2025 13:07		AR\AJ	Ok
11	PEM	PEM	PD091524.D	09 Dec 2025 13:20		AR\AJ	Ok
12	PSTDCCC050	PSTDCCC050	PD091525.D	09 Dec 2025 14:05		AR\AJ	Ok,M
13	PB170776BS	PB170776BS	PD091526.D	09 Dec 2025 14:22		AR\AJ	Ok,M
14	Q3795-05	RBR 200059	PD091527.D	09 Dec 2025 14:50		AR\AJ	Ok,M
15	I.BLK	I.BLK	PD091528.D	09 Dec 2025 16:33		AR\AJ	Ok
16	PSTDCCC050	PSTDCCC050	PD091529.D	09 Dec 2025 16:47		AR\AJ	Ok,M

M : Manual Integration

SOP ID: M3541-ASE Extraction-15

Clean Up SOP #: Florisil

Matrix : Solid

Weigh By: EH

Balance check: EH

Balance ID: EX-SC-2

pH Strip Lot#: N/A

Extraction By: RJ

Filter By: RJ

pH Meter ID: N/A

Hood ID: 3,7

Extraction Start Date : 12/02/2025

Extraction Start Time : 08:16

Extraction End Date : 12/02/2025

Extraction End Time : 11:30

Concentration By: EH

Supervisor By : RUPESH

Extraction Method: ☐ Separatory Funnel ☐ Continious Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☒ Soxhlet

Standarded Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	500 PPB	PP25070
Surrogate	1.0ML	200 PPB	PP25072
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2660
Baked Na2SO4	N/A	EP2663
Sand	N/A	E3951
Hexane	N/A	E3988
Florisil	N/A	E3976
9:1 Hexane:Acetone Mixture	N/A	EP2644
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS721.

KD Bath ID: N/A

Envap ID: N-EVAP-02

KD Bath Temperature: N/A

Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
12/2/25	RS (Ext-Lab)	Test/PCB Lab
11:35	Preparation Group	Analysis Group

Analytical Method: M3541-ASE Extraction-15

Concentration Date: 12/02/2025

Sample ID	Client Sample ID	Test	g/mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB170776BL	PBLK776	Pesticide-TCL	30.02	N/A	ritesh	Evelyn	10			U1-3
PB170776BS	PLCS776	Pesticide-TCL	30.01	N/A	ritesh	Evelyn	10			4
Q3742-01	SB-1125-23	Pesticide-TCL	30.06	N/A	ritesh	Evelyn	10	E		5
Q3742-02	SB-1125-19	Pesticide-TCL	30.02	N/A	ritesh	Evelyn	10	E		6
Q3742-03	SB-1125-8	Pesticide-TCL	30.09	N/A	ritesh	Evelyn	10	E		U2-1
Q3742-04	SB-1125-9	Pesticide-TCL	30.04	N/A	ritesh	Evelyn	10	E		2
Q3742-05	SB-1125-21	Pesticide-TCL	30.08	N/A	ritesh	Evelyn	10	E		3
Q3742-06	SB-1125-13	Pesticide-TCL	30.05	N/A	ritesh	Evelyn	10	E		4
Q3742-07	SB-1125-18	Pesticide-TCL	30.03	N/A	ritesh	Evelyn	10	E		5
Q3742-07MS	SB-1125-18MS	Pesticide-TCL	30.07	N/A	ritesh	Evelyn	10	E		6
Q3742-07MS D	SB-1125-18MSD	Pesticide-TCL	30.04	N/A	ritesh	Evelyn	10	E		U1-2
Q3742-08	SB-1125-20	Pesticide-TCL	30.05	N/A	ritesh	Evelyn	10	E		3
Q3742-09	SB-1125-22	Pesticide-TCL	30.03	N/A	ritesh	Evelyn	10	E		4
Q3742-10	SB-1125-2	Pesticide-TCL	30.06	N/A	ritesh	Evelyn	10	E		5
Q3742-16	SB-1125-1	Pesticide-TCL	30.05	N/A	ritesh	Evelyn	10	E		6
Q3742-17	SB-1125-25	Pesticide-TCL	30.06	N/A	ritesh	Evelyn	10	E		U2-1
Q3742-18	SB-1125-24	Pesticide-TCL	30.04	N/A	ritesh	Evelyn	10	E		2
Q3746-01	ROW	Pesticide-TCL	30.09	N/A	ritesh	Evelyn	10	E		3
Q3748-01	AU-05-12-1-2025	Pesticide-TCL	30.02	N/A	ritesh	Evelyn	10	E		4
Q3750-01	FENCE WC-1	Pesticide-TCL	30.06	N/A	ritesh	Evelyn	10	E		5
Q3750-04	FENCE WC-2	Pesticide-TCL	30.03	N/A	ritesh	Evelyn	10	E		6

* Extracts relinquished on the same date as received.

RS
12/2

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q3742P WorkList ID : 193410 Department : Extraction Date : 12-02-2025 08:11:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3742-01	SB-1125-23	Solid	Pesticide-TCL	Cool 4 deg C	REMI01		11/25/6025	8081B
Q3742-02	SB-1125-19	Solid	Pesticide-TCL	Cool 4 deg C	REMI01		11/25/2025	8081B
Q3742-03	SB-1125-8	Solid	Pesticide-TCL	Cool 4 deg C	REMI01		11/25/2025	8081B
Q3742-04	SB-1125-9	Solid	Pesticide-TCL	Cool 4 deg C	REMI01		11/25/2025	8081B
Q3742-05	SB-1125-21	Solid	Pesticide-TCL	Cool 4 deg C	REMI01		11/25/2025	8081B
Q3742-06	SB-1125-13	Solid	Pesticide-TCL	Cool 4 deg C	REMI01		11/25/2025	8081B
Q3742-07	SB-1125-18	Solid	Pesticide-TCL	Cool 4 deg C	REMI01		11/25/2025	8081B
Q3742-08	SB-1125-20	Solid	Pesticide-TCL	Cool 4 deg C	REMI01		11/25/2025	8081B
Q3742-09	SB-1125-22	Solid	Pesticide-TCL	Cool 4 deg C	REMI01		11/25/2025	8081B
Q3742-10	SB-1125-2	Solid	Pesticide-TCL	Cool 4 deg C	REMI01		11/25/2025	8081B
Q3742-16	SB-1125-1	Solid	Pesticide-TCL	Cool 4 deg C	REMI01		11/25/2025	8081B
Q3742-17	SB-1125-25	Solid	Pesticide-TCL	Cool 4 deg C	REMI01		11/25/2025	8081B
Q3742-18	SB-1125-24	Solid	Pesticide-TCL	Cool 4 deg C	REMI01		11/25/2025	8081B
Q3746-01	ROW	Solid	Pesticide-TCL	Cool 4 deg C	PSEG05	A11	12/02/2025	8081B
Q3748-01	AU-05-12-1-2025	Solid	Pesticide-TCL	Cool 4 deg C	PSEG05	A12	12/01/2025	8081B
Q3750-01	FENCE WC-1	Solid	Pesticide-TCL	Cool 4 deg C	PSEG05	A11	12/01/2025	8081B
Q3750-04	FENCE WC-2	Solid	Pesticide-TCL	Cool 4 deg C	PSEG05	A11	12/01/2025	8081B

Date/Time 12/02/25 8:11
 Raw Sample Received by: R3 C (6-21)-(261)
 Raw Sample Relinquished by: [Signature]

Date/Time 12/02/25 8:40
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: R3 C (6-21)-(261)

LAB CHRONICLE

OrderID:	Q3742	OrderDate:	12/1/2025 9:06:00 AM
Client:	Remington & Vernick	Project:	Saddler Property
Contact:	Justin Zarzecki	Location:	--Select--,A21

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3742-01	SB-1125-23	SOIL			11/26/25			11/26/25
			PCB	8082A		12/02/25	12/02/25	
			Pesticide-TCL	8081B		12/02/25	12/02/25	
			EPH_NF	NJEPH		12/02/25	12/02/25	
Q3742-02	SB-1125-19	SOIL			11/25/25			11/26/25
			PCB	8082A		12/02/25	12/02/25	
			Pesticide-TCL	8081B		12/02/25	12/02/25	
			EPH_NF	NJEPH		12/02/25	12/02/25	
Q3742-03	SB-1125-8	SOIL			11/25/25			11/26/25
			PCB	8082A		12/02/25	12/02/25	
			EPH_F2	NJEPH		12/04/25	12/04/25	
Q3742-04	SB-1125-9	SOIL			11/25/25			11/26/25
			PCB	8082A		12/02/25	12/02/25	
			EPH_F2	NJEPH		12/04/25	12/04/25	
Q3742-05	SB-1125-21	SOIL			11/25/25			11/26/25
			PCB	8082A		12/02/25	12/02/25	
			Pesticide-TCL	8081B		12/02/25	12/02/25	
			EPH_NF	NJEPH		12/02/25	12/02/25	
Q3742-06	SB-1125-13	SOIL			11/25/25			11/26/25
			PCB	8082A		12/02/25	12/02/25	
			Pesticide-TCL	8081B		12/02/25	12/02/25	
			EPH_NF	NJEPH		12/02/25	12/02/25	
Q3742-07	SB-1125-18	SOIL			11/25/25			11/26/25
			PCB	8082A		12/02/25	12/02/25	

LAB CHRONICLE

Q3742-08	SB-1125-20	SOIL	Pesticide-TCL	8081B	12/02/25	12/02/25	11/25/25	11/26/25
			EPH_NF	NJEPH	12/02/25	12/02/25		
Q3742-09	SB-1125-22	SOIL	PCB	8082A	12/02/25	12/02/25	11/25/25	11/26/25
			Pesticide-TCL	8081B	12/02/25	12/02/25		
			EPH_NF	NJEPH	12/02/25	12/02/25		
Q3742-10	SB-1125-2	SOIL	PCB	8082A	12/02/25	12/02/25	11/25/25	11/26/25
			Pesticide-TCL	8081B	12/02/25	12/02/25		
			EPH_NF	NJEPH	12/02/25	12/02/25		
Q3742-11	SB-1125-3A	Solid	PCB	8082A	12/02/25	12/02/25	11/25/25	11/26/25
			Pesticide-TCL	8081B	12/02/25	12/02/25		
Q3742-12	SB-1125-3B	Solid	EPH_F2	NJEPH	12/04/25	12/05/25	11/25/25	11/26/25
Q3742-13	SB-1125-3C	Solid	EPH_F2	NJEPH	12/04/25	12/05/25	11/25/25	11/26/25
			EPH_F2	NJEPH	12/04/25	12/09/25		
Q3742-14	SB-1125-3D	Solid	EPH_F2	NJEPH	12/04/25	12/05/25	11/25/25	11/26/25
Q3742-15	SB-1125-3E	Solid	EPH_F2	NJEPH	12/04/25	12/05/25	11/25/25	11/26/25
Q3742-16	SB-1125-1	SOIL	PCB	8082A	12/02/25	12/02/25	11/25/25	11/26/25
			Pesticide-TCL	8081B	12/02/25	12/02/25		
			EPH_NF	NJEPH	12/02/25	12/02/25		
Q3742-17	SB-1125-25	SOIL	PCB	8082A	12/02/25	12/02/25	11/25/25	11/26/25
			Pesticide-TCL	8081B	12/02/25	12/02/25		

LAB CHRONICLE							
Q3742-18	SB-1125-24	SOIL	EPH_NF	NJEPH	12/02/25	12/02/25	
					11/25/25		11/26/25
			PCB	8082A	12/02/25	12/02/25	
			Pesticide-TCL	8081B	12/02/25	12/02/25	
			EPH_NF	NJEPH	12/02/25	12/02/25	

A

C

D

E

F

G

H

I

J

K

L