

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120225\
 Data File : BP026207.D
 Acq On : 02 Dec 2025 14:31
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 02 14:56:51 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	76	-0.01
2	1,4-Dioxane	40.000	38.698	3.3	76	0.00
3	Pyridine	40.000	38.199	4.5	73	0.00
4	n-Nitrosodimethylamine	40.000	38.373	4.1	73	0.00
5 S	2-Fluorophenol	80.000	77.599	3.0	73	0.00
6	Aniline	40.000	37.310	6.7	70	0.00
7 S	Phenol-d6	80.000	76.590	4.3	71	-0.01
8	2-Chlorophenol	40.000	39.554	1.1	74	0.00
9	Benzaldehyde	40.000	41.516	-3.8	71	-0.01
10 C	Phenol	40.000	38.137	4.7	70	0.00
11	bis(2-Chloroethyl)ether	40.000	38.896	2.8	72	0.00
12	1,3-Dichlorobenzene	40.000	40.187	-0.5	76	-0.01
13 C	1,4-Dichlorobenzene	40.000	39.963	0.1	75	0.00
14	1,2-Dichlorobenzene	40.000	40.405	-1.0	76	0.00
15	Benzyl Alcohol	40.000	37.721	5.7	70	-0.01
16	2,2'-oxybis(1-Chloropropane	40.000	39.447	1.4	73	-0.01
17	2-Methylphenol	40.000	38.542	3.6	71	-0.02
18	Hexachloroethane	40.000	41.042	-2.6	77	-0.02
19 P	n-Nitroso-di-n-propylamine	40.000	38.105	4.7	69	-0.02
20	3+4-Methylphenols	40.000	37.520	6.2	70	-0.02
21 I	Naphthalene-d8	20.000	20.000	0.0	73	-0.02
22	Acetophenone	40.000	39.027	2.4	71	-0.02
23 S	Nitrobenzene-d5	80.000	78.766	1.5	71	-0.02
24	Nitrobenzene	40.000	39.217	2.0	71	-0.02
25	Isophorone	40.000	38.058	4.9	68	-0.02
26 C	2-Nitrophenol	40.000	40.176	-0.4	70	-0.02
27	2,4-Dimethylphenol	40.000	35.268	11.8	64	-0.02
28	bis(2-Chloroethoxy)methane	40.000	38.462	3.8	70	-0.02
29 C	2,4-Dichlorophenol	40.000	39.804	0.5	71	-0.02
30	1,2,4-Trichlorobenzene	40.000	41.020	-2.6	75	-0.02
31	Naphthalene	40.000	39.766	0.6	73	-0.02
32	Benzoic acid	40.000	33.149	17.1	60	-0.04
33	4-Chloroaniline	40.000	37.978	5.1	68	-0.02
34 C	Hexachlorobutadiene	40.000	42.427	-6.1	78	-0.03
35	Caprolactam	40.000	36.096	9.8	65	-0.02
36 C	4-Chloro-3-methylphenol	40.000	38.780	3.0	69	-0.02
37	2-Methylnaphthalene	40.000	39.695	0.8	72	-0.02
38	1-Methylnaphthalene	40.000	39.994	0.0	72	-0.02
39 I	Acenaphthene-d10	20.000	20.000	0.0	71	-0.04
40	1,2,4,5-Tetrachlorobenzene	40.000	41.192	-3.0	75	-0.04
41 P	Hexachlorocyclopentadiene	40.000	32.423	18.9	58	-0.03
42 S	2,4,6-Tribromophenol	80.000	85.041	-6.3	75	-0.03
43 C	2,4,6-Trichlorophenol	40.000	40.148	-0.4	71	-0.02
44	2,4,5-Trichlorophenol	40.000	39.824	0.4	71	-0.03
45 S	2-Fluorobiphenyl	80.000	81.152	-1.4	74	-0.04
46	1,1'-Biphenyl	40.000	39.061	2.3	71	-0.03
47	2-Chloronaphthalene	40.000	39.419	1.5	71	-0.04

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120225\
 Data File : BP026207.D
 Acq On : 02 Dec 2025 14:31
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 02 14:56:51 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.134	4.7	66	-0.03
49	Acenaphthylene	40.000	39.697	0.8	71	-0.03
50	Dimethylphthalate	40.000	39.565	1.1	71	-0.03
51	2,6-Dinitrotoluene	40.000	41.395	-3.5	69	-0.03
52 C	Acenaphthene	40.000	39.207	2.0	70	-0.03
53	3-Nitroaniline	40.000	37.717	5.7	65	-0.04
54 P	2,4-Dinitrophenol	40.000	36.442	8.9	63	-0.04
55	Dibenzofuran	40.000	40.218	-0.5	72	-0.03
56 P	4-Nitrophenol	40.000	35.534	11.2	63	-0.04
57	2,4-Dinitrotoluene	40.000	40.813	-2.0	69	-0.04
58	Fluorene	40.000	40.006	-0.0	73	-0.04
59	2,3,4,6-Tetrachlorophenol	40.000	40.699	-1.7	71	-0.03
60	Diethylphthalate	40.000	40.491	-1.2	72	-0.04
61	4-Chlorophenyl-phenylether	40.000	40.834	-2.1	73	-0.03
62	4-Nitroaniline	40.000	37.537	6.2	66	-0.03
63	Azobenzene	40.000	39.747	0.6	70	-0.03
64 I	Phenanthrene-d10	20.000	20.000	0.0	74	-0.04
65	4,6-Dinitro-2-methylphenol	40.000	38.420	3.9	67	-0.04
66 c	n-Nitrosodiphenylamine	40.000	39.239	1.9	72	-0.03
67	4-Bromophenyl-phenylether	40.000	40.619	-1.5	72	-0.03
68	Hexachlorobenzene	40.000	41.070	-2.7	74	-0.04
69	Atrazine	40.000	40.473	-1.2	73	-0.03
70 C	Pentachlorophenol	40.000	39.732	0.7	72	-0.04
71	Phenanthrene	40.000	39.741	0.6	74	-0.04
72	Anthracene	40.000	40.636	-1.6	74	-0.04
73	Carbazole	40.000	39.676	0.8	72	-0.02
74	Di-n-butylphthalate	40.000	41.156	-2.9	72	-0.04
75 C	Fluoranthene	40.000	41.429	-3.6	76	-0.02
76 I	Chrysene-d12	20.000	20.000	0.0	75	-0.03
77	Benzidine	40.000	37.124	7.2	72	-0.02
78	Pyrene	40.000	41.020	-2.6	76	-0.02
79 S	Terphenyl-d14	80.000	86.877	-8.6	81	-0.02
80	Butylbenzylphthalate	40.000	42.275	-5.7	76	-0.02
81	Benzo(a)anthracene	40.000	39.698	0.8	75	-0.03
82	3,3'-Dichlorobenzidine	40.000	37.103	7.2	70	-0.04
83	Chrysene	40.000	40.245	-0.6	76	-0.02
84	Bis(2-ethylhexyl)phthalate	40.000	43.400	-8.5	76	-0.04
85 c	Di-n-octyl phthalate	40.000	40.254	-0.6	73	-0.05
86 I	Perylene-d12	20.000	20.000	0.0	62	-0.06
87	Indeno(1,2,3-cd)pyrene	40.000	36.972	7.6	57	-0.11
88	Benzo(b)fluoranthene	40.000	42.607	-6.5	65	-0.06
89	Benzo(k)fluoranthene	40.000	44.379	-10.9	69	-0.06
90 C	Benzo(a)pyrene	40.000	40.888	-2.2	62	-0.06
91	Dibenzo(a,h)anthracene	40.000	37.068	7.3	57	-0.11
92	Benzo(g,h,i)perylene	40.000	37.204	7.0	57	-0.13

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120225\
Data File : BP026207.D
Acq On : 02 Dec 2025 14:31
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Dec 02 14:56:51 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 19 01:25:31 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0