

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF110525

Review By	rahul	Review On	11/5/2025 2:56:50 PM
Supervise By	Jagrut	Supervise On	11/6/2025 10:21:33 AM
SubDirectory	BF110525	HP Acquire Method	BNA_F
		HP Processing Method	BF110525

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13181,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF144168.D	05 Nov 2025 12:21		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF144169.D	05 Nov 2025 12:50		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF144170.D	05 Nov 2025 13:20		RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BF144171.D	05 Nov 2025 13:50		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF144172.D	05 Nov 2025 14:20		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF144173.D	05 Nov 2025 14:50	Compound #41 kept on QR	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF144174.D	05 Nov 2025 15:49		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF144175.D	05 Nov 2025 16:19		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF144176.D	05 Nov 2025 16:49		RC/JU	Ok
10	SSTDICV040	ICVBF110525	BF144177.D	05 Nov 2025 17:21		RC/JU	Ok,M
11	PB170372BL	PB170372BL	BF144178.D	05 Nov 2025 18:10		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF120225

Review By	Jagrut	Review On	12/3/2025 10:09:52 AM
Supervise By	Sohil	Supervise On	12/3/2025 3:31:56 PM
SubDirectory	BF120225	HP Acquire Method	BNA_F
		HP Processing Method	BF110525

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13184,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF144397.D	02 Dec 2025 10:31		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF144398.D	02 Dec 2025 11:00		RC/JU	Ok
3	PB170788BL	PB170788BL	BF144399.D	02 Dec 2025 13:16		RC/JU	Ok
4	PB170788BS	PB170788BS	BF144400.D	02 Dec 2025 13:47		RC/JU	Ok
5	Q3742-01	SB-1125-23	BF144401.D	02 Dec 2025 14:21		RC/JU	Ok,M
6	Q3742-03	SB-1125-8	BF144402.D	02 Dec 2025 14:51	Hit of # 09 (Below CRQL)	RC/JU	Not Ok
7	Q3742-07	SB-1125-18	BF144403.D	02 Dec 2025 15:20		RC/JU	Ok,M
8	Q3742-08	SB-1125-20	BF144404.D	02 Dec 2025 15:50		RC/JU	Ok,M
9	Q3742-09	SB-1125-22	BF144405.D	02 Dec 2025 16:19		RC/JU	Ok,M
10	Q3742-17	SB-1125-25	BF144406.D	02 Dec 2025 16:49		RC/JU	Ok,M
11	Q3742-18	SB-1125-24	BF144407.D	02 Dec 2025 17:19		RC/JU	Ok,M
12	Q3742-16	SB-1125-1	BF144408.D	02 Dec 2025 17:49		RC/JU	Ok
13	Q3746-01	ROW	BF144409.D	02 Dec 2025 18:19		RC/JU	Ok
14	Q3750-01	FENCE WC-1	BF144410.D	02 Dec 2025 18:49	Need 5X Dilution	RC/JU	Dilution
15	Q3750-04	FENCE WC-2	BF144411.D	02 Dec 2025 19:18		RC/JU	Ok,M
16	Q3748-01	AU-05-12-1-2025	BF144412.D	02 Dec 2025 19:48	Internal Standard Fail	RC/JU	Ok
17	Q3719-13	SB-1125-17A	BF144413.D	02 Dec 2025 20:18	Internal Standard Fail	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP111825

Review By	Jagrut	Review On	11/19/2025 10:53:24 AM
Supervise By	Sohil	Supervise On	11/19/2025 3:40:33 PM
SubDirectory	BP111825	HP Acquire Method	BNA_P
		HP Processing Method	BP111825

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6910,SP6911,SP6912,SP6913,SP6914,SP6915,SP6916,SP6917
CCC	SP6913
Internal Standard/PEM	S13183,10ul/1000ul sample
ICV/I.BLK	SP6928
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP026142.D	18 Nov 2025 13:45		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP026143.D	18 Nov 2025 14:31		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP026144.D	18 Nov 2025 15:12		RC/JU	Ok
4	SSTDICC010	SSTDICC010	BP026145.D	18 Nov 2025 15:54		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BP026146.D	18 Nov 2025 16:35		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BP026147.D	18 Nov 2025 17:57		RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BP026148.D	18 Nov 2025 18:38		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BP026149.D	18 Nov 2025 20:00		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BP026150.D	18 Nov 2025 21:22		RC/JU	Ok,M
10	SSTDICV040	ICVBP111825	BP026151.D	18 Nov 2025 23:25		RC/JU	Ok,M
11	PB170493TB	PB170493TB	BP026152.D	19 Nov 2025 00:06		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP120225

Review By	Jagrut	Review On	12/3/2025 10:12:44 AM
Supervise By	Sohil	Supervise On	12/3/2025 3:31:04 PM
SubDirectory	BP120225	HP Acquire Method	BNA_P
		HP Processing Method	BP111825

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6910,SP6911,SP6912,SP6913,SP6914,SP6915,SP6916,SP6917
CCC	SP6913
Internal Standard/PEM	S13184,10ul/1000ul sample
ICV/I.BLK	SP6928
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP026206.D	02 Dec 2025 13:50		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP026207.D	02 Dec 2025 14:31		RC/JU	Ok
3	PB170782BL	PB170782BL	BP026208.D	02 Dec 2025 15:12	Surrogate fail & internal standard fail	RC/JU	Not Ok
4	PB170782BS	PB170782BS	BP026209.D	02 Dec 2025 15:53		RC/JU	Ok
5	PB170782BSD	PB170782BSD	BP026210.D	02 Dec 2025 16:34		RC/JU	Ok
6	PB170782BL	PB170782BL	BP026211.D	02 Dec 2025 17:15		RC/JU	Ok
7	SP6937	SP6937	BP026212.D	02 Dec 2025 17:56	8270/625.1 SPIKE	RC/JU	Ok,M
8	Q3743-01	TW-1125-1	BP026213.D	02 Dec 2025 18:37		RC/JU	Ok
9	Q3743-02	TW-1125-2	BP026214.D	02 Dec 2025 19:18		RC/JU	Ok
10	Q3742-02	SB-1125-19	BP026215.D	02 Dec 2025 19:59		RC/JU	Ok,M
11	Q3742-02MS	SB-1125-19MS	BP026216.D	02 Dec 2025 20:40		RC/JU	Ok
12	Q3742-02MSD	SB-1125-19MSD	BP026217.D	02 Dec 2025 21:21		RC/JU	Ok
13	Q3742-04	SB-1125-9	BP026218.D	02 Dec 2025 22:02		RC/JU	Ok,M
14	Q3742-03	SB-1125-8	BP026219.D	02 Dec 2025 22:43		RC/JU	Ok,M
15	Q3742-05	SB-1125-21	BP026220.D	02 Dec 2025 23:24		RC/JU	Ok
16	Q3742-06	SB-1125-13	BP026221.D	03 Dec 2025 00:06		RC/JU	Ok
17	Q3742-10	SB-1125-2	BP026222.D	03 Dec 2025 00:47		RC/JU	Ok,M

M : Manual Integration