

Report of Analysis

Client: Remington & Vernick
Project: Saddler Property
Client Sample ID: SB-1125-24
Lab Sample ID: Q3742-18
Analytical Method: 8260D
Sample Wt/Vol: 4.82 g

Level: LOW
Final Vol: 5000 uL

Date Collected: 11/25/25
Date Received: 11/26/25
SDG No.: Q3742
Matrix: SOIL
% Solid: 62.8
Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	1.90	U	1	1.90	8.30	ug/Kg	12/04/25 13:32	VY120425
74-87-3	Chloromethane	1.90	U	1	1.90	8.30	ug/Kg	12/04/25 13:32	VY120425
75-01-4	Vinyl Chloride	1.30	U	1	1.30	8.30	ug/Kg	12/04/25 13:32	VY120425
74-83-9	Bromomethane	1.80	U	1	1.80	8.30	ug/Kg	12/04/25 13:32	VY120425
75-00-3	Chloroethane	2.10	U	1	2.10	8.30	ug/Kg	12/04/25 13:32	VY120425
75-69-4	Trichlorofluoromethane	2.00	U	1	2.00	8.30	ug/Kg	12/04/25 13:32	VY120425
76-13-1	1,1,2-Trichlorotrifluoroethane	1.80	U	1	1.80	8.30	ug/Kg	12/04/25 13:32	VY120425
75-35-4	1,1-Dichloroethene	1.70	U	1	1.70	8.30	ug/Kg	12/04/25 13:32	VY120425
67-64-1	Acetone	33.2	J	1	7.80	41.3	ug/Kg	12/04/25 13:32	VY120425
75-15-0	Carbon Disulfide	1.80	U	1	1.80	8.30	ug/Kg	12/04/25 13:32	VY120425
1634-04-4	Methyl tert-butyl Ether	1.20	U	1	1.20	8.30	ug/Kg	12/04/25 13:32	VY120425
79-20-9	Methyl Acetate	2.50	U	1	2.50	8.30	ug/Kg	12/04/25 13:32	VY120425
75-09-2	Methylene Chloride	5.80	U	1	5.80	16.5	ug/Kg	12/04/25 13:32	VY120425
156-60-5	trans-1,2-Dichloroethene	1.40	U	1	1.40	8.30	ug/Kg	12/04/25 13:32	VY120425
75-34-3	1,1-Dichloroethane	1.30	U	1	1.30	8.30	ug/Kg	12/04/25 13:32	VY120425
110-82-7	Cyclohexane	1.30	U	1	1.30	8.30	ug/Kg	12/04/25 13:32	VY120425
78-93-3	2-Butanone	10.8	U	1	10.8	41.3	ug/Kg	12/04/25 13:32	VY120425
56-23-5	Carbon Tetrachloride	1.60	U	1	1.60	8.30	ug/Kg	12/04/25 13:32	VY120425
156-59-2	cis-1,2-Dichloroethene	1.20	U	1	1.20	8.30	ug/Kg	12/04/25 13:32	VY120425
74-97-5	Bromochloromethane	1.90	U	1	1.90	8.30	ug/Kg	12/04/25 13:32	VY120425
67-66-3	Chloroform	1.40	U	1	1.40	8.30	ug/Kg	12/04/25 13:32	VY120425
71-55-6	1,1,1-Trichloroethane	1.50	U	1	1.50	8.30	ug/Kg	12/04/25 13:32	VY120425
108-87-2	Methylcyclohexane	1.50	U	1	1.50	8.30	ug/Kg	12/04/25 13:32	VY120425
71-43-2	Benzene	1.30	U	1	1.30	8.30	ug/Kg	12/04/25 13:32	VY120425
107-06-2	1,2-Dichloroethane	1.30	U	1	1.30	8.30	ug/Kg	12/04/25 13:32	VY120425
79-01-6	Trichloroethene	1.30	U	1	1.30	8.30	ug/Kg	12/04/25 13:32	VY120425
78-87-5	1,2-Dichloropropane	1.50	U	1	1.50	8.30	ug/Kg	12/04/25 13:32	VY120425
75-27-4	Bromodichloromethane	1.30	U	1	1.30	8.30	ug/Kg	12/04/25 13:32	VY120425
108-10-1	4-Methyl-2-Pentanone	5.90	U	1	5.90	41.3	ug/Kg	12/04/25 13:32	VY120425
108-88-3	Toluene	1.30	U	1	1.30	8.30	ug/Kg	12/04/25 13:32	VY120425
10061-02-6	t-1,3-Dichloropropene	1.10	U	1	1.10	8.30	ug/Kg	12/04/25 13:32	VY120425
10061-01-5	cis-1,3-Dichloropropene	1.00	U	1	1.00	8.30	ug/Kg	12/04/25 13:32	VY120425
79-00-5	1,1,2-Trichloroethane	1.50	U	1	1.50	8.30	ug/Kg	12/04/25 13:32	VY120425
591-78-6	2-Hexanone	6.10	U	1	6.10	41.3	ug/Kg	12/04/25 13:32	VY120425
124-48-1	Dibromochloromethane	1.40	U	1	1.40	8.30	ug/Kg	12/04/25 13:32	VY120425
106-93-4	1,2-Dibromoethane	1.50	U	1	1.50	8.30	ug/Kg	12/04/25 13:32	VY120425
127-18-4	Tetrachloroethene	1.70	U	1	1.70	8.30	ug/Kg	12/04/25 13:32	VY120425
108-90-7	Chlorobenzene	1.50	U	1	1.50	8.30	ug/Kg	12/04/25 13:32	VY120425
100-41-4	Ethyl Benzene	1.10	U	1	1.10	8.30	ug/Kg	12/04/25 13:32	VY120425
179601-23-1	m/p-Xylenes	2.00	U	1	2.00	16.5	ug/Kg	12/04/25 13:32	VY120425
95-47-6	o-Xylene	1.40	U	1	1.40	8.30	ug/Kg	12/04/25 13:32	VY120425
100-42-5	Styrene	1.20	U	1	1.20	8.30	ug/Kg	12/04/25 13:32	VY120425
75-25-2	Bromoform	1.40	U	1	1.40	8.30	ug/Kg	12/04/25 13:32	VY120425

Report of Analysis

Client:	Remington & Vernick	Date Collected:	11/25/25
Project:	Saddler Property	Date Received:	11/26/25
Client Sample ID:	SB-1125-24	SDG No.:	Q3742
Lab Sample ID:	Q3742-18	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	62.8
Sample Wt/Vol:	4.82 g	Test:	VOC-TCLVOA-10
	Level: LOW		
	Final Vol: 5000 uL		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	1.30	U	1	1.30	8.30	ug/Kg	12/04/25 13:32	VY120425
79-34-5	1,1,2,2-Tetrachloroethane	2.00	U	1	2.00	8.30	ug/Kg	12/04/25 13:32	VY120425
541-73-1	1,3-Dichlorobenzene	2.80	U	1	2.80	8.30	ug/Kg	12/04/25 13:32	VY120425
106-46-7	1,4-Dichlorobenzene	2.60	U	1	2.60	8.30	ug/Kg	12/04/25 13:32	VY120425
95-50-1	1,2-Dichlorobenzene	2.40	U	1	2.40	8.30	ug/Kg	12/04/25 13:32	VY120425
96-12-8	1,2-Dibromo-3-Chloropropane	3.00	U	1	3.00	8.30	ug/Kg	12/04/25 13:32	VY120425
120-82-1	1,2,4-Trichlorobenzene	4.90	U	1	4.90	8.30	ug/Kg	12/04/25 13:32	VY120425
87-61-6	1,2,3-Trichlorobenzene	5.30	U	1	5.30	8.30	ug/Kg	12/04/25 13:32	VY120425
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	52.8			63 - 155	106%	SPK: 50		
1868-53-7	Dibromofluoromethane	49.0			70 - 134	98%	SPK: 50		
2037-26-5	Toluene-d8	46.3			74 - 123	93%	SPK: 50		
460-00-4	4-Bromofluorobenzene	41.5			52 - 138	83%	SPK: 50		
INTERNAL STANDARDS									
		Area Count							
363-72-4	Pentafluorobenzene	754000							
540-36-3	1,4-Difluorobenzene	1320000							
3114-55-4	Chlorobenzene-d5	1140000							
3855-82-1	1,4-Dichlorobenzene-d4	498000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products