

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY112625

Review By	Semsettin Yesilyurt	Review On	12/1/2025 9:50:11 AM
Supervise By	John Carlone	Supervise On	12/1/2025 2:45:01 PM
SubDirectory	VY112625	HP Acquire Method	MSVOA_Y HP Processing Method 82y112625s.m

STD. NAME	STD REF.#
Tune/Reschk	VP136432
Initial Calibration Stds	VP136433,VP136434,VP136435,VP136436,VP136437,VP136438
CCC	VP136440
Internal Standard/PEM	VP136146
ICV/I.BLK	VP136439
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023808.D	26 Nov 2025 07:11		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY023809.D	26 Nov 2025 08:04		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY023810.D	26 Nov 2025 08:27		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY023811.D	26 Nov 2025 08:49		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY023812.D	26 Nov 2025 09:17		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY023813.D	26 Nov 2025 09:39		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY023814.D	26 Nov 2025 10:02		SY/MD	Ok,M
8	VIBLK	VIBLK	VY023815.D	26 Nov 2025 10:27		SY/MD	Ok
9	VSTDICV050	ICVVY112625	VY023816.D	26 Nov 2025 10:50		SY/MD	Ok,M
10	VY1126SBL01	VY1126SBL01	VY023817.D	26 Nov 2025 11:13		SY/MD	Ok
11	VY1126SBS01	VY1126SBS01	VY023818.D	26 Nov 2025 11:47		SY/MD	Ok,M
12	VY1126SBSD01	VY1126SBSD01	VY023819.D	26 Nov 2025 12:09		SY/MD	Ok,M
13	Q3604-04	S-1	VY023820.D	26 Nov 2025 12:34	not reported,vial A	SY/MD	Not Ok
14	Q3604-05	S-2	VY023821.D	26 Nov 2025 12:57	not reported,vial A	SY/MD	Not Ok
15	VIBLK	VIBLK	VY023822.D	26 Nov 2025 13:21		SY/MD	Ok
16	Q3721-01	EO-2-112525	VY023823.D	26 Nov 2025 13:44	vial A	SY/MD	Ok
17	Q3723-01	CC-1-112525	VY023824.D	26 Nov 2025 14:08	vial A	SY/MD	Ok
18	Q3725-04	STOCKPILE-SAMPLES	VY023825.D	26 Nov 2025 14:31	vial A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY112625

Review By	Semsettin Yesilyurt	Review On	12/1/2025 9:50:11 AM		
Supervise By	John Carlone	Supervise On	12/1/2025 2:45:01 PM		
SubDirectory	VY112625	HP Acquire Method	MSVOA_Y	HP Processing Method	82y112625s.m

STD. NAME	STD REF.#
Tune/Reschk	VP136432
Initial Calibration Stds	VP136433,VP136434,VP136435,VP136436,VP136437,VP136438
CCC	VP136440
Internal Standard/PEM	VP136146
ICV/I.BLK	VP136439
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	VIBLK	VIBLK	VY023826.D	26 Nov 2025 14:54		SY/MD	Ok
20	VSTDCCC050	VSTDCCC050EC	VY023827.D	26 Nov 2025 15:17		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY120125

Review By	John Carlone	Review On	12/2/2025 9:00:18 AM
Supervise By	Maresh Dadoda	Supervise On	12/2/2025 10:16:40 AM
SubDirectory	VY120125	HP Acquire Method	MSVOA_Y HP Processing Method 82y112625s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136454
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136455,VP136456 VP136146

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023828.D	01 Dec 2025 07:24		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY023829.D	01 Dec 2025 07:55		SY/MD	Ok,M
3	VY1201SBL01	VY1201SBL01	VY023830.D	01 Dec 2025 09:13		SY/MD	Ok
4	VY1201SBS01	VY1201SBS01	VY023831.D	01 Dec 2025 09:37		SY/MD	Ok,M
5	VY1201SBSD01	VY1201SBSD01	VY023832.D	01 Dec 2025 10:00		SY/MD	Ok,M
6	Q3719-09RE	SB-1125-15ARE	VY023833.D	01 Dec 2025 10:38	Internal standard fail;Surrogate fail,vial B	SY/MD	Confirms
7	Q3719-13	SB-1125-17A	VY023834.D	01 Dec 2025 11:25	vial B	SY/MD	Ok
8	Q3735-01	BU-3-112625	VY023835.D	01 Dec 2025 12:02	Internal Standard Fail,vial A	SY/MD	ReRun
9	Q3740-01	WC1	VY023836.D	01 Dec 2025 12:26	vial A	SY/MD	Ok
10	Q3741-02	B50-SP13-112025	VY023837.D	01 Dec 2025 12:49	vial A	SY/MD	Ok
11	Q3741-03	B50-SP14-112025	VY023838.D	01 Dec 2025 13:12	vial A	SY/MD	Ok
12	Q3741-04	B50-SP9-112425	VY023839.D	01 Dec 2025 13:36	Surrogate fail,vial A	SY/MD	ReRun
13	Q3741-05	B50-SP11-112525	VY023840.D	01 Dec 2025 13:59	Internal standard fail;Surrogate fail,vial A	SY/MD	ReRun
14	Q3741-01	B50-SP3-111925	VY023841.D	01 Dec 2025 14:23	vial A	SY/MD	Ok
15	Q3741-06	B50-SP4-112525	VY023842.D	01 Dec 2025 14:46	Internal standard fail;Surrogate fail,vial A	SY/MD	ReRun
16	Q3742-01	SB-1125-23	VY023843.D	01 Dec 2025 15:10	vial A	SY/MD	Ok
17	Q3742-02	SB-1125-19	VY023844.D	01 Dec 2025 15:33	vial A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY120125

Review By	John Carlone	Review On	12/2/2025 9:00:18 AM		
Supervise By	Mahesh Dadoda	Supervise On	12/2/2025 10:16:40 AM		
SubDirectory	VY120125	HP Acquire Method	MSVOA_Y	HP Processing Method	82y112625s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136454
CCC	VP136455,VP136456
Internal Standard/PEM	VP136146
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	Q3742-03	SB-1125-8	VY023845.D	01 Dec 2025 15:57	vial A	SY/MD	Ok
19	Q3742-04	SB-1125-9	VY023846.D	01 Dec 2025 16:20	vial A	SY/MD	Ok
20	VSTDCCC050	VSTDCCC050EC	VY023847.D	01 Dec 2025 17:47		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY120225

Review By	Semsettin Yesilyurt	Review On	12/3/2025 12:40:53 PM
Supervise By	Maresh Dadoda	Supervise On	12/3/2025 2:29:19 PM
SubDirectory	VY120225	HP Acquire Method	MSVOA_Y HP Processing Method 82y112625s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	AP136460
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136461,VP136462 VP136146

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023848.D	02 Dec 2025 06:50		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY023849.D	02 Dec 2025 09:14	CCAL failed high for com.#16	SY/MD	Ok,M
3	VY1202SBL01	VY1202SBL01	VY023850.D	02 Dec 2025 10:24		SY/MD	Ok
4	VY1202SBS01	VY1202SBS01	VY023851.D	02 Dec 2025 10:53		SY/MD	Ok,M
5	VY1202SBSD01	VY1202SBSD01	VY023852.D	02 Dec 2025 11:16		SY/MD	Ok,M
6	Q3742-05	SB-1125-21	VY023853.D	02 Dec 2025 12:08	vial B	SY/MD	Ok
7	Q3735-01	BU-3-112625	VY023854.D	02 Dec 2025 12:31	vial B	SY/MD	Ok
8	Q3741-04RE	B50-SP9-112425RE	VY023855.D	02 Dec 2025 12:55	Internal standard fail;Surrogate fail,vial B	SY/MD	Confirms
9	Q3741-05RE	B50-SP11-112525RE	VY023856.D	02 Dec 2025 13:18	Surrogate fail,vial B	SY/MD	Confirms
10	Q3741-06RE	B50-SP4-112525RE	VY023857.D	02 Dec 2025 13:42	Internal standard fail;Not match with first run,vial B	SY/MD	Confirms
11	Q3746-01	ROW	VY023858.D	02 Dec 2025 14:05	vial A	SY/MD	Ok
12	Q3742-06	SB-1125-13	VY023859.D	02 Dec 2025 14:29	vial A	SY/MD	Ok
13	VIBLK	VIBLK	VY023860.D	02 Dec 2025 15:33		SY/MD	Ok
14	Q3742-08	SB-1125-20	VY023861.D	02 Dec 2025 15:57	vial A	SY/MD	Ok
15	Q3742-09	SB-1125-22	VY023862.D	02 Dec 2025 16:20	Internal Standard Fail; CCC Failed High,vial A	SY/MD	ReRun
16	VSTDCCC050	VSTDCCC050EC	VY023863.D	02 Dec 2025 17:20		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY120325

Review By	John Carlone	Review On	12/4/2025 8:43:11 AM
Supervise By	Semsettin Yesilyurt	Supervise On	12/4/2025 2:37:23 PM
SubDirectory	VY120325	HP Acquire Method	MSVOA_Y HP Processing Method 82y120325s.m

STD. NAME	STD REF.#
Tune/Reschk	VP136489
Initial Calibration Stds	VP136490,VP136491,VP136492,VP136493,VP136494,VP136495
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP136496
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023864.D	03 Dec 2025 07:40		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY023865.D	03 Dec 2025 08:11	Need ICAL	SY/MD	Not Ok
3	VIBLK	VIBLK	VY023866.D	03 Dec 2025 13:19		SY/MD	Ok
4	VSTDICC005	VSTDICC005	VY023867.D	03 Dec 2025 13:42	Low spiked	SY/MD	Not Ok
5	VSTDICC010	VSTDICC010	VY023868.D	03 Dec 2025 14:05	Method failed for Comp. #13	SY/MD	Ok,M
6	VSTDICC020	VSTDICC020	VY023869.D	03 Dec 2025 14:27		SY/MD	Ok,M
7	VSTDICCC050	VSTDICCC050	VY023870.D	03 Dec 2025 14:50	Comp. #10,11,20,56,58 are on Linear Regression	SY/MD	Ok,M
8	VSTDICC100	VSTDICC100	VY023871.D	03 Dec 2025 15:12		SY/MD	Ok,M
9	VSTDICC150	VSTDICC150	VY023872.D	03 Dec 2025 15:35		SY/MD	Ok,M
10	VIBLK	VIBLK	VY023873.D	03 Dec 2025 16:00		SY/MD	Ok
11	VSTDICC005	VSTDICC005	VY023874.D	03 Dec 2025 16:23		SY/MD	Ok,M
12	VSTDICV050	ICVVY120325	VY023875.D	03 Dec 2025 17:08	ICV marginally failed for comp. #16,28	SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY120425

Review By	John Carlone	Review On	12/5/2025 11:05:32 AM
Supervise By	Semsettin Yesilyurt	Supervise On	12/5/2025 2:15:05 PM
SubDirectory	VY120425	HP Acquire Method	MSVOA_Y HP Processing Method 82y120325s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136508
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136509,VP136510 VP136146

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023876.D	04 Dec 2025 07:00		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY023877.D	04 Dec 2025 07:31		SY/MD	Ok,M
3	VY1204SBL01	VY1204SBL01	VY023878.D	04 Dec 2025 08:40		SY/MD	Ok
4	VY1204SBS01	VY1204SBS01	VY023879.D	04 Dec 2025 09:04		SY/MD	Ok,M
5	VY1204SBSD01	VY1204SBSD01	VY023880.D	04 Dec 2025 09:27		SY/MD	Ok,M
6	Q3750-03	FENCE WC-1-VOC	VY023881.D	04 Dec 2025 09:52	vial A	SY/MD	Ok
7	Q3750-06	FENCE WC-2-VOC	VY023882.D	04 Dec 2025 10:15	Internal Standard Fail,vial A	SY/MD	ReRun
8	Q3767-01	346	VY023883.D	04 Dec 2025 10:39	vial A	SY/MD	Ok
9	Q3768-01	20071	VY023884.D	04 Dec 2025 11:02	Surrogate Fail;Internal Standard Fail,confirmation with MeOH,vial A	SY/MD	Ok
10	VIBLK	VIBLK	VY023885.D	04 Dec 2025 11:25		SY/MD	Ok
11	VIBLK	VIBLK	VY023886.D	04 Dec 2025 11:58		SY/MD	Ok
12	Q3750-06RE	FENCE WC-2-VOCRE	VY023887.D	04 Dec 2025 12:22	Internal Standard Fail,vial B	SY/MD	Confirms
13	Q3742-16	SB-1125-1	VY023888.D	04 Dec 2025 12:45	vial A	SY/MD	Ok
14	Q3742-17	SB-1125-25	VY023889.D	04 Dec 2025 13:09	vial A	SY/MD	Ok
15	Q3742-18	SB-1125-24	VY023890.D	04 Dec 2025 13:32	vial A	SY/MD	Ok
16	Q3753-02	MOO-25-334-337	VY023891.D	04 Dec 2025 13:56	vial A	SY/MD	Ok
17	Q3753-04	AR520-0002	VY023892.D	04 Dec 2025 14:19	vial A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY120425

Review By	John Carlone	Review On	12/5/2025 11:05:32 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	12/5/2025 2:15:05 PM		
SubDirectory	VY120425	HP Acquire Method	MSVOA_Y	HP Processing Method	82y120325s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP136508			
CCC		VP136509,VP136510			
Internal Standard/PEM		VP136146			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q3742-09	SB-1125-22	VY023893.D	04 Dec 2025 14:43	vial B	SY/MD	Ok
19	Q3742-07	SB-1125-18	VY023894.D	04 Dec 2025 15:06	vial B	SY/MD	Ok
20	Q3742-10	SB-1125-2	VY023895.D	04 Dec 2025 15:29	vial B	SY/MD	Ok
21	VSTDCCC050	VSTDCCC050EC	VY023896.D	04 Dec 2025 15:52		SY/MD	Ok,M

M : Manual Integration