

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: REMI01
Lab Code: ACE SDG NO.: Q3742
Lab File ID: VY023829.D Date Analyzed: 12/01/2025
Instrument ID: MSVOA_Y Time Analyzed: 07:55
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	565841	7.71	855261	8.62	736930	11.42
UPPER LIMIT	1131680	8.213	1710520	9.122	1473860	11.92
LOWER LIMIT	282921	7.213	427631	8.122	368465	10.92
EPA SAMPLE NO.						
SB-1125-23	656114	7.71	1207708	8.62	1048441	11.41
SB-1125-19	639201	7.71	1174775	8.62	1006083	11.41
SB-1125-8	608834	7.71	1113703	8.62	955695	11.41
SB-1125-9	620754	7.71	1130173	8.62	961054	11.41
VY1201SBL01	719047	7.71	1221329	8.62	995220	11.42
VY1201SBS01	535284	7.71	821167	8.62	693280	11.42
VY1201SBSD01	526036	7.71	818117	8.62	695554	11.42

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.