

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: REMI01
 Lab Code: ACE SDG NO.: Q3742
 Lab File ID: VY023877.D Date Analyzed: 12/04/2025
 Instrument ID: MSVOA_Y Time Analyzed: 07:31
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	520627	7.72	774985	8.62	669312	11.42
UPPER LIMIT	1041250	8.219	1549970	9.122	1338620	11.92
LOWER LIMIT	260314	7.219	387493	8.122	334656	10.92
EPA SAMPLE NO.						
SB-1125-18	781859	7.71	1291434	8.62	1105691	11.42
SB-1125-22	803856	7.71	1376820	8.62	1230434	11.42
SB-1125-2	804667	7.71	1357896	8.62	1169651	11.41
SB-1125-1	757757	7.71	1302856	8.62	1139174	11.42
SB-1125-25	676192	7.71	1168984	8.62	1032423	11.42
SB-1125-24	754176	7.71	1322794	8.62	1144208	11.42
VY1204SBL01	753303	7.71	1347495	8.62	1163906	11.42
VY1204SBS01	494998	7.71	762810	8.62	650156	11.42

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.