

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: REMI01
 Lab Code: ACE SDG NO.: Q3742
 Lab File ID: VY023849.D Date Analyzed: 12/02/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:14
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	345634	13.353				
UPPER LIMIT	691268	13.853				
LOWER LIMIT	172817	12.853				
EPA SAMPLE NO.						
SB-1125-21	346373	13.35				
SB-1125-13	379120	13.35				
SB-1125-20	471581	13.35				
VY1202SBL01	454228	13.35				
VY1202SBS01	339534	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.