

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: REMI01
 Lab Code: ACE SDG NO.: Q3742
 Lab File ID: VY023877.D Date Analyzed: 12/04/2025
 Instrument ID: MSVOA_Y Time Analyzed: 07:31
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	343252	13.353				
UPPER LIMIT	686504	13.853				
LOWER LIMIT	171626	12.853				
EPA SAMPLE NO.						
SB-1125-18	482687	13.35				
SB-1125-22	551526	13.35				
SB-1125-2	504866	13.35				
SB-1125-1	483490	13.35				
SB-1125-25	459524	13.35				
SB-1125-24	497891	13.35				
VY1204SBL01	486950	13.35				
VY1204SBS01	339898	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.