

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3742

Lab File ID: VY023808.D

BFB Injection Date: 11/26/2025

Instrument ID: MSVOA_Y

BFB Injection Time: 07:11

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.5
75	30.0 - 60.0% of mass 95	51.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	0.9 (1.1) 1
174	50.0 - 100.0% of mass 95	87.3
175	5.0 - 9.0% of mass 174	6.4 (7.4) 1
176	95.0 - 101.0% of mass 174	83.8 (96) 1
177	5.0 - 9.0% of mass 176	5.6 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VY023809.D	11/26/2025	08:04
VSTDIC010	VSTDIC010	VY023810.D	11/26/2025	08:27
VSTDIC020	VSTDIC020	VY023811.D	11/26/2025	08:49
VSTDIC050	VSTDIC050	VY023812.D	11/26/2025	09:17
VSTDIC100	VSTDIC100	VY023813.D	11/26/2025	09:39
VSTDIC150	VSTDIC150	VY023814.D	11/26/2025	10:02