

Initial and Continuing Calibration Blank Summary

Client: Remington & Vernick

SDG No.: Q3743

Project: Saddler Property

RunNo.: LB138097

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	0.0019	0.0025	J	0.00096	0.005	12/03/2025
Sample ID: CCB1 Cyanide	mg/L	0.002	0.0025	J	0.00096	0.005	12/03/2025
Sample ID: CCB2 Cyanide	mg/L	0.0019	0.0025	J	0.00096	0.005	12/03/2025
Sample ID: CCB3 Cyanide	mg/L	0.0015	0.0025	J	0.00096	0.005	12/03/2025
Sample ID: CCB4 Cyanide	mg/L	0.002	0.0025	J	0.00096	0.005	12/03/2025
Sample ID: CCB5 Cyanide	mg/L	0.0018	0.0025	J	0.00096	0.005	12/03/2025