

Duplicate Sample Summary

Client:	Remington & Vernick	SDG No.:	Q3743
Project:	Saddler Property	Sample ID:	Q3743-02
Client ID:	TW-1125-2DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.025		0.025		1	0		12/03/2025

Duplicate Sample Summary

Client:	Remington & Vernick	SDG No.:	Q3743
Project:	Saddler Property	Sample ID:	Q3743-02
Client ID:	TW-1125-2MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.054		0.054		1	0		12/03/2025