

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB138105

Review By	Janvi	Review On	12/4/2025 10:27:33 AM
Supervise By	jaswal	Supervise On	12/5/2025 11:29:34 AM
STD. NAME	STD REF.#		
ICAL Standard	MP88152,MP88162,MP88160,MP88158,MP88157,MP88156,MP88155,MP88154,MP88153		
ICV Standard	MP88163,MP88162		
CCV Standard	MP88164		
ICSA Standard	MP88165,MP88166		
CRI Standard	MP88162		
LCS Standard			
Chk Standard	MP88168,MP88169		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE	TUNE	12/03/25 09:17		Jaswal	OK
2	S0	S0	CAL1	12/03/25 10:05		Jaswal	OK
3	S2	S2	CAL3	12/03/25 10:08		Jaswal	OK
4	S3	S3	CAL4	12/03/25 10:12		Jaswal	OK
5	S4	S4	CAL5	12/03/25 10:15		Jaswal	OK
6	S5	S5	CAL6	12/03/25 10:18		Jaswal	OK
7	S6	S6	CAL7	12/03/25 10:21		Jaswal	OK
8	S7	S7	CAL8	12/03/25 10:23		Jaswal	OK
9	S8	S8	CAL9	12/03/25 10:26		Jaswal	OK
10	ICV01	ICV01	ICV	12/03/25 10:59		Jaswal	OK
11	LLICV01	LLICV01	LLICV	12/03/25 11:10		Jaswal	OK
12	ICB01	ICB01	ICB	12/03/25 11:13		Jaswal	OK
13	ICSA01	ICSA01	ICSA	12/03/25 11:18		Jaswal	OK
14	ICSAB01	ICSAB01	ICSAB	12/03/25 11:21		Jaswal	OK
15	CCV01	CCV01	CCV	12/03/25 11:24		Jaswal	OK
16	CCB01	CCB01	CCB	12/03/25 11:27		Jaswal	OK
17	CRI	CRI	CRDL	12/03/25 11:42		Jaswal	OK
18	PB170780BL	PB170780BL	MB	12/03/25 11:45		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB138105

Review By	Janvi	Review On	12/4/2025 10:27:33 AM
Supervise By	jaswal	Supervise On	12/5/2025 11:29:34 AM
STD. NAME	STD REF.#		
ICAL Standard	MP88152,MP88162,MP88160,MP88158,MP88157,MP88156,MP88155,MP88154,MP88153		
ICV Standard	MP88163,MP88162		
CCV Standard	MP88164		
ICSA Standard	MP88165,MP88166		
CRI Standard	MP88162		
LCS Standard			
Chk Standard	MP88168,MP88169		

19	PB170780BS	PB170780BS	LCS	12/03/25 11:49		Jaswal	OK
20	Q3743-01DL	TW-1125-1DL	SAM	12/03/25 11:52		Jaswal	OK
21	Q3743-02DL	TW-1125-2DL	SAM	12/03/25 11:55		Jaswal	OK
22	Q3743-02DUPDL	TW-1125-2DUPDL	DUP	12/03/25 11:58		Jaswal	OK
23	Q3743-02LDL	TW-1125-2LDL	SD	12/03/25 12:01		Jaswal	OK
24	Q3743-02MSDL	TW-1125-2MSDL	MS	12/03/25 12:04		Jaswal	OK
25	Q3743-02MSDDL	TW-1125-2MSDDL	MSD	12/03/25 12:07		Jaswal	OK
26	Q3743-02ADL	TW-1125-2ADL	PS	12/03/25 12:10	0.1 ML OF MP88170 AND 0.2ML OF EACH MP88171,MP88172,MP88 173 IN 10ML SAMPLE BEFORE DILUTION.	Jaswal	OK
27	CCV02	CCV02	CCV	12/03/25 12:20		Jaswal	OK
28	CCB02	CCB02	CCB	12/03/25 12:27		Jaswal	OK
29	Q3749-01DL	AUD-25-0204DL	SAM	12/03/25 12:51	K High , INT_STD 89Y(2) Fail	Jaswal	Dilution
30	Q3749-01DL2	AUD-25-0204DL2	SAM	12/03/25 12:57	25X For K , 25X For INT_STD 89Y(2)	Jaswal	Confirms
31	CCV03	CCV03	CCV	12/03/25 13:00		Jaswal	OK
32	CCB03	CCB03	CCB	12/03/25 13:05		Jaswal	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB138109

Review By	mohan	Review On	12/5/2025 1:12:25 PM
Supervise By	jaswal	Supervise On	12/5/2025 8:54:44 PM

STD. NAME	STD REF.#
ICAL Standard	MP88196,MP88197,MP88198,MP88199,MP88200,MP88201
ICV Standard	MP88202
CCV Standard	MP88204
ICSA Standard	
CRI Standard	MP88206
LCS Standard	
Chk Standard	MP88203,MP88205,MP88207,MP88209

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	12/04/25 09:24		mohan	OK
2	S0.2	S0.2	CAL2	12/04/25 09:26		mohan	OK
3	S2.5	S2.5	CAL3	12/04/25 09:28		mohan	OK
4	S5	S5	CAL4	12/04/25 09:31		mohan	OK
5	S7.5	S7.5	CAL5	12/04/25 09:33		mohan	OK
6	S10	S10	CAL6	12/04/25 09:41		mohan	OK
7	ICV137	ICV137	ICV	12/04/25 09:45		mohan	OK
8	ICB137	ICB137	ICB	12/04/25 09:47		mohan	OK
9	CCV59	CCV59	CCV	12/04/25 09:51		mohan	OK
10	CCB59	CCB59	CCB	12/04/25 09:54		mohan	OK
11	CRA	CRA	CRDL	12/04/25 09:56		mohan	OK
12	HighStd	HighStd	HIGH STD	12/04/25 09:58		mohan	OK
13	ChkStd	ChkStd	SAM	12/04/25 10:00		mohan	OK
14	PB170812BL	PB170812BL	MB	12/04/25 10:03		mohan	OK
15	PB170812BS	PB170812BS	LCS	12/04/25 10:05		mohan	OK
16	Q3750-07	FENCE WC-1	SAM	12/04/25 10:07		mohan	OK
17	Q3750-08	FENCE WC-2	SAM	12/04/25 10:10		mohan	OK
18	Q3753-03	MOO-25-334-337	SAM	12/04/25 10:12		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB138109

Review By	mohan	Review On	12/5/2025 1:12:25 PM
Supervise By	jaswal	Supervise On	12/5/2025 8:54:44 PM
STD. NAME	STD REF.#		
ICAL Standard	MP88196,MP88197,MP88198,MP88199,MP88200,MP88201		
ICV Standard	MP88202		
CCV Standard	MP88204		
ICSA Standard			
CRI Standard	MP88206		
LCS Standard			
Chk Standard	MP88203,MP88205,MP88207,MP88209		

19	Q3753-05	AR520-0002	SAM	12/04/25 10:14		mohan	OK
20	Q3753-05DUP	AR520-0002DUP	DUP	12/04/25 10:16		mohan	OK
21	CCV60	CCV60	CCV	12/04/25 10:19		mohan	OK
22	CCB60	CCB60	CCB	12/04/25 10:21		mohan	OK
23	Q3753-05MS	AR520-0002MS	MS	12/04/25 10:23		mohan	OK
24	Q3753-05MSD	AR520-0002MSD	MSD	12/04/25 10:25		mohan	OK
25	PB170813BL	PB170813BL	MB	12/04/25 10:28		mohan	OK
26	Q3743-01	TW-1125-1	SAM	12/04/25 10:37	Hg is high, Need dilution	mohan	Dilution
27	Q3743-01DUP	TW-1125-1DUP	DUP	12/04/25 10:40	Hg is high, Need dilution	mohan	Dilution
28	Q3743-01MS	TW-1125-1MS	MS	12/04/25 10:42	Hg is high, Need dilution	mohan	Dilution
29	Q3743-01MSD	TW-1125-1MSD	MSD	12/04/25 10:44	Hg is high, Need dilution	mohan	Dilution
30	Q3743-02	TW-1125-2	SAM	12/04/25 10:47		mohan	OK
31	Q3749-01	AUD-25-0204	SAM	12/04/25 10:49		mohan	OK
32	CCV61	CCV61	CCV	12/04/25 10:51		mohan	OK
33	CCB61	CCB61	CCB	12/04/25 10:54		mohan	OK
34	Q3771-02	Comp	SAM	12/04/25 10:56		mohan	OK
35	PB170785TB	PB170785TB	MB	12/04/25 10:58		mohan	OK
36	Q3753-05L	AR520-0002L	SD	12/04/25 11:00		mohan	OK
37	Q3753-05A	AR520-0002A	PS	12/04/25 11:03		mohan	OK
38	Q3743-01L	TW-1125-1L	SD	12/04/25 11:05		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB138109

Review By	mohan	Review On	12/5/2025 1:12:25 PM
Supervise By	jaswal	Supervise On	12/5/2025 8:54:44 PM

STD. NAME	STD REF.#
ICAL Standard	MP88196,MP88197,MP88198,MP88199,MP88200,MP88201
ICV Standard	MP88202
CCV Standard	MP88204
ICSA Standard	
CRI Standard	MP88206
LCS Standard	
Chk Standard	MP88203,MP88205,MP88207,MP88209

39	Q3743-01A	TW-1125-1A	PS	12/04/25 11:07	Hg is high, Need dilution	mohan	Dilution
40	CCV62	CCV62	CCV	12/04/25 11:10		mohan	OK
41	CCB62	CCB62	CCB	12/04/25 11:12		mohan	OK
42	Q3743-01DL	TW-1125-1DL	SAM	12/04/25 11:14	5x For Hg	mohan	Confirms
43	Q3743-01DUPDL	TW-1125-1DUPDL	DUP	12/04/25 11:16	5x For Hg	mohan	Confirms
44	Q3743-01MSDL	TW-1125-1MSDL	MS	12/04/25 11:19	5x For Hg	mohan	Confirms
45	Q3743-01MSDDL	TW-1125-1MSDDL	MSD	12/04/25 11:21	5x For Hg	mohan	Confirms
46	Q3743-01LDL	TW-1125-1LDL	SD	12/04/25 11:23	Not Required	mohan	Not Ok
47	Q3743-01ADL	TW-1125-1ADL	PS	12/04/25 11:26	5x For Hg	mohan	Confirms
48	PB170813BS	PB170813BS	LCS	12/04/25 12:16		mohan	OK
49	CCV63	CCV63	CCV	12/04/25 12:19		mohan	OK
50	CCB63	CCB63	CCB	12/04/25 12:21		mohan	OK