

Hit Summary Sheet
SW-846

SDG No.: Q3743 **Order ID:** Q3743
Client: Remington & Vernick **Project ID:** Saddler Property

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : TW-1125-1								
Q3743-01	TW-1125-1	Water	Aluminum	156000	D	9.70	100	ug/L
Q3743-01	TW-1125-1	Water	Antimony	4.25	JD	0.55	10.0	ug/L
Q3743-01	TW-1125-1	Water	Arsenic	323	D	0.45	5.00	ug/L
Q3743-01	TW-1125-1	Water	Barium	2260	D	1.05	50.0	ug/L
Q3743-01	TW-1125-1	Water	Beryllium	15.2	D	1.60	5.00	ug/L
Q3743-01	TW-1125-1	Water	Cadmium	17.4	D	1.70	5.00	ug/L
Q3743-01	TW-1125-1	Water	Calcium	144000	D	229	2500	ug/L
Q3743-01	TW-1125-1	Water	Chromium	2560	D	1.05	10.0	ug/L
Q3743-01	TW-1125-1	Water	Cobalt	210	D	0.35	5.00	ug/L
Q3743-01	TW-1125-1	Water	Copper	956	D	1.50	10.0	ug/L
Q3743-01	TW-1125-1	Water	Iron	440000	D	39.1	250	ug/L
Q3743-01	TW-1125-1	Water	Lead	2360	D	1.05	5.00	ug/L
Q3743-01	TW-1125-1	Water	Magnesium	96000	D	97.5	2500	ug/L
Q3743-01	TW-1125-1	Water	Manganese	14000	D	2.15	5.00	ug/L
Q3743-01	TW-1125-1	Water	Nickel	424	D	1.35	5.00	ug/L
Q3743-01	TW-1125-1	Water	Mercury	11.9	D	0.38	1.00	ug/L
Q3743-01	TW-1125-1	Water	Potassium	26100	D	182	2500	ug/L
Q3743-01	TW-1125-1	Water	Selenium	14.7	JD	14.5	25.0	ug/L
Q3743-01	TW-1125-1	Water	Silver	4.55	JD	0.30	5.00	ug/L
Q3743-01	TW-1125-1	Water	Sodium	13400	D	640	2500	ug/L
Q3743-01	TW-1125-1	Water	Thallium	4.15	JD	0.30	5.00	ug/L
Q3743-01	TW-1125-1	Water	Vanadium	368	D	0.39	25.0	ug/L
Q3743-01	TW-1125-1	Water	Zinc	5950	D	6.25	25.0	ug/L
Client ID : TW-1125-2								
Q3743-02	TW-1125-2	Water	Aluminum	48700	D	9.70	100	ug/L
Q3743-02	TW-1125-2	Water	Antimony	3.15	JD	0.55	10.0	ug/L
Q3743-02	TW-1125-2	Water	Arsenic	310	D	0.45	5.00	ug/L
Q3743-02	TW-1125-2	Water	Barium	893	D	1.05	50.0	ug/L
Q3743-02	TW-1125-2	Water	Beryllium	5.05	D	1.60	5.00	ug/L
Q3743-02	TW-1125-2	Water	Cadmium	7.75	D	1.70	5.00	ug/L
Q3743-02	TW-1125-2	Water	Calcium	131000	D	229	2500	ug/L
Q3743-02	TW-1125-2	Water	Chromium	1400	D	1.05	10.0	ug/L
Q3743-02	TW-1125-2	Water	Cobalt	145	D	0.35	5.00	ug/L
Q3743-02	TW-1125-2	Water	Copper	462	D	1.50	10.0	ug/L
Q3743-02	TW-1125-2	Water	Iron	199000	D	39.1	250	ug/L
Q3743-02	TW-1125-2	Water	Lead	1100	D	1.05	5.00	ug/L
Q3743-02	TW-1125-2	Water	Magnesium	64300	D	97.5	2500	ug/L

Hit Summary Sheet
SW-846

SDG No.: Q3743

Order ID: Q3743

Client: Remington & Vernick

Project ID: Saddler Property

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q3743-02	TW-1125-2	Water	Manganese	4750	D	2.15	5.00	ug/L
Q3743-02	TW-1125-2	Water	Nickel	180	D	1.35	5.00	ug/L
Q3743-02	TW-1125-2	Water	Mercury	5.92		0.076	0.20	ug/L
Q3743-02	TW-1125-2	Water	Potassium	16900	D	182	2500	ug/L
Q3743-02	TW-1125-2	Water	Silver	3.05	JD	0.30	5.00	ug/L
Q3743-02	TW-1125-2	Water	Sodium	32700	D	640	2500	ug/L
Q3743-02	TW-1125-2	Water	Thallium	2.15	JD	0.30	5.00	ug/L
Q3743-02	TW-1125-2	Water	Vanadium	94.5	D	0.39	25.0	ug/L
Q3743-02	TW-1125-2	Water	Zinc	3510	D	6.25	25.0	ug/L



SAMPLE DATA

Report of Analysis

Client: Remington & Vernick
Project: Saddler Property
Client Sample ID: TW-1125-1
Lab Sample ID: Q3743-01
Level (low/med): low

Date Collected: 11/26/25
Date Received: 11/26/25
SDG No.: Q3743
Matrix: Water
% Solid: 0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	156000	D	5	9.70	100	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7440-36-0	Antimony	4.25	JD	5	0.55	10.0	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7440-38-2	Arsenic	323	D	5	0.45	5.00	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7440-39-3	Barium	2260	D	5	1.05	50.0	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7440-41-7	Beryllium	15.2	D	5	1.60	5.00	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7440-43-9	Cadmium	17.4	D	5	1.70	5.00	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7440-70-2	Calcium	144000	DN	5	229	2500	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7440-47-3	Chromium	2560	D	5	1.05	10.0	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7440-48-4	Cobalt	210	D	5	0.35	5.00	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7440-50-8	Copper	956	D	5	1.50	10.0	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7439-89-6	Iron	440000	D	5	39.1	250	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7439-92-1	Lead	2360	D	5	1.05	5.00	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7439-95-4	Magnesium	96000	D	5	97.5	2500	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7439-96-5	Manganese	14000	D	5	2.15	5.00	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7439-97-6	Mercury	11.9	DN	5	0.38	1.00	ug/L	12/04/25 06:30	12/04/25 11:14	7470A	M7470A
7440-02-0	Nickel	424	D	5	1.35	5.00	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7440-09-7	Potassium	26100	D	5	182	2500	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7782-49-2	Selenium	14.7	JD	5	14.5	25.0	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7440-22-4	Silver	4.55	JDN	5	0.30	5.00	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7440-23-5	Sodium	13400	D	5	640	2500	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7440-28-0	Thallium	4.15	JD	5	0.30	5.00	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7440-62-2	Vanadium	368	D	5	0.39	25.0	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A
7440-66-6	Zinc	5950	D	5	6.25	25.0	ug/L	12/02/25 09:30	12/03/25 11:52	6020B	3010A

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: Remington & Vernick
Project: Saddler Property
Client Sample ID: TW-1125-2
Lab Sample ID: Q3743-02
Level (low/med): low

Date Collected: 11/26/25
Date Received: 11/26/25
SDG No.: Q3743
Matrix: Water
% Solid: 0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	48700	D	5	9.70	100	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7440-36-0	Antimony	3.15	JD	5	0.55	10.0	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7440-38-2	Arsenic	310	D	5	0.45	5.00	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7440-39-3	Barium	893	D	5	1.05	50.0	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7440-41-7	Beryllium	5.05	D	5	1.60	5.00	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7440-43-9	Cadmium	7.75	D	5	1.70	5.00	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7440-70-2	Calcium	131000	DN	5	229	2500	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7440-47-3	Chromium	1400	D	5	1.05	10.0	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7440-48-4	Cobalt	145	D	5	0.35	5.00	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7440-50-8	Copper	462	D	5	1.50	10.0	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7439-89-6	Iron	199000	D	5	39.1	250	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7439-92-1	Lead	1100	D	5	1.05	5.00	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7439-95-4	Magnesium	64300	D	5	97.5	2500	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7439-96-5	Manganese	4750	D	5	2.15	5.00	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7439-97-6	Mercury	5.92	N	1	0.076	0.20	ug/L	12/04/25 06:30	12/04/25 10:47	7470A	M7470A
7440-02-0	Nickel	180	D	5	1.35	5.00	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7440-09-7	Potassium	16900	D	5	182	2500	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7782-49-2	Selenium	14.5	UD	5	14.5	25.0	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7440-22-4	Silver	3.05	JDN	5	0.30	5.00	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7440-23-5	Sodium	32700	D	5	640	2500	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7440-28-0	Thallium	2.15	JD	5	0.30	5.00	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7440-62-2	Vanadium	94.5	D	5	0.39	25.0	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A
7440-66-6	Zinc	3510	D	5	6.25	25.0	ug/L	12/02/25 09:30	12/03/25 11:55	6020B	3010A

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB137	Mercury	0.076	+/-0.2	U	0.20	CV	12/04/2025	09:47	LB138109

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick **SDG No.:** Q3743
Contract: REMI01 **Lab Code:** ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB59	Mercury	0.076	+/-0.2	U	0.20	CV	12/04/2025	09:54	LB138109
CCB60	Mercury	0.076	+/-0.2	U	0.20	CV	12/04/2025	10:21	LB138109
CCB61	Mercury	0.076	+/-0.2	U	0.20	CV	12/04/2025	10:54	LB138109
CCB62	Mercury	0.076	+/-0.2	U	0.20	CV	12/04/2025	11:12	LB138109
CCB63	Mercury	0.076	+/-0.2	U	0.20	CV	12/04/2025	12:21	LB138109

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	4.93	+/-10	J	20.0	P	12/03/2025	11:13	LB138105
	Antimony	0.11	+/-1	U	2.00	P	12/03/2025	11:13	LB138105
	Arsenic	0.089	+/-0.5	U	1.00	P	12/03/2025	11:13	LB138105
	Barium	0.21	+/-5	U	10.0	P	12/03/2025	11:13	LB138105
	Beryllium	0.32	+/-0.5	U	1.00	P	12/03/2025	11:13	LB138105
	Cadmium	0.34	+/-0.5	U	1.00	P	12/03/2025	11:13	LB138105
	Calcium	45.7	+/-250	U	500	P	12/03/2025	11:13	LB138105
	Chromium	0.21	+/-1	U	2.00	P	12/03/2025	11:13	LB138105
	Cobalt	0.070	+/-0.5	U	1.00	P	12/03/2025	11:13	LB138105
	Copper	0.30	+/-1	U	2.00	P	12/03/2025	11:13	LB138105
	Iron	7.81	+/-25	U	50.0	P	12/03/2025	11:13	LB138105
	Lead	0.21	+/-0.5	U	1.00	P	12/03/2025	11:13	LB138105
	Magnesium	19.5	+/-250	U	500	P	12/03/2025	11:13	LB138105
	Manganese	0.43	+/-0.5	U	1.00	P	12/03/2025	11:13	LB138105
	Nickel	0.27	+/-0.5	U	1.00	P	12/03/2025	11:13	LB138105
	Potassium	36.4	+/-250	U	500	P	12/03/2025	11:13	LB138105
	Selenium	2.90	+/-2.5	U	5.00	P	12/03/2025	11:13	LB138105
	Silver	0.060	+/-0.5	U	1.00	P	12/03/2025	11:13	LB138105
	Sodium	128	+/-250	U	500	P	12/03/2025	11:13	LB138105
	Thallium	0.060	+/-0.5	U	1.00	P	12/03/2025	11:13	LB138105
	Vanadium	0.077	+/-2.5	U	5.00	P	12/03/2025	11:13	LB138105
	Zinc	1.25	+/-2.5	U	5.00	P	12/03/2025	11:13	LB138105

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	4.94	+/-10	J	20.0	P	12/03/2025	11:27	LB138105
	Antimony	0.37	+/-1	J	2.00	P	12/03/2025	11:27	LB138105
	Arsenic	0.089	+/-0.5	U	1.00	P	12/03/2025	11:27	LB138105
	Barium	0.21	+/-5	U	10.0	P	12/03/2025	11:27	LB138105
	Beryllium	0.32	+/-0.5	U	1.00	P	12/03/2025	11:27	LB138105
	Cadmium	0.34	+/-0.5	U	1.00	P	12/03/2025	11:27	LB138105
	Calcium	45.7	+/-250	U	500	P	12/03/2025	11:27	LB138105
	Chromium	0.21	+/-1	U	2.00	P	12/03/2025	11:27	LB138105
	Cobalt	0.070	+/-0.5	U	1.00	P	12/03/2025	11:27	LB138105
	Copper	0.37	+/-1	J	2.00	P	12/03/2025	11:27	LB138105
	Iron	10.5	+/-25	J	50.0	P	12/03/2025	11:27	LB138105
	Lead	0.21	+/-0.5	U	1.00	P	12/03/2025	11:27	LB138105
	Magnesium	19.5	+/-250	U	500	P	12/03/2025	11:27	LB138105
	Manganese	0.43	+/-0.5	U	1.00	P	12/03/2025	11:27	LB138105
	Nickel	0.27	+/-0.5	U	1.00	P	12/03/2025	11:27	LB138105
	Potassium	36.4	+/-250	U	500	P	12/03/2025	11:27	LB138105
	Selenium	2.90	+/-2.5	U	5.00	P	12/03/2025	11:27	LB138105
	Silver	0.060	+/-0.5	J	1.00	P	12/03/2025	11:27	LB138105
	Sodium	128	+/-250	U	500	P	12/03/2025	11:27	LB138105
	Thallium	0.10	+/-0.5	J	1.00	P	12/03/2025	11:27	LB138105
	Vanadium	0.077	+/-2.5	U	5.00	P	12/03/2025	11:27	LB138105
	Zinc	1.25	+/-2.5	U	5.00	P	12/03/2025	11:27	LB138105
CCB02	Aluminum	3.56	+/-10	J	20.0	P	12/03/2025	12:27	LB138105
	Antimony	0.20	+/-1	J	2.00	P	12/03/2025	12:27	LB138105
	Arsenic	0.089	+/-0.5	U	1.00	P	12/03/2025	12:27	LB138105
	Barium	0.21	+/-5	U	10.0	P	12/03/2025	12:27	LB138105
	Beryllium	0.32	+/-0.5	U	1.00	P	12/03/2025	12:27	LB138105
	Cadmium	0.34	+/-0.5	U	1.00	P	12/03/2025	12:27	LB138105
	Calcium	45.7	+/-250	U	500	P	12/03/2025	12:27	LB138105
	Chromium	0.21	+/-1	U	2.00	P	12/03/2025	12:27	LB138105
	Cobalt	0.070	+/-0.5	U	1.00	P	12/03/2025	12:27	LB138105
	Copper	0.41	+/-1	J	2.00	P	12/03/2025	12:27	LB138105
	Iron	12.2	+/-25	J	50.0	P	12/03/2025	12:27	LB138105
	Lead	0.21	+/-0.5	U	1.00	P	12/03/2025	12:27	LB138105
	Magnesium	19.5	+/-250	U	500	P	12/03/2025	12:27	LB138105
	Manganese	0.43	+/-0.5	U	1.00	P	12/03/2025	12:27	LB138105
	Nickel	0.27	+/-0.5	U	1.00	P	12/03/2025	12:27	LB138105
	Potassium	36.4	+/-250	U	500	P	12/03/2025	12:27	LB138105
	Selenium	2.90	+/-2.5	U	5.00	P	12/03/2025	12:27	LB138105

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	0.060	+/-0.5	U	1.00	P	12/03/2025	12:27	LB138105
	Sodium	128	+/-250	U	500	P	12/03/2025	12:27	LB138105
	Thallium	0.070	+/-0.5	J	1.00	P	12/03/2025	12:27	LB138105
	Vanadium	0.077	+/-2.5	U	5.00	P	12/03/2025	12:27	LB138105
	Zinc	1.25	+/-2.5	U	5.00	P	12/03/2025	12:27	LB138105
CCB03	Aluminum	2.78	+/-10	J	20.0	P	12/03/2025	13:05	LB138105
	Antimony	0.18	+/-1	J	2.00	P	12/03/2025	13:05	LB138105
	Arsenic	0.089	+/-0.5	U	1.00	P	12/03/2025	13:05	LB138105
	Barium	0.21	+/-5	U	10.0	P	12/03/2025	13:05	LB138105
	Beryllium	0.32	+/-0.5	U	1.00	P	12/03/2025	13:05	LB138105
	Cadmium	0.34	+/-0.5	U	1.00	P	12/03/2025	13:05	LB138105
	Calcium	45.7	+/-250	U	500	P	12/03/2025	13:05	LB138105
	Chromium	0.21	+/-1	U	2.00	P	12/03/2025	13:05	LB138105
	Cobalt	0.070	+/-0.5	U	1.00	P	12/03/2025	13:05	LB138105
	Copper	0.30	+/-1	U	2.00	P	12/03/2025	13:05	LB138105
	Iron	7.81	+/-25	U	50.0	P	12/03/2025	13:05	LB138105
	Lead	0.21	+/-0.5	U	1.00	P	12/03/2025	13:05	LB138105
	Magnesium	19.5	+/-250	U	500	P	12/03/2025	13:05	LB138105
	Manganese	0.43	+/-0.5	U	1.00	P	12/03/2025	13:05	LB138105
	Nickel	0.27	+/-0.5	U	1.00	P	12/03/2025	13:05	LB138105
	Potassium	36.4	+/-250	U	500	P	12/03/2025	13:05	LB138105
	Selenium	3.01	+/-2.5	J*	5.00	P	12/03/2025	13:05	LB138105
	Silver	0.060	+/-0.5	U	1.00	P	12/03/2025	13:05	LB138105
	Sodium	128	+/-250	U	500	P	12/03/2025	13:05	LB138105
	Thallium	0.060	+/-0.5	U	1.00	P	12/03/2025	13:05	LB138105
	Vanadium	0.077	+/-2.5	U	5.00	P	12/03/2025	13:05	LB138105
	Zinc	1.25	+/-2.5	U	5.00	P	12/03/2025	13:05	LB138105

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3743

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170813BL		WATER		Batch Number:	PB170813		Prep Date:	12/04/2025	
	Mercury	0.076	<0.2	U	0.20	CV	12/04/2025	10:28	LB138109

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3743

Instrument: P7

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170780BL	WATER			Batch Number:	PB170780		Prep Date:	12/02/2025	
	Aluminum	7.07	<10	J	20.0	P	12/03/2025	11:45	LB138105
	Antimony	0.11	<1	U	2.00	P	12/03/2025	11:45	LB138105
	Arsenic	0.089	<0.5	U	1.00	P	12/03/2025	11:45	LB138105
	Barium	0.21	<5	U	10.0	P	12/03/2025	11:45	LB138105
	Beryllium	0.32	<0.5	U	1.00	P	12/03/2025	11:45	LB138105
	Cadmium	0.34	<0.5	U	1.00	P	12/03/2025	11:45	LB138105
	Calcium	45.7	<250	U	500	P	12/03/2025	11:45	LB138105
	Chromium	0.21	<1	U	2.00	P	12/03/2025	11:45	LB138105
	Cobalt	0.070	<0.5	U	1.00	P	12/03/2025	11:45	LB138105
	Copper	0.30	<1	U	2.00	P	12/03/2025	11:45	LB138105
	Iron	7.81	<25	U	50.0	P	12/03/2025	11:45	LB138105
	Lead	0.21	<0.5	U	1.00	P	12/03/2025	11:45	LB138105
	Magnesium	19.5	<250	U	500	P	12/03/2025	11:45	LB138105
	Manganese	0.43	<0.5	U	1.00	P	12/03/2025	11:45	LB138105
	Nickel	0.27	<0.5	U	1.00	P	12/03/2025	11:45	LB138105
	Potassium	36.4	<250	U	500	P	12/03/2025	11:45	LB138105
	Selenium	2.90	<2.5	U	5.00	P	12/03/2025	11:45	LB138105
	Silver	0.060	<0.5	U	1.00	P	12/03/2025	11:45	LB138105
	Sodium	128	<250	U	500	P	12/03/2025	11:45	LB138105
	Thallium	0.060	<0.5	U	1.00	P	12/03/2025	11:45	LB138105
	Vanadium	0.077	<2.5	U	5.00	P	12/03/2025	11:45	LB138105
	Zinc	1.25	<2.5	U	5.00	P	12/03/2025	11:45	LB138105



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV137	Mercury	4.20	4.0	105	90 - 110	CV	12/04/2025	09:45	LB138109

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV59	Mercury	5.14	5.0	103	90 - 110	CV	12/04/2025	09:51	LB138109
CCV60	Mercury	5.16	5.0	103	90 - 110	CV	12/04/2025	10:19	LB138109
CCV61	Mercury	5.09	5.0	102	90 - 110	CV	12/04/2025	10:51	LB138109
CCV62	Mercury	4.99	5.0	100	90 - 110	CV	12/04/2025	11:10	LB138109
CCV63	Mercury	5.39	5.0	108	90 - 110	CV	12/04/2025	12:19	LB138109

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	1490	1600	93	90 - 110	P	12/03/2025	10:59	LB138105
	Antimony	422	400	105	90 - 110	P	12/03/2025	10:59	LB138105
	Arsenic	407	400	102	90 - 110	P	12/03/2025	10:59	LB138105
	Barium	1610	1600	101	90 - 110	P	12/03/2025	10:59	LB138105
	Beryllium	42.1	40.0	105	90 - 110	P	12/03/2025	10:59	LB138105
	Cadmium	209	200	104	90 - 110	P	12/03/2025	10:59	LB138105
	Calcium	4040	4000	101	90 - 110	P	12/03/2025	10:59	LB138105
	Chromium	164	160	102	90 - 110	P	12/03/2025	10:59	LB138105
	Cobalt	413	400	103	90 - 110	P	12/03/2025	10:59	LB138105
	Copper	186	200	93	90 - 110	P	12/03/2025	10:59	LB138105
	Iron	849	800	106	90 - 110	P	12/03/2025	10:59	LB138105
	Lead	391	400	98	90 - 110	P	12/03/2025	10:59	LB138105
	Magnesium	3740	4000	94	90 - 110	P	12/03/2025	10:59	LB138105
	Manganese	393	400	98	90 - 110	P	12/03/2025	10:59	LB138105
	Nickel	408	400	102	90 - 110	P	12/03/2025	10:59	LB138105
	Potassium	4060	4000	102	90 - 110	P	12/03/2025	10:59	LB138105
	Selenium	439	400	110	90 - 110	P	12/03/2025	10:59	LB138105
	Silver	187	200	94	90 - 110	P	12/03/2025	10:59	LB138105
	Sodium	3960	4000	99	90 - 110	P	12/03/2025	10:59	LB138105
	Thallium	381	400	95	90 - 110	P	12/03/2025	10:59	LB138105
	Vanadium	402	400	100	90 - 110	P	12/03/2025	10:59	LB138105
	Zinc	389	400	97	90 - 110	P	12/03/2025	10:59	LB138105

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	21.5	20.0	108	80 - 120	P	12/03/2025	11:10	LB138105
	Antimony	2.07	2.0	104	80 - 120	P	12/03/2025	11:10	LB138105
	Arsenic	1.19	1.0	119	80 - 120	P	12/03/2025	11:10	LB138105
	Barium	9.74	10.0	97	80 - 120	P	12/03/2025	11:10	LB138105
	Beryllium	1.08	1.0	108	80 - 120	P	12/03/2025	11:10	LB138105
	Cadmium	0.80	1.0	80	80 - 120	P	12/03/2025	11:10	LB138105
	Calcium	496	500	99	80 - 120	P	12/03/2025	11:10	LB138105
	Chromium	1.97	2.0	98	80 - 120	P	12/03/2025	11:10	LB138105
	Cobalt	1.08	1.0	108	80 - 120	P	12/03/2025	11:10	LB138105
	Copper	1.97	2.0	98	80 - 120	P	12/03/2025	11:10	LB138105
	Iron	56.0	50.0	112	80 - 120	P	12/03/2025	11:10	LB138105
	Lead	0.88	1.0	88	80 - 120	P	12/03/2025	11:10	LB138105
	Magnesium	479	500	96	80 - 120	P	12/03/2025	11:10	LB138105
	Manganese	1.16	1.0	116	80 - 120	P	12/03/2025	11:10	LB138105
	Nickel	0.93	1.0	93	80 - 120	P	12/03/2025	11:10	LB138105
	Potassium	490	500	98	80 - 120	P	12/03/2025	11:10	LB138105
	Selenium	5.63	5.0	113	80 - 120	P	12/03/2025	11:10	LB138105
	Silver	1.11	1.0	111	80 - 120	P	12/03/2025	11:10	LB138105
	Sodium	484	500	97	80 - 120	P	12/03/2025	11:10	LB138105
	Thallium	1.00	1.0	100	80 - 120	P	12/03/2025	11:10	LB138105
	Vanadium	5.04	5.0	101	80 - 120	P	12/03/2025	11:10	LB138105
	Zinc	5.55	5.0	111	80 - 120	P	12/03/2025	11:10	LB138105

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	50300	50000	101	90 - 110	P	12/03/2025	11:24	LB138105
	Antimony	525	500	105	90 - 110	P	12/03/2025	11:24	LB138105
	Arsenic	513	500	103	90 - 110	P	12/03/2025	11:24	LB138105
	Barium	2660	2500	106	90 - 110	P	12/03/2025	11:24	LB138105
	Beryllium	507	500	101	90 - 110	P	12/03/2025	11:24	LB138105
	Cadmium	509	500	102	90 - 110	P	12/03/2025	11:24	LB138105
	Calcium	254000	250000	101	90 - 110	P	12/03/2025	11:24	LB138105
	Chromium	530	500	106	90 - 110	P	12/03/2025	11:24	LB138105
	Cobalt	524	500	105	90 - 110	P	12/03/2025	11:24	LB138105
	Copper	5090	5000	102	90 - 110	P	12/03/2025	11:24	LB138105
	Iron	129000	125000	104	90 - 110	P	12/03/2025	11:24	LB138105
	Lead	2650	2500	106	90 - 110	P	12/03/2025	11:24	LB138105
	Magnesium	251000	250000	100	90 - 110	P	12/03/2025	11:24	LB138105
	Manganese	5160	5000	103	90 - 110	P	12/03/2025	11:24	LB138105
	Nickel	507	500	101	90 - 110	P	12/03/2025	11:24	LB138105
	Potassium	123000	125000	99	90 - 110	P	12/03/2025	11:24	LB138105
	Selenium	491	500	98	90 - 110	P	12/03/2025	11:24	LB138105
	Silver	511	500	102	90 - 110	P	12/03/2025	11:24	LB138105
	Sodium	256000	250000	102	90 - 110	P	12/03/2025	11:24	LB138105
	Thallium	542	500	108	90 - 110	P	12/03/2025	11:24	LB138105
	Vanadium	522	500	104	90 - 110	P	12/03/2025	11:24	LB138105
	Zinc	5010	5000	100	90 - 110	P	12/03/2025	11:24	LB138105
CCV02	Aluminum	50400	50000	101	90 - 110	P	12/03/2025	12:20	LB138105
	Antimony	522	500	104	90 - 110	P	12/03/2025	12:20	LB138105
	Arsenic	513	500	103	90 - 110	P	12/03/2025	12:20	LB138105
	Barium	2610	2500	104	90 - 110	P	12/03/2025	12:20	LB138105
	Beryllium	497	500	100	90 - 110	P	12/03/2025	12:20	LB138105
	Cadmium	505	500	101	90 - 110	P	12/03/2025	12:20	LB138105
	Calcium	250000	250000	100	90 - 110	P	12/03/2025	12:20	LB138105
	Chromium	529	500	106	90 - 110	P	12/03/2025	12:20	LB138105
	Cobalt	523	500	104	90 - 110	P	12/03/2025	12:20	LB138105
	Copper	5080	5000	102	90 - 110	P	12/03/2025	12:20	LB138105
	Iron	130000	125000	104	90 - 110	P	12/03/2025	12:20	LB138105
	Lead	2630	2500	105	90 - 110	P	12/03/2025	12:20	LB138105

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick
Contract: REMI01
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

SDG No.: Q3743
Lab Code: ACE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	253000	250000	101	90 - 110	P	12/03/2025	12:20	LB138105
	Manganese	5190	5000	104	90 - 110	P	12/03/2025	12:20	LB138105
	Nickel	510	500	102	90 - 110	P	12/03/2025	12:20	LB138105
	Potassium	124000	125000	99	90 - 110	P	12/03/2025	12:20	LB138105
	Selenium	479	500	96	90 - 110	P	12/03/2025	12:20	LB138105
	Silver	512	500	102	90 - 110	P	12/03/2025	12:20	LB138105
	Sodium	258000	250000	103	90 - 110	P	12/03/2025	12:20	LB138105
	Thallium	545	500	109	90 - 110	P	12/03/2025	12:20	LB138105
	Vanadium	521	500	104	90 - 110	P	12/03/2025	12:20	LB138105
	Zinc	4970	5000	99	90 - 110	P	12/03/2025	12:20	LB138105
CCV03	Aluminum	48800	50000	98	90 - 110	P	12/03/2025	13:00	LB138105
	Antimony	517	500	104	90 - 110	P	12/03/2025	13:00	LB138105
	Arsenic	511	500	102	90 - 110	P	12/03/2025	13:00	LB138105
	Barium	2560	2500	103	90 - 110	P	12/03/2025	13:00	LB138105
	Beryllium	511	500	102	90 - 110	P	12/03/2025	13:00	LB138105
	Cadmium	502	500	100	90 - 110	P	12/03/2025	13:00	LB138105
	Calcium	251000	250000	100	90 - 110	P	12/03/2025	13:00	LB138105
	Chromium	527	500	105	90 - 110	P	12/03/2025	13:00	LB138105
	Cobalt	520	500	104	90 - 110	P	12/03/2025	13:00	LB138105
	Copper	5130	5000	103	90 - 110	P	12/03/2025	13:00	LB138105
	Iron	130000	125000	104	90 - 110	P	12/03/2025	13:00	LB138105
	Lead	2570	2500	103	90 - 110	P	12/03/2025	13:00	LB138105
	Magnesium	243000	250000	97	90 - 110	P	12/03/2025	13:00	LB138105
	Manganese	5200	5000	104	90 - 110	P	12/03/2025	13:00	LB138105
	Nickel	508	500	102	90 - 110	P	12/03/2025	13:00	LB138105
	Potassium	123000	125000	99	90 - 110	P	12/03/2025	13:00	LB138105
	Selenium	500	500	100	90 - 110	P	12/03/2025	13:00	LB138105
	Silver	509	500	102	90 - 110	P	12/03/2025	13:00	LB138105
	Sodium	250000	250000	100	90 - 110	P	12/03/2025	13:00	LB138105
	Thallium	530	500	106	90 - 110	P	12/03/2025	13:00	LB138105
	Vanadium	518	500	104	90 - 110	P	12/03/2025	13:00	LB138105
	Zinc	4980	5000	100	90 - 110	P	12/03/2025	13:00	LB138105

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

Initial Calibration Source:

Continuing Calibration Source:

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI	Aluminum	22.5	20.0	112	70 - 130	P	12/03/2025	11:42	LB138105
	Antimony	2.11	2.0	106	70 - 130	P	12/03/2025	11:42	LB138105
	Arsenic	1.16	1.0	116	70 - 130	P	12/03/2025	11:42	LB138105
	Barium	9.58	10.0	96	70 - 130	P	12/03/2025	11:42	LB138105
	Beryllium	1.09	1.0	109	70 - 130	P	12/03/2025	11:42	LB138105
	Cadmium	0.80	1.0	80	70 - 130	P	12/03/2025	11:42	LB138105
	Calcium	513	500	103	70 - 130	P	12/03/2025	11:42	LB138105
	Chromium	2.03	2.0	102	70 - 130	P	12/03/2025	11:42	LB138105
	Cobalt	1.07	1.0	107	50 - 150	P	12/03/2025	11:42	LB138105
	Copper	2.02	2.0	101	70 - 130	P	12/03/2025	11:42	LB138105
	Iron	58.1	50.0	116	70 - 130	P	12/03/2025	11:42	LB138105
	Lead	0.94	1.0	94	70 - 130	P	12/03/2025	11:42	LB138105
	Magnesium	475	500	95	70 - 130	P	12/03/2025	11:42	LB138105
	Manganese	1.19	1.0	119	50 - 150	P	12/03/2025	11:42	LB138105
	Nickel	0.89	1.0	89	70 - 130	P	12/03/2025	11:42	LB138105
	Potassium	505	500	101	70 - 130	P	12/03/2025	11:42	LB138105
	Selenium	6.32	5.0	126	70 - 130	P	12/03/2025	11:42	LB138105
	Silver	1.09	1.0	109	70 - 130	P	12/03/2025	11:42	LB138105
	Sodium	494	500	99	70 - 130	P	12/03/2025	11:42	LB138105
	Thallium	1.06	1.0	106	70 - 130	P	12/03/2025	11:42	LB138105
	Vanadium	4.96	5.0	99	70 - 130	P	12/03/2025	11:42	LB138105
	Zinc	5.57	5.0	111	50 - 150	P	12/03/2025	11:42	LB138105
CRA	Mercury	0.24	0.2	118	70 - 130	CV	12/04/2025	09:56	LB138109

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Remington & Vernick</u>	SDG No.: <u>Q3743</u>
Contract: <u>REMI01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P7</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	90000	100000	90	0	0	12/03/2025	11:18	LB138105
	Antimony	1.13	1.5	75	-2.5	5.5	12/03/2025	11:18	LB138105
	Arsenic	0.85	0.1	850	-1.9	2.1	12/03/2025	11:18	LB138105
	Barium	1.44	1.2	120	-18.8	21.2	12/03/2025	11:18	LB138105
	Beryllium	0.30			-2	2	12/03/2025	11:18	LB138105
	Cadmium	0.070	0.7	10	-1.3	2.7	12/03/2025	11:18	LB138105
	Calcium	97900	100000	98	0	0	12/03/2025	11:18	LB138105
	Chromium	20.1	21.0	96	17	25	12/03/2025	11:18	LB138105
	Cobalt	1.27	1.0	127	-1	3	12/03/2025	11:18	LB138105
	Copper	7.09	8.0	89	4	12	12/03/2025	11:18	LB138105
	Iron	98900	100000	99	0	0	12/03/2025	11:18	LB138105
	Lead	4.40	4.0	110	2	6	12/03/2025	11:18	LB138105
	Magnesium	91100	100000	91	0	0	12/03/2025	11:18	LB138105
	Manganese	7.44	7.0	106	5	9	12/03/2025	11:18	LB138105
	Nickel	5.82	6.0	97	4	8	12/03/2025	11:18	LB138105
	Potassium	97000	100000	97	0	0	12/03/2025	11:18	LB138105
	Selenium	0.28	0.3	93	-9.7	10	12/03/2025	11:18	LB138105
	Silver	0.070			-2	2	12/03/2025	11:18	LB138105
	Sodium	99100	100000	99	0	0	12/03/2025	11:18	LB138105
	Thallium	0.070			-2	2	12/03/2025	11:18	LB138105
	Vanadium	0.23	0.5	46	-9.5	10.5	12/03/2025	11:18	LB138105
	Zinc	10.6	11.0	96	1	21	12/03/2025	11:18	LB138105
ICSAB01	Aluminum	89500	100000	90	0	0	12/03/2025	11:21	LB138105
	Antimony	20.8	22.0	94	18	26	12/03/2025	11:21	LB138105
	Arsenic	20.7	19.0	109	16.2	21.9	12/03/2025	11:21	LB138105
	Barium	21.2	22.0	96	2	42	12/03/2025	11:21	LB138105
	Beryllium	20.7	19.0	109	16.2	21.9	12/03/2025	11:21	LB138105
	Cadmium	20.4	20.0	102	17	23	12/03/2025	11:21	LB138105
	Calcium	95900	100000	96	0	0	12/03/2025	11:21	LB138105
	Chromium	41.0	40.0	102	34	46	12/03/2025	11:21	LB138105
	Cobalt	21.4	20.0	107	17	23	12/03/2025	11:21	LB138105
	Copper	25.5	25.0	102	21	29	12/03/2025	11:21	LB138105
	Iron	97800	100000	98	0	0	12/03/2025	11:21	LB138105
	Lead	23.5	25.0	94	21.3	28.8	12/03/2025	11:21	LB138105
	Magnesium	90200	100000	90	0	0	12/03/2025	11:21	LB138105
	Manganese	26.0	27.0	96	23	31.1	12/03/2025	11:21	LB138105
	Nickel	25.9	24.0	108	20.4	27.6	12/03/2025	11:21	LB138105
	Potassium	96700	100000	97	0	0	12/03/2025	11:21	LB138105
	Selenium	20.4	19.0	107	9	29	12/03/2025	11:21	LB138105
	Silver	19.4	18.0	108	15.3	20.7	12/03/2025	11:21	LB138105
	Sodium	98100	100000	98	0	0	12/03/2025	11:21	LB138105
	Thallium	19.8	21.0	94	17.9	24.2	12/03/2025	11:21	LB138105

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P7

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	20.6	19.0	108	9	29	12/03/2025	11:21	LB138105
	Zinc	29.9	29.0	103	19	39	12/03/2025	11:21	LB138105



METAL QC DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Remington & Vernick</u>	level: <u>low</u>	sdg no.: <u>Q3743</u>
contract: <u>REMI01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3743-01</u>	client id: <u>TW-1125-1MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3743-01MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	17.9	D	11.9	D	4.0	149	N	CV

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Remington & Vernick</u>	level: <u>low</u>	sdg no.: <u>Q3743</u>
contract: <u>REMI01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3743-01</u>	client id: <u>TW-1125-1MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3743-01MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	17.6	D	11.9	D	4.0	142	N	CV

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Remington & Vernick</u>	level: <u>low</u>	sdg no.: <u>Q3743</u>
contract: <u>REMI01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3743-02</u>	client id: <u>TW-1125-2MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3743-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	ug/L	75 - 125	55900	D	48700	D	10000	73		P
Antimony	ug/L	75 - 125	463	D	3.15	JD	500	92		P
Arsenic	ug/L	75 - 125	746	D	310	D	500	87		P
Barium	ug/L	75 - 125	2940	D	893	D	2500	82		P
Beryllium	ug/L	75 - 125	466	D	5.05	D	500	92		P
Cadmium	ug/L	75 - 125	485	D	7.75	D	500	96		P
Calcium	ug/L	75 - 125	168000	D	131000	D	50000	74	N	P
Chromium	ug/L	75 - 125	1800	D	1400	D	500	79		P
Cobalt	ug/L	75 - 125	600	D	145	D	500	91		P
Copper	ug/L	75 - 125	4720	D	462	D	5000	85		P
Iron	ug/L	75 - 125	214000	D	199000	D	25000	61		P
Lead	ug/L	75 - 125	3420	D	1100	D	2500	93		P
Magnesium	ug/L	75 - 125	105000	D	64300	D	50000	82		P
Manganese	ug/L	75 - 125	8780	D	4750	D	5000	81		P
Nickel	ug/L	75 - 125	633	D	180	D	500	91		P
Potassium	ug/L	75 - 125	36400	D	16900	D	25000	78		P
Selenium	ug/L	75 - 125	464	D	25.0	UD	500	91		P
Silver	ug/L	75 - 125	79.2	D	3.05	JD	500	15	N	P
Sodium	ug/L	75 - 125	75300	D	32700	D	50000	85		P
Thallium	ug/L	75 - 125	433	D	2.15	JD	500	86		P
Vanadium	ug/L	75 - 125	556	D	94.5	D	500	92		P
Zinc	ug/L	75 - 125	7690	D	3510	D	5000	83		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Remington & Vernick</u>	level: <u>low</u>	sdg no.: <u>Q3743</u>
contract: <u>REMI01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3743-02</u>	client id: <u>TW-1125-2MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3743-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	ug/L	75 - 125	57300	D	48700	D	10000	87		P
Antimony	ug/L	75 - 125	462	D	3.15	JD	500	92		P
Arsenic	ug/L	75 - 125	763	D	310	D	500	91		P
Barium	ug/L	75 - 125	2910	D	893	D	2500	81		P
Beryllium	ug/L	75 - 125	473	D	5.05	D	500	94		P
Cadmium	ug/L	75 - 125	480	D	7.75	D	500	94		P
Calcium	ug/L	75 - 125	172000	D	131000	D	50000	83		P
Chromium	ug/L	75 - 125	1820	D	1400	D	500	84		P
Cobalt	ug/L	75 - 125	606	D	145	D	500	92		P
Copper	ug/L	75 - 125	4790	D	462	D	5000	86		P
Iron	ug/L	75 - 125	218000	D	199000	D	25000	75		P
Lead	ug/L	75 - 125	3430	D	1100	D	2500	93		P
Magnesium	ug/L	75 - 125	106000	D	64300	D	50000	83		P
Manganese	ug/L	75 - 125	8990	D	4750	D	5000	85		P
Nickel	ug/L	75 - 125	644	D	180	D	500	93		P
Potassium	ug/L	75 - 125	37400	D	16900	D	25000	82		P
Selenium	ug/L	75 - 125	482	D	25.0	UD	500	95		P
Silver	ug/L	75 - 125	78.6	D	3.05	JD	500	15	N	P
Sodium	ug/L	75 - 125	76600	D	32700	D	50000	88		P
Thallium	ug/L	75 - 125	436	D	2.15	JD	500	87		P
Vanadium	ug/L	75 - 125	559	D	94.5	D	500	93		P
Zinc	ug/L	75 - 125	7820	D	3510	D	5000	86		P

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

Matrix: Water

Level: LOW

Client ID: TW-1125-1A

Sample ID: Q3743-01 **Spiked ID:** Q3743-01A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	16.1	D	11.9	D	20.0	21	N	CV

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

Matrix: Water

Level: LOW

Client ID: TW-1125-2A

Sample ID: Q3743-02

Spiked ID: Q3743-02A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Calcium	ug/L	75 - 125	176000	D	131000	D	50000	90		P
Silver	ug/L	75 - 125	82.2	D	3.05	JD	500	16	N	P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Remington & Vernick</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3743</u>
Contract:	<u>REMI01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3743-01</u>	Client ID:	<u>TW-1125-1DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3743-01DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	11.9	D	12.5	D	5		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Remington & Vernick</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3743</u>
Contract:	<u>REMI01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3743-01MS</u>	Client ID:	<u>TW-1125-1MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3743-01MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	17.9	D	17.6	D	2		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Remington & Vernick</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3743</u>
Contract:	<u>REMI01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3743-02</u>	Client ID:	<u>TW-1125-2DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3743-02DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	48700	D	50900	D	4		P
Antimony	ug/L	20	3.15	JD	2.70	JD	15		P
Arsenic	ug/L	20	310	D	323	D	4		P
Barium	ug/L	20	893	D	897	D	0		P
Beryllium	ug/L	20	5.05	D	5.05	D	0		P
Cadmium	ug/L	20	7.75	D	8.30	D	7		P
Calcium	ug/L	20	131000	D	132000	D	1		P
Chromium	ug/L	20	1400	D	1460	D	4		P
Cobalt	ug/L	20	145	D	150	D	3		P
Copper	ug/L	20	462	D	479	D	4		P
Iron	ug/L	20	199000	D	204000	D	2		P
Lead	ug/L	20	1100	D	1130	D	3		P
Magnesium	ug/L	20	64300	D	65900	D	2		P
Manganese	ug/L	20	4750	D	4960	D	4		P
Nickel	ug/L	20	180	D	190	D	5		P
Potassium	ug/L	20	16900	D	17600	D	4		P
Selenium	ug/L	20	25.0	UD	25.0	UD			P
Silver	ug/L	20	3.05	JD	3.05	JD	0		P
Sodium	ug/L	20	32700	D	33400	D	2		P
Thallium	ug/L	20	2.15	JD	2.20	JD	2		P
Vanadium	ug/L	20	94.5	D	97.6	D	3		P
Zinc	ug/L	20	3510	D	3640	D	4		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Remington & Vernick **Level:** LOW **SDG No.:** Q3743
Contract: REMI01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q3743-02MS **Client ID:** TW-1125-2MSD
Percent Solids for Sample: NA **Duplicate ID** Q3743-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	55900	D	57300	D	2		P
Antimony	ug/L	20	463	D	462	D	0		P
Arsenic	ug/L	20	746	D	763	D	2		P
Barium	ug/L	20	2940	D	2910	D	1		P
Beryllium	ug/L	20	466	D	473	D	1		P
Cadmium	ug/L	20	485	D	480	D	1		P
Calcium	ug/L	20	168000	D	172000	D	2		P
Chromium	ug/L	20	1800	D	1820	D	1		P
Cobalt	ug/L	20	600	D	606	D	1		P
Copper	ug/L	20	4720	D	4790	D	1		P
Iron	ug/L	20	214000	D	218000	D	2		P
Lead	ug/L	20	3420	D	3430	D	0		P
Magnesium	ug/L	20	105000	D	106000	D	1		P
Manganese	ug/L	20	8780	D	8990	D	2		P
Nickel	ug/L	20	633	D	644	D	2		P
Potassium	ug/L	20	36400	D	37400	D	3		P
Selenium	ug/L	20	464	D	482	D	4		P
Silver	ug/L	20	79.2	D	78.6	D	1		P
Sodium	ug/L	20	75300	D	76600	D	2		P
Thallium	ug/L	20	433	D	436	D	1		P
Vanadium	ug/L	20	556	D	559	D	1		P
Zinc	ug/L	20	7690	D	7820	D	2		P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client:	<u>Remington & Vernick</u>	SDG No.:	<u>Q3743</u>
Contract:	<u>REMI01</u>	Lab Code:	<u>ACE</u>

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170780BS							
Aluminum	ug/L	10000	9440		94	80 - 120	P
Antimony	ug/L	500	505		101	80 - 120	P
Arsenic	ug/L	500	498		100	80 - 120	P
Barium	ug/L	2500	2480		99	80 - 120	P
Beryllium	ug/L	500	483		97	80 - 120	P
Cadmium	ug/L	500	499		100	80 - 120	P
Calcium	ug/L	50000	48600		97	80 - 120	P
Chromium	ug/L	500	500		100	80 - 120	P
Cobalt	ug/L	500	503		101	80 - 120	P
Copper	ug/L	5000	5020		100	80 - 120	P
Iron	ug/L	25000	26200		105	80 - 120	P
Lead	ug/L	2500	2470		99	80 - 120	P
Magnesium	ug/L	50000	47400		95	80 - 120	P
Manganese	ug/L	5000	5010		100	80 - 120	P
Nickel	ug/L	500	495		99	80 - 120	P
Potassium	ug/L	25000	23800		95	80 - 120	P
Selenium	ug/L	500	501		100	80 - 120	P
Silver	ug/L	500	506		101	80 - 120	P
Sodium	ug/L	50000	48900		98	80 - 120	P
Thallium	ug/L	500	506		101	80 - 120	P
Vanadium	ug/L	500	492		98	80 - 120	P
Zinc	ug/L	5000	5130		103	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170813BS Mercury	ug/L	4.0	4.37		109	80 - 120	CV

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Remington & Vernick

Contract: REMI01

Lab Code: ACE

SDG No.: Q3743

Instrument ID: P7

Start Date : 12/03/2025

Run Number: LB138105

End Date : 12/03/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
S0	S0	1005	100		100		100		100		100	
S2	S2	1008	99		98		98		99		100	
S3	S3	1012	101		99		98		98		100	
S4	S4	1015	95		95		95		95		97	
S5	S5	1018	90		90		91		90		92	
S6	S6	1021	88		85		86		84		88	
S7	S7	1023	81		77		78		77		81	
S8	S8	1026	93		89		89		84		93	
ICV01	ICV01	1059	101		98		97		97		98	
LLICV01	LLICV01	1110	105		100		99		100		100	
ICB01	ICB01	1113	101		97		97		99		98	
ICSA01	ICSA01	1118	101		96		95		92		98	
ICSAB01	ICSAB01	1121	102		99		99		95		101	
CCV01	CCV01	1124	87		86		88		84		90	
CCB01	CCB01	1127	103		101		102		103		103	
CRI	CRI	1142	100		98		99		99		100	
PB170780BL	PB170780BL	1145	101		99		100		100		101	
PB170780BS	PB170780BS	1149	91		87		88		87		90	
Q3743-01	TW-1125-1	1152	101		98		106		96		101	
Q3743-02	TW-1125-2	1155	102		98		101		97		100	
Q3743-02DUP	TW-1125-2DU	1158	102		100		101		98		101	
Q3743-02L	TW-1125-2L	1201	101		99		99		99		99	
Q3743-02MS	TW-1125-2MS	1204	96		97		100		95		97	
Q3743-02MSD	TW-1125-2MS	1207	98		99		103		97		100	
Q3743-02A	TW-1125-2A	1210	100		100		104		100		101	
CCV02	CCV02	1220	91		87		88		84		89	
CCB02	CCB02	1227	101		98		97		99		99	

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Remington & Vernick

Contract: REMI01

Lab Code: ACE

SDG No.: Q3743

Instrument ID: P7

Start Date : 12/03/2025

Run Number: LB138105

End Date : 12/03/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q	Element 165Ho	Q
S0	S0	1005	100		100		100		100		100	
S2	S2	1008	99		100		100		100		101	
S3	S3	1012	97		99		98		98		98	
S4	S4	1015	95		96		95		95		97	
S5	S5	1018	90		91		90		90		93	
S6	S6	1021	85		87		85		87		90	
S7	S7	1023	78		78		76		80		82	
S8	S8	1026	91		92		85		93		95	
ICV01	ICV01	1059	97		99		97		98		99	
LLICV01	LLICV01	1110	97		100		98		99		100	
ICB01	ICB01	1113	95		99		99		98		100	
ICSA01	ICSA01	1118	94		97		94		97		99	
ICSAB01	ICSAB01	1121	97		99		95		98		101	
CCV01	CCV01	1124	86		88		86		89		90	
CCB01	CCB01	1127	98		101		102		100		104	
CRI	CRI	1142	96		100		100		99		101	
PB170780BL	PB170780BL	1145	96		101		99		98		102	
PB170780BS	PB170780BS	1149	85		86		86		89		91	
Q3743-01	TW-1125-1	1152	97		107		95		98		100	
Q3743-02	TW-1125-2	1155	97		102		97		98		101	
Q3743-02DUP	TW-1125-2DU	1158	95		100		95		96		100	
Q3743-02L	TW-1125-2L	1201	97		101		100		100		102	
Q3743-02MS	TW-1125-2MS	1204	95		102		96		98		101	
Q3743-02MSD	TW-1125-2MS	1207	97		103		98		100		103	
Q3743-02A	TW-1125-2A	1210	98		105		99		99		103	
CCV02	CCV02	1220	85		86		83		88		92	
CCB02	CCB02	1227	95		99		98		98		101	

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Remington & Vernick

Contract: REMI01

Lab Code: ACE

SDG No.: Q3743

Instrument ID: P7

Start Date : 12/03/2025

Run Number: LB138105

End Date : 12/03/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1005	100		100							
S2	S2	1008	99		98							
S3	S3	1012	99		99							
S4	S4	1015	97		98							
S5	S5	1018	92		92							
S6	S6	1021	88		87							
S7	S7	1023	80		81							
S8	S8	1026	94		86							
ICV01	ICV01	1059	100		101							
LLICV01	LLICV01	1110	102		103							
ICB01	ICB01	1113	99		100							
ICSA01	ICSA01	1118	100		98							
ICSAB01	ICSAB01	1121	103		101							
CCV01	CCV01	1124	92		88							
CCB01	CCB01	1127	104		107							
CRI	CRI	1142	103		102							
PB170780BL	PB170780BL	1145	101		104							
PB170780BS	PB170780BS	1149	92		92							
Q3743-01	TW-1125-1	1152	101		101							
Q3743-02	TW-1125-2	1155	102		103							
Q3743-02DUP	TW-1125-2DUP	1158	104		101							
Q3743-02L	TW-1125-2L	1201	102		102							
Q3743-02MS	TW-1125-2MS	1204	101		99							
Q3743-02MSD	TW-1125-2MSD	1207	103		102							
Q3743-02A	TW-1125-2A	1210	101		102							
CCV02	CCV02	1220	91		88							
CCB02	CCB02	1227	100		100							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Remington & Vernick

Contract: REMI01

Lab Code: ACE

SDG No.: Q3743

Instrument ID: P7

Start Date : 12/03/2025

Run Number: LB138105

End Date : 12/03/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1005	100									
S2	S2	1008	99									
S3	S3	1012	98									
S4	S4	1015	94									
S5	S5	1018	92									
S6	S6	1021	87									
S7	S7	1023	80									
S8	S8	1026	87									
ICV01	ICV01	1059	100									
LLICV01	LLICV01	1110	102									
ICB01	ICB01	1113	99									
ICSA01	ICSA01	1118	97									
ICSAB01	ICSAB01	1121	99									
CCV01	CCV01	1124	86									
CCB01	CCB01	1127	102									
CRI	CRI	1142	102									
PB170780BL	PB170780BL	1145	104									
PB170780BS	PB170780BS	1149	89									
Q3743-01	TW-1125-1	1152	100									
Q3743-02	TW-1125-2	1155	101									
Q3743-02DUP	TW-1125-2DUP	1158	97									
Q3743-02L	TW-1125-2L	1201	103									
Q3743-02MS	TW-1125-2MS	1204	99									
Q3743-02MSD	TW-1125-2MSD	1207	101									
Q3743-02A	TW-1125-2A	1210	100									
CCV02	CCV02	1220	87									
CCB02	CCB02	1227	101									

Metals

-9 -

ICP SERIAL DILUTIONS

SAMPLE NO.

TW-1125-1L

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE **Lb No.:** lb138109 **Lab Sample ID :** Q3743-01L **SDG No.:** Q3743

Matrix (soil/water): Water **Level (low/med):** LOW

Concentration Units: ug/L

Analyte	Initial Sample Result (I) C	Serial Dilution Result (S) C	% Differ- ence	Q	M
Mercury	13.3 OR	18.7	41		CV

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

TW-1125-2L

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **Lb No.:** lb138105 **Lab Sample ID :** Q3743-02L **SDG No.:** Q3743
Matrix (soil/water): Water **Level (low/med):** LOW
Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Differ-ence	Q	M
Aluminum	48700	D	48300	D	1		P
Antimony	3.15	JD	3.50	JD	11		P
Arsenic	310	D	323	D	4		P
Barium	893	D	909	D	2		P
Beryllium	5.05	D	25.0	UD	6		P
Cadmium	7.75	D	25.0	UD	26		P
Calcium	131000	D	135000	D	3		P
Chromium	1400	D	1410	D	1		P
Cobalt	145	D	150	D	3		P
Copper	462	D	473	D	2		P
Iron	199000	D	205000	D	3		P
Lead	1100	D	1100	D	0		P
Magnesium	64300	D	63700	D	1		P
Manganese	4750	D	4750	D	0		P
Nickel	180	D	184	D	2		P
Potassium	16900	D	17200	D	2		P
Selenium	25.0	UD	125	UD			P
Silver	3.05	JD	3.25	JD	7		P
Sodium	32700	D	32100	D	2		P
Thallium	2.15	JD	2.75	JD	28		P
Vanadium	94.5	D	94.8	JD	0		P
Zinc	3510	D	3610	D	3		P

metals
- 14 -
ANALYSIS RUN LOG

Client: Remington & Vernick

Contract: REMI01

Lab code: ACE

Sdg no.: Q3743

Instrument id number: _____

Method: _____

Run number: LB138105

Start date: 12/03/2025

End date: 12/03/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1005	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S2	S2	1	1008	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S3	S3	1	1012	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S4	S4	1	1015	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S5	S5	1	1018	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S6	S6	1	1021	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S7	S7	1	1023	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S8	S8	1	1026	Al,Ca,Fe,K,Mg,Na
ICV01	ICV01	1	1059	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
LLICV01	LLICV01	1	1110	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICB01	ICB01	1	1113	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSA01	ICSA01	1	1118	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSAB01	ICSAB01	1	1121	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV01	CCV01	1	1124	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB01	CCB01	1	1127	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CRI	CRI	1	1142	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
PB170780BL	PB170780BL	1	1145	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
PB170780BS	PB170780BS	1	1149	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3743-01	TW-1125-1	5	1152	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3743-02	TW-1125-2	5	1155	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3743-02DUP	TW-1125-2DUP	5	1158	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3743-02L	TW-1125-2L	25	1201	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3743-02MS	TW-1125-2MS	5	1204	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3743-02MSD	TW-1125-2MSD	5	1207	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3743-02A	TW-1125-2A	5	1210	Ag,Ca
CCV02	CCV02	1	1220	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB02	CCB02	1	1227	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV03	CCV03	1	1300	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB03	CCB03	1	1305	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn

metals
- 14 -
ANALYSIS RUN LOG

Client: Remington & Vernick

Contract: REMI01

Lab code: ACE

Sdg no.: Q3743

Instrument id number: _____

Method: _____

Run number: LB138109

Start date: 12/04/2025

End date: 12/04/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	0924	HG
S0.2	S0.2	1	0926	HG
S2.5	S2.5	1	0928	HG
S5	S5	1	0931	HG
S7.5	S7.5	1	0933	HG
S10	S10	1	0941	HG
ICV137	ICV137	1	0945	HG
ICB137	ICB137	1	0947	HG
CCV59	CCV59	1	0951	HG
CCB59	CCB59	1	0954	HG
CRA	CRA	1	0956	HG
CCV60	CCV60	1	1019	HG
CCB60	CCB60	1	1021	HG
PB170813BL	PB170813BL	1	1028	HG
Q3743-02	TW-1125-2	1	1047	HG
CCV61	CCV61	1	1051	HG
CCB61	CCB61	1	1054	HG
Q3743-01L	TW-1125-1L	5	1105	HG
CCV62	CCV62	1	1110	HG
CCB62	CCB62	1	1112	HG
Q3743-01	TW-1125-1	5	1114	HG
Q3743-01DUP	TW-1125-1DUP	5	1116	HG
Q3743-01MS	TW-1125-1MS	5	1119	HG
Q3743-01MSD	TW-1125-1MSD	5	1121	HG
Q3743-01A	TW-1125-1A	5	1126	HG
PB170813BS	PB170813BS	1	1216	HG
CCV63	CCV63	1	1219	HG
CCB63	CCB63	1	1221	HG



METAL PREPARATION & INSTRUMENT DATA

LAB CHRONICLE

OrderID:	Q3743	OrderDate:	12/1/2025 9:45:00 AM
Client:	Remington & Vernick	Project:	Saddler Property
Contact:	Justin Zarzecki	Location:	A12,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3743-01	TW-1125-1	Water			11/26/25			11/26/25
			Mercury	7470A		12/04/25	12/04/25	
			Metals ICP-TAL	6020B		12/02/25	12/03/25	
Q3743-02	TW-1125-2	Water			11/26/25			11/26/25
			Mercury	7470A		12/04/25	12/04/25	
			Metals ICP-TAL	6020B		12/02/25	12/03/25	



METAL PREPARATION & ANALYICAL SUMMARY

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

Method: _____

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB170780							
PB170780BL	PB170780BL	MB	WATER	12/02/2025	50.0	50.0	
PB170780BS	PB170780BS	LCS	WATER	12/02/2025	50.0	50.0	
Q3743-01	TW-1125-1	SAM	WATER	12/02/2025	50.0	50.0	
Q3743-02	TW-1125-2	SAM	WATER	12/02/2025	50.0	50.0	
Q3743-02DUP	TW-1125-2DUP	DUP	WATER	12/02/2025	50.0	50.0	
Q3743-02MS	TW-1125-2MS	MS	WATER	12/02/2025	50.0	50.0	
Q3743-02MSD	TW-1125-2MSD	MSD	WATER	12/02/2025	50.0	50.0	

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

Method: _____

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB170813							
PB170813BL	PB170813BL	MB	WATER	12/04/2025	30.0	30.0	
PB170813BS	PB170813BS	LCS	WATER	12/04/2025	30.0	30.0	
Q3743-01	TW-1125-1	SAM	WATER	12/04/2025	30.0	30.0	
Q3743-01DUP	TW-1125-1DUP	DUP	WATER	12/04/2025	30.0	30.0	
Q3743-01MS	TW-1125-1MS	MS	WATER	12/04/2025	30.0	30.0	
Q3743-01MSD	TW-1125-1MSD	MSD	WATER	12/04/2025	30.0	30.0	
Q3743-02	TW-1125-2	SAM	WATER	12/04/2025	30.0	30.0	

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB138105

Review By	Janvi	Review On	12/4/2025 10:27:33 AM
Supervise By	jaswal	Supervise On	12/5/2025 11:29:34 AM
STD. NAME	STD REF.#		
ICAL Standard	MP88152,MP88162,MP88160,MP88158,MP88157,MP88156,MP88155,MP88154,MP88153		
ICV Standard	MP88163,MP88162		
CCV Standard	MP88164		
ICSA Standard	MP88165,MP88166		
CRI Standard	MP88162		
LCS Standard			
Chk Standard	MP88168,MP88169		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE	TUNE	12/03/25 09:17		Jaswal	OK
2	S0	S0	CAL1	12/03/25 10:05		Jaswal	OK
3	S2	S2	CAL3	12/03/25 10:08		Jaswal	OK
4	S3	S3	CAL4	12/03/25 10:12		Jaswal	OK
5	S4	S4	CAL5	12/03/25 10:15		Jaswal	OK
6	S5	S5	CAL6	12/03/25 10:18		Jaswal	OK
7	S6	S6	CAL7	12/03/25 10:21		Jaswal	OK
8	S7	S7	CAL8	12/03/25 10:23		Jaswal	OK
9	S8	S8	CAL9	12/03/25 10:26		Jaswal	OK
10	ICV01	ICV01	ICV	12/03/25 10:59		Jaswal	OK
11	LLICV01	LLICV01	LLICV	12/03/25 11:10		Jaswal	OK
12	ICB01	ICB01	ICB	12/03/25 11:13		Jaswal	OK
13	ICSA01	ICSA01	ICSA	12/03/25 11:18		Jaswal	OK
14	ICSAB01	ICSAB01	ICSAB	12/03/25 11:21		Jaswal	OK
15	CCV01	CCV01	CCV	12/03/25 11:24		Jaswal	OK
16	CCB01	CCB01	CCB	12/03/25 11:27		Jaswal	OK
17	CRI	CRI	CRDL	12/03/25 11:42		Jaswal	OK
18	PB170780BL	PB170780BL	MB	12/03/25 11:45		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB138105

Review By	Janvi	Review On	12/4/2025 10:27:33 AM
Supervise By	jaswal	Supervise On	12/5/2025 11:29:34 AM
STD. NAME	STD REF.#		
ICAL Standard	MP88152,MP88162,MP88160,MP88158,MP88157,MP88156,MP88155,MP88154,MP88153		
ICV Standard	MP88163,MP88162		
CCV Standard	MP88164		
ICSA Standard	MP88165,MP88166		
CRI Standard	MP88162		
LCS Standard			
Chk Standard	MP88168,MP88169		

19	PB170780BS	PB170780BS	LCS	12/03/25 11:49		Jaswal	OK
20	Q3743-01DL	TW-1125-1DL	SAM	12/03/25 11:52		Jaswal	OK
21	Q3743-02DL	TW-1125-2DL	SAM	12/03/25 11:55		Jaswal	OK
22	Q3743-02DUPDL	TW-1125-2DUPDL	DUP	12/03/25 11:58		Jaswal	OK
23	Q3743-02LDL	TW-1125-2LDL	SD	12/03/25 12:01		Jaswal	OK
24	Q3743-02MSDL	TW-1125-2MSDL	MS	12/03/25 12:04		Jaswal	OK
25	Q3743-02MSDDL	TW-1125-2MSDDL	MSD	12/03/25 12:07		Jaswal	OK
26	Q3743-02ADL	TW-1125-2ADL	PS	12/03/25 12:10	0.1 ML OF MP88170 AND 0.2ML OF EACH MP88171,MP88172,MP88 173 IN 10ML SAMPLE BEFORE DILUTION.	Jaswal	OK
27	CCV02	CCV02	CCV	12/03/25 12:20		Jaswal	OK
28	CCB02	CCB02	CCB	12/03/25 12:27		Jaswal	OK
29	Q3749-01DL	AUD-25-0204DL	SAM	12/03/25 12:51	K High , INT_STD 89Y(2) Fail	Jaswal	Dilution
30	Q3749-01DL2	AUD-25-0204DL2	SAM	12/03/25 12:57	25X For K , 25X For INT_STD 89Y(2)	Jaswal	Confirms
31	CCV03	CCV03	CCV	12/03/25 13:00		Jaswal	OK
32	CCB03	CCB03	CCB	12/03/25 13:05		Jaswal	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB138109

Review By	mohan	Review On	12/5/2025 1:12:25 PM
Supervise By	jaswal	Supervise On	12/5/2025 8:54:44 PM
STD. NAME	STD REF.#		
ICAL Standard	MP88196,MP88197,MP88198,MP88199,MP88200,MP88201		
ICV Standard	MP88202		
CCV Standard	MP88204		
ICSA Standard			
CRI Standard	MP88206		
LCS Standard			
Chk Standard	MP88203,MP88205,MP88207,MP88209		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	12/04/25 09:24		mohan	OK
2	S0.2	S0.2	CAL2	12/04/25 09:26		mohan	OK
3	S2.5	S2.5	CAL3	12/04/25 09:28		mohan	OK
4	S5	S5	CAL4	12/04/25 09:31		mohan	OK
5	S7.5	S7.5	CAL5	12/04/25 09:33		mohan	OK
6	S10	S10	CAL6	12/04/25 09:41		mohan	OK
7	ICV137	ICV137	ICV	12/04/25 09:45		mohan	OK
8	ICB137	ICB137	ICB	12/04/25 09:47		mohan	OK
9	CCV59	CCV59	CCV	12/04/25 09:51		mohan	OK
10	CCB59	CCB59	CCB	12/04/25 09:54		mohan	OK
11	CRA	CRA	CRDL	12/04/25 09:56		mohan	OK
12	HighStd	HighStd	HIGH STD	12/04/25 09:58		mohan	OK
13	ChkStd	ChkStd	SAM	12/04/25 10:00		mohan	OK
14	PB170812BL	PB170812BL	MB	12/04/25 10:03		mohan	OK
15	PB170812BS	PB170812BS	LCS	12/04/25 10:05		mohan	OK
16	Q3750-07	FENCE WC-1	SAM	12/04/25 10:07		mohan	OK
17	Q3750-08	FENCE WC-2	SAM	12/04/25 10:10		mohan	OK
18	Q3753-03	MOO-25-334-337	SAM	12/04/25 10:12		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB138109

Review By	mohan	Review On	12/5/2025 1:12:25 PM
Supervise By	jaswal	Supervise On	12/5/2025 8:54:44 PM
STD. NAME	STD REF.#		
ICAL Standard	MP88196,MP88197,MP88198,MP88199,MP88200,MP88201		
ICV Standard	MP88202		
CCV Standard	MP88204		
ICSA Standard			
CRI Standard	MP88206		
LCS Standard			
Chk Standard	MP88203,MP88205,MP88207,MP88209		

19	Q3753-05	AR520-0002	SAM	12/04/25 10:14		mohan	OK
20	Q3753-05DUP	AR520-0002DUP	DUP	12/04/25 10:16		mohan	OK
21	CCV60	CCV60	CCV	12/04/25 10:19		mohan	OK
22	CCB60	CCB60	CCB	12/04/25 10:21		mohan	OK
23	Q3753-05MS	AR520-0002MS	MS	12/04/25 10:23		mohan	OK
24	Q3753-05MSD	AR520-0002MSD	MSD	12/04/25 10:25		mohan	OK
25	PB170813BL	PB170813BL	MB	12/04/25 10:28		mohan	OK
26	Q3743-01	TW-1125-1	SAM	12/04/25 10:37	Hg is high, Need dilution	mohan	Dilution
27	Q3743-01DUP	TW-1125-1DUP	DUP	12/04/25 10:40	Hg is high, Need dilution	mohan	Dilution
28	Q3743-01MS	TW-1125-1MS	MS	12/04/25 10:42	Hg is high, Need dilution	mohan	Dilution
29	Q3743-01MSD	TW-1125-1MSD	MSD	12/04/25 10:44	Hg is high, Need dilution	mohan	Dilution
30	Q3743-02	TW-1125-2	SAM	12/04/25 10:47		mohan	OK
31	Q3749-01	AUD-25-0204	SAM	12/04/25 10:49		mohan	OK
32	CCV61	CCV61	CCV	12/04/25 10:51		mohan	OK
33	CCB61	CCB61	CCB	12/04/25 10:54		mohan	OK
34	Q3771-02	Comp	SAM	12/04/25 10:56		mohan	OK
35	PB170785TB	PB170785TB	MB	12/04/25 10:58		mohan	OK
36	Q3753-05L	AR520-0002L	SD	12/04/25 11:00		mohan	OK
37	Q3753-05A	AR520-0002A	PS	12/04/25 11:03		mohan	OK
38	Q3743-01L	TW-1125-1L	SD	12/04/25 11:05		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB138109

Review By	mohan	Review On	12/5/2025 1:12:25 PM
Supervise By	jaswal	Supervise On	12/5/2025 8:54:44 PM

STD. NAME	STD REF.#
ICAL Standard	MP88196,MP88197,MP88198,MP88199,MP88200,MP88201
ICV Standard	MP88202
CCV Standard	MP88204
ICSA Standard	
CRI Standard	MP88206
LCS Standard	
Chk Standard	MP88203,MP88205,MP88207,MP88209

39	Q3743-01A	TW-1125-1A	PS	12/04/25 11:07	Hg is high, Need dilution	mohan	Dilution
40	CCV62	CCV62	CCV	12/04/25 11:10		mohan	OK
41	CCB62	CCB62	CCB	12/04/25 11:12		mohan	OK
42	Q3743-01DL	TW-1125-1DL	SAM	12/04/25 11:14	5x For Hg	mohan	Confirms
43	Q3743-01DUPDL	TW-1125-1DUPDL	DUP	12/04/25 11:16	5x For Hg	mohan	Confirms
44	Q3743-01MSDL	TW-1125-1MSDL	MS	12/04/25 11:19	5x For Hg	mohan	Confirms
45	Q3743-01MSDDL	TW-1125-1MSDDL	MSD	12/04/25 11:21	5x For Hg	mohan	Confirms
46	Q3743-01LDL	TW-1125-1LDL	SD	12/04/25 11:23	Not Required	mohan	Not Ok
47	Q3743-01ADL	TW-1125-1ADL	PS	12/04/25 11:26	5x For Hg	mohan	Confirms
48	PB170813BS	PB170813BS	LCS	12/04/25 12:16		mohan	OK
49	CCV63	CCV63	CCV	12/04/25 12:19		mohan	OK
50	CCB63	CCB63	CCB	12/04/25 12:21		mohan	OK

SOP ID : M3010A-Digestion-17

SDG No : N/A

Matrix : WATER

Pipette ID: ICP A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M6069

Hood ID : #3

Block ID: 1. HOT BLOCK #1 2. N/A

Start Digest Date: 12/02/2025 Time : 09:30 Temp : 96 °C

End Digest Date: 12/02/2025 Time : 11:35 Temp : 96 °C

Digestion tube ID: M5595

Block thermometer ID: MET-DIG. #1

Dig Technician Signature: SKG

Supervisor Signature: [Signature]

Temp : 1. 96°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
Spike Sol 1	0.50	MP88170
Spike Sol 2	1.00	MP88171
Spike Sol 3	1.00	MP88172
Spike Sol 4	1.00	MP88173
Spike Sol 5	1.00mL	MP88174

Chemical Used	ML/SAMPLE USED	Lot Number
Conc. HNO3	3.00	M6187
1:1 HCL	5.00	MP87148
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

HOT BLOCK#1 CELL#50 96C
X*, MP88172,MP88173,MP88174

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
12/02/25 12:35	SKGmetdis	SD (metals lab)
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
PB170780BL	PBW780	<2	50	50	Colorless	Colorless	Clear	Clear	N/A	1
PB170780BS	LCS780	<2	50	50	Colorless	Colorless	Clear	Clear	MP88170,MP88171,X*	2
Q3743-01	TW-1125-1	<2	50	50	Brown	ligh	Cloudy	Clear	N/A	3
Q3743-02	TW-1125-2	<2	50	50	Brown	light	Cloudy	Clear	N/A	4
Q3743-02MS	TW-1125-2MS	<2	50	50	Brown	ligh	Cloudy	Clear	MP88170,MP88171,X*	6
Q3743-02MSD	TW-1125-2MSD	<2	50	50	Brown	ligh	Cloudy	Clear	MP88170,MP88171,X*	7
Q3743-02DUP	TW-1125-2DUP	<2	50	50	Brown	ligh	Cloudy	Clear	N/A	5
Q3749-01	AUD-25-0204	<2	50	50	Pink	Light Pink	Clear	Clear	N/A	8

SOP ID : M7470A-Mercury-20

SDG No : NA

Matrix : WATER

Pipette ID: HG A

Balance ID : N/A

Filter paper ID : NA

pH Strip ID : M6069

Hood ID : #1

Block ID: 1. HG HOT BLOCK#3 2. N/A

Start Digest Date: 12/04/2025 Time : 06:30 Temp : 93 °C

End Digest Date: 12/04/2025 Time : 08:30 Temp : 94 °C

Digestion tube ID: M5595

Block thermometer ID: HG-DIG#3

Dig Technician Signature: *MB*

Supervisor Signature: *12*

Temp : 1. 93°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
ICV	30mL	MP88202
CCV	30mL	MP88204
CRA	30mL	MP88206
Blank Spike	0.48mL	MP88195
Matrix Spike	0.48mL	MP88195

Chemical Used	ML/SAMPLE USED	Lot Number
HNO3/H2SO4(1:2)	2.25mL	MP87633
KMnO4 (5%)	4.5mL	MP87634
K2S2O8 (5%)	2.4mL	MP87635
Hydroxylamine HCL (12%)	1.8mL	MP87636
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

LAB SAMPLE ID	CLIENT SAMPLE ID	Wt(g)/Vol(ml)	Comment
0.0 ppb	S0	30mL	MP88196
0.05 ppb	S0.05	N/A	N/A
0.2 ppb	S0.2	30mL	MP88197
2.5 ppb	S2.5	30mL	MP88198
5.0 ppb	S5.0	30mL	MP88199
7.5 ppb	S7.5	30mL	MP88200
10.0 ppb	S10.0	30mL	MP88201
ICV	ICV	30mL	MP88202
ICB	ICB	30mL	MP88203
CCV	CCV	30mL	MP88204
CCB	CCB	30mL	MP88205
CRI	CRI	30mL	MP88206
CHK STD	CHK STD	30mL	MP88207

Extraction Conformance/Non-Conformance Comments:

N/A		
Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
12/4/25 @ 8:55	<i>MB - N/A Lab</i>	<i>MB - Metal Lab</i>
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Vol (ml)	Final Vol (ml)	pH	Comment	Prep Pos
PB170813BL	PBW813	30	30	<2	N/A	3-19
PB170813BS	LCS813	30	30	<2	MP88195	20
Q3743-01DUP	TW-1125-1DUP	30	30	<2	N/A	22
Q3743-01MS	TW-1125-1MS	30	30	<2	MP88195	23
Q3743-01MSD	TW-1125-1MSD	30	30	<2	MP88195	24
Q3743-01	TW-1125-1	30	30	<2	N/A	21
Q3743-02	TW-1125-2	30	30	<2	N/A	25
Q3749-01	AUD-25-0204	30	30	<2	N/A	26
Q3771-02	COMP	30	30	<2	N/A	27