



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB137	Mercury	0.076	+/-0.2	U	0.20	CV	12/04/2025	09:47	LB138109

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB59	Mercury	0.076	+/-0.2	U	0.20	CV	12/04/2025	09:54	LB138109
CCB60	Mercury	0.076	+/-0.2	U	0.20	CV	12/04/2025	10:21	LB138109
CCB61	Mercury	0.076	+/-0.2	U	0.20	CV	12/04/2025	10:54	LB138109
CCB62	Mercury	0.076	+/-0.2	U	0.20	CV	12/04/2025	11:12	LB138109
CCB63	Mercury	0.076	+/-0.2	U	0.20	CV	12/04/2025	12:21	LB138109

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	4.93	+/-10	J	20.0	P	12/03/2025	11:13	LB138105
	Antimony	0.11	+/-1	U	2.00	P	12/03/2025	11:13	LB138105
	Arsenic	0.089	+/-0.5	U	1.00	P	12/03/2025	11:13	LB138105
	Barium	0.21	+/-5	U	10.0	P	12/03/2025	11:13	LB138105
	Beryllium	0.32	+/-0.5	U	1.00	P	12/03/2025	11:13	LB138105
	Cadmium	0.34	+/-0.5	U	1.00	P	12/03/2025	11:13	LB138105
	Calcium	45.7	+/-250	U	500	P	12/03/2025	11:13	LB138105
	Chromium	0.21	+/-1	U	2.00	P	12/03/2025	11:13	LB138105
	Cobalt	0.070	+/-0.5	U	1.00	P	12/03/2025	11:13	LB138105
	Copper	0.30	+/-1	U	2.00	P	12/03/2025	11:13	LB138105
	Iron	7.81	+/-25	U	50.0	P	12/03/2025	11:13	LB138105
	Lead	0.21	+/-0.5	U	1.00	P	12/03/2025	11:13	LB138105
	Magnesium	19.5	+/-250	U	500	P	12/03/2025	11:13	LB138105
	Manganese	0.43	+/-0.5	U	1.00	P	12/03/2025	11:13	LB138105
	Nickel	0.27	+/-0.5	U	1.00	P	12/03/2025	11:13	LB138105
	Potassium	36.4	+/-250	U	500	P	12/03/2025	11:13	LB138105
	Selenium	2.90	+/-2.5	U	5.00	P	12/03/2025	11:13	LB138105
	Silver	0.060	+/-0.5	U	1.00	P	12/03/2025	11:13	LB138105
	Sodium	128	+/-250	U	500	P	12/03/2025	11:13	LB138105
	Thallium	0.060	+/-0.5	U	1.00	P	12/03/2025	11:13	LB138105
	Vanadium	0.077	+/-2.5	U	5.00	P	12/03/2025	11:13	LB138105
	Zinc	1.25	+/-2.5	U	5.00	P	12/03/2025	11:13	LB138105

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	4.94	+/-10	J	20.0	P	12/03/2025	11:27	LB138105
	Antimony	0.37	+/-1	J	2.00	P	12/03/2025	11:27	LB138105
	Arsenic	0.089	+/-0.5	U	1.00	P	12/03/2025	11:27	LB138105
	Barium	0.21	+/-5	U	10.0	P	12/03/2025	11:27	LB138105
	Beryllium	0.32	+/-0.5	U	1.00	P	12/03/2025	11:27	LB138105
	Cadmium	0.34	+/-0.5	U	1.00	P	12/03/2025	11:27	LB138105
	Calcium	45.7	+/-250	U	500	P	12/03/2025	11:27	LB138105
	Chromium	0.21	+/-1	U	2.00	P	12/03/2025	11:27	LB138105
	Cobalt	0.070	+/-0.5	U	1.00	P	12/03/2025	11:27	LB138105
	Copper	0.37	+/-1	J	2.00	P	12/03/2025	11:27	LB138105
	Iron	10.5	+/-25	J	50.0	P	12/03/2025	11:27	LB138105
	Lead	0.21	+/-0.5	U	1.00	P	12/03/2025	11:27	LB138105
	Magnesium	19.5	+/-250	U	500	P	12/03/2025	11:27	LB138105
	Manganese	0.43	+/-0.5	U	1.00	P	12/03/2025	11:27	LB138105
	Nickel	0.27	+/-0.5	U	1.00	P	12/03/2025	11:27	LB138105
	Potassium	36.4	+/-250	U	500	P	12/03/2025	11:27	LB138105
	Selenium	2.90	+/-2.5	U	5.00	P	12/03/2025	11:27	LB138105
	Silver	0.060	+/-0.5	J	1.00	P	12/03/2025	11:27	LB138105
	Sodium	128	+/-250	U	500	P	12/03/2025	11:27	LB138105
	Thallium	0.10	+/-0.5	J	1.00	P	12/03/2025	11:27	LB138105
	Vanadium	0.077	+/-2.5	U	5.00	P	12/03/2025	11:27	LB138105
	Zinc	1.25	+/-2.5	U	5.00	P	12/03/2025	11:27	LB138105
CCB02	Aluminum	3.56	+/-10	J	20.0	P	12/03/2025	12:27	LB138105
	Antimony	0.20	+/-1	J	2.00	P	12/03/2025	12:27	LB138105
	Arsenic	0.089	+/-0.5	U	1.00	P	12/03/2025	12:27	LB138105
	Barium	0.21	+/-5	U	10.0	P	12/03/2025	12:27	LB138105
	Beryllium	0.32	+/-0.5	U	1.00	P	12/03/2025	12:27	LB138105
	Cadmium	0.34	+/-0.5	U	1.00	P	12/03/2025	12:27	LB138105
	Calcium	45.7	+/-250	U	500	P	12/03/2025	12:27	LB138105
	Chromium	0.21	+/-1	U	2.00	P	12/03/2025	12:27	LB138105
	Cobalt	0.070	+/-0.5	U	1.00	P	12/03/2025	12:27	LB138105
	Copper	0.41	+/-1	J	2.00	P	12/03/2025	12:27	LB138105
	Iron	12.2	+/-25	J	50.0	P	12/03/2025	12:27	LB138105
	Lead	0.21	+/-0.5	U	1.00	P	12/03/2025	12:27	LB138105
	Magnesium	19.5	+/-250	U	500	P	12/03/2025	12:27	LB138105
	Manganese	0.43	+/-0.5	U	1.00	P	12/03/2025	12:27	LB138105
	Nickel	0.27	+/-0.5	U	1.00	P	12/03/2025	12:27	LB138105
	Potassium	36.4	+/-250	U	500	P	12/03/2025	12:27	LB138105
	Selenium	2.90	+/-2.5	U	5.00	P	12/03/2025	12:27	LB138105

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3743

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	0.060	+/-0.5	U	1.00	P	12/03/2025	12:27	LB138105
	Sodium	128	+/-250	U	500	P	12/03/2025	12:27	LB138105
	Thallium	0.070	+/-0.5	J	1.00	P	12/03/2025	12:27	LB138105
	Vanadium	0.077	+/-2.5	U	5.00	P	12/03/2025	12:27	LB138105
	Zinc	1.25	+/-2.5	U	5.00	P	12/03/2025	12:27	LB138105
CCB03	Aluminum	2.78	+/-10	J	20.0	P	12/03/2025	13:05	LB138105
	Antimony	0.18	+/-1	J	2.00	P	12/03/2025	13:05	LB138105
	Arsenic	0.089	+/-0.5	U	1.00	P	12/03/2025	13:05	LB138105
	Barium	0.21	+/-5	U	10.0	P	12/03/2025	13:05	LB138105
	Beryllium	0.32	+/-0.5	U	1.00	P	12/03/2025	13:05	LB138105
	Cadmium	0.34	+/-0.5	U	1.00	P	12/03/2025	13:05	LB138105
	Calcium	45.7	+/-250	U	500	P	12/03/2025	13:05	LB138105
	Chromium	0.21	+/-1	U	2.00	P	12/03/2025	13:05	LB138105
	Cobalt	0.070	+/-0.5	U	1.00	P	12/03/2025	13:05	LB138105
	Copper	0.30	+/-1	U	2.00	P	12/03/2025	13:05	LB138105
	Iron	7.81	+/-25	U	50.0	P	12/03/2025	13:05	LB138105
	Lead	0.21	+/-0.5	U	1.00	P	12/03/2025	13:05	LB138105
	Magnesium	19.5	+/-250	U	500	P	12/03/2025	13:05	LB138105
	Manganese	0.43	+/-0.5	U	1.00	P	12/03/2025	13:05	LB138105
	Nickel	0.27	+/-0.5	U	1.00	P	12/03/2025	13:05	LB138105
	Potassium	36.4	+/-250	U	500	P	12/03/2025	13:05	LB138105
	Selenium	3.01	+/-2.5	J*	5.00	P	12/03/2025	13:05	LB138105
	Silver	0.060	+/-0.5	U	1.00	P	12/03/2025	13:05	LB138105
	Sodium	128	+/-250	U	500	P	12/03/2025	13:05	LB138105
	Thallium	0.060	+/-0.5	U	1.00	P	12/03/2025	13:05	LB138105
	Vanadium	0.077	+/-2.5	U	5.00	P	12/03/2025	13:05	LB138105
	Zinc	1.25	+/-2.5	U	5.00	P	12/03/2025	13:05	LB138105