

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Remington & Vernick

Contract: REMI01

Lab Code: ACE

SDG No.: Q3743

Instrument ID: P7

Start Date : 12/03/2025

Run Number: LB138105

End Date : 12/03/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
S0	S0	1005	100		100		100		100		100	
S2	S2	1008	99		98		98		99		100	
S3	S3	1012	101		99		98		98		100	
S4	S4	1015	95		95		95		95		97	
S5	S5	1018	90		90		91		90		92	
S6	S6	1021	88		85		86		84		88	
S7	S7	1023	81		77		78		77		81	
S8	S8	1026	93		89		89		84		93	
ICV01	ICV01	1059	101		98		97		97		98	
LLICV01	LLICV01	1110	105		100		99		100		100	
ICB01	ICB01	1113	101		97		97		99		98	
ICSA01	ICSA01	1118	101		96		95		92		98	
ICSAB01	ICSAB01	1121	102		99		99		95		101	
CCV01	CCV01	1124	87		86		88		84		90	
CCB01	CCB01	1127	103		101		102		103		103	
CRI	CRI	1142	100		98		99		99		100	
PB170780BL	PB170780BL	1145	101		99		100		100		101	
PB170780BS	PB170780BS	1149	91		87		88		87		90	
Q3743-01	TW-1125-1	1152	101		98		106		96		101	
Q3743-02	TW-1125-2	1155	102		98		101		97		100	
Q3743-02DUP	TW-1125-2DU	1158	102		100		101		98		101	
Q3743-02L	TW-1125-2L	1201	101		99		99		99		99	
Q3743-02MS	TW-1125-2MS	1204	96		97		100		95		97	
Q3743-02MSD	TW-1125-2MS	1207	98		99		103		97		100	
Q3743-02A	TW-1125-2A	1210	100		100		104		100		101	
CCV02	CCV02	1220	91		87		88		84		89	
CCB02	CCB02	1227	101		98		97		99		99	

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Remington & Vernick

Contract: REMI01

Lab Code: ACE

SDG No.: Q3743

Instrument ID: P7

Start Date : 12/03/2025

Run Number: LB138105

End Date : 12/03/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q	Element 165Ho	Q
S0	S0	1005	100		100		100		100		100	
S2	S2	1008	99		100		100		100		101	
S3	S3	1012	97		99		98		98		98	
S4	S4	1015	95		96		95		95		97	
S5	S5	1018	90		91		90		90		93	
S6	S6	1021	85		87		85		87		90	
S7	S7	1023	78		78		76		80		82	
S8	S8	1026	91		92		85		93		95	
ICV01	ICV01	1059	97		99		97		98		99	
LLICV01	LLICV01	1110	97		100		98		99		100	
ICB01	ICB01	1113	95		99		99		98		100	
ICSA01	ICSA01	1118	94		97		94		97		99	
ICSAB01	ICSAB01	1121	97		99		95		98		101	
CCV01	CCV01	1124	86		88		86		89		90	
CCB01	CCB01	1127	98		101		102		100		104	
CRI	CRI	1142	96		100		100		99		101	
PB170780BL	PB170780BL	1145	96		101		99		98		102	
PB170780BS	PB170780BS	1149	85		86		86		89		91	
Q3743-01	TW-1125-1	1152	97		107		95		98		100	
Q3743-02	TW-1125-2	1155	97		102		97		98		101	
Q3743-02DUP	TW-1125-2DU	1158	95		100		95		96		100	
Q3743-02L	TW-1125-2L	1201	97		101		100		100		102	
Q3743-02MS	TW-1125-2MS	1204	95		102		96		98		101	
Q3743-02MSD	TW-1125-2MS	1207	97		103		98		100		103	
Q3743-02A	TW-1125-2A	1210	98		105		99		99		103	
CCV02	CCV02	1220	85		86		83		88		92	
CCB02	CCB02	1227	95		99		98		98		101	